

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/06/2024 11/06/2024

Client Sample No.: CCAL03 Date Analyzed: 11/06/2024

Lab Sample No.: HSTDCCC750 Data File : PS028284.D Time Analyzed: 23:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.274	9.177	9.377	764.460	712.500	7.3
2,4-D	8.391	8.294	8.494	746.130	705.000	5.8
2,4-DCAA	7.260	7.162	7.362	788.410	750.000	5.1

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/06/2024 11/06/2024

Client Sample No.: CCAL03 Date Analyzed: 11/06/2024

Lab Sample No.: HSTDCCC750 Data File : PS028284.D Time Analyzed: 23:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.921	9.823	10.023	729.370	712.500	2.4
2,4-D	9.013	8.915	9.115	714.340	705.000	1.3
2,4-DCAA	7.763	7.665	7.865	773.710	750.000	3.2