

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM **Case No.:** P4578 **SAS No.:** P4578 **SDG NO.:** P4578

GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/06/2024 11/06/2024

Client Sample No.: CCAL04 **Date Analyzed:** 11/07/2024

Lab Sample No.: HSTDCCC750 **Data File :** PS028297.D **Time Analyzed:** 04:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.270	9.177	9.377	773.230	712.500	8.5
2,4-D	8.388	8.294	8.494	749.690	705.000	6.3
2,4-DCAA	7.258	7.162	7.362	772.740	750.000	3.0

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/06/2024 11/06/2024

Client Sample No.: CCAL04 Date Analyzed: 11/07/2024

Lab Sample No.: HSTDCCC750 Data File : PS028297.D Time Analyzed: 04:40

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.918	9.823	10.023	718.100	712.500	0.8
2,4-D	9.011	8.915	9.115	700.360	705.000	-0.7
2,4-DCAA	7.761	7.665	7.865	760.410	750.000	1.4