

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/06/2024 11/06/2024

Client Sample No.: CCAL05 Date Analyzed: 11/07/2024

Lab Sample No.: HSTDCCC750 Data File : PS028335.D Time Analyzed: 20:41

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.265	9.177	9.377	752.310	712.500	5.6
2,4-D	8.384	8.294	8.494	729.800	705.000	3.5
2,4-DCAA	7.254	7.162	7.362	762.350	750.000	1.6

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/06/2024 11/06/2024

Client Sample No.: CCAL05 Date Analyzed: 11/07/2024

Lab Sample No.: HSTDCCC750 Data File : PS028335.D Time Analyzed: 20:41

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.914	9.823	10.023	631.720	712.500	-11.3
2,4-D	9.008	8.915	9.115	614.020	705.000	-12.9
2,4-DCAA	7.758	7.665	7.865	672.470	750.000	-10.3