

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/06/2024 11/06/2024

Client Sample No.: CCAL06 Date Analyzed: 11/08/2024

Lab Sample No.: HSTDCCC750 Data File : PS028347.D Time Analyzed: 01:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.263	9.177	9.377	757.670	712.500	6.3
2,4-D	8.383	8.294	8.494	735.490	705.000	4.3
2,4-DCAA	7.252	7.162	7.362	768.470	750.000	2.5

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/06/2024 11/06/2024

Client Sample No.: CCAL06 Date Analyzed: 11/08/2024

Lab Sample No.: HSTDCCC750 Data File : PS028347.D Time Analyzed: 01:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.914	9.823	10.023	618.860	712.500	-13.1
2,4-D	9.008	8.915	9.115	603.860	705.000	-14.3
2,4-DCAA	7.758	7.665	7.865	659.600	750.000	-12.1