

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

Continuing Calib Date: 10/29/2024 Initial Calibration Date(s): 10/28/2024 10/28/2024

Continuing Calib Time: 16:53 Initial Calibration Time(s): 14:43 15:36

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.06	9.05	8.95	9.15	-0.01
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.92	4.92	4.82	5.02	0.00
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endrin	6.58	6.57	6.47	6.67	-0.01
Methoxychlor	7.50	7.50	7.40	7.60	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM **Case No.:** P4578 **SAS No.:** P4578 **SDG NO.:** P4578

Continuing Calib Date: 10/29/2024 **Initial Calibration Date(s):** 10/28/2024 10/28/2024

Continuing Calib Time: 16:53 **Initial Calibration Time(s):** 14:43 15:36

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.92	7.92	7.82	8.02	0.00
Tetrachloro-m-xylene	2.78	2.78	2.68	2.88	0.00
gamma-BHC (Lindane)	3.61	3.61	3.51	3.71	0.00
Heptachlor	3.95	3.95	3.85	4.05	0.00
Heptachlor epoxide	4.73	4.73	4.63	4.83	0.00
Endrin	5.64	5.64	5.54	5.74	0.00
Methoxychlor	6.62	6.62	6.52	6.72	0.01