

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL01 Date Analyzed: 10/29/2024

Lab Sample No.: PSTDCCC050 Data File : PL092695.D Time Analyzed: 10:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.053	8.954	9.154	49.010	50.000	-2.0
Endrin	6.575	6.474	6.674	47.250	50.000	-5.5
gamma-BHC (Lindane)	4.328	4.228	4.428	49.670	50.000	-0.7
Heptachlor	4.916	4.817	5.017	48.700	50.000	-2.6
Heptachlor epoxide	5.684	5.584	5.784	50.730	50.000	1.5
Methoxychlor	7.498	7.399	7.599	48.570	50.000	-2.9
Tetrachloro-m-xylene	3.540	3.440	3.640	49.870	50.000	-0.3

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL01 Date Analyzed: 10/29/2024

Lab Sample No.: PSTDCCC050 Data File : PL092695.D Time Analyzed: 10:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.915	7.816	8.016	51.500	50.000	3.0
Endrin	5.641	5.541	5.741	51.690	50.000	3.4
gamma-BHC (Lindane)	3.611	3.511	3.711	51.790	50.000	3.6
Heptachlor	3.950	3.849	4.049	50.790	50.000	1.6
Heptachlor epoxide	4.732	4.632	4.832	51.090	50.000	2.2
Methoxychlor	6.614	6.515	6.715	49.430	50.000	-1.1
Tetrachloro-m-xylene	2.779	2.678	2.878	50.880	50.000	1.8