

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL02 Date Analyzed: 10/29/2024

Lab Sample No.: PSTDCCC050 Data File : PL092711.D Time Analyzed: 14:07

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.054	8.954	9.154	50.330	50.000	0.7
Endrin	6.575	6.474	6.674	46.800	50.000	-6.4
gamma-BHC (Lindane)	4.329	4.228	4.428	51.390	50.000	2.8
Heptachlor	4.917	4.817	5.017	49.780	50.000	-0.4
Heptachlor epoxide	5.685	5.584	5.784	50.060	50.000	0.1
Methoxychlor	7.500	7.399	7.599	46.810	50.000	-6.4
Tetrachloro-m-xylene	3.541	3.440	3.640	51.450	50.000	2.9

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL02 Date Analyzed: 10/29/2024

Lab Sample No.: PSTDCCC050 Data File : PL092711.D Time Analyzed: 14:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.916	7.816	8.016	53.470	50.000	6.9
Endrin	5.641	5.541	5.741	52.150	50.000	4.3
gamma-BHC (Lindane)	3.610	3.511	3.711	53.770	50.000	7.5
Heptachlor	3.949	3.849	4.049	52.090	50.000	4.2
Heptachlor epoxide	4.732	4.632	4.832	53.380	50.000	6.8
Methoxychlor	6.614	6.515	6.715	49.000	50.000	-2.0
Tetrachloro-m-xylene	2.778	2.678	2.878	52.950	50.000	5.9