

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102924\
 Data File : PL092711.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Oct 2024 14:07
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 10/30/2024
 Supervised By :Ankita Jodhani 11/04/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 30 00:51:19 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102824.M
 Quant Title : GC Extractables
 QLast Update : Mon Oct 28 18:58:23 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.541	2.778	126.1E6	144.1E6	51.450	52.945
28) SA Decachlor...	9.054	7.916	96870332	145.9E6	50.334	53.468
Target Compounds						
2) A alpha-BHC	3.996	3.280	175.2E6	215.4E6	51.670	53.986
3) MA gamma-BHC...	4.329	3.610	167.5E6	207.5E6	51.385	53.771
4) MA Heptachlor	4.917	3.949	149.5E6	196.6E6	49.783	52.091
5) MB Aldrin	5.258	4.229	150.6E6	196.8E6	50.184m	53.760
6) B beta-BHC	4.527	3.910	73253824	86231673	50.679	52.381
7) B delta-BHC	4.773	4.139	160.4E6	206.3E6	51.247	53.679
8) B Heptachlo...	5.685	4.732	138.2E6	178.4E6	50.056	53.377
9) A Endosulfan I	6.070	5.101	125.1E6	161.2E6	50.005	52.800
10) B gamma-Chl...	5.941	4.982	134.0E6	178.7E6	50.347	53.144
11) B alpha-Chl...	6.020	5.045	133.0E6	176.3E6	50.234	52.998
12) B 4,4'-DDE	6.193	5.234	119.9E6	172.8E6	50.594	53.648
13) MA Dieldrin	6.345	5.366	131.8E6	178.8E6	50.014	53.536
14) MA Endrin	6.575	5.641	106.5E6	150.9E6	46.799	52.154
15) B Endosulfa...	6.795	5.934	117.4E6	152.0E6	49.234	53.648m
16) A 4,4'-DDD	6.711	5.789	102.2E6	141.1E6	53.243	56.657
17) MA 4,4'-DDT	7.023	6.040	97499557	131.2E6	47.107	48.922
18) B Endrin al...	6.925	6.115	95349863	120.6E6	50.773	52.264
19) B Endosulfa...	7.159	6.338	107.0E6	142.3E6	49.283	52.747
20) A Methoxychlor	7.500	6.614	53358109	69970481	46.805	48.998
21) B Endrin ke...	7.643	6.843	122.4E6	166.3E6	50.519	54.205
22) Mirex	8.117	7.022	96842515	134.4E6	48.868	51.526m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102924\
Data File : PL092711.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Oct 2024 14:07
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 10/30/2024

Supervised By :Ankita Jodhani 11/04/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 30 00:51:19 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102824.M
Quant Title : GC Extractables
QLast Update : Mon Oct 28 18:58:23 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

