

Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:		
Project:	Amtrak Sawtooth Bridges 2024		Date Received:		
Client Sample ID:	PB164445BL		SDG No.:	P4578	
Lab Sample ID:	PB164445BL		Matrix:	TCLP	
Analytical Method:	SW8081		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL092697.D	1	10/26/24 12:50	10/29/24 10:43	PB164445

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.00050	U	0.00050	0.0050	ug/L
76-44-8	Heptachlor	0.00050	U	0.00050	0.0050	ug/L
1024-57-3	Heptachlor epoxide	0.00090	U	0.00090	0.0050	ug/L
72-20-8	Endrin	0.00040	U	0.00040	0.0050	ug/L
72-43-5	Methoxychlor	0.0011	U	0.0011	0.0050	ug/L
8001-35-2	Toxaphene	0.015	U	0.015	0.10	ug/L
57-74-9	Chlordane	0.0082	U	0.0082	0.050	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	19.2		30 (43) - 150 (140)	96%	SPK: 20
877-09-8	Tetrachloro-m-xylene	18.3		30 (77) - 150 (126)	91%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit