

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4578

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

DataFile: BE101427.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB164533BS	Benzaldehyde	50	11.3	ug/L	23				20 (10)	160 (162)	
	Phenol	50	53.6	ug/L	107				20 (66)	160 (118)	
	bis(2-Chloroethyl)ether	50	44.7	ug/L	89				70 (62)	130 (103)	
	2-Chlorophenol	50	51.9	ug/L	104				70 (70)	130 (117)	
	2-Methylphenol	50	49.8	ug/L	100				70 (69)	130 (109)	
	2,2-oxybis(1-Chloropropane)	50	49.2	ug/L	98				70 (65)	130 (100)	
	Acetophenone	50	46.8	ug/L	94				70 (60)	130 (104)	
	3+4-Methylphenols	50	49.1	ug/L	98				20 (67)	160 (106)	
	N-Nitroso-di-n-propylamine	50	47.2	ug/L	94				70 (57)	130 (107)	
	Hexachloroethane	50	47.3	ug/L	95				20 (76)	160 (118)	
	Nitrobenzene	50	45.9	ug/L	92				70 (58)	130 (106)	
	Isophorone	50	46.2	ug/L	92				70 (61)	130 (102)	
	2-Nitrophenol	50	50.7	ug/L	101				70 (70)	130 (115)	
	2,4-Dimethylphenol	50	60.5	ug/L	121				70 (42)	130 (142)	
	bis(2-Chloroethoxy)methane	50	47.8	ug/L	96				70 (58)	130 (109)	
	2,4-Dichlorophenol	50	50.9	ug/L	102				70 (66)	130 (115)	
	Naphthalene	50	45.8	ug/L	92				70 (64)	130 (107)	
	4-Chloroaniline	50	33.1	ug/L	66		*		70 (10)	130 (85)	
	Hexachlorobutadiene	50	46.5	ug/L	93				70 (69)	130 (101)	
	Caprolactam	50	43.5	ug/L	87				20 (58)	160 (128)	
	4-Chloro-3-methylphenol	50	45.1	ug/L	90				70 (65)	130 (114)	
	2-Methylnaphthalene	50	45.9	ug/L	92				70 (64)	130 (107)	
	Hexachlorocyclopentadiene	100	200	ug/L	200		*		20 (36)	160 (160)	
	2,4,6-Trichlorophenol	50	52.0	ug/L	104				70 (61)	130 (110)	
	2,4,5-Trichlorophenol	50	49.8	ug/L	100				70 (70)	130 (106)	
	1,1-Biphenyl	50	47.9	ug/L	96				70 (72)	130 (98)	
	2-Chloronaphthalene	50	47.3	ug/L	95				70 (59)	130 (106)	
	2-Nitroaniline	50	51.8	ug/L	104				70 (73)	130 (114)	
	Dimethylphthalate	50	47.7	ug/L	95				70 (64)	130 (103)	
	Acenaphthylene	50	51.2	ug/L	102				70 (79)	130 (103)	
	2,6-Dinitrotoluene	50	46.1	ug/L	92				70 (64)	130 (110)	
	3-Nitroaniline	50	37.3	ug/L	75				70 (28)	130 (100)	
	Acenaphthene	50	48.4	ug/L	97				70 (59)	130 (113)	
	2,4-Dinitrophenol	100	82.1	ug/L	82				20 (36)	160 (166)	
	4-Nitrophenol	100	100	ug/L	100				20 (45)	160 (147)	
	Dibenzofuran	50	48.1	ug/L	96				70 (65)	130 (106)	
	2,4-Dinitrotoluene	50	48.6	ug/L	97				70 (60)	130 (115)	
	Diethylphthalate	50	47.4	ug/L	95				70 (63)	130 (105)	
	4-Chlorophenyl-phenylether	50	46.8	ug/L	94				70 (61)	130 (104)	
	Fluorene	50	47.2	ug/L	94				70 (64)	130 (107)	
	4-Nitroaniline	50	47.0	ug/L	94				70 (55)	130 (125)	
	4,6-Dinitro-2-methylphenol	50	48.1	ug/L	96				70 (62)	130 (132)	
	N-Nitrosodiphenylamine	50	49.3	ug/L	99				70 (61)	130 (109)	
	4-Bromophenyl-phenylether	50	47.3	ug/L	95				70 (73)	130 (103)	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4578

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

DataFile: BE101427.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB164533BS	Hexachlorobenzene	50	46.9	ug/L	94				70 (73)	130 (106)	
	Atrazine	50	61.2	ug/L	122				70 (76)	130 (120)	
	Pentachlorophenol	100	100	ug/L	100				20 (47)	160 (114)	
	Phenanthrene	50	49.7	ug/L	99				70 (62)	130 (109)	
	Anthracene	50	52.6	ug/L	105				70 (65)	130 (110)	
	Carbazole	50	48.6	ug/L	97				70 (62)	130 (106)	
	Di-n-butylphthalate	50	51.0	ug/L	102				70 (64)	130 (106)	
	Fluoranthene	50	48.9	ug/L	98				70 (64)	130 (110)	
	Pyrene	50	47.2	ug/L	94				70 (71)	130 (103)	
	Butylbenzylphthalate	50	49.1	ug/L	98				70 (61)	130 (105)	
	3,3-Dichlorobenzidine	50	40.2	ug/L	80				70 (43)	130 (108)	
	Benzo(a)anthracene	50	50.1	ug/L	100				70 (62)	130 (107)	
	Chrysene	50	49.7	ug/L	99				70 (61)	130 (108)	
	bis(2-Ethylhexyl)phthalate	50	52.7	ug/L	105				70 (59)	130 (110)	
	Di-n-octyl phthalate	50	55.7	ug/L	111				70 (52)	130 (139)	
	Benzo(b)fluoranthene	50	47.3	ug/L	95				70 (77)	130 (113)	
	Benzo(k)fluoranthene	50	51.0	ug/L	102				70 (77)	130 (105)	
	Benzo(a)pyrene	50	52.9	ug/L	106				70 (72)	130 (131)	
	Indeno(1,2,3-cd)pyrene	50	51.1	ug/L	102				70 (72)	130 (105)	
	Dibenz(a,h)anthracene	50	51.4	ug/L	103				70 (78)	130 (115)	
	Benzo(g,h,i)perylene	50	45.8	ug/L	92				70 (75)	130 (118)	
	1,2,4,5-Tetrachlorobenzene	50	49.1	ug/L	98				70 (72)	130 (101)	
	1,4-Dioxane	50	37.1	ug/L	74				20 (38)	160 (125)	
	2,3,4,6-Tetrachlorophenol	50	50.8	ug/L	102				70 (63)	130 (116)	