

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/31/2024

Lab File ID: BE101425.D Time Analyzed: 10:18

Instrument ID: BNA_E GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	136647	7.573	695845	10.34	498092	14.18
UPPER LIMIT	273294	8.073	1391690	10.84	996184	14.683
LOWER LIMIT	68323.5	7.073	347923	9.84	249046	13.683
EPA SAMPLE NO.						
01 PB164533BL	111412	7.57	479387	10.34	322455	14.18
02 PB164533BS	115147	7.57	538934	10.34	360623	14.18
03 WB-305-SW-1	150057	7.57	662069	10.34	470146	14.18
04 PB164533BSD	113034	7.57	526465	10.34	360052	14.18

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.