

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/29/2024

Lab File ID: BF140113.D Time Analyzed: 16:01

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	154336	6.892	582321	8.18	326390	9.93
UPPER LIMIT	308672	7.392	1164640	8.675	652780	10.428
LOWER LIMIT	77168	6.392	291161	7.675	163195	9.428
EPA SAMPLE NO.						
01 WB-305-TOP-1	121877	6.89	473789	8.17	240944	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.