

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/31/2024

Lab File ID: BF140176.D Time Analyzed: 09:22

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	122098	6.886	468668	8.17	269255	9.93
UPPER LIMIT	244196	7.386	937336	8.669	538510	10.433
LOWER LIMIT	61049	6.386	234334	7.669	134628	9.433
EPA SAMPLE NO.						
01 WB-305-SW	185491	6.89	676798	8.16	334897	9.93

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.