

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/29/2024

Lab File ID: BF140113.D Time Analyzed: 16:01

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	563401	11.416	296186	14.074	329400	15.592
UPPER LIMIT	1126800	11.916	592372	14.574	658800	16.092
LOWER LIMIT	281701	10.916	148093	13.574	164700	15.092
EPA SAMPLE NO.						
01 WB-305-TOP-1	345918	11.41	236214	14.05	244867	15.55

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.