

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/30/2024

Lab File ID: BF140152.D Time Analyzed: 12:50

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	475664	11.416	284421	14.068	273128	15.574
UPPER LIMIT	951328	11.916	568842	14.568	546256	16.074
LOWER LIMIT	237832	10.916	142211	13.568	136564	15.074
EPA SAMPLE NO.						
01 WB-305-TOP	363381	11.42	172346	14.07	220971	15.58

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.