

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/01/2024

Lab File ID: BF140200.D Time Analyzed: 09:22

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	606919	11.422	388668	14.08	308087	15.58
UPPER LIMIT	1213840	11.922	777336	14.58	616174	16.08
LOWER LIMIT	303460	10.922	194334	13.58	154044	15.08
EPA SAMPLE NO.						
01 WB-305-SWRE	893182	11.42	535712	14.07	445221	15.57
02 WB-305-SW-1RE	750159	11.42	460215	14.07	381772	15.58

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.