



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Cover Page

Order ID : P4578

Project ID : Amtrak Sawtooth Bridges 2024

Client : Portal Partners Tri-Venture

Lab Sample Number

P4578-01
P4578-02
P4578-03
P4578-04
P4578-05
P4578-06
P4578-07
P4578-08
P4578-09

Client Sample Number

WB-305-SW
WB-305-SW-1
WB-305-TOP
WB-305-TOP-1
WB-305-BOT
WB-305-BOT
WB-305-BOT-1
WB-305-BOT-1
TB-10252024

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 10/30/2024

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Portal Partners Tri-Venture

Project Name: Amtrak Sawtooth Bridges 2024

Project # N/A

Chemtech Project # P4578

Test Name: TCLP BNA

A. Number of Samples and Date of Receipt:

6 Solid samples were received on 10/25/2024.

3 Water samples were received on 10/25/2024.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Corrosivity, EPH, EPH, Hexavalent Chromium, Ignitability, Mercury, Metals ICP-TAL, METALS-TAL, PCB, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SVOC-TCL BNA -20, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL, Trivalent Chromium, VOC-TCLVOA-10 and VOC-TCLVOA-10. This data package contains results for TCLP BNA.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um df The analysis of TCLP BNA was based on method 8270E and extraction was done based on method 3510 and TCLP extraction method was 1311.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS recoveries met the requirements for all compounds .

The MSD recoveries met the acceptable requirements .

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

E. Additional Comments:

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature _____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following "Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: P4578

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication, airbill, sample management lab chronicle, login page) ✓

Check chain-of-custody for proper relinquish/return of samples ✓

Is the chain of custody signed and complete ✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts ✓

Collect information for each project id from server. Were all requirements followed ✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page ✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody ✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results ✓

Do requested analyses on Chain of Custody agree with the log-in page ✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody ✓

Were the samples received within hold time ✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle ✓

ANALYTICAL:

Was method requirement followed? ✓

Was client requirement followed? ✓

Does the case narrative summarize all QC failure? ✓

All runlogs and manual integration are reviewed for requirements ✓

All manual calculations and /or hand notations verified ✓

LAB CHRONICLE

OrderID:	P4578	OrderDate:	10/25/2024 3:58:00 PM					
Client:	Portal Partners Tri-Venture	Project:	Amtrak Sawtooth Bridges 2024					
Contact:	Joseph Krupansky	Location:	K63,VOA Ref. #2 Soil,VOA Ref. #3 Water					
<hr/>								
LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P4578-06	WB-305-BOT	TCLP			10/25/24			10/25/24
			TCLP BNA	8270E		10/26/24	10/28/24	
P4578-08	WB-305-BOT-1	TCLP			10/25/24			10/25/24
			TCLP BNA	8270E		10/26/24	10/28/24	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: P4578

Client: Portal Partners Tri-Venture

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :				0.000				
			Total Svoc :		0.00			
			Total Concentration:		0.00			



QC

SUMMARY

Surrogate Summary

SW-846

SDG No.: P4578

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P4545-02MS	V NJ-215MS	2-Fluorophenol	150	135	90		15 (10)	110 (139)
		Phenol-d6	150	129	86		15 (10)	110 (134)
		Nitrobenzene-d5	100	103	103		30 (49)	130 (133)
		2-Fluorobiphenyl	100	98.8	99		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	159	106		15 (44)	110 (137)
		Terphenyl-d14	100	105	105		30 (48)	130 (125)
P4545-02MSD	V NJ-215MSD	2-Fluorophenol	150	136	91		15 (10)	110 (139)
		Phenol-d6	150	132	88		15 (10)	110 (134)
		Nitrobenzene-d5	100	105	105		30 (49)	130 (133)
		2-Fluorobiphenyl	100	101	101		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	162	108		15 (44)	110 (137)
		Terphenyl-d14	100	105	105		30 (48)	130 (125)
P4578-06	WB-305-BOT	2-Fluorophenol	150	130	87		15 (10)	110 (139)
		Phenol-d6	150	120	80		15 (10)	110 (134)
		Nitrobenzene-d5	100	105	105		30 (49)	130 (133)
		2-Fluorobiphenyl	100	101	101		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	159	106		15 (44)	110 (137)
		Terphenyl-d14	100	111	111		30 (48)	130 (125)
P4578-08	WB-305-BOT-1	2-Fluorophenol	150	121	81		15 (10)	110 (139)
		Phenol-d6	150	112	75		15 (10)	110 (134)
		Nitrobenzene-d5	100	96.4	96		30 (49)	130 (133)
		2-Fluorobiphenyl	100	93.8	94		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	147	98		15 (44)	110 (137)
		Terphenyl-d14	100	102	102		30 (48)	130 (125)
PB164393TB	PB164393TB	2-Fluorophenol	150	122	82		15 (10)	110 (139)
		Phenol-d6	150	119	79		15 (10)	110 (134)
		Nitrobenzene-d5	100	92.2	92		30 (49)	130 (133)
		2-Fluorobiphenyl	100	87.9	88		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	137	91		15 (44)	110 (137)
		Terphenyl-d14	100	88.8	89		30 (48)	130 (125)
PB164444BL	PB164444BL	2-Fluorophenol	150	120	80		15 (10)	110 (139)
		Phenol-d6	150	114	76		15 (10)	110 (134)
		Nitrobenzene-d5	100	90.9	91		30 (49)	130 (133)
		2-Fluorobiphenyl	100	87.1	87		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	133	89		15 (44)	110 (137)
		Terphenyl-d14	100	89.1	89		30 (48)	130 (125)
PB164444BS	PB164444BS	2-Fluorophenol	150	120	80		15 (10)	110 (139)
		Phenol-d6	150	118	79		15 (10)	110 (134)
		Nitrobenzene-d5	100	92.6	93		30 (49)	130 (133)
		2-Fluorobiphenyl	100	87.2	87		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	141	94		15 (44)	110 (137)
		Terphenyl-d14	100	95.7	96		30 (48)	130 (125)

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4578

Client: Portal Partners Tri-Venture

Analytical Method: SW8270E

Parameter	Spike	Sample			Rec	RPD	RPD	Limits		
		Result	Result	Units				Qual	Low	High
Lab Sample ID:	P4545-02MS		Client Sample ID:	VNJ-215MS				DataFile:	BF140092.D	
Pyridine	500	0	230	ug/L	46				20 (10)	160 (109)
1,4-Dichlorobenzene	500	0	430	ug/L	86				70 (55)	130 (125)
2-Methylphenol	500	0	490	ug/L	98				70 (37)	130 (126)
3+4-Methylphenols	500	0	480	ug/L	96				20 (31)	160 (127)
Hexachloroethane	500	0	420	ug/L	84				20 (49)	160 (110)
Nitrobenzene	500	0	470	ug/L	94				70 (62)	130 (112)
Hexachlorobutadiene	500	0	460	ug/L	92				70 (52)	130 (125)
2,4,6-Trichlorophenol	500	0	550	ug/L	110				70 (78)	130 (112)
2,4,5-Trichlorophenol	500	0	540	ug/L	108				70 (71)	130 (111)
2,4-Dinitrotoluene	500	0	600	ug/L	120				70 (50)	130 (142)
Hexachlorobenzene	500	0	540	ug/L	108				70 (72)	130 (115)
Pentachlorophenol	1000	0	780	ug/L	78				20 (25)	160 (139)

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4578

Client: Portal Partners Tri-Venture

Analytical Method: SW8270E

Parameter	Spike	Sample			Rec	RPD	RPD	Limits		
		Result	Result	Units				Low	High	RPD
Lab Sample ID:	P4545-02MSD	Client Sample ID:	VNJ-215MSD				DataFile:	BF140093.D		
Pyridine	500	0	250	ug/L	50	8		20 (10)	160 (109)	20 (20)
1,4-Dichlorobenzene	500	0	440	ug/L	88	2		70 (55)	130 (125)	20 (20)
2-Methylphenol	500	0	500	ug/L	100	2		70 (37)	130 (126)	20 (20)
3+4-Methylphenols	500	0	500	ug/L	100	4		20 (31)	160 (127)	20 (20)
Hexachloroethane	500	0	430	ug/L	86	2		20 (49)	160 (110)	20 (20)
Nitrobenzene	500	0	470	ug/L	94	0		70 (62)	130 (112)	20 (20)
Hexachlorobutadiene	500	0	460	ug/L	92	0		70 (52)	130 (125)	20 (20)
2,4,6-Trichlorophenol	500	0	560	ug/L	112	2		70 (78)	130 (112)	20 (20)
2,4,5-Trichlorophenol	500	0	540	ug/L	108	0		70 (71)	130 (111)	20 (20)
2,4-Dinitrotoluene	500	0	610	ug/L	122	2		70 (50)	130 (142)	20 (20)
Hexachlorobenzene	500	0	560	ug/L	112	4		70 (72)	130 (115)	20 (20)
Pentachlorophenol	1000	0	850	ug/L	85	9		20 (25)	160 (139)	20 (20)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4578

Client: Portal Partners Tri-Venture

Analytical Method: 8270E DataFile: BF140077.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Qual	Limits		
									Low	High	
PB164444BS	Pyridine	50	38.6	ug/L	77				20 (29)	160 (97)	
	1,4-Dichlorobenzene	50	43.7	ug/L	87				70 (76)	130 (103)	
	2-Methylphenol	50	44.4	ug/L	89				70 (69)	130 (109)	
	3+4-Methylphenols	50	43.0	ug/L	86				20 (67)	160 (106)	
	Hexachloroethane	50	44.2	ug/L	88				20 (76)	160 (118)	
	Nitrobenzene	50	43.5	ug/L	87				70 (58)	130 (106)	
	Hexachlorobutadiene	50	44.9	ug/L	90				70 (69)	130 (101)	
	2,4,6-Trichlorophenol	50	45.9	ug/L	92				70 (61)	130 (110)	
	2,4,5-Trichlorophenol	50	44.6	ug/L	89				70 (70)	130 (106)	
	2,4-Dinitrotoluene	50	52.4	ug/L	105				70 (60)	130 (115)	
	Hexachlorobenzene	50	45.0	ug/L	90				70 (73)	130 (106)	
	Pentachlorophenol	100	89.7	ug/L	90				20 (47)	160 (114)	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB164444BL

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM Case No.: P4578

SAS No.: P4578 SDG NO.: P4578

Lab File ID: BF140076.D

Lab Sample ID: PB164444BL

Instrument ID: BNA_F

Date Extracted: 10/26/2024

Matrix: (soil/water) water

Date Analyzed: 10/28/2024

Level: (low/med) LOW

Time Analyzed: 10:01

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB164393TB	PB164393TB	BF140078.D	10/28/2024
WB-305-BOT	P4578-06	BF140084.D	10/28/2024
WB-305-BOT-1	P4578-08	BF140085.D	10/28/2024
PB164444BS	PB164444BS	BF140077.D	10/28/2024
VNJ-215MS	P4545-02MS	BF140092.D	10/28/2024
VNJ-215MSD	P4545-02MSD	BF140093.D	10/28/2024

COMMENTS:



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM

SAS No.: P4578 SDG NO.: P4578

Lab File ID: BF139843.D

DFTPP Injection Date: 10/18/2024

Instrument ID: BNA_F

DFTPP Injection Time: 09:22

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	43.8
68	Less than 2.0% of mass 69	0.7 (1.9) 1
69	Mass 69 relative abundance	37.8
70	Less than 2.0% of mass 69	0.3 (0.7) 1
127	10.0 - 80.0% of mass 198	46.9
197	Less than 2.0% of mass 198	0.8
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	26.9
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	14
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	17.9 (19.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF139844.D	10/18/2024	10:27
SSTDICC005	SSTDICC005	BF139845.D	10/18/2024	10:55
SSTDICC010	SSTDICC010	BF139846.D	10/18/2024	11:23
SSTDICC020	SSTDICC020	BF139847.D	10/18/2024	11:52
SSTDICCC040	SSTDICCC040	BF139848.D	10/18/2024	12:20
SSTDICC050	SSTDICC050	BF139849.D	10/18/2024	12:49
SSTDICC060	SSTDICC060	BF139850.D	10/18/2024	13:17
SSTDICC080	SSTDICC080	BF139851.D	10/18/2024	13:46



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM

SAS No.: P4578 SDG NO.: P4578

Lab File ID: BF140074.D

DFTPP Injection Date: 10/28/2024

Instrument ID: BNA_F

DFTPP Injection Time: 09:04

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	41.1
68	Less than 2.0% of mass 69	0.6 (1.6) 1
69	Mass 69 relative abundance	37.2
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	46.3
197	Less than 2.0% of mass 198	0.8
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.5
275	10.0 - 60.0% of mass 198	27.6
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	15.5
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.8 (18.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF140075.D	10/28/2024	09:32
PB164444BL	PB164444BL	BF140076.D	10/28/2024	10:01
PB164444BS	PB164444BS	BF140077.D	10/28/2024	10:29
PB164393TB	PB164393TB	BF140078.D	10/28/2024	10:58
WB-305-BOT	P4578-06	BF140084.D	10/28/2024	13:51
WB-305-BOT-1	P4578-08	BF140085.D	10/28/2024	14:19
VNJ-215MS	P4545-02MS	BF140092.D	10/28/2024	17:35
VNJ-215MSD	P4545-02MSD	BF140093.D	10/28/2024	18:03



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH
Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG No.: P4578
EPA Sample No.: SSTDCCC040 Date Analyzed: 10/28/2024
Lab File ID: BF140075.D Time Analyzed: 09:32
Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	132517	6.892	497360	8.18	274065	9.93
UPPER LIMIT	265034	7.392	994720	8.675	548130	10.427
LOWER LIMIT	66258.5	6.392	248680	7.675	137033	9.427
EPA SAMPLE NO.						
01 PB164444BL	132330	6.89	518019	8.17	298114	9.92
02 PB164444BS	127868	6.89	495282	8.18	281323	9.93
03 PB164393TB	128211	6.89	508163	8.17	288600	9.92
04 VNJ-215MS	113612	6.89	441889	8.17	241647	9.93
05 VNJ-215MSD	111820	6.89	432287	8.17	236921	9.93
06 WB-305-BOT	112580	6.89	438516	8.17	246067	9.92
07 WB-305-BOT-1	111257	6.89	441878	8.17	245603	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name :	CHEMTECH						
Lab Code :	CHEM	Case No. :	P4578	SAS No. :	P4578	SDG NO. :	P4578
EPA Sample No. :	SSTDCCC040		Date Analyzed:	10/28/2024			
Lab File ID:	BF140075.D		Time Analyzed:	09:32			
Instrument ID:	BNA_F		GC Column:	DB-U1	ID:	0.18	(mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	494703	11.416	250901	14.051	263501	15.527
	989406	11.916	501802	14.551	527002	16.027
	247352	10.916	125451	13.551	131751	15.027
EPA SAMPLE NO.						
01 PB164444BL	550598	11.41	341644	14.05	261224	15.53
02 PB164444BS	511374	11.42	262593	14.05	265684	15.53
03 PB164393TB	536321	11.41	334195	14.05	261975	15.53
04 VNJ-215MS	409717	11.41	203129	14.05	263019	15.53
05 VNJ-215MSD	398949	11.41	204155	14.05	256308	15.53
06 WB-305-BOT	437262	11.41	219498	14.05	258448	15.53
07 WB-305-BOT-1	437798	11.41	215751	14.05	230540	15.53

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



SAMPLE

DATA



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	10/25/24	
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	10/25/24	
Client Sample ID:	WB-305-BOT			SDG No.:	P4578	
Lab Sample ID:	P4578-06			Matrix:	TCLP	
Analytical Method:	SW8270			% Solid:	0	
Sample Wt/Vol:	100	Units:	mL	Final Vol:	1000	uL
Soil Aliquot Vol:	uL			Test:	TCLP BNA	
Extraction Type :	Decanted : N			Level :	LOW	
Injection Volume :	GPC Factor : 1.0			GPC Cleanup :	N	PH :
Prep Method :	SW3541					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140084.D	1	10/26/24 12:10	10/28/24 13:51	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	15.5	U	15.5	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	8.40	U	8.40	50.0	ug/L
95-48-7	2-Methylphenol	11.3	U	11.3	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.5	U	11.5	100	ug/L
67-72-1	Hexachloroethane	10.1	U	10.1	50.0	ug/L
98-95-3	Nitrobenzene	12.7	U	12.7	50.0	ug/L
87-68-3	Hexachlorobutadiene	12.7	U	12.7	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	8.90	U	8.90	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	10.1	U	10.1	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	15.2	U	15.2	50.0	ug/L
118-74-1	Hexachlorobenzene	11.4	U	11.4	50.0	ug/L
87-86-5	Pentachlorophenol	18.5	U	18.5	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	130		15 (10) - 110 (139)	87%	SPK: 150
13127-88-3	Phenol-d6	120		15 (10) - 110 (134)	80%	SPK: 150
4165-60-0	Nitrobenzene-d5	105		30 (49) - 130 (133)	105%	SPK: 100
321-60-8	2-Fluorobiphenyl	101		30 (52) - 130 (132)	101%	SPK: 100
118-79-6	2,4,6-Tribromophenol	159		15 (44) - 110 (137)	106%	SPK: 150
1718-51-0	Terphenyl-d14	111		30 (48) - 130 (125)	111%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	113000	6.887			
1146-65-2	Naphthalene-d8	439000	8.169			
15067-26-2	Acenaphthene-d10	246000	9.922			
1517-22-2	Phenanthrene-d10	437000	11.41			
1719-03-5	Chrysene-d12	219000	14.045			
1520-96-3	Perylene-d12	258000	15.527			



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	10/25/24	
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	10/25/24	
Client Sample ID:	WB-305-BOT			SDG No.:	P4578	
Lab Sample ID:	P4578-06			Matrix:	TCLP	
Analytical Method:	SW8270			% Solid:	0	
Sample Wt/Vol:	100	Units:	mL	Final Vol:	1000	uL
Soil Aliquot Vol:	uL			Test:	TCLP BNA	
Extraction Type :				Decanted :	N	Level :
Injection Volume :				GPC Factor :	1.0	GPC Cleanup : N PH :
Prep Method :	SW3541					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140084.D	1	10/26/24 12:10	10/28/24 13:51	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140084.D
 Acq On : 28 Oct 2024 13:51
 Operator : RC/JU
 Sample : P4578-06
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WB-305-BOT

Quant Time: Oct 28 14:22:20 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.887	152	112580	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	438516	20.000	ng	0.00
39) Acenaphthene-d10	9.922	164	246067	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	437262	20.000	ng	0.00
76) Chrysene-d12	14.045	240	219498	20.000	ng	0.00
86) Perylene-d12	15.527	264	258448	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	938458	130.498	ng	0.02
7) Phenol-d6	6.510	99	1115782	119.796	ng	0.00
23) Nitrobenzene-d5	7.445	82	834614	105.486	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	367109	159.499	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1500601	100.787	ng	0.00
79) Terphenyl-d14	12.998	244	1489277	110.559	ng	0.00




Target Compounds	Qvalue
(#= qualifier out of range (m) = manual integration (+) = signals summed	

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140084.D
 Acq On : 28 Oct 2024 13:51
 Operator : RC/JU
 Sample : P4578-06
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WB-305-BOT

Quant Time: Oct 28 14:22:20 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.887 min Scan# 8
Delta R.T. -0.005 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51


Instrument :
BNA_F
ClientSampleId :
WB-305-BOT


Tgt Ion:152 Resp: 112580
Ion Ratio Lower Upper
152 100
150 159.0 130.2 195.2
115 65.7 51.4 77.2



#5
2-Fluorophenol
Concen: 130.498 ng
RT: 5.516 min Scan# 582
Delta R.T. 0.018 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51



Tgt Ion:112 Resp: 938458
Ion Ratio Lower Upper
112 100
64 65.4 53.0 79.6
63 33.9 27.0 40.4






#7
Phenol-d6
Concen: 119.796 ng
RT: 6.510 min Scan# 7
Instrument: BNA_F
Delta R.T. 0.000 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51
ClientSampleId : WB-305-BOT


Tgt Ion: 99 Resp: 1115782
Ion Ratio Lower Upper
99 100
42 19.6 16.7 25.1
71 35.6 27.7 41.5




#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.169 min Scan# 1033
Delta R.T. -0.006 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51

Tgt Ion:136 Resp: 438516
Ion Ratio Lower Upper
136 100
137 10.9 8.6 12.8
54 9.6 8.4 12.6
68 6.0 5.1 7.7









#23
Nitrobenzene-d5
Concen: 105.486 ng
RT: 7.445 min Scan# 911
Instrument: BNA_F
Delta R.T. -0.006 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51
ClientSampleId : WB-305-BOT





Tgt Ion: 82 Resp: 834614
Ion Ratio Lower Upper
82 100
128 41.9 33.4 50.0
54 56.8 47.8 71.8




#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.922 min Scan# 1331
Delta R.T. -0.005 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51





Tgt Ion:164 Resp: 246067
Ion Ratio Lower Upper
164 100
162 101.7 81.0 121.4
160 44.9 35.4 53.0






#42
2,4,6-Tribromophenol
Concen: 159.499 ng
RT: 10.710 min Scan# 14
Instrument: BNA_F
Delta R.T. -0.006 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51
ClientSampleId : WB-305-BOT


Tgt Ion:330 Resp: 367109
Ion Ratio Lower Upper
330 100
332 96.4 78.1 117.1
141 29.4 26.6 39.8



#45
2-Fluorobiphenyl
Concen: 100.787 ng
RT: 9.245 min Scan# 1216
Delta R.T. -0.006 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51

Tgt Ion:172 Resp: 1500601
Ion Ratio Lower Upper
172 100
171 35.9 28.6 43.0
170 23.9 19.1 28.7







#64
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 11.410 min Scan# 1
 Delta R.T. -0.006 min
 Lab File: BF140084.D
 Acq: 28 Oct 2024 13:51


Instrument : BNA_F
 ClientSampleId : WB-305-BOT


Tgt Ion:188 Resp: 437262
 Ion Ratio Lower Upper
 188 100
 94 9.5 7.9 11.9
 80 10.3 9.0 13.4



#76
 Chrysene-d12
 Concen: 20.000 ng
 RT: 14.045 min Scan# 2032
 Delta R.T. -0.006 min
 Lab File: BF140084.D
 Acq: 28 Oct 2024 13:51



Tgt Ion:240 Resp: 219498
 Ion Ratio Lower Upper
 240 100
 120 11.6 9.4 14.2
 236 25.4 20.9 31.3






#79
Terphenyl-d14
Concen: 110.559 ng
RT: 12.998 min Scan# 1
Instrument : BNA_F
Delta R.T. 0.000 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51
ClientSampleId : WB-305-BOT

Tgt Ion:244 Resp: 1489277
Ion Ratio Lower Upper
244 100
212 6.6 5.7 8.5
122 9.8 8.6 13.0



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.527 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51

Tgt Ion:264 Resp: 258448
Ion Ratio Lower Upper
264 100
260 24.0 19.4 29.2
265 21.8 17.4 26.0





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	10/25/24	
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	10/25/24	
Client Sample ID:	WB-305-BOT-1			SDG No.:	P4578	
Lab Sample ID:	P4578-08			Matrix:	TCLP	
Analytical Method:	SW8270			% Solid:	0	
Sample Wt/Vol:	100	Units:	mL	Final Vol:	1000	uL
Soil Aliquot Vol:	uL			Test:	TCLP BNA	
Extraction Type :				Decanted :	N	Level :
Injection Volume :				GPC Factor :	1.0	GPC Cleanup : N PH :
Prep Method :	SW3541					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140085.D	1	10/26/24 12:10	10/28/24 14:19	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	15.5	U	15.5	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	8.40	U	8.40	50.0	ug/L
95-48-7	2-Methylphenol	11.3	U	11.3	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.5	U	11.5	100	ug/L
67-72-1	Hexachloroethane	10.1	U	10.1	50.0	ug/L
98-95-3	Nitrobenzene	12.7	U	12.7	50.0	ug/L
87-68-3	Hexachlorobutadiene	12.7	U	12.7	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	8.90	U	8.90	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	10.1	U	10.1	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	15.2	U	15.2	50.0	ug/L
118-74-1	Hexachlorobenzene	11.4	U	11.4	50.0	ug/L
87-86-5	Pentachlorophenol	18.5	U	18.5	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	121		15 (10) - 110 (139)	81%	SPK: 150
13127-88-3	Phenol-d6	112		15 (10) - 110 (134)	75%	SPK: 150
4165-60-0	Nitrobenzene-d5	96.4		30 (49) - 130 (133)	96%	SPK: 100
321-60-8	2-Fluorobiphenyl	93.8		30 (52) - 130 (132)	94%	SPK: 100
118-79-6	2,4,6-Tribromophenol	147		15 (44) - 110 (137)	98%	SPK: 150
1718-51-0	Terphenyl-d14	102		30 (48) - 130 (125)	102%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	111000	6.887			
1146-65-2	Naphthalene-d8	442000	8.169			
15067-26-2	Acenaphthene-d10	246000	9.922			
1517-22-2	Phenanthrene-d10	438000	11.41			
1719-03-5	Chrysene-d12	216000	14.045			
1520-96-3	Perylene-d12	231000	15.527			



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	10/25/24	
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	10/25/24	
Client Sample ID:	WB-305-BOT-1			SDG No.:	P4578	
Lab Sample ID:	P4578-08			Matrix:	TCLP	
Analytical Method:	SW8270			% Solid:	0	
Sample Wt/Vol:	100	Units:	mL	Final Vol:	1000	uL
Soil Aliquot Vol:	uL			Test:	TCLP BNA	
Extraction Type :				Decanted :	N	Level :
Injection Volume :				GPC Factor :	1.0	GPC Cleanup : N PH :
Prep Method :	SW3541					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140085.D	1	10/26/24 12:10	10/28/24 14:19	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140085.D
 Acq On : 28 Oct 2024 14:19
 Operator : RC/JU
 Sample : P4578-08
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WB-305-BOT-1

Quant Time: Oct 28 14:45:06 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.887	152	111257	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	441878	20.000	ng	0.00
39) Acenaphthene-d10	9.922	164	245603	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	437798	20.000	ng	0.00
76) Chrysene-d12	14.045	240	215751	20.000	ng	0.00
86) Perylene-d12	15.527	264	230540	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	863113	121.448	ng	0.02
7) Phenol-d6	6.510	99	1033565	112.289	ng	0.00
23) Nitrobenzene-d5	7.445	82	768626	96.407	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	338552	147.370	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1393721	93.785	ng	0.00
79) Terphenyl-d14	12.998	244	1353675	102.238	ng	0.00




Target Compounds	Qvalue
(#= qualifier out of range (m) = manual integration (+) = signals summed	

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140085.D
 Acq On : 28 Oct 2024 14:19
 Operator : RC/JU
 Sample : P4578-08
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WB-305-BOT-1

Quant Time: Oct 28 14:45:06 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.887 min Scan# 8
Delta R.T. -0.005 min
Lab File: BF140085.D
Acq: 28 Oct 2024 14:19


Instrument :
BNA_F
ClientSampleId :
WB-305-BOT-1


Tgt Ion:152 Resp: 111257
Ion Ratio Lower Upper
152 100
150 157.9 130.2 195.2
115 64.3 51.4 77.2



#5
2-Fluorophenol
Concen: 121.448 ng
RT: 5.516 min Scan# 582
Delta R.T. 0.018 min
Lab File: BF140085.D
Acq: 28 Oct 2024 14:19

Tgt Ion:112 Resp: 863113
Ion Ratio Lower Upper
112 100
64 64.4 53.0 79.6
63 33.1 27.0 40.4







#7
 Phenol-d6
 Concen: 112.289 ng
 RT: 6.510 min Scan# 7
 Delta R.T. 0.000 min
 Lab File: BF140085.D
 Acq: 28 Oct 2024 14:19


Instrument : BNA_F
 ClientSampleId : WB-305-BOT-1


Tgt Ion: 99 Resp: 1033565
 Ion Ratio Lower Upper
 99 100
 42 19.3 16.7 25.1
 71 34.9 27.7 41.5




#21
 Naphthalene-d8
 Concen: 20.000 ng
 RT: 8.169 min Scan# 1033
 Delta R.T. -0.006 min
 Lab File: BF140085.D
 Acq: 28 Oct 2024 14:19

Tgt Ion:136 Resp: 441878
 Ion Ratio Lower Upper
 136 100
 137 10.9 8.6 12.8
 54 9.5 8.4 12.6
 68 6.3 5.1 7.7









#23
Nitrobenzene-d5
Concen: 96.407 ng
RT: 7.445 min Scan# 911
Instrument: BNA_F
Delta R.T. -0.006 min
Lab File: BF140085.D
ClientSampleId : WB-305-BOT-1
Acq: 28 Oct 2024 14:19





Tgt Ion: 82 Resp: 768626
Ion Ratio Lower Upper
82 100
128 42.3 33.4 50.0
54 56.8 47.8 71.8




#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.922 min Scan# 1331
Delta R.T. -0.005 min
Lab File: BF140085.D
Acq: 28 Oct 2024 14:19



Tgt Ion:164 Resp: 245603
Ion Ratio Lower Upper
164 100
162 100.9 81.0 121.4
160 44.5 35.4 53.0







#42
2,4,6-Tribromophenol
Concen: 147.370 ng
RT: 10.710 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF140085.D
Acq: 28 Oct 2024 14:19


Instrument : BNA_F
ClientSampleId : WB-305-BOT-1


Tgt Ion:330 Resp: 338552
Ion Ratio Lower Upper
330 100
332 94.9 78.1 117.1
141 29.4 26.6 39.8



#45
2-Fluorobiphenyl
Concen: 93.785 ng
RT: 9.245 min Scan# 1216
Delta R.T. -0.006 min
Lab File: BF140085.D
Acq: 28 Oct 2024 14:19

Tgt Ion:172 Resp: 1393721
Ion Ratio Lower Upper
172 100
171 35.6 28.6 43.0
170 23.7 19.1 28.7







#64
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 11.410 min Scan# 1
 Delta R.T. -0.006 min
 Lab File: BF140085.D
 Acq: 28 Oct 2024 14:19


Instrument : BNA_F
 ClientSampleId : WB-305-BOT-1


Tgt Ion:188 Resp: 437798
 Ion Ratio Lower Upper
 188 100
 94 9.8 7.9 11.9
 80 10.4 9.0 13.4



#76
 Chrysene-d12
 Concen: 20.000 ng
 RT: 14.045 min Scan# 2032
 Delta R.T. -0.006 min
 Lab File: BF140085.D
 Acq: 28 Oct 2024 14:19



Tgt Ion:240 Resp: 215751
 Ion Ratio Lower Upper
 240 100
 120 11.4 9.4 14.2
 236 25.6 20.9 31.3






#79
Terphenyl-d14
Concen: 102.238 ng
RT: 12.998 min Scan# 1
Instrument: BNA_F
Delta R.T. 0.000 min
Lab File: BF140085.D
ClientSampleId : WB-305-BOT-1
Acq: 28 Oct 2024 14:19

Tgt Ion:244 Resp: 1353675
Ion Ratio Lower Upper
244 100
212 6.8 5.7 8.5
122 9.8 8.6 13.0



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.527 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BF140085.D
Acq: 28 Oct 2024 14:19

Tgt Ion:264 Resp: 230540
Ion Ratio Lower Upper
264 100
260 24.3 19.4 29.2
265 21.5 17.4 26.0





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	10/26/24	
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	10/26/24	
Client Sample ID:	PB164393TB			SDG No.:	P4578	
Lab Sample ID:	PB164393TB			Matrix:	TCLP	
Analytical Method:	SW8270			% Solid:	0	
Sample Wt/Vol:	100	Units:	mL	Final Vol:	1000	uL
Soil Aliquot Vol:	uL			Test:	TCLP BNA	
Extraction Type :				Decanted :	N	Level :
Injection Volume :				GPC Factor :	1.0	GPC Cleanup : N PH :
Prep Method :	SW3541					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140078.D	1	10/26/24 12:10	10/28/24 10:58	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	15.5	U	15.5	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	8.40	U	8.40	50.0	ug/L
95-48-7	2-Methylphenol	11.3	U	11.3	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.5	U	11.5	100	ug/L
67-72-1	Hexachloroethane	10.1	U	10.1	50.0	ug/L
98-95-3	Nitrobenzene	12.7	U	12.7	50.0	ug/L
87-68-3	Hexachlorobutadiene	12.7	U	12.7	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	8.90	U	8.90	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	10.1	U	10.1	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	15.2	U	15.2	50.0	ug/L
118-74-1	Hexachlorobenzene	11.4	U	11.4	50.0	ug/L
87-86-5	Pentachlorophenol	18.5	U	18.5	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	122		15 (10) - 110 (139)	82%	SPK: 150
13127-88-3	Phenol-d6	119		15 (10) - 110 (134)	79%	SPK: 150
4165-60-0	Nitrobenzene-d5	92.2		30 (49) - 130 (133)	92%	SPK: 100
321-60-8	2-Fluorobiphenyl	87.9		30 (52) - 130 (132)	88%	SPK: 100
118-79-6	2,4,6-Tribromophenol	137		15 (44) - 110 (137)	91%	SPK: 150
1718-51-0	Terphenyl-d14	88.8		30 (48) - 130 (125)	89%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	128000	6.887			
1146-65-2	Naphthalene-d8	508000	8.169			
15067-26-2	Acenaphthene-d10	289000	9.922			
1517-22-2	Phenanthrene-d10	536000	11.41			
1719-03-5	Chrysene-d12	334000	14.051			
1520-96-3	Perylene-d12	262000	15.527			



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	10/26/24	
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	10/26/24	
Client Sample ID:	PB164393TB			SDG No.:	P4578	
Lab Sample ID:	PB164393TB			Matrix:	TCLP	
Analytical Method:	SW8270			% Solid:	0	
Sample Wt/Vol:	100	Units:	mL	Final Vol:	1000	uL
Soil Aliquot Vol:	uL			Test:	TCLP BNA	
Extraction Type :				Decanted :	N	Level :
Injection Volume :				GPC Factor :	1.0	GPC Cleanup : N PH :
Prep Method :	SW3541					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140078.D	1	10/26/24 12:10	10/28/24 10:58	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140078.D
 Acq On : 28 Oct 2024 10:58
 Operator : RC/JU
 Sample : PB164393TB
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB164393TB

Quant Time: Oct 28 11:28:25 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.887	152	128211	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	508163	20.000	ng	0.00
39) Acenaphthene-d10	9.922	164	288600	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	536321	20.000	ng	0.00
76) Chrysene-d12	14.051	240	334195	20.000	ng	0.00
86) Perylene-d12	15.527	264	261975	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	1001524	122.289	ng	0.02
7) Phenol-d6	6.510	99	1264624	119.224	ng	0.00
23) Nitrobenzene-d5	7.446	82	845367	92.201	ng	0.00
42) 2,4,6-Tribromophenol	10.716	330	368703	136.583	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1534211	87.858	ng	0.00
79) Terphenyl-d14	12.998	244	1821703	88.823	ng	0.00




Target Compounds	Qvalue
(#= qualifier out of range (m) = manual integration (+) = signals summed	

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140078.D
Acq On : 28 Oct 2024 10:58
Operator : RC/JU
Sample : PB164393TB
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB164393TB

Quant Time: Oct 28 11:28:25 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.887 min Scan# 8
Delta R.T. -0.005 min
Lab File: BF140078.D
Acq: 28 Oct 2024 10:58


Instrument :
BNA_F
ClientSampleId :
PB164393TB


Tgt Ion:152 Resp: 128211
Ion Ratio Lower Upper
152 100
150 156.8 130.2 195.2
115 64.7 51.4 77.2



#5
2-Fluorophenol
Concen: 122.289 ng
RT: 5.516 min Scan# 582
Delta R.T. 0.018 min
Lab File: BF140078.D
Acq: 28 Oct 2024 10:58



Tgt Ion:112 Resp: 1001524
Ion Ratio Lower Upper
112 100
64 64.3 53.0 79.6
63 33.5 27.0 40.4






#7
Phenol-d6
Concen: 119.224 ng
RT: 6.510 min Scan# 7
Instrument: BNA_F
Delta R.T. 0.000 min
Lab File: BF140078.D
Acq: 28 Oct 2024 10:58
ClientSampleId : PB164393TB







Tgt Ion: 99 Resp: 1264624
Ion Ratio Lower Upper
99 100
42 19.9 16.7 25.1
71 36.0 27.7 41.5



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.169 min Scan# 1033
Delta R.T. -0.006 min
Lab File: BF140078.D
Acq: 28 Oct 2024 10:58

Tgt Ion:136 Resp: 508163
Ion Ratio Lower Upper
136 100
137 10.9 8.6 12.8
54 10.1 8.4 12.6
68 6.0 5.1 7.7






#23

Nitrobenzene-d5
Concen: 92.201 ngRT: 7.446 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF140078.D
Acq: 28 Oct 2024 10:58Instrument :
BNA_F
ClientSampleId :
PB164393TB


Tgt Ion: 82 Resp: 845367
Ion Ratio Lower Upper
82 100
128 42.3 33.4 50.0
54 56.9 47.8 71.8




#39

Acenaphthene-d10
Concen: 20.000 ng
RT: 9.922 min Scan# 1331
Delta R.T. -0.005 min
Lab File: BF140078.D
Acq: 28 Oct 2024 10:58

Tgt Ion:164 Resp: 288600
Ion Ratio Lower Upper
164 100
162 101.6 81.0 121.4
160 45.2 35.4 53.0







#42
2,4,6-Tribromophenol
Concen: 136.583 ng
RT: 10.716 min Scan# 1466
Delta R.T. 0.000 min
Lab File: BF140078.D
Acq: 28 Oct 2024 10:58


Instrument : BNA_F
ClientSampleId : PB164393TB


Tgt Ion:330 Resp: 368703
Ion Ratio Lower Upper
330 100
332 97.6 78.1 117.1
141 29.6 26.6 39.8



#45
2-Fluorobiphenyl
Concen: 87.858 ng
RT: 9.245 min Scan# 1216
Delta R.T. -0.006 min
Lab File: BF140078.D
Acq: 28 Oct 2024 10:58

Tgt Ion:172 Resp: 1534211
Ion Ratio Lower Upper
172 100
171 35.4 28.6 43.0
170 23.8 19.1 28.7







#64
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 11.410 min Scan# 1
 Delta R.T. -0.006 min
 Lab File: BF140078.D
 Acq: 28 Oct 2024 10:58


Instrument : BNA_F
 ClientSampleId : PB164393TB



Tgt Ion:188 Resp: 536321
 Ion Ratio Lower Upper
 188 100
 94 9.4 7.9 11.9
 80 10.6 9.0 13.4



#76
 Chrysene-d12
 Concen: 20.000 ng
 RT: 14.051 min Scan# 2033
 Delta R.T. 0.000 min
 Lab File: BF140078.D
 Acq: 28 Oct 2024 10:58

Tgt Ion:240 Resp: 334195
 Ion Ratio Lower Upper
 240 100
 120 10.5 9.4 14.2
 236 26.0 20.9 31.3







CALIBRATION

SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

6C

SEMICVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECHContract: PORT06Lab Code: CHEM Case No.: P4578SAS No.: P4578 SDG No.: P4578Instrument ID: BNA_FCalibration Date(s): 10/18/2024 10/18/2024Calibration Time(s): 10:27 13:46

LAB FILE ID:		RRF2.5 = BF139844.D		RRF005 = BF139845.D		RRF010 = BF139846.D			
		RRF020 = BF139847.D		RRF040 = BF139848.D		RRF050 = BF139849.D			
COMPOUND		RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Pyridine		1.622	1.570	1.522	1.504	1.386	1.476	7.3	
2-Fluorophenol		1.465	1.398	1.331	1.278	1.179	1.278	10.1	
Phenol-d6		1.900	1.818	1.709	1.647	1.538	1.655	10.1	
1,4-Dichlorobenzene		1.723	1.641	1.558	1.487	1.392	1.498	10.2	
2-Methylphenol		1.264	1.164	1.134	1.114	1.053	1.114	7.7	
3+4-Methylphenols		1.678	1.573	1.520	1.418	1.309	1.424	12.4	
Nitrobenzene-d5		0.365	0.365	0.372	0.371	0.354	0.361	2.9	
Hexachloroethane		0.583	0.571	0.549	0.541	0.507	0.534	7.1	
Nitrobenzene		0.417	0.402	0.408	0.403	0.383	0.396	4.1	
Hexachlorobutadiene		0.224	0.206	0.203	0.197	0.185	0.197	8.0	
2,4,6-Trichlorophenol		0.405	0.382	0.400	0.387	0.363	0.382	4.8	
2-Fluorobiphenyl		1.512	1.396	1.280	1.162	1.088	1.210	16.0	
2,4,5-Trichlorophenol		0.426	0.425	0.398	0.401	0.381	0.394	6.6	
2,4-Dinitrotoluene		0.280	0.314	0.337	0.356	0.344	0.332	7.9	
2,4,6-Tribromophenol		0.201	0.196	0.193	0.188	0.179	0.187	5.5	
Hexachlorobenzene		0.258	0.245	0.237	0.231	0.217	0.232	7.2	
Pentachlorophenol		0.118	0.137	0.148	0.152	0.145	0.141	8.0	
Terphenyl-d14		1.381	1.335	1.340	1.244	1.154	1.227	11.0	

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF101824.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Oct 18 15:07:50 2024
 Response Via : Initial Calibration

Calibration Files

2.5 =BF139844.D 5 =BF139845.D 10 =BF139846.D 20 =BF139847.D 40 =BF139848.D 50 =BF139849.D 60 =BF139850.D 80 =BF139851.D

	Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD	
<hr/>												
1) I	1,4-Dichlorobenzene					-----ISTD-----						
2)	1,4-Dioxane	0.651	0.625	0.603	0.591	0.571	0.581	0.548	0.596	5.74		
3)	Pyridine	1.622	1.570	1.522	1.504	1.386	1.406	1.326	1.476	7.26		
4)	n-Nitrosodimethylamine	0.815	0.800	0.799	0.806	0.782	0.794	0.757	0.793	2.39		
5) S	2-Fluorophenol	1.465	1.398	1.331	1.278	1.179	1.186	1.106	1.278	10.10		
6)	Aniline	1.673	1.649	1.618	1.582	1.471	1.479	1.324	1.542	8.05		
7) S	Phenol-d6	1.900	1.818	1.709	1.647	1.538	1.539	1.432	1.655	10.06		
8)	2-Chlorophenol	1.503	1.422	1.358	1.310	1.212	1.221	1.116	1.306	10.24		
9)	Benzaldehyde				1.137	1.042	0.940	0.873	0.853	0.740	0.931	15.22
10) C	Phenol	1.952	1.832	1.760	1.712	1.583	1.601	1.502	1.706	9.19		
11)	bis(2-Chloroethyl)ether	1.470	1.423	1.359	1.294	1.251	1.249	1.183	1.319	7.82		
12)	1,3-Dichlorobenzene	1.718	1.656	1.544	1.499	1.392	1.391	1.294	1.499	10.19		
13) C	1,4-Dichlorobenzene	1.723	1.641	1.558	1.487	1.392	1.391	1.291	1.498	10.19		
14)	1,2-Dichlorobenzene	1.660	1.579	1.478	1.379	1.273	1.267	1.149	1.398	13.15		
15)	Benzyl Alcohol	1.355	1.299	1.257	1.213	1.154	1.146	1.071	1.214	8.07		
16)	2,2'-oxybis(1,4-phenylene)	2.524	2.409	2.353	2.255	2.117	2.115	1.964	2.248	8.69		
17)	2-Methylphenol	1.264	1.164	1.134	1.114	1.053	1.063	1.004	1.114	7.69		
18)	Hexachloroethane	0.583	0.571	0.549	0.541	0.507	0.511	0.477	0.534	7.06		
19) P	n-Nitroso-di-n-butylamine	1.105	1.141	1.066	1.025	0.974	0.921	0.927	0.869	1.004	9.63	
20)	3+4-Methylphenols		1.678	1.573	1.520	1.418	1.309	1.300	1.173	1.424	12.41	
<hr/>												
21) I	Naphthalene-d8			-----ISTD-----								
22)	Acetophenone	0.578	0.533	0.516	0.481	0.452	0.454	0.414	0.490	11.43		
23) S	Nitrobenzene-d5	0.365	0.365	0.372	0.371	0.354	0.357	0.342	0.361	2.92		
24)	Nitrobenzene	0.417	0.402	0.408	0.403	0.383	0.388	0.370	0.396	4.10		
25)	Isophorone	0.755	0.709	0.702	0.679	0.652	0.660	0.637	0.685	5.90		
26) C	2-Nitrophenol	0.124	0.137	0.150	0.157	0.158	0.161	0.157	0.149	9.22		
27)	2,4-Dimethylphenol	0.285	0.259	0.256	0.248	0.237	0.234	0.223	0.249	8.07		
28)	bis(2-Chloroethyl)ether	0.468	0.440	0.432	0.412	0.394	0.390	0.369	0.415	8.22		
29) C	2,4-Dichlorophenol	0.308	0.297	0.293	0.283	0.272	0.274	0.257	0.283	6.17		
30)	1,2,4-Trichlorobenzene	0.351	0.331	0.325	0.315	0.298	0.298	0.279	0.314	7.68		
31)	Naphthalene	1.209	1.121	1.091	1.023	0.958	0.944	0.875	1.031	11.23		
32)	Benzoic acid		0.175	0.207	0.220	0.231	0.235	0.235	0.217	10.69		
33)	4-Chloroaniline	0.397	0.382	0.368	0.351	0.332	0.326	0.306	0.352	9.26		
34) C	Hexachlorobutane	0.224	0.206	0.203	0.197	0.185	0.187	0.177	0.197	7.96		
35)	Caprolactam	0.092	0.092	0.092	0.092	0.088	0.088	0.086	0.090	2.85		
36) C	4-Chloro-3-methylphenol	0.341	0.328	0.324	0.315	0.299	0.303	0.287	0.314	6.03		
37)	2-Methylnaphthalene	0.740	0.689	0.666	0.621	0.587	0.583	0.537	0.632	11.11		
38)	1-Methylnaphthalene		0.726	0.683	0.655	0.612	0.569	0.567	0.526	0.620	11.55	

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\

Method File : 8270-BF101824.M

39) I	Acenaphthene-d10	-----ISTD-----	
40)	1,2,4,5-Tetrac...	0.609 0.579 0.562 0.537 0.512 0.506 0.472 0.540	8.72
41) P	Hexachlorocycl...	0.185 0.193 0.198 0.198 0.186 0.185 0.169 0.188	5.28
42) S	2,4,6-Tribromo...	0.201 0.196 0.193 0.188 0.179 0.180 0.173 0.187	5.47
43) C	2,4,6-Trichlor...	0.405 0.382 0.400 0.387 0.363 0.383 0.354 0.382	4.80
44)	2,4,5-Trichlor...	0.426 0.425 0.398 0.401 0.381 0.369 0.359 0.394	6.58
45) S	2-Fluorobiphenyl	1.512 1.396 1.280 1.162 1.088 1.060 0.973 1.210	16.05
46)	1,1'-Biphenyl	1.640 1.536 1.483 1.379 1.285 1.262 1.161 1.392	12.18
47)	2-Chloronaphth...	1.288 1.225 1.164 1.111 1.048 1.040 0.973 1.121	9.95
48)	2-Nitroaniline	0.299 0.318 0.353 0.365 0.354 0.362 0.349 0.343	7.20
49)	Acenaphthylene	1.854 1.756 1.717 1.615 1.507 1.505 1.394 1.621	10.06
50)	Dimethylphthalate	1.410 1.344 1.283 1.237 1.172 1.163 1.110 1.246	8.60
51)	2,6-Dinitrotol...	0.256 0.266 0.278 0.280 0.273 0.274 0.260 0.270	3.41
52) C	Acenaphthene	1.200 1.136 1.089 1.044 0.977 0.983 0.913 1.049	9.55
53)	3-Nitroaniline	0.270 0.275 0.296 0.292 0.276 0.282 0.256 0.278	4.84
54) P	2,4-Dinitrophenol	0.056 0.077 0.101 0.101 0.113 0.112 0.093	23.89
55)	Dibenzofuran	1.767 1.659 1.579 1.486 1.389 1.375 1.278 1.505	11.52
56) P	4-Nitrophenol	0.187 0.207 0.225 0.232 0.219 0.220 0.208 0.214	6.84
57)	2,4-Dinitrotol...	0.280 0.314 0.337 0.356 0.344 0.351 0.338 0.332	7.94
58)	Fluorene	1.409 1.309 1.201 1.110 1.029 1.021 0.944 1.146	14.68
59)	2,3,4,6-Tetrac...	0.334 0.322 0.326 0.310 0.296 0.293 0.281 0.309	6.33
60)	Diethylphthalate	1.366 1.308 1.267 1.214 1.165 1.161 1.080 1.223	7.99
61)	4-Chlorophenyl...	0.698 0.638 0.617 0.569 0.526 0.520 0.484 0.579	13.14
62)	4-Nitroaniline	0.249 0.259 0.270 0.275 0.263 0.269 0.254 0.263	3.65
63)	Azobenzene	1.459 1.385 1.355 1.306 1.216 1.212 1.143 1.297	8.62
64) I	Phenanthrene-d10	-----ISTD-----	
65)	4,6-Dinitro-2....	0.055 0.072 0.087 0.090 0.092 0.093 0.082	18.60
66) c	n-Nitrosodiphe...	0.672 0.641 0.621 0.595 0.568 0.561 0.533 0.599	8.14
67)	4-Bromophenyl....	0.230 0.217 0.211 0.202 0.197 0.196 0.189 0.206	6.98
68)	Hexachlorobenzene	0.258 0.245 0.237 0.231 0.217 0.221 0.212 0.232	7.20
69)	Atrazine	0.189 0.175 0.156 0.175 0.135 0.153 0.151 0.162	11.36
70) C	Pentachlorophenol	0.118 0.137 0.148 0.152 0.145 0.144 0.140 0.141	8.04
71)	Phenanthrene	1.121 1.035 0.994 0.933 0.863 0.860 0.807 0.945	11.79
72)	Anthracene	1.082 1.014 0.972 0.913 0.851 0.833 0.787 0.922	11.53
73)	Carbazole	1.002 0.964 0.923 0.846 0.776 0.772 0.717 0.857	12.64
74)	Di-n-butylphth...	1.104 1.071 1.062 1.014 0.923 0.909 0.850 0.990	9.77
75) C	Fluoranthene	1.149 1.108 1.036 0.943 0.842 0.835 0.772 0.955	15.34
76) I	Chrysene-d12	-----ISTD-----	
77)	Benzidine	0.457 0.461 0.293 0.366 0.276 0.203 0.246 0.329	30.86
78)	Pyrene	1.892 1.828 1.900 1.805 1.685 1.649 1.496 1.751	8.43
79) S	Terphenyl-d14	1.381 1.335 1.340 1.244 1.154 1.121 1.018 1.227	10.96
80)	Butylbenzylpht...	0.490 0.513 0.536 0.555 0.535 0.531 0.514 0.525	3.99
81)	Benzo(a)anthra...	1.425 1.355 1.329 1.331 1.267 1.237 1.169 1.302	6.48
82)	3,3'-Dichlorob...	0.368 0.372 0.390 0.387 0.374 0.382 0.384 0.380	2.15
83)	Chrysene	1.325 1.244 1.234 1.167 1.134 1.151 1.101 1.194	6.52
84)	Bis(2-ethylhex...	0.520 0.539 0.577 0.626 0.620 0.626 0.614 0.589	7.48
85) c	Di-n-octyl pht...	0.786 0.930 1.135 1.182 1.207 1.186 1.071	16.14

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\


Method File : 8270-BF101824.M

86)	I	Perylene-d12	- - - - - ISTD - - - - -										
87)		Indeno(1,2,3-c...)	1.209	1.261	1.307	1.352	1.299	1.326	1.253	1.287			3.77
88)		Benzo(b)fluora...	1.317	1.240	1.319	1.196	1.105	1.239	1.111	1.218			7.15
89)		Benzo(k)fluora...	1.213	1.177	0.992	1.066	1.030	0.929	0.947	1.051			10.42
90)	C	Benzo(a)pyrene	1.030	1.024	1.018	1.025	0.974	0.999	0.947	1.002			3.12
91)		Dibenzo(a,h)an...	1.021	1.064	1.103	1.120	1.083	1.085	1.036	1.073			3.31
92)		Benzo(g,h,i)pe...	1.030	1.046	1.090	1.128	1.081	1.095	1.035	1.072			3.37

(#) = Out of Range

2,4-Dinitrophenol

Response Ratio



$$\text{Response} = 1.220\text{e-001} * \text{Amt} - 3.983\text{e-002}$$

Coef of Det (r^2) = 0.997175 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF101824.M

Calibration Table Last Updated: Fri Oct 18 15:07:50 2024

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139844.D
 Acq On : 18 Oct 2024 10:27
 Operator : RC/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Quant Time: Oct 18 14:15:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 14:11:20 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.887	152	190003	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	748539	20.000	ng	0.00
39) Acenaphthene-d10	9.928	164	423999	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	760140	20.000	ng	0.00
76) Chrysene-d12	14.051	240	519381	20.000	ng	0.00
86) Perylene-d12	15.527	264	410408	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112		0d	0.000	ng
7) Phenol-d6	0.000	99		0d	0.000	ng
23) Nitrobenzene-d5	0.000	82		0d	0.000	ng
42) 2,4,6-Tribromophenol	0.000	330		0d	0.000	ng
45) 2-Fluorobiphenyl	0.000	172		0d	0.000	ng
79) Terphenyl-d14	0.000	244		0d	0.000	ng
Target Compounds						
19) n-Nitroso-di-n-propyla...	7.287	70	26252	2.754	ng	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139844.D
Acq On : 18 Oct 2024 10:27
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5



Quant Time: Oct 18 14:15:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.887 min Scan# 8
Instrument : BNA_F
Delta R.T. -0.005 min
Lab File: BF139844.D
ClientSampleId : SSTDICC2.5
Acq: 18 Oct 2024 10:27


Tgt Ion:152 Resp: 190003
Ion Ratio Lower Upper
152 100
150 158.9 130.2 195.2
115 64.4 51.4 77.2



#5
2-Fluorophenol
Concen: 0.000 ng
Expected RT: 5.50 min

Lab File: BF139844.D
Acq: 18 Oct 2024 10:27


Tgt Ion: 112
Sig Exp Ratio
112 100
64 66.3
63 33.7



#7
Phenol-d6
Concen: 0.000 ng
Expected RT: 6.51 min
Lab File: BF139844.D
Acq: 18 Oct 2024 10:27


Instrument:
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion: 99
Sig Exp Ratio
99 100
42 20.9
71 34.6



#19
n-Nitroso-di-n-propylamine
Concen: 2.754 ng
RT: 7.287 min Scan# 883
Delta R.T. -0.017 min
Lab File: BF139844.D
Acq: 18 Oct 2024 10:27

Tgt Ion: 70 Resp: 26252
Ion Ratio Lower Upper
70 100
42 70.4 51.5 77.3
101 10.6 7.5 11.3
130 18.9 15.5 23.3






#21
 Naphthalene-d8
 Concen: 20.000 ng
 RT: 8.169 min Scan# 1
 Delta R.T. -0.006 min
 Lab File: BF139844.D
 Acq: 18 Oct 2024 10:27

Instrument : BNA_F
 ClientSampleId : SSTDICC2.5

Tgt Ion:136 Resp: 748539

Ion Ratio Lower Upper


136	100		
137	10.9	8.6	12.8
54	10.6	8.4	12.6
68	6.5	5.1	7.7






#23
 Nitrobenzene-d5
 Concen: 0.000 ng
 Expected RT: 7.45 min

Lab File: BF139844.D
 Acq: 18 Oct 2024 10:27

Tgt Ion: 82
 Sig Exp Ratio
 82 100
 128 41.7
 54 59.8





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.928 min Scan# 1

Delta R.T. 0.001 min

Lab File: BF139844.D

Acq: 18 Oct 2024 10:27

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5



Tgt Ion:164 Resp: 423999

Ion Ratio Lower Upper

164 100

162 99.5 81.0 121.4

160 43.8 35.4 53.0



#42

2,4,6-Tribromophenol

Concen: 0.000 ng

Expected RT: 10.72 min

Lab File: BF139844.D

Acq: 18 Oct 2024 10:27


Tgt Ion: 330






Sig Exp Ratio

330 100

332 97.6

141 33.2





#45
2-Fluorobiphenyl
Concen: 0.000 ng
Expected RT: 9.25 min


Lab File: BF139844.D
Acq: 18 Oct 2024 10:27


Tgt Ion: 172
Sig Exp Ratio
172 100
171 35.8
170 23.9

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1584
Delta R.T. -0.006 min
Lab File: BF139844.D
Acq: 18 Oct 2024 10:27



Tgt Ion:188 Resp: 760140
Ion Ratio Lower Upper
188 100
94 10.4 7.9 11.9
80 11.6 9.0 13.4






#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.051 min Scan# 2
Instrument: BNA_F
Delta R.T. 0.000 min
Lab File: BF139844.D
ClientSampleId : SSTDICC2.5
Acq: 18 Oct 2024 10:27

Tgt Ion:240 Resp: 519381
Ion Ratio Lower Upper
240 100
120 11.9 9.4 14.2
236 26.4 20.9 31.3



#79
Terphenyl-d14
Concen: 0.000 ng
Expected RT: 13.00 min
Lab File: BF139844.D
Acq: 18 Oct 2024 10:27


Tgt Ion: 244
Sig Exp Ratio
244 100
212 7.1
122 10.8



#86
Perylene-d₁₂
Concen: 20.000 ng
RT: 15.527 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139844.D
Acq: 18 Oct 2024 10:27

Instrument : BNA_F
ClientSampleId : SSTDICC2.5

Tgt Ion:264 Resp: 410408
Ion Ratio Lower Upper
264 100
260 23.9 19.4 29.2
265 21.5 17.4 26.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139845.D
 Acq On : 18 Oct 2024 10:55
 Operator : RC/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005

Quant Time: Oct 18 14:16:29 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 14:11:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.887	152	189112	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	737062	20.000	ng	0.00
39) Acenaphthene-d10	9.928	164	414324	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	741430	20.000	ng	0.00
76) Chrysene-d12	14.051	240	477223	20.000	ng	0.00
86) Perylene-d12	15.527	264	377685	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	138495	11.465	ng	0.00
7) Phenol-d6	6.493	99	179648	11.482	ng	-0.02
23) Nitrobenzene-d5	7.445	82	134439	10.109	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	41682	10.755	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	313171	12.492	ng	0.00
79) Terphenyl-d14	12.992	244	329424	11.248	ng	0.00
Target Compounds						
				Qvalue		
2) 1,4-Dioxane	2.687	88	30759	5.461	ng	97
3) Pyridine	3.452	79	76703	5.494	ng	98
4) n-Nitrosodimethylamine	3.375	42	38552	5.140	ng	# 96
6) Aniline	6.545	93	79083	5.424	ng	98
8) 2-Chlorophenol	6.669	128	71065	5.754	ng	98
10) Phenol	6.510	94	92287	5.620	ng	95
11) bis(2-Chloroethyl)ether	6.616	93	69504	5.575	ng	98
12) 1,3-Dichlorobenzene	6.828	146	81239	5.731	ng	99
13) 1,4-Dichlorobenzene	6.904	146	81442	5.750	ng	99
14) 1,2-Dichlorobenzene	7.057	146	78458	5.936	ng	97
15) Benzyl Alcohol	7.016	79	64060	5.582	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.163	45	119350	5.614	ng	100
17) 2-Methylphenol	7.122	107	59770	5.676	ng	96
18) Hexachloroethane	7.404	117	27566	5.458	ng	96
19) n-Nitroso-di-n-propyla...	7.287	70	53958	5.686	ng	97
20) 3+4-Methylphenols	7.275	107	79329	5.890	ng	92
22) Acetophenone	7.287	105	106522	5.900	ng	98
24) Nitrobenzene	7.463	77	76831	5.269	ng	96
25) Isophorone	7.698	82	139084	5.510	ng	100
26) 2-Nitrophenol	7.781	139	22867	4.157	ng	95
27) 2,4-Dimethylphenol	7.810	122	52430	5.716	ng	97
28) bis(2-Chloroethoxy)met...	7.916	93	86298	5.643	ng	99
29) 2,4-Dichlorophenol	8.016	162	56772	5.434	ng	99
30) 1,2,4-Trichlorobenzene	8.110	180	64640	5.592	ng	99
31) Naphthalene	8.192	128	222725	5.859	ng	99
33) 4-Chloroaniline	8.234	127	73196	5.647	ng	99
34) Hexachlorobutadiene	8.310	225	41186	5.670	ng	97
35) Caprolactam	8.557	113	16989	5.114	ng	98
36) 4-Chloro-3-methylphenol	8.704	107	62903	5.438	ng	99
37) 2-Methylnaphthalene	8.881	142	136292	5.854	ng	99
38) 1-Methylnaphthalene	8.981	142	133747	5.856	ng	100
40) 1,2,4,5-Tetrachloroben...	9.045	216	63046	5.640	ng	100
41) Hexachlorocyclopentadiene	9.039	237	19121	4.916	ng	97
43) 2,4,6-Trichlorophenol	9.151	196	41997	5.307	ng	98
44) 2,4,5-Trichlorophenol	9.187	196	44086	5.399	ng	99
46) 1,1'-Biphenyl	9.345	154	169920	5.891	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139845.D
 Acq On : 18 Oct 2024 10:55
 Operator : RC/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005

Quant Time: Oct 18 14:16:29 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 14:11:20 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
47) 2-Chloronaphthalene	9.369	162	133456	5.745	ng	99
48) 2-Nitroaniline	9.457	65	30999	4.363	ng	99
49) Acenaphthylene	9.786	152	192014	5.717	ng	99
50) Dimethylphthalate	9.639	163	146096	5.661	ng	100
51) 2,6-Dinitrotoluene	9.698	165	26490	4.741	ng	96
52) Acenaphthene	9.957	154	124288	5.721	ng	100
53) 3-Nitroaniline	9.863	138	27995	4.858	ng	# 94
55) Dibenzofuran	10.128	168	182985	5.870	ng	98
56) 4-Nitrophenol	10.004	139	19402	4.376	ng	89
57) 2,4-Dinitrotoluene	10.098	165	29038	4.226	ng	92
58) Fluorene	10.469	166	145913	6.146	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.239	232	34636	5.411	ng	97
60) Diethylphthalate	10.339	149	141466	5.583	ng	99
61) 4-Chlorophenyl-phenyle...	10.469	204	72331	6.033	ng	96
62) 4-Nitroaniline	10.469	138	25747	4.731	ng	96
63) Azobenzene	10.622	77	151158	5.627	ng	99
66) n-Nitrosodiphenylamine	10.581	169	124472	5.609	ng	100
67) 4-Bromophenyl-phenylether	10.957	248	42707	5.591	ng	94
68) Hexachlorobenzene	11.016	284	47895	5.575	ng	98
69) Atrazine	11.104	200	35075	5.835	ng	99
70) Pentachlorophenol	11.210	266	21781	4.178	ng	93
71) Phenanthrene	11.433	178	207716	5.931	ng	98
72) Anthracene	11.486	178	200621	5.871	ng	99
73) Carbazole	11.639	167	185743	5.845	ng	99
74) Di-n-butylphthalate	11.975	149	204647	5.574	ng	99
75) Fluoranthene	12.622	202	213038	6.017	ng	100
77) Benzidine	12.745	184	54501	7.022	ng	99
78) Pyrene	12.851	202	225742	5.404	ng	99
80) Butylbenzylphthalate	13.474	149	58502	4.671	ng	100
81) Benzo(a)anthracene	14.039	228	170026	5.473	ng	99
82) 3,3'-Dichlorobenzidine	14.004	252	43937	4.851	ng	100
83) Chrysene	14.074	228	158121	5.551	ng	98
84) Bis(2-ethylhexyl)phtha...	14.033	149	62085	4.419	ng	97
87) Indeno(1,2,3-cd)pyrene	17.010	276	114185	4.699	ng	98
88) Benzo(b)fluoranthene	15.092	252	124353	5.405	ng	100
89) Benzo(k)fluoranthene	15.121	252	114525	5.772	ng	99
90) Benzo(a)pyrene	15.463	252	97243	5.137	ng	100
91) Dibenzo(a,h)anthracene	17.033	278	96416	4.757	ng	99
92) Benzo(g,h,i)perylene	17.457	276	97292	4.805	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139845.D
 Acq On : 18 Oct 2024 10:55
 Operator : RC/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005

Quant Time: Oct 18 14:16:29 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 14:11:20 2024
 Response via : Initial Calibration







#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.887 min Scan# 8
 Delta R.T. -0.005 min
 Lab File: BF139845.D
 Acq: 18 Oct 2024 10:55


Instrument : BNA_F
 ClientSampleId : SSTDICC005


Tgt Ion:152 Resp: 189112
 Ion Ratio Lower Upper
 152 100
 150 157.4 130.2 195.2
 115 65.7 51.4 77.2



#2
 1,4-Dioxane
 Concen: 5.461 ng
 RT: 2.687 min Scan# 101
 Delta R.T. 0.000 min
 Lab File: BF139845.D
 Acq: 18 Oct 2024 10:55


Tgt Ion: 88 Resp: 30759
 Ion Ratio Lower Upper
 88 100
 58 89.3 73.3 109.9
 43 35.6 27.1 40.7







#3
 Pyridine
 Concen: 5.494 ng
 RT: 3.452 min Scan# 2
 Delta R.T. 0.006 min
 Lab File: BF139845.D
 Acq: 18 Oct 2024 10:55


Instrument : BNA_F
 ClientSampleId : SSTDICC005




Tgt Ion: 79 Resp: 76703
 Ion Ratio Lower Upper
 79 100
 52 86.7 68.5 102.7
 51 42.2 32.3 48.5




#4
 n-Nitrosodimethylamine
 Concen: 5.140 ng
 RT: 3.375 min Scan# 218
 Delta R.T. -0.023 min
 Lab File: BF139845.D
 Acq: 18 Oct 2024 10:55



Tgt Ion: 42 Resp: 38552
 Ion Ratio Lower Upper
 42 100
 74 115.9 96.2 144.4
 44 11.0 6.2 9.2#







#5
2-Fluorophenol
Concen: 11.465 ng
RT: 5.493 min Scan# 5
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55


Instrument : BNA_F
ClientSampleId : SSTDICC005


Tgt Ion:112 Resp: 138495
Ion Ratio Lower Upper
112 100
64 65.5 53.0 79.6
63 33.2 27.0 40.4



#6
Aniline
Concen: 5.424 ng
RT: 6.545 min Scan# 757
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55



Tgt Ion: 93 Resp: 79083
Ion Ratio Lower Upper
93 100
66 41.4 34.0 51.0
65 21.1 17.4 26.0






#7
Phenol-d6
Concen: 11.482 ng
RT: 6.493 min Scan# 7
Instrument : BNA_F
Delta R.T. -0.017 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55
ClientSampleId : SSTDICC005


Tgt Ion: 99 Resp: 179648
Ion Ratio Lower Upper
99 100
42 20.7 16.7 25.1
71 34.3 27.7 41.5



#8
2-Chlorophenol
Concen: 5.754 ng
RT: 6.669 min Scan# 778
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:128 Resp: 71065
Ion Ratio Lower Upper
128 100
130 33.3 12.3 52.3
64 52.8 34.6 74.6







#10
 Phenol
 Concen: 5.620 ng
 RT: 6.510 min Scan# 7
 Delta R.T. -0.012 min
 Lab File: BF139845.D
 Acq: 18 Oct 2024 10:55


Instrument : BNA_F
 ClientSampleId : SSTDICC005


Tgt Ion: 94 Resp: 92287
 Ion Ratio Lower Upper
 94 100
 65 29.4 11.1 51.1
 66 38.6 22.7 62.7




#11
 bis(2-Chloroethyl)ether
 Concen: 5.575 ng
 RT: 6.616 min Scan# 769
 Delta R.T. -0.012 min
 Lab File: BF139845.D
 Acq: 18 Oct 2024 10:55

Tgt Ion: 93 Resp: 69504
 Ion Ratio Lower Upper
 93 100
 63 87.5 65.0 105.0
 95 32.6 12.4 52.4









#12
1,3-Dichlorobenzene
Concen: 5.731 ng
RT: 6.828 min Scan# 8
Instrument : BNA_F
Delta R.T. -0.006 min
Lab File: BF139845.D
ClientSampleId : SSTDICC005
Acq: 18 Oct 2024 10:55





Tgt Ion:146 Resp: 81239
Ion Ratio Lower Upper
146 100
148 64.8 51.0 76.4
75 33.2 26.4 39.6






#13
1,4-Dichlorobenzene
Concen: 5.750 ng
RT: 6.904 min Scan# 818
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55



Tgt Ion:146 Resp: 81442
Ion Ratio Lower Upper
146 100
148 63.7 51.8 77.8
111 43.5 34.6 51.8





#14

1,2-Dichlorobenzene

Concen: 5.936 ng

RT: 7.057 min Scan# 8

Delta R.T. -0.006 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:146 Resp: 78458

Ion Ratio Lower Upper

146 100

148 63.0 51.4 77.2

111 43.3 36.7 55.1



#15

Benzyl Alcohol

Concen: 5.582 ng

RT: 7.016 min Scan# 837

Delta R.T. -0.012 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55


Tgt Ion: 79 Resp: 64060


Ion Ratio Lower Upper

79 100

108 75.6 61.4 92.0



77 62.3 49.9 74.9






#16
2,2'-oxybis(1-Chloropropane)
Concen: 5.614 ng
RT: 7.163 min Scan# 8
Instrument: BNA_F
Delta R.T. -0.006 min
Lab File: BF139845.D
ClientSampleId : SSTDICC005
Acq: 18 Oct 2024 10:55




Tgt Ion: 45 Resp: 119350
Ion Ratio Lower Upper
45 100
77 11.8 0.0 31.5
79 9.1 0.0 29.1



#17
2-Methylphenol
Concen: 5.676 ng
RT: 7.122 min Scan# 855
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:107 Resp: 59770
Ion Ratio Lower Upper
107 100
108 111.2 91.3 136.9
77 43.6 37.1 55.7
79 42.5 37.0 55.6





#18

Hexachloroethane

Concen: 5.458 ng

RT: 7.404 min Scan# 9

Delta R.T. 0.000 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion: 117 Resp: 27566

Ion Ratio Lower Upper

117 100

119 99.1 77.0 115.4

201 101.0 76.2 114.4



#19

n-Nitroso-di-n-propylamine

Concen: 5.686 ng

RT: 7.287 min Scan# 883

Delta R.T. -0.017 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55

Tgt Ion: 70 Resp: 53958


Ion Ratio Lower Upper




70 100

42 68.0 51.5 77.3

101 9.2 7.5 11.3

130 19.0 15.5 23.3









#20
3+4-Methylphenols
Concen: 5.890 ng
RT: 7.275 min Scan# 8
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument : BNA_F
ClientSampleId : SSTDICC005

Tgt Ion:107 Resp: 79329


Ion	Ratio	Lower	Upper
107	100		
108	89.7	69.8	109.8
77	54.7	50.5	90.5
79	28.9	10.2	50.2




#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.169 min Scan# 1033
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:136 Resp: 737062

Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.6	12.8
54	10.7	8.4	12.6
68	6.9	5.1	7.7







#22
 Acetophenone
 Concen: 5.900 ng
 RT: 7.287 min Scan# 8
 Delta R.T. -0.012 min
 Lab File: BF139845.D
 Acq: 18 Oct 2024 10:55

Instrument : BNA_F
 ClientSampleId : SSTDICC005

Tgt Ion:105 Resp: 106522

Ion Ratio Lower Upper

105	100		
71	8.6	5.9	8.9
51	36.4	29.4	44.2
120	20.4	17.5	26.3







#23
 Nitrobenzene-d5
 Concen: 10.109 ng
 RT: 7.445 min Scan# 910
 Delta R.T. -0.006 min
 Lab File: BF139845.D
 Acq: 18 Oct 2024 10:55

Tgt Ion: 82 Resp: 134439

Ion Ratio Lower Upper

82	100		
128	42.0	33.4	50.0
54	60.5	47.8	71.8





#24

Nitrobenzene

Concen: 5.269 ng

RT: 7.463 min Scan# 9

Instrument :

BNA_F

Delta R.T. -0.012 min

Lab File: BF139845.D

ClientSampleId :

Acq: 18 Oct 2024 10:55

SSTDICC005





Tgt Ion: 77 Resp: 76831

Ion Ratio Lower Upper

77 100

123 41.4 35.7 53.5

65 14.0 11.5 17.3



#25

Isophorone

Concen: 5.510 ng

RT: 7.698 min Scan# 953

Delta R.T. -0.012 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55


Tgt Ion: 82 Resp: 139084


Ion Ratio Lower Upper

82 100

95 6.8 5.5 8.3



138 16.5 13.2 19.8






#26
2-Nitrophenol
Concen: 4.157 ng
RT: 7.781 min Scan# 9
Instrument: BNA_F
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55
ClientSampleId : SSTDICC005




Tgt Ion:139 Resp: 22867
Ion Ratio Lower Upper
139 100
109 33.3 25.1 37.7
65 63.8 47.5 71.3



#27
2,4-Dimethylphenol
Concen: 5.716 ng
RT: 7.810 min Scan# 972
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:122 Resp: 52430
Ion Ratio Lower Upper
122 100
107 112.5 91.9 137.9
121 54.7 46.2 69.4





#28

bis(2-Chloroethoxy)methane

Concen: 5.643 ng

RT: 7.916 min Scan# 9

Delta R.T. -0.006 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55

Instrument:

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion: 93 Resp: 86298

Ion Ratio Lower Upper

93 100

95 32.5 26.0 39.0

123 10.6 9.8 14.8



#29

2,4-Dichlorophenol

Concen: 5.434 ng

RT: 8.016 min Scan# 1007

Delta R.T. -0.006 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55


Tgt Ion: 162 Resp: 56772


Ion Ratio Lower Upper

162 100

164 63.6 43.9 83.9

98 39.2 19.9 59.9







#30
 1,2,4-Trichlorobenzene
 Concen: 5.592 ng
 RT: 8.110 min Scan# 1
 Delta R.T. -0.006 min
 Lab File: BF139845.D
 Acq: 18 Oct 2024 10:55


Instrument : BNA_F
 ClientSampleId : SSTDICC005







Tgt Ion:180 Resp: 64640
 Ion Ratio Lower Upper
 180 100
 182 98.8 80.5 120.7
 145 31.8 25.4 38.0



#31
 Naphthalene
 Concen: 5.859 ng
 RT: 8.192 min Scan# 1037
 Delta R.T. -0.006 min
 Lab File: BF139845.D
 Acq: 18 Oct 2024 10:55

Tgt Ion:128 Resp: 222725
 Ion Ratio Lower Upper
 128 100
 129 10.7 8.9 13.3
 127 12.9 10.6 16.0





#33

4-Chloroaniline

Concen: 5.647 ng

RT: 8.234 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55

Instrument : BNA_F

ClientSampleId : SSTDICC005

Tgt Ion:127 Resp: 73196


Ion Ratio Lower Upper

127 100

129 32.9 25.4 38.2

65 34.5 28.4 42.6

92 20.6 16.6 25.0



#34

Hexachlorobutadiene

Concen: 5.670 ng

RT: 8.310 min Scan# 1057

Delta R.T. -0.006 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55


Tgt Ion:225 Resp: 41186




Ion Ratio Lower Upper

225 100

223 62.1 50.6 76.0

227 61.9 51.8 77.8





#35

Caprolactam

Concen: 5.114 ng

RT: 8.557 min Scan# 1

Delta R.T. -0.053 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:113 Resp: 16989

Ion Ratio Lower Upper

113 100

55 212.2 189.2 229.2

56 156.7 132.5 172.5



#36

4-Chloro-3-methylphenol

Concen: 5.438 ng

RT: 8.704 min Scan# 1124

Delta R.T. -0.006 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55


Tgt Ion:107 Resp: 62903






Ion Ratio Lower Upper

107 100

144 24.3 19.1 28.7

142 72.0 57.2 85.8





#37

2-Methylnaphthalene

Concen: 5.854 ng

RT: 8.881 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55

Instrument :

BNA_F

ClientSampleId :


SSTDICC005

Tgt Ion:142 Resp: 136292

Ion Ratio Lower Upper

142 100

141 87.8 70.8 106.2



#38

1-Methylnaphthalene

Concen: 5.856 ng

RT: 8.981 min Scan# 1171

Delta R.T. -0.006 min

Lab File: BF139845.D


Acq: 18 Oct 2024 10:55




Tgt Ion:142 Resp: 133747

Ion Ratio Lower Upper

142 100

141 91.7 73.5 110.3





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.928 min Scan# 1

Delta R.T. 0.001 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:164 Resp: 414324

Ion Ratio Lower Upper

164 100

162 101.3 81.0 121.4

160 44.1 35.4 53.0



#40

1,2,4,5-Tetrachlorobenzene

Concen: 5.640 ng

RT: 9.045 min Scan# 1182

Delta R.T. -0.006 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55

Tgt Ion:216 Resp: 63046


Ion Ratio Lower Upper


216 100

214 79.3 63.4 95.0

179 21.4 17.2 25.8

108 18.7 15.6 23.4





#41

Hexachlorocyclopentadiene

Concen: 4.916 ng

RT: 9.039 min Scan# 1

Instrument:

Delta R.T. 0.000 min

BNA_F

Lab File: BF139845.D

ClientSampleId :

Acq: 18 Oct 2024 10:55

SSTDICC005



Tgt Ion:237 Resp: 19121

Ion Ratio Lower Upper

237 100

235 65.2 42.9 82.9

272 12.4 0.0 32.3



#42

2,4,6-Tribromophenol

Concen: 10.755 ng

RT: 10.710 min Scan# 1465

Delta R.T. -0.006 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55


Tgt Ion:330 Resp: 41682




Ion Ratio Lower Upper

330 100

332 96.2 78.1 117.1

141 32.3 26.6 39.8





#43

2,4,6-Trichlorophenol

Concen: 5.307 ng

RT: 9.151 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:196 Resp: 41997

Ion Ratio Lower Upper

196 100

198 92.8 76.2 114.2

200 28.8 24.0 36.0



#44

2,4,5-Trichlorophenol

Concen: 5.399 ng

RT: 9.187 min Scan# 1206

Delta R.T. -0.006 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55

Tgt Ion:196 Resp: 44086


Ion Ratio Lower Upper


196 100

198 96.0 75.8 113.6

97 51.2 41.8 62.6

132 29.1 22.9 34.3







#45
2-Fluorobiphenyl
Concen: 12.492 ng
RT: 9.245 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55


Instrument : BNA_F
ClientSampleId : SSTDICC005


Tgt Ion:172 Resp: 313171
Ion Ratio Lower Upper
172 100
171 35.6 28.6 43.0
170 23.5 19.1 28.7



#46
1,1'-Biphenyl
Concen: 5.891 ng
RT: 9.345 min Scan# 1233
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:154 Resp: 169920
Ion Ratio Lower Upper
154 100
153 39.7 20.3 60.3
76 14.1 0.0 33.9





#47

2-Chloronaphthalene

Concen: 5.745 ng

RT: 9.369 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55

Instrument :

BNA_F

ClientSampleId :

SSTDICC005



Tgt Ion:162 Resp: 133456

Ion Ratio Lower Upper

162 100

127 41.4 33.8 50.6

164 32.7 26.6 39.8



#48

2-Nitroaniline

Concen: 4.363 ng

RT: 9.457 min Scan# 1252

Delta R.T. -0.006 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55


Tgt Ion: 65 Resp: 30999




Ion Ratio Lower Upper

65 100

92 65.6 50.9 76.3

138 91.1 73.4 110.0





#49

Acenaphthylene
Concen: 5.717 ng
RT: 9.786 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :

BNA_F





ClientSampleId :

SSTDICC005

Tgt Ion:152 Resp: 192014

Ion Ratio Lower Upper

152	100		
151	19.5	15.9	23.9
153	13.4	10.7	16.1




#50




Dimethylphthalate
Concen: 5.661 ng
RT: 9.639 min Scan# 1283
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:163 Resp: 146096

Ion Ratio Lower Upper





163	100		
194	4.3	3.4	5.0
164	10.1	8.2	12.2






#51
2,6-Dinitrotoluene
Concen: 4.741 ng
RT: 9.698 min Scan# 1
Instrument : BNA_F
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55
ClientSampleId : SSTDICC005


Tgt Ion:165 Resp: 26490
Ion Ratio Lower Upper
165 100
63 66.6 50.6 75.8
89 61.8 47.5 71.3



#52
Acenaphthene
Concen: 5.721 ng
RT: 9.957 min Scan# 1337
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55


Tgt Ion:154 Resp: 124288
Ion Ratio Lower Upper
154 100
153 112.9 89.8 134.6
152 53.0 42.4 63.6








#53
3-Nitroaniline
Concen: 4.858 ng
RT: 9.863 min Scan# 1
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55


Instrument : BNA_F
ClientSampleId : SSTDICC005





Tgt Ion:138 Resp: 27995
Ion Ratio Lower Upper
138 100
108 17.3 10.2 15.2#
92 128.5 98.3 147.5




#55
Dibenzofuran
Concen: 5.870 ng
RT: 10.128 min Scan# 1366
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55





Tgt Ion:168 Resp: 182985
Ion Ratio Lower Upper
168 100
139 38.5 31.8 47.6
169 14.3 11.6 17.4






#56
4-Nitrophenol
Concen: 4.376 ng
RT: 10.004 min Scan# 1
Instrument : BNA_F
Delta R.T. -0.017 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55
ClientSampleId : SSTDICC005


Tgt Ion:139 Resp: 19402
Ion Ratio Lower Upper
139 100
109 75.6 53.6 93.6
65 124.8 87.8 127.8



#57
2,4-Dinitrotoluene
Concen: 4.226 ng
RT: 10.098 min Scan# 1361
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55



Tgt Ion:165 Resp: 29038
Ion Ratio Lower Upper
165 100
63 55.0 40.1 60.1
89 83.3 60.4 90.6
182 2.5 2.0 3.0






#58
Fluorene
Concen: 6.146 ng
RT: 10.469 min Scan# 1
Instrument : BNA_F
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55
ClientSampleId : SSTDICC005


Tgt Ion:166 Resp: 145913
Ion Ratio Lower Upper
166 100
165 97.5 78.5 117.7
167 13.2 10.9 16.3



#59
2,3,4,6-Tetrachlorophenol
Concen: 5.411 ng
RT: 10.239 min Scan# 1385
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55



Tgt Ion:232 Resp: 34636
Ion Ratio Lower Upper
232 100
131 43.7 37.0 55.6
130 2.3 1.9 2.9
166 29.5 24.2 36.2






#60
Diethylphthalate
Concen: 5.583 ng
RT: 10.339 min Scan# 1
Instrument : BNA_F
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55
ClientSampleId : SSTDICC005


Tgt Ion:149 Resp: 141466
Ion Ratio Lower Upper
149 100
177 22.0 17.9 26.9
150 12.0 10.1 15.1



#61
4-Chlorophenyl-phenylether
Concen: 6.033 ng
RT: 10.469 min Scan# 1424
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55



Tgt Ion:204 Resp: 72331
Ion Ratio Lower Upper
204 100
206 33.3 26.3 39.5
141 56.2 48.6 73.0






#62
4-Nitroaniline
Concen: 4.731 ng
RT: 10.469 min Scan# 1
Instrument: BNA_F
Delta R.T. -0.017 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55
ClientSampleId : SSTDICC005


Tgt Ion:138 Resp: 25747
Ion Ratio Lower Upper
138 100
92 57.8 35.7 75.7
108 81.3 57.5 97.5



#63
Azobenzene
Concen: 5.627 ng
RT: 10.622 min Scan# 1450
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion: 77 Resp: 151158
Ion Ratio Lower Upper
77 100
182 23.2 3.3 43.3
105 21.5 1.2 41.2
51 32.4 13.3 53.3







#64
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 11.410 min Scan# 1
 Delta R.T. -0.006 min
 Lab File: BF139845.D
 Acq: 18 Oct 2024 10:55


Instrument : BNA_F
 ClientSampleId : SSTDICC005




Tgt Ion:188 Resp: 741430
 Ion Ratio Lower Upper
 188 100
 94 10.2 7.9 11.9
 80 11.4 9.0 13.4



#66
 n-Nitrosodiphenylamine
 Concen: 5.609 ng
 RT: 10.581 min Scan# 1443
 Delta R.T. -0.006 min
 Lab File: BF139845.D
 Acq: 18 Oct 2024 10:55

Tgt Ion:169 Resp: 124472
 Ion Ratio Lower Upper
 169 100
 168 61.9 49.8 74.6
 167 35.0 28.0 42.0





#67

4-Bromophenyl-phenylether

Concen: 5.591 ng

RT: 10.957 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:248 Resp: 42707

Ion Ratio Lower Upper

248 100

250 93.5 79.0 118.6

141 71.1 61.7 92.5



#68

Hexachlorobenzene

Concen: 5.575 ng

RT: 11.016 min Scan# 1517

Delta R.T. -0.006 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55


Tgt Ion:284 Resp: 47895


Ion Ratio Lower Upper

284 100

142 35.4 26.9 40.3

249 31.8 25.3 37.9







#69
Atrazine
Concen: 5.835 ng
RT: 11.104 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55


Instrument : BNA_F
ClientSampleId : SSTDICC005


Tgt Ion:200 Resp: 35075
Ion Ratio Lower Upper
200 100
173 26.4 7.7 47.7
215 48.2 27.7 67.7



#70
Pentachlorophenol
Concen: 4.178 ng
RT: 11.210 min Scan# 1550
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55



Tgt Ion:266 Resp: 21781
Ion Ratio Lower Upper
266 100
268 68.6 50.4 75.6
264 68.9 50.6 75.8






#71
Phenanthrene
Concen: 5.931 ng
RT: 11.433 min Scan# 1
Instrument: BNA_F
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55
ClientSampleId : SSTDICC005


Tgt Ion:178 Resp: 207716
Ion Ratio Lower Upper
178 100
176 18.6 15.8 23.6
179 14.9 12.6 18.8



#72
Anthracene
Concen: 5.871 ng
RT: 11.486 min Scan# 1597
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:178 Resp: 200621
Ion Ratio Lower Upper
178 100
176 18.4 15.3 22.9
179 15.0 12.4 18.6







#73
Carbazole
Concen: 5.845 ng
RT: 11.639 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55


Instrument : BNA_F
ClientSampleId : SSTDICC005


Tgt Ion:167 Resp: 185743
Ion Ratio Lower Upper
167 100
166 21.4 17.0 25.6
139 12.7 10.8 16.2



#74
Di-n-butylphthalate
Concen: 5.574 ng
RT: 11.975 min Scan# 1680
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:149 Resp: 204647
Ion Ratio Lower Upper
149 100
150 8.9 7.5 11.3
104 5.0 4.1 6.1







#75
Fluoranthene
Concen: 6.017 ng
RT: 12.622 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55


Instrument : BNA_F
ClientSampleId : SSTDICC005


Tgt Ion:202 Resp: 213038
Ion Ratio Lower Upper
202 100
101 12.2 0.0 31.9
203 17.5 0.0 37.5



#76
Chrysene-d₁₂
Concen: 20.000 ng
RT: 14.051 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:240 Resp: 477223
Ion Ratio Lower Upper
240 100
120 11.1 9.4 14.2
236 26.2 20.9 31.3







#77
Benzidine
Concen: 7.022 ng
RT: 12.745 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55


Instrument : BNA_F
ClientSampleId : SSTDICC005


Tgt Ion:184 Resp: 54501
Ion Ratio Lower Upper
184 100
185 15.5 12.0 18.0
183 11.4 9.4 14.2



#78
Pyrene
Concen: 5.404 ng
RT: 12.851 min Scan# 1829
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:202 Resp: 225742
Ion Ratio Lower Upper
202 100
200 21.3 17.2 25.8
203 17.4 14.2 21.4







#79
 Terphenyl-d14
 Concen: 11.248 ng
 RT: 12.992 min Scan# 1
 Delta R.T. -0.006 min
 Lab File: BF139845.D
 Acq: 18 Oct 2024 10:55


Instrument : BNA_F
 ClientSampleId : SSTDICC005




Tgt Ion:244 Resp: 329424
 Ion Ratio Lower Upper
 244 100
 212 7.0 5.7 8.5
 122 11.6 8.6 13.0



#80
 Butylbenzylphthalate
 Concen: 4.671 ng
 RT: 13.474 min Scan# 1935
 Delta R.T. 0.000 min
 Lab File: BF139845.D
 Acq: 18 Oct 2024 10:55

Tgt Ion:149 Resp: 58502
 Ion Ratio Lower Upper
 149 100
 91 74.4 59.4 89.2
 206 20.3 16.4 24.6





#81

Benzo(a)anthracene

Concen: 5.473 ng

RT: 14.039 min Scan# 2

Delta R.T. 0.000 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:228 Resp: 170026

Ion Ratio Lower Upper

228 100

226 26.6 21.6 32.4

229 19.9 16.1 24.1



#82

3,3'-Dichlorobenzidine

Concen: 4.851 ng

RT: 14.004 min Scan# 2025

Delta R.T. 0.000 min

Lab File: BF139845.D

Acq: 18 Oct 2024 10:55


Tgt Ion:252 Resp: 43937


Ion Ratio Lower Upper

252 100

254 63.3 50.8 76.2

126 11.1 8.9 13.3







#83
Chrysene
Concen: 5.551 ng
RT: 14.074 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55


Instrument : BNA_F
ClientSampleId : SSTDICC005


Tgt Ion:228 Resp: 158121
Ion Ratio Lower Upper
228 100
226 29.3 24.1 36.1
229 18.6 15.8 23.6



#84
Bis(2-ethylhexyl)phthalate
Concen: 4.419 ng
RT: 14.033 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55



Tgt Ion:149 Resp: 62085
Ion Ratio Lower Upper
149 100
167 24.8 21.4 32.0
279 3.7 3.0 4.4






#86
Perylene-d₁₂
Concen: 20.000 ng
RT: 15.527 min Scan# 2
Instrument: BNA_F
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55
ClientSampleId : SSTDICC005


Tgt Ion:264 Resp: 377685
Ion Ratio Lower Upper
264 100
260 24.5 19.4 29.2
265 21.5 17.4 26.0



#87
Indeno(1,2,3-cd)pyrene
Concen: 4.699 ng
RT: 17.010 min Scan# 2536
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:276 Resp: 114185
Ion Ratio Lower Upper
276 100
138 27.1 20.4 30.6
277 25.3 20.2 30.4







#88
 Benzo(b)fluoranthene
 Concen: 5.405 ng
 RT: 15.092 min Scan# 2
 Delta R.T. -0.006 min
 Lab File: BF139845.D
 Acq: 18 Oct 2024 10:55


Instrument : BNA_F
 ClientSampleId : SSTDICC005


Tgt Ion:252 Resp: 124353
 Ion Ratio Lower Upper
 252 100
 253 22.2 17.6 26.4
 125 9.7 7.9 11.9



#89
 Benzo(k)fluoranthene
 Concen: 5.772 ng
 RT: 15.121 min Scan# 2215
 Delta R.T. -0.006 min
 Lab File: BF139845.D
 Acq: 18 Oct 2024 10:55



Tgt Ion:252 Resp: 114525
 Ion Ratio Lower Upper
 252 100
 253 21.4 17.6 26.4
 125 10.0 7.9 11.9






#90
Benzo(a)pyrene
Concen: 5.137 ng
RT: 15.463 min Scan# 2
Instrument: BNA_F
Delta R.T. -0.006 min
Lab File: BF139845.D
ClientSampleId : SSTDICC005
Acq: 18 Oct 2024 10:55

Tgt Ion:252 Resp: 97243
Ion Ratio Lower Upper
252 100
253 21.6 17.3 25.9
125 10.9 8.7 13.1



#91
Dibenzo(a,h)anthracene
Concen: 4.757 ng
RT: 17.033 min Scan# 2540
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:278 Resp: 96416
Ion Ratio Lower Upper
278 100
139 16.6 13.1 19.7
279 24.4 19.0 28.4



#92

Benzo(g,h,i)perylene

Concen: 4.805 ng

RT: 17.457 min Scan# 2 Instrument :

Delta R.T. -0.023 min BNA_F

Lab File: BF139845.D ClientSampleId :

Acq: 18 Oct 2024 10:55 SSTDICC005





Tgt Ion:276 Resp: 97292

Ion Ratio Lower Upper

276 100

277 23.1 19.0 28.6

138 21.3 17.4 26.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139846.D
 Acq On : 18 Oct 2024 11:23
 Operator : RC/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010

Quant Time: Oct 18 14:17:15 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 14:11:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.887	152	188831	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	733621	20.000	ng	0.00
39) Acenaphthene-d10	9.928	164	411556	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	732661	20.000	ng	0.00
76) Chrysene-d12	14.051	240	452443	20.000	ng	0.00
86) Perylene-d12	15.527	264	375863	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	264054	21.891	ng	0.00
7) Phenol-d6	6.498	99	343334	21.977	ng	-0.01
23) Nitrobenzene-d5	7.445	82	268128	20.257	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	80655	20.952	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	574563	23.073	ng	0.00
79) Terphenyl-d14	12.998	244	603787	21.745	ng	0.00
Target Compounds						
				Qvalue		
2) 1,4-Dioxane	2.681	88	58969	10.485	ng	97
3) Pyridine	3.440	79	148260	10.636	ng	97
4) n-Nitrosodimethylamine	3.375	42	75504	10.081	ng	96
6) Aniline	6.546	93	155680	10.692	ng	98
8) 2-Chlorophenol	6.669	128	134290	10.890	ng	99
9) Benzaldehyde	6.440	77	107330	12.213	ng	98
10) Phenol	6.510	94	172965	10.549	ng	98
11) bis(2-Chloroethyl)ether	6.616	93	134367	10.794	ng	100
12) 1,3-Dichlorobenzene	6.828	146	156338	11.045	ng	99
13) 1,4-Dichlorobenzene	6.904	146	154957	10.957	ng	99
14) 1,2-Dichlorobenzene	7.057	146	149129	11.299	ng	98
15) Benzyl Alcohol	7.022	79	122669	10.704	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.163	45	227443	10.715	ng	99
17) 2-Methylphenol	7.128	107	109901	10.452	ng	98
18) Hexachloroethane	7.404	117	53892	10.686	ng	99
19) n-Nitroso-di-n-propyla...	7.293	70	100635	10.621	ng	98
20) 3+4-Methylphenols	7.275	107	148497	11.041	ng	# 90
22) Acetophenone	7.293	105	195595	10.885	ng	96
24) Nitrobenzene	7.463	77	147356	10.154	ng	98
25) Isophorone	7.698	82	260247	10.359	ng	98
26) 2-Nitrophenol	7.781	139	50159	9.162	ng	100
27) 2,4-Dimethylphenol	7.810	122	94947	10.400	ng	99
28) bis(2-Chloroethoxy)met...	7.916	93	161423	10.605	ng	99
29) 2,4-Dichlorophenol	8.016	162	109102	10.492	ng	100
30) 1,2,4-Trichlorobenzene	8.110	180	121232	10.537	ng	99
31) Naphthalene	8.192	128	411053	10.864	ng	99
32) Benzoic acid	7.869	122	64294	8.075	ng	99
33) 4-Chloroaniline	8.234	127	140021	10.853	ng	100
34) Hexachlorobutadiene	8.310	225	75740	10.477	ng	99
35) Caprolactam	8.569	113	33908	10.256	ng	99
36) 4-Chloro-3-methylphenol	8.704	107	120339	10.452	ng	98
37) 2-Methylnaphthalene	8.881	142	252554	10.899	ng	99
38) 1-Methylnaphthalene	8.981	142	250609	11.023	ng	99
40) 1,2,4,5-Tetrachloroben...	9.045	216	119227	10.738	ng	99
41) Hexachlorocyclopentadiene	9.039	237	39754	10.290	ng	98
43) 2,4,6-Trichlorophenol	9.151	196	78578	9.996	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139846.D
 Acq On : 18 Oct 2024 11:23
 Operator : RC/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010

Quant Time: Oct 18 14:17:15 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 14:11:20 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.187	196	87482	10.787	ng	100
46) 1,1'-Biphenyl	9.345	154	316146	11.035	ng	99
47) 2-Chloronaphthalene	9.369	162	252140	10.927	ng	98
48) 2-Nitroaniline	9.457	65	65425	9.270	ng	98
49) Acenaphthylene	9.787	152	361398	10.833	ng	99
50) Dimethylphthalate	9.639	163	276645	10.792	ng	99
51) 2,6-Dinitrotoluene	9.698	165	54824	9.877	ng	95
52) Acenaphthene	9.957	154	233749	10.832	ng	98
53) 3-Nitroaniline	9.869	138	56574	9.884	ng	# 98
54) 2,4-Dinitrophenol	9.969	184	11577	11.138	ng	# 1
55) Dibenzofuran	10.128	168	341446	11.028	ng	98
56) 4-Nitrophenol	10.010	139	42688	9.694	ng	98
57) 2,4-Dinitrotoluene	10.098	165	64666	9.474	ng	94
58) Fluorene	10.475	166	269323	11.420	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.239	232	66170	10.408	ng	97
60) Diethylphthalate	10.345	149	269178	10.694	ng	100
61) 4-Chlorophenyl-phenyle...	10.469	204	131191	11.016	ng	97
62) 4-Nitroaniline	10.475	138	53288	9.857	ng	98
63) Azobenzene	10.622	77	284960	10.679	ng	100
65) 4,6-Dinitro-2-methylph...	10.504	198	20120	6.732	ng	97
66) n-Nitrosodiphenylamine	10.581	169	234741	10.705	ng	99
67) 4-Bromophenyl-phenylether	10.957	248	79420	10.522	ng	98
68) Hexachlorobenzene	11.016	284	89927	10.594	ng	98
69) Atrazine	11.104	200	64267	10.819	ng	99
70) Pentachlorophenol	11.210	266	50361	9.775	ng	100
71) Phenanthrene	11.433	178	379233	10.958	ng	99
72) Anthracene	11.486	178	371391	10.999	ng	99
73) Carbazole	11.639	167	353253	11.249	ng	99
74) Di-n-butylphthalate	11.975	149	392482	10.818	ng	99
75) Fluoranthene	12.622	202	405919	11.602	ng	99
77) Benzidine	12.745	184	104253	14.168	ng	99
78) Pyrene	12.851	202	413606	10.444	ng	99
80) Butylbenzylphthalate	13.474	149	116053	9.774	ng	100
81) Benzo(a)anthracene	14.039	228	306594	10.410	ng	99
82) 3,3'-Dichlorobenzidine	14.004	252	84205	9.806	ng	99
83) Chrysene	14.074	228	281467	10.423	ng	99
84) Bis(2-ethylhexyl)phtha...	14.033	149	121932	9.153	ng	99
85) Di-n-octyl phthalate	14.651	149	177824	7.339	ng	99
87) Indeno(1,2,3-cd)pyrene	17.010	276	236914	9.798	ng	99
88) Benzo(b)fluoranthene	15.092	252	232988	10.176	ng	99
89) Benzo(k)fluoranthene	15.121	252	221219	11.203	ng	99
90) Benzo(a)pyrene	15.463	252	192493	10.217	ng	100
91) Dibenzo(a,h)anthracene	17.033	278	199946	9.912	ng	99
92) Benzo(g,h,i)perylene	17.457	276	196609	9.758	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139846.D
Acq On : 18 Oct 2024 11:23
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Quant Time: Oct 18 14:17:15 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration







#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.887 min Scan# 8
 Delta R.T. -0.005 min
 Lab File: BF139846.D
 Acq: 18 Oct 2024 11:23


Instrument : BNA_F
 ClientSampleId : SSTDICC010


Tgt Ion:152 Resp: 188831
 Ion Ratio Lower Upper
 152 100
 150 157.0 130.2 195.2
 115 64.1 51.4 77.2



#2
 1,4-Dioxane
 Concen: 10.485 ng
 RT: 2.681 min Scan# 100
 Delta R.T. -0.006 min
 Lab File: BF139846.D
 Acq: 18 Oct 2024 11:23


Tgt Ion: 88 Resp: 58969
 Ion Ratio Lower Upper
 88 100
 58 87.5 73.3 109.9
 43 34.3 27.1 40.7







#3
 Pyridine
 Concen: 10.636 ng
 RT: 3.440 min Scan# 2
 Delta R.T. -0.006 min
 Lab File: BF139846.D
 Acq: 18 Oct 2024 11:23


Instrument : BNA_F
 ClientSampleId : SSTDICC010





Tgt Ion: 79 Resp: 148260
 Ion Ratio Lower Upper
 79 100
 52 88.5 68.5 102.7
 51 41.6 32.3 48.5




#4
 n-Nitrosodimethylamine
 Concen: 10.081 ng
 RT: 3.375 min Scan# 218
 Delta R.T. -0.023 min
 Lab File: BF139846.D
 Acq: 18 Oct 2024 11:23



Tgt Ion: 42 Resp: 75504
 Ion Ratio Lower Upper
 42 100
 74 115.8 96.2 144.4
 44 8.2 6.2 9.2







#5
2-Fluorophenol
Concen: 21.891 ng
RT: 5.493 min Scan# 5
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23


Instrument : BNA_F
ClientSampleId : SSTDICC010


Tgt Ion:112 Resp: 264054
Ion Ratio Lower Upper
112 100
64 65.6 53.0 79.6
63 34.0 27.0 40.4



#6
Aniline
Concen: 10.692 ng
RT: 6.546 min Scan# 757
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion: 93 Resp: 155680
Ion Ratio Lower Upper
93 100
66 41.1 34.0 51.0
65 21.2 17.4 26.0







#7
 Phenol-d6
 Concen: 21.977 ng
 RT: 6.498 min Scan# 7
 Delta R.T. -0.012 min
 Lab File: BF139846.D
 Acq: 18 Oct 2024 11:23


Instrument : BNA_F
 ClientSampleId : SSTDICC010


Tgt Ion: 99 Resp: 343334
 Ion Ratio Lower Upper
 99 100
 42 20.3 16.7 25.1
 71 32.7 27.7 41.5



#8
 2-Chlorophenol
 Concen: 10.890 ng
 RT: 6.669 min Scan# 778
 Delta R.T. -0.006 min
 Lab File: BF139846.D
 Acq: 18 Oct 2024 11:23


Tgt Ion:128 Resp: 134290
 Ion Ratio Lower Upper
 128 100
 130 32.4 12.3 52.3
 64 54.1 34.6 74.6








#9
 Benzaldehyde
 Concen: 12.213 ng
 RT: 6.440 min Scan# 7
 Delta R.T. 0.001 min
 Lab File: BF139846.D
 Acq: 18 Oct 2024 11:23


Instrument : BNA_F
 ClientSampleId : SSTDICC010





Tgt Ion: 77 Resp: 107330
 Ion Ratio Lower Upper
 77 100
 105 94.7 75.5 115.5
 106 94.6 71.7 111.7









#10
 Phenol
 Concen: 10.549 ng
 RT: 6.510 min Scan# 751
 Delta R.T. -0.012 min
 Lab File: BF139846.D
 Acq: 18 Oct 2024 11:23



Tgt Ion: 94 Resp: 172965
 Ion Ratio Lower Upper
 94 100
 65 30.1 11.1 51.1
 66 41.0 22.7 62.7





#11

bis(2-Chloroethyl)ether

Concen: 10.794 ng

RT: 6.616 min Scan# 7

Delta R.T. -0.012 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23

Instrument :

BNA_F

ClientSampleId :

SSTDICC010


Tgt Ion: 93 Resp: 134367

Ion Ratio Lower Upper

93 100

63 85.2 65.0 105.0

95 31.8 12.4 52.4



#12

1,3-Dichlorobenzene

Concen: 11.045 ng

RT: 6.828 min Scan# 805

Delta R.T. -0.006 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23


Tgt Ion:146 Resp: 156338




Ion Ratio Lower Upper

146 100

148 62.9 51.0 76.4

75 31.8 26.4 39.6





#13

1,4-Dichlorobenzene
Concen: 10.957 ng
RT: 6.904 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :

BNA_F





ClientSampleId :

SSTDICC010

Tgt Ion:146 Resp: 154957

Ion Ratio Lower Upper

146	100		
148	63.7	51.8	77.8
111	43.7	34.6	51.8




#14




1,2-Dichlorobenzene
Concen: 11.299 ng
RT: 7.057 min Scan# 844
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:146 Resp: 149129

Ion Ratio Lower Upper

146	100		
148	63.1	51.4	77.2
111	44.0	36.7	55.1





#15

Benzyl Alcohol

Concen: 10.704 ng

RT: 7.022 min Scan# 8

Delta R.T. -0.006 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23

Instrument:

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion: 79 Resp: 122669

Ion Ratio Lower Upper

79 100

108 77.7 61.4 92.0

77 62.8 49.9 74.9



#16

2,2'-oxybis(1-Chloropropane)

Concen: 10.715 ng

RT: 7.163 min Scan# 862

Delta R.T. -0.006 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23


Tgt Ion: 45 Resp: 227443




Ion Ratio Lower Upper

45 100

77 11.2 0.0 31.5

79 8.7 0.0 29.1





#17

2-Methylphenol

Concen: 10.452 ng

RT: 7.128 min Scan# 8

Instrument :

BNA_F

Delta R.T. -0.006 min

Lab File: BF139846.D

ClientSampleId :

Acq: 18 Oct 2024 11:23

SSTDICC010

Tgt Ion:107 Resp: 109901

Ion Ratio Lower Upper

107 100





108 116.1 91.3 136.9

77 44.4 37.1 55.7

79 45.6 37.0 55.6

Abundance

7.128



#18

Hexachloroethane

Concen: 10.686 ng

RT: 7.404 min Scan# 903

Delta R.T. 0.000 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23

Tgt Ion:117 Resp: 53892


Ion Ratio Lower Upper


117 100

119 94.7 77.0 115.4

201 96.4 76.2 114.4

Abundance







#19
n-Nitroso-di-n-propylamine
Concen: 10.621 ng
RT: 7.293 min Scan# 8
Instrument : BNA_F
Delta R.T. -0.012 min
Lab File: BF139846.D
ClientSampleId : SSTDICC010
Acq: 18 Oct 2024 11:23

Tgt Ion: 70 Resp: 100635

Ion Ratio Lower Upper

	70	100
42	66.4	51.5
101	9.6	7.5
130	19.4	15.5
		23.3






#20
3+4-Methylphenols
Concen: 11.041 ng
RT: 7.275 min Scan# 881
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23




Tgt Ion:107 Resp: 148497

Ion Ratio Lower Upper

	107	100
108	90.1	69.8
77	49.6	50.5
79	28.8	10.2
		50.2







#23

Nitrobenzene-d5

Concen: 20.257 ng

RT: 7.445 min Scan# 913

Delta R.T. -0.006 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion: 82 Resp: 268128

Ion Ratio Lower Upper

82 100

128 41.3 33.4 50.0

54 59.5 47.8 71.8



#24

Nitrobenzene

Concen: 10.154 ng

RT: 7.463 min Scan# 913

Delta R.T. -0.012 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23


Tgt Ion: 77 Resp: 147356


Ion Ratio Lower Upper

77 100

123 43.0 35.7 53.5



65 14.6 11.5 17.3






#25
Isophorone
Concen: 10.359 ng
RT: 7.698 min Scan# 9
Instrument : BNA_F
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23
ClientSampleId : SSTDICC010




Tgt Ion: 82 Resp: 260247
Ion Ratio Lower Upper
82 100
95 6.6 5.5 8.3
138 15.7 13.2 19.8



#26
2-Nitrophenol
Concen: 9.162 ng
RT: 7.781 min Scan# 967
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23




Tgt Ion:139 Resp: 50159
Ion Ratio Lower Upper
139 100
109 30.9 25.1 37.7
65 59.2 47.5 71.3







#27
2,4-Dimethylphenol
Concen: 10.400 ng
RT: 7.810 min Scan# 9
Instrument : BNA_F
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23
ClientSampleId : SSTDICC010




Tgt Ion:122 Resp: 94947
Ion Ratio Lower Upper
122 100
107 114.1 91.9 137.9
121 58.0 46.2 69.4



#28
bis(2-Chloroethoxy)methane
Concen: 10.605 ng
RT: 7.916 min Scan# 990
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion: 93 Resp: 161423
Ion Ratio Lower Upper
93 100
95 33.0 26.0 39.0
123 11.2 9.8 14.8





#29

2,4-Dichlorophenol
Concen: 10.492 ng
RT: 8.016 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:162 Resp: 109102

Ion Ratio Lower Upper

162 100

164 64.1 43.9 83.9

98 39.8 19.9 59.9



#30

1,2,4-Trichlorobenzene
Concen: 10.537 ng
RT: 8.110 min Scan# 1023
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23


Tgt Ion:180 Resp: 121232




Ion Ratio Lower Upper

180 100

182 101.5 80.5 120.7

145 32.7 25.4 38.0





#31

Naphthalene

Concen: 10.864 ng

RT: 8.192 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:128 Resp: 411053

Ion Ratio Lower Upper

128 100

129 10.9 8.9 13.3

127 13.0 10.6 16.0



#32

Benzoic acid

Concen: 8.075 ng

RT: 7.869 min Scan# 982

Delta R.T. -0.053 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23


Tgt Ion:122 Resp: 64294


Ion Ratio Lower Upper

122 100

105 126.9 107.6 147.6

77 99.2 80.2 120.2







#33
4-Chloroaniline
Concen: 10.853 ng
RT: 8.234 min Scan# 1
Instrument: BNA_F
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23
ClientSampleId : SSTDICC010

Tgt Ion:127 Resp: 140021
Ion Ratio Lower Upper


127	100		
129	31.8	25.4	38.2
65	35.4	28.4	42.6
92	20.4	16.6	25.0






#34
Hexachlorobutadiene
Concen: 10.477 ng
RT: 8.310 min Scan# 1057
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:225 Resp: 75740
Ion Ratio Lower Upper





225	100		
223	63.0	50.6	76.0
227	64.3	51.8	77.8






#35
Caprolactam
Concen: 10.256 ng
RT: 8.569 min Scan# 1
Instrument: BNA_F
Delta R.T. -0.041 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23
ClientSampleId : SSTDICC010


Tgt Ion:113 Resp: 33908
Ion Ratio Lower Upper
113 100
55 208.3 189.2 229.2
56 151.4 132.5 172.5





#36
4-Chloro-3-methylphenol
Concen: 10.452 ng
RT: 8.704 min Scan# 1124
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:107 Resp: 120339
Ion Ratio Lower Upper
107 100
144 25.2 19.1 28.7
142 72.5 57.2 85.8






Abundance Scan 1154 (8.881 min): BF139846.D\data.ms



Abundance Scan 1171 (8.981 min): BF139846.D\data.ms



#37

2-Methylnaphthalene

Concen: 10.899 ng

RT: 8.881 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23

Instrument :

BNA_F

ClientSampleId :


SSTDICC010

Tgt Ion:142 Resp: 252554

Ion Ratio Lower Upper

142 100

141 89.7 70.8 106.2



#38

1-Methylnaphthalene

Concen: 11.023 ng

RT: 8.981 min Scan# 1171

Delta R.T. -0.006 min

Lab File: BF139846.D


Acq: 18 Oct 2024 11:23




Tgt Ion:142 Resp: 250609

Ion Ratio Lower Upper

142 100

141 91.3 73.5 110.3





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.928 min Scan# 1

Delta R.T. 0.001 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23

Instrument :

BNA_F

ClientSampleId :

SSTDICC010




Tgt Ion:164 Resp: 411556

Ion Ratio Lower Upper

164 100

162 100.2 81.0 121.4

160 43.4 35.4 53.0



#40

1,2,4,5-Tetrachlorobenzene

Concen: 10.738 ng

RT: 9.045 min Scan# 1182

Delta R.T. -0.006 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23

Tgt Ion:216 Resp: 119227


Ion Ratio Lower Upper

216 100

214 78.1 63.4 95.0

179 21.2 17.2 25.8


108 18.7 15.6 23.4



Abundance Scan 1182 (9.045 min): BF139846.D\data.ms (-1)

Sub

m/z-->



#41

Hexachlorocyclopentadiene

Concen: 10.290 ng

RT: 9.039 min Scan# 1

Instrument:

BNA_F

Delta R.T. 0.000 min

Lab File: BF139846.D

ClientSampleId :

Acq: 18 Oct 2024 11:23

SSTDICC010

Tgt Ion:237 Resp: 39754

Ion Ratio Lower Upper

237 100

235 61.3 42.9 82.9

272 12.3 0.0 32.3

Abundance

30000

20000

10000

0

9.039

Time-->

#42

2,4,6-Tribromophenol

Concen: 20.952 ng

RT: 10.710 min Scan# 1465

Delta R.T. -0.006 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23

Tgt Ion:330 Resp: 80655

Ion Ratio Lower Upper

330 100

332 95.9 78.1 117.1

141 34.2 26.6 39.8

Abundance

60000

40000




20000

0

10.710

Time-->

BF139846.D 8270-BF101824.M Sat Oct 19 03:56:43 2024



#43

2,4,6-Trichlorophenol

Concen: 9.996 ng

RT: 9.151 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:196 Resp: 78578

Ion Ratio Lower Upper

196 100

198 95.7 76.2 114.2

200 29.8 24.0 36.0



#44

2,4,5-Trichlorophenol

Concen: 10.787 ng

RT: 9.187 min Scan# 1206

Delta R.T. -0.006 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23

Tgt Ion:196 Resp: 87482


Ion Ratio Lower Upper


196 100

198 94.9 75.8 113.6

97 52.0 41.8 62.6

132 28.0 22.9 34.3







#45
2-Fluorobiphenyl
Concen: 23.073 ng
RT: 9.245 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23


Instrument : BNA_F
ClientSampleId : SSTDICC010




Tgt Ion:172 Resp: 574563
Ion Ratio Lower Upper
172 100
171 35.2 28.6 43.0
170 23.5 19.1 28.7



#46
1,1'-Biphenyl
Concen: 11.035 ng
RT: 9.345 min Scan# 1233
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:154 Resp: 316146
Ion Ratio Lower Upper
154 100
153 39.4 20.3 60.3
76 13.9 0.0 33.9





#47

2-Chloronaphthalene

Concen: 10.927 ng

RT: 9.369 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:162 Resp: 252140

Ion Ratio Lower Upper

162 100

127 40.9 33.8 50.6

164 31.8 26.6 39.8



#48

2-Nitroaniline

Concen: 9.270 ng

RT: 9.457 min Scan# 1252

Delta R.T. -0.006 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23


Tgt Ion: 65 Resp: 65425




Ion Ratio Lower Upper

65 100

92 65.5 50.9 76.3

138 92.7 73.4 110.0





#49

Acenaphthylene

Concen: 10.833 ng

RT: 9.787 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23

Instrument:

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:152 Resp: 361398

Ion Ratio Lower Upper

152 100

151 19.9 15.9 23.9

153 12.7 10.7 16.1



#50

Dimethylphthalate

Concen: 10.792 ng

RT: 9.639 min Scan# 1283

Delta R.T. -0.012 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23


Tgt Ion:163 Resp: 276645




Ion Ratio Lower Upper

163 100

194 3.7 3.4 5.0





164 10.0 8.2 12.2






#51
2,6-Dinitrotoluene
Concen: 9.877 ng
RT: 9.698 min Scan# 1
Instrument : BNA_F
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23
ClientSampleId : SSTDICC010




Tgt Ion:165 Resp: 54824
Ion Ratio Lower Upper
165 100
63 66.0 50.6 75.8
89 64.1 47.5 71.3



#52
Acenaphthene
Concen: 10.832 ng
RT: 9.957 min Scan# 1337
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:154 Resp: 233749
Ion Ratio Lower Upper
154 100
153 109.8 89.8 134.6
152 52.9 42.4 63.6








#53
3-Nitroaniline
Concen: 9.884 ng
RT: 9.869 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23



Instrument : BNA_F
ClientSampleId : SSTDICC010


Tgt Ion:138 Resp: 56574
Ion Ratio Lower Upper
138 100
108 15.7 10.2 15.2#
92 124.7 98.3 147.5



#54
2,4-Dinitrophenol
Concen: 11.138 ng
RT: 9.969 min Scan# 1339
Delta R.T. -0.005 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23



Tgt Ion:184 Resp: 11577
Ion Ratio Lower Upper
184 100
63 142.4 84.7 127.1#
154 870.2 338.2 507.2#






#55
Dibenzofuran
Concen: 11.028 ng
RT: 10.128 min Scan# 1
Instrument: BNA_F
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23
ClientSampleId : SSTDICC010


Tgt Ion:168 Resp: 341446
Ion Ratio Lower Upper
168 100
139 38.2 31.8 47.6
169 13.9 11.6 17.4



#56
4-Nitrophenol
Concen: 9.694 ng
RT: 10.010 min Scan# 1346
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23



Tgt Ion:139 Resp: 42688
Ion Ratio Lower Upper
139 100
109 74.3 53.6 93.6
65 111.6 87.8 127.8






#57
2,4-Dinitrotoluene
Concen: 9.474 ng
RT: 10.098 min Scan# 1
Instrument : BNA_F
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23
ClientSampleId : SSTDICC010


Tgt Ion:165 Resp: 64666
Ion Ratio Lower Upper
165 100
63 53.7 40.1 60.1
89 81.9 60.4 90.6
182 2.9 2.0 3.0



#58
Fluorene
Concen: 11.420 ng
RT: 10.475 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:166 Resp: 269323
Ion Ratio Lower Upper
166 100
165 97.4 78.5 117.7
167 13.4 10.9 16.3







#59
2,3,4,6-Tetrachlorophenol
Concen: 10.408 ng
RT: 10.239 min Scan# 1
Instrument : BNA_F
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23
ClientSampleId : SSTDICC010

Tgt Ion:232 Resp: 66170
Ion Ratio Lower Upper

232	100
131	44.4
130	2.3
166	29.0


Abundance




#60
Diethylphthalate
Concen: 10.694 ng
RT: 10.345 min Scan# 1403
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:149 Resp: 269178
Ion Ratio Lower Upper

149	100
177	22.4
150	12.3





#61

4-Chlorophenyl-phenylether

Concen: 11.016 ng

RT: 10.469 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23

Instrument :

BNA_F

ClientSampleId :

SSTDICC010



Tgt Ion:204 Resp: 131191

Ion Ratio Lower Upper

204 100

206 32.5 26.3 39.5

141 58.0 48.6 73.0



#62

4-Nitroaniline

Concen: 9.857 ng

RT: 10.475 min Scan# 1425

Delta R.T. -0.012 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23


Tgt Ion:138 Resp: 53288


Ion Ratio Lower Upper

138 100

92 54.6 35.7 75.7

108 75.6 57.5 97.5





#63
Azobenzene
Concen: 10.679 ng
RT: 10.622 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument : BNA_F
ClientSampleId : SSTDICC010



Tgt Ion: 77 Resp: 284960

Ion Ratio Lower Upper

	77	100
182	23.3	3.3
105	21.3	1.2
51	33.8	13.3

	43.3
	53.3

Abundance



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1584
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23


Tgt Ion:188 Resp: 732661


Ion Ratio Lower Upper

	188	100
94	10.5	7.9
80	11.5	9.0

	11.9
	13.4



Abundance






#65
4,6-Dinitro-2-methylphenol
Concen: 6.732 ng
RT: 10.504 min Scan# 1
Instrument: BNA_F
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23
ClientSampleId : SSTDICC010




Tgt Ion:198 Resp: 20120
Ion Ratio Lower Upper
198 100
51 61.0 38.2 78.2
105 49.9 27.9 67.9



#66
n-Nitrosodiphenylamine
Concen: 10.705 ng
RT: 10.581 min Scan# 1443
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:169 Resp: 234741
Ion Ratio Lower Upper
169 100
168 63.0 49.8 74.6
167 34.3 28.0 42.0





#67

4-Bromophenyl-phenylether

Concen: 10.522 ng

RT: 10.957 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:248 Resp: 79420

Ion Ratio Lower Upper

248 100

250 96.8 79.0 118.6

141 74.5 61.7 92.5



#68

Hexachlorobenzene

Concen: 10.594 ng

RT: 11.016 min Scan# 1517

Delta R.T. -0.006 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23


Tgt Ion:284 Resp: 89927


Ion Ratio Lower Upper

284 100

142 35.1 26.9 40.3

249 32.2 25.3 37.9







#69
Atrazine
Concen: 10.819 ng
RT: 11.104 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23


Instrument : BNA_F
ClientSampleId : SSTDICC010


Tgt Ion:200 Resp: 64267
Ion Ratio Lower Upper
200 100
173 27.4 7.7 47.7
215 46.8 27.7 67.7



#70
Pentachlorophenol
Concen: 9.775 ng
RT: 11.210 min Scan# 1550
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:266 Resp: 50361
Ion Ratio Lower Upper
266 100
268 63.2 50.4 75.6
264 62.8 50.6 75.8







#71
 Phenanthrene
 Concen: 10.958 ng
 RT: 11.433 min Scan# 1
 Delta R.T. -0.006 min
 Lab File: BF139846.D
 Acq: 18 Oct 2024 11:23


Instrument : BNA_F
 ClientSampleId : SSTDICC010


Tgt Ion:178 Resp: 379233
 Ion Ratio Lower Upper
 178 100
 176 19.4 15.8 23.6
 179 15.4 12.6 18.8



#72
 Anthracene
 Concen: 10.999 ng
 RT: 11.486 min Scan# 1597
 Delta R.T. -0.006 min
 Lab File: BF139846.D
 Acq: 18 Oct 2024 11:23



Tgt Ion:178 Resp: 371391
 Ion Ratio Lower Upper
 178 100
 176 18.7 15.3 22.9
 179 15.0 12.4 18.6






#73
Carbazole
Concen: 11.249 ng
RT: 11.639 min Scan# 1
Instrument : BNA_F
Delta R.T. 0.000 min
Lab File: BF139846.D
ClientSampleId : SSTDICC010
Acq: 18 Oct 2024 11:23


Tgt Ion:167 Resp: 353253
Ion Ratio Lower Upper
167 100
166 21.0 17.0 25.6
139 12.8 10.8 16.2



#74
Di-n-butylphthalate
Concen: 10.818 ng
RT: 11.975 min Scan# 1680
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:149 Resp: 392482
Ion Ratio Lower Upper
149 100
150 9.2 7.5 11.3
104 4.9 4.1 6.1







#75
Fluoranthene
Concen: 11.602 ng
RT: 12.622 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23


Instrument : BNA_F
ClientSampleId : SSTDICC010


Tgt Ion:202 Resp: 405919
Ion Ratio Lower Upper
202 100
101 12.5 0.0 31.9
203 17.2 0.0 37.5



#76
Chrysene-d₁₂
Concen: 20.000 ng
RT: 14.051 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23



Tgt Ion:240 Resp: 452443
Ion Ratio Lower Upper
240 100
120 11.4 9.4 14.2
236 26.0 20.9 31.3






#77
Benzidine
Concen: 14.168 ng
RT: 12.745 min Scan# 1
Instrument : BNA_F
Delta R.T. 0.000 min
Lab File: BF139846.D
ClientSampleId : SSTDICC010
Acq: 18 Oct 2024 11:23


Tgt Ion:184 Resp: 104253
Ion Ratio Lower Upper
184 100
185 15.0 12.0 18.0
183 11.4 9.4 14.2



#78
Pyrene
Concen: 10.444 ng
RT: 12.851 min Scan# 1829
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:202 Resp: 413606
Ion Ratio Lower Upper
202 100
200 21.0 17.2 25.8
203 17.4 14.2 21.4







#79
Terphenyl-d14
Concen: 21.745 ng
RT: 12.998 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23


Instrument : BNA_F
ClientSampleId : SSTDICC010




Tgt Ion:244 Resp: 603787
Ion Ratio Lower Upper
244 100
212 6.8 5.7 8.5
122 10.6 8.6 13.0



#80
Butylbenzylphthalate
Concen: 9.774 ng
RT: 13.474 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:149 Resp: 116053
Ion Ratio Lower Upper
149 100
91 74.4 59.4 89.2
206 20.1 16.4 24.6





#81

Benzo(a)anthracene

Concen: 10.410 ng

RT: 14.039 min Scan# 2

Instrument: BNA_F

Delta R.T. 0.000 min

Lab File: BF139846.D ClientSampleId :

Acq: 18 Oct 2024 11:23 SSTDICC010




Tgt Ion:228 Resp: 306594

Ion Ratio Lower Upper

228 100

226 26.7 21.6 32.4

229 19.5 16.1 24.1



#82

3,3'-Dichlorobenzidine

Concen: 9.806 ng

RT: 14.004 min Scan# 2025

Delta R.T. 0.000 min

Lab File: BF139846.D

Acq: 18 Oct 2024 11:23


Tgt Ion:252 Resp: 84205


Ion Ratio Lower Upper

252 100

254 64.7 50.8 76.2

126 11.3 8.9 13.3







#83
Chrysene
Concen: 10.423 ng
RT: 14.074 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23


Instrument : BNA_F
ClientSampleId : SSTDICC010


Tgt Ion:228 Resp: 281467
Ion Ratio Lower Upper
228 100
226 29.4 24.1 36.1
229 19.6 15.8 23.6



#84
Bis(2-ethylhexyl)phthalate
Concen: 9.153 ng
RT: 14.033 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:149 Resp: 121932
Ion Ratio Lower Upper
149 100
167 26.0 21.4 32.0
279 3.8 3.0 4.4







#85
 Di-n-octyl phthalate
 Concen: 7.339 ng
 RT: 14.651 min Scan# 2
 Delta R.T. -0.000 min
 Lab File: BF139846.D
 Acq: 18 Oct 2024 11:23


Instrument : BNA_F
 ClientSampleId : SSTDICC010


Tgt Ion:149 Resp: 177824
 Ion Ratio Lower Upper
 149 100
 167 1.4 1.2 1.8
 43 11.8 9.8 14.6



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.527 min Scan# 2284
 Delta R.T. 0.000 min
 Lab File: BF139846.D
 Acq: 18 Oct 2024 11:23

Tgt Ion:264 Resp: 375863
 Ion Ratio Lower Upper
 264 100
 260 23.9 19.4 29.2
 265 21.3 17.4 26.0







#87
Indeno(1,2,3-cd)pyrene
Concen: 9.798 ng
RT: 17.010 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23


Instrument: BNA_F
ClientSampleId: SSTDICC010


Tgt Ion:276 Resp: 236914
Ion Ratio Lower Upper
276 100
138 25.8 20.4 30.6
277 25.0 20.2 30.4



#88
Benzo(b)fluoranthene
Concen: 10.176 ng
RT: 15.092 min Scan# 2210
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:252 Resp: 232988
Ion Ratio Lower Upper
252 100
253 21.4 17.6 26.4
125 10.3 7.9 11.9







#89
 Benzo(k)fluoranthene
 Concen: 11.203 ng
 RT: 15.121 min Scan# 2
 Delta R.T. -0.006 min
 Lab File: BF139846.D
 Acq: 18 Oct 2024 11:23


Instrument : BNA_F
 ClientSampleId : SSTDICC010


Tgt Ion:252 Resp: 221219
 Ion Ratio Lower Upper
 252 100
 253 21.5 17.6 26.4
 125 10.2 7.9 11.9



#90
 Benzo(a)pyrene
 Concen: 10.217 ng
 RT: 15.463 min Scan# 2273
 Delta R.T. -0.006 min
 Lab File: BF139846.D
 Acq: 18 Oct 2024 11:23

Tgt Ion:252 Resp: 192493
 Ion Ratio Lower Upper
 252 100
 253 21.7 17.3 25.9
 125 11.3 8.7 13.1







#91
 Dibenzo(a,h)anthracene
 Concen: 9.912 ng
 RT: 17.033 min Scan# 2
 Delta R.T. -0.012 min
 Lab File: BF139846.D
 Acq: 18 Oct 2024 11:23


Instrument : BNA_F
ClientSampleId : SSTDICC010

Tgt Ion:278 Resp: 199946
 Ion Ratio Lower Upper
 278 100
 139 16.0 13.1 19.7
 279 23.2 19.0 28.4



#92
 Benzo(g,h,i)perylene
 Concen: 9.758 ng
 RT: 17.457 min Scan# 2612
 Delta R.T. -0.023 min
 Lab File: BF139846.D
 Acq: 18 Oct 2024 11:23

Tgt Ion:276 Resp: 196609
 Ion Ratio Lower Upper
 276 100
 277 23.3 19.0 28.6
 138 21.4 17.4 26.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139847.D
 Acq On : 18 Oct 2024 11:52
 Operator : RC/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC020

Quant Time: Oct 18 14:18:00 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 14:11:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.893	152	192582	20.000	ng	0.00
21) Naphthalene-d8	8.175	136	735124	20.000	ng	0.00
39) Acenaphthene-d10	9.928	164	413023	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	724973	20.000	ng	0.00
76) Chrysene-d12	14.051	240	402251	20.000	ng	0.00
86) Perylene-d12	15.527	264	371811	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.498	112	512648	41.673	ng	0.00
7) Phenol-d6	6.504	99	658086	41.304	ng	0.00
23) Nitrobenzene-d5	7.451	82	546550	41.206	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	159165	41.199	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1057523	42.316	ng	0.00
79) Terphenyl-d14	12.998	244	1078101	43.673	ng	0.00
Target Compounds						
				Qvalue		
2) 1,4-Dioxane	2.681	88	116188	20.257	ng	98
3) Pyridine	3.440	79	293019	20.611	ng	98
4) n-Nitrosodimethylamine	3.381	42	153824	20.138	ng	99
6) Aniline	6.551	93	311561	20.982	ng	99
8) 2-Chlorophenol	6.669	128	261502	20.794	ng	98
9) Benzaldehyde	6.440	77	200586	22.379	ng	98
10) Phenol	6.516	94	338933	20.269	ng	98
11) bis(2-Chloroethyl)ether	6.622	93	261690	20.612	ng	99
12) 1,3-Dichlorobenzene	6.828	146	297432	20.603	ng	99
13) 1,4-Dichlorobenzene	6.910	146	300107	20.808	ng	98
14) 1,2-Dichlorobenzene	7.063	146	284645	21.146	ng	99
15) Benzyl Alcohol	7.022	79	242106	20.715	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.163	45	453174	20.933	ng	99
17) 2-Methylphenol	7.128	107	218455	20.371	ng	99
18) Hexachloroethane	7.404	117	105775	20.566	ng	98
19) n-Nitroso-di-n-propyla...	7.293	70	197390	20.427	ng	97
20) 3+4-Methylphenols	7.281	107	292795	21.346	ng	96
22) Acetophenone	7.293	105	379127	21.056	ng	99
24) Nitrobenzene	7.469	77	299625	20.604	ng	99
25) Isophorone	7.704	82	516119	20.502	ng	100
26) 2-Nitrophenol	7.787	139	110497	20.141	ng	97
27) 2,4-Dimethylphenol	7.816	122	187930	20.544	ng	98
28) bis(2-Chloroethoxy)met...	7.916	93	317497	20.816	ng	99
29) 2,4-Dichlorophenol	8.016	162	215409	20.673	ng	98
30) 1,2,4-Trichlorobenzene	8.110	180	238559	20.691	ng	100
31) Naphthalene	8.192	128	801971	21.153	ng	99
32) Benzoic acid	7.892	122	152028	19.054	ng	99
33) 4-Chloroaniline	8.239	127	270435	20.919	ng	98
34) Hexachlorobutadiene	8.310	225	149527	20.641	ng	100
35) Caprolactam	8.587	113	67624	20.411	ng	100
36) 4-Chloro-3-methylphenol	8.704	107	238461	20.668	ng	100
37) 2-Methylnaphthalene	8.881	142	489343	21.075	ng	100
38) 1-Methylnaphthalene	8.981	142	481797	21.149	ng	100
40) 1,2,4,5-Tetrachloroben...	9.045	216	232025	20.823	ng	99
41) Hexachlorocyclopentadiene	9.039	237	81579	21.041	ng	99
43) 2,4,6-Trichlorophenol	9.157	196	165112	20.930	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139847.D
 Acq On : 18 Oct 2024 11:52
 Operator : RC/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC020

Quant Time: Oct 18 14:18:00 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 14:11:20 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.192	196	164517	20.213	ng	98
46) 1,1'-Biphenyl	9.345	154	612342	21.297	ng	99
47) 2-Chloronaphthalene	9.375	162	480916	20.767	ng	99
48) 2-Nitroaniline	9.463	65	146002	20.614	ng	98
49) Acenaphthylene	9.786	152	709161	21.182	ng	99
50) Dimethylphthalate	9.645	163	529745	20.592	ng	100
51) 2,6-Dinitrotoluene	9.704	165	114726	20.596	ng	98
52) Acenaphthene	9.963	154	449770	20.768	ng	99
53) 3-Nitroaniline	9.869	138	122186	21.271	ng	99
54) 2,4-Dinitrophenol	9.975	184	31801	19.148	ng	# 67
55) Dibenzofuran	10.134	168	652249	20.991	ng	98
56) 4-Nitrophenol	10.010	139	93060	21.057	ng	95
57) 2,4-Dinitrotoluene	10.104	165	139381	20.348	ng	98
58) Fluorene	10.475	166	496012	20.958	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.245	232	134842	21.133	ng	98
60) Diethylphthalate	10.345	149	523458	20.723	ng	100
61) 4-Chlorophenyl-phenyle...	10.469	204	254950	21.331	ng	97
62) 4-Nitroaniline	10.481	138	111683	20.586	ng	98
63) Azobenzene	10.628	77	559837	20.906	ng	99
65) 4,6-Dinitro-2-methylph...	10.510	198	52180	17.645	ng	100
66) n-Nitrosodiphenylamine	10.581	169	449925	20.736	ng	99
67) 4-Bromophenyl-phenylether	10.957	248	153196	20.512	ng	98
68) Hexachlorobenzene	11.022	284	172091	20.488	ng	99
69) Atrazine	11.104	200	113089	19.240	ng	99
70) Pentachlorophenol	11.210	266	107595	21.106	ng	99
71) Phenanthrene	11.439	178	720363	21.036	ng	100
72) Anthracene	11.486	178	704417	21.083	ng	99
73) Carbazole	11.639	167	669328	21.540	ng	100
74) Di-n-butylphthalate	11.975	149	769779	21.442	ng	99
75) Fluoranthene	12.622	202	751337	21.703	ng	99
77) Benzidine	12.745	184	118035	18.042	ng	100
78) Pyrene	12.851	202	764077	21.700	ng	100
80) Butylbenzylphthalate	13.474	149	215683	20.432	ng	98
81) Benzo(a)anthracene	14.039	228	534525	20.413	ng	100
82) 3,3'-Dichlorobenzidine	14.004	252	156770	20.534	ng	100
83) Chrysene	14.074	228	496447	20.678	ng	99
84) Bis(2-ethylhexyl)phtha...	14.033	149	232200	19.606	ng	99
85) Di-n-octyl phthalate	14.651	149	373952	17.359	ng	100
87) Indeno(1,2,3-cd)pyrene	17.015	276	485858	20.312	ng	100
88) Benzo(b)fluoranthene	15.092	252	490566	21.659	ng	100
89) Benzo(k)fluoranthene	15.121	252	368835	18.883	ng	99
90) Benzo(a)pyrene	15.463	252	378617	20.316	ng	100
91) Dibenzo(a,h)anthracene	17.039	278	410285	20.561	ng	99
92) Benzo(g,h,i)perylene	17.468	276	405140	20.326	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139847.D
 Acq On : 18 Oct 2024 11:52
 Operator : RC/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC020

Quant Time: Oct 18 14:18:00 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 14:11:20 2024
 Response via : Initial Calibration







#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.893 min Scan# 8
 Delta R.T. 0.001 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52


Instrument : BNA_F
 ClientSampleId : SSTDICC020


Tgt Ion:152 Resp: 192582
 Ion Ratio Lower Upper
 152 100
 150 156.0 130.2 195.2
 115 62.6 51.4 77.2



#2
 1,4-Dioxane
 Concen: 20.257 ng
 RT: 2.681 min Scan# 100
 Delta R.T. -0.006 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52

Tgt Ion: 88 Resp: 116188
 Ion Ratio Lower Upper
 88 100
 58 88.9 73.3 109.9
 43 33.6 27.1 40.7







#3
 Pyridine
 Concen: 20.611 ng
 RT: 3.440 min Scan# 2
 Delta R.T. -0.006 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52


Instrument : BNA_F
 ClientSampleId : SSTDICC020


Tgt Ion: 79 Resp: 293019
 Ion Ratio Lower Upper
 79 100
 52 87.3 68.5 102.7
 51 41.6 32.3 48.5



#4
 n-Nitrosodimethylamine
 Concen: 20.138 ng
 RT: 3.381 min Scan# 219
 Delta R.T. -0.017 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52


Tgt Ion: 42 Resp: 153824
 Ion Ratio Lower Upper
 42 100
 74 118.8 96.2 144.4
 44 8.0 6.2 9.2







#5
2-Fluorophenol
Concen: 41.673 ng
RT: 5.498 min Scan# 5
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52


Instrument : BNA_F
ClientSampleId : SSTDICC020





Tgt Ion:112 Resp: 512648
Ion Ratio Lower Upper
112 100
64 65.1 53.0 79.6
63 33.1 27.0 40.4




#6
Aniline
Concen: 20.982 ng
RT: 6.551 min Scan# 758
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52



Tgt Ion: 93 Resp: 311561
Ion Ratio Lower Upper
93 100
66 42.0 34.0 51.0
65 21.2 17.4 26.0







#7
 Phenol-d6
 Concen: 41.304 ng
 RT: 6.504 min Scan# 7
 Delta R.T. -0.006 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52


Instrument : BNA_F
 ClientSampleId : SSTDICC020


Tgt Ion: 99 Resp: 658086
 Ion Ratio Lower Upper
 99 100
 42 20.5 16.7 25.1
 71 33.8 27.7 41.5



#8
 2-Chlorophenol
 Concen: 20.794 ng
 RT: 6.669 min Scan# 778
 Delta R.T. -0.006 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52

Tgt Ion:128 Resp: 261502
 Ion Ratio Lower Upper
 128 100
 130 32.8 12.3 52.3
 64 56.8 34.6 74.6







#9
 Benzaldehyde
 Concen: 22.379 ng
 RT: 6.440 min Scan# 7
 Delta R.T. 0.001 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52


Instrument : BNA_F
 ClientSampleId : SSTDICC020


Tgt Ion: 77 Resp: 200586
 Ion Ratio Lower Upper
 77 100
 105 97.1 75.5 115.5
 106 94.0 71.7 111.7




#10
 Phenol
 Concen: 20.269 ng
 RT: 6.516 min Scan# 752
 Delta R.T. -0.006 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52

Tgt Ion: 94 Resp: 338933
 Ion Ratio Lower Upper
 94 100
 65 30.3 11.1 51.1
 66 40.8 22.7 62.7








#11
bis(2-Chloroethyl)ether
Concen: 20.612 ng
RT: 6.622 min Scan# 7
Instrument : BNA_F
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52
ClientSampleId : SSTDICC020




Tgt Ion: 93 Resp: 261690
Ion Ratio Lower Upper
93 100
63 86.3 65.0 105.0
95 32.9 12.4 52.4






#12
1,3-Dichlorobenzene
Concen: 20.603 ng
RT: 6.828 min Scan# 805
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52



Tgt Ion:146 Resp: 297432
Ion Ratio Lower Upper
146 100
148 63.8 51.0 76.4
75 33.9 26.4 39.6





#13

1,4-Dichlorobenzene

Concen: 20.808 ng

RT: 6.910 min Scan# 8

Delta R.T. 0.000 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52

Instrument : BNA_F

ClientSampleId : SSTDICC020





Tgt Ion:146 Resp: 300107

Ion Ratio Lower Upper

146 100

148 63.1 51.8 77.8

111 42.1 34.6 51.8



#14

1,2-Dichlorobenzene

Concen: 21.146 ng

RT: 7.063 min Scan# 845

Delta R.T. 0.000 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52


Tgt Ion:146 Resp: 284645







Ion Ratio Lower Upper

146 100

148 64.2 51.4 77.2

111 44.2 36.7 55.1






#15

Benzyl Alcohol
Concen: 20.715 ngRT: 7.022 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion: 79 Resp: 242106
Ion Ratio Lower Upper
79 100
108 77.2 61.4 92.0
77 62.8 49.9 74.9






#16

2,2'-oxybis(1-Chloropropane)
Concen: 20.933 ng
RT: 7.163 min Scan# 862
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion: 45 Resp: 453174
Ion Ratio Lower Upper
45 100
77 11.1 0.0 31.5
79 8.5 0.0 29.1





#17

2-Methylphenol

Concen: 20.371 ng

RT: 7.128 min Scan# 8

Delta R.T. -0.006 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52

Instrument:

BNA_F

ClientSampleId :

SSTDICC020

Tgt Ion:107 Resp: 218455

Ion Ratio Lower Upper

107 100





108 115.0 91.3 136.9

77 46.6 37.1 55.7

79 45.5 37.0 55.6

Abundance

7.128



#18

Hexachloroethane

Concen: 20.566 ng

RT: 7.404 min Scan# 903

Delta R.T. 0.000 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52

Tgt Ion:117 Resp: 105775

Ion Ratio Lower Upper


117 100


119 97.3 77.0 115.4

201 98.4 76.2 114.4

Abundance

7.404







#19
n-Nitroso-di-n-propylamine
Concen: 20.427 ng
RT: 7.293 min Scan# 8
Instrument : BNA_F
Delta R.T. -0.012 min
Lab File: BF139847.D
ClientSampleId : SSTDICC020
Acq: 18 Oct 2024 11:52

Tgt Ion: 70 Resp: 197390

Ion Ratio Lower Upper

	70	100
42	67.2	51.5
101	9.2	7.5
130	19.1	15.5
		23.3





#20
3+4-Methylphenols
Concen: 21.346 ng
RT: 7.281 min Scan# 882
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:107 Resp: 292795

Ion Ratio Lower Upper

	107	100
108	89.6	69.8
77	61.4	50.5
79	29.4	10.2
		50.2





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.175 min Scan# 1

Instrument:

BNA_F

Delta R.T. 0.000 min

Lab File: BF139847.D

ClientSampleId :

Acq: 18 Oct 2024 11:52

SSTDICC020

Tgt Ion:136 Resp: 735124

Ion Ratio Lower Upper

136 100

137 10.9 8.6 12.8

54 10.3 8.4 12.6

68 6.5 5.1 7.7

Abundance

500000

400000


300000

200000

100000

0



Time--> 8.10 8.15 8.20



#23
 Nitrobenzene-d5
 Concen: 41.206 ng
 RT: 7.451 min Scan# 911
 Delta R.T. 0.000 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52


Instrument : BNA_F
 ClientSampleId : SSTDICC020




Tgt Ion: 82 Resp: 546550
 Ion Ratio Lower Upper
 82 100
 128 42.3 33.4 50.0
 54 59.2 47.8 71.8



#24
 Nitrobenzene
 Concen: 20.604 ng
 RT: 7.469 min Scan# 914
 Delta R.T. -0.006 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52

Tgt Ion: 77 Resp: 299625
 Ion Ratio Lower Upper
 77 100
 123 44.2 35.7 53.5
 65 14.7 11.5 17.3









#25
 Isophorone
 Concen: 20.502 ng
 RT: 7.704 min Scan# 9
 Delta R.T. -0.006 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52


Instrument: BNA_F
ClientSampleId: SSTDICC020




Tgt Ion: 82 Resp: 516119
 Ion Ratio Lower Upper
 82 100
 95 7.1 5.5 8.3
 138 16.3 13.2 19.8



#26
 2-Nitrophenol
 Concen: 20.141 ng
 RT: 7.787 min Scan# 968
 Delta R.T. 0.000 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52

Tgt Ion:139 Resp: 110497
 Ion Ratio Lower Upper
 139 100
 109 30.0 25.1 37.7
 65 56.5 47.5 71.3





#27

2,4-Dimethylphenol

Concen: 20.544 ng

RT: 7.816 min Scan# 9

Delta R.T. 0.000 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC020





Tgt Ion:122 Resp: 187930

Ion Ratio Lower Upper

122 100

107 112.6 91.9 137.9

121 57.1 46.2 69.4



#28

bis(2-Chloroethoxy)methane

Concen: 20.816 ng

RT: 7.916 min Scan# 990

Delta R.T. -0.006 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52


Tgt Ion: 93 Resp: 317497




Ion Ratio Lower Upper

93 100

95 32.9 26.0 39.0

123 11.1 9.8 14.8





#29

2,4-Dichlorophenol

Concen: 20.673 ng

RT: 8.016 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC020





Tgt Ion:162 Resp: 215409

Ion Ratio Lower Upper

162 100

164 62.7 43.9 83.9

98 41.1 19.9 59.9



#30

1,2,4-Trichlorobenzene

Concen: 20.691 ng

RT: 8.110 min Scan# 1023

Delta R.T. -0.006 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52


Tgt Ion:180 Resp: 238559




Ion Ratio Lower Upper

180 100

182 100.6 80.5 120.7

145 32.5 25.4 38.0





#31

Naphthalene
Concen: 21.153 ng
RT: 8.192 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :

BNA_F





ClientSampleId :

SSTDICC020

Tgt Ion:128 Resp: 801971

Ion Ratio Lower Upper

128	100		
129	11.3	8.9	13.3
127	13.1	10.6	16.0




#32


Benzoic acid
Concen: 19.054 ng
RT: 7.892 min Scan# 986
Delta R.T. -0.030 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:122 Resp: 152028

Ion Ratio Lower Upper

122	100		
105	127.9	107.6	147.6
77	98.2	80.2	120.2







#33
4-Chloroaniline
Concen: 20.919 ng
RT: 8.239 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument : BNA_F
ClientSampleId : SSTDICC020

Tgt Ion:127 Resp: 270435
Ion Ratio Lower Upper

127	100
129	33.0
65	34.3
92	20.5


Abundance






#34
Hexachlorobutadiene
Concen: 20.641 ng
RT: 8.310 min Scan# 1057
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:225 Resp: 149527
Ion Ratio Lower Upper

225	100
223	63.4
227	64.6









#35
 Caprolactam
 Concen: 20.411 ng
 RT: 8.587 min Scan# 1
 Delta R.T. -0.023 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52


Instrument : BNA_F
 ClientSampleId : SSTDICC020




Tgt Ion:113 Resp: 67624
 Ion Ratio Lower Upper
 113 100
 55 209.6 189.2 229.2
 56 152.6 132.5 172.5



#36
 4-Chloro-3-methylphenol
 Concen: 20.668 ng
 RT: 8.704 min Scan# 1124
 Delta R.T. -0.006 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52

Tgt Ion:107 Resp: 238461
 Ion Ratio Lower Upper
 107 100
 144 23.9 19.1 28.7
 142 71.9 57.2 85.8





#37

2-Methylnaphthalene

Concen: 21.075 ng

RT: 8.881 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52

Instrument :

BNA_F

ClientSampleId :





SSTDICC020

Tgt Ion:142 Resp: 489343

Ion Ratio Lower Upper

142 100

141 88.4 70.8 106.2



#38

1-Methylnaphthalene

Concen: 21.149 ng

RT: 8.981 min Scan# 1171

Delta R.T. -0.006 min

Lab File: BF139847.D


Acq: 18 Oct 2024 11:52




Tgt Ion:142 Resp: 481797

Ion Ratio Lower Upper

142 100

141 91.7 73.5 110.3





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.928 min Scan# 1

Delta R.T. 0.001 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC020

Tgt Ion:164 Resp: 413023





Ion Ratio Lower Upper

164 100

162 99.7 81.0 121.4

160 43.8 35.4 53.0

Abundance



#40

1,2,4,5-Tetrachlorobenzene

Concen: 20.823 ng

RT: 9.045 min Scan# 1182

Delta R.T. -0.006 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52

Tgt Ion:216 Resp: 232025

Ion Ratio Lower Upper


216 100


214 78.2 63.4 95.0

179 20.8 17.2 25.8

108 18.7 15.6 23.4

Abundance





#41

Hexachlorocyclopentadiene

Concen: 21.041 ng

RT: 9.039 min Scan# 1

Instrument:

Delta R.T. 0.000 min

BNA_F

Lab File: BF139847.D

ClientSampleId :

Acq: 18 Oct 2024 11:52

SSTDICC020



Tgt Ion:237 Resp: 81579

Ion Ratio Lower Upper

237 100

235 62.3 42.9 82.9

272 12.7 0.0 32.3



#42

2,4,6-Tribromophenol

Concen: 41.199 ng

RT: 10.710 min Scan# 1465

Delta R.T. -0.006 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52


Tgt Ion:330 Resp: 159165




Ion Ratio Lower Upper

330 100

332 96.4 78.1 117.1

141 32.9 26.6 39.8





#43

2,4,6-Trichlorophenol

Concen: 20.930 ng

RT: 9.157 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC020





Tgt Ion:196 Resp: 165112

Ion Ratio Lower Upper

196 100

198 95.4 76.2 114.2

200 30.8 24.0 36.0



#44

2,4,5-Trichlorophenol

Concen: 20.213 ng

RT: 9.192 min Scan# 1207

Delta R.T. 0.000 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52

Tgt Ion:196 Resp: 164517


Ion Ratio Lower Upper




196 100

198 96.2 75.8 113.6

97 49.9 41.8 62.6

132 28.4 22.9 34.3









#45
2-Fluorobiphenyl
Concen: 42.316 ng
RT: 9.245 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52


Instrument : BNA_F
ClientSampleId : SSTDICC020




Tgt Ion:172 Resp: 1057523
Ion Ratio Lower Upper
172 100
171 35.0 28.6 43.0
170 23.6 19.1 28.7



#46
1,1'-Biphenyl
Concen: 21.297 ng
RT: 9.345 min Scan# 1233
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:154 Resp: 612342
Ion Ratio Lower Upper
154 100
153 39.6 20.3 60.3
76 14.4 0.0 33.9





#47

2-Chloronaphthalene

Concen: 20.767 ng

RT: 9.375 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC020




Tgt Ion:162 Resp: 480916

Ion Ratio Lower Upper

162 100

127 41.7 33.8 50.6

164 32.8 26.6 39.8



#48

2-Nitroaniline

Concen: 20.614 ng

RT: 9.463 min Scan# 1253

Delta R.T. 0.000 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52


Tgt Ion: 65 Resp: 146002




Ion Ratio Lower Upper

65 100

92 64.0 50.9 76.3

138 94.4 73.4 110.0





#49

Acenaphthylene

Concen: 21.182 ng

RT: 9.786 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC020





Tgt Ion:152 Resp: 709161

Ion Ratio Lower Upper

152 100

151 20.0 15.9 23.9

153 13.0 10.7 16.1



#50

Dimethylphthalate

Concen: 20.592 ng

RT: 9.645 min Scan# 1284

Delta R.T. -0.006 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52


Tgt Ion:163 Resp: 529745




Ion Ratio Lower Upper

163 100

194 4.1 3.4 5.0

164 10.1 8.2 12.2





#51

2,6-Dinitrotoluene

Concen: 20.596 ng

RT: 9.704 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC020

Tgt Ion:165 Resp: 114726

Ion Ratio Lower Upper

165 100

63 64.6





89 58.4

50.6

75.8

47.5

71.3



#52

Acenaphthene

Concen: 20.768 ng

RT: 9.963 min Scan# 1338

Delta R.T. 0.000 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52

Tgt Ion:154 Resp: 449770

Ion Ratio Lower Upper

154 100

153 110.4


152 52.5




89.8

134.6

42.4

63.6









#53
3-Nitroaniline
Concen: 21.271 ng
RT: 9.869 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52


Instrument: BNA_F
ClientSampleId: SSTDICC020


Tgt Ion:138 Resp: 122186
Ion Ratio Lower Upper
138 100
108 13.5 10.2 15.2
92 121.9 98.3 147.5



#54
2,4-Dinitrophenol
Concen: 19.148 ng
RT: 9.975 min Scan# 1340
Delta R.T. 0.001 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52



Tgt Ion:184 Resp: 31801
Ion Ratio Lower Upper
184 100
63 91.3 84.7 127.1
154 331.4 338.2 507.2#






#55
Dibenzofuran
Concen: 20.991 ng
RT: 10.134 min Scan# 1
Instrument: BNA_F
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52
ClientSampleId : SSTDICC020


Tgt Ion:168 Resp: 652249
Ion Ratio Lower Upper
168 100
139 38.0 31.8 47.6
169 14.2 11.6 17.4



#56
4-Nitrophenol
Concen: 21.057 ng
RT: 10.010 min Scan# 1346
Delta R.T. -0.012 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52



Tgt Ion:139 Resp: 93060
Ion Ratio Lower Upper
139 100
109 74.2 53.6 93.6
65 115.7 87.8 127.8






#57
2,4-Dinitrotoluene
Concen: 20.348 ng
RT: 10.104 min Scan# 1
Instrument : BNA_F
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52
ClientSampleId : SSTDICC020


Tgt Ion:165 Resp: 139381
Ion Ratio Lower Upper
165 100
63 50.6 40.1 60.1
89 77.9 60.4 90.6
182 2.8 2.0 3.0



#58
Fluorene
Concen: 20.958 ng
RT: 10.475 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:166 Resp: 496012
Ion Ratio Lower Upper
166 100
165 98.3 78.5 117.7
167 13.5 10.9 16.3







#59
2,3,4,6-Tetrachlorophenol
Concen: 21.133 ng
RT: 10.245 min Scan# 1
Instrument : BNA_F
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52
ClientSampleId : SSTDICC020

Tgt Ion:232 Resp: 134842
Ion Ratio Lower Upper

232	100
131	45.1
130	2.5
166	29.2


Abundance




#60
Diethylphthalate
Concen: 20.723 ng
RT: 10.345 min Scan# 1403
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:149 Resp: 523458
Ion Ratio Lower Upper

149	100
177	22.5
150	12.7





#61

4-Chlorophenyl-phenylether

Concen: 21.331 ng

RT: 10.469 min Scan# 1424

Delta R.T. 0.000 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC020



Tgt Ion:204 Resp: 254950

Ion Ratio Lower Upper

204 100

206 32.3 26.3 39.5

141 58.3 48.6 73.0



#62

4-Nitroaniline

Concen: 20.586 ng

RT: 10.481 min Scan# 1426

Delta R.T. -0.006 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52


Tgt Ion:138 Resp: 111683


Ion Ratio Lower Upper

138 100

92 53.4 35.7 75.7

108 76.5 57.5 97.5





#63

Azobenzene

Concen: 20.906 ng

RT: 10.628 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC020

Tgt Ion: 77 Resp: 559837

Ion Ratio Lower Upper

77 100

182 24.4 3.3 43.3

105 21.3 1.2 41.2

51 33.0 13.3 53.3

Abundance

400000

300000


200000

100000

0

Time-->

10.628



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.410 min Scan# 1584

Delta R.T. -0.006 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52

Tgt Ion:188 Resp: 724973

Ion Ratio Lower Upper

188 100

94 10.5 7.9 11.9

80 11.7 9.0 13.4

Abundance

500000

400000

300000


200000

100000

0

Time-->

11.410



#65

4,6-Dinitro-2-methylphenol

Concen: 17.645 ng

RT: 10.510 min Scan# 1431

Delta R.T. -0.006 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC020


Tgt Ion:198 Resp: 52180

Ion Ratio Lower Upper

198 100

51 58.5 38.2 78.2

105 48.3 27.9 67.9



#66

n-Nitrosodiphenylamine

Concen: 20.736 ng

RT: 10.581 min Scan# 1443

Delta R.T. -0.006 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52




Tgt Ion:169 Resp: 449925




Ion Ratio Lower Upper

169 100

168 63.1 49.8 74.6

167 34.4 28.0 42.0





#67

4-Bromophenyl-phenylether

Concen: 20.512 ng

RT: 10.957 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC020





Tgt Ion:248 Resp: 153196

Ion Ratio Lower Upper

248 100

250 97.4 79.0 118.6

141 74.3 61.7 92.5



#68

Hexachlorobenzene

Concen: 20.488 ng

RT: 11.022 min Scan# 1518

Delta R.T. 0.000 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52


Tgt Ion:284 Resp: 172091


Ion Ratio Lower Upper

284 100

142 32.7 26.9 40.3

249 31.5 25.3 37.9







#69
Atrazine
Concen: 19.240 ng
RT: 11.104 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52


Instrument : BNA_F
ClientSampleId : SSTDICC020


Tgt Ion:200 Resp: 113089
Ion Ratio Lower Upper
200 100
173 27.2 7.7 47.7
215 46.9 27.7 67.7



#70
Pentachlorophenol
Concen: 21.106 ng
RT: 11.210 min Scan# 1550
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:266 Resp: 107595
Ion Ratio Lower Upper
266 100
268 63.4 50.4 75.6
264 62.8 50.6 75.8







#71
 Phenanthrene
 Concen: 21.036 ng
 RT: 11.439 min Scan# 1
 Delta R.T. 0.000 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52


Instrument : BNA_F
 ClientSampleId : SSTDICC020


Tgt Ion:178 Resp: 720363
 Ion Ratio Lower Upper
 178 100
 176 19.5 15.8 23.6
 179 15.5 12.6 18.8



#72
 Anthracene
 Concen: 21.083 ng
 RT: 11.486 min Scan# 1597
 Delta R.T. -0.006 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52



Tgt Ion:178 Resp: 704417
 Ion Ratio Lower Upper
 178 100
 176 18.7 15.3 22.9
 179 15.3 12.4 18.6






#73
Carbazole
Concen: 21.540 ng
RT: 11.639 min Scan# 1
Instrument : BNA_F
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52
ClientSampleId : SSTDICC020


Tgt Ion:167 Resp: 669328
Ion Ratio Lower Upper
167 100
166 21.3 17.0 25.6
139 13.1 10.8 16.2



#74
Di-n-butylphthalate
Concen: 21.442 ng
RT: 11.975 min Scan# 1680
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:149 Resp: 769779
Ion Ratio Lower Upper
149 100
150 9.0 7.5 11.3
104 4.8 4.1 6.1







#75
Fluoranthene
Concen: 21.703 ng
RT: 12.622 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52


Instrument : BNA_F
ClientSampleId : SSTDICC020


Tgt Ion:202 Resp: 751337
Ion Ratio Lower Upper
202 100
101 12.6 0.0 31.9
203 17.7 0.0 37.5



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.051 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:240 Resp: 402251
Ion Ratio Lower Upper
240 100
120 11.3 9.4 14.2
236 26.2 20.9 31.3







#77
Benzidine
Concen: 18.042 ng
RT: 12.745 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52


Instrument : BNA_F
ClientSampleId : SSTDICC020


Tgt Ion:184 Resp: 118035
Ion Ratio Lower Upper
184 100
185 15.0 12.0 18.0
183 11.8 9.4 14.2



#78
Pyrene
Concen: 21.700 ng
RT: 12.851 min Scan# 1829
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:202 Resp: 764077
Ion Ratio Lower Upper
202 100
200 21.5 17.2 25.8
203 17.4 14.2 21.4







#79
Terphenyl-d14
Concen: 43.673 ng
RT: 12.998 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52


Instrument : BNA_F
ClientSampleId : SSTDICC020




Tgt Ion:244 Resp: 1078101
Ion Ratio Lower Upper
244 100
212 6.8 5.7 8.5
122 10.7 8.6 13.0



#80
Butylbenzylphthalate
Concen: 20.432 ng
RT: 13.474 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:149 Resp: 215683
Ion Ratio Lower Upper
149 100
91 76.6 59.4 89.2
206 20.9 16.4 24.6





#81

Benzo(a)anthracene

Concen: 20.413 ng

RT: 14.039 min Scan# 2

Delta R.T. 0.000 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC020





Tgt Ion:228 Resp: 534525

Ion Ratio Lower Upper

228 100

226 26.8 21.6 32.4

229 19.8 16.1 24.1



#82

3,3'-Dichlorobenzidine

Concen: 20.534 ng

RT: 14.004 min Scan# 2025

Delta R.T. 0.000 min

Lab File: BF139847.D

Acq: 18 Oct 2024 11:52


Tgt Ion:252 Resp: 156770


Ion Ratio Lower Upper

252 100

254 63.2 50.8 76.2

126 11.3 8.9 13.3







#83
Chrysene
Concen: 20.678 ng
RT: 14.074 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52


Instrument : BNA_F
ClientSampleId : SSTDICC020


Tgt Ion:228 Resp: 496447
Ion Ratio Lower Upper
228 100
226 29.9 24.1 36.1
229 20.1 15.8 23.6



#84
Bis(2-ethylhexyl)phthalate
Concen: 19.606 ng
RT: 14.033 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:149 Resp: 232200
Ion Ratio Lower Upper
149 100
167 26.3 21.4 32.0
279 4.0 3.0 4.4







#85
 Di-n-octyl phthalate
 Concen: 17.359 ng
 RT: 14.651 min Scan# 2
 Delta R.T. -0.000 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52


Instrument : BNA_F
 ClientSampleId : SSTDICC020


Tgt Ion:149 Resp: 373952
 Ion Ratio Lower Upper
 149 100
 167 1.4 1.2 1.8
 43 12.2 9.8 14.6



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.527 min Scan# 2284
 Delta R.T. 0.000 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52

Tgt Ion:264 Resp: 371811
 Ion Ratio Lower Upper
 264 100
 260 23.6 19.4 29.2
 265 21.1 17.4 26.0







#87
 Indeno(1,2,3-cd)pyrene
 Concen: 20.312 ng
 RT: 17.015 min Scan# 2
 Delta R.T. -0.006 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52


Instrument : BNA_F
 ClientSampleId : SSTDICC020


Tgt Ion:276 Resp: 485858
 Ion Ratio Lower Upper
 276 100
 138 25.9 20.4 30.6
 277 25.2 20.2 30.4



#88
 Benzo(b)fluoranthene
 Concen: 21.659 ng
 RT: 15.092 min Scan# 2210
 Delta R.T. -0.006 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52

Tgt Ion:252 Resp: 490566
 Ion Ratio Lower Upper
 252 100
 253 22.2 17.6 26.4
 125 10.2 7.9 11.9







#89
 Benzo(k)fluoranthene
 Concen: 18.883 ng
 RT: 15.121 min Scan# 2
 Delta R.T. -0.006 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52


Instrument : BNA_F
 ClientSampleId : SSTDICC020


Tgt Ion:252 Resp: 368835
 Ion Ratio Lower Upper
 252 100
 253 22.3 17.6 26.4
 125 10.2 7.9 11.9



#90
 Benzo(a)pyrene
 Concen: 20.316 ng
 RT: 15.463 min Scan# 2273
 Delta R.T. -0.006 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52

Tgt Ion:252 Resp: 378617
 Ion Ratio Lower Upper
 252 100
 253 21.7 17.3 25.9
 125 11.2 8.7 13.1







#91
 Dibenzo(a,h)anthracene
 Concen: 20.561 ng
 RT: 17.039 min Scan# 2
 Delta R.T. -0.006 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52


Instrument : BNA_F
 ClientSampleId : SSTDICC020

Tgt Ion:278 Resp: 410285
 Ion Ratio Lower Upper
 278 100
 139 15.4 13.1 19.7
 279 23.8 19.0 28.4



#92
 Benzo(g,h,i)perylene
 Concen: 20.326 ng
 RT: 17.468 min Scan# 2614
 Delta R.T. -0.012 min
 Lab File: BF139847.D
 Acq: 18 Oct 2024 11:52

Tgt Ion:276 Resp: 405140
 Ion Ratio Lower Upper
 276 100
 277 23.9 19.0 28.6
 138 21.2 17.4 26.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139848.D
 Acq On : 18 Oct 2024 12:20
 Operator : RC/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICCC040

Quant Time: Oct 18 14:18:47 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 14:11:20 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.892	152	181899	20.000	ng	0.00
21) Naphthalene-d8	8.175	136	688880	20.000	ng	0.00
39) Acenaphthene-d10	9.927	164	379753	20.000	ng	0.00
64) Phenanthrene-d10	11.416	188	663459	20.000	ng	0.00
76) Chrysene-d12	14.051	240	345800	20.000	ng	0.00
86) Perylene-d12	15.527	264	365076	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.498	112	929569	80.002	ng	0.00
7) Phenol-d6	6.510	99	1197998	79.607	ng	0.00
23) Nitrobenzene-d5	7.451	82	1021808	82.209	ng	0.00
42) 2,4,6-Tribromophenol	10.716	330	285205	80.292	ng	0.00
45) 2-Fluorobiphenyl	9.251	172	1764495	76.791	ng	0.00
79) Terphenyl-d14	12.998	244	1721125	81.103	ng	0.00
Target Compounds						
				Qvalue		
2) 1,4-Dioxane	2.687	88	214958	39.679	ng	100
3) Pyridine	3.446	79	547145	40.746	ng	100
4) n-Nitrosodimethylamine	3.398	42	293138	40.631	ng	100
6) Aniline	6.551	93	575402	41.026	ng	100
8) 2-Chlorophenol	6.675	128	476614	40.124	ng	100
9) Benzaldehyde	6.439	77	341959	40.393	ng	100
10) Phenol	6.522	94	622748m	39.429	ng	
11) bis(2-Chloroethyl)ether	6.628	93	470765	39.258	ng	100
12) 1,3-Dichlorobenzene	6.834	146	545252	39.988	ng	100
13) 1,4-Dichlorobenzene	6.910	146	541068	39.718	ng	100
14) 1,2-Dichlorobenzene	7.063	146	501857	39.473	ng	100
15) Benzyl Alcohol	7.028	79	441461	39.991	ng	100
16) 2,2'-oxybis(1-Chloropr...	7.169	45	820216	40.113	ng	100
17) 2-Methylphenol	7.133	107	405109	39.996	ng	100
18) Hexachloroethane	7.404	117	196637	40.478	ng	100
19) n-Nitroso-di-n-propyla...	7.304	70	354454	38.835	ng	100
20) 3+4-Methylphenols	7.286	107	515857	39.818	ng	100
22) Acetophenone	7.298	105	663304	39.312	ng	100
24) Nitrobenzene	7.475	77	555115	40.735	ng	100
25) Isophorone	7.710	82	935400	39.651	ng	100
26) 2-Nitrophenol	7.786	139	216916	42.193	ng	100
27) 2,4-Dimethylphenol	7.816	122	341985	39.894	ng	100
28) bis(2-Chloroethoxy)met...	7.922	93	567188	39.683	ng	100
29) 2,4-Dichlorophenol	8.022	162	390330	39.976	ng	100
30) 1,2,4-Trichlorobenzene	8.116	180	434274	40.195	ng	100
31) Naphthalene	8.198	128	1409796	39.681	ng	100
32) Benzoic acid	7.922	122	302443	40.451	ng	100
33) 4-Chloroaniline	8.239	127	483982	39.950	ng	100
34) Hexachlorobutadiene	8.316	225	271315	39.967	ng	100
35) Caprolactam	8.610	113	126585	40.772	ng	100
36) 4-Chloro-3-methylphenol	8.710	107	433727	40.116	ng	100
37) 2-Methylnaphthalene	8.886	142	856111	39.345	ng	100
38) 1-Methylnaphthalene	8.986	142	842638	39.471	ng	100
40) 1,2,4,5-Tetrachloroben...	9.051	216	407511	39.776	ng	100
41) Hexachlorocyclopentadiene	9.039	237	150540	42.229	ng	100
43) 2,4,6-Trichlorophenol	9.157	196	293577	40.474	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139848.D
 Acq On : 18 Oct 2024 12:20
 Operator : RC/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICCC040

Quant Time: Oct 18 14:18:47 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 14:11:20 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.192	196	304213	40.651	ng	100
46) 1,1'-Biphenyl	9.351	154	1047094	39.608	ng	100
47) 2-Chloronaphthalene	9.375	162	843452	39.613	ng	100
48) 2-Nitroaniline	9.463	65	277165	42.562	ng	100
49) Acenaphthylene	9.792	152	1226273	39.837	ng	100
50) Dimethylphthalate	9.651	163	939670	39.726	ng	100
51) 2,6-Dinitrotoluene	9.710	165	212956	41.580	ng	100
52) Acenaphthene	9.963	154	792756	39.811	ng	100
53) 3-Nitroaniline	9.874	138	222098	42.051	ng	100
54) 2,4-Dinitrophenol	9.974	184	76948	39.739	ng	100
55) Dibenzofuran	10.133	168	1128670	39.505	ng	100
56) 4-Nitrophenol	10.022	139	175913	43.292	ng	100
57) 2,4-Dinitrotoluene	10.110	165	270604	42.965	ng	100
58) Fluorene	10.474	166	843048	38.742	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.245	232	235175	40.087	ng	100
60) Diethylphthalate	10.351	149	922384	39.715	ng	100
61) 4-Chlorophenyl-phenyle...	10.469	204	432393	39.347	ng	100
62) 4-Nitroaniline	10.486	138	208926	41.884	ng	100
63) Azobenzene	10.627	77	991989	40.289	ng	100
65) 4,6-Dinitro-2-methylph...	10.516	198	115824	42.799	ng	100
66) n-Nitrosodiphenylamine	10.586	169	789753	39.772	ng	100
67) 4-Bromophenyl-phenylether	10.957	248	268509	39.285	ng	100
68) Hexachlorobenzene	11.021	284	306919	39.927	ng	100
69) Atrazine	11.110	200	232078	43.145	ng	100
70) Pentachlorophenol	11.210	266	201559	43.204	ng	100
71) Phenanthrene	11.439	178	1238155	39.508	ng	100
72) Anthracene	11.492	178	1211718	39.628	ng	100
73) Carbazole	11.639	167	1122566	39.475	ng	100
74) Di-n-butylphthalate	11.974	149	1345878	40.965	ng	100
75) Fluoranthene	12.627	202	1251202	39.493	ng	100
77) Benzidine	12.745	184	252893	44.967	ng	100
78) Pyrene	12.857	202	1248309	41.241	ng	100
80) Butylbenzylphthalate	13.474	149	383831	42.297	ng	100
81) Benzo(a)anthracene	14.039	228	920698	40.901	ng	100
82) 3,3'-Dichlorobenzidine	14.004	252	267769	40.798	ng	100
83) Chrysene	14.080	228	806950	39.098	ng	100
84) Bis(2-ethylhexyl)phtha...	14.033	149	432750	42.504	ng	100
85) Di-n-octyl phthalate	14.651	149	785179	42.399	ng	100
87) Indeno(1,2,3-cd)pyrene	17.021	276	987345	42.039	ng	100
88) Benzo(b)fluoranthene	15.098	252	873168	39.263	ng	100
89) Benzo(k)fluoranthene	15.127	252	778470	40.589	ng	100
90) Benzo(a)pyrene	15.468	252	748507	40.905	ng	100
91) Dibenzo(a,h)anthracene	17.045	278	818044	41.752	ng	100
92) Benzo(g,h,i)perylene	17.480	276	823727	42.090	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139848.D
 Acq On : 18 Oct 2024 12:20
 Operator : RC/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1




Quant Time: Oct 18 14:18:47 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 14:11:20 2024
 Response via : Initial Calibration

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024








#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.892 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#2
1,4-Dioxane
Concen: 39.679 ng
RT: 2.687 min Scan# 101
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion: 88 Resp: 214958
Ion Ratio Lower Upper
88 100
58 91.6 73.3 109.9
43 33.9 27.1 40.7





#3
Pyridine
Concen: 40.746 ng
RT: 3.446 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20




Instrument : BNA_F
ClientSampleId : SSTDICCC040




Tgt Ion: 79 Resp: 547145
Ion Ratio Lower Upper
79 100
52 85.6 68.5 102.7
51 40.4 32.3 48.5

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024






#4
n-Nitrosodimethylamine
Concen: 40.631 ng
RT: 3.398 min Scan# 222
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20



Tgt Ion: 42 Resp: 293138
Ion Ratio Lower Upper
42 100
74 120.3 96.2 144.4
44 7.7 6.2 9.2









#5
2-Fluorophenol
Concen: 80.002 ng
RT: 5.498 min Scan# 5
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Instrument : BNA_F
ClientSampleId : SSTDICCC040


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#6
Aniline
Concen: 41.026 ng
RT: 6.551 min Scan# 758
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion: 93 Resp: 575402
Ion Ratio Lower Upper
93 100
66 42.5 34.0 51.0
65 21.7 17.4 26.0







#7
 Phenol-d6
 Concen: 79.607 ng
 RT: 6.510 min Scan# 7
 Delta R.T. 0.000 min
 Lab File: BF139848.D
 Acq: 18 Oct 2024 12:20

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024



#8
 2-Chlorophenol
 Concen: 40.124 ng
 RT: 6.675 min Scan# 779
 Delta R.T. 0.000 min
 Lab File: BF139848.D
 Acq: 18 Oct 2024 12:20


Tgt Ion:128 Resp: 476614
 Ion Ratio Lower Upper
 128 100
 130 32.3 12.3 52.3
 64 54.6 34.6 74.6





#9
Benzaldehyde
Concen: 40.393 ng
RT: 6.439 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20




Instrument : BNA_F
ClientSampleId : SSTDICCC040




Tgt Ion: 77 Resp: 341959
Ion Ratio Lower Upper
77 100
105 95.5 75.5 115.5
106 91.7 71.7 111.7

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024









#10
Phenol
Concen: 39.429 ng m
RT: 6.522 min Scan# 753
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20



Tgt Ion: 94 Resp: 622748
Ion Ratio Lower Upper
94 100
65 31.1 11.1 51.1
66 42.7 22.7 62.7





#11

bis(2-Chloroethyl)ether

Concen: 39.258 ng

RT: 6.628 min Scan# 7

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20


Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#12

1,3-Dichlorobenzene

Concen: 39.988 ng

RT: 6.834 min Scan# 806


Delta R.T. 0.000 min




Lab File: BF139848.D

Acq: 18 Oct 2024 12:20

Tgt Ion:146 Resp: 545252

Ion Ratio	Lower	Upper
146	100	
148	63.7	51.0
75	33.0	26.4





#13

1,4-Dichlorobenzene

Concen: 39.718 ng

RT: 6.910 min Scan# 819

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20




Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#14

1,2-Dichlorobenzene

Concen: 39.473 ng

RT: 7.063 min Scan# 845

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20


Tgt Ion:146 Resp: 501857




Ion Ratio Lower Upper

146 100

148 64.3 51.4 77.2

111 45.9 36.7 55.1





#15

Benzyl Alcohol

Concen: 39.991 ng

RT: 7.028 min Scan# 8

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20




Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#16

2,2'-oxybis(1-Chloropropane)

Concen: 40.113 ng

RT: 7.169 min Scan# 863

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20


Tgt Ion: 45 Resp: 820216




Ion Ratio Lower Upper

45 100

77 11.5 0.0 31.5

79 9.1 0.0 29.1





#17

2-Methylphenol

Concen: 39.996 ng

RT: 7.133 min Scan# 8

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20




Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#18

Hexachloroethane

Concen: 40.478 ng

RT: 7.404 min Scan# 903

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20


Tgt Ion:117 Resp: 196637


Ion Ratio Lower Upper

117 100

119 96.2 77.0 115.4

201 95.3 76.2 114.4







#19
n-Nitroso-di-n-propylamine
Concen: 38.835 ng
RT: 7.304 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#20
3+4-Methylphenols
Concen: 39.818 ng
RT: 7.286 min Scan# 883
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:107 Resp: 515857
Ion Ratio Lower Upper
107 100
108 89.8 69.8 109.8
77 70.5 50.5 90.5
79 30.2 10.2 50.2





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.175 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20




Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#22

Acetophenone

Concen: 39.312 ng

RT: 7.298 min Scan# 885

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20

Tgt Ion:105 Resp: 663304


Ion Ratio Lower Upper




105 100

71 7.4 5.9 8.9

51 36.8 29.4 44.2

120 21.9 17.5 26.3





#23

Nitrobenzene-d5

Concen: 82.209 ng

RT: 7.451 min Scan# 911

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20




Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#24

Nitrobenzene

Concen: 40.735 ng

RT: 7.475 min Scan# 915

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20


Tgt Ion: 77 Resp: 555115




Ion Ratio Lower Upper

77 100

123 44.6 35.7 53.5

65 14.4 11.5 17.3





#25

Isophorone

Concen: 39.651 ng

RT: 7.710 min Scan# 9

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20




Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#26

2-Nitrophenol

Concen: 42.193 ng

RT: 7.786 min Scan# 968

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20


Tgt Ion:139 Resp: 216916




Ion Ratio Lower Upper

139 100

109 31.4 25.1 37.7

65 59.4 47.5 71.3





#27

2,4-Dimethylphenol

Concen: 39.894 ng

RT: 7.816 min Scan# 9

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20




Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#28

bis(2-Chloroethoxy)methane

Concen: 39.683 ng

RT: 7.922 min Scan# 991

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20


Tgt Ion: 93 Resp: 567188







Ion Ratio Lower Upper

93 100

95 32.5 26.0 39.0

123 12.3 9.8 14.8





#29


2,4-Dichlorophenol
Concen: 39.976 ng

RT: 8.022 min Scan# 1


Delta R.T. 0.000 min




Lab File: BF139848.D

Acq: 18 Oct 2024 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040**Manual Integrations
APPROVED**Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024

#30

1,2,4-Trichlorobenzene
Concen: 40.195 ng
RT: 8.116 min Scan# 1024
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20Tgt Ion:180 Resp: 434274
Ion Ratio Lower Upper
180 100
182 100.6 80.5 120.7
145 31.7 25.4 38.0



#31

Naphthalene

Concen: 39.681 ng

RT: 8.198 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20

Instrument :

BNA_F





ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#32

Benzoic acid

Concen: 40.451 ng

RT: 7.922 min Scan# 991


Delta R.T. -0.000 min



Lab File: BF139848.D

Acq: 18 Oct 2024 12:20

Tgt Ion:122 Resp: 302443

Ion	Ratio	Lower	Upper
122	100		
105	127.6	107.6	147.6
77	100.2	80.2	120.2





#33

4-Chloroaniline

Concen: 39.950 ng

RT: 8.239 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**
Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024

Tgt Ion:127 Resp: 483981

Ion Ratio Lower Upper


127 100

129 31.8 25.4 38.2

65 35.5 28.4 42.6

92 20.8 16.6 25.0

Abundance



#34

Hexachlorobutadiene

Concen: 39.967 ng

RT: 8.316 min Scan# 1058

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20

Tgt Ion:225 Resp: 271315


Ion Ratio Lower Upper


225 100

223 63.3 50.6 76.0


227 64.8 51.8 77.8

Abundance






Abundance Scan 1108 (8.610 min): BF139848.D\data.ms



Abundance Scan 1108 (8.610 min): BF139848.D\data.ms (-1)



#35

Caprolactam

Concen: 40.772 ng

RT: 8.610 min Scan# 1108

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

Tgt Ion:113 Resp: 126589

Ion Ratio Lower Upper

113 100



55 209.2 189.2 229.2

56 152.5 132.5 172.5


Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/21/2024


Supervised By :mohammad ahmed 10/21/2024



Abundance Scan 1125 (8.710 min): BF139848.D\data.ms



Abundance Scan 1125 (8.710 min): BF139848.D\data.ms (-1)



#36

4-Chloro-3-methylphenol

Concen: 40.116 ng

RT: 8.710 min Scan# 1125

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20


Tgt Ion:107 Resp: 433727




Ion Ratio Lower Upper

107 100

144 23.9 19.1 28.7

142 71.5 57.2 85.8





#37

2-Methylnaphthalene

Concen: 39.345 ng

RT: 8.886 min Scan# 1155

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20





Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#38

1-Methylnaphthalene

Concen: 39.471 ng

RT: 8.986 min Scan# 1172

Delta R.T. 0.000 min

Lab File: BF139848.D


Acq: 18 Oct 2024 12:20




Tgt Ion:142 Resp: 842638

Ion Ratio Lower Upper

142 100

141 91.9 73.5 110.3





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.927 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20





Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#40

1,2,4,5-Tetrachlorobenzene


Concen: 39.776 ng



RT: 9.051 min Scan# 1183

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20

 Tgt Ion:216 Resp: 407511
 Ion Ratio Lower Upper
 216 100
 214 79.2 63.4 95.0
 179 21.5 17.2 25.8
 108 19.5 15.6 23.4




#41

Hexachlorocyclopentadiene

Concen: 42.229 ng

RT: 9.039 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20


Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#42

2,4,6-Tribromophenol

Concen: 80.292 ng

RT: 10.716 min Scan# 1466

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20


Tgt Ion:330 Resp: 285205




Ion Ratio Lower Upper

330 100

332 97.6 78.1 117.1

141 33.2 26.6 39.8





#43

2,4,6-Trichlorophenol

Concen: 40.474 ng

RT: 9.157 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20




Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#44

2,4,5-Trichlorophenol

Concen: 40.651 ng

RT: 9.192 min Scan# 1207

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20

Tgt Ion:196 Resp: 304213



Ion Ratio Lower Upper




196 100

198 94.7 75.8 113.6

97 52.2 41.8 62.6

132 28.6 22.9 34.3





#45

2-Fluorobiphenyl

Concen: 76.791 ng

RT: 9.251 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20





Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#46

1,1'-Biphenyl

Concen: 39.608 ng

RT: 9.351 min Scan# 1234

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20


Tgt Ion:154 Resp: 1047094




Ion Ratio Lower Upper

154 100

153 40.3 20.3 60.3

76 13.9 0.0 33.9





#47

2-Chloronaphthalene

Concen: 39.613 ng

RT: 9.375 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20




Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#48

2-Nitroaniline

Concen: 42.562 ng

RT: 9.463 min Scan# 1253

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20

Tgt Ion: 65 Resp: 277165

Ion Ratio Lower Upper

65 100

92 63.6 50.9 76.3

138 91.7 73.4 110.0

Abundance




Time--> 9.30 9.35 9.40

9.375

Abundance

Time--> 9.40 9.45 9.50 9.55

9.463



#49

Acenaphthylene

Concen: 39.837 ng

RT: 9.792 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20





Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#50

Dimethylphthalate

Concen: 39.726 ng

RT: 9.651 min Scan# 1285

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20


Tgt Ion:163 Resp: 939670




Ion Ratio Lower Upper

163 100

194 4.2 3.4 5.0

164 10.2 8.2 12.2





#51

2,6-Dinitrotoluene

Concen: 41.580 ng

RT: 9.710 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20





Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#52

Acenaphthene


Concen: 39.811 ng




RT: 9.963 min Scan# 1338

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20

 Tgt Ion:154 Resp: 792756
 Ion Ratio Lower Upper
 154 100
 153 112.2 89.8 134.6
 152 53.0 42.4 63.6




#53

3-Nitroaniline

Concen: 42.051 ng

RT: 9.874 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20




Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#54

2,4-Dinitrophenol

Concen: 39.739 ng

RT: 9.974 min Scan# 1340

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20


Tgt Ion:184 Resp: 76948


Ion Ratio Lower Upper

184 100

63 105.9 84.7 127.1

154 422.7 338.2 507.2







#55
Dibenzofuran
Concen: 39.505 ng
RT: 10.133 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Instrument : BNA_F
ClientSampleId : SSTDICCC040


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#56
4-Nitrophenol
Concen: 43.292 ng
RT: 10.022 min Scan# 1348
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:139 Resp: 175913
Ion Ratio Lower Upper
139 100
109 73.6 53.6 93.6
65 107.8 87.8 127.8







#57
2,4-Dinitrotoluene
Concen: 42.965 ng
RT: 10.110 min Scan# 1
Instrument : BNA_F
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20
ClientSampleId : SSTDICCC040


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#58
Fluorene
Concen: 38.742 ng
RT: 10.474 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:166 Resp: 843048
Ion Ratio Lower Upper
166 100
165 98.1 78.5 117.7
167 13.6 10.9 16.3






#59
2,3,4,6-Tetrachlorophenol
Concen: 40.087 ng
RT: 10.245 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040


Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#60
Diethylphthalate
Concen: 39.715 ng
RT: 10.351 min Scan# 1404
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:149 Resp: 922384
Ion Ratio Lower Upper
149 100
177 22.4 17.9 26.9
150 12.6 10.1 15.1






#61
4-Chlorophenyl-phenylether
Concen: 39.347 ng
RT: 10.469 min Scan# 1424
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Instrument : BNA_F
ClientSampleId : SSTDICCC040


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#62
4-Nitroaniline
Concen: 41.884 ng
RT: 10.486 min Scan# 1427
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:138 Resp: 208926
Ion Ratio Lower Upper
138 100
92 55.7 35.7 75.7
108 77.5 57.5 97.5





#63

Azobenzene

Concen: 40.289 ng

RT: 10.627 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20

Instrument :

BNA_F


ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.416 min Scan# 1585

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20


Tgt Ion:188 Resp: 663459


Ion Ratio Lower Upper

188 100

94 9.9 7.9 11.9

80 11.2 9.0 13.4







#65
4,6-Dinitro-2-methylphenol
Concen: 42.799 ng
RT: 10.516 min Scan# 1432
Delta R.T. -0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#66
n-Nitrosodiphenylamine
Concen: 39.772 ng
RT: 10.586 min Scan# 1444
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:169 Resp: 789753
Ion Ratio Lower Upper
169 100
168 62.2 49.8 74.6
167 35.0 28.0 42.0









#67
4-Bromophenyl-phenylether
Concen: 39.285 ng
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#68
Hexachlorobenzene
Concen: 39.927 ng
RT: 11.021 min Scan# 1518
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:284 Resp: 306919
Ion Ratio Lower Upper
284 100
142 33.6 26.9 40.3
249 31.6 25.3 37.9







#69
Atrazine
Concen: 43.145 ng
RT: 11.110 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Instrument : BNA_F
ClientSampleId : SSTDICCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#70
Pentachlorophenol
Concen: 43.204 ng
RT: 11.210 min Scan# 1550
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:266 Resp: 201559
Ion Ratio Lower Upper
266 100
268 63.0 50.4 75.6
264 63.2 50.6 75.8





#71

Phenanthrene

Concen: 39.508 ng

RT: 11.439 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

Tgt Ion:178 Resp: 1238151

Ion Ratio Lower Upper

178 100



176 19.7 15.8 23.6

179 15.7 12.6 18.8

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#72

Anthracene

Concen: 39.628 ng

RT: 11.492 min Scan# 1598

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20


Tgt Ion:178 Resp: 1211718




Ion Ratio Lower Upper

178 100

176 19.1 15.3 22.9

179 15.5 12.4 18.6









#73
Carbazole
Concen: 39.475 ng
RT: 11.639 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#74
Di-n-butylphthalate
Concen: 40.965 ng
RT: 11.974 min Scan# 1680
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:149 Resp: 1345878
Ion Ratio Lower Upper
149 100
150 9.4 7.5 11.3
104 5.1 4.1 6.1









#75
Fluoranthene
Concen: 39.493 ng
RT: 12.627 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Instrument : BNA_F
ClientSampleId : SSTDICCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.051 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:240 Resp: 345800
Ion Ratio Lower Upper
240 100
120 11.8 9.4 14.2
236 26.1 20.9 31.3







#77
Benzidine
Concen: 44.967 ng
RT: 12.745 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Instrument : BNA_F
ClientSampleId : SSTDICCC040


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#78
Pyrene
Concen: 41.241 ng
RT: 12.857 min Scan# 1830
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:202 Resp: 1248309
Ion Ratio Lower Upper
202 100
200 21.5 17.2 25.8
203 17.8 14.2 21.4







#79
Terphenyl-d14
Concen: 81.103 ng
RT: 12.998 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Instrument : BNA_F
ClientSampleId : SSTDICCC040


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#80
Butylbenzylphthalate
Concen: 42.297 ng
RT: 13.474 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:149 Resp: 383831
Ion Ratio Lower Upper
149 100
91 74.3 59.4 89.2
206 20.5 16.4 24.6





#81

Benzo(a)anthracene

Concen: 40.901 ng

RT: 14.039 min Scan# 2

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20





Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#82

3,3'-Dichlorobenzidine

Concen: 40.798 ng


RT: 14.004 min Scan# 2025




Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20

Tgt	Ion	Ion Ratio	Resp:	Lower	Upper
252	100				
254	63.5	50.8	267769	76.2	
126	11.1	8.9		13.3	





#83

Chrysene

Concen: 39.098 ng

RT: 14.080 min Scan# 2

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20




Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#84

Bis(2-ethylhexyl)phthalate

Concen: 42.504 ng

RT: 14.033 min Scan# 2030

Delta R.T. 0.000 min


Lab File: BF139848.D


Acq: 18 Oct 2024 12:20

Tgt Ion:149 Resp: 432750

Ion Ratio Lower Upper

Tgt Ion	Ratio	Lower	Upper
149	100		
167	26.7	21.4	32.0
279	3.7	3.0	4.4







#85
 Di-n-octyl phthalate
 Concen: 42.399 ng
 RT: 14.651 min Scan# 2
 Delta R.T. -0.000 min
 Lab File: BF139848.D
 Acq: 18 Oct 2024 12:20

Instrument : BNA_F
 ClientSampleId : SSTDICCC040


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024



#86
 Perylene-d₁₂
 Concen: 20.000 ng
 RT: 15.527 min Scan# 2284
 Delta R.T. 0.000 min
 Lab File: BF139848.D
 Acq: 18 Oct 2024 12:20

Tgt Ion:264 Resp: 365076
 Ion Ratio Lower Upper
 264 100
 260 24.3 19.4 29.2
 265 21.7 17.4 26.0





#87

Indeno(1,2,3-cd)pyrene

Concen: 42.039 ng

RT: 17.021 min Scan# 2

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20





Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#88

Benzo(b)fluoranthene

Concen: 39.263 ng

RT: 15.098 min Scan# 2211

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20


Tgt Ion:252 Resp: 873168


Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4

125 9.9 7.9 11.9





#89

Benzo(k)fluoranthene

Concen: 40.589 ng

RT: 15.127 min Scan# 2

Delta R.T. 0.000 min

Lab File: BF139848.D

Acq: 18 Oct 2024 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

Tgt Ion:252 Resp: 778470

Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4


125 9.9 7.9 11.9


Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024







#91
 Dibenzo(a,h)anthracene
 Concen: 41.752 ng
 RT: 17.045 min Scan# 2
 Delta R.T. 0.000 min
 Lab File: BF139848.D
 Acq: 18 Oct 2024 12:20

Instrument : BNA_F
 ClientSampleId : SSTDICCC040


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024



#92
 Benzo(g,h,i)perylene
 Concen: 42.090 ng
 RT: 17.480 min Scan# 2616
 Delta R.T. 0.000 min
 Lab File: BF139848.D
 Acq: 18 Oct 2024 12:20

Tgt Ion:276 Resp: 823727
 Ion Ratio Lower Upper
 276 100
 277 23.8 19.0 28.6
 138 21.7 17.4 26.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139849.D
 Acq On : 18 Oct 2024 12:49
 Operator : RC/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Quant Time: Oct 18 14:19:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 14:11:20 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.892	152	197235	20.000	ng	0.00
21) Naphthalene-d8	8.175	136	739665	20.000	ng	0.00
39) Acenaphthene-d10	9.927	164	405194	20.000	ng	0.00
64) Phenanthrene-d10	11.416	188	694873	20.000	ng	0.00
76) Chrysene-d12	14.051	240	349496	20.000	ng	0.00
86) Perylene-d12	15.527	264	390757	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.498	112	1162820	92.295	ng	0.00
7) Phenol-d6	6.510	99	1517079	92.972	ng	0.00
23) Nitrobenzene-d5	7.457	82	1310450	98.193	ng	0.00
42) 2,4,6-Tribromophenol	10.716	330	362330	95.600	ng	0.00
45) 2-Fluorobiphenyl	9.251	172	2204460	89.915	ng	0.00
79) Terphenyl-d14	12.998	244	2015874	93.988	ng	0.00
Target Compounds						
				Qvalue		
2) 1,4-Dioxane	2.687	88	281572	47.933	ng	99
3) Pyridine	3.446	79	683175	46.920	ng	99
4) n-Nitrosodimethylamine	3.404	42	385387	49.264	ng	97
6) Aniline	6.557	93	725204	47.686	ng	98
8) 2-Chlorophenol	6.675	128	597810	46.414	ng	98
9) Benzaldehyde	6.439	77	430623	46.911	ng	100
10) Phenol	6.528	94	780406m	45.569	ng	
11) bis(2-Chloroethyl)ether	6.628	93	617014	47.453	ng	99
12) 1,3-Dichlorobenzene	6.834	146	686600	46.439	ng	99
13) 1,4-Dichlorobenzene	6.910	146	686616	46.483	ng	99
14) 1,2-Dichlorobenzene	7.063	146	627851	45.543	ng	99
15) Benzyl Alcohol	7.028	79	569012	47.538	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.169	45	1043921	47.084	ng	100
17) 2-Methylphenol	7.134	107	519035	47.259	ng	99
18) Hexachloroethane	7.404	117	250100	47.480	ng	100
19) n-Nitroso-di-n-propyla...	7.310	70	454173	45.892	ng	99
20) 3+4-Methylphenols	7.292	107	645650	45.961	ng	# 88
22) Acetophenone	7.304	105	835908	46.140	ng	97
24) Nitrobenzene	7.475	77	707970	48.385	ng	100
25) Isophorone	7.710	82	1205277	47.583	ng	99
26) 2-Nitrophenol	7.786	139	292634	53.013	ng	99
27) 2,4-Dimethylphenol	7.822	122	439040	47.699	ng	98
28) bis(2-Chloroethoxy)met...	7.922	93	728968	47.500	ng	100
29) 2,4-Dichlorophenol	8.028	162	502930	47.971	ng	98
30) 1,2,4-Trichlorobenzene	8.116	180	550757	47.476	ng	99
31) Naphthalene	8.198	128	1770616	46.415	ng	100
32) Benzoic acid	7.933	122	426532	53.131	ng	100
33) 4-Chloroaniline	8.245	127	613822	47.189	ng	99
34) Hexachlorobutadiene	8.316	225	343002	47.058	ng	100
35) Caprolactam	8.616	113	162260	48.675	ng	98
36) 4-Chloro-3-methylphenol	8.716	107	553487	47.678	ng	99
37) 2-Methylnaphthalene	8.886	142	1085491	46.462	ng	99
38) 1-Methylnaphthalene	8.986	142	1052879	45.933	ng	100
40) 1,2,4,5-Tetrachloroben...	9.051	216	518960	47.474	ng	99
41) Hexachlorocyclopentadiene	9.039	237	188234	49.488	ng	99
43) 2,4,6-Trichlorophenol	9.157	196	367373	47.468	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139849.D
 Acq On : 18 Oct 2024 12:49
 Operator : RC/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Quant Time: Oct 18 14:19:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 14:11:20 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.192	196	385963	48.336	ng	98
46) 1,1'-Biphenyl	9.351	154	1301562	46.143	ng	99
47) 2-Chloronaphthalene	9.375	162	1061417	46.720	ng	99
48) 2-Nitroaniline	9.469	65	358283	51.564	ng	97
49) Acenaphthylene	9.792	152	1526583	46.479	ng	99
50) Dimethylphthalate	9.651	163	1187472	47.050	ng	99
51) 2,6-Dinitrotoluene	9.710	165	277023	50.693	ng	99
52) Acenaphthene	9.963	154	989603	46.576	ng	100
53) 3-Nitroaniline	9.880	138	279467	49.591	ng	100
54) 2,4-Dinitrophenol	9.980	184	102786	48.105	ng	# 18
55) Dibenzofuran	10.133	168	1407211	46.162	ng	99
56) 4-Nitrophenol	10.022	139	221410	51.068	ng	97
57) 2,4-Dinitrotoluene	10.110	165	348628	51.878	ng	97
58) Fluorene	10.480	166	1042268	44.889	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.251	232	300306	47.975	ng	99
60) Diethylphthalate	10.351	149	1180411	47.634	ng	99
61) 4-Chlorophenyl-phenyle...	10.469	204	532654	45.427	ng	98
62) 4-Nitroaniline	10.492	138	266320	50.037	ng	99
63) Azobenzene	10.633	77	1231644	46.881	ng	98
65) 4,6-Dinitro-2-methylph...	10.522	198	156461	55.202	ng	94
66) n-Nitrosodiphenylamine	10.586	169	987295	47.472	ng	100
67) 4-Bromophenyl-phenylether	10.963	248	341790	47.746	ng	97
68) Hexachlorobenzene	11.027	284	376968	46.823	ng	98
69) Atrazine	11.110	200	234998	41.713	ng	99
70) Pentachlorophenol	11.216	266	252064	51.587	ng	98
71) Phenanthrene	11.439	178	1500010	45.700	ng	99
72) Anthracene	11.492	178	1478655	46.172	ng	99
73) Carbazole	11.645	167	1348098	45.263	ng	99
74) Di-n-butylphthalate	11.974	149	1602928	46.583	ng	100
75) Fluoranthene	12.627	202	1462936	44.089	ng	99
77) Benzidine	12.745	184	240940	42.388	ng	99
78) Pyrene	12.857	202	1471893	48.113	ng	100
80) Butylbenzylphthalate	13.474	149	467242	50.944	ng	100
81) Benzo(a)anthracene	14.045	228	1107235	48.667	ng	99
82) 3,3'-Dichlorobenzidine	14.004	252	326706	49.251	ng	98
83) Chrysene	14.080	228	990517	47.484	ng	99
84) Bis(2-ethylhexyl)phtha...	14.033	149	541828	52.655	ng	100
85) Di-n-octyl phthalate	14.651	149	1033054	55.195	ng	100
87) Indeno(1,2,3-cd)pyrene	17.027	276	1268552	50.463	ng	100
88) Benzo(b)fluoranthene	15.098	252	1079928	45.369	ng	100
89) Benzo(k)fluoranthene	15.127	252	1006425	49.026	ng	99
90) Benzo(a)pyrene	15.468	252	951346	48.573	ng	99
91) Dibenzo(a,h)anthracene	17.045	278	1058119	50.456	ng	100
92) Benzo(g,h,i)perylene	17.480	276	1055618	50.394	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139849.D
 Acq On : 18 Oct 2024 12:49
 Operator : RC/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1


Quant Time: Oct 18 14:19:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 14:11:20 2024
 Response via : Initial Calibration

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.892 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49




Instrument : BNA_F
ClientSampleId : SSTDICC050




Tgt Ion:152 Resp: 197239
Ion Ratio Lower Upper
152 100
150 160.0 130.2 195.2
115 63.6 51.4 77.2

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#2
1,4-Dioxane
Concen: 47.933 ng
RT: 2.687 min Scan# 101
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49




Tgt Ion: 88 Resp: 281572
Ion Ratio Lower Upper
88 100
58 91.1 73.3 109.9
43 33.2 27.1 40.7





#3
Pyridine
Concen: 46.920 ng
RT: 3.446 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49




Instrument : BNA_F
ClientSampleId : SSTDICC050




Tgt Ion: 79 Resp: 68317
Ion Ratio Lower Upper
79 100
52 86.6 68.5 102.7
51 41.3 32.3 48.5

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#4
n-Nitrosodimethylamine
Concen: 49.264 ng
RT: 3.404 min Scan# 223
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49




Tgt Ion: 42 Resp: 385387
Ion Ratio Lower Upper
42 100
74 116.3 96.2 144.4
44 7.2 6.2 9.2





#5
2-Fluorophenol
Concen: 92.295 ng
RT: 5.498 min Scan# 5
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49



Instrument : BNA_F
ClientSampleId : SSTDICC050




Tgt Ion:112 Resp: 1162820
Ion Ratio Lower Upper
112 100
64 67.1 53.0 79.6
63 34.7 27.0 40.4

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#6
Aniline
Concen: 47.686 ng
RT: 6.557 min Scan# 759
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49



Tgt Ion: 93 Resp: 725204
Ion Ratio Lower Upper
93 100
66 41.3 34.0 51.0
65 21.3 17.4 26.0







#7
 Phenol-d6
 Concen: 92.972 ng
 RT: 6.510 min Scan# 7
 Delta R.T. 0.000 min
 Lab File: BF139849.D
 Acq: 18 Oct 2024 12:49

Instrument : BNA_F
 ClientSampleId : SSTDICC050


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024



#8
 2-Chlorophenol
 Concen: 46.414 ng
 RT: 6.675 min Scan# 779
 Delta R.T. 0.000 min
 Lab File: BF139849.D
 Acq: 18 Oct 2024 12:49


Tgt Ion:128 Resp: 597810
 Ion Ratio Lower Upper
 128 100
 130 32.6 12.3 52.3
 64 57.0 34.6 74.6





#9
Benzaldehyde
Concen: 46.911 ng
RT: 6.439 min Scan# 739
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49



Instrument : BNA_F
ClientSampleId : SSTDICC050




Tgt Ion: 77 Resp: 43062
Ion Ratio Lower Upper
77 100
105 95.2 75.5 115.5
106 92.0 71.7 111.7

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#10
Phenol
Concen: 45.569 ng m
RT: 6.528 min Scan# 754
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49




Tgt Ion: 94 Resp: 780406
Ion Ratio Lower Upper
94 100
65 31.6 11.1 51.1
66 43.3 22.7 62.7





#11
bis(2-Chloroethyl)ether
Concen: 47.453 ng
RT: 6.628 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49



Instrument : BNA_F
ClientSampleId : SSTDICC050




Tgt Ion: 93 Resp: 61701
Ion Ratio Lower Upper
93 100
63 85.8 65.0 105.0
95 32.2 12.4 52.4

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024









#12
1,3-Dichlorobenzene
Concen: 46.439 ng
RT: 6.834 min Scan# 806
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49



Tgt Ion:146 Resp: 686600
Ion Ratio Lower Upper
146 100
148 64.7 51.0 76.4
75 33.3 26.4 39.6





#13

1,4-Dichlorobenzene

Concen: 46.483 ng

RT: 6.910 min Scan# 819

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Instrument :

BNA_F

ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#14

1,2-Dichlorobenzene

Concen: 45.543 ng


RT: 7.063 min Scan# 845




Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Tgt	Ion:146	Resp:	627851
Ion	Ratio	Lower	Upper
146	100		
148	63.7	51.4	77.2
111	46.3	36.7	55.1





#15

Benzyl Alcohol

Concen: 47.538 ng

RT: 7.028 min Scan# 8

Delta R.T. 0.000 min




Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Instrument : BNA_F

ClientSampleId : SSTDICC050

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#16

2,2'-oxybis(1-Chloropropane)

Concen: 47.084 ng

RT: 7.169 min Scan# 863


Delta R.T. 0.000 min




Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Tgt Ion: 45 Resp: 1043921

Ion	Ratio	Lower	Upper
45	100		
77	11.6	0.0	31.5
79	9.0	0.0	29.1





#17

2-Methylphenol

Concen: 47.259 ng

RT: 7.134 min Scan# 8

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49




Instrument :

BNA_F

ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#18

Hexachloroethane

Concen: 47.480 ng

RT: 7.404 min Scan# 903

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Tgt Ion:117 Resp: 250100


Ion Ratio Lower Upper

117 100

119 96.0 77.0 115.4

201 95.3 76.2 114.4







#19
n-Nitroso-di-n-propylamine
Concen: 45.892 ng
RT: 7.310 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Instrument :
BNA_F
ClientSampleId :
SSTDICC050


Manual Integrations APPROVED







Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#20
3+4-Methylphenols
Concen: 45.961 ng
RT: 7.292 min Scan# 884
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:107 Resp: 645650
Ion Ratio Lower Upper
107 100
108 90.8 69.8 109.8
77 95.4 50.5 90.5#
79 31.2 10.2 50.2





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.175 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC050

Tgt Ion:136 Resp: 73966

Ion Ratio Lower Upper

136 100

137 11.2 8.6 12.8

54 10.6 8.4 12.6


68 6.7 5.1 7.7

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#22

Acetophenone

Concen: 46.140 ng

RT: 7.304 min Scan# 886

Delta R.T. 0.006 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Tgt Ion:105 Resp: 835908


Ion Ratio Lower Upper




105 100

71 7.9 5.9 8.9

51 33.7 29.4 44.2

120 22.0 17.5 26.3





#23

Nitrobenzene-d5

Concen: 98.193 ng

RT: 7.457 min Scan# 912

Delta R.T. 0.006 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49




Instrument :

BNA_F

ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#24

Nitrobenzene

Concen: 48.385 ng

RT: 7.475 min Scan# 915

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Tgt Ion: 77 Resp: 707970




Ion Ratio Lower Upper

77 100

123 44.2 35.7 53.5

65 14.5 11.5 17.3





#25

Isophorone

Concen: 47.583 ng

RT: 7.710 min Scan# 9

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49





Instrument :

BNA_F

ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#26

2-Nitrophenol

Concen: 53.013 ng

RT: 7.786 min Scan# 968

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Tgt Ion:139 Resp: 292634




Ion Ratio Lower Upper

139 100

109 31.4 25.1 37.7

65 58.4 47.5 71.3





#27

2,4-Dimethylphenol

Concen: 47.699 ng

RT: 7.822 min Scan# 9

Delta R.T. 0.006 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49




Instrument :

BNA_F

ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#28

bis(2-Chloroethoxy)methane

Concen: 47.500 ng

RT: 7.922 min Scan# 991

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Tgt Ion: 93 Resp: 728968




Ion Ratio Lower Upper

93 100

95 32.4 26.0 39.0

123 12.0 9.8 14.8





#29

2,4-Dichlorophenol

Concen: 47.971 ng

RT: 8.028 min Scan# 1

Delta R.T. 0.006 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC050

Tgt Ion:162 Resp: 502930

Ion Ratio Lower Upper

162 100




164 65.3 43.9 83.9

98 38.6 19.9 59.9

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#30

1,2,4-Trichlorobenzene

Concen: 47.476 ng

RT: 8.116 min Scan# 1024

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Tgt Ion:180 Resp: 550757




Ion Ratio Lower Upper

180 100

182 99.7 80.5 120.7

145 32.4 25.4 38.0





#31

Naphthalene

Concen: 46.415 ng

RT: 8.198 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Instrument :

BNA_F





ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#32

Benzoic acid

Concen: 53.131 ng


RT: 7.933 min Scan# 993







Delta R.T. 0.011 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Tgt	Ion:122	Resp:	426532
Ion	Ratio	Lower	Upper
122	100		
105	127.2	107.6	147.6
77	100.6	80.2	120.2





#33

4-Chloroaniline

Concen: 47.189 ng

RT: 8.245 min Scan# 1

Delta R.T. 0.006 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**
Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024

Tgt Ion:127 Resp: 613821


Ion Ratio Lower Upper

127 100

129 32.7 25.4 38.2

65 34.6 28.4 42.6

92 20.5 16.6 25.0



#34

Hexachlorobutadiene

Concen: 47.058 ng

RT: 8.316 min Scan# 1058

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Tgt Ion:225 Resp: 343002


Ion Ratio Lower Upper

225 100



223 63.7 50.6 76.0

227 65.1 51.8 77.8







Abundance Scan 1109 (8.616 min): BF139849.D\data.ms



Abundance Scan 1126 (8.716 min): BF139849.D\data.ms



#35

Caprolactam

Concen: 48.675 ng

RT: 8.616 min Scan# 1108

Delta R.T. 0.006 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC050

Tgt Ion:113 Resp: 162260

Ion Ratio Lower Upper

113 100


55 208.2 189.2 229.2

56 157.0 132.5 172.5

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#36

4-Chloro-3-methylphenol

Concen: 47.678 ng

RT: 8.716 min Scan# 1126

Delta R.T. 0.006 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Tgt Ion:107 Resp: 553487




Ion Ratio Lower Upper

107 100

144 24.3 19.1 28.7

142 72.6 57.2 85.8





#37

2-Methylnaphthalene

Concen: 46.462 ng

RT: 8.886 min Scan# 1155

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49





Instrument :

BNA_F

ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#38

1-Methylnaphthalene

Concen: 45.933 ng

RT: 8.986 min Scan# 1172

Delta R.T. 0.000 min

Lab File: BF139849.D


Acq: 18 Oct 2024 12:49




Tgt Ion:142 Resp: 1052879

Ion Ratio Lower Upper

142 100

141 91.8 73.5 110.3





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.927 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Instrument :

BNA_F





ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#40

1,2,4,5-Tetrachlorobenzene

Concen: 47.474 ng


RT: 9.051 min Scan# 1183


Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Tgt	Ion:216	Resp:	518960
Ion	Ratio	Lower	Upper
216	100		
214	78.3	63.4	95.0
179	21.5	17.2	25.8
108	19.4	15.6	23.4





#41

Hexachlorocyclopentadiene

Concen: 49.488 ng

RT: 9.039 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC050

Tgt Ion:237 Resp: 188234

Ion Ratio Lower Upper

237 100



235 63.8 42.9 82.9

272 12.0 0.0 32.3

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#42

2,4,6-Tribromophenol

Concen: 95.600 ng

RT: 10.716 min Scan# 1466

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Tgt Ion:330 Resp: 362330







Ion Ratio Lower Upper

330 100

332 98.1 78.1 117.1

141 32.6 26.6 39.8





#43

2,4,6-Trichlorophenol

Concen: 47.468 ng

RT: 9.157 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Instrument :

BNA_F

ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#44

2,4,5-Trichlorophenol

Concen: 48.336 ng

RT: 9.192 min Scan# 1207

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Tgt Ion:196 Resp: 385963


Ion Ratio Lower Upper




196 100

198 96.2 75.8 113.6

97 54.3 41.8 62.6

132 28.5 22.9 34.3





#45

2-Fluorobiphenyl

Concen: 89.915 ng

RT: 9.251 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49





Instrument :

BNA_F

ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#46

1,1'-Biphenyl

Concen: 46.143 ng

RT: 9.351 min Scan# 1234

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Tgt Ion:154 Resp: 1301562




Ion Ratio Lower Upper

154 100

153 40.7 20.3 60.3

76 14.1 0.0 33.9





#47

2-Chloronaphthalene

Concen: 46.720 ng

RT: 9.375 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49





Instrument :

BNA_F

ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#48

2-Nitroaniline

Concen: 51.564 ng

RT: 9.469 min Scan# 1254

Delta R.T. 0.006 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Tgt Ion: 65 Resp: 358283







Ion Ratio Lower Upper

65 100

92 64.6 50.9 76.3

138 96.2 73.4 110.0





#49

Acenaphthylene

Concen: 46.479 ng

RT: 9.792 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Instrument :

BNA_F


ClientSampleId :

SSTDICC050

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#50

Dimethylphthalate

Concen: 47.050 ng

RT: 9.651 min Scan# 1285

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Tgt Ion:163 Resp: 1187472







Ion Ratio Lower Upper

163 100

194 3.9 3.4 5.0

164 10.4 8.2 12.2





#51

2,6-Dinitrotoluene

Concen: 50.693 ng

RT: 9.710 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Instrument :

BNA_F


ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#52

Acenaphthene

Concen: 46.576 ng

RT: 9.963 min Scan# 1338

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Tgt Ion:154 Resp: 989603




Ion Ratio Lower Upper

154 100

153 112.4 89.8 134.6

152 53.4 42.4 63.6








#53
3-Nitroaniline
Concen: 49.591 ng
RT: 9.880 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Instrument : BNA_F
ClientSampleId : SSTDICC050


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#54
2,4-Dinitrophenol
Concen: 48.105 ng
RT: 9.980 min Scan# 1341
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:184 Resp: 102786
Ion Ratio Lower Upper
184 100
63 84.8 84.7 127.1
154 182.9 338.2 507.2#





#55

Dibenzofuran

Concen: 46.162 ng

RT: 10.133 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49




Instrument :

BNA_F

ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#56

4-Nitrophenol

Concen: 51.068 ng

RT: 10.022 min Scan# 1348

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Tgt Ion:139 Resp: 221410


Ion Ratio Lower Upper

139 100

109 74.0 53.6 93.6

65 112.1 87.8 127.8





#57

2,4-Dinitrotoluene

Concen: 51.878 ng

RT: 10.110 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Instrument :

BNA_F



ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024

 Tgt Ion:165 Resp: 348623
 Ion Ratio Lower Upper

 165 100
 63 52.2 40.1 60.1
 89 78.4 60.4 90.6
 182 2.6 2.0 3.0


#58

Fluorene

Concen: 44.889 ng


RT: 10.480 min Scan# 1426


Delta R.T. 0.006 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Tgt Ion:166 Resp: 1042268

 Ion Ratio Lower Upper
 166 100
 165 98.6 78.5 117.7
 167 13.5 10.9 16.3




#59

2,3,4,6-Tetrachlorophenol

Concen: 47.975 ng

RT: 10.251 min Scan# 1

Delta R.T. 0.006 min

Lab File: BF139849.D


Acq: 18 Oct 2024 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**
Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024

#60

Diethylphthalate

Concen: 47.634 ng

RT: 10.351 min Scan# 1404

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Tgt Ion:149 Resp: 1180411


Ion Ratio Lower Upper

149 100

177 21.9 17.9 26.9

150 12.5 10.1 15.1







#61
4-Chlorophenyl-phenylether
Concen: 45.427 ng
RT: 10.469 min Scan# 1424
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Instrument : BNA_F
ClientSampleId : SSTDICC050


Manual Integrations
APPROVED







Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#62
4-Nitroaniline
Concen: 50.037 ng
RT: 10.492 min Scan# 1428
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:138 Resp: 266320
Ion Ratio Lower Upper
138 100
92 55.4 35.7 75.7
108 78.6 57.5 97.5





#63

Azobenzene

Concen: 46.881 ng

RT: 10.633 min Scan# 1

Delta R.T. 0.006 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Instrument :

BNA_F


ClientSampleId :

SSTDICC050

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.416 min Scan# 1585

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Tgt Ion:188 Resp: 694873



Ion Ratio Lower Upper

188 100

94 10.2 7.9 11.9

80 11.1 9.0 13.4






#65
4,6-Dinitro-2-methylphenol
Concen: 55.202 ng
RT: 10.522 min Scan# 1433
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Instrument : BNA_F
ClientSampleId : SSTDICC050


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#66
n-Nitrosodiphenylamine
Concen: 47.472 ng
RT: 10.586 min Scan# 1444
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:169 Resp: 987295
Ion Ratio Lower Upper
169 100
168 62.4 49.8 74.6
167 34.6 28.0 42.0









#67
4-Bromophenyl-phenylether
Concen: 47.746 ng
RT: 10.963 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Instrument :
BNA_F
ClientSampleId :
SSTDICC050


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#68
Hexachlorobenzene
Concen: 46.823 ng
RT: 11.027 min Scan# 1519
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:284 Resp: 376968
Ion Ratio Lower Upper
284 100
142 32.2 26.9 40.3
249 31.3 25.3 37.9







#69
Atrazine
Concen: 41.713 ng
RT: 11.110 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Instrument : BNA_F
ClientSampleId : SSTDICC050


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#70
Pentachlorophenol
Concen: 51.587 ng
RT: 11.216 min Scan# 1551
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:266 Resp: 252064
Ion Ratio Lower Upper
266 100
268 64.2 50.4 75.6
264 61.7 50.6 75.8





#71

Phenanthrene

Concen: 45.700 ng

RT: 11.439 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Instrument :

BNA_F



ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#72

Anthracene

Concen: 46.172 ng


RT: 11.492 min Scan# 1598




Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Tgt	Ion:178	Resp:	1478655
Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.3	22.9
179	15.8	12.4	18.6









#73
Carbazole
Concen: 45.263 ng
RT: 11.645 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Instrument :
BNA_F
ClientSampleId :
SSTDICC050


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#74
Di-n-butylphthalate
Concen: 46.583 ng
RT: 11.974 min Scan# 1680
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:149 Resp: 1602928
Ion Ratio Lower Upper
149 100
150 9.5 7.5 11.3
104 5.2 4.1 6.1







#75
Fluoranthene
Concen: 44.089 ng
RT: 12.627 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Instrument : BNA_F
ClientSampleId : SSTDICC050


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#76
Chrysene-d₁₂
Concen: 20.000 ng
RT: 14.051 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:240 Resp: 349496
Ion Ratio Lower Upper
240 100
120 12.0 9.4 14.2
236 26.0 20.9 31.3







#77
Benzidine
Concen: 42.388 ng
RT: 12.745 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Instrument : BNA_F
ClientSampleId : SSTDICC050


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#78
Pyrene
Concen: 48.113 ng
RT: 12.857 min Scan# 1830
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:202 Resp: 1471893
Ion Ratio Lower Upper
202 100
200 21.8 17.2 25.8
203 17.8 14.2 21.4





#79

Terphenyl-d14

Concen: 93.988 ng

RT: 12.998 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

Instrument :

BNA_F



ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#80

Butylbenzylphthalate

Concen: 50.944 ng

RT: 13.474 min Scan# 1935

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Tgt Ion:149 Resp: 467242







Ion Ratio Lower Upper

149 100

91 74.5 59.4 89.2

206 20.2 16.4 24.6





#81

Benzo(a)anthracene

Concen: 48.667 ng

RT: 14.045 min Scan# 2

Delta R.T. 0.006 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Instrument :

BNA_F

ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#82

3,3'-Dichlorobenzidine

Concen: 49.251 ng

RT: 14.004 min Scan# 2025

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Tgt Ion:252 Resp: 326706




Ion Ratio Lower Upper

252 100

254 65.6 50.8 76.2

126 11.0 8.9 13.3





#83

Chrysene

Concen: 47.484 ng

RT: 14.080 min Scan# 2

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49





Instrument :

BNA_F

ClientSampleId :

SSTDICC050

Manual Integrations
APPROVED

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#84

Bis(2-ethylhexyl)phthalate

Concen: 52.655 ng

RT: 14.033 min Scan# 2030

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Tgt Ion:149 Resp: 541828


Ion Ratio Lower Upper

149 100


167 26.7 21.4 32.0

279 4.1 3.0 4.4








Abundance Scan 2135 (14.651 min): BF139849.D\data.ms (-)




Abundance Scan 2135 (14.651 min): BF139849.D\data.ms (-)



Abundance Scan 2284 (15.527 min): BF139849.D\data.ms (-)



Abundance Scan 2284 (15.527 min): BF139849.D\data.ms (-)



#85

Di-n-octyl phthalate

Concen: 55.195 ng

RT: 14.651 min Scan# 2

Instrument :

BNA_F

Delta R.T. -0.000 min

ClientSampleId :

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

SSTDICC050

Tgt Ion:149 Resp: 1033054

Ion Ratio Lower Upper

149 100


167 1.5 1.2 1.8

43 12.4 9.8 14.6

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.527 min Scan# 2284

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Tgt Ion:264 Resp: 390757







Ion Ratio Lower Upper

264 100

260 23.7 19.4 29.2

265 21.7 17.4 26.0





#87

Indeno(1,2,3-cd)pyrene

Concen: 50.463 ng

RT: 17.027 min Scan# 2

Delta R.T. 0.006 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Instrument :

BNA_F

ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#88

Benzo(b)fluoranthene

Concen: 45.369 ng

RT: 15.098 min Scan# 2211

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49


Tgt Ion:252 Resp: 1079928


Ion Ratio Lower Upper

252 100

253 21.8 17.6 26.4

125 10.0 7.9 11.9





#89

Benzo(k)fluoranthene

Concen: 49.026 ng

RT: 15.127 min Scan# 2

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49



Instrument :

BNA_F

ClientSampleId :

SSTDICC050

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#90

Benzo(a)pyrene


Concen: 48.573 ng


RT: 15.468 min Scan# 2274

Delta R.T. 0.000 min

Lab File: BF139849.D

Acq: 18 Oct 2024 12:49

 Tgt Ion:252 Resp: 951346
 Ion Ratio Lower Upper
 252 100
 253 22.2 17.3 25.9
 125 11.1 8.7 13.1






#91
 Dibenzo(a,h)anthracene
 Concen: 50.456 ng
 RT: 17.045 min Scan# 2
 Delta R.T. 0.000 min
 Lab File: BF139849.D
 Acq: 18 Oct 2024 12:49

Instrument : BNA_F
 ClientSampleId : SSTDICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024



#92
 Benzo(g,h,i)perylene
 Concen: 50.394 ng
 RT: 17.480 min Scan# 2616
 Delta R.T. 0.000 min
 Lab File: BF139849.D
 Acq: 18 Oct 2024 12:49

Tgt Ion:276 Resp: 1055618
 Ion Ratio Lower Upper
 276 100
 277 23.5 19.0 28.6
 138 22.4 17.4 26.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139850.D
 Acq On : 18 Oct 2024 13:17
 Operator : RC/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Quant Time: Oct 18 14:20:24 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 14:11:20 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.893	152	189346	20.000	ng	0.00
21) Naphthalene-d8	8.175	136	712033	20.000	ng	0.00
39) Acenaphthene-d10	9.928	164	389299	20.000	ng	0.00
64) Phenanthrene-d10	11.416	188	667372	20.000	ng	0.00
76) Chrysene-d12	14.057	240	337032	20.000	ng	0.00
86) Perylene-d12	15.527	264	377203	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.499	112	1347774	111.432	ng	0.00
7) Phenol-d6	6.516	99	1748088	111.592	ng	0.00
23) Nitrobenzene-d5	7.457	82	1523975	118.624	ng	0.00
42) 2,4,6-Tribromophenol	10.722	330	420891	115.586	ng	0.00
45) 2-Fluorobiphenyl	9.251	172	2475821	105.106	ng	0.00
79) Terphenyl-d14	13.004	244	2266127	109.563	ng	0.00
Target Compounds						
				Qvalue		
2) 1,4-Dioxane	2.687	88	330127	58.541	ng	98
3) Pyridine	3.446	79	798529	57.127	ng	99
4) n-Nitrosodimethylamine	3.410	42	451209	60.081	ng	98
6) Aniline	6.551	93	840113	57.544	ng	100
8) 2-Chlorophenol	6.675	128	693549	56.091	ng	96
9) Benzaldehyde	6.440	77	484510	54.980	ng	99
10) Phenol	6.528	94	909456m	55.317	ng	
11) bis(2-Chloroethyl)ether	6.628	93	709494	56.838	ng	99
12) 1,3-Dichlorobenzene	6.834	146	790296	55.679	ng	98
13) 1,4-Dichlorobenzene	6.910	146	790397	55.738	ng	99
14) 1,2-Dichlorobenzene	7.063	146	719470	54.363	ng	99
15) Benzyl Alcohol	7.028	79	650971	56.651	ng	98
16) 2,2'-oxybis(1-Chloropr...	7.169	45	1201584	56.453	ng	99
17) 2-Methylphenol	7.134	107	603882	57.276	ng	99
18) Hexachloroethane	7.404	117	290250	57.398	ng	100
19) n-Nitroso-di-n-propyla...	7.310	70	526516	55.418	ng	99
20) 3+4-Methylphenols	7.293	107	738542	54.764	ng	93
22) Acetophenone	7.304	105	969975	55.618	ng	99
24) Nitrobenzene	7.475	77	828015	58.785	ng	98
25) Isophorone	7.716	82	1409786	57.816	ng	100
26) 2-Nitrophenol	7.787	139	344300	64.793	ng	100
27) 2,4-Dimethylphenol	7.822	122	500332	56.468	ng	99
28) bis(2-Chloroethoxy)met...	7.922	93	832755	56.369	ng	99
29) 2,4-Dichlorophenol	8.028	162	584871	57.952	ng	99
30) 1,2,4-Trichlorobenzene	8.116	180	635946	56.947	ng	99
31) Naphthalene	8.198	128	2016330	54.907	ng	99
32) Benzoic acid	7.945	122	502169	64.980	ng	98
33) 4-Chloroaniline	8.245	127	695566	55.548	ng	99
34) Hexachlorobutadiene	8.316	225	399483	56.934	ng	99
35) Caprolactam	8.622	113	188948	58.880	ng	97
36) 4-Chloro-3-methylphenol	8.716	107	646732	57.872	ng	100
37) 2-Methylnaphthalene	8.887	142	1245977	55.401	ng	99
38) 1-Methylnaphthalene	8.987	142	1211981	54.926	ng	100
40) 1,2,4,5-Tetrachloroben...	9.051	216	591461	56.315	ng	100
41) Hexachlorocyclopentadiene	9.039	237	216567	59.262	ng	100
43) 2,4,6-Trichlorophenol	9.157	196	447702	60.209	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139850.D
 Acq On : 18 Oct 2024 13:17
 Operator : RC/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC060

Quant Time: Oct 18 14:20:24 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 14:11:20 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.198	196	431080	56.191	ng	99
46) 1,1'-Biphenyl	9.351	154	1474193	54.396	ng	99
47) 2-Chloronaphthalene	9.375	162	1214856	55.657	ng	98
48) 2-Nitroaniline	9.469	65	423080	63.375	ng	97
49) Acenaphthylene	9.792	152	1758248	55.718	ng	99
50) Dimethylphthalate	9.657	163	1358812	56.037	ng	100
51) 2,6-Dinitrotoluene	9.710	165	320076	60.963	ng	97
52) Acenaphthene	9.969	154	1148095	56.242	ng	98
53) 3-Nitroaniline	9.881	138	328790	60.726	ng	98
54) 2,4-Dinitrophenol	9.981	184	131611	61.938	ng	# 29
55) Dibenzofuran	10.139	168	1605572	54.819	ng	100
56) 4-Nitrophenol	10.028	139	256728	61.631	ng	100
57) 2,4-Dinitrotoluene	10.116	165	410361	63.558	ng	99
58) Fluorene	10.481	166	1192018	53.435	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.251	232	342475	56.946	ng	99
60) Diethylphthalate	10.351	149	1355992	56.954	ng	99
61) 4-Chlorophenyl-phenyle...	10.469	204	606730	53.858	ng	98
62) 4-Nitroaniline	10.498	138	314525	61.507	ng	98
63) Azobenzene	10.633	77	1415762	56.090	ng	99
65) 4,6-Dinitro-2-methylph...	10.522	198	184974	67.951	ng	99
66) n-Nitrosodiphenylamine	10.592	169	1122831	56.214	ng	100
67) 4-Bromophenyl-phenylether	10.963	248	392570	57.100	ng	98
68) Hexachlorobenzene	11.028	284	442187	57.187	ng	98
69) Atrazine	11.116	200	307149	56.767	ng	99
70) Pentachlorophenol	11.216	266	289014	61.587	ng	99
71) Phenanthrene	11.439	178	1721763	54.618	ng	99
72) Anthracene	11.492	178	1667221	54.206	ng	99
73) Carbazole	11.645	167	1545085	54.015	ng	99
74) Di-n-butylphthalate	11.975	149	1819929	55.069	ng	100
75) Fluoranthene	12.627	202	1671405	52.448	ng	99
77) Benzidine	12.745	184	205548	37.499	ng	99
78) Pyrene	12.857	202	1667489	56.522	ng	99
80) Butylbenzylphthalate	13.474	149	537039	60.719	ng	98
81) Benzo(a)anthracene	14.045	228	1250382	56.992	ng	100
82) 3,3'-Dichlorobenzidine	14.004	252	386072	60.353	ng	99
83) Chrysene	14.080	228	1163808	57.855	ng	100
84) Bis(2-ethylhexyl)phtha...	14.033	149	632594	63.749	ng	100
85) Di-n-octyl phthalate	14.651	149	1220737	67.634	ng	100
87) Indeno(1,2,3-cd)pyrene	17.027	276	1500084	61.817	ng	99
88) Benzo(b)fluoranthene	15.098	252	1402395	61.033	ng	100
89) Benzo(k)fluoranthene	15.127	252	1051303	53.053	ng	100
90) Benzo(a)pyrene	15.468	252	1129942	59.764	ng	100
91) Dibenzo(a,h)anthracene	17.051	278	1228124	60.667	ng	99
92) Benzo(g,h,i)perylene	17.486	276	1238955	61.272	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139850.D
 Acq On : 18 Oct 2024 13:17
 Operator : RC/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 8 Sample Multiplier: 1


Quant Time: Oct 18 14:20:24 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 14:11:20 2024
 Response via : Initial Calibration

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024







#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.893 min Scan# 8
 Delta R.T. 0.001 min
 Lab File: BF139850.D
 Acq: 18 Oct 2024 13:17

Instrument : BNA_F
 ClientSampleId : SSTDICC060


Manual Integrations
APPROVED







Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024



#2
 1,4-Dioxane
 Concen: 58.541 ng
 RT: 2.687 min Scan# 101
 Delta R.T. 0.000 min
 Lab File: BF139850.D
 Acq: 18 Oct 2024 13:17

Tgt Ion: 88 Resp: 330127
 Ion Ratio Lower Upper
 88 100
 58 89.7 73.3 109.9
 43 33.0 27.1 40.7





#3

Pyridine

Concen: 57.127 ng

RT: 3.446 min Scan# 2

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICC060

Tgt Ion: 79 Resp: 798529

Ion Ratio Lower Upper

79 100


52 86.8 68.5 102.7

51 41.4 32.3 48.5

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#4

n-Nitrosodimethylamine

Concen: 60.081 ng

RT: 3.410 min Scan# 224

Delta R.T. 0.012 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17


Tgt Ion: 42 Resp: 451209


Ion Ratio Lower Upper

42 100

74 118.0 96.2 144.4


44 7.1 6.2 9.2





#5
2-Fluorophenol
Concen: 111.432 ng
RT: 5.499 min Scan# 5
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17



Instrument : BNA_F
ClientSampleId : SSTDICC060




Tgt Ion:112 Resp: 134777
Ion Ratio Lower Upper
112 100
64 67.5 53.0 79.6
63 34.5 27.0 40.4

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#6
Aniline
Concen: 57.544 ng
RT: 6.551 min Scan# 758
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17



Tgt Ion: 93 Resp: 840113
Ion Ratio Lower Upper
93 100
66 42.2 34.0 51.0
65 21.7 17.4 26.0







#7
Phenol-d6
Concen: 111.592 ng
RT: 6.516 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Instrument : BNA_F
ClientSampleId : SSTDICC060


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#8
2-Chlorophenol
Concen: 56.091 ng
RT: 6.675 min Scan# 779
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17


Tgt Ion:128 Resp: 693549
Ion Ratio Lower Upper
128 100
130 32.5 12.3 52.3
64 58.5 34.6 74.6





#9
Benzaldehyde
Concen: 54.980 ng
RT: 6.440 min Scan# 7
Delta R.T. 0.001 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17




Instrument : BNA_F
ClientSampleId : SSTDICC060




Tgt Ion: 77 Resp: 484510
Ion Ratio Lower Upper
77 100
105 95.6 75.5 115.5
106 92.7 71.7 111.7

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#10
Phenol
Concen: 55.317 ng m
RT: 6.528 min Scan# 754
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17




Tgt Ion: 94 Resp: 909456
Ion Ratio Lower Upper
94 100
65 32.2 11.1 51.1
66 44.9 22.7 62.7





#11
bis(2-Chloroethyl)ether
Concen: 56.838 ng
RT: 6.628 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17




Instrument : BNA_F
ClientSampleId : SSTDICC060




Tgt Ion: 93 Resp: 70949
Ion Ratio Lower Upper
93 100
63 86.7 65.0 105.0
95 32.4 12.4 52.4

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024









#12
1,3-Dichlorobenzene
Concen: 55.679 ng
RT: 6.834 min Scan# 806
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17



Tgt Ion:146 Resp: 790296
Ion Ratio Lower Upper
146 100
148 65.1 51.0 76.4
75 34.3 26.4 39.6





#13

1,4-Dichlorobenzene

Concen: 55.738 ng

RT: 6.910 min Scan# 8

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17


Instrument :

BNA_F

ClientSampleId :

SSTDICC060

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#14

1,2-Dichlorobenzene

Concen: 54.363 ng

RT: 7.063 min Scan# 845

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17


Tgt Ion:146 Resp: 719470




Ion Ratio Lower Upper

146 100

148 63.8 51.4 77.2

111 46.8 36.7 55.1





#15

Benzyl Alcohol

Concen: 56.651 ng

RT: 7.028 min Scan# 8

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17




Instrument :

BNA_F

ClientSampleId :

SSTDICC060

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#16

2,2'-oxybis(1-Chloropropane)

Concen: 56.453 ng

RT: 7.169 min Scan# 863

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17


Tgt Ion: 45 Resp: 1201584




Ion Ratio Lower Upper

45 100

77 11.9 0.0 31.5

79 9.3 0.0 29.1








#17
2-Methylphenol
Concen: 57.276 ng
RT: 7.134 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDICC060


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#18
Hexachloroethane
Concen: 57.398 ng
RT: 7.404 min Scan# 903
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:117 Resp: 290250
Ion Ratio Lower Upper
117 100
119 96.0 77.0 115.4
201 94.9 76.2 114.4







#19
n-Nitroso-di-n-propylamine
Concen: 55.418 ng
RT: 7.310 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDICC060


Manual Integrations APPROVED







Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#20
3+4-Methylphenols
Concen: 54.764 ng
RT: 7.293 min Scan# 884
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:107 Resp: 738542
Ion Ratio Lower Upper
107 100
108 89.8 69.8 109.8
77 86.6 50.5 90.5
79 30.9 10.2 50.2





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.175 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17

Instrument :

BNA_F


ClientSampleId :

SSTDICC060

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#22

Acetophenone

Concen: 55.618 ng


RT: 7.304 min Scan# 886







Delta R.T. 0.006 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17

Tgt	Ion:105	Resp:	969975
Ion	Ratio	Lower	Upper
105	100		
71	7.0	5.9	8.9
51	35.8	29.4	44.2
120	21.5	17.5	26.3





#23

Nitrobenzene-d5

Concen: 118.624 ng

RT: 7.457 min Scan# 912

Delta R.T. 0.006 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17

Instrument :

BNA_F


ClientSampleId :

SSTDICC060

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#24

Nitrobenzene

Concen: 58.785 ng

RT: 7.475 min Scan# 915


Delta R.T. 0.000 min


Lab File: BF139850.D

Acq: 18 Oct 2024 13:17



Tgt Ion: 77 Resp: 828015

Ion	Ratio	Lower	Upper
77	100		
123	43.4	35.7	53.5
65	14.9	11.5	17.3






Abundance Scan 956 (7.716 min): BF139850.D\data.ms



Abundance Scan 968 (7.787 min): BF139850.D\data.ms



#25

Isophorone

Concen: 57.816 ng

RT: 7.716 min Scan# 9

Delta R.T. 0.006 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17

Instrument :


BNA_F

ClientSampleId :

SSTDICC060

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#26

2-Nitrophenol

Concen: 64.793 ng

RT: 7.787 min Scan# 968

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17


Tgt Ion:139 Resp: 344300




Ion Ratio Lower Upper

139 100

109 31.6 25.1 37.7

65 59.3 47.5 71.3





#27

2,4-Dimethylphenol

Concen: 56.468 ng

RT: 7.822 min Scan# 9

Delta R.T. 0.006 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17




Instrument :

BNA_F

ClientSampleId :

SSTDICC060

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#28

bis(2-Chloroethoxy)methane

Concen: 56.369 ng

RT: 7.922 min Scan# 991

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17


Tgt Ion: 93 Resp: 832755







Ion Ratio Lower Upper

93 100

95 32.9 26.0 39.0

123 11.8 9.8 14.8






#29


2,4-Dichlorophenol
Concen: 57.952 ngRT: 8.028 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17Instrument :
BNA_F
ClientSampleId :
SSTDICC060




Tgt Ion:162 Resp: 58487:
Ion Ratio Lower Upper
162 100
164 63.9 43.9 83.9
98 38.9 19.9 59.9

**Manual Integrations
APPROVED**Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024

#30
1,2,4-Trichlorobenzene
Concen: 56.947 ng
RT: 8.116 min Scan# 1024
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:180 Resp: 635946
Ion Ratio Lower Upper
180 100
182 99.7 80.5 120.7
145 32.6 25.4 38.0





#31

Naphthalene

Concen: 54.907 ng

RT: 8.198 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17

Instrument :

BNA_F





ClientSampleId :

SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#32

Benzoic acid

Concen: 64.980 ng


RT: 7.945 min Scan# 995







Delta R.T. 0.023 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17

Tgt Ion:122 Resp: 502169
 Ion Ratio Lower Upper
 122 100
 105 127.1 107.6 147.6
 77 96.2 80.2 120.2





#33

4-Chloroaniline

Concen: 55.548 ng

RT: 8.245 min Scan# 1

Delta R.T. 0.006 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICC060

**Manual Integrations
APPROVED**
Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024

Tgt Ion:127 Resp: 695560

Ion Ratio Lower Upper


127 100

129 32.3 25.4 38.2

65 35.7 28.4 42.6

92 20.7 16.6 25.0

Abundance



#34

Hexachlorobutadiene

Concen: 56.934 ng

RT: 8.316 min Scan# 1058

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17

Tgt Ion:225 Resp: 399483


Ion Ratio Lower Upper


225 100

223 62.8 50.6 76.0



227 63.6 51.8 77.8

Abundance







Abundance Scan 1110 (8.622 min): BF139850.D\data.ms



Abundance Scan 1126 (8.716 min): BF139850.D\data.ms



#35

Caprolactam

Concen: 58.880 ng

RT: 8.622 min Scan# 1

Delta R.T. 0.012 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICC060

Tgt Ion:113 Resp: 188943

Ion Ratio Lower Upper

113 100


55 213.0 189.2 229.2

56 156.6 132.5 172.5

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#36

4-Chloro-3-methylphenol

Concen: 57.872 ng

RT: 8.716 min Scan# 1126

Delta R.T. 0.006 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17


Tgt Ion:107 Resp: 646732


Ion Ratio Lower Upper

107 100


144 24.3 19.1 28.7

142 71.2 57.2 85.8








Abundance Scan 1155 (8.887 min): BF139850.D\data.ms




Abundance Scan 1155 (8.887 min): BF139850.D\data.ms (-1)



Abundance Scan 1172 (8.987 min): BF139850.D\data.ms



Abundance Scan 1172 (8.987 min): BF139850.D\data.ms (-1)



#37

2-Methylnaphthalene

Concen: 55.401 ng

RT: 8.887 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17

Instrument :

BNA_F


ClientSampleId :

SSTDICC060

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#38

1-Methylnaphthalene

Concen: 54.926 ng

RT: 8.987 min Scan# 1172

Delta R.T. 0.000 min

Lab File: BF139850.D


Acq: 18 Oct 2024 13:17







Tgt Ion:142 Resp: 1211981

Ion Ratio Lower Upper

142 100

141 91.6 73.5 110.3





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.928 min Scan# 1

Delta R.T. 0.001 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17

Instrument :

BNA_F


ClientSampleId :

SSTDICC060

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#40

1,2,4,5-Tetrachlorobenzene

Concen: 56.315 ng


RT: 9.051 min Scan# 1183







Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17

Tgt	Ion:216	Resp:	591461
Ion	Ratio	Lower	Upper
216	100		
214	79.6	63.4	95.0
179	21.5	17.2	25.8
108	19.6	15.6	23.4





#41

Hexachlorocyclopentadiene

Concen: 59.262 ng

RT: 9.039 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17


Instrument :

BNA_F

ClientSampleId :

SSTDICC060

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#42

2,4,6-Tribromophenol

Concen: 115.586 ng

RT: 10.722 min Scan# 1467

Delta R.T. 0.006 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17


Tgt Ion:330 Resp: 420891




Ion Ratio Lower Upper

330 100

332 95.9 78.1 117.1

141 32.0 26.6 39.8





#43

2,4,6-Trichlorophenol

Concen: 60.209 ng

RT: 9.157 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17




Instrument :


BNA_F




ClientSampleId :

SSTDICC060

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024

 #44
 2,4,5-Trichlorophenol
 Concen: 56.191 ng
 RT: 9.198 min Scan# 1208
 Delta R.T. 0.006 min
 Lab File: BF139850.D
 Acq: 18 Oct 2024 13:17

 Tgt Ion:196 Resp: 431080
 Ion Ratio Lower Upper
 196 100
 198 94.5 75.8 113.6
 97 50.5 41.8 62.6
 132 28.0 22.9 34.3




#45

2-Fluorobiphenyl

Concen: 105.106 ng

RT: 9.251 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17




Instrument :

BNA_F

ClientSampleId :

SSTDICC060

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#46

1,1'-Biphenyl

Concen: 54.396 ng

RT: 9.351 min Scan# 1234



Delta R.T. 0.000 min




Lab File: BF139850.D

Acq: 18 Oct 2024 13:17

Tgt Ion:154 Resp: 1474193

Ion	Ratio	Lower	Upper
154	100		
153	40.9	20.3	60.3
76	14.5	0.0	33.9





#47

2-Chloronaphthalene

Concen: 55.657 ng

RT: 9.375 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17




Instrument :

BNA_F

ClientSampleId :

SSTDICC060

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#48

2-Nitroaniline

Concen: 63.375 ng

RT: 9.469 min Scan# 1254

Delta R.T. 0.006 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17



Tgt Ion: 65 Resp: 423080




Ion Ratio Lower Upper

65 100

92 64.6 50.9 76.3

138 95.8 73.4 110.0





#49

Acenaphthylene

Concen: 55.718 ng

RT: 9.792 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17

Instrument :

BNA_F





ClientSampleId :

SSTDICC060

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#50

Dimethylphthalate

Concen: 56.037 ng

RT: 9.657 min Scan# 1286

Delta R.T. 0.006 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17


Tgt Ion:163 Resp: 1358812



Ion Ratio Lower Upper

163 100

194 4.0 3.4 5.0

164 10.1 8.2 12.2






#51
2,6-Dinitrotoluene
Concen: 60.963 ng
RT: 9.710 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDICC060


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#52
Acenaphthene
Concen: 56.242 ng
RT: 9.969 min Scan# 1339
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:154 Resp: 1148095
Ion Ratio Lower Upper
154 100
153 109.6 89.8 134.6
152 52.5 42.4 63.6






Abundance Scan 1324 (9.881 min): BF139850.D\data.ms



Abundance Scan 1324 (9.881 min): BF139850.D\data.ms (-1)



#53

3-Nitroaniline

Concen: 60.726 ng

RT: 9.881 min Scan# 1

Delta R.T. 0.006 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17

Instrument :

BNA_F


ClientSampleId :

SSTDICC060


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/21/2024


Supervised By :mohammad ahmed 10/21/2024



Abundance Scan 1341 (9.981 min): BF139850.D\data.ms



Abundance Scan 1341 (9.981 min): BF139850.D\data.ms (-1)



#54

2,4-Dinitrophenol

Concen: 61.938 ng

RT: 9.981 min Scan# 1341

Delta R.T. 0.007 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17

Tgt Ion:184 Resp: 131611

Ion Ratio Lower Upper

184 100

63 88.4 84.7 127.1

154 215.7 338.2 507.2#

Abundance Scan 1340 (9.974 min): BF139848.D\data.ms (-1)


Abundance Scan 1341 (9.981 min): BF139850.D\data.ms

Abundance Scan 1341 (9.981 min): BF139850.D\data.ms (-1)


Time-->

Time-->

Time-->






#55
Dibenzofuran
Concen: 54.819 ng
RT: 10.139 min Scan# 1
Instrument : BNA_F
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17
ClientSampleId : SSTDICC060




Tgt Ion:168 Resp: 160557:
Ion Ratio Lower Upper
168 100
139 39.7 31.8 47.6
169 14.4 11.6 17.4

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024






#56
4-Nitrophenol
Concen: 61.631 ng
RT: 10.028 min Scan# 1349
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17



Tgt Ion:139 Resp: 256728
Ion Ratio Lower Upper
139 100
109 74.0 53.6 93.6
65 108.1 87.8 127.8





#57

2,4-Dinitrotoluene

Concen: 63.558 ng

RT: 10.116 min Scan# 1

Delta R.T. 0.006 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17




Instrument :

BNA_F

ClientSampleId :

SSTDICC060

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#58

Fluorene

Concen: 53.435 ng

RT: 10.481 min Scan# 1426

Delta R.T. 0.006 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17





Tgt Ion:166 Resp: 1192018


Ion Ratio Lower Upper

166 100

165 98.0 78.5 117.7

167 13.5 10.9 16.3







#59
2,3,4,6-Tetrachlorophenol
Concen: 56.946 ng
RT: 10.251 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDICC060


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#60
Diethylphthalate
Concen: 56.954 ng
RT: 10.351 min Scan# 1404
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:149 Resp: 1355992
Ion Ratio Lower Upper
149 100
177 22.0 17.9 26.9
150 12.8 10.1 15.1







#61
4-Chlorophenyl-phenylether
Concen: 53.858 ng
RT: 10.469 min Scan# 1424
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Instrument : BNA_F
ClientSampleId : SSTDICC060


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#62
4-Nitroaniline
Concen: 61.507 ng
RT: 10.498 min Scan# 1429
Delta R.T. 0.012 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:138 Resp: 314525
Ion Ratio Lower Upper
138 100
92 53.4 35.7 75.7
108 78.2 57.5 97.5





#63

Azobenzene

Concen: 56.090 ng

RT: 10.633 min Scan# 141576

Delta R.T. 0.006 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17

Instrument :

BNA_F



ClientSampleId :

SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.416 min Scan# 1585

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17


Tgt Ion:188 Resp: 667372


Ion Ratio Lower Upper

188 100

94 10.0 7.9 11.9

80 10.9 9.0 13.4







#65
4,6-Dinitro-2-methylphenol
Concen: 67.951 ng
RT: 10.522 min Scan# 1433
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Instrument : BNA_F
ClientSampleId : SSTDICC060


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#66
n-Nitrosodiphenylamine
Concen: 56.214 ng
RT: 10.592 min Scan# 1445
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17


Tgt Ion:169 Resp: 1122831
Ion Ratio Lower Upper
169 100
168 62.5 49.8 74.6
167 34.8 28.0 42.0





#67
4-Bromophenyl-phenylether
Concen: 57.100 ng
RT: 10.963 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17




Instrument :
BNA_F
ClientSampleId :
SSTDICC060




Tgt Ion:248 Resp: 392570
Ion Ratio Lower Upper
248 100
250 98.7 79.0 118.6
141 72.9 61.7 92.5

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#68
Hexachlorobenzene
Concen: 57.187 ng
RT: 11.028 min Scan# 1519
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17



Tgt Ion:284 Resp: 442187
Ion Ratio Lower Upper
284 100
142 31.2 26.9 40.3
249 31.3 25.3 37.9






#69
Atrazine
Concen: 56.767 ng
RT: 11.116 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Instrument : BNA_F
ClientSampleId : SSTDICC060


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#70
Pentachlorophenol
Concen: 61.587 ng
RT: 11.216 min Scan# 1551
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:266 Resp: 289014
Ion Ratio Lower Upper
266 100
268 63.7 50.4 75.6
264 63.0 50.6 75.8







#71
Phenanthrene
Concen: 54.618 ng
RT: 11.439 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Instrument : BNA_F
ClientSampleId : SSTDICC060


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#72
Anthracene
Concen: 54.206 ng
RT: 11.492 min Scan# 1598
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:178 Resp: 1667221
Ion Ratio Lower Upper
178 100
176 19.6 15.3 22.9
179 15.9 12.4 18.6









#73
Carbazole
Concen: 54.015 ng
RT: 11.645 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Instrument : BNA_F
ClientSampleId : SSTDICC060


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#74
Di-n-butylphthalate
Concen: 55.069 ng
RT: 11.975 min Scan# 1680
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:149 Resp: 1819929
Ion Ratio Lower Upper
149 100
150 9.5 7.5 11.3
104 5.3 4.1 6.1







#75
Fluoranthene
Concen: 52.448 ng
RT: 12.627 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Instrument : BNA_F
ClientSampleId : SSTDICC060


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#76
Chrysene-d₁₂
Concen: 20.000 ng
RT: 14.057 min Scan# 2034
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:240 Resp: 337032
Ion Ratio Lower Upper
240 100
120 11.0 9.4 14.2
236 26.7 20.9 31.3







#77
Benzidine
Concen: 37.499 ng
RT: 12.745 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Instrument : BNA_F
ClientSampleId : SSTDICC060


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#78
Pyrene
Concen: 56.522 ng
RT: 12.857 min Scan# 1830
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:202 Resp: 1667489
Ion Ratio Lower Upper
202 100
200 21.9 17.2 25.8
203 18.0 14.2 21.4







#79
Terphenyl-d14
Concen: 109.563 ng
RT: 13.004 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Instrument : BNA_F
ClientSampleId : SSTDICC060


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#80
Butylbenzylphthalate
Concen: 60.719 ng
RT: 13.474 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:149 Resp: 537039
Ion Ratio Lower Upper
149 100
91 76.1 59.4 89.2
206 20.7 16.4 24.6





#81

Benzo(a)anthracene

Concen: 56.992 ng

RT: 14.045 min Scan# 2

Delta R.T. 0.006 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17





Instrument :

BNA_F

ClientSampleId :

SSTDICC060

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#82

3,3'-Dichlorobenzidine

Concen: 60.353 ng

RT: 14.004 min Scan# 2025

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17


Tgt Ion:252 Resp: 386072




Ion Ratio Lower Upper

252 100

254 64.3 50.8 76.2

126 11.2 8.9 13.3





#83

Chrysene

Concen: 57.855 ng

RT: 14.080 min Scan# 2

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17





Instrument :

BNA_F

ClientSampleId :

SSTDICC060

Manual Integrations
APPROVED

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#84

Bis(2-ethylhexyl)phthalate

Concen: 63.749 ng

RT: 14.033 min Scan# 2030

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17


Tgt Ion:149 Resp: 632594







Ion Ratio Lower Upper

149 100

167 26.7 21.4 32.0

279 3.9 3.0 4.4









#85

Di-n-octyl phthalate
Concen: 67.634 ng
RT: 14.651 min Scan# 2
Instrument : BNA_F
Delta R.T. -0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

ClientSampleId : SSTDICC060


Manual Integrations APPROVED







Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#86

Perylene-d12
Concen: 20.000 ng
RT: 15.527 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17





#87

Indeno(1,2,3-cd)pyrene

Concen: 61.817 ng

RT: 17.027 min Scan# 2

Delta R.T. 0.006 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17


Instrument :

BNA_F

ClientSampleId :

SSTDICC060

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#88

Benzo(b)fluoranthene

Concen: 61.033 ng

RT: 15.098 min Scan# 2211

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17


Tgt Ion:252 Resp: 1402395


Ion Ratio Lower Upper

252 100

253 22.2 17.6 26.4

125 10.1 7.9 11.9





#89

Benzo(k)fluoranthene

Concen: 53.053 ng

RT: 15.127 min Scan# 2

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17



Instrument :

BNA_F

ClientSampleId :

SSTDICC060

Manual Integrations
APPROVED

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#90

Benzo(a)pyrene


Concen: 59.764 ng


RT: 15.468 min Scan# 2274

Delta R.T. 0.000 min

Lab File: BF139850.D

Acq: 18 Oct 2024 13:17

 Tgt Ion:252 Resp: 1129942
 Ion Ratio Lower Upper
 252 100
 253 21.7 17.3 25.9
 125 11.0 8.7 13.1






#91
Dibenzo(a,h)anthracene
Concen: 60.667 ng
RT: 17.051 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Instrument : BNA_F
ClientSampleId : SSTDICC060


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#92
Benzo(g,h,i)perylene
Concen: 61.272 ng
RT: 17.486 min Scan# 2617
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:276 Resp: 1238955
Ion Ratio Lower Upper
276 100
277 23.6 19.0 28.6
138 21.4 17.4 26.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139851.D
 Acq On : 18 Oct 2024 13:46
 Operator : RC/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC080

Quant Time: Oct 18 14:21:12 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 14:11:20 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.892	152	187399	20.000	ng	0.00
21) Naphthalene-d8	8.175	136	693285	20.000	ng	0.00
39) Acenaphthene-d10	9.933	164	377723	20.000	ng	0.00
64) Phenanthrene-d10	11.416	188	633721	20.000	ng	0.00
76) Chrysene-d12	14.057	240	323513	20.000	ng	0.00
86) Perylene-d12	15.533	264	374491	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.504	112	1657822	138.491	ng	0.00
7) Phenol-d6	6.522	99	2147173	138.492	ng	0.01
23) Nitrobenzene-d5	7.463	82	1897374	151.683	ng	0.01
42) 2,4,6-Tribromophenol	10.722	330	522350	147.845	ng	0.00
45) 2-Fluorobiphenyl	9.251	172	2941186	128.689	ng	0.00
79) Terphenyl-d14	13.004	244	2634676	132.704	ng	0.00
Target Compounds						
				Qvalue		
2) 1,4-Dioxane	2.693	88	410806	73.604	ng	99
3) Pyridine	3.452	79	993777	71.834	ng	99
4) n-Nitrosodimethylamine	3.422	42	567647	76.370	ng	98
6) Aniline	6.598	93	992375m	68.680	ng	
8) 2-Chlorophenol	6.681	128	836233	68.333	ng	98
9) Benzaldehyde	6.445	77	554983	63.632	ng	99
10) Phenol	6.534	94	1125558	69.172	ng	95
11) bis(2-Chloroethyl)ether	6.634	93	886786	71.780	ng	99
12) 1,3-Dichlorobenzene	6.834	146	969627	69.024	ng	98
13) 1,4-Dichlorobenzene	6.910	146	968088	68.978	ng	99
14) 1,2-Dichlorobenzene	7.069	146	861333	65.758	ng	99
15) Benzyl Alcohol	7.034	79	803101	70.616	ng	98
16) 2,2'-oxybis(1-Chloropr...	7.169	45	1472196	69.886	ng	99
17) 2-Methylphenol	7.140	107	752459	72.109	ng	99
18) Hexachloroethane	7.410	117	357637	71.459	ng	97
19) n-Nitroso-di-n-propyla...	7.316	70	651048	69.238	ng	99
20) 3+4-Methylphenols	7.298	107	878987	65.856	ng	# 85
22) Acetophenone	7.310	105	1149288	67.682	ng	98
24) Nitrobenzene	7.481	77	1025571	74.779	ng	99
25) Isophorone	7.722	82	1767234	74.436	ng	100
26) 2-Nitrophenol	7.792	139	434755	84.029	ng	99
27) 2,4-Dimethylphenol	7.822	122	619136	71.765	ng	99
28) bis(2-Chloroethoxy)met...	7.928	93	1022457	71.081	ng	100
29) 2,4-Dichlorophenol	8.028	162	711892	72.445	ng	99
30) 1,2,4-Trichlorobenzene	8.116	180	774108	71.194	ng	99
31) Naphthalene	8.198	128	2427578	67.894	ng	99
32) Benzoic acid	7.963	122	651887	86.634	ng	99
33) 4-Chloroaniline	8.251	127	849601	69.684	ng	99
34) Hexachlorobutadiene	8.316	225	490203	71.753	ng	99
35) Caprolactam	8.634	113	239184	76.550	ng	99
36) 4-Chloro-3-methylphenol	8.722	107	794640	73.031	ng	99
37) 2-Methylnaphthalene	8.886	142	1487841	67.944	ng	100
38) 1-Methylnaphthalene	8.986	142	1457890	67.858	ng	99
40) 1,2,4,5-Tetrachloroben...	9.051	216	712936	69.962	ng	100
41) Hexachlorocyclopentadiene	9.039	237	256007	72.201	ng	100
43) 2,4,6-Trichlorophenol	9.163	196	535533	74.228	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139851.D
 Acq On : 18 Oct 2024 13:46
 Operator : RC/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC080

Quant Time: Oct 18 14:21:12 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 14:11:20 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.198	196	542663	72.904	ng	99
46) 1,1'-Biphenyl	9.357	154	1753952	66.703	ng	99
47) 2-Chloronaphthalene	9.381	162	1470012	69.411	ng	99
48) 2-Nitroaniline	9.469	65	527552	81.447	ng	99
49) Acenaphthylene	9.792	152	2106357	68.796	ng	99
50) Dimethylphthalate	9.657	163	1676969	71.277	ng	99
51) 2,6-Dinitrotoluene	9.716	165	393234	77.193	ng	99
52) Acenaphthene	9.969	154	1378724	69.610	ng	99
53) 3-Nitroaniline	9.886	138	387157	73.697	ng	96
54) 2,4-Dinitrophenol	9.986	184	169167	79.932	ng	# 1
55) Dibenzofuran	10.139	168	1930402	67.930	ng	99
56) 4-Nitrophenol	10.033	139	314170	77.732	ng	99
57) 2,4-Dinitrotoluene	10.116	165	510703	81.523	ng	97
58) Fluorene	10.480	166	1426794	65.920	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.251	232	424598	72.764	ng	99
60) Diethylphthalate	10.357	149	1631842	70.641	ng	100
61) 4-Chlorophenyl-phenyle...	10.475	204	730565	66.837	ng	100
62) 4-Nitroaniline	10.504	138	383383	77.270	ng	98
63) Azobenzene	10.633	77	1727533	70.539	ng	99
65) 4,6-Dinitro-2-methylph...	10.528	198	235323	91.037	ng	98
66) n-Nitrosodiphenylamine	10.592	169	1350839	71.220	ng	99
67) 4-Bromophenyl-phenylether	10.963	248	477995	73.216	ng	99
68) Hexachlorobenzene	11.028	284	536511	73.070	ng	98
69) Atrazine	11.116	200	382294	74.406	ng	99
70) Pentachlorophenol	11.216	266	354105	79.464	ng	99
71) Phenanthrene	11.445	178	2045577	68.336	ng	100
72) Anthracene	11.492	178	1995750	68.333	ng	99
73) Carbazole	11.645	167	1818340	66.943	ng	99
74) Di-n-butylphthalate	11.980	149	2153517	68.623	ng	100
75) Fluoranthene	12.627	202	1955847	64.632	ng	99
77) Benzidine	12.745	184	318678	60.567	ng	99
78) Pyrene	12.857	202	1935829	68.360	ng	99
80) Butylbenzylphthalate	13.474	149	664532	78.274	ng	98
81) Benzo(a)anthracene	14.045	228	1512920	71.840	ng	99
82) 3,3'-Dichlorobenzidine	14.010	252	496976	80.937	ng	98
83) Chrysene	14.080	228	1424522	73.774	ng	99
84) Bis(2-ethylhexyl)phtha...	14.033	149	794387	83.399	ng	99
85) Di-n-octyl phthalate	14.651	149	1534331	88.561	ng	100
87) Indeno(1,2,3-cd)pyrene	17.033	276	1877616	77.935	ng	99
88) Benzo(b)fluoranthene	15.098	252	1664856	72.980	ng	99
89) Benzo(k)fluoranthene	15.133	252	1419158	72.134	ng	99
90) Benzo(a)pyrene	15.474	252	1419026	75.598	ng	100
91) Dibenzo(a,h)anthracene	17.057	278	1552091	77.226	ng	99
92) Benzo(g,h,i)perylene	17.492	276	1550622	77.240	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139851.D
 Acq On : 18 Oct 2024 13:46
 Operator : RC/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1


Quant Time: Oct 18 14:21:12 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 14:11:20 2024
 Response via : Initial Calibration

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC080


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024








#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.892 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46




Tgt Ion:152 Resp: 187395
Ion Ratio Lower Upper
152 100
150 161.2 130.2 195.2
115 64.4 51.4 77.2

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024






#2
1,4-Dioxane
Concen: 73.604 ng
RT: 2.693 min Scan# 102
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46



Tgt Ion: 88 Resp: 410806
Ion Ratio Lower Upper
88 100
58 91.2 73.3 109.9
43 33.5 27.1 40.7









#3

Pyridine
Concen: 71.834 ng
RT: 3.452 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#4

n-Nitrosodimethylamine
Concen: 76.370 ng
RT: 3.422 min Scan# 226
Delta R.T. 0.024 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46


Tgt Ion: 42 Resp: 567647
Ion Ratio Lower Upper
42 100
74 117.8 96.2 144.4
44 7.1 6.2 9.2





#5
2-Fluorophenol
Concen: 138.491 ng
RT: 5.504 min Scan# 5
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46




Instrument : BNA_F
ClientSampleId : SSTDICC080




Tgt Ion:112 Resp: 165782
Ion Ratio Lower Upper
112 100
64 67.0 53.0 79.6
63 35.0 27.0 40.4

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#6
Aniline
Concen: 68.680 ng m
RT: 6.598 min Scan# 766
Delta R.T. 0.047 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46



Tgt Ion: 93 Resp: 992375
Ion Ratio Lower Upper
93 100
66 41.8 34.0 51.0
65 21.1 17.4 26.0







#7
Phenol-d6
Concen: 138.492 ng
RT: 6.522 min Scan# 7
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Instrument : BNA_F
ClientSampleId : SSTDICC080


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#8
2-Chlorophenol
Concen: 68.333 ng
RT: 6.681 min Scan# 780
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:128 Resp: 836233
Ion Ratio Lower Upper
128 100
130 32.8 12.3 52.3
64 56.2 34.6 74.6







#9
Benzaldehyde
Concen: 63.632 ng
RT: 6.445 min Scan# 7
Instrument : BNA_F
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

ClientSampleId : SSTDICC080


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#10
Phenol
Concen: 69.172 ng
RT: 6.534 min Scan# 755
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46


Tgt Ion: 94 Resp: 1125558
Ion Ratio Lower Upper
94 100
65 33.1 11.1 51.1
66 47.1 22.7 62.7





#11
bis(2-Chloroethyl)ether
Concen: 71.780 ng
RT: 6.634 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46




Instrument : BNA_F
ClientSampleId : SSTDICC080




Tgt Ion: 93 Resp: 886780
Ion Ratio Lower Upper
93 100
63 85.6 65.0 105.0
95 32.2 12.4 52.4

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024









#12
1,3-Dichlorobenzene
Concen: 69.024 ng
RT: 6.834 min Scan# 806
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46



Tgt Ion:146 Resp: 969627
Ion Ratio Lower Upper
146 100
148 64.1 51.0 76.4
75 35.0 26.4 39.6





#13

1,4-Dichlorobenzene

Concen: 68.978 ng

RT: 6.910 min Scan# 8

Delta R.T. 0.000 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC080

Tgt Ion:146 Resp: 968083

Ion Ratio Lower Upper

146 100

148 64.0 51.8 77.8


111 44.3 34.6 51.8

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#14

1,2-Dichlorobenzene

Concen: 65.758 ng

RT: 7.069 min Scan# 846

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46


Tgt Ion:146 Resp: 861333







Ion Ratio Lower Upper

146 100

148 63.5 51.4 77.2

111 46.5 36.7 55.1





#15




Benzyl Alcohol
Concen: 70.616 ng




RT: 7.034 min Scan# 8

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46

Instrument :
BNA_F
ClientSampleId :
SSTDICC080**Manual Integrations
APPROVED**Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#17

2-Methylphenol

Concen: 72.109 ng

RT: 7.140 min Scan# 8

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46




Instrument :

BNA_F

ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#18

Hexachloroethane

Concen: 71.459 ng

RT: 7.410 min Scan# 904

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46



Tgt Ion:117 Resp: 357637


Ion Ratio Lower Upper

117 100

119 95.6 77.0 115.4

201 99.9 76.2 114.4







#19
n-Nitroso-di-n-propylamine
Concen: 69.238 ng
RT: 7.316 min Scan# 8
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Instrument : BNA_F
ClientSampleId : SSTDICC080


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#20
3+4-Methylphenols
Concen: 65.856 ng
RT: 7.298 min Scan# 885
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:107 Resp: 878987
Ion Ratio Lower Upper
107 100
108 91.7 69.8 109.8
77 100.5 50.5 90.5#
79 31.7 10.2 50.2





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.175 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46




Instrument :

BNA_F

ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#22

Acetophenone

Concen: 67.682 ng

RT: 7.310 min Scan# 887

Delta R.T. 0.012 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46

Tgt Ion:105 Resp: 1149288


Ion Ratio Lower Upper




105 100

71 7.6 5.9 8.9

51 35.3 29.4 44.2

120 21.8 17.5 26.3





#23

Nitrobenzene-d5

Concen: 151.683 ng

RT: 7.463 min Scan# 913

Delta R.T. 0.012 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46




Instrument :

BNA_F

ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#24

Nitrobenzene

Concen: 74.779 ng

RT: 7.481 min Scan# 916

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46


Tgt Ion: 77 Resp: 1025571


Ion Ratio Lower Upper

77 100

123 44.0 35.7 53.5


65 14.6 11.5 17.3





#25
Isophorone
Concen: 74.436 ng
RT: 7.722 min Scan# 9
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46



Instrument : BNA_F
ClientSampleId : SSTDICC080




Tgt Ion: 82 Resp: 1767234
Ion Ratio Lower Upper
82 100
95 7.2 5.5 8.3
138 16.5 13.2 19.8

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024






#26
2-Nitrophenol
Concen: 84.029 ng
RT: 7.792 min Scan# 969
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46



Tgt Ion:139 Resp: 434755
Ion Ratio Lower Upper
139 100
109 31.6 25.1 37.7
65 57.8 47.5 71.3





#27

2,4-Dimethylphenol

Concen: 71.765 ng

RT: 7.822 min Scan# 9

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46

Instrument :

BNA_F




ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#28

bis(2-Chloroethoxy)methane

Concen: 71.081 ng

RT: 7.928 min Scan# 992

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46


Tgt Ion: 93 Resp: 1022457







Ion Ratio Lower Upper

93 100

95 32.7 26.0 39.0

123 12.2 9.8 14.8





#29

2,4-Dichlorophenol

Concen: 72.445 ng

RT: 8.028 min Scan# 1

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC080

Tgt Ion:162 Resp: 71189

Ion Ratio Lower Upper

162 100


164 64.5 43.9 83.9

98 40.8 19.9 59.9

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#30

1,2,4-Trichlorobenzene

Concen: 71.194 ng

RT: 8.116 min Scan# 1024

Delta R.T. 0.000 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46


Tgt Ion:180 Resp: 774108




Ion Ratio Lower Upper

180 100

182 99.8 80.5 120.7

145 33.0 25.4 38.0





#31

Naphthalene

Concen: 67.894 ng

RT: 8.198 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46

Instrument :

BNA_F





ClientSampleId :

SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#32

Benzoic acid

Concen: 86.634 ng

RT: 7.963 min Scan# 998

Delta R.T. 0.041 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46


Tgt Ion:122 Resp: 651887







Ion Ratio Lower Upper

122 100

105 126.8 107.6 147.6

77 98.1 80.2 120.2





#33

4-Chloroaniline

Concen: 69.684 ng

RT: 8.251 min Scan# 1

Delta R.T. 0.012 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46

Instrument :

BNA_F


ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**
Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024

Tgt Ion:127 Resp: 849601

Ion	Ratio	Lower	Upper
127	100		
129	32.5	25.4	38.2
65	35.3	28.4	42.6
92	20.7	16.6	25.0



#34

Hexachlorobutadiene

Concen: 71.753 ng

RT: 8.316 min Scan# 1058


Delta R.T. 0.000 min


Lab File: BF139851.D

Acq: 18 Oct 2024 13:46



Tgt Ion:225 Resp: 490203

Ion	Ratio	Lower	Upper
225	100		
223	63.2	50.6	76.0
227	64.0	51.8	77.8







Abundance Scan 1112 (8.634 min): BF139851.D\data.ms



Abundance Scan 1127 (8.722 min): BF139851.D\data.ms



#35

Caprolactam

Concen: 76.550 ng

RT: 8.634 min Scan# 1

Delta R.T. 0.024 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC080

Tgt Ion:113 Resp: 239184

Ion Ratio Lower Upper

113 100


55 206.7 189.2 229.2

56 152.7 132.5 172.5

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#36

4-Chloro-3-methylphenol

Concen: 73.031 ng

RT: 8.722 min Scan# 1127

Delta R.T. 0.012 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46


Tgt Ion:107 Resp: 794640




Ion Ratio Lower Upper

107 100

144 24.4 19.1 28.7

142 72.3 57.2 85.8





#37

2-Methylnaphthalene

Concen: 67.944 ng

RT: 8.886 min Scan# 1155

Delta R.T. 0.000 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46





Instrument :

BNA_F

ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#38

1-Methylnaphthalene


Concen: 67.858 ng







RT: 8.986 min Scan# 1172

Delta R.T. 0.000 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46

 Tgt Ion:142 Resp: 1457890
 Ion Ratio Lower Upper
 142 100
 141 91.4 73.5 110.3




#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.933 min Scan# 1

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46

Instrument :

BNA_F


ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#40

1,2,4,5-Tetrachlorobenzene

Concen: 69.962 ng


RT: 9.051 min Scan# 1183




Delta R.T. 0.000 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46

Tgt	Ion:216	Resp:	712936
Ion	Ratio	Lower	Upper
216	100		
214	79.7	63.4	95.0
179	21.7	17.2	25.8
108	19.5	15.6	23.4





#41

Hexachlorocyclopentadiene

Concen: 72.201 ng

RT: 9.039 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46





Instrument :

BNA_F

ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#42

2,4,6-Tribromophenol

Concen: 147.845 ng

RT: 10.722 min Scan# 1467

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46


Tgt Ion:330 Resp: 522350







Ion Ratio Lower Upper

330 100

332 96.0 78.1 117.1

141 31.2 26.6 39.8





#43

2,4,6-Trichlorophenol

Concen: 74.228 ng

RT: 9.163 min Scan# 1

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46


Instrument :

BNA_F

ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#44

2,4,5-Trichlorophenol

Concen: 72.904 ng


RT: 9.198 min Scan# 1208




Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46

Tgt	Ion:196	Resp:	542663
Ion	Ratio	Lower	Upper
196	100		
198	95.2	75.8	113.6
97	54.3	41.8	62.6
132	29.0	22.9	34.3





#45

2-Fluorobiphenyl

Concen: 128.689 ng

RT: 9.251 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46





Instrument :

BNA_F

ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#46

1,1'-Biphenyl

Concen: 66.703 ng

RT: 9.357 min Scan# 1235

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46


Tgt Ion:154 Resp: 1753952


Ion Ratio Lower Upper

154 100

153 41.3 20.3 60.3

76 13.8 0.0 33.9





#47

2-Chloronaphthalene

Concen: 69.411 ng

RT: 9.381 min Scan# 1

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46



Instrument :

BNA_F

ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#48

2-Nitroaniline

Concen: 81.447 ng

RT: 9.469 min Scan# 1254

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46


Tgt Ion: 65 Resp: 527552




Ion Ratio Lower Upper

65 100

92 64.1 50.9 76.3

138 93.3 73.4 110.0





#49

Acenaphthylene

Concen: 68.796 ng

RT: 9.792 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46

Instrument :

BNA_F





ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#50

Dimethylphthalate

Concen: 71.277 ng

RT: 9.657 min Scan# 1286

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46


Tgt Ion:163 Resp: 1676969


Ion Ratio Lower Upper

163 100

194 4.2 3.4 5.0

164 10.5 8.2 12.2







#51
2,6-Dinitrotoluene
Concen: 77.193 ng
RT: 9.716 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Instrument : BNA_F
ClientSampleId : SSTDICC080


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#52
Acenaphthene
Concen: 69.610 ng
RT: 9.969 min Scan# 1339
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:154 Resp: 1378724
Ion Ratio Lower Upper
154 100
153 110.4 89.8 134.6
152 53.4 42.4 63.6





#53

3-Nitroaniline

Concen: 73.697 ng

RT: 9.886 min Scan# 1

Delta R.T. 0.012 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46

Instrument :

BNA_F





ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#54

2,4-Dinitrophenol

Concen: 79.932 ng

RT: 9.986 min Scan# 1342

Delta R.T. 0.012 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46


Tgt Ion:184 Resp: 169167




Ion Ratio Lower Upper

184 100

63 77.7 84.7 127.1#

154 102.0 338.2 507.2#





#55

Dibenzofuran

Concen: 67.930 ng

RT: 10.139 min Scan# 1

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46





Instrument :

BNA_F

ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#56

4-Nitrophenol


Concen: 77.732 ng


RT: 10.033 min Scan# 1350

Delta R.T. 0.012 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46







#57
2,4-Dinitrotoluene
Concen: 81.523 ng
RT: 10.116 min Scan# 1 Instrument :
Delta R.T. 0.006 min BNA_F
Lab File: BF139851.D ClientSampleId :
Acq: 18 Oct 2024 13:46 SSTDICC080


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



```
#58  
Fluorene  
Concen: 65.920 ng  
RT: 10.480 min Scan# 1426  
Delta R.T. 0.006 min  
Lab File: BF139851.D  
Acq: 18 Oct 2024 13:46
```

Tgt	Ion:166	Resp:	1426794
Ion	Ratio	Lower	Upper
166	100		
165	98.6	78.5	117.7
167	13.5	10.9	16.3






#59
2,3,4,6-Tetrachlorophenol
Concen: 72.764 ng
RT: 10.251 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Instrument :
BNA_F
ClientSampleId :
SSTDICC080


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#60
Diethylphthalate
Concen: 70.641 ng
RT: 10.357 min Scan# 1405
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:149 Resp: 1631842
Ion Ratio Lower Upper
149 100
177 22.2 17.9 26.9
150 12.8 10.1 15.1






#61
4-Chlorophenyl-phenylether
Concen: 66.837 ng
RT: 10.475 min Scan# 1425
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Instrument : BNA_F
ClientSampleId : SSTDICC080


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#62
4-Nitroaniline
Concen: 77.270 ng
RT: 10.504 min Scan# 1430
Delta R.T. 0.018 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:138 Resp: 383383
Ion Ratio Lower Upper
138 100
92 53.7 35.7 75.7
108 78.3 57.5 97.5





#63

Azobenzene

Concen: 70.539 ng

RT: 10.633 min Scan# 1452

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46

Instrument :

BNA_F


ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.416 min Scan# 1585

Delta R.T. 0.000 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46


Tgt Ion:188 Resp: 633721


Ion Ratio Lower Upper

188 100

94 10.7 7.9 11.9

80 11.1 9.0 13.4







#65
4,6-Dinitro-2-methylphenol
Concen: 91.037 ng
RT: 10.528 min Scan# 1434
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Instrument : BNA_F
ClientSampleId : SSTDICC080


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#66
n-Nitrosodiphenylamine
Concen: 71.220 ng
RT: 10.592 min Scan# 1445
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:169 Resp: 1350839
Ion Ratio Lower Upper
169 100
168 63.4 49.8 74.6
167 35.5 28.0 42.0









#67
4-Bromophenyl-phenylether
Concen: 73.216 ng
RT: 10.963 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Instrument :
BNA_F
ClientSampleId :
SSTDICC080


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#68
Hexachlorobenzene
Concen: 73.070 ng
RT: 11.028 min Scan# 1519
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:284 Resp: 536511
Ion Ratio Lower Upper
284 100
142 31.9 26.9 40.3
249 31.3 25.3 37.9







#69
Atrazine
Concen: 74.406 ng
RT: 11.116 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Instrument : BNA_F
ClientSampleId : SSTDICC080


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#70
Pentachlorophenol
Concen: 79.464 ng
RT: 11.216 min Scan# 1551
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:266 Resp: 354105
Ion Ratio Lower Upper
266 100
268 63.1 50.4 75.6
264 62.5 50.6 75.8





#71

Phenanthrene

Concen: 68.336 ng

RT: 11.445 min Scan# 1

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46

Instrument :

BNA_F



ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#72

Anthracene

Concen: 68.333 ng


RT: 11.492 min Scan# 1598




Delta R.T. 0.000 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46

Tgt	Ion:178	Resp:	1995750
Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.3	22.9
179	16.0	12.4	18.6









#73
Carbazole
Concen: 66.943 ng
RT: 11.645 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Instrument : BNA_F
ClientSampleId : SSTDICC080


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#74
Di-n-butylphthalate
Concen: 68.623 ng
RT: 11.980 min Scan# 1681
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:149 Resp: 2153517
Ion Ratio Lower Upper
149 100
150 9.5 7.5 11.3
104 5.2 4.1 6.1







#75
Fluoranthene
Concen: 64.632 ng
RT: 12.627 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Instrument : BNA_F
ClientSampleId : SSTDICC080


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.057 min Scan# 2034
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:240 Resp: 323513
Ion Ratio Lower Upper
240 100
120 10.6 9.4 14.2
236 26.1 20.9 31.3





#77

Benzidine

Concen: 60.567 ng

RT: 12.745 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46

Instrument :

BNA_F




ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#78

Pyrene

Concen: 68.360 ng

RT: 12.857 min Scan# 1830

Delta R.T. 0.000 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46



Tgt Ion:202 Resp: 1935829


Ion Ratio Lower Upper

202 100

200 21.7 17.2 25.8

203 18.2 14.2 21.4







#79
Terphenyl-d14
Concen: 132.704 ng
RT: 13.004 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Instrument : BNA_F
ClientSampleId : SSTDICC080


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#80
Butylbenzylphthalate
Concen: 78.274 ng
RT: 13.474 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:149 Resp: 664532
Ion Ratio Lower Upper
149 100
91 76.5 59.4 89.2
206 20.5 16.4 24.6





#81

Benzo(a)anthracene

Concen: 71.840 ng

RT: 14.045 min Scan# 2

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46





Instrument :

BNA_F

ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#82

3,3'-Dichlorobenzidine

Concen: 80.937 ng


RT: 14.010 min Scan# 2026




Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46

Tgt	Ion	Ion Ratio	Resp:	Lower	Upper
252	100				
254	65.7	50.8	496976	76.2	
126	11.0	8.9		13.3	





#83

Chrysene

Concen: 73.774 ng

RT: 14.080 min Scan# 2

Delta R.T. 0.000 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46

Instrument :

BNA_F





ClientSampleId :

SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024

Supervised By :mohammad ahmed 10/21/2024



#84

Bis(2-ethylhexyl)phthalate

Concen: 83.399 ng

RT: 14.033 min Scan# 2030

Delta R.T. 0.000 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46


Tgt Ion:149 Resp: 794387




Ion Ratio Lower Upper

149 100

167 26.3 21.4 32.0

279 4.0 3.0 4.4





#85

Di-n-octyl phthalate

Concen: 88.561 ng

RT: 14.651 min Scan# 2

Delta R.T. -0.000 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46





Instrument :

BNA_F

ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#86

Perylene-d₁₂


Concen: 20.000 ng







RT: 15.533 min Scan# 2285

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46





#87

Indeno(1,2,3-cd)pyrene

Concen: 77.935 ng

RT: 17.033 min Scan# 2

Delta R.T. 0.012 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46


Instrument :

BNA_F

ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#88

Benzo(b)fluoranthene

Concen: 72.980 ng

RT: 15.098 min Scan# 2211

Delta R.T. 0.000 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46


Tgt Ion:252 Resp: 1664856




Ion Ratio Lower Upper

252 100

253 22.4 17.6 26.4

125 10.4 7.9 11.9





#89

Benzo(k)fluoranthene

Concen: 72.134 ng

RT: 15.133 min Scan# 2

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46





Instrument :

BNA_F

ClientSampleId :

SSTDICC080

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/21/2024
 Supervised By :mohammad ahmed 10/21/2024


#90

Benzo(a)pyrene

Concen: 75.598 ng

RT: 15.474 min Scan# 2275

Delta R.T. 0.006 min

Lab File: BF139851.D

Acq: 18 Oct 2024 13:46


Tgt Ion:252 Resp: 1419026


Ion Ratio Lower Upper

252 100

253 21.8 17.3 25.9

125 10.7 8.7 13.1







#91
Dibenzo(a,h)anthracene
Concen: 77.226 ng
RT: 17.057 min Scan# 2
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Instrument : BNA_F
ClientSampleId : SSTDICC080


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#92
Benzo(g,h,i)perylene
Concen: 77.240 ng
RT: 17.492 min Scan# 2618
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:276 Resp: 1550622
Ion Ratio Lower Upper
276 100
277 23.9 19.0 28.6
138 21.6 17.4 26.0





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>CHEMTECH</u>		Contract:	<u>PORT06</u>	
Lab Code:	<u>CHEM</u>	Case No.:	<u>P4578</u>	SAS No.:	<u>P4578</u>
Instrument ID:	<u>BNA_F</u>		Calibration Date/Time:	<u>10/28/2024</u>	<u>09:32</u>
Lab File ID:	<u>BF140075.D</u>		Init. Calib. Date(s):	<u>10/18/2024</u>	<u>10/18/2024</u>
EPA Sample No.:	<u>SSTDCCC040</u>		Init. Calib. Time(s):	<u>10:27</u>	<u>13:46</u>
GC Column:	<u>DB-UI</u>	ID: <u>0.18</u>	(mm)		

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.476	1.412		-4.3	
2-Fluorophenol	1.278	1.198		-6.3	
Phenol-d6	1.655	1.537		-7.1	
1,4-Dichlorobenzene	1.498	1.486		-0.8	20.0
2-Methylphenol	1.114	1.050		-5.7	
3+4-Methylphenols	1.424	1.320		-7.3	
Nitrobenzene-d5	0.361	0.377		4.4	
Hexachloroethane	0.534	0.526		-1.5	
Nitrobenzene	0.396	0.397		0.3	
Hexachlorobutadiene	0.197	0.200		1.5	20.0
2,4,6-Trichlorophenol	0.382	0.375		-1.8	20.0
2-Fluorobiphenyl	1.210	1.195		-1.2	
2,4,5-Trichlorophenol	0.394	0.392		-0.5	
2,4-Dinitrotoluene	0.332	0.382		15.1	
2,4,6-Tribromophenol	0.187	0.197		5.3	
Hexachlorobenzene	0.232	0.233		0.4	
Pentachlorophenol	0.141	0.144		2.1	20.0
Terphenyl-d14	1.227	1.250		1.9	

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140075.D
 Acq On : 28 Oct 2024 09:32
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Oct 28 09:55:45 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.892	152	132517	20.000	ng	0.00
21) Naphthalene-d8	8.175	136	497360	20.000	ng	0.00
39) Acenaphthene-d10	9.927	164	274065	20.000	ng	0.00
64) Phenanthrene-d10	11.416	188	494703	20.000	ng	0.00
76) Chrysene-d12	14.051	240	250901	20.000	ng	0.00
86) Perylene-d12	15.527	264	263501	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.504	112	635250	75.045	ng	0.00
7) Phenol-d6	6.516	99	814577	74.300	ng	0.00
23) Nitrobenzene-d5	7.451	82	750063	83.584	ng	0.00
42) 2,4,6-Tribromophenol	10.716	330	216066	84.285	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1310048	79.000	ng	0.00
79) Terphenyl-d14	12.998	244	1254654	81.484	ng	0.00
Target Compounds						
				Qvalue		
2) 1,4-Dioxane	2.669	88	148154	37.538	ng	96
3) Pyridine	3.440	79	374133	38.244	ng	98
4) n-Nitrosodimethylamine	3.387	42	197677	37.610	ng	96
6) Aniline	6.551	93	388187	37.992	ng	99
8) 2-Chlorophenol	6.675	128	328283	37.936	ng	100
9) Benzaldehyde	6.439	77	234409	38.007	ng	99
10) Phenol	6.528	94	427262m	37.802	ng	
11) bis(2-Chloroethyl)ether	6.628	93	324860	37.186	ng	100
12) 1,3-Dichlorobenzene	6.833	146	395383	39.802	ng	99
13) 1,4-Dichlorobenzene	6.910	146	393761	39.676	ng	99
14) 1,2-Dichlorobenzene	7.063	146	364568	39.360	ng	99
15) Benzyl Alcohol	7.028	79	307563	38.244	ng	98
16) 2,2'-oxybis(1-Chloropr...	7.163	45	538199	36.130	ng	98
17) 2-Methylphenol	7.133	107	278276	37.712	ng	99
18) Hexachloroethane	7.404	117	139330	39.369	ng	98
19) n-Nitroso-di-n-propyla...	7.304	70	243835	36.671	ng	98
20) 3+4-Methylphenols	7.292	107	349881	37.070	ng	# 83
22) Acetophenone	7.298	105	466951	38.331	ng	99
24) Nitrobenzene	7.475	77	394888	40.136	ng	100
25) Isophorone	7.710	82	654467	38.425	ng	99
26) 2-Nitrophenol	7.786	139	151084	40.704	ng	98
27) 2,4-Dimethylphenol	7.816	122	220819	35.678	ng	97
28) bis(2-Chloroethoxy)met...	7.922	93	392131	38.000	ng	100
29) 2,4-Dichlorophenol	8.028	162	277268	39.331	ng	99
30) 1,2,4-Trichlorobenzene	8.116	180	313301	40.164	ng	100
31) Naphthalene	8.198	128	1010972	39.413	ng	100
32) Benzoic acid	7.928	122	184222	34.127	ng	99
33) 4-Chloroaniline	8.239	127	341100	38.998	ng	99
34) Hexachlorobutadiene	8.310	225	198499	40.501	ng	100
35) Caprolactam	8.616	113	88634	39.542	ng	92
36) 4-Chloro-3-methylphenol	8.716	107	305431	39.128	ng	98
37) 2-Methylnaphthalene	8.886	142	616924	39.271	ng	99
38) 1-Methylnaphthalene	8.986	142	607270	39.400	ng	100
40) 1,2,4,5-Tetrachloroben...	9.051	216	303883	41.099	ng	99
41) Hexachlorocyclopentadiene	9.039	237	99773	38.781	ng	100
43) 2,4,6-Trichlorophenol	9.157	196	205507	39.258	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140075.D
 Acq On : 28 Oct 2024 09:32
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Oct 28 09:55:45 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.198	196	214705	39.754	ng	98
46) 1,1'-Biphenyl	9.351	154	756187	39.635	ng	100
47) 2-Chloronaphthalene	9.375	162	620778	40.398	ng	99
48) 2-Nitroaniline	9.469	65	201129	42.796	ng	94
49) Acenaphthylene	9.792	152	891367	40.124	ng	99
50) Dimethylphthalate	9.651	163	682031	39.953	ng	99
51) 2,6-Dinitrotoluene	9.710	165	161527	43.701	ng	98
52) Acenaphthene	9.963	154	579769	40.343	ng	99
53) 3-Nitroaniline	9.874	138	160653	42.148	ng	98
54) 2,4-Dinitrophenol	9.980	184	64847	45.309	ng	# 1
55) Dibenzofuran	10.133	168	813269	39.443	ng	99
56) 4-Nitrophenol	10.027	139	120058	40.940	ng	97
57) 2,4-Dinitrotoluene	10.110	165	209655	46.125	ng	99
58) Fluorene	10.474	166	633488	40.338	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.251	232	170444	40.257	ng	98
60) Diethylphthalate	10.345	149	675530	40.303	ng	100
61) 4-Chlorophenyl-phenyle...	10.469	204	324430	40.907	ng	99
62) 4-Nitroaniline	10.492	138	156727	43.535	ng	98
63) Azobenzene	10.627	77	685497	38.577	ng	98
65) 4,6-Dinitro-2-methylph...	10.522	198	96169	47.659	ng	92
66) n-Nitrosodiphenylamine	10.586	169	569092	38.436	ng	100
67) 4-Bromophenyl-phenylether	10.957	248	205221	40.268	ng	98
68) Hexachlorobenzene	11.021	284	230272	40.175	ng	99
69) Atrazine	11.110	200	155137	38.680	ng	99
70) Pentachlorophenol	11.216	266	142153	40.865	ng	98
71) Phenanthrene	11.439	178	899645	38.500	ng	99
72) Anthracene	11.492	178	889039	38.994	ng	100
73) Carbazole	11.645	167	803004	37.870	ng	99
74) Di-n-butylphthalate	11.974	149	957414	39.082	ng	99
75) Fluoranthene	12.627	202	901406	38.158	ng	99
77) Benzidine	12.745	184	186877	45.296	ng	99
78) Pyrene	12.857	202	894517	40.730	ng	99
80) Butylbenzylphthalate	13.468	149	271701	41.265	ng	99
81) Benzo(a)anthracene	14.039	228	665269	40.732	ng	99
82) 3,3'-Dichlorobenzidine	14.004	252	211560	44.426	ng	99
83) Chrysene	14.080	228	587960	39.262	ng	100
84) Bis(2-ethylhexyl)phtha...	14.027	149	334928	45.339	ng	99
85) Di-n-octyl phthalate	14.639	149	547586	40.754	ng	98
87) Indeno(1,2,3-cd)pyrene	17.033	276	684787	40.396	ng	98
88) Benzo(b)fluoranthene	15.098	252	615487	38.345	ng	99
89) Benzo(k)fluoranthene	15.127	252	563702	40.721	ng	99
90) Benzo(a)pyrene	15.468	252	533309	40.379	ng	99
91) Dibenzo(a,h)anthracene	17.051	278	563984	39.881	ng	98
92) Benzo(g,h,i)perylene	17.486	276	583161	41.284	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140075.D
 Acq On : 28 Oct 2024 09:32
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1




Quant Time: Oct 28 09:55:45 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024









#1

1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.892 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#2

1,4-Dioxane
Concen: 37.538 ng
RT: 2.669 min Scan# 98
Delta R.T. -0.018 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32


Tgt Ion: 88 Resp: 148154
Ion Ratio Lower Upper
88 100
58 87.9 73.3 109.9
43 32.5 27.1 40.7





#3
 Pyridine
 Concen: 38.244 ng
 RT: 3.440 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BF140075.D
 Acq: 28 Oct 2024 09:32



Instrument : BNA_F
 ClientSampleId : SSTDCCC040



Tgt Ion: 79 Resp: 374133
 Ion Ratio Lower Upper
 79 100
 52 83.2 68.5 102.7
 51 40.7 32.3 48.5



Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024






#4
 n-Nitrosodimethylamine
 Concen: 37.610 ng
 RT: 3.387 min Scan# 220
 Delta R.T. -0.012 min
 Lab File: BF140075.D
 Acq: 28 Oct 2024 09:32

Tgt Ion: 42 Resp: 197677
 Ion Ratio Lower Upper
 42 100
 74 115.6 96.2 144.4
 44 8.1 6.2 9.2



Sub






#5
2-Fluorophenol
Concen: 75.045 ng
RT: 5.504 min Scan# 5
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Instrument : BNA_F
ClientSampleId : SSTDCCC040


Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#6
Aniline
Concen: 37.992 ng
RT: 6.551 min Scan# 758
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion: 93 Resp: 388187
Ion Ratio Lower Upper
93 100
66 41.5 34.0 51.0
65 21.6 17.4 26.0





#7

Phenol-d6

Concen: 74.300 ng

RT: 6.516 min Scan# 7

Delta R.T. 0.006 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32


Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024


#8

2-Chlorophenol


Concen: 37.936 ng


RT: 6.675 min Scan# 779

Delta R.T. 0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32



 Tgt Ion:128 Resp: 328283
 Ion Ratio Lower Upper
 128 100
 130 32.5 12.3 52.3
 64 54.8 34.6 74.6




#9
Benzaldehyde
Concen: 38.007 ng
RT: 6.439 min Scan# 7
Instrument: BNA_F
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32
ClientSampleId : SSTDCCC040


Manual Integrations APPROVED







Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#10
Phenol
Concen: 37.802 ng
RT: 6.528 min Scan# 754
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion: 94 Resp: 427262
Ion Ratio Lower Upper
94 100
65 31.7 11.1 51.1
66 43.6 22.7 62.7






#11
bis(2-Chloroethyl)ether
Concen: 37.186 ng
RT: 6.628 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations APPROVED







Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#12
1,3-Dichlorobenzene
Concen: 39.802 ng
RT: 6.833 min Scan# 806
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:146 Resp: 395383
Ion Ratio Lower Upper
146 100
148 64.0 51.0 76.4
75 32.4 26.4 39.6





#13

1,4-Dichlorobenzene

Concen: 39.676 ng

RT: 6.910 min Scan# 819

Delta R.T. -0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tgt Ion:146 Resp: 39376

Ion Ratio Lower Upper

146 100

148 64.2 51.8 77.8

111 42.7 34.6 51.8

**Manual Integrations
APPROVED**
Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

#14

1,2-Dichlorobenzene

Concen: 39.360 ng

RT: 7.063 min Scan# 845

Delta R.T. -0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32


Tgt Ion:146 Resp: 364568




Ion Ratio Lower Upper

146 100

148 63.2 51.4 77.2

111 44.9 36.7 55.1





#15

Benzyl Alcohol

Concen: 38.244 ng

RT: 7.028 min Scan# 8

Delta R.T. 0.000 min

Lab File: BF140075.D




Acq: 28 Oct 2024 09:32

Instrument : BNA_F

ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024


#16

2,2'-oxybis(1-Chloropropane)

Concen: 36.130 ng

RT: 7.163 min Scan# 862

Delta R.T. -0.006 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32


Tgt Ion: 45 Resp: 538199




Ion Ratio Lower Upper

45 100

77 12.5 0.0 31.5

79 9.4 0.0 29.1





#17

2-Methylphenol

Concen: 37.712 ng

RT: 7.133 min Scan# 8

Delta R.T. -0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Instrument :

BNA_F




ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#18

Hexachloroethane

Concen: 39.369 ng

RT: 7.404 min Scan# 903

Delta R.T. 0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32


Tgt Ion:117 Resp: 139330




Ion Ratio Lower Upper

117 100

119 96.9 77.0 115.4

201 99.5 76.2 114.4









#19
n-Nitroso-di-n-propylamine
Concen: 36.671 ng
RT: 7.304 min Scan# 884
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations APPROVED







Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#20
3+4-Methylphenols
Concen: 37.070 ng
RT: 7.292 min Scan# 884
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:107 Resp: 349881
Ion Ratio Lower Upper
107 100
108 90.4 69.8 109.8
77 107.5 50.5 90.5#
79 31.3 10.2 50.2





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.175 min Scan# 1

Delta R.T. 0.000 min


Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Instrument : BNA_F

ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**
Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

#22

Acetophenone

Concen: 38.331 ng


RT: 7.298 min Scan# 885




Delta R.T. -0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Tgt	Ion:105	Resp:	466951
Ion	Ratio	Lower	Upper
105	100		
71	7.7	5.9	8.9
51	37.2	29.4	44.2
120	21.4	17.5	26.3





#23

Nitrobenzene-d5

Concen: 83.584 ng

RT: 7.451 min Scan# 911

Delta R.T. 0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Instrument :

BNA_F




ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#24

Nitrobenzene

Concen: 40.136 ng

RT: 7.475 min Scan# 915

Delta R.T. 0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32


Tgt Ion: 77 Resp: 394888


Ion Ratio Lower Upper

77 100


123 44.3 35.7 53.5

65 14.3 11.5 17.3





Abundance Scan 955 (7.710 min): BF140075.D\data.ms



#25

Isophorone

Concen: 38.425 ng

RT: 7.710 min Scan# 955

Delta R.T. -0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Instrument:

BNA_F



ClientSampleId :

SSTDCCC040


**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



Abundance Scan 968 (7.786 min): BF140075.D\data.ms



#26

2-Nitrophenol

Concen: 40.704 ng

RT: 7.786 min Scan# 968

Delta R.T. -0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32


Tgt Ion:139 Resp: 151084




Ion Ratio Lower Upper

139 100

109 31.1 25.1 37.7

65 57.4 47.5 71.3





#27

2,4-Dimethylphenol

Concen: 35.678 ng

RT: 7.816 min Scan# 9

Delta R.T. 0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32




Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024


#28

bis(2-Chloroethoxy)methane

Concen: 38.000 ng

RT: 7.922 min Scan# 991

Delta R.T. -0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32


Tgt Ion: 93 Resp: 392131




Ion Ratio Lower Upper

93 100

95 32.7 26.0 39.0

123 12.0 9.8 14.8





#29

2,4-Dichlorophenol

Concen: 39.331 ng

RT: 8.028 min Scan# 1008

Delta R.T. 0.006 min




Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Instrument : BNA_F
ClientSampleId : SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#30

1,2,4-Trichlorobenzene

Concen: 40.164 ng

RT: 8.116 min Scan# 1024

Delta R.T. 0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32


Tgt Ion:180 Resp: 313301


Ion Ratio Lower Upper

180 100

182 100.6 80.5 120.7


145 32.0 25.4 38.0





#31
Naphthalene
Concen: 39.413 ng
RT: 8.198 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:128 Resp: 101097:
Ion Ratio Lower Upper
128 100
129 11.1 8.9 13.3
127 13.3 10.6 16.0

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024









#32
Benzoic acid
Concen: 34.127 ng
RT: 7.928 min Scan# 992
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32



Tgt Ion:122 Resp: 184222
Ion Ratio Lower Upper
122 100
105 129.3 107.6 147.6
77 98.6 80.2 120.2





#33

4-Chloroaniline

Concen: 38.998 ng

RT: 8.239 min Scan# 1045

Delta R.T. -0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Instrument:

BNA_F

ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024

Tgt Ion:127 Resp: 341100


Ion Ratio Lower Upper

127 100

129 32.3 25.4 38.2

65 34.5 28.4 42.6

92 20.6 16.6 25.0



#34

Hexachlorobutadiene

Concen: 40.501 ng

RT: 8.310 min Scan# 1057

Delta R.T. -0.006 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32


Tgt Ion:225 Resp: 198499







Ion Ratio Lower Upper

225 100

223 63.6 50.6 76.0

227 64.9 51.8 77.8





#35

Caprolactam

Concen: 39.542 ng

RT: 8.616 min Scan# 1108

Delta R.T. 0.006 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:113 Resp: 88634

Ion Ratio Lower Upper

113 100

55 193.0 189.2 229.2


56 146.0 132.5 172.5

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#36

4-Chloro-3-methylphenol

Concen: 39.128 ng

RT: 8.716 min Scan# 1126

Delta R.T. 0.006 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32


Tgt Ion:107 Resp: 305431


Ion Ratio Lower Upper

107 100


144 24.4 19.1 28.7

142 73.2 57.2 85.8





Abundance Scan 1155 (8.886 min): BF140075.D\data.ms



#37

2-Methylnaphthalene

Concen: 39.271 ng

RT: 8.886 min Scan# 1

Delta R.T. -0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:142 Resp: 616924

Ion Ratio Lower Upper


142 100

141 89.2 70.8 106.2


Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



Abundance Scan 1172 (8.986 min): BF140075.D\data.ms



#38

1-Methylnaphthalene

Concen: 39.400 ng

RT: 8.986 min Scan# 1172

Delta R.T. 0.000 min

Lab File: BF140075.D



Acq: 28 Oct 2024 09:32




Tgt Ion:142 Resp: 607270

Ion Ratio Lower Upper

142 100

141 92.0 73.5 110.3





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.927 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Instrument :

BNA_F





ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#40

1,2,4,5-Tetrachlorobenzene

Concen: 41.099 ng


RT: 9.051 min Scan# 1183


Delta R.T. 0.000 min

Lab File: BF140075.D



Acq: 28 Oct 2024 09:32

Tgt	Ion:216	Resp:	303883
Ion	Ratio	Lower	Upper
216	100		
214	80.2	63.4	95.0
179	20.9	17.2	25.8
108	18.9	15.6	23.4







Abundance Scan 1181 (9.039 min): BF140075.D\data.ms



Abundance Scan 1466 (10.716 min): BF140075.D\data.ms



#41

Hexachlorocyclopentadiene

Concen: 38.781 ng

RT: 9.039 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:237 Resp: 99773


Ion Ratio Lower Upper

237 100

235 62.8 42.9 82.9

272 12.8 0.0 32.3

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

#42

2,4,6-Tribromophenol

Concen: 84.285 ng

RT: 10.716 min Scan# 1466

Delta R.T. 0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32


Tgt Ion:330 Resp: 216066







Ion Ratio Lower Upper

330 100

332 96.6 78.1 117.1

141 31.7 26.6 39.8





#43

2,4,6-Trichlorophenol

Concen: 39.258 ng

RT: 9.157 min Scan# 1201

Delta R.T. -0.000 min

Lab File: BF140075.D





Acq: 28 Oct 2024 09:32

Instrument:

BNA_F

ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**
Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

#44

2,4,5-Trichlorophenol

Concen: 39.754 ng


RT: 9.198 min Scan# 1208







Delta R.T. 0.006 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Tgt	Ion:196	Resp:	214705
Ion	Ratio	Lower	Upper
196	100		
198	95.8	75.8	113.6
97	50.4	41.8	62.6
132	27.9	22.9	34.3





#45

2-Fluorobiphenyl

Concen: 79.000 ng

RT: 9.245 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32


Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024


#46

1,1'-Biphenyl



Concen: 39.635 ng




RT: 9.351 min Scan# 1234

Delta R.T. -0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32





#47

2-Chloronaphthalene

Concen: 40.398 ng

RT: 9.375 min Scan# 1238

Delta R.T. -0.000 min





Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Instrument : BNA_F

ClientSampleId : SSTDCCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024


#48

2-Nitroaniline

Concen: 42.796 ng


RT: 9.469 min Scan# 1254







Delta R.T. 0.006 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Tgt	Ion	Ion Ratio	Resp:	Lower	Upper
65	100				
92	65.6	65.6	201129	50.9	76.3
138	99.2	99.2	201129	73.4	110.0






#49


Acenaphthylene
Concen: 40.124 ngRT: 9.792 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32Instrument :
BNA_F
ClientSampleId :
SSTDCCC040







Tgt	Ion	Ion Ratio	Resp:	Lower	Upper
	152	100			
	151	20.1	891367	15.9	23.9
	153	13.1		10.7	16.1

**Manual Integrations
APPROVED**Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024


#50
Dimethylphthalate
Concen: 39.953 ng
RT: 9.651 min Scan# 1285
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt	Ion	Ion Ratio	Resp:	Lower	Upper
	163	100			
	194	4.1	682031	3.4	5.0
	164	9.8		8.2	12.2









#51

2,6-Dinitrotoluene
Concen: 43.701 ngRT: 9.710 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32Instrument :
BNA_F
ClientSampleId :
SSTDCCC040**Manual Integrations
APPROVED**Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

#52
Acenaphthene
Concen: 40.343 ng
RT: 9.963 min Scan# 1338
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:154 Resp: 579769
Ion Ratio Lower Upper
154 100
153 112.4 89.8 134.6
152 54.4 42.4 63.6





#53

3-Nitroaniline

Concen: 42.148 ng

RT: 9.874 min Scan# 1323

Delta R.T. -0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Instrument :

BNA_F





ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#54

2,4-Dinitrophenol

Concen: 45.309 ng


RT: 9.980 min Scan# 1341


Delta R.T. 0.006 min

Lab File: BF140075.D


Acq: 28 Oct 2024 09:32

Tgt	Ion	Ion Ratio	Resp:	Lower	Upper
184	100				
63	78.4		64847	84.7	127.1
154	125.1		#	338.2	507.2







#55
Dibenzofuran
Concen: 39.443 ng
RT: 10.133 min Scan# 1
Instrument: BNA_F
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32
ClientSampleId : SSTDCCC040




Tgt Ion:168 Resp: 813269
Ion Ratio Lower Upper
168 100
139 40.0 31.8 47.6
169 14.2 11.6 17.4

Manual Integrations APPROVED




Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024





#56
4-Nitrophenol
Concen: 40.940 ng
RT: 10.027 min Scan# 1349
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32



Tgt Ion:139 Resp: 120058
Ion Ratio Lower Upper
139 100
109 76.0 53.6 93.6
65 111.5 87.8 127.8





#57

2,4-Dinitrotoluene

Concen: 46.125 ng

RT: 10.110 min Scan# 1

Delta R.T. -0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:165 Resp: 209655

Ion Ratio Lower Upper

165 100

63 49.7 40.1 60.1


89 76.1 60.4 90.6

182 2.7 2.0 3.0

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#58

Fluorene

Concen: 40.338 ng

RT: 10.474 min Scan# 1425

Delta R.T. -0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32


Tgt Ion:166 Resp: 633488




Ion Ratio Lower Upper

166 100

165 97.7 78.5 117.7

167 13.6 10.9 16.3








#59
2,3,4,6-Tetrachlorophenol
Concen: 40.257 ng
RT: 10.251 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#60
Diethylphthalate
Concen: 40.303 ng
RT: 10.345 min Scan# 1403
Delta R.T. -0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:149 Resp: 675530
Ion Ratio Lower Upper
149 100
177 22.2 17.9 26.9
150 12.6 10.1 15.1





#61



4-Chlorophenyl-phenylether
Concen: 40.907 ng
RT: 10.469 min Scan# 1424
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024


Tgt	Ion:204	Resp:	324430
	Ion Ratio	Lower	Upper
204	100		
206	32.5	26.3	39.5
141	59.6	48.6	73.0




#62

4-Nitroaniline
Concen: 43.535 ng
RT: 10.492 min Scan# 1428
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt	Ion:138	Resp:	156727
	Ion Ratio	Lower	Upper
138	100		
92	52.9	35.7	75.7
108	78.6	57.5	97.5





#63

Azobenzene

Concen: 38.577 ng

RT: 10.627 min Scan# 1451

Delta R.T. 0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Instrument:

BNA_F


ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.416 min Scan# 1585

Delta R.T. 0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32


Tgt Ion:188 Resp: 494703




Ion Ratio Lower Upper

188 100

94 9.0 7.9 11.9

80 9.9 9.0 13.4









#65
4,6-Dinitro-2-methylphenol
Concen: 47.659 ng
RT: 10.522 min Scan# 1433
Delta R.T. 0.005 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#66
n-Nitrosodiphenylamine
Concen: 38.436 ng
RT: 10.586 min Scan# 1444
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:169 Resp: 569092
Ion Ratio Lower Upper
169 100
168 62.5 49.8 74.6
167 35.2 28.0 42.0






#67
4-Bromophenyl-phenylether
Concen: 40.268 ng
RT: 10.957 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#68
Hexachlorobenzene
Concen: 40.175 ng
RT: 11.021 min Scan# 1518
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:284 Resp: 230272
Ion Ratio Lower Upper
284 100
142 32.7 26.9 40.3
249 31.8 25.3 37.9






#69
Atrazine
Concen: 38.680 ng
RT: 11.110 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Instrument : BNA_F
ClientSampleId : SSTDCCC040


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#70
Pentachlorophenol
Concen: 40.865 ng
RT: 11.216 min Scan# 1551
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:266 Resp: 142153
Ion Ratio Lower Upper
266 100
268 65.1 50.4 75.6
264 63.6 50.6 75.8





#71

Phenanthrene

Concen: 38.500 ng

RT: 11.439 min Scan# 1

Delta R.T. -0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Instrument:

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:178 Resp: 899645

Ion Ratio Lower Upper

178 100




176 19.5 15.8 23.6

179 15.4 12.6 18.8

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#72

Anthracene

Concen: 38.994 ng

RT: 11.492 min Scan# 1598

Delta R.T. 0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32



Tgt Ion:178 Resp: 889039



Ion Ratio Lower Upper

178 100

176 19.0 15.3 22.9

179 15.4 12.4 18.6






#73


Carbazole
Concen: 37.870 ngRT: 11.645 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:	Ion Ratio	Resp:	803004
167	100		
166	21.6	Lower	17.0
139	13.3	Upper	25.6
			10.8
			16.2

**Manual Integrations
APPROVED**Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

#74
Di-n-butylphthalate
Concen: 39.082 ng
RT: 11.974 min Scan# 1680
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:	Ion Ratio	Resp:	957414
149	100		
150	9.1	Lower	7.5
104	4.9	Upper	11.3
			4.1
			6.1







#75
Fluoranthene
Concen: 38.158 ng
RT: 12.627 min Scan# 1
Instrument: BNA_F
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32


Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#76
Chrysene-d₁₂
Concen: 20.000 ng
RT: 14.051 min Scan# 2033
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:240 Resp: 250901
Ion Ratio Lower Upper
240 100
120 11.0 9.4 14.2
236 25.6 20.9 31.3







#77

Benzidine
Concen: 45.296 ngRT: 12.745 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32Instrument :
BNA_F
ClientSampleId :
SSTDCCC040**Manual Integrations
APPROVED**Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

#78

Pyrene
Concen: 40.730 ng
RT: 12.857 min Scan# 1830
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32Tgt Ion:202 Resp: 894517
Ion Ratio Lower Upper
202 100
200 21.3 17.2 25.8
203 17.5 14.2 21.4





#79
Terphenyl-d14
Concen: 81.484 ng
RT: 12.998 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Instrument : BNA_F
ClientSampleId : SSTDCCC040


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#80
Butylbenzylphthalate
Concen: 41.265 ng
RT: 13.468 min Scan# 1934
Delta R.T. -0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:149 Resp: 271701
Ion Ratio Lower Upper
149 100
91 75.1 59.4 89.2
206 20.0 16.4 24.6





#81

Benzo(a)anthracene

Concen: 40.732 ng

RT: 14.039 min Scan# 2031

Delta R.T. -0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Instrument :

BNA_F





ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#82

3,3'-Dichlorobenzidine

Concen: 44.426 ng


RT: 14.004 min Scan# 2025




Delta R.T. 0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Tgt	Ion	Ion Ratio	Resp:	Lower	Upper
252	100				
254	64.1	50.8	211560	76.2	
126	10.5	8.9		13.3	





#83

Chrysene

Concen: 39.262 ng

RT: 14.080 min Scan# 2038

Delta R.T. 0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32




Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024


#84

Bis(2-ethylhexyl)phthalate

Concen: 45.339 ng

RT: 14.027 min Scan# 2029

Delta R.T. -0.006 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32



Tgt Ion:149 Resp: 334928



Ion Ratio Lower Upper

149 100

167 26.2 21.4 32.0

279 4.0 3.0 4.4






#85

Di-n-octyl phthalate
Concen: 40.754 ng
RT: 14.639 min Scan# 2133
Delta R.T. -0.012 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024









#86

Perylene-d12
Concen: 20.000 ng
RT: 15.527 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:264 Resp: 263501
Ion Ratio Lower Upper
264 100
260 24.1 19.4 29.2
265 22.0 17.4 26.0





#87

Indeno(1,2,3-cd)pyrene

Concen: 40.396 ng

RT: 17.033 min Scan# 2

Delta R.T. 0.012 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:276 Resp: 684782

Ion Ratio Lower Upper

276 100


138 23.3 20.4 30.6

277 25.3 20.2 30.4

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#88

Benzo(b)fluoranthene

Concen: 38.345 ng

RT: 15.098 min Scan# 2211

Delta R.T. 0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32


Tgt Ion:252 Resp: 615487




Ion Ratio Lower Upper

252 100

253 21.8 17.6 26.4

125 9.0 7.9 11.9





#89

Benzo(k)fluoranthene

Concen: 40.721 ng

RT: 15.127 min Scan# 2216

Delta R.T. -0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32




Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024


#90

Benzo(a)pyrene

Concen: 40.379 ng

RT: 15.468 min Scan# 2274

Delta R.T. 0.000 min

Lab File: BF140075.D

Acq: 28 Oct 2024 09:32


Tgt Ion:252 Resp: 533309



Ion Ratio Lower Upper

252 100

253 21.9 17.3 25.9

125 10.4 8.7 13.1






#91
Dibenzo(a,h)anthracene
Concen: 39.881 ng
RT: 17.051 min Scan# 2

Instrument : BNA_F
ClientSampleId : SSTDCCC040

Tgt Ion:	Ion Ratio	Resp:	Lower	Upper
278	100	563984		
139	14.5		13.1	19.7
279	23.9		19.0	28.4


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#92
Benzo(g,h,i)perylene
Concen: 41.284 ng
RT: 17.486 min Scan# 2617
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:	Ion Ratio	Resp:	Lower	Upper
276	100	583161		
277	23.6		19.0	28.6
138	20.4		17.4	26.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140075.D
 Acq On : 28 Oct 2024 09:32
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 28 09:55:45 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	73	0.00
2	1,4-Dioxane	40.000	37.538	6.2	69	-0.02
3	Pyridine	40.000	38.244	4.4	68	0.00
4	n-Nitrosodimethylamine	40.000	37.610	6.0	67	-0.01
5 S	2-Fluorophenol	80.000	75.045	6.2	68	0.00
6	Aniline	40.000	37.992	5.0	67	0.00
7 S	Phenol-d6	80.000	74.300	7.1	68	0.00
8	2-Chlorophenol	40.000	37.936	5.2	69	0.00
9	Benzaldehyde	40.000	38.007	5.0	69	0.00
10 C	Phenol	40.000	37.802	5.5	69	0.00
11	bis(2-Chloroethyl)ether	40.000	37.186	7.0	69	0.00
12	1,3-Dichlorobenzene	40.000	39.802	0.5	73	0.00
13 C	1,4-Dichlorobenzene	40.000	39.676	0.8	73	0.00
14	1,2-Dichlorobenzene	40.000	39.360	1.6	73	0.00
15	Benzyl Alcohol	40.000	38.244	4.4	70	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.130	9.7	66	0.00
17	2-Methylphenol	40.000	37.712	5.7	69	0.00
18	Hexachloroethane	40.000	39.369	1.6	71	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.671	8.3	69	0.00
20	3+4-Methylphenols	40.000	37.070	7.3	68	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	72	0.00
22	Acetophenone	40.000	38.331	4.2	70	0.00
23 S	Nitrobenzene-d5	80.000	83.584	-4.5	73	0.00
24	Nitrobenzene	40.000	40.136	-0.3	71	0.00
25	Isophorone	40.000	38.425	3.9	70	0.00
26 C	2-Nitrophenol	40.000	40.704	-1.8	70	0.00
27	2,4-Dimethylphenol	40.000	35.678	10.8	65	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.000	5.0	69	0.00
29 C	2,4-Dichlorophenol	40.000	39.331	1.7	71	0.00
30	1,2,4-Trichlorobenzene	40.000	40.164	-0.4	72	0.00
31	Naphthalene	40.000	39.413	1.5	72	0.00
32	Benzoic acid	40.000	34.127	14.7	61	0.00
33	4-Chloroaniline	40.000	38.998	2.5	70	0.00
34 C	Hexachlorobutadiene	40.000	40.501	-1.3	73	0.00
35	Caprolactam	40.000	39.542	1.1	70	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.128	2.2	70	0.00
37	2-Methylnaphthalene	40.000	39.271	1.8	72	0.00
38	1-Methylnaphthalene	40.000	39.400	1.5	72	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	72	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.099	-2.7	75	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.781	3.0	66	0.00
42 S	2,4,6-Tribromophenol	80.000	84.285	-5.4	76	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.258	1.9	70	0.00
44	2,4,5-Trichlorophenol	40.000	39.754	0.6	71	0.00
45 S	2-Fluorobiphenyl	80.000	79.000	1.3	74	0.00
46	1,1'-Biphenyl	40.000	39.635	0.9	72	0.00
47	2-Chloronaphthalene	40.000	40.398	-1.0	74	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140075.D
 Acq On : 28 Oct 2024 09:32
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 28 09:55:45 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.796	-7.0	73	0.00
49	Acenaphthylene	40.000	40.124	-0.3	73	0.00
50	Dimethylphthalate	40.000	39.953	0.1	73	0.00
51	2,6-Dinitrotoluene	40.000	43.701	-9.3	76	0.00
52 C	Acenaphthene	40.000	40.343	-0.9	73	0.00
53	3-Nitroaniline	40.000	42.148	-5.4	72	0.00
54 P	2,4-Dinitrophenol	40.000	45.309	-13.3	84	0.00
55	Dibenzofuran	40.000	39.443	1.4	72	0.00
56 P	4-Nitrophenol	40.000	40.940	-2.3	68	0.00
57	2,4-Dinitrotoluene	40.000	46.125	-15.3	77	0.00
58	Fluorene	40.000	40.338	-0.8	75	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.257	-0.6	72	0.00
60	Diethylphthalate	40.000	40.303	-0.8	73	0.00
61	4-Chlorophenyl-phenylether	40.000	40.907	-2.3	75	0.00
62	4-Nitroaniline	40.000	43.535	-8.8	75	0.00
63	Azobenzene	40.000	38.577	3.6	69	0.00
64 I	Phanthrene-d10	20.000	20.000	0.0	75	0.00
65	4,6-Dinitro-2-methylphenol	40.000	47.659	-19.1	83	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.436	3.9	72	0.00
67	4-Bromophenyl-phenylether	40.000	40.268	-0.7	76	0.00
68	Hexachlorobenzene	40.000	40.175	-0.4	75	0.00
69	Atrazine	40.000	38.680	3.3	67	0.00
70 C	Pentachlorophenol	40.000	40.865	-2.2	71	0.00
71	Phanthrene	40.000	38.500	3.8	73	0.00
72	Anthracene	40.000	38.994	2.5	73	0.00
73	Carbazole	40.000	37.870	5.3	72	0.00
74	Di-n-butylphthalate	40.000	39.082	2.3	71	0.00
75 C	Fluoranthene	40.000	38.158	4.6	72	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	73	0.00
77	Benzidine	40.000	45.296	-13.2	74	0.00
78	Pyrene	40.000	40.730	-1.8	72	0.00
79 S	Terphenyl-d14	80.000	81.484	-1.9	73	0.00
80	Butylbenzylphthalate	40.000	41.265	-3.2	71	0.00
81	Benzo(a)anthracene	40.000	40.732	-1.8	72	0.00
82	3,3'-Dichlorobenzidine	40.000	44.426	-11.1	79	0.00
83	Chrysene	40.000	39.262	1.8	73	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	45.339	-13.3	77	0.00
85 c	Di-n-octyl phthalate	40.000	40.754	-1.9	70	-0.01
86 I	Perylene-d12	20.000	20.000	0.0	72	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.396	-1.0	69	0.01
88	Benzo(b)fluoranthene	40.000	38.345	4.1	70	0.00
89	Benzo(k)fluoranthene	40.000	40.721	-1.8	72	0.00
90 C	Benzo(a)pyrene	40.000	40.379	-0.9	71	0.00
91	Dibenzo(a,h)anthracene	40.000	39.881	0.3	69	0.00
92	Benzo(g,h,i)perylene	40.000	41.284	-3.2	71	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140075.D
Acq On : 28 Oct 2024 09:32
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 28 09:55:45 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	Dev(min)
----------	--------	-------	------	------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140075.D
 Acq On : 28 Oct 2024 09:32
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 28 09:55:45 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	73	0.00
2	1,4-Dioxane	40.000	37.538	6.2	69	-0.02
3	Pyridine	40.000	38.244	4.4	68	0.00
4	n-Nitrosodimethylamine	40.000	37.610	6.0	67	-0.01
5 S	2-Fluorophenol	80.000	75.045	6.2	68	0.00
6	Aniline	40.000	37.992	5.0	67	0.00
7 S	Phenol-d6	80.000	74.300	7.1	68	0.00
8	2-Chlorophenol	40.000	37.936	5.2	69	0.00
9	Benzaldehyde	40.000	38.007	5.0	69	0.00
10 C	Phenol	40.000	37.802	5.5	69	0.00
11	bis(2-Chloroethyl)ether	40.000	37.186	7.0	69	0.00
12	1,3-Dichlorobenzene	40.000	39.802	0.5	73	0.00
13 C	1,4-Dichlorobenzene	40.000	39.676	0.8	73	0.00
14	1,2-Dichlorobenzene	40.000	39.360	1.6	73	0.00
15	Benzyl Alcohol	40.000	38.244	4.4	70	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.130	9.7	66	0.00
17	2-Methylphenol	40.000	37.712	5.7	69	0.00
18	Hexachloroethane	40.000	39.369	1.6	71	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.671	8.3	69	0.00
20	3+4-Methylphenols	40.000	37.070	7.3	68	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	72	0.00
22	Acetophenone	40.000	38.331	4.2	70	0.00
23 S	Nitrobenzene-d5	80.000	83.584	-4.5	73	0.00
24	Nitrobenzene	40.000	40.136	-0.3	71	0.00
25	Isophorone	40.000	38.425	3.9	70	0.00
26 C	2-Nitrophenol	40.000	40.704	-1.8	70	0.00
27	2,4-Dimethylphenol	40.000	35.678	10.8	65	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.000	5.0	69	0.00
29 C	2,4-Dichlorophenol	40.000	39.331	1.7	71	0.00
30	1,2,4-Trichlorobenzene	40.000	40.164	-0.4	72	0.00
31	Naphthalene	40.000	39.413	1.5	72	0.00
32	Benzoic acid	40.000	34.127	14.7	61	0.00
33	4-Chloroaniline	40.000	38.998	2.5	70	0.00
34 C	Hexachlorobutadiene	40.000	40.501	-1.3	73	0.00
35	Caprolactam	40.000	39.542	1.1	70	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.128	2.2	70	0.00
37	2-Methylnaphthalene	40.000	39.271	1.8	72	0.00
38	1-Methylnaphthalene	40.000	39.400	1.5	72	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	72	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.099	-2.7	75	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.781	3.0	66	0.00
42 S	2,4,6-Tribromophenol	80.000	84.285	-5.4	76	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.258	1.9	70	0.00
44	2,4,5-Trichlorophenol	40.000	39.754	0.6	71	0.00
45 S	2-Fluorobiphenyl	80.000	79.000	1.3	74	0.00
46	1,1'-Biphenyl	40.000	39.635	0.9	72	0.00
47	2-Chloronaphthalene	40.000	40.398	-1.0	74	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140075.D
 Acq On : 28 Oct 2024 09:32
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 28 09:55:45 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.796	-7.0	73	0.00
49	Acenaphthylene	40.000	40.124	-0.3	73	0.00
50	Dimethylphthalate	40.000	39.953	0.1	73	0.00
51	2,6-Dinitrotoluene	40.000	43.701	-9.3	76	0.00
52 C	Acenaphthene	40.000	40.343	-0.9	73	0.00
53	3-Nitroaniline	40.000	42.148	-5.4	72	0.00
54 P	2,4-Dinitrophenol	40.000	45.309	-13.3	84	0.00
55	Dibenzofuran	40.000	39.443	1.4	72	0.00
56 P	4-Nitrophenol	40.000	40.940	-2.3	68	0.00
57	2,4-Dinitrotoluene	40.000	46.125	-15.3	77	0.00
58	Fluorene	40.000	40.338	-0.8	75	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.257	-0.6	72	0.00
60	Diethylphthalate	40.000	40.303	-0.8	73	0.00
61	4-Chlorophenyl-phenylether	40.000	40.907	-2.3	75	0.00
62	4-Nitroaniline	40.000	43.535	-8.8	75	0.00
63	Azobenzene	40.000	38.577	3.6	69	0.00
64 I	Phanthrene-d10	20.000	20.000	0.0	75	0.00
65	4,6-Dinitro-2-methylphenol	40.000	47.659	-19.1	83	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.436	3.9	72	0.00
67	4-Bromophenyl-phenylether	40.000	40.268	-0.7	76	0.00
68	Hexachlorobenzene	40.000	40.175	-0.4	75	0.00
69	Atrazine	40.000	38.680	3.3	67	0.00
70 C	Pentachlorophenol	40.000	40.865	-2.2	71	0.00
71	Phanthrene	40.000	38.500	3.8	73	0.00
72	Anthracene	40.000	38.994	2.5	73	0.00
73	Carbazole	40.000	37.870	5.3	72	0.00
74	Di-n-butylphthalate	40.000	39.082	2.3	71	0.00
75 C	Fluoranthene	40.000	38.158	4.6	72	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	73	0.00
77	Benzidine	40.000	45.296	-13.2	74	0.00
78	Pyrene	40.000	40.730	-1.8	72	0.00
79 S	Terphenyl-d14	80.000	81.484	-1.9	73	0.00
80	Butylbenzylphthalate	40.000	41.265	-3.2	71	0.00
81	Benzo(a)anthracene	40.000	40.732	-1.8	72	0.00
82	3,3'-Dichlorobenzidine	40.000	44.426	-11.1	79	0.00
83	Chrysene	40.000	39.262	1.8	73	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	45.339	-13.3	77	0.00
85 c	Di-n-octyl phthalate	40.000	40.754	-1.9	70	-0.01
86 I	Perylene-d12	20.000	20.000	0.0	72	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.396	-1.0	69	0.01
88	Benzo(b)fluoranthene	40.000	38.345	4.1	70	0.00
89	Benzo(k)fluoranthene	40.000	40.721	-1.8	72	0.00
90 C	Benzo(a)pyrene	40.000	40.379	-0.9	71	0.00
91	Dibenzo(a,h)anthracene	40.000	39.881	0.3	69	0.00
92	Benzo(g,h,i)perylene	40.000	41.284	-3.2	71	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140075.D
Acq On : 28 Oct 2024 09:32
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 28 09:55:45 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	Dev(min)
----------	--------	-------	------	------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0



QC SAMPLE


DATA

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139843.D
 Acq On : 18 Oct 2024 09:22
 Operator : RC/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Oct 18 15:07:50 2024



AutoFind: Scans 1627, 1628, 1629; Background Corrected with Scan 1621

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.8	92701	PASS
68	69	0.00	2	1.9	1486	PASS
69	198	0.00	100	37.8	79959	PASS
70	69	0.00	2	0.7	529	PASS
127	198	10	80	46.9	99176	PASS
197	198	0.00	2	0.8	1650	PASS
198	198	100	100	100.0	211520	PASS
199	198	5	9	6.7	14078	PASS
275	198	10	60	26.9	56867	PASS
365	198	1	100	3.4	7148	PASS
441	198	0.01	100	14.0	29512	PASS
442	442	50	100	100.0	196979	PASS
443	442	15	24	19.2	37896	PASS

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/18/2024	BNA_F	<u>BF139843.D</u>
Compound Name	Response	Retention Time
DDT	585942	13.574
DDD	33558	13.274
DDE	421	12.939
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
33979	619921	5.48

Instrument :
BNA_F
ClientSampleId :
DFTPP

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139843.D
 Acq On : 18 Oct 2024 09:22
 Operator : RC/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Oct 18 11:03:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 11:03:04 2024
 Response via : Initial Calibration

Abundance


Ion 265.70 (265.40 to 266.40): BF139843.D\data.ms
 Ion 268.00 (267.70 to 268.70): BF139843.D\data.ms
 Ion 264.00 (263.70 to 264.70): BF139843.D\data.ms

11.210 Dailing = 1.27


S E

Time--> 10.20 10.30 10.40 10.50 10.60 10.70 10.80 10.90 11.00 11.10 11.20 11.30 11.40 11.50 11.60 11.70 11.80 11.90 12.00 12.10 12.20 12.30

Scan 1550 (11.210 min): BF139843.D\data.ms



Scan 1550 (11.210 min): BF139848.D\data.ms (-1544) (-)



TIC: BF139843.D\data.ms

(70) Pentachlorophenol (C)

11.210min (-0.006) 60102.24 ng

response 277191

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.00	63.60
264.00	62.50	63.30
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
 Data File : BF139843.D
 Acq On : 18 Oct 2024 09:22
 Operator : RC/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Oct 18 11:03:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 11:03:04 2024
 Response via : Initial Calibration

Abundance

Ion 184.00 (183.70 to 184.70): BF139843.D\data.ms
 Ion 185.00 (184.70 to 185.70): BF139843.D\data.ms
 Ion 183.00 (182.70 to 183.70): BF139843.D\data.ms

12.75^tailing = 0.89

\$ E

Time--> 11.70 11.80 11.90 12.00 12.10 12.20 12.30 12.40 12.50 12.60 12.70 12.80 12.90 13.00 13.10 13.20 13.30 13.40 13.50 13.60 13.70 13.80 13.90

Abundance

Scan 1812 (12.751 min): BF139843.D\data.ms

184.2

39.1 52.0 65.1 77.1 92.1 102.1 117.1 130.1 139.1 156.1 167.1 207.1

m/z--> 30 40 50 60 70 80 90 100 110 120 130 140 150 160 170 180 190 200 210 220 230 240

Abundance

Scan 1811 (12.745 min): BF139848.D\data.ms (-1804) (-)

184.1

39.0 51.0 65.1 77.0 92.1 102.0 117.1 130.1 139.1 156.1 167.1 208.0 237.0

m/z--> 30 40 50 60 70 80 90 100 110 120 130 140 150 160 170 180 190 200 210 220 230 240

TIC: BF139843.D\data.ms

(77) Benzidine

12.751min (+ 0.006) 86231.01 ng

response 1303015


Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	14.39
183.00	11.70	11.82
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140074.D
 Acq On : 28 Oct 2024 09:04
 Operator : RC/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Oct 18 15:07:50 2024



AutoFind: Scans 1626, 1627, 1628; Background Corrected with Scan 1620

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.7	78035	PASS
68	69	0.00	2	1.6	1158	PASS
69	198	0.00	100	38.6	70576	PASS
70	69	0.00	2	0.5	343	PASS
127	198	10	80	48.1	87888	PASS
197	198	0.00	2	0.8	1489	PASS
198	198	100	100	100.0	182720	PASS
199	198	5	9	6.8	12368	PASS
275	198	10	60	28.6	52275	PASS
365	198	1	100	3.5	6329	PASS
441	198	0.01	100	16.1	29440	PASS
442	442	50	100	100.0	189696	PASS
443	442	15	24	18.8	35661	PASS


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140074.D
 Acq On : 28 Oct 2024 09:04
 Operator : RC/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Oct 28 11:24:41 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration



Abundance

Ion 265.70 (265.40 to 266.40): BF140074.D\data.ms
 Ion 268.00 (267.70 to 268.70): BF140074.D\data.ms
 Ion 264.00 (263.70 to 264.70): BF140074.D\data.ms



Abundance

Scan 1551 (11.216 min): BF140074.D\data.ms



TIC: BF140074.D\data.ms

(70) Pentachlorophenol (C)

11.216min (+ 0.006) 128927.16 ng




response 209420

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.00	64.64
264.00	63.20	63.11
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140074.D
 Acq On : 28 Oct 2024 09:04
 Operator : RC/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Oct 28 11:24:41 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration



(77) Benzidine

12.751min (+ 0.006) 156036.69 ng

response 1051967

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.00	14.13
183.00	11.80	11.80
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/28/2024	BNA_F	<u>BF140074.D</u>
Compound Name	Response	Retention Time
DDT	521702	13.574
DDD	8598	13.304
DDE	1230	12.933
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
9828	531530	1.85

Instrument :
BNA_F
ClientSampleId :
DFTPP



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	
Client Sample ID:	PB164444BL			SDG No.:	P4578
Lab Sample ID:	PB164444BL			Matrix:	TCLP
Analytical Method:	SW8270			% Solid:	0
Sample Wt/Vol:	1000	Units:	mL	Final Vol:	1000 uL
Soil Aliquot Vol:			uL	Test:	TCLP BNA
Extraction Type :		Decanted :	N	Level :	LOW
Injection Volume :		GPC Factor :	1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140076.D	1	10/26/24 12:10	10/28/24 10:01	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	1.60	U	1.60	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.84	U	0.84	5.00	ug/L
95-48-7	2-Methylphenol	1.10	U	1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	1.20	U	1.20	10.0	ug/L
67-72-1	Hexachloroethane	1.00	U	1.00	5.00	ug/L
98-95-3	Nitrobenzene	1.30	U	1.30	5.00	ug/L
87-68-3	Hexachlorobutadiene	1.30	U	1.30	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	0.89	U	0.89	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	1.00	U	1.00	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	1.50	U	1.50	5.00	ug/L
118-74-1	Hexachlorobenzene	1.10	U	1.10	5.00	ug/L
87-86-5	Pentachlorophenol	1.90	U	1.90	10.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	120		15 (10) - 110 (139)	80%	SPK: 150
13127-88-3	Phenol-d6	114		15 (10) - 110 (134)	76%	SPK: 150
4165-60-0	Nitrobenzene-d5	90.9		30 (49) - 130 (133)	91%	SPK: 100
321-60-8	2-Fluorobiphenyl	87.1		30 (52) - 130 (132)	87%	SPK: 100
118-79-6	2,4,6-Tribromophenol	133		15 (44) - 110 (137)	89%	SPK: 150
1718-51-0	Terphenyl-d14	89.1		30 (48) - 130 (125)	89%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	132000	6.886			
1146-65-2	Naphthalene-d8	518000	8.169			
15067-26-2	Acenaphthene-d10	298000	9.922			
1517-22-2	Phenanthrene-d10	551000	11.41			
1719-03-5	Chrysene-d12	342000	14.045			
1520-96-3	Perylene-d12	261000	15.527			



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	
Client Sample ID:	PB164444BL			SDG No.:	P4578
Lab Sample ID:	PB164444BL			Matrix:	TCLP
Analytical Method:	SW8270			% Solid:	0
Sample Wt/Vol:	1000	Units:	mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL			Test:	TCLP BNA
Extraction Type :				Decanted :	N
Injection Volume :				Level :	LOW
Prep Method :	GPC Factor : 1.0			GPC Cleanup :	N PH :
	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140076.D	1	10/26/24 12:10	10/28/24 10:01	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140076.D
 Acq On : 28 Oct 2024 10:01
 Operator : RC/JU
 Sample : PB164444BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB164444BL

Quant Time: Oct 28 11:22:30 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.886	152	132330	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	518019	20.000	ng	0.00
39) Acenaphthene-d10	9.922	164	298114	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	550598	20.000	ng	0.00
76) Chrysene-d12	14.045	240	341644	20.000	ng	0.00
86) Perylene-d12	15.527	264	261224	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	1015752	120.165	ng	0.02
7) Phenol-d6	6.510	99	1244979	113.718	ng	0.00
23) Nitrobenzene-d5	7.445	82	849954	90.938	ng	0.00
42) 2,4,6-Tribromophenol	10.716	330	370207	132.764	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1571487	87.121	ng	0.00
79) Terphenyl-d14	12.998	244	1868766	89.131	ng	0.00


Target Compounds	Qvalue
(#= qualifier out of range (m)= manual integration (+)= signals summed	

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140076.D
 Acq On : 28 Oct 2024 10:01
 Operator : RC/JU
 Sample : PB164444BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB164444BL



Quant Time: Oct 28 11:22:30 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.886 min Scan# 8
Instrument: BNA_F
Delta R.T. -0.006 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01
ClientSampleId : PB164444BL


Tgt Ion:152 Resp: 132330
Ion Ratio Lower Upper
152 100
150 156.9 130.2 195.2
115 63.9 51.4 77.2



#5
2-Fluorophenol
Concen: 120.165 ng
RT: 5.516 min Scan# 582
Delta R.T. 0.018 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01

Tgt Ion:112 Resp: 1015752
Ion Ratio Lower Upper
112 100
64 65.1 53.0 79.6
63 34.3 27.0 40.4







#7
Phenol-d6
Concen: 113.718 ng
RT: 6.510 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01


Instrument : BNA_F
ClientSampleId : PB164444BL


Tgt Ion: 99 Resp: 1244979
Ion Ratio Lower Upper
99 100
42 20.0 16.7 25.1
71 36.5 27.7 41.5



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.169 min Scan# 1033
Delta R.T. -0.006 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01


Tgt Ion:136 Resp: 518019
Ion Ratio Lower Upper
136 100
137 10.7 8.6 12.8
54 10.0 8.4 12.6
68 6.1 5.1 7.7








#23
Nitrobenzene-d5
Concen: 90.938 ng
RT: 7.445 min Scan# 911
Delta R.T. -0.006 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01


Instrument : BNA_F
ClientSampleId : PB164444BL





Tgt Ion: 82 Resp: 849954
Ion Ratio Lower Upper
82 100
128 41.5 33.4 50.0
54 57.6 47.8 71.8




#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.922 min Scan# 1331
Delta R.T. -0.005 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01





Tgt Ion:164 Resp: 298114
Ion Ratio Lower Upper
164 100
162 100.2 81.0 121.4
160 43.8 35.4 53.0






#42
2,4,6-Tribromophenol
Concen: 132.764 ng
RT: 10.716 min Scan# 14
Instrument: BNA_F
Delta R.T. 0.000 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01
ClientSampleId : PB164444BL


Tgt Ion:330 Resp: 370207
Ion Ratio Lower Upper
330 100
332 95.2 78.1 117.1
141 30.0 26.6 39.8



#45
2-Fluorobiphenyl
Concen: 87.121 ng
RT: 9.245 min Scan# 1216
Delta R.T. -0.006 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01

Tgt Ion:172 Resp: 1571487
Ion Ratio Lower Upper
172 100
171 35.9 28.6 43.0
170 23.9 19.1 28.7





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.410 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF140076.D

Acq: 28 Oct 2024 10:01

Instrument:

BNA_F

ClientSampleId :

PB164444BL

Tgt Ion:188 Resp: 550598

Ion Ratio Lower Upper

188 100

94 9.8 7.9 11.9

80 10.6 9.0 13.4

Abundance

400000 11.410


300000

200000

100000

0

Time--> 11.30 11.40 11.50



#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.045 min Scan# 2032

Delta R.T. -0.006 min

Lab File: BF140076.D

Acq: 28 Oct 2024 10:01

Tgt Ion:240 Resp: 341644

Ion Ratio Lower Upper

240 100

120 11.6 9.4 14.2

236 26.2 20.9 31.3

Abundance

200000 14.045


150000

100000

50000

0



Time--> 14.00 14.045 14.20



#79
Terphenyl-d14
Concen: 89.131 ng
RT: 12.998 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01


Instrument : BNA_F
ClientSampleId : PB164444BL

Tgt Ion:244 Resp: 1868766
Ion Ratio Lower Upper
244 100
212 6.9 5.7 8.5
122 10.0 8.6 13.0



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.527 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01

Tgt Ion:264 Resp: 261224
Ion Ratio Lower Upper
264 100
260 23.3 19.4 29.2
265 21.3 17.4 26.0





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	
Client Sample ID:	PB164444BS			SDG No.:	P4578
Lab Sample ID:	PB164444BS			Matrix:	TCLP
Analytical Method:	SW8270			% Solid:	0
Sample Wt/Vol:	1000	Units:	mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL			Test:	TCLP BNA
Extraction Type :				Decanted :	N
Injection Volume :				Level :	LOW
Prep Method :	GPC Factor : 1.0			GPC Cleanup :	N PH :
SW3510C					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140077.D	1	10/26/24 12:10	10/28/24 10:29	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	38.6		1.60	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	43.7		0.84	5.00	ug/L
95-48-7	2-Methylphenol	44.4		1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	43.0		1.20	10.0	ug/L
67-72-1	Hexachloroethane	44.2		1.00	5.00	ug/L
98-95-3	Nitrobenzene	43.5		1.30	5.00	ug/L
87-68-3	Hexachlorobutadiene	44.9		1.30	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	45.9		0.89	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	44.6		1.00	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	52.4		1.50	5.00	ug/L
118-74-1	Hexachlorobenzene	45.0		1.10	5.00	ug/L
87-86-5	Pentachlorophenol	89.7	E	1.90	10.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	120		15 (10) - 110 (139)	80%	SPK: 150
13127-88-3	Phenol-d6	118		15 (10) - 110 (134)	79%	SPK: 150
4165-60-0	Nitrobenzene-d5	92.6		30 (49) - 130 (133)	93%	SPK: 100
321-60-8	2-Fluorobiphenyl	87.2		30 (52) - 130 (132)	87%	SPK: 100
118-79-6	2,4,6-Tribromophenol	141		15 (44) - 110 (137)	94%	SPK: 150
1718-51-0	Terphenyl-d14	95.7		30 (48) - 130 (125)	96%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	128000		6.887		
1146-65-2	Naphthalene-d8	495000		8.175		
15067-26-2	Acenaphthene-d10	281000		9.928		
1517-22-2	Phenanthrene-d10	511000		11.416		
1719-03-5	Chrysene-d12	263000		14.051		
1520-96-3	Perylene-d12	266000		15.527		



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	
Client Sample ID:	PB164444BS			SDG No.:	P4578
Lab Sample ID:	PB164444BS			Matrix:	TCLP
Analytical Method:	SW8270			% Solid:	0
Sample Wt/Vol:	1000	Units:	mL	Final Vol:	1000 uL
Soil Aliquot Vol:			uL	Test:	TCLP BNA
Extraction Type :		Decanted :	N	Level :	LOW
Injection Volume :		GPC Factor :	1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140077.D	1	10/26/24 12:10	10/28/24 10:29	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140077.D
 Acq On : 28 Oct 2024 10:29
 Operator : RC/JU
 Sample : PB164444BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB164444BS

Quant Time: Oct 28 11:23:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.887	152	127868	20.000	ng	0.00
21) Naphthalene-d8	8.175	136	495282	20.000	ng	0.00
39) Acenaphthene-d10	9.928	164	281323	20.000	ng	0.00
64) Phenanthrene-d10	11.416	188	511374	20.000	ng	0.00
76) Chrysene-d12	14.051	240	262593	20.000	ng	0.00
86) Perylene-d12	15.527	264	265684	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	983498	120.410	ng	0.02
7) Phenol-d6	6.516	99	1248557	118.025	ng	0.00
23) Nitrobenzene-d5	7.451	82	827399	92.589	ng	0.00
42) 2,4,6-Tribromophenol	10.722	330	372077	141.398	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1485083	87.245	ng	0.00
79) Terphenyl-d14	12.998	244	1542921	95.744	ng	0.00
Target Compounds						
				Qvalue		
2) 1,4-Dioxane	2.781	88	135793	35.657	ng	97
3) Pyridine	3.534	79	364598	38.624	ng	98
4) n-Nitrosodimethylamine	3.481	42	210444	41.494	ng	# 99
6) Aniline	6.551	93	381952	38.741	ng	100
8) 2-Chlorophenol	6.675	128	380511	45.570	ng	98
9) Benzaldehyde	6.440	77	75047	12.611	ng	96
10) Phenol	6.528	94	480470m	44.055	ng	
11) bis(2-Chloroethyl)ether	6.622	93	369586	43.843	ng	99
12) 1,3-Dichlorobenzene	6.834	146	409351	42.706	ng	99
13) 1,4-Dichlorobenzene	6.910	146	418487	43.700	ng	99
14) 1,2-Dichlorobenzene	7.057	146	399902	44.744	ng	98
15) Benzyl Alcohol	7.028	79	358350	46.179	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.163	45	620620	43.177	ng	98
17) 2-Methylphenol	7.139	107	315983	44.379	ng	98
18) Hexachloroethane	7.404	117	150834	44.169	ng	96
19) n-Nitroso-di-n-propyla...	7.304	70	281684	43.903	ng	97
20) 3+4-Methylphenols	7.292	107	391158	42.951	ng	# 77
22) Acetophenone	7.298	105	532350	43.883	ng	99
24) Nitrobenzene	7.469	77	425814	43.461	ng	99
25) Isophorone	7.710	82	773146	45.583	ng	99
26) 2-Nitrophenol	7.786	139	188880	51.101	ng	98
27) 2,4-Dimethylphenol	7.822	122	320582	52.015	ng	99
28) bis(2-Chloroethoxy)met...	7.916	93	456094	44.384	ng	99
29) 2,4-Dichlorophenol	8.028	162	318680	45.395	ng	98
30) 1,2,4-Trichlorobenzene	8.110	180	336007	43.256	ng	98
31) Naphthalene	8.192	128	1127006	44.121	ng	99
32) Benzoic acid	7.939	122	216873	40.344	ng	99
33) 4-Chloroaniline	8.239	127	159322	18.292	ng	98
34) Hexachlorobutadiene	8.310	225	219251	44.923	ng	98
35) Caprolactam	8.610	113	103223m	46.244	ng	
36) 4-Chloro-3-methylphenol	8.716	107	357353	45.972	ng	100
37) 2-Methylnaphthalene	8.886	142	713865	45.632	ng	99
38) 1-Methylnaphthalene	8.986	142	669536	43.622	ng	100
40) 1,2,4,5-Tetrachloroben...	9.051	216	338977	44.663	ng	99
41) Hexachlorocyclopentadiene	9.039	237	437259	165.576	ng	99
43) 2,4,6-Trichlorophenol	9.163	196	246720	45.915	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140077.D
 Acq On : 28 Oct 2024 10:29
 Operator : RC/JU
 Sample : PB164444BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB164444BS

Quant Time: Oct 28 11:23:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.204	196	247409	44.627	ng	98
46) 1,1'-Biphenyl	9.351	154	859069	43.865	ng	99
47) 2-Chloronaphthalene	9.375	162	677574	42.957	ng	100
48) 2-Nitroaniline	9.469	65	236689	49.063	ng	95
49) Acenaphthylene	9.792	152	1088844	47.749	ng	100
50) Dimethylphthalate	9.651	163	804892	45.933	ng	100
51) 2,6-Dinitrotoluene	9.710	165	179884	47.412	ng	99
52) Acenaphthene	9.963	154	735816	49.881	ng	99
53) 3-Nitroaniline	9.875	138	109911	28.091	ng	98
54) 2,4-Dinitrophenol	9.986	184	177437	109.904	ng	# 1
55) Dibenzofuran	10.133	168	947573	44.771	ng	100
56) 4-Nitrophenol	10.033	139	286527	95.186	ng	96
57) 2,4-Dinitrotoluene	10.116	165	244414	52.385	ng	99
58) Fluorene	10.480	166	727905	45.154	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.251	232	212306	48.851	ng	97
60) Diethylphthalate	10.351	149	791021	45.976	ng	100
61) 4-Chlorophenyl-phenyle...	10.469	204	369072	45.336	ng	99
62) 4-Nitroaniline	10.498	138	180917	48.958	ng	97
63) Azobenzene	10.627	77	806460	44.214	ng	98
65) 4,6-Dinitro-2-methylph...	10.522	198	126328	60.564	ng	95
66) n-Nitrosodiphenylamine	10.586	169	680576	44.467	ng	100
67) 4-Bromophenyl-phenylether	10.957	248	237977	45.173	ng	98
68) Hexachlorobenzene	11.027	284	266322	44.950	ng	96
69) Atrazine	11.116	200	224215	54.080	ng	99
70) Pentachlorophenol	11.216	266	322659	89.731	ng	99
71) Phenanthrene	11.439	178	1074904	44.500	ng	100
72) Anthracene	11.492	178	1107771	47.004	ng	100
73) Carbazole	11.645	167	952915	43.475	ng	99
74) Di-n-butylphthalate	11.974	149	1167048	46.086	ng	99
75) Fluoranthene	12.627	202	1073772	43.973	ng	99
77) Benzidine	12.745	184	237729	55.056	ng	99
78) Pyrene	12.857	202	1070018	46.552	ng	100
80) Butylbenzylphthalate	13.468	149	358899	52.081	ng	99
81) Benzo(a)anthracene	14.045	228	824539	48.236	ng	100
82) 3,3'-Dichlorobenzidine	14.004	252	161964	32.497	ng	99
83) Chrysene	14.080	228	734278	46.850	ng	100
84) Bis(2-ethylhexyl)phtha...	14.027	149	433668	56.091	ng	100
85) Di-n-octyl phthalate	14.639	149	699522	49.743	ng	97
87) Indeno(1,2,3-cd)pyrene	17.033	276	839150	49.096	ng	98
88) Benzo(b)fluoranthene	15.098	252	705155	43.570	ng	99
89) Benzo(k)fluoranthene	15.127	252	711830	50.999	ng	99
90) Benzo(a)pyrene	15.468	252	674445	50.645	ng	99
91) Dibenzo(a,h)anthracene	17.051	278	679796	47.676	ng	98
92) Benzo(g,h,i)perylene	17.486	276	644416	45.246	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140077.D
 Acq On : 28 Oct 2024 10:29
 Operator : RC/JU
 Sample : PB164444BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Quant Time: Oct 28 11:23:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Instrument :
 BNA_F
 ClientSampleId :
 PB164444BS


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024








#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.887 min Scan# 8
Delta R.T. -0.005 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29




Tgt Ion:152 Resp: 127868
Ion Ratio Lower Upper
152 100
150 159.6 130.2 195.2
115 65.7 51.4 77.2

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024





#2
1,4-Dioxane
Concen: 35.657 ng
RT: 2.781 min Scan# 117
Delta R.T. 0.094 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29



Tgt Ion: 88 Resp: 135793
Ion Ratio Lower Upper
88 100
58 88.2 73.3 109.9
43 32.6 27.1 40.7





#3

Pyridine

Concen: 38.624 ng

RT: 3.534 min Scan# 245

Delta R.T. 0.088 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Instrument :

BNA_F

ClientSampleId :

PB164444BS

Tgt Ion: 79 Resp: 364598

Ion Ratio Lower Upper

79 100


52 82.9 68.5 102.7

51 40.0 32.3 48.5

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#4

n-Nitrosodimethylamine

Concen: 41.494 ng

RT: 3.481 min Scan# 236

Delta R.T. 0.082 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29


Tgt Ion: 42 Resp: 210444


Ion Ratio Lower Upper

42 100

74 119.2 96.2 144.4


44 9.6 6.2 9.2#





#5
2-Fluorophenol
Concen: 120.410 ng
RT: 5.516 min Scan# 5
Delta R.T. 0.018 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29



Instrument : BNA_F
ClientSampleId : PB164444BS




Tgt Ion:112 Resp: 983498
Ion Ratio Lower Upper
112 100
64 65.8 53.0 79.6
63 34.5 27.0 40.4

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#6
Aniline
Concen: 38.741 ng
RT: 6.551 min Scan# 758
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29



Tgt Ion: 93 Resp: 381952
Ion Ratio Lower Upper
93 100
66 42.5 34.0 51.0
65 21.6 17.4 26.0







#7
Phenol-d6
Concen: 118.025 ng
RT: 6.516 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Instrument : BNA_F
ClientSampleId : PB164444BS


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#8
2-Chlorophenol
Concen: 45.570 ng
RT: 6.675 min Scan# 779
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:128 Resp: 380511
Ion Ratio Lower Upper
128 100
130 32.8 12.3 52.3
64 52.7 34.6 74.6







#9
Benzaldehyde
Concen: 12.611 ng
RT: 6.440 min Scan# 7
Delta R.T. 0.001 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Instrument : BNA_F
ClientSampleId : PB164444BS


Manual Integrations
APPROVED







Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#10
Phenol
Concen: 44.055 ng m
RT: 6.528 min Scan# 754
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion: 94 Resp: 480470
Ion Ratio Lower Upper
94 100
65 32.0 11.1 51.1
66 47.1 22.7 62.7






#11
bis(2-Chloroethyl)ether
Concen: 43.843 ng
RT: 6.622 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Instrument :
BNA_F
ClientSampleId :
PB164444BS


Manual Integrations APPROVED







Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#12
1,3-Dichlorobenzene
Concen: 42.706 ng
RT: 6.834 min Scan# 806
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:146 Resp: 409351
Ion Ratio Lower Upper
146 100
148 63.5 51.0 76.4
75 32.1 26.4 39.6





#13

1,4-Dichlorobenzene

Concen: 43.700 ng

RT: 6.910 min Scan# 819

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29


Instrument :

BNA_F

ClientSampleId :

PB164444BS

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024


#14

1,2-Dichlorobenzene

Concen: 44.744 ng

RT: 7.057 min Scan# 844


Delta R.T. -0.006 min







Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Tgt Ion:146 Resp: 399902

Ion	Ratio	Lower	Upper
146	100		
148	62.6	51.4	77.2
111	45.4	36.7	55.1





#15

Benzyl Alcohol

Concen: 46.179 ng

RT: 7.028 min Scan# 8

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Instrument : BNA_F

ClientSampleId : PB164444BS

Tgt Ion: 79 Resp: 358356

Ion Ratio Lower Upper

79 100


108 75.1 61.4 92.0

77 62.3 49.9 74.9

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#16

2,2'-oxybis(1-Chloropropane)

Concen: 43.177 ng

RT: 7.163 min Scan# 862

Delta R.T. -0.006 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29


Tgt Ion: 45 Resp: 620620




Ion Ratio Lower Upper

45 100

77 12.6 0.0 31.5

79 9.7 0.0 29.1








#17
2-Methylphenol
Concen: 44.379 ng
RT: 7.139 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Instrument :
BNA_F
ClientSampleId :
PB164444BS


Manual Integrations APPROVED







Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#18
Hexachloroethane
Concen: 44.169 ng
RT: 7.404 min Scan# 903
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:117 Resp: 150834
Ion Ratio Lower Upper
117 100
119 95.0 77.0 115.4
201 102.8 76.2 114.4





#19

n-Nitroso-di-n-propylamine

Concen: 43.903 ng

RT: 7.304 min Scan# 886

Delta R.T. 0.000 min

Lab File: BF140077.D


Acq: 28 Oct 2024 10:29

Instrument :

BNA_F

ClientSampleId :

PB164444BS

**Manual Integrations
APPROVED**
Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

#20

3+4-Methylphenols



Concen: 42.951 ng




RT: 7.292 min Scan# 884

Delta R.T. 0.006 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.175 min Scan# 1034

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Instrument:

BNA_F




ClientSampleId :

PB164444BS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#22

Acetophenone

Concen: 43.883 ng

RT: 7.298 min Scan# 885

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Tgt Ion:105 Resp: 532350


Ion Ratio Lower Upper


105 100

71 7.9 5.9 8.9

51 36.6 29.4 44.2

120 21.5 17.5 26.3





#23

Nitrobenzene-d5



Concen: 92.589 ng

RT: 7.451 min Scan# 911

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29



#24

Nitrobenzene


Concen: 43.461 ng

RT: 7.469 min Scan# 914

Delta R.T. -0.006 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29





Tgt Ion: 77 Resp: 425814

Ion Ratio Lower Upper

77 100

123 44.2 35.7 53.5

65 14.5 11.5 17.3



Tgt Ion: 82 Resp: 827395

Ion Ratio Lower Upper

82 100



128 42.3 33.4 50.0




54 57.7 47.8 71.8

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024





#25

Isophorone

Concen: 45.583 ng

RT: 7.710 min Scan# 955

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Instrument :

BNA_F




ClientSampleId :

PB164444BS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#26

2-Nitrophenol

Concen: 51.101 ng

RT: 7.786 min Scan# 968

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29


Tgt Ion:139 Resp: 188880




Ion Ratio Lower Upper

139 100

109 31.2 25.1 37.7

65 57.5 47.5 71.3





#27

2,4-Dimethylphenol

Concen: 52.015 ng

RT: 7.822 min Scan# 9

Delta R.T. 0.006 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Instrument :

BNA_F




ClientSampleId :

PB164444BS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#28

bis(2-Chloroethoxy)methane

Concen: 44.384 ng

RT: 7.916 min Scan# 990

Delta R.T. -0.006 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29


Tgt Ion: 93 Resp: 456094




Ion Ratio Lower Upper

93 100

95 32.7 26.0 39.0

123 11.9 9.8 14.8





#29

2,4-Dichlorophenol




Concen: 45.395 ng

RT: 8.028 min Scan# 1023

Delta R.T. 0.006 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Instrument: BNA_F
ClientSampleId: PB164444BS
**Manual Integrations
APPROVED**
Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

#30

1,2,4-Trichlorobenzene

Concen: 43.256 ng

RT: 8.110 min Scan# 1023

Delta R.T. -0.006 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29


Tgt Ion:180 Resp: 336007







Ion Ratio Lower Upper

180 100

182 98.4 80.5 120.7

145 32.4 25.4 38.0





#31

Naphthalene

Concen: 44.121 ng

RT: 8.192 min Scan# 1037

Delta R.T. -0.006 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Instrument:

BNA_F


ClientSampleId :

PB164444BS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#32

Benzoic acid

Concen: 40.344 ng


RT: 7.939 min Scan# 994







Delta R.T. 0.017 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Tgt	Ion:122	Resp:	216873
Ion Ratio	Lower	Upper	
122	100		
105	126.8	107.6	147.6
77	99.3	80.2	120.2





#33

4-Chloroaniline

Concen: 18.292 ng

RT: 8.239 min Scan# 1045

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Instrument :

BNA_F


ClientSampleId :

PB164444BS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#34

Hexachlorobutadiene

Concen: 44.923 ng


RT: 8.310 min Scan# 1057







Delta R.T. -0.006 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Tgt	Ion:225	Resp:	219251
Ion	Ratio	Lower	Upper
225	100		
223	62.2	50.6	76.0
227	63.2	51.8	77.8





#35

Caprolactam

Concen: 46.244 ng m

RT: 8.610 min Scan# 1108

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Instrument:

BNA_F


ClientSampleId :

PB164444BS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#36

4-Chloro-3-methylphenol

Concen: 45.972 ng

RT: 8.716 min Scan# 1126

Delta R.T. 0.006 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29


Tgt Ion:107 Resp: 357353


Ion Ratio Lower Upper

107 100

144 24.0 19.1 28.7

142 71.2 57.2 85.8








#37
2-Methylnaphthalene
Concen: 45.632 ng
RT: 8.886 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024











#38
1-Methylnaphthalene
Concen: 43.622 ng
RT: 8.986 min Scan# 1172
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

```

Tgt Ion:142 Resp: 669536
Ion Ratio Lower Upper
142 100
141 91.4 73.5 110.3

```





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.928 min Scan# 1

Delta R.T. 0.001 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Instrument :

BNA_F

ClientSampleId :

PB164444BS

Tgt Ion:164 Resp: 281323

Ion Ratio Lower Upper

164 100



162 100.9 81.0 121.4

160 44.3 35.4 53.0

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#40

1,2,4,5-Tetrachlorobenzene

Concen: 44.663 ng

RT: 9.051 min Scan# 1183

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Tgt Ion:216 Resp: 338977


Ion Ratio Lower Upper


216 100

214 79.1 63.4 95.0

179 21.1 17.2 25.8

108 21.4 15.6 23.4







#41
Hexachlorocyclopentadiene
Concen: 165.576 ng
RT: 9.039 min Scan# 1
Instrument : BNA_F
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

ClientSampleId : PB164444BS


Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#42
2,4,6-Tribromophenol
Concen: 141.398 ng
RT: 10.722 min Scan# 1467
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:330 Resp: 372077
Ion Ratio Lower Upper
330 100
332 96.2 78.1 117.1
141 30.1 26.6 39.8





#43

2,4,6-Trichlorophenol

Concen: 45.915 ng

RT: 9.163 min Scan# 1209

Delta R.T. 0.006 min

Lab File: BF140077.D


Acq: 28 Oct 2024 10:29

Instrument :

BNA_F

ClientSampleId :

PB164444BS

Manual Integrations
APPROVED
Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

#44

2,4,5-Trichlorophenol

Concen: 44.627 ng


RT: 9.204 min Scan# 1209







Delta R.T. 0.012 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Tgt	Ion:196	Resp:	247409
Ion	Ratio	Lower	Upper
196	100		
198	96.3	75.8	113.6
97	50.5	41.8	62.6
132	27.3	22.9	34.3





#45

2-Fluorobiphenyl

Concen: 87.245 ng

RT: 9.245 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Instrument:

BNA_F


ClientSampleId :

PB164444BS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#46

1,1'-Biphenyl



Concen: 43.865 ng




RT: 9.351 min Scan# 1234

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29





#47

2-Chloronaphthalene

Concen: 42.957 ng

RT: 9.375 min Scan# 1238

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29




Instrument :

BNA_F

ClientSampleId :

PB164444BS

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024


#48

2-Nitroaniline

Concen: 49.063 ng

RT: 9.469 min Scan# 1254

Delta R.T. 0.006 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29



Tgt Ion: 65 Resp: 236689

Ion Ratio Lower Upper

65 100

92 66.1 50.9 76.3

138 97.9 73.4 110.0



Abundance

Time-->







9.30 9.35 9.40 9.45

Time-->

9.40 9.45 9.50 9.55

Time-->

9.40 9.45 9.50 9.55



#49

Acenaphthylene
Concen: 47.749 ng

RT: 9.792 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Instrument :

BNA_F


ClientSampleId :

PB164444BS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#50

Dimethylphthalate

Concen: 45.933 ng

RT: 9.651 min Scan# 1285

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29


Tgt Ion:163 Resp: 804892







Ion Ratio Lower Upper

163 100

194 4.1 3.4 5.0

164 10.1 8.2 12.2





#51

2,6-Dinitrotoluene

Concen: 47.412 ng

RT: 9.710 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29


Instrument :

BNA_F

ClientSampleId :

PB164444BS

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024


#52

Acenaphthene

Concen: 49.881 ng

RT: 9.963 min Scan# 1338

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29


Tgt Ion:154 Resp: 735816







Ion Ratio Lower Upper

154 100


153 111.6 89.8 134.6

152 54.0 42.4 63.6







#53

3-Nitroaniline
Concen: 28.091 ngRT: 9.875 min Scan# 1323
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29Instrument :
BNA_F
ClientSampleId :
PB164444BS**Manual Integrations
APPROVED**Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024


#54

2,4-Dinitrophenol
Concen: 109.904 ng
RT: 9.986 min Scan# 1342
Delta R.T. 0.012 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29Tgt Ion:184 Resp: 177437
Ion Ratio Lower Upper
184 100
63 73.3 84.7 127.1#
154 64.0 338.2 507.2#



#55
Dibenzofuran
Concen: 44.771 ng
RT: 10.133 min Scan# 1367
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29



Instrument : BNA_F
ClientSampleId : PB164444BS




Tgt Ion:168 Resp: 947573
Ion Ratio Lower Upper
168 100
139 40.0 31.8 47.6
169 14.3 11.6 17.4

Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#56
4-Nitrophenol
Concen: 95.186 ng
RT: 10.033 min Scan# 1350
Delta R.T. 0.012 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29



Tgt Ion:139 Resp: 286527
Ion Ratio Lower Upper
139 100
109 76.9 53.6 93.6
65 112.6 87.8 127.8







#57
2,4-Dinitrotoluene
Concen: 52.385 ng
RT: 10.116 min Scan# 1 Instrument :
Delta R.T. 0.006 min BNA_F
Lab File: BF140077.D ClientSampleId :
Acq: 28 Oct 2024 10:29 PB164444BS


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#58
Fluorene
Concen: 45.154 ng
RT: 10.480 min Scan# 1426
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt	Ion:166	Resp:	727905
Ion	Ratio	Lower	Upper
166	100		
165	97.4	78.5	117.7
167	13.2	10.9	16.3






#59
2,3,4,6-Tetrachlorophenol
Concen: 48.851 ng
RT: 10.251 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Instrument :
BNA_F
ClientSampleId :
PB164444BS


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#60
Diethylphthalate
Concen: 45.976 ng
RT: 10.351 min Scan# 1404
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:149 Resp: 791021
Ion Ratio Lower Upper
149 100
177 22.2 17.9 26.9
150 12.7 10.1 15.1







#61
4-Chlorophenyl-phenylether
Concen: 45.336 ng
RT: 10.469 min Scan# 1424
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Instrument : BNA_F
ClientSampleId : PB164444BS


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#62
4-Nitroaniline
Concen: 48.958 ng
RT: 10.498 min Scan# 1429
Delta R.T. 0.012 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:138 Resp: 180917
Ion Ratio Lower Upper
138 100
92 51.3 35.7 75.7
108 77.4 57.5 97.5





#63

Azobenzene

Concen: 44.214 ng

RT: 10.627 min Scan# 1451

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Instrument :

BNA_F

ClientSampleId :

PB164444BS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024

Tgt Ion: 77 Resp: 806466


Ion Ratio Lower Upper

77 100

182 24.9 3.3 43.3

105 21.4 1.2 41.2

51 32.5 13.3 53.3



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.416 min Scan# 1585

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29


Tgt Ion:188 Resp: 511374







Ion Ratio Lower Upper

188 100

94 9.4 7.9 11.9

80 10.3 9.0 13.4






#65
4,6-Dinitro-2-methylphenol
Concen: 60.564 ng
RT: 10.522 min Scan# 1444
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Instrument :
BNA_F
ClientSampleId :
PB164444BS


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#66
n-Nitrosodiphenylamine
Concen: 44.467 ng
RT: 10.586 min Scan# 1444
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:169 Resp: 680576
Ion Ratio Lower Upper
169 100
168 62.5 49.8 74.6
167 34.8 28.0 42.0









#67
4-Bromophenyl-phenylether
Concen: 45.173 ng
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Instrument :
BNA_F
ClientSampleId :
PB164444BS


Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#68
Hexachlorobenzene
Concen: 44.950 ng
RT: 11.027 min Scan# 1519
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:284 Resp: 266322
Ion Ratio Lower Upper
284 100
142 29.5 26.9 40.3
249 31.2 25.3 37.9






#69
Atrazine
Concen: 54.080 ng
RT: 11.116 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Instrument : BNA_F
ClientSampleId : PB164444BS


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#70
Pentachlorophenol
Concen: 89.731 ng
RT: 11.216 min Scan# 1551
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:266 Resp: 322659
Ion Ratio Lower Upper
266 100
268 62.7 50.4 75.6
264 62.1 50.6 75.8





#71

Phenanthrene

Concen: 44.500 ng

RT: 11.439 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Instrument:

BNA_F




ClientSampleId :

PB164444BS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#72

Anthracene

Concen: 47.004 ng



RT: 11.492 min Scan# 1598







Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Tgt	Ion:178	Resp:	1107771
Ion	Ratio	Lower	Upper
178	100		
176	19.2	15.3	22.9
179	15.6	12.4	18.6





#73

Carbazole
Concen: 43.475 ng

RT: 11.645 min Scan# 1


Delta R.T. 0.006 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29


Instrument :
BNA_F
ClientSampleId :
PB164444BS


Tgt Ion:	Ion Ratio	Resp:	952915
		Lower	Upper
167	100		
166	21.6	17.0	25.6
139	13.2	10.8	16.2

**Manual Integrations
APPROVED**Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

#74
Di-n-butylphthalate
Concen: 46.086 ng
RT: 11.974 min Scan# 1680
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:	Ion Ratio	Resp:	1167048
		Lower	Upper
149	100		
150	9.1	7.5	11.3
104	5.0	4.1	6.1







#75
Fluoranthene
Concen: 43.973 ng
RT: 12.627 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Instrument : BNA_F
ClientSampleId : PB164444BS


Manual Integrations APPROVED







Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#76
Chrysene-d₁₂
Concen: 20.000 ng
RT: 14.051 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:240 Resp: 262593
Ion Ratio Lower Upper
240 100
120 11.1 9.4 14.2
236 26.5 20.9 31.3





#77

Benzidine

Concen: 55.056 ng

RT: 12.745 min Scan# 1811

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Instrument:

BNA_F


ClientSampleId :

PB164444BS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#78

Pyrene

Concen: 46.552 ng

RT: 12.857 min Scan# 1830

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29


Tgt Ion:202 Resp: 1070018



Ion Ratio Lower Upper

202 100

200 21.8 17.2 25.8

203 17.7 14.2 21.4





#79

Terphenyl-d14

Concen: 95.744 ng

RT: 12.998 min Scan# 1854

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Instrument:

BNA_F


ClientSampleId :

PB164444BS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#80

Butylbenzylphthalate

Concen: 52.081 ng

RT: 13.468 min Scan# 1934

Delta R.T. -0.006 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29


Tgt Ion:149 Resp: 358899







Ion Ratio Lower Upper

149 100

91 74.9 59.4 89.2

206 20.0 16.4 24.6





#81

Benzo(a)anthracene

Concen: 48.236 ng

RT: 14.045 min Scan# 2031

Delta R.T. 0.006 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29


Instrument :

BNA_F

ClientSampleId :

PB164444BS

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024


#82

3,3'-Dichlorobenzidine

Concen: 32.497 ng

RT: 14.004 min Scan# 2025

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29


Tgt Ion:252 Resp: 161964







Ion Ratio Lower Upper

252 100

254 64.2 50.8 76.2

126 10.5 8.9 13.3





#83

Chrysene

Concen: 46.850 ng

RT: 14.080 min Scan# 2038

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29


Instrument :

BNA_F

ClientSampleId :

PB164444BS

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024


#84

Bis(2-ethylhexyl)phthalate

Concen: 56.091 ng

RT: 14.027 min Scan# 2029

Delta R.T. -0.006 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29


Tgt Ion:149 Resp: 433668



Ion Ratio Lower Upper

149 100

167 26.6 21.4 32.0

279 4.0 3.0 4.4





#85

Di-n-octyl phthalate

Concen: 49.743 ng

RT: 14.639 min Scan# 2

Delta R.T. -0.012 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Instrument :

BNA_F


ClientSampleId :

PB164444BS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.527 min Scan# 2284

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29


Tgt Ion:264 Resp: 265684







Ion Ratio Lower Upper

264 100

260 24.0 19.4 29.2

265 21.8 17.4 26.0





#87

Indeno(1,2,3-cd)pyrene

Concen: 49.096 ng

RT: 17.033 min Scan# 2

Delta R.T. 0.012 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Instrument :

BNA_F


ClientSampleId :

PB164444BS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#88

Benzo(b)fluoranthene

Concen: 43.570 ng


RT: 15.098 min Scan# 2211


Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Tgt	Ion	252	Resp:	705155
Ion	Ratio	Lower	Upper	
252	100			
253	21.9	17.6	26.4	
125	9.4	7.9	11.9	





#89

Benzo(k)fluoranthene

Concen: 50.999 ng

RT: 15.127 min Scan# 2216

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29

Instrument :

BNA_F

ClientSampleId :

PB164444BS

Tgt Ion:252 Resp: 711830

Ion Ratio Lower Upper

252 100



253 22.3 17.6 26.4

125 9.1 7.9 11.9

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#90

Benzo(a)pyrene

Concen: 50.645 ng

RT: 15.468 min Scan# 2274

Delta R.T. 0.000 min

Lab File: BF140077.D

Acq: 28 Oct 2024 10:29


Tgt Ion:252 Resp: 674445


Ion Ratio Lower Upper

252 100

253 22.1 17.3 25.9

125 10.7 8.7 13.1







#91
Dibenzo(a,h)anthracene
Concen: 47.676 ng
RT: 17.051 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Instrument : BNA_F
ClientSampleId : PB164444BS


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#92
Benzo(g,h,i)perylene
Concen: 45.246 ng
RT: 17.486 min Scan# 2617
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:276 Resp: 644416
Ion Ratio Lower Upper
276 100
277 23.7 19.0 28.6
138 20.0 17.4 26.0





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	10/24/24	
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	10/24/24	
Client Sample ID:	VNJ-215MS			SDG No.:	P4578	
Lab Sample ID:	P4545-02MS			Matrix:	TCLP	
Analytical Method:	SW8270			% Solid:	0	
Sample Wt/Vol:	100	Units:	mL	Final Vol:	1000	uL
Soil Aliquot Vol:	uL			Test:	TCLP BNA	
Extraction Type :				Decanted :	N	Level :
Injection Volume :				GPC Factor :	1.0	GPC Cleanup : N PH :
Prep Method :	SW3510C					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140092.D	1	10/26/24 12:10	10/28/24 17:35	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	230		15.5	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	430		8.40	50.0	ug/L
95-48-7	2-Methylphenol	490		11.3	50.0	ug/L
65794-96-9	3+4-Methylphenols	480		11.5	100	ug/L
67-72-1	Hexachloroethane	420		10.1	50.0	ug/L
98-95-3	Nitrobenzene	470		12.7	50.0	ug/L
87-68-3	Hexachlorobutadiene	460		12.7	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	550		8.90	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	540		10.1	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	600		15.2	50.0	ug/L
118-74-1	Hexachlorobenzene	540		11.4	50.0	ug/L
87-86-5	Pentachlorophenol	780		18.5	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	135		15 (10) - 110 (139)	90%	SPK: 150
13127-88-3	Phenol-d6	129		15 (10) - 110 (134)	86%	SPK: 150
4165-60-0	Nitrobenzene-d5	103		30 (49) - 130 (133)	103%	SPK: 100
321-60-8	2-Fluorobiphenyl	98.8		30 (52) - 130 (132)	99%	SPK: 100
118-79-6	2,4,6-Tribromophenol	159		15 (44) - 110 (137)	106%	SPK: 150
1718-51-0	Terphenyl-d14	105		30 (48) - 130 (125)	105%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	114000	6.892			
1146-65-2	Naphthalene-d8	442000	8.169			
15067-26-2	Acenaphthene-d10	242000	9.928			
1517-22-2	Phenanthrene-d10	410000	11.41			
1719-03-5	Chrysene-d12	203000	14.051			
1520-96-3	Perylene-d12	263000	15.527			



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	10/24/24	
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	10/24/24	
Client Sample ID:	V NJ-215MS			SDG No.:	P4578	
Lab Sample ID:	P4545-02MS			Matrix:	TCLP	
Analytical Method:	SW8270			% Solid:	0	
Sample Wt/Vol:	100	Units:	mL	Final Vol:	1000	uL
Soil Aliquot Vol:	uL			Test:	TCLP BNA	
Extraction Type :				Decanted :	N	Level :
Injection Volume :				GPC Factor :	1.0	GPC Cleanup : N PH :
Prep Method :	SW3510C					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140092.D	1	10/26/24 12:10	10/28/24 17:35	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140092.D
 Acq On : 28 Oct 2024 17:35
 Operator : RC/JU
 Sample : P4545-02MS
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 VNJ-215MS

Quant Time: Oct 28 18:05:02 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
 10/30/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
Internal Standards							10/29/2024
1) 1,4-Dichlorobenzene-d4	6.892	152	113612	20.000	ng	0.00	Supervised By :mohammad ahmed
21) Naphthalene-d8	8.169	136	441889	20.000	ng	0.00	
39) Acenaphthene-d10	9.928	164	241647	20.000	ng	0.00	
64) Phenanthrene-d10	11.410	188	409717	20.000	ng	0.00	
76) Chrysene-d12	14.051	240	203129	20.000	ng	0.00	
86) Perylene-d12	15.527	264	263019	20.000	ng	0.00	10/30/2024

System Monitoring Compounds

5) 2-Fluorophenol	5.522	112	977769	134.729	ng	0.02
7) Phenol-d6	6.516	99	1208635	128.587	ng	0.00
23) Nitrobenzene-d5	7.451	82	821761	103.069	ng	0.00
42) 2,4,6-Tribromophenol	10.716	330	358712	158.702	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1444688	98.807	ng	0.00
79) Terphenyl-d14	12.998	244	1306207	104.783	ng	0.00

Target Compounds

				Qvalue	
2) 1,4-Dioxane	2.881	88	126476	37.378	ng # 82
3) Pyridine	3.587	79	189602	22.606	ng 98
4) n-Nitrosodimethylamine	3.516	42	187502	41.610	ng # 97
6) Aniline	6.557	93	172468	19.688	ng 99
8) 2-Chlorophenol	6.675	128	370207	49.899	ng 99
9) Benzaldehyde	6.440	77	47671	9.016	ng 99
10) Phenol	6.528	94	445105m	45.933	ng
11) bis(2-Chloroethyl)ether	6.628	93	351685	46.955	ng 99
12) 1,3-Dichlorobenzene	6.834	146	361312	42.425	ng 99
13) 1,4-Dichlorobenzene	6.910	146	367844	43.232	ng 99
14) 1,2-Dichlorobenzene	7.057	146	355662	44.788	ng 99
15) Benzyl Alcohol	7.028	79	347790	50.442	ng 99
16) 2,2'-oxybis(1-Chloropr...	7.163	45	584140	45.739	ng 97
17) 2-Methylphenol	7.134	107	310465	49.075	ng 99
18) Hexachloroethane	7.404	117	128123	42.227	ng 96
19) n-Nitroso-di-n-propyla...	7.304	70	276089	48.431	ng 95
20) 3+4-Methylphenols	7.287	107	391991	48.443	ng 94
22) Acetophenone	7.298	105	508513	46.983	ng 98
24) Nitrobenzene	7.469	77	408928	46.780	ng 100
25) Isophorone	7.710	82	765851	50.609	ng 99
26) 2-Nitrophenol	7.787	139	189690	57.521	ng 97
27) 2,4-Dimethylphenol	7.816	122	314864	57.260	ng 98
28) bis(2-Chloroethoxy)met...	7.916	93	447113	48.767	ng 99
29) 2,4-Dichlorophenol	8.022	162	320220	51.126	ng 100
30) 1,2,4-Trichlorobenzene	8.110	180	316171	45.620	ng 100
31) Naphthalene	8.192	128	1066174	46.783	ng 100
32) Benzoic acid	7.898	122	84400	17.598	ng 95
33) 4-Chloroaniline	8.234	127	133599	17.192	ng 98
34) Hexachlorobutadiene	8.310	225	199456	45.805	ng 99
35) Caprolactam	8.604	113	83595m	41.975	ng
36) 4-Chloro-3-methylphenol	8.716	107	363875	52.467	ng 99
37) 2-Methylnaphthalene	8.881	142	696991	49.937	ng 100
38) 1-Methylnaphthalene	8.981	142	647449	47.280	ng 99
40) 1,2,4,5-Tetrachloroben...	9.051	216	331367	50.829	ng 99
41) Hexachlorocyclopentadiene	9.034	237	362304	159.719	ng 100
43) 2,4,6-Trichlorophenol	9.157	196	253954	55.021	ng 99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140092.D
 Acq On : 28 Oct 2024 17:35
 Operator : RC/JU
 Sample : P4545-02MS
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 VNJ-215MS

Quant Time: Oct 28 18:05:02 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
44) 2,4,5-Trichlorophenol	9.204	196	258939	54.376	ng	98	10/29/2024
46) 1,1'-Biphenyl	9.345	154	856147	50.894	ng	100	Supervised By :mohammad ahmed
47) 2-Chloronaphthalene	9.375	162	675337	49.845	ng	100	
48) 2-Nitroaniline	9.463	65	241309	58.234	ng	97	
49) Acenaphthylene	9.786	152	1080746	55.175	ng	99	
50) Dimethylphthalate	9.645	163	822973	54.677	ng	99	
51) 2,6-Dinitrotoluene	9.710	165	178715	54.837	ng	96	10/30/2024
52) Acenaphthene	9.963	154	661593	52.213	ng	99	
53) 3-Nitroaniline	9.875	138	131503	39.128	ng	98	
54) 2,4-Dinitrophenol	9.980	184	60419	47.508	ng	#	1
55) Dibenzofuran	10.133	168	947629	52.125	ng	100	
56) 4-Nitrophenol	10.033	139	254819	98.551	ng	98	
57) 2,4-Dinitrotoluene	10.110	165	238568	59.527	ng	99	
58) Fluorene	10.475	166	716535	51.747	ng	99	
59) 2,3,4,6-Tetrachlorophenol	10.245	232	201940	54.095	ng	98	
60) Diethylphthalate	10.345	149	792061	53.595	ng	99	
61) 4-Chlorophenyl-phenyle...	10.469	204	368778	52.737	ng	98	
62) 4-Nitroaniline	10.492	138	167633	52.812	ng	98	
63) Azobenzene	10.628	77	807556	51.543	ng	97	
65) 4,6-Dinitro-2-methylph...	10.516	198	60965	36.479	ng	94	
66) n-Nitrosodiphenylamine	10.586	169	670666	54.692	ng	100	
67) 4-Bromophenyl-phenylether	10.957	248	233236	55.258	ng	97	
68) Hexachlorobenzene	11.022	284	255815	53.889	ng	98	
69) Atrazine	11.110	200	212035	63.831	ng	99	
70) Pentachlorophenol	11.216	266	224019	77.757	ng	99	
71) Phenanthrene	11.439	178	1031591	53.303	ng	100	
72) Anthracene	11.492	178	1041472	55.155	ng	100	
73) Carbazole	11.639	167	882304	50.241	ng	100	
74) Di-n-butylphthalate	11.969	149	1119955	55.199	ng	100	
75) Fluoranthene	12.627	202	912791	46.655	ng	99	
77) Benzidine	12.745	184	17706	5.301	ng	98	
78) Pyrene	12.851	202	903442	50.811	ng	100	
80) Butylbenzylphthalate	13.469	149	345144	64.747	ng	99	
81) Benzo(a)anthracene	14.039	228	724089	54.760	ng	100	
82) 3,3'-Dichlorobenzidine	14.004	252	185837	48.202	ng	100	
83) Chrysene	14.080	228	701255	57.841	ng	99	
84) Bis(2-ethylhexyl)phtha...	14.027	149	456891	76.394	ng	100	
85) Di-n-octyl phthalate	14.639	149	793362	72.931	ng	98	
87) Indeno(1,2,3-cd)pyrene	17.027	276	775966	45.859	ng	98	
88) Benzo(b)fluoranthene	15.098	252	831851	51.919	ng	99	
89) Benzo(k)fluoranthene	15.127	252	782283	56.615	ng	99	
90) Benzo(a)pyrene	15.468	252	780503	59.203	ng	99	
91) Dibenzo(a,h)anthracene	17.045	278	651067	46.124	ng	99	
92) Benzo(g,h,i)perylene	17.474	276	529084	37.525	ng	99	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140092.D
 Acq On : 28 Oct 2024 17:35
 Operator : RC/JU
 Sample : P4545-02MS
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Quant Time: Oct 28 18:05:02 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Instrument :
 BNA_F
 ClientSampleId :
 VNJ-215MS

Manual Integrations APPROVED


Reviewed By :Yogesh Patel




Supervised By :mohammad ahmed

10/30/2024

10/29/2024
 Supervised By :mohammad ahmed

10/30/2024





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.892 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS




Manual Integrations APPROVED

Reviewed By :Yogesh
Patel

Supervised By :mohammad
ahmed
10/30/2024


10/29/2024
Supervised By :mohammad
ahmed


10/30/2024



#2
1,4-Dioxane
Concen: 37.378 ng
RT: 2.881 min Scan# 134
Delta R.T. 0.194 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion: 88 Resp: 126476
Ion Ratio Lower Upper
88 100
58 89.6 73.3 109.9
43 0.0 27.1 40.7#





#3
Pyridine
Concen: 22.606 ng
RT: 3.587 min Scan# 21
Delta R.T. 0.141 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Instrument : BNA_F
ClientSampleId : VNJ-215MS



Manual Integrations APPROVED

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/2024


10/29/2024
Supervised By :mohammad ahmed







10/30/2024



#4
n-Nitrosodimethylamine
Concen: 41.610 ng
RT: 3.516 min Scan# 242
Delta R.T. 0.118 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion: 42 Resp: 187502
Ion Ratio Lower Upper
42 100
74 123.8 96.2 144.4
44 11.0 6.2 9.2#





#5

2-Fluorophenol

Concen: 134.729 ng

RT: 5.522 min Scan# 5

Delta R.T. 0.024 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument :

BNA_F

ClientSampleId :

VNJ-215MS

Tgt Ion:112 Resp: 977769

Ion Ratio Lower Upper

112 100

64 65.0 53.0 79.6


63 33.4 27.0 40.4

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/202410/29/2024
Supervised By :mohammad ahmed

10/30/2024



#6

Aniline

Concen: 19.688 ng

RT: 6.557 min Scan# 759

Delta R.T. 0.006 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35


Tgt Ion: 93 Resp: 172468



Ion Ratio Lower Upper

93 100

66 42.0 34.0 51.0

65 21.4 17.4 26.0





#7

Phenol-d6
Concen: 128.587 ng
RT: 6.516 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS


Manual Integrations APPROVED

Reviewed By :Yogesh
Patel

Supervised By :mohammad
ahmed
10/30/2024

10/29/2024
Supervised By :mohammad
ahmed


10/30/2024




#8

2-Chlorophenol
Concen: 49.899 ng
RT: 6.675 min Scan# 779
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:128 Resp: 370207
Ion Ratio Lower Upper
128 100
130 32.8 12.3 52.3
64 53.5 34.6 74.6





#9
Benzaldehyde
Concen: 9.016 ng
RT: 6.440 min Scan# 7
Delta R.T. 0.001 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Instrument : BNA_F
ClientSampleId : VNJ-215MS



Manual Integrations
APPROVED

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/2024


10/29/2024
Supervised By :mohammad ahmed







10/30/2024



#10
Phenol
Concen: 45.933 ng
RT: 6.528 min Scan# 754
Delta R.T. 0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion: 94 Resp: 445105
Ion Ratio Lower Upper
94 100
65 32.3 11.1 51.1
66 45.7 22.7 62.7





#11
bis(2-Chloroethyl)ether
Concen: 46.955 ng
RT: 6.628 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS


Manual Integrations APPROVED

Reviewed By :Yogesh
Patel

Supervised By :mohammad
ahmed
10/30/2024


10/29/2024
Supervised By :mohammad
ahmed







10/30/2024



#12
1,3-Dichlorobenzene
Concen: 42.425 ng
RT: 6.834 min Scan# 806
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:146 Resp: 361312
Ion Ratio Lower Upper
146 100
148 64.1 51.0 76.4
75 31.4 26.4 39.6






#13

1,4-Dichlorobenzene
Concen: 43.232 ngRT: 6.910 min Scan# 819
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35Instrument :
BNA_F
ClientSampleId :
VNJ-215MS


Tgt Ion:146 Resp: 367844


Ion Ratio Lower Upper
146 100
148 64.2 51.8 77.8
111 42.6 34.6 51.8**Manual Integrations
APPROVED**Reviewed By :Yogesh
PatelSupervised By :mohammad
ahmed
10/30/202410/29/2024
Supervised By :mohammad
ahmed

10/30/2024



#14

1,2-Dichlorobenzene
Concen: 44.788 ng
RT: 7.057 min Scan# 844
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35Tgt Ion:146 Resp: 355662
Ion Ratio Lower Upper
146 100
148 64.4 51.4 77.2
111 45.0 36.7 55.1



#15

Benzyl Alcohol
Concen: 50.442 ng

RT: 7.028 min Scan# 8

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument : BNA_F

ClientSampleId :



VNJ-215MS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/202410/29/2024
Supervised By :mohammad ahmed

10/30/2024



#16

2,2'-oxybis(1-Chloropropane)

Concen: 45.739 ng

RT: 7.163 min Scan# 862

Delta R.T. -0.006 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35


Tgt Ion: 45 Resp: 584140




Ion Ratio Lower Upper

45 100

77 13.0 0.0 31.5

79 9.6 0.0 29.1





#17

2-Methylphenol

Concen: 49.075 ng

RT: 7.134 min Scan# 8

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument :

BNA_F

ClientSampleId :




VNJ-215MS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/202410/29/2024
Supervised By :mohammad ahmed

10/30/2024



#18

Hexachloroethane

Concen: 42.227 ng

RT: 7.404 min Scan# 903

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35


Tgt Ion:117 Resp: 128123




Ion Ratio Lower Upper

117 100

119 96.3 77.0 115.4

201 103.7 76.2 114.4





#19
n-Nitroso-di-n-propylamine
Concen: 48.431 ng
RT: 7.304 min Scan# 886
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS





Manual Integrations APPROVED

Reviewed By :Yogesh
Patel

Supervised By :mohammad
ahmed
10/30/2024


10/29/2024
Supervised By :mohammad
ahmed







10/30/2024



#20
3+4-Methylphenols
Concen: 48.443 ng
RT: 7.287 min Scan# 883
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:107 Resp: 391991
Ion Ratio Lower Upper
107 100
108 89.9 69.8 109.8
77 83.7 50.5 90.5
79 29.5 10.2 50.2





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.169 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument :

BNA_F

ClientSampleId :


VNJ-215MS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/202410/29/2024
Supervised By :mohammad ahmed

10/30/2024



#22

Acetophenone

Concen: 46.983 ng


RT: 7.298 min Scan# 885







Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Tgt	Ion:105	Resp:	508513
Ion	Ratio	Lower	Upper
105	100		
71	8.1	5.9	8.9
51	34.9	29.4	44.2
120	21.7	17.5	26.3





#23

Nitrobenzene-d5

Concen: 103.069 ng

RT: 7.451 min Scan# 911

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument :

BNA_F

ClientSampleId :


VNJ-215MS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/202410/29/2024
Supervised By :mohammad ahmed

10/30/2024



#24

Nitrobenzene

Concen: 46.780 ng

RT: 7.469 min Scan# 914

Delta R.T. -0.006 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35


Tgt Ion: 77 Resp: 408928







Ion Ratio Lower Upper

77 100

123 44.8 35.7 53.5

65 14.4 11.5 17.3







#25


Isophorone
Concen: 50.609 ngRT: 7.710 min Scan# 91
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35Instrument :
BNA_F
ClientSampleId :
VNJ-215MS**Manual Integrations
APPROVED**Reviewed By :Yogesh
PatelSupervised By :mohammad
ahmed
10/30/202410/29/2024
Supervised By :mohammad
ahmed

10/30/2024



#26

2-Nitrophenol
Concen: 57.521 ng
RT: 7.787 min Scan# 968
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35Tgt Ion:139 Resp: 189690
Ion Ratio Lower Upper
139 100
109 30.8 25.1 37.7
65 56.2 47.5 71.3



#27
2,4-Dimethylphenol
Concen: 57.260 ng
RT: 7.816 min Scan# 973
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS


Manual Integrations APPROVED

Reviewed By :Yogesh
Patel

Supervised By :mohammad
ahmed
10/30/2024

10/29/2024
Supervised By :mohammad
ahmed







10/30/2024



#28
bis(2-Chloroethoxy)methane
Concen: 48.767 ng
RT: 7.916 min Scan# 990
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion: 93 Resp: 447113
Ion Ratio Lower Upper
93 100
95 32.8 26.0 39.0
123 11.2 9.8 14.8





#29

2,4-Dichlorophenol
Concen: 51.126 ng

RT: 8.022 min Scan# 1008

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument : BNA_F

ClientSampleId : VNJ-215MS

Tgt Ion:162 Resp: 320220

Ion Ratio Lower Upper

162 100

164 63.7 43.9 83.9


98 40.2 19.9 59.9

Manual Integrations APPROVED

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/202410/29/2024
Supervised By :mohammad ahmed

10/30/2024



#30

1,2,4-Trichlorobenzene
Concen: 45.620 ng
RT: 8.110 min Scan# 1023
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35


Tgt Ion:180 Resp: 316171


Ion Ratio Lower Upper

180 100

182 100.3 80.5 120.7

145 32.3 25.4 38.0





#31

Naphthalene

Concen: 46.783 ng

RT: 8.192 min Scan# 1066174

Delta R.T. -0.006 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument:

BNA_F

ClientSampleId :


VNJ-215MS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/202410/29/2024
Supervised By :mohammad ahmed

10/30/2024



#32

Benzoic acid

Concen: 17.598 ng


RT: 7.898 min Scan# 987







Delta R.T. -0.024 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Tgt	Ion:122	Resp:	84400
Ion	Ratio	Lower	Upper
122	100		
105	122.6	107.6	147.6
77	94.4	80.2	120.2





#33

4-Chloroaniline

Concen: 17.192 ng

RT: 8.234 min Scan# 1044

Delta R.T. -0.006 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument:

BNA_F

ClientSampleId :


VNJ-215MS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/202410/29/2024
Supervised By :mohammad ahmed

10/30/2024



#34

Hexachlorobutadiene

Concen: 45.805 ng


RT: 8.310 min Scan# 1057







Delta R.T. -0.006 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Tgt	Ion:225	Resp:	199456
Ion	Ratio	Lower	Upper
225	100		
223	62.7	50.6	76.0
227	63.7	51.8	77.8





#35

Caprolactam

Concen: 41.975 ng m

RT: 8.604 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument:

BNA_F

ClientSampleId :

VNJ-215MS

Tgt Ion:113 Resp: 83595

Ion Ratio Lower Upper

113 100

55 199.8 189.2 229.2

56 147.9 132.5 172.5


Manual Integrations

APPROVED

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/202410/29/2024
Supervised By :mohammad ahmed

10/30/2024



#36

4-Chloro-3-methylphenol

Concen: 52.467 ng

RT: 8.716 min Scan# 1126

Delta R.T. 0.006 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35


Tgt Ion:107 Resp: 363875







Ion Ratio Lower Upper

107 100

144 24.4 19.1 28.7

142 72.4 57.2 85.8





#37

2-Methylnaphthalene

Concen: 49.937 ng

RT: 8.881 min Scan# 1155

Delta R.T. -0.006 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument :

BNA_F

ClientSampleId :


VNJ-215MS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/202410/29/2024
Supervised By :mohammad ahmed

10/30/2024



#38

1-Methylnaphthalene

Concen: 47.280 ng

RT: 8.981 min Scan# 1171

Delta R.T. -0.006 min

Lab File: BF140092.D


Acq: 28 Oct 2024 17:35




Tgt Ion:142 Resp: 647449

Ion Ratio Lower Upper

142 100

141 92.6 73.5 110.3





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.928 min Scan# 1

Delta R.T. 0.001 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument : BNA_F
ClientSampleId : VNJ-215MS

Tgt Ion:164 Resp: 24164

Ion Ratio Lower Upper

164 100

162 101.1 81.0 121.4





160 44.2 35.4 53.0

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/202410/29/2024
Supervised By :mohammad ahmed

10/30/2024



#40

1,2,4,5-Tetrachlorobenzene

Concen: 50.829 ng

RT: 9.051 min Scan# 1183

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Tgt Ion:216 Resp: 331367


Ion Ratio Lower Upper


216 100

214 78.7 63.4 95.0

179 21.2 17.2 25.8

108 21.1 15.6 23.4





#41
Hexachlorocyclopentadiene
Concen: 159.719 ng
RT: 9.034 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS



Manual Integrations APPROVED

Reviewed By :Yogesh
Patel

Supervised By :mohammad
ahmed
10/30/2024


10/29/2024
Supervised By :mohammad
ahmed

10/30/2024



#42
2,4,6-Tribromophenol
Concen: 158.702 ng
RT: 10.716 min Scan# 1466
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:330 Resp: 358712
Ion Ratio Lower Upper
330 100
332 96.9 78.1 117.1
141 29.1 26.6 39.8



#43

2,4,6-Trichlorophenol

Concen: 55.021 ng

RT: 9.157 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument :

BNA_F

ClientSampleId :

VNJ-215MS

Tgt Ion:196 Resp: 253954

Ion Ratio Lower Upper


196 100

198 93.7 76.2 114.2

200 29.8 24.0 36.0

Manual Integrations**APPROVED**Reviewed By :Yogesh
PatelSupervised By :mohammad
ahmed
10/30/202410/29/2024
Supervised By :mohammad
ahmed

10/30/2024



#44

2,4,5-Trichlorophenol

Concen: 54.376 ng

RT: 9.204 min Scan# 1209

Delta R.T. 0.012 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Tgt Ion:196 Resp: 258939




Ion Ratio Lower Upper







196 100

198 94.2 75.8 113.6

97 49.0 41.8 62.6

132 26.5 22.9 34.3






#45

2-Fluorobiphenyl
Concen: 98.807 ngRT: 9.245 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Tgt Ion:	Ion Ratio	Resp:	Lower	Upper
172	100			
171	35.9	1444688	28.6	43.0
170	23.9		19.1	28.7

**Manual Integrations
APPROVED**Reviewed By :Yogesh
PatelSupervised By :mohammad
ahmed
10/30/202410/29/2024
Supervised By :mohammad
ahmed


10/30/2024









#46

1,1'-Biphenyl
Concen: 50.894 ng
RT: 9.345 min Scan# 1233
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:	Ion Ratio	Resp:	Lower	Upper
154	100			
153	40.5	856147	20.3	60.3
76	13.7		0.0	33.9





#47

2-Chloronaphthalene

Concen: 49.845 ng

RT: 9.375 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument :

BNA_F

ClientSampleId :

VNJ-215MS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/202410/29/2024
Supervised By :mohammad ahmed

10/30/2024

#48

2-Nitroaniline

Concen: 58.234 ng

RT: 9.463 min Scan# 1253

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Tgt Ion: 65 Resp: 241309


Ion Ratio Lower Upper







65 100

92 65.7 50.9 76.3

138 94.2 73.4 110.0

Abundance







#49







Acenaphthylene
Concen: 55.175 ngRT: 9.786 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35Instrument :
BNA_F
ClientSampleId :
VNJ-215MS**Manual Integrations
APPROVED**Reviewed By :Yogesh
PatelSupervised By :mohammad
ahmed
10/30/202410/29/2024
Supervised By :mohammad
ahmed

10/30/2024



#50

Dimethylphthalate
Concen: 54.677 ng
RT: 9.645 min Scan# 1284
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35Tgt Ion:163 Resp: 822973
Ion Ratio Lower Upper
163 100
194 4.1 3.4 5.0
164 10.5 8.2 12.2



#51

2,6-Dinitrotoluene

Concen: 54.837 ng

RT: 9.710 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument :


BNA_F

ClientSampleId :

VNJ-215MS

**Manual Integrations
APPROVED**
Reviewed By :Yogesh
PatelSupervised By :mohammad
ahmed
10/30/202410/29/2024
Supervised By :mohammad
ahmed

10/30/2024



#52

Acenaphthene

Concen: 52.213 ng

RT: 9.963 min Scan# 1338

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35


Tgt Ion:154 Resp: 661593


Ion Ratio Lower Upper

154 100

153 113.2 89.8 134.6

152 53.7 42.4 63.6





#53
3-Nitroaniline
Concen: 39.128 ng
RT: 9.875 min Scan# 131503
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Instrument : BNA_F
ClientSampleId : VNJ-215MS



Manual Integrations
APPROVED

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/2024


10/29/2024
Supervised By :mohammad ahmed


10/30/2024



#54
2,4-Dinitrophenol
Concen: 47.508 ng
RT: 9.980 min Scan# 1341
Delta R.T. 0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:184 Resp: 60419
Ion Ratio Lower Upper
184 100
63 71.1 84.7 127.1#
154 103.2 338.2 507.2#





#55
Dibenzofuran
Concen: 52.125 ng
RT: 10.133 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Instrument : BNA_F
ClientSampleId : VNJ-215MS



Manual Integrations APPROVED

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/2024


10/29/2024
Supervised By :mohammad ahmed

10/30/2024



#56
4-Nitrophenol
Concen: 98.551 ng
RT: 10.033 min Scan# 1350
Delta R.T. 0.012 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:139 Resp: 254819
Ion Ratio Lower Upper
139 100
109 73.1 53.6 93.6
65 104.9 87.8 127.8



#57
 2,4-Dinitrotoluene
 Concen: 59.527 ng
 RT: 10.110 min Scan# 1
 Delta R.T. 0.000 min
 Lab File: BF140092.D
 Acq: 28 Oct 2024 17:35

Instrument : BNA_F
 ClientSampleId : VNJ-215MS

Tgt	Ion:165	Resp:	238568
	Ion Ratio	Lower	Upper
165	100		
63	49.4	40.1	60.1
89	74.2	60.4	90.6
182	2.8	2.0	3.0


Manual Integrations APPROVED

Reviewed By :Yogesh Patel


Supervised By :mohammad ahmed
 10/30/2024

10/29/2024
 Supervised By :mohammad ahmed

10/30/2024




Time--> 10.00 10.10 10.20




Time--> 10.40 10.50 10.60

Tgt	Ion:166	Resp:	716535
	Ion Ratio	Lower	Upper
166	100		
165	98.9	78.5	117.7
167	13.5	10.9	16.3




Time--> 10.40 10.50 10.60


Time--> 10.40 10.50 10.60




Time--> 40 60 80 100 120 140 160 180




Time--> 40 60 80 100 120 140 160 180




Time--> 40 60 80 100 120 140 160 180




Time--> 10.40 10.50 10.60



Time--> 10.40 10.50 10.60



Time--> 10.40 10.50 10.60



#59
2,3,4,6-Tetrachlorophenol
Concen: 54.095 ng
RT: 10.245 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS


Manual Integrations APPROVED

Reviewed By :Yogesh
Patel

Supervised By :mohammad
ahmed
10/30/2024


10/29/2024
Supervised By :mohammad
ahmed


10/30/2024



#60
Diethylphthalate
Concen: 53.595 ng
RT: 10.345 min Scan# 1403
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:149 Resp: 792061
Ion Ratio Lower Upper
149 100
177 22.1 17.9 26.9
150 12.8 10.1 15.1





#61
4-Chlorophenyl-phenylether
Concen: 52.737 ng
RT: 10.469 min Scan# 1424
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS



Manual Integrations APPROVED

Reviewed By :Yogesh
Patel

Supervised By :mohammad
ahmed
10/30/2024


10/29/2024
Supervised By :mohammad
ahmed


10/30/2024



#62
4-Nitroaniline
Concen: 52.812 ng
RT: 10.492 min Scan# 1428
Delta R.T. 0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:138 Resp: 167633
Ion Ratio Lower Upper
138 100
92 52.5 35.7 75.7
108 77.6 57.5 97.5





#63

Azobenzene

Concen: 51.543 ng

RT: 10.628 min Scan# 1451

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument:

BNA_F

ClientSampleId :



VNJ-215MS

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/202410/29/2024
Supervised By :mohammad ahmed

10/30/2024



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.410 min Scan# 1584

Delta R.T. -0.006 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35


Tgt Ion:188 Resp: 409717



Ion Ratio Lower Upper

188 100

94 10.2 7.9 11.9

80 10.9 9.0 13.4





#65
4,6-Dinitro-2-methylphenol
Concen: 36.479 ng
RT: 10.516 min Scan# 1432
Delta R.T. -0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS


Manual Integrations APPROVED

Reviewed By :Yogesh
Patel

Supervised By :mohammad
ahmed
10/30/2024


10/29/2024
Supervised By :mohammad
ahmed




10/30/2024



#66
n-Nitrosodiphenylamine
Concen: 54.692 ng
RT: 10.586 min Scan# 1444
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:169 Resp: 670666
Ion Ratio Lower Upper
169 100
168 62.2 49.8 74.6
167 34.5 28.0 42.0





#67
4-Bromophenyl-phenylether
Concen: 55.258 ng
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS




Manual Integrations APPROVED

Reviewed By :Yogesh
Patel

Supervised By :mohammad
ahmed
10/30/2024


10/29/2024
Supervised By :mohammad
ahmed


10/30/2024



#68
Hexachlorobenzene
Concen: 53.889 ng
RT: 11.022 min Scan# 1518
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:284 Resp: 255815
Ion Ratio Lower Upper
284 100
142 32.4 26.9 40.3
249 32.2 25.3 37.9





#69
Atrazine
Concen: 63.831 ng
RT: 11.110 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Instrument : BNA_F
ClientSampleId : VNJ-215MS



Manual Integrations APPROVED

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/2024


10/29/2024
Supervised By :mohammad ahmed




10/30/2024



#70
Pentachlorophenol
Concen: 77.757 ng
RT: 11.216 min Scan# 1551
Delta R.T. 0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:266 Resp: 224019
Ion Ratio Lower Upper
266 100
268 64.2 50.4 75.6
264 62.9 50.6 75.8





#71

Phenanthrene

Concen: 53.303 ng

RT: 11.439 min Scan# 1589

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument:

BNA_F

ClientSampleId :

VNJ-215MS

Tgt Ion:178 Resp: 103159:

Ion Ratio Lower Upper

178 100

176 19.5 15.8 23.6





179 15.7 12.6 18.8

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/202410/29/2024
Supervised By :mohammad ahmed

10/30/2024



#72

Anthracene

Concen: 55.155 ng

RT: 11.492 min Scan# 1598

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35


Tgt Ion:178 Resp: 1041472


Ion Ratio Lower Upper

178 100

176 19.1 15.3 22.9

179 15.4 12.4 18.6





#73
Carbazole
Concen: 50.241 ng
RT: 11.639 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Instrument : BNA_F
ClientSampleId : VNJ-215MS

Tgt Ion:167 Resp: 882304
Ion Ratio Lower Upper
167 100
166 21.5 17.0 25.6
139 13.4 10.8 16.2



Manual Integrations APPROVED

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/2024


10/29/2024
Supervised By :mohammad ahmed


10/30/2024



#74
Di-n-butylphthalate
Concen: 55.199 ng
RT: 11.969 min Scan# 1679
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:149 Resp: 1119955
Ion Ratio Lower Upper
149 100
150 9.4 7.5 11.3
104 5.1 4.1 6.1





#75
Fluoranthene
Concen: 46.655 ng
RT: 12.627 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Instrument : BNA_F
ClientSampleId : VNJ-215MS



Manual Integrations
APPROVED

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/2024


10/29/2024
Supervised By :mohammad ahmed


10/30/2024



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.051 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:240 Resp: 203129
Ion Ratio Lower Upper
240 100
120 10.6 9.4 14.2
236 25.9 20.9 31.3





#77
Benzidine
Concen: 5.301 ng
RT: 12.745 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Instrument : BNA_F
ClientSampleId : VNJ-215MS

Tgt Ion:184 Resp: 17706
Ion Ratio Lower Upper
184 100
185 16.3 12.0 18.0
183 11.7 9.4 14.2



Manual Integrations APPROVED

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/2024


10/29/2024
Supervised By :mohammad ahmed



10/30/2024



#78
Pyrene
Concen: 50.811 ng
RT: 12.851 min Scan# 1829
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:202 Resp: 903442
Ion Ratio Lower Upper
202 100
200 21.4 17.2 25.8
203 17.6 14.2 21.4





#79

Terphenyl-d14
Concen: 104.783 ng

RT: 12.998 min Scan# 1854

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument:

BNA_F

ClientSampleId:

VNJ-215MS

**Manual Integrations
APPROVED**Reviewed By :Yogesh
PatelSupervised By :mohammad
ahmed
10/30/202410/29/2024
Supervised By :mohammad
ahmed

10/30/2024

Tgt Ion:244 Resp: 1306201

Ion Ratio Lower Upper


244 100

212 6.8 5.7 8.5

122 9.9 8.6 13.0

Abundance

12.998



#80

Butylbenzylphthalate

Concen: 64.747 ng

RT: 13.469 min Scan# 1934

Delta R.T. -0.006 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Tgt Ion:149 Resp: 345144

Ion Ratio Lower Upper


149 100


91 75.1 59.4 89.2

206 20.6 16.4 24.6

Abundance

13.469





#81

Benzo(a)anthracene

Concen: 54.760 ng

RT: 14.039 min Scan# 2031

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument :

BNA_F

ClientSampleId :

VNJ-215MS

Tgt Ion:228 Resp: 724089

Ion Ratio Lower Upper

228 100

226 26.8 21.6 32.4



229 19.8 16.1 24.1

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/202410/29/2024
Supervised By :mohammad ahmed

10/30/2024



#82

3,3'-Dichlorobenzidine

Concen: 48.202 ng

RT: 14.004 min Scan# 2025

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35


Tgt Ion:252 Resp: 185837


Ion Ratio Lower Upper

252 100

254 63.6 50.8 76.2

126 10.7 8.9 13.3





#83

Chrysene

Concen: 57.841 ng

RT: 14.080 min Scan# 2038

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument :

BNA_F

ClientSampleId :

VNJ-215MS

Tgt Ion:228 Resp: 701255

Ion Ratio Lower Upper

228 100

226 29.7 24.1 36.1



229 19.9 15.8 23.6

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/202410/29/2024
Supervised By :mohammad ahmed

10/30/2024



#84

Bis(2-ethylhexyl)phthalate

Concen: 76.394 ng

RT: 14.027 min Scan# 2029

Delta R.T. -0.006 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35


Tgt Ion:149 Resp: 456891



Ion Ratio Lower Upper

149 100

167 26.8 21.4 32.0

279 4.1 3.0 4.4





#85

Di-n-octyl phthalate

Concen: 72.931 ng

RT: 14.639 min Scan# 2

Delta R.T. -0.012 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument :

BNA_F

ClientSampleId :

VNJ-215MS

Tgt Ion:149 Resp: 793362

Ion Ratio Lower Upper

149 100

167 1.6 1.2 1.8


43 11.2 9.8 14.6

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/202410/29/2024
Supervised By :mohammad ahmed

10/30/2024



#86

Perylene-d₁₂

Concen: 20.000 ng

RT: 15.527 min Scan# 2284

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35


Tgt Ion:264 Resp: 263019



Ion Ratio Lower Upper

264 100

260 23.4 19.4 29.2

265 21.7 17.4 26.0





#87

Indeno(1,2,3-cd)pyrene

Concen: 45.859 ng

RT: 17.027 min Scan# 2

Delta R.T. 0.006 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument :

BNA_F

ClientSampleId :

VNJ-215MS

Tgt Ion:276 Resp: 775966

Ion Ratio Lower Upper

276 100

138 24.1 20.4 30.6


277 25.1 20.2 30.4

Manual Integrations APPROVED

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/202410/29/2024
Supervised By :mohammad ahmed

10/30/2024



#88

Benzo(b)fluoranthene

Concen: 51.919 ng

RT: 15.098 min Scan# 2211

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35


Tgt Ion:252 Resp: 831851



Ion Ratio Lower Upper

252 100

253 21.7 17.6 26.4

125 9.3 7.9 11.9





#89

Benzo(k)fluoranthene

Concen: 56.615 ng

RT: 15.127 min Scan# 2

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35

Instrument :

BNA_F

ClientSampleId :


VNJ-215MS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/202410/29/2024
Supervised By :mohammad ahmed

10/30/2024



#90

Benzo(a)pyrene

Concen: 59.203 ng

RT: 15.468 min Scan# 2274

Delta R.T. 0.000 min

Lab File: BF140092.D

Acq: 28 Oct 2024 17:35


Tgt Ion:252 Resp: 780503



Ion Ratio Lower Upper

252 100

253 22.1 17.3 25.9

125 10.4 8.7 13.1





#91
Dibenzo(a,h)anthracene
Concen: 46.124 ng
RT: 17.045 min Scan# 2542
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Instrument : BNA_F
ClientSampleId : VNJ-215MS


Manual Integrations APPROVED

Reviewed By :Yogesh Patel

Supervised By :mohammad ahmed
10/30/2024


10/29/2024
Supervised By :mohammad ahmed

10/30/2024



#92
Benzo(g,h,i)perylene
Concen: 37.525 ng
RT: 17.474 min Scan# 2615
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:276 Resp: 529084
Ion Ratio Lower Upper
276 100
277 23.3 19.0 28.6
138 21.2 17.4 26.0





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	10/24/24	
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	10/24/24	
Client Sample ID:	VNJ-215MSD			SDG No.:	P4578	
Lab Sample ID:	P4545-02MSD			Matrix:	TCLP	
Analytical Method:	SW8270			% Solid:	0	
Sample Wt/Vol:	100	Units:	mL	Final Vol:	1000	uL
Soil Aliquot Vol:	uL			Test:	TCLP BNA	
Extraction Type :				Decanted :	N	Level :
Injection Volume :				GPC Factor :	1.0	GPC Cleanup : N PH :
Prep Method :	SW3510C					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140093.D	1	10/26/24 12:10	10/28/24 18:03	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	250		15.5	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	440		8.40	50.0	ug/L
95-48-7	2-Methylphenol	500		11.3	50.0	ug/L
65794-96-9	3+4-Methylphenols	500		11.5	100	ug/L
67-72-1	Hexachloroethane	430		10.1	50.0	ug/L
98-95-3	Nitrobenzene	470		12.7	50.0	ug/L
87-68-3	Hexachlorobutadiene	460		12.7	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	560		8.90	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	540		10.1	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	610		15.2	50.0	ug/L
118-74-1	Hexachlorobenzene	560		11.4	50.0	ug/L
87-86-5	Pentachlorophenol	850	E	18.5	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	136		15 (10) - 110 (139)	91%	SPK: 150
13127-88-3	Phenol-d6	132		15 (10) - 110 (134)	88%	SPK: 150
4165-60-0	Nitrobenzene-d5	105		30 (49) - 130 (133)	105%	SPK: 100
321-60-8	2-Fluorobiphenyl	101		30 (52) - 130 (132)	101%	SPK: 100
118-79-6	2,4,6-Tribromophenol	162		15 (44) - 110 (137)	108%	SPK: 150
1718-51-0	Terphenyl-d14	105		30 (48) - 130 (125)	105%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	112000		6.886		
1146-65-2	Naphthalene-d8	432000		8.169		
15067-26-2	Acenaphthene-d10	237000		9.927		
1517-22-2	Phenanthrene-d10	399000		11.41		
1719-03-5	Chrysene-d12	204000		14.051		
1520-96-3	Perylene-d12	256000		15.527		



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	10/24/24	
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	10/24/24	
Client Sample ID:	V NJ-215MSD			SDG No.:	P4578	
Lab Sample ID:	P4545-02MSD			Matrix:	TCLP	
Analytical Method:	SW8270			% Solid:	0	
Sample Wt/Vol:	100	Units:	mL	Final Vol:	1000	uL
Soil Aliquot Vol:	uL			Test:	TCLP BNA	
Extraction Type :				Decanted :	N	Level :
Injection Volume :				GPC Factor :	1.0	GPC Cleanup : N PH :
Prep Method :	SW3510C					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140093.D	1	10/26/24 12:10	10/28/24 18:03	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140093.D
 Acq On : 28 Oct 2024 18:03
 Operator : RC/JU
 Sample : P4545-02MSD
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 VNJ-215MSD

Quant Time: Oct 28 18:37:04 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.886	152	111820	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	432287	20.000	ng	0.00
39) Acenaphthene-d10	9.927	164	236921	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	398949	20.000	ng	0.00
76) Chrysene-d12	14.051	240	204155	20.000	ng	0.00
86) Perylene-d12	15.527	264	256308	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.522	112	974076	136.371	ng	0.02
7) Phenol-d6	6.516	99	1219104	131.779	ng	0.00
23) Nitrobenzene-d5	7.451	82	820015	105.134	ng	0.00
42) 2,4,6-Tribromophenol	10.716	330	358251	161.659	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1443612	100.702	ng	0.00
79) Terphenyl-d14	12.998	244	1318745	105.257	ng	0.00
Target Compounds						
				Qvalue		
2) 1,4-Dioxane	2.881	88	128884	38.700	ng	# 81
3) Pyridine	3.587	79	206806	25.053	ng	98
4) n-Nitrosodimethylamine	3.516	42	189912	42.820	ng	# 98
6) Aniline	6.551	93	182449	21.161	ng	99
8) 2-Chlorophenol	6.675	128	372961	51.076	ng	97
9) Benzaldehyde	6.439	77	44949	8.637	ng	95
10) Phenol	6.528	94	448413m	47.016	ng	
11) bis(2-Chloroethyl)ether	6.628	93	356154	48.313	ng	99
12) 1,3-Dichlorobenzene	6.834	146	359207	42.853	ng	99
13) 1,4-Dichlorobenzene	6.904	146	367348	43.866	ng	99
14) 1,2-Dichlorobenzene	7.057	146	355848	45.529	ng	99
15) Benzyl Alcohol	7.028	79	349195	51.458	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.157	45	577398	45.935	ng	96
17) 2-Methylphenol	7.134	107	313004	50.269	ng	99
18) Hexachloroethane	7.404	117	129514	43.369	ng	95
19) n-Nitroso-di-n-propyla...	7.298	70	273178	48.688	ng	98
20) 3+4-Methylphenols	7.286	107	394617	49.549	ng	92
22) Acetophenone	7.298	105	510185	48.185	ng	97
24) Nitrobenzene	7.469	77	405962	47.472	ng	100
25) Isophorone	7.710	82	761382	51.431	ng	99
26) 2-Nitrophenol	7.786	139	190721	59.118	ng	96
27) 2,4-Dimethylphenol	7.816	122	316975	58.924	ng	99
28) bis(2-Chloroethoxy)met...	7.916	93	451618	50.352	ng	99
29) 2,4-Dichlorophenol	8.022	162	320207	52.260	ng	100
30) 1,2,4-Trichlorobenzene	8.110	180	314131	46.333	ng	100
31) Naphthalene	8.192	128	1064132	47.730	ng	100
32) Benzoic acid	7.904	122	123617	26.347	ng	98
33) 4-Chloroaniline	8.233	127	150937	19.854	ng	99
34) Hexachlorobutadiene	8.310	225	196299	46.081	ng	99
35) Caprolactam	8.604	113	85333m	43.800	ng	
36) 4-Chloro-3-methylphenol	8.716	107	365021	53.801	ng	98
37) 2-Methylnaphthalene	8.880	142	701854	51.402	ng	99
38) 1-Methylnaphthalene	8.980	142	652303	48.693	ng	100
40) 1,2,4,5-Tetrachloroben...	9.045	216	331879	51.923	ng	99
41) Hexachlorocyclopentadiene	9.033	237	348918	156.886	ng	99
43) 2,4,6-Trichlorophenol	9.157	196	253091	55.928	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140093.D
 Acq On : 28 Oct 2024 18:03
 Operator : RC/JU
 Sample : P4545-02MSD
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 VNJ-215MSD

Quant Time: Oct 28 18:37:04 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.204	196	253581	54.313	ng	98
46) 1,1'-Biphenyl	9.345	154	857737	52.006	ng	100
47) 2-Chloronaphthalene	9.375	162	680498	51.228	ng	99
48) 2-Nitroaniline	9.463	65	244814	60.258	ng	99
49) Acenaphthylene	9.786	152	1071538	55.796	ng	99
50) Dimethylphthalate	9.645	163	828465	56.139	ng	100
51) 2,6-Dinitrotoluene	9.710	165	177848	55.660	ng	95
52) Acenaphthene	9.963	154	668819	53.836	ng	99
53) 3-Nitroaniline	9.875	138	128514	39.002	ng	97
54) 2,4-Dinitrophenol	9.980	184	66396	52.460	ng	# 1
55) Dibenzofuran	10.133	168	950618	53.332	ng	99
56) 4-Nitrophenol	10.033	139	257304	101.497	ng	99
57) 2,4-Dinitrotoluene	10.110	165	238589	60.720	ng	99
58) Fluorene	10.475	166	727096	53.557	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.245	232	204109	55.766	ng	99
60) Diethylphthalate	10.345	149	792166	54.672	ng	99
61) 4-Chlorophenyl-phenyle...	10.469	204	374542	54.630	ng	97
62) 4-Nitroaniline	10.492	138	166610	53.537	ng	98
63) Azobenzene	10.627	77	802948	52.271	ng	97
65) 4,6-Dinitro-2-methylph...	10.516	198	66402	40.805	ng	95
66) n-Nitrosodiphenylamine	10.586	169	673533	56.408	ng	99
67) 4-Bromophenyl-phenylether	10.957	248	231386	56.299	ng	98
68) Hexachlorobenzene	11.022	284	257174	55.638	ng	98
69) Atrazine	11.110	200	211444	65.372	ng	99
70) Pentachlorophenol	11.216	266	237118	84.525	ng	99
71) Phenanthrene	11.439	178	1027335	54.516	ng	100
72) Anthracene	11.492	178	1041878	56.666	ng	100
73) Carbazole	11.639	167	877488	51.316	ng	100
74) Di-n-butylphthalate	11.969	149	1110467	56.209	ng	100
75) Fluoranthene	12.621	202	920478	48.318	ng	100
77) Benzidine	12.745	184	27530	8.201	ng	98
78) Pyrene	12.851	202	902116	50.481	ng	99
80) Butylbenzylphthalate	13.468	149	344594	64.319	ng	100
81) Benzo(a)anthracene	14.039	228	746125	56.143	ng	99
82) 3,3'-Dichlorobenzidine	14.004	252	188100	48.544	ng	99
83) Chrysene	14.074	228	707350	58.050	ng	100
84) Bis(2-ethylhexyl)phtha...	14.027	149	458667	76.306	ng	100
85) Di-n-octyl phthalate	14.639	149	799157	73.095	ng	97
87) Indeno(1,2,3-cd)pyrene	17.021	276	748376	45.387	ng	99
88) Benzo(b)fluoranthene	15.092	252	856677	54.869	ng	100
89) Benzo(k)fluoranthene	15.127	252	777981	57.778	ng	99
90) Benzo(a)pyrene	15.468	252	787766	61.319	ng	100
91) Dibenzo(a,h)anthracene	17.045	278	626364	45.536	ng	98
92) Benzo(g,h,i)perylene	17.474	276	506321	36.850	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
 Data File : BF140093.D
 Acq On : 28 Oct 2024 18:03
 Operator : RC/JU
 Sample : P4545-02MSD
 Misc :
 ALS Vial : 20 Sample Multiplier: 1




Quant Time: Oct 28 18:37:04 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Instrument :
 BNA_F
 ClientSampleId :
 VNJ-215MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024








#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.886 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD


Manual Integrations APPROVED







Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#2
1,4-Dioxane
Concen: 38.700 ng
RT: 2.881 min Scan# 134
Delta R.T. 0.194 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion: 88 Resp: 128884
Ion Ratio Lower Upper
88 100
58 87.7 73.3 109.9
43 0.0 27.1 40.7#





#3

Pyridine

Concen: 25.053 ng

RT: 3.587 min Scan# 2

Delta R.T. 0.141 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03

Instrument :

BNA_F

ClientSampleId :

VNJ-215MSD

Tgt Ion: 79 Resp: 206806

Ion Ratio Lower Upper

79 100


52 83.6 68.5 102.7

51 39.7 32.3 48.5

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#4

n-Nitrosodimethylamine

Concen: 42.820 ng

RT: 3.516 min Scan# 242

Delta R.T. 0.118 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03


Tgt Ion: 42 Resp: 189912




Ion Ratio Lower Upper

42 100

74 121.6 96.2 144.4

44 10.3 6.2 9.2#









#5
2-Fluorophenol
Concen: 136.371 ng
RT: 5.522 min Scan# 51
Delta R.T. 0.024 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Instrument : BNA_F
ClientSampleId : VNJ-215MSD


Manual Integrations APPROVED



Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#6
Aniline
Concen: 21.161 ng
RT: 6.551 min Scan# 758
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion: 93 Resp: 182449
Ion Ratio Lower Upper
93 100
66 43.1 34.0 51.0
65 22.1 17.4 26.0






#7
 Phenol-d6
 Concen: 131.779 ng
 RT: 6.516 min Scan# 7
 Delta R.T. 0.006 min
 Lab File: BF140093.D
 Acq: 28 Oct 2024 18:03

Instrument :
 BNA_F
 ClientSampleId :
 VNJ-215MSD


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024



#8
 2-Chlorophenol
 Concen: 51.076 ng
 RT: 6.675 min Scan# 779
 Delta R.T. 0.000 min
 Lab File: BF140093.D
 Acq: 28 Oct 2024 18:03

Tgt Ion:128 Resp: 372961
 Ion Ratio Lower Upper
 128 100
 130 32.7 12.3 52.3
 64 51.8 34.6 74.6







#9
Benzaldehyde
Concen: 8.637 ng
RT: 6.439 min Scan# 7
Instrument: BNA_F
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03


Manual Integrations
APPROVED







Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#10
Phenol
Concen: 47.016 ng
RT: 6.528 min Scan# 754
Delta R.T. 0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion: 94 Resp: 448413
Ion Ratio Lower Upper
94 100
65 31.6 11.1 51.1
66 45.3 22.7 62.7






#11
bis(2-Chloroethyl)ether
Concen: 48.313 ng
RT: 6.628 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD


Manual Integrations APPROVED







Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#12
1,3-Dichlorobenzene
Concen: 42.853 ng
RT: 6.834 min Scan# 806
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:146 Resp: 359207
Ion Ratio Lower Upper
146 100
148 63.9 51.0 76.4
75 31.6 26.4 39.6





#13

1,4-Dichlorobenzene

Concen: 43.866 ng

RT: 6.904 min Scan# 8

Delta R.T. -0.006 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03

Instrument :

BNA_F

ClientSampleId :

VNJ-215MSD

Tgt Ion:146 Resp: 367348

Ion Ratio Lower Upper

146 100

148 63.6 51.8 77.8


111 43.3 34.6 51.8

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#14

1,2-Dichlorobenzene

Concen: 45.529 ng

RT: 7.057 min Scan# 844

Delta R.T. -0.006 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03


Tgt Ion:146 Resp: 355848




Ion Ratio Lower Upper

146 100

148 63.9 51.4 77.2

111 45.0 36.7 55.1





#15

Benzyl Alcohol

Concen: 51.458 ng

RT: 7.028 min Scan# 8

Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03

Instrument :

BNA_F




ClientSampleId :

VNJ-215MSD

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#16

2,2'-oxybis(1-Chloropropane)

Concen: 45.935 ng

RT: 7.157 min Scan# 861

Delta R.T. -0.012 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03



Tgt Ion: 45 Resp: 577398




Ion Ratio Lower Upper

45 100

77 13.2 0.0 31.5

79 10.0 0.0 29.1





#17

2-Methylphenol

Concen: 50.269 ng

RT: 7.134 min Scan# 8

Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03

Instrument :

BNA_F




ClientSampleId :

VNJ-215MSD

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#18

Hexachloroethane

Concen: 43.369 ng

RT: 7.404 min Scan# 903

Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03


Tgt Ion:117 Resp: 129514







Ion Ratio Lower Upper

117 100

119 94.7 77.0 115.4

201 104.2 76.2 114.4






#19

n-Nitroso-di-n-propylamine
Concen: 48.688 ng
RT: 7.298 min Scan# 885
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024









#20

3+4-Methylphenols
Concen: 49.549 ng
RT: 7.286 min Scan# 883
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt	Ion	107	394617
	Ion Ratio	Lower	Upper
107	100		
108	89.6	69.8	109.8
77	88.5	50.5	90.5
79	29.4	10.2	50.2





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.169 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03

Instrument :

BNA_F


ClientSampleId :

VNJ-215MSD

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#22

Acetophenone

Concen: 48.185 ng


RT: 7.298 min Scan# 885




Delta R.T. 0.000 min

Lab File: BF140093.D



Acq: 28 Oct 2024 18:03

Tgt	Ion:105	Resp:	510185
Ion	Ratio	Lower	Upper
105	100		
71	8.7	5.9	8.9
51	34.6	29.4	44.2
120	21.4	17.5	26.3








#23


Nitrobenzene-d5
Concen: 105.134 ngRT: 7.451 min Scan# 911
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD**Manual Integrations
APPROVED**Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

#24



Nitrobenzene
Concen: 47.472 ng
RT: 7.469 min Scan# 914
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03Tgt Ion: 77 Resp: 405962
Ion Ratio Lower Upper
77 100
123 44.7 35.7 53.5
65 14.5 11.5 17.3




Abundance Scan 955 (7.710 min): BF140093.D\data.ms




Abundance Scan 955 (7.710 min): BF140093.D\data.ms (-90)



Abundance Scan 968 (7.786 min): BF140093.D\data.ms



Abundance Scan 968 (7.786 min): BF140093.D\data.ms (-91)



#25

Isophorone

Concen: 51.431 ng

RT: 7.710 min Scan# 9

Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03

Instrument :

BNA_F

ClientSampleId :

VNJ-215MSD

Tgt Ion: 82 Resp: 761382

Ion Ratio Lower Upper

82 100


95 7.3 5.5 8.3

138 16.8 13.2 19.8

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#26

2-Nitrophenol

Concen: 59.118 ng

RT: 7.786 min Scan# 968

Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03


Tgt Ion:139 Resp: 190721




Ion Ratio Lower Upper

139 100

109 30.4 25.1 37.7

65 55.5 47.5 71.3





#27

2,4-Dimethylphenol

Concen: 58.924 ng

RT: 7.816 min Scan# 9

Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03

Instrument :

BNA_F




ClientSampleId :

VNJ-215MSD

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#28

bis(2-Chloroethoxy)methane

Concen: 50.352 ng

RT: 7.916 min Scan# 990

Delta R.T. -0.006 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03


Tgt Ion: 93 Resp: 451618







Ion Ratio Lower Upper

93 100

95 32.6 26.0 39.0

123 11.5 9.8 14.8





#29

2,4-Dichlorophenol

Concen: 52.260 ng

RT: 8.022 min Scan# 1008

Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03


Instrument :

BNA_F

ClientSampleId :


VNJ-215MSD







**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024


#30
 1,2,4-Trichlorobenzene
 Concen: 46.333 ng
 RT: 8.110 min Scan# 1023
 Delta R.T. -0.006 min
 Lab File: BF140093.D
 Acq: 28 Oct 2024 18:03

Tgt Ion:180 Resp: 314131
 Ion Ratio Lower Upper
 180 100
 182 100.4 80.5 120.7
 145 31.8 25.4 38.0





#31

Naphthalene

Concen: 47.730 ng

RT: 8.192 min Scan# 106413

Delta R.T. -0.006 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03

Instrument :

BNA_F


ClientSampleId :

VNJ-215MSD

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#32

Benzoic acid

Concen: 26.347 ng

RT: 7.904 min Scan# 988

Delta R.T. -0.018 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03


Tgt Ion:122 Resp: 123617







Ion Ratio Lower Upper

122 100

105 125.8 107.6 147.6

77 97.8 80.2 120.2





#33

4-Chloroaniline

Concen: 19.854 ng

RT: 8.233 min Scan# 1045

Delta R.T. -0.006 min

Lab File: BF140093.D


Acq: 28 Oct 2024 18:03

Instrument:

BNA_F

ClientSampleId :

VNJ-215MSD

**Manual Integrations
APPROVED**
Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

#34

Hexachlorobutadiene

Concen: 46.081 ng


RT: 8.310 min Scan# 1057







Delta R.T. -0.006 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03

Tgt	Ion:225	Resp:	196299
Ion	Ratio	Lower	Upper
225	100		
223	63.5	50.6	76.0
227	64.1	51.8	77.8





#35

Caprolactam

Concen: 43.800 ng m

RT: 8.604 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03

Instrument :

BNA_F


ClientSampleId :

VNJ-215MSD

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#36

4-Chloro-3-methylphenol

Concen: 53.801 ng

RT: 8.716 min Scan# 1126

Delta R.T. 0.006 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03


Tgt Ion:107 Resp: 365021




Ion Ratio Lower Upper

107 100

144 24.6 19.1 28.7

142 72.9 57.2 85.8





#37

2-Methylnaphthalene

Concen: 51.402 ng

RT: 8.880 min Scan# 1155

Delta R.T. -0.006 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03




Instrument :

BNA_F

ClientSampleId :

VNJ-215MSD

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024


#38

1-Methylnaphthalene

Concen: 48.693 ng

RT: 8.980 min Scan# 1171

Delta R.T. -0.006 min

Lab File: BF140093.D


Acq: 28 Oct 2024 18:03




Tgt Ion:142 Resp: 652303

Ion Ratio Lower Upper

142 100

141 91.8 73.5 110.3





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.927 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03

Instrument :

BNA_F




ClientSampleId :

VNJ-215MSD

**Manual Integrations
APPROVED**



Reviewed By :Yogesh Patel 10/29/2024


Supervised By :mohammad ahmed 10/30/2024



#40
1,2,4,5-Tetrachlorobenzene
Concen: 51.923 ng
RT: 9.045 min Scan# 1182
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:216 Resp: 331879
Ion Ratio Lower Upper
216 100
214 79.1 63.4 95.0
179 21.1 17.2 25.8
108 21.1 15.6 23.4







#41
Hexachlorocyclopentadiene
Concen: 156.886 ng
RT: 9.033 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD


Manual Integrations APPROVED







Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#42
2,4,6-Tribromophenol
Concen: 161.659 ng
RT: 10.716 min Scan# 1466
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:330 Resp: 358251
Ion Ratio Lower Upper
330 100
332 95.9 78.1 117.1
141 29.3 26.6 39.8












#43


2,4,6-Trichlorophenol
Concen: 55.928 ngRT: 9.157 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD**Manual Integrations
APPROVED**Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

#44


2,4,5-Trichlorophenol
Concen: 54.313 ng
RT: 9.204 min Scan# 1209
Delta R.T. 0.012 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03Tgt Ion:196 Resp: 253581
Ion Ratio Lower Upper
196 100
198 96.3 75.8 113.6
97 50.3 41.8 62.6
132 26.9 22.9 34.3






#45

2-Fluorobiphenyl
Concen: 100.702 ngRT: 9.245 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD**Manual Integrations
APPROVED**Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

#46

1,1'-Biphenyl
Concen: 52.006 ng
RT: 9.345 min Scan# 1233
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03Tgt Ion:154 Resp: 857737
Ion Ratio Lower Upper
154 100
153 40.3 20.3 60.3
76 13.5 0.0 33.9



#47

2-Chloronaphthalene

Concen: 51.228 ng

RT: 9.375 min Scan# 1238

Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03




Instrument :

BNA_F

ClientSampleId :

VNJ-215MSD

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024


#48

2-Nitroaniline

Concen: 60.258 ng

RT: 9.463 min Scan# 1253

Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03


Tgt Ion: 65 Resp: 244814







Ion Ratio Lower Upper

65 100

92 63.9 50.9 76.3

138 93.8 73.4 110.0







#49







Acenaphthylene
Concen: 55.796 ngRT: 9.786 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Tgt Ion:152 Resp: 1071538

Ion Ratio Lower Upper
152 100
151 20.3 15.9 23.9
153 13.5 10.7 16.1**Manual Integrations
APPROVED**Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

#50

Dimethylphthalate
Concen: 56.139 ng
RT: 9.645 min Scan# 1284
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03Tgt Ion:163 Resp: 828465
Ion Ratio Lower Upper
163 100
194 4.0 3.4 5.0
164 10.2 8.2 12.2



#51

2,6-Dinitrotoluene

Concen: 55.660 ng

RT: 9.710 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03

Instrument :

BNA_F


ClientSampleId :

VNJ-215MSD

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#52

Acenaphthene

Concen: 53.836 ng

RT: 9.963 min Scan# 1338

Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03


Tgt Ion:154 Resp: 668819







Ion Ratio Lower Upper

154 100

153 110.8 89.8 134.6

152 53.8 42.4 63.6





#53

3-Nitroaniline
Concen: 39.002 ng

RT: 9.875 min Scan# 1323


Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD


Tgt	Ion:138	Resp:	128514
Ion Ratio	Lower	Upper	
138	100		
108	11.3	10.2	15.2
92	119.0	98.3	147.5


**Manual Integrations
APPROVED**Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

#54

2,4-Dinitrophenol
Concen: 52.460 ng
RT: 9.980 min Scan# 1341
Delta R.T. 0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03


Tgt	Ion:184	Resp:	66396
Ion Ratio	Lower	Upper	
184	100		
63	70.8	84.7	127.1#
154	98.9	338.2	507.2#








#55
Dibenzofuran
Concen: 53.332 ng
RT: 10.133 min Scan# 1367
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Instrument : BNA_F
ClientSampleId : VNJ-215MSD




Tgt Ion:168 Resp: 950618
Ion Ratio Lower Upper
168 100
139 39.2 31.8 47.6
169 14.4 11.6 17.4



Manual Integrations APPROVED
Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#56
4-Nitrophenol
Concen: 101.497 ng
RT: 10.033 min Scan# 1350
Delta R.T. 0.012 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03



Tgt Ion:139 Resp: 257304
Ion Ratio Lower Upper
139 100
109 74.2 53.6 93.6
65 105.7 87.8 127.8






#57
2,4-Dinitrotoluene
Concen: 60.720 ng
RT: 10.110 min Scan# 1
Instrument: BNA_F
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03
ClientSampleId : VNJ-215MSD



Manual Integrations APPROVED




Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#58
Fluorene
Concen: 53.557 ng
RT: 10.475 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:166 Resp: 727096
Ion Ratio Lower Upper
166 100
165 97.3 78.5 117.7
167 13.4 10.9 16.3








#59
2,3,4,6-Tetrachlorophenol
Concen: 55.766 ng
RT: 10.245 min Scan# 1386
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Instrument : BNA_F
ClientSampleId : VNJ-215MSD


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#60
Diethylphthalate
Concen: 54.672 ng
RT: 10.345 min Scan# 1403
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:149 Resp: 792166
Ion Ratio Lower Upper
149 100
177 21.8 17.9 26.9
150 12.6 10.1 15.1







#61
4-Chlorophenyl-phenylether
Concen: 54.630 ng
RT: 10.469 min Scan# 1424
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD


Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#62
4-Nitroaniline
Concen: 53.537 ng
RT: 10.492 min Scan# 1428
Delta R.T. 0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:138 Resp: 166610
Ion Ratio Lower Upper
138 100
92 51.9 35.7 75.7
108 77.9 57.5 97.5





#63

Azobenzene

Concen: 52.271 ng

RT: 10.627 min Scan# 1451

Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03

Instrument :

BNA_F


ClientSampleId :

VNJ-215MSD

**Manual Integrations
APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.410 min Scan# 1584

Delta R.T. -0.006 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03


Tgt Ion:188 Resp: 398949




Ion Ratio Lower Upper

188 100

94 10.1 7.9 11.9

80 10.6 9.0 13.4









#65
4,6-Dinitro-2-methylphenol
Concen: 40.805 ng
RT: 10.516 min Scan# 1432
Delta R.T. -0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#66
n-Nitrosodiphenylamine
Concen: 56.408 ng
RT: 10.586 min Scan# 1444
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:169 Resp: 673533
Ion Ratio Lower Upper
169 100
168 61.7 49.8 74.6
167 34.4 28.0 42.0









#67
4-Bromophenyl-phenylether
Concen: 56.299 ng
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#68
Hexachlorobenzene
Concen: 55.638 ng
RT: 11.022 min Scan# 1518
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:284 Resp: 257174
Ion Ratio Lower Upper
284 100
142 32.0 26.9 40.3
249 31.6 25.3 37.9







#69
Atrazine
Concen: 65.372 ng
RT: 11.110 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Instrument : BNA_F
ClientSampleId : VNJ-215MSD


Manual Integrations
APPROVED




Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#70
Pentachlorophenol
Concen: 84.525 ng
RT: 11.216 min Scan# 1551
Delta R.T. 0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:266 Resp: 237118
Ion Ratio Lower Upper
266 100
268 62.5 50.4 75.6
264 61.7 50.6 75.8





#71

Phenanthrene

Concen: 54.516 ng

RT: 11.439 min Scan# 1589

Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03





Instrument:

BNA_F

ClientSampleId :

VNJ-215MSD

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024


#72

Anthracene

Concen: 56.666 ng

RT: 11.492 min Scan# 1598

Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03


Tgt Ion:178 Resp: 1041878


Ion Ratio Lower Upper

178 100

176 19.1 15.3 22.9

179 15.6 12.4 18.6







#73
Carbazole
Concen: 51.316 ng
RT: 11.639 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Instrument : BNA_F
ClientSampleId : VNJ-215MSD

Tgt Ion:167 Resp: 877488
Ion Ratio Lower Upper
167 100
166 21.4 17.0 25.6
139 13.5 10.8 16.2


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#74
Di-n-butylphthalate
Concen: 56.209 ng
RT: 11.969 min Scan# 1679
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:149 Resp: 1110467
Ion Ratio Lower Upper
149 100
150 9.4 7.5 11.3
104 5.1 4.1 6.1







#75
Fluoranthene
Concen: 48.318 ng
RT: 12.621 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Instrument : BNA_F
ClientSampleId : VNJ-215MSD

Tgt Ion:	Ion Ratio	Resp:	920478
202	100		
101	12.0	Lower	31.9
203	17.7	Upper	37.5


Manual Integrations APPROVED







Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.051 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:	Ion Ratio	Resp:	204155
240	100		
120	10.4	Lower	14.2
236	26.1	Upper	31.3





#77

Benzidine

Concen: 8.201 ng

RT: 12.745 min Scan# 1

Delta R.T. -0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03

Instrument :

BNA_F

ClientSampleId :

VNJ-215MSD

Tgt Ion:184 Resp: 27530

Ion Ratio Lower Upper

184 100


185 16.3 12.0 18.0

183 11.6 9.4 14.2

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#78

Pyrene

Concen: 50.481 ng

RT: 12.851 min Scan# 1829

Delta R.T. -0.006 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03


Tgt Ion:202 Resp: 902116


Ion Ratio Lower Upper

202 100

200 21.6 17.2 25.8

203 17.4 14.2 21.4







#79
Terphenyl-d14
Concen: 105.257 ng
RT: 12.998 min Scan# 131874
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Instrument : BNA_F
ClientSampleId : VNJ-215MSD


Manual Integrations APPROVED




Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#80
Butylbenzylphthalate
Concen: 64.319 ng
RT: 13.468 min Scan# 1934
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:149 Resp: 344594
Ion Ratio Lower Upper
149 100
91 74.3 59.4 89.2
206 20.6 16.4 24.6





#81

Benzo(a)anthracene

Concen: 56.143 ng

RT: 14.039 min Scan# 2031

Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03





Instrument :

BNA_F

ClientSampleId :

VNJ-215MSD

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024


#82

3,3'-Dichlorobenzidine


Concen: 48.544 ng




RT: 14.004 min Scan# 2025

Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03

 Tgt Ion:252 Resp: 188100
 Ion Ratio Lower Upper
 252 100
 254 64.2 50.8 76.2
 126 10.6 8.9 13.3




#83

Chrysene

Concen: 58.050 ng

RT: 14.074 min Scan# 2037

Delta R.T. -0.006 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03




Instrument :

BNA_F

ClientSampleId :

VNJ-215MSD

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024


#84

Bis(2-ethylhexyl)phthalate

Concen: 76.306 ng

RT: 14.027 min Scan# 2029

Delta R.T. -0.006 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03


Tgt Ion:149 Resp: 458667



Ion Ratio Lower Upper

149 100

167 26.7 21.4 32.0

279 4.1 3.0 4.4





#85

Di-n-octyl phthalate

Concen: 73.095 ng

RT: 14.639 min Scan# 2

Delta R.T. -0.012 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03

Instrument :

BNA_F

ClientSampleId :

VNJ-215MSD

Tgt Ion:149 Resp: 799157

Ion Ratio Lower Upper

149 100


167 1.6 1.2 1.8

43 10.9 9.8 14.6

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#86

Perylene-d₁₂

Concen: 20.000 ng

RT: 15.527 min Scan# 2284

Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03


Tgt Ion:264 Resp: 256308







Ion Ratio Lower Upper

264 100

260 24.0 19.4 29.2

265 21.0 17.4 26.0





#87

Indeno(1,2,3-cd)pyrene

Concen: 45.387 ng

RT: 17.021 min Scan# 2

Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03

Instrument :

BNA_F

ClientSampleId :

VNJ-215MSD

Tgt Ion:276 Resp: 748376

Ion Ratio Lower Upper

276 100


138 24.2 20.4 30.6

277 25.2 20.2 30.4

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/29/2024

Supervised By :mohammad ahmed 10/30/2024



#88

Benzo(b)fluoranthene

Concen: 54.869 ng

RT: 15.092 min Scan# 2210

Delta R.T. -0.006 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03


Tgt Ion:252 Resp: 856677







Ion Ratio Lower Upper

252 100

253 22.1 17.6 26.4

125 10.2 7.9 11.9





#89

Benzo(k)fluoranthene

Concen: 57.778 ng

RT: 15.127 min Scan# 2216

Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03


Instrument :

BNA_F

ClientSampleId :

VNJ-215MSD

**Manual Integrations
APPROVED**

 Reviewed By :Yogesh Patel 10/29/2024
 Supervised By :mohammad ahmed 10/30/2024


#90

Benzo(a)pyrene

Concen: 61.319 ng

RT: 15.468 min Scan# 2274

Delta R.T. 0.000 min

Lab File: BF140093.D

Acq: 28 Oct 2024 18:03


Tgt Ion:252 Resp: 787766


Ion Ratio Lower Upper

252 100

253 21.7 17.3 25.9

125 10.6 8.7 13.1







#91
Dibenzo(a,h)anthracene
Concen: 45.536 ng
RT: 17.045 min Scan# 2542
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Instrument : BNA_F
ClientSampleId : VNJ-215MSD


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#92
Benzo(g,h,i)perylene
Concen: 36.850 ng
RT: 17.474 min Scan# 2615
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:276 Resp: 506321
Ion Ratio Lower Upper
276 100
277 23.8 19.0 28.6
138 21.2 17.4 26.0





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BF101824	Instrument	BNA_f
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDICCC040	BF139848.D	Phenol	yogesh	10/21/2024 6:33:45 AM	mohammad	10/21/2024 6:38:35 AM	Peak Integrated by Software
SSTDICC050	BF139849.D	Phenol	yogesh	10/21/2024 6:33:46 AM	mohammad	10/21/2024 6:38:35 AM	Peak Integrated by Software
SSTDICC060	BF139850.D	Phenol	yogesh	10/21/2024 6:33:47 AM	mohammad	10/21/2024 6:38:35 AM	Peak Integrated by Software
SSTDICC080	BF139851.D	Aniline	yogesh	10/21/2024 6:33:49 AM	mohammad	10/21/2024 6:38:35 AM	Peak Integrated by Software
SSTDICV040	BF139852.D	Phenol	yogesh	10/21/2024 6:33:50 AM	mohammad	10/21/2024 6:38:35 AM	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BF102824	Instrument	BNA_f
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDCCC040	BF140075.D	Phenol	yogesh	10/29/2024 1:35:05 AM	mohammad	10/29/2024 1:38:39 AM	Peak Integrated by Software
PB164444BS	BF140077.D	Caprolactam	yogesh	10/29/2024 1:35:06 AM	mohammad	10/29/2024 1:38:39 AM	Peak Integrated by Software
PB164444BS	BF140077.D	Phenol	yogesh	10/29/2024 1:35:06 AM	mohammad	10/29/2024 1:38:39 AM	Peak Integrated by Software
P4545-02MS	BF140092.D	Caprolactam	yogesh	10/29/2024 4:52:56 AM	mohammad	10/29/2024 1:38:39 AM	Peak Integrated by Software
P4545-02MS	BF140092.D	Phenol	yogesh	10/29/2024 4:52:56 AM	mohammad	10/29/2024 1:38:39 AM	Peak Integrated by Software
P4545-02MSD	BF140093.D	Caprolactam	yogesh	10/29/2024 1:35:09 AM	mohammad	10/29/2024 1:38:39 AM	Peak Integrated by Software
P4545-02MSD	BF140093.D	Phenol	yogesh	10/29/2024 1:35:09 AM	mohammad	10/29/2024 1:38:39 AM	Peak Integrated by Software

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF101824

Review By	yogesh	Review On	10/21/2024 6:34:01 AM
Supervise By	mohammad	Supervise On	10/21/2024 6:38:35 AM
SubDirectory	BF101824	HP Acquire Method	BNA_F
HP Processing Method	bf101824		
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	SP6573 SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6624 S12322,10ul/1000ul sample SP6559		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF139843.D	18 Oct 2024 09:22	RC/JU	Ok
2	SSTDICC2.5	BF139844.D	18 Oct 2024 10:27	RC/JU	Ok
3	SSTDICC005	BF139845.D	18 Oct 2024 10:55	RC/JU	Ok
4	SSTDICC010	BF139846.D	18 Oct 2024 11:23	RC/JU	Ok
5	SSTDICC020	BF139847.D	18 Oct 2024 11:52	RC/JU	Ok
6	SSTDICCC040	BF139848.D	18 Oct 2024 12:20	RC/JU	Ok,M
7	SSTDICC050	BF139849.D	18 Oct 2024 12:49	RC/JU	Ok,M
8	SSTDICC060	BF139850.D	18 Oct 2024 13:17	RC/JU	Ok,M
9	SSTDICC080	BF139851.D	18 Oct 2024 13:46	RC/JU	Ok,M
10	SSTDICV040	BF139852.D	18 Oct 2024 14:19	RC/JU	Ok,M
11	PB164211BL	BF139853.D	18 Oct 2024 14:48	RC/JU	Ok
12	P4405-01	BF139854.D	18 Oct 2024 15:21	RC/JU	Ok
13	P4431-01	BF139855.D	18 Oct 2024 15:50	RC/JU	Ok,M
14	P4421-01	BF139856.D	18 Oct 2024 16:18	RC/JU	Ok,M
15	P4422-01	BF139857.D	18 Oct 2024 16:46	RC/JU	Ok
16	P4425-01	BF139858.D	18 Oct 2024 17:15	RC/JU	Ok
17	P4425-03	BF139859.D	18 Oct 2024 17:44	RC/JU	Ok
18	P4425-05	BF139860.D	18 Oct 2024 18:12	RC/JU	Ok
19	P4425-07	BF139861.D	18 Oct 2024 18:41	RC/JU	ReRun
20	P4425-09	BF139862.D	18 Oct 2024 19:09	RC/JU	ReRun
21	P4426-03	BF139863.D	18 Oct 2024 19:37	RC/JU	ReRun

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF101824

Review By	yogesh	Review On	10/21/2024 6:34:01 AM
Supervise By	mohammad	Supervise On	10/21/2024 6:38:35 AM
SubDirectory	BF101824	HP Acquire Method	BNA_F
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	SP6573 SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6624 S12322,10ul/1000ul sample SP6559		

22	P4426-07	BF139864.D	18 Oct 2024 20:05	RC/JU	ReRun
23	P4426-17	BF139865.D	18 Oct 2024 20:34	RC/JU	ReRun
24	P4426-11	BF139866.D	18 Oct 2024 21:02	RC/JU	ReRun

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF102824

Review By	yogesh	Review On	10/29/2024 1:35:27 AM
Supervise By	mohammad	Supervise On	10/29/2024 1:38:39 AM
SubDirectory	BF102824	HP Acquire Method	BNA_F
HP Processing Method	bf101824		
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	SP6573 SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6624 S12323,10ul/1000ul sample SP6559		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF140074.D	28 Oct 2024 09:04	RC/JU	Ok
2	SSTDCCC040	BF140075.D	28 Oct 2024 09:32	RC/JU	Ok,M
3	PB164444BL	BF140076.D	28 Oct 2024 10:01	RC/JU	Ok
4	PB164444BS	BF140077.D	28 Oct 2024 10:29	RC/JU	Ok,M
5	PB164393TB	BF140078.D	28 Oct 2024 10:58	RC/JU	Ok
6	PB164444BL	BF140079.D	28 Oct 2024 11:31	RC/JU	Not Ok
7	P4547-04	BF140080.D	28 Oct 2024 11:59	RC/JU	Ok
8	P4547-08	BF140081.D	28 Oct 2024 12:27	RC/JU	Ok
9	P4561-04	BF140082.D	28 Oct 2024 12:55	RC/JU	Ok
10	P4561-08	BF140083.D	28 Oct 2024 13:23	RC/JU	Ok
11	P4578-06	BF140084.D	28 Oct 2024 13:51	RC/JU	Ok
12	P4578-08	BF140085.D	28 Oct 2024 14:19	RC/JU	Ok
13	P4567-04	BF140086.D	28 Oct 2024 14:47	RC/JU	Ok
14	P4567-08	BF140087.D	28 Oct 2024 15:15	RC/JU	Ok
15	P4567-12	BF140088.D	28 Oct 2024 15:43	RC/JU	Ok
16	P4574-03	BF140089.D	28 Oct 2024 16:11	RC/JU	Ok
17	P4574-06	BF140090.D	28 Oct 2024 16:39	RC/JU	Ok
18	P4545-02	BF140091.D	28 Oct 2024 17:07	RC/JU	Ok
19	P4545-02MS	BF140092.D	28 Oct 2024 17:35	RC/JU	Ok,M
20	P4545-02MSD	BF140093.D	28 Oct 2024 18:03	RC/JU	Ok,M
21	P4567-01	BF140094.D	28 Oct 2024 18:31	RC/JU	Ok



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF102824

Review By	yogesh	Review On	10/29/2024 1:35:27 AM
Supervise By	mohammad	Supervise On	10/29/2024 1:38:39 AM
SubDirectory	BF102824	HP Acquire Method	BNA_F
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	SP6573 SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6624 S12323,10ul/1000ul sample SP6559		

22	P4574-04	BF140095.D	28 Oct 2024 18:59	RC/JU	Ok
23	P4577-01	BF140096.D	28 Oct 2024 19:27	RC/JU	Ok,NS
24	P4566-01	BF140097.D	28 Oct 2024 19:55	RC/JU	Ok,NS
25	P4575-01	BF140098.D	28 Oct 2024 20:23	RC/JU	Ok,NS

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF101824

Review By	yogesh	Review On	10/21/2024 6:34:01 AM		
Supervise By	mohammad	Supervise On	10/21/2024 6:38:35 AM		
SubDirectory	BF101824	HP Acquire Method	BNA_F	HP Processing Method	bf101824
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621				
CCC	SP6624				
Internal Standard/PEM	S12322,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF139843.D	18 Oct 2024 09:22		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF139844.D	18 Oct 2024 10:27		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF139845.D	18 Oct 2024 10:55	Compound #9,32,54,65,85 removed from 5 ppm	RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF139846.D	18 Oct 2024 11:23		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF139847.D	18 Oct 2024 11:52	Compound #54 Kept on LR	RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF139848.D	18 Oct 2024 12:20	The Calibration is Good For 8270 DOD Except com#77 and good for 625.1 Method	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BF139849.D	18 Oct 2024 12:49	Com#77(Benzidine) Failed in the calibration for both DOD and NON-DOD	RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BF139850.D	18 Oct 2024 13:17		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BF139851.D	18 Oct 2024 13:46		RC/JU	Ok,M
10	SSTDICV040	SSTDICV040	BF139852.D	18 Oct 2024 14:19		RC/JU	Ok,M
11	PB164211BL	PB164211BL	BF139853.D	18 Oct 2024 14:48		RC/JU	Ok
12	P4405-01	MH-121	BF139854.D	18 Oct 2024 15:21		RC/JU	Ok
13	P4431-01	72-11934	BF139855.D	18 Oct 2024 15:50		RC/JU	Ok,M
14	P4421-01	EO-02-101624	BF139856.D	18 Oct 2024 16:18		RC/JU	Ok,M
15	P4422-01	EO-01-101624	BF139857.D	18 Oct 2024 16:46		RC/JU	Ok
16	P4425-01	TP-1	BF139858.D	18 Oct 2024 17:15		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF101824

Review By	yogesh	Review On	10/21/2024 6:34:01 AM		
Supervise By	mohammad	Supervise On	10/21/2024 6:38:35 AM		
SubDirectory	BF101824	HP Acquire Method	BNA_F	HP Processing Method	bf101824
STD. NAME	STD REF.#				
Tune/Reschk Initial Calibration Stds	SP6573 SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621				
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6624 S12322,10ul/1000ul sample SP6559				

17	P4425-03	TP-2	BF139859.D	18 Oct 2024 17:44		RC/JU	Ok
18	P4425-05	TP-3	BF139860.D	18 Oct 2024 18:12		RC/JU	Ok
19	P4425-07	TP-4	BF139861.D	18 Oct 2024 18:41	Internal Standrad Fail	RC/JU	ReRun
20	P4425-09	TP-5	BF139862.D	18 Oct 2024 19:09	Internal Standrad Fail	RC/JU	ReRun
21	P4426-03	PAD-2	BF139863.D	18 Oct 2024 19:37	Internal Standrad Fail	RC/JU	ReRun
22	P4426-07	PAD-4	BF139864.D	18 Oct 2024 20:05	Internal Standrad Fail	RC/JU	ReRun
23	P4426-17	PAD-9	BF139865.D	18 Oct 2024 20:34	Internal Standard Fail	RC/JU	ReRun
24	P4426-11	PAD-6	BF139866.D	18 Oct 2024 21:02	Internal Standard Fail	RC/JU	ReRun

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF102824

Review By	yogesh	Review On	10/29/2024 1:35:27 AM		
Supervise By	mohammad	Supervise On	10/29/2024 1:38:39 AM		
SubDirectory	BF102824	HP Acquire Method	BNA_F	HP Processing Method	bf101824
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621				
CCC	SP6624				
Internal Standard/PEM	S12323,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140074.D	28 Oct 2024 09:04		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140075.D	28 Oct 2024 09:32		RC/JU	Ok,M
3	PB164444BL	PB164444BL	BF140076.D	28 Oct 2024 10:01		RC/JU	Ok
4	PB164444BS	PB164444BS	BF140077.D	28 Oct 2024 10:29		RC/JU	Ok,M
5	PB164393TB	PB164393TB	BF140078.D	28 Oct 2024 10:58		RC/JU	Ok
6	PB164444BL	PB164444BL	BF140079.D	28 Oct 2024 11:31	Already analysed with ok status	RC/JU	Not Ok
7	P4547-04	BP-F-21	BF140080.D	28 Oct 2024 11:59		RC/JU	Ok
8	P4547-08	BP-F-20	BF140081.D	28 Oct 2024 12:27		RC/JU	Ok
9	P4561-04	BP-F-19	BF140082.D	28 Oct 2024 12:55		RC/JU	Ok
10	P4561-08	BP-F-18	BF140083.D	28 Oct 2024 13:23		RC/JU	Ok
11	P4578-06	WB-305-BOT	BF140084.D	28 Oct 2024 13:51		RC/JU	Ok
12	P4578-08	WB-305-BOT-1	BF140085.D	28 Oct 2024 14:19		RC/JU	Ok
13	P4567-04	WC-1	BF140086.D	28 Oct 2024 14:47		RC/JU	Ok
14	P4567-08	WC-2	BF140087.D	28 Oct 2024 15:15		RC/JU	Ok
15	P4567-12	WC-3	BF140088.D	28 Oct 2024 15:43		RC/JU	Ok
16	P4574-03	GRAVEL-1	BF140089.D	28 Oct 2024 16:11		RC/JU	Ok
17	P4574-06	GRAVEL-2	BF140090.D	28 Oct 2024 16:39		RC/JU	Ok
18	P4545-02	VNJ-215	BF140091.D	28 Oct 2024 17:07		RC/JU	Ok



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF102824

Review By	yogesh	Review On	10/29/2024 1:35:27 AM		
Supervise By	mohammad	Supervise On	10/29/2024 1:38:39 AM		
SubDirectory	BF102824	HP Acquire Method	BNA_F	HP Processing Method	bf101824
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621				
CCC	SP6624				
Internal Standard/PEM	S12323,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P4545-02MS	VNJ-215MS	BF140092.D	28 Oct 2024 17:35		RC/JU	Ok,M
20	P4545-02MSD	VNJ-215MSD	BF140093.D	28 Oct 2024 18:03		RC/JU	Ok,M
21	P4567-01	WC-1	BF140094.D	28 Oct 2024 18:31		RC/JU	Ok
22	P4574-04	GRAVEL-2	BF140095.D	28 Oct 2024 18:59		RC/JU	Ok
23	P4577-01	TR-05-102524	BF140096.D	28 Oct 2024 19:27		RC/JU	Ok,NS
24	P4566-01	HD-01-102524	BF140097.D	28 Oct 2024 19:55		RC/JU	Ok,NS
25	P4575-01	PL-02-102424	BF140098.D	28 Oct 2024 20:23		RC/JU	Ok,NS

M : Manual Integration



SOP ID : M1311-TCLP-15
SDG No : N/A
Weigh By : JP
Balance ID : WC SC-4
pH Meter ID : WC PH METER-1
Extraction By : JP
Filter By : JP
Pipette ID : WC
Tumbler ID : T-1 / T-2
TCLP Filter ID : 114771

Start Prep Date : 10/25/2024 Time : 17:40
End Prep Date : 10/26/2024 Time : 10:37
Combination Ratio : 20
ZHE Cleaning Batch : N/A
Initial Room Temperature: 23 °C
Final Room Temperature: 22 °C
TCLP Technician Signature : JP
Supervisor By : 12

Standard Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
TCLP-FLUID-1	N/A	WP108622
HCL-TCLP,1N	N/A	WP108584
HNO3-TCLP,1N	N/A	WP108585
pH Strips	N/A	W1931,W1934,W2350,W2755
pH Strips	N/A	W1937,W1938,W1939,W1940,W1941,W1942
1 Liter Amber	N/A	23091
120ml Plastic bottle	N/A	21029
1:1 HNO3	MP81119	N/A

Extraction Conformance/Non-Conformance Comments:

Matrix spikes are added after filtration and before preservation. Tumbler T-1 / T-2 CHECKED,30 RPM. Particle size reduction is not required. p4578-08 is used for MS-MSD.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
10/26/24 11:45	80 TCLP Room	JP 1EF+
	Preparation Group	Analysis Group JAN 2024

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
P4452-03	ETGI-285	01	100.03	2000	N/A	N/A	N/A	11.0	1.5	T-1
P4545-02	VNJ-215	02	100.02	2000	N/A	N/A	N/A	5.8	1.5	T-1
P4547-04	BP-F-21	03	100.01	2000	N/A	N/A	N/A	4.5	1.0	T-1
P4547-08	BP-F-20	04	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-1
P4561-04	BP-F-19	05	100.03	2000	N/A	N/A	N/A	5.6	1.5	T-1
P4561-08	BP-F-18	06	100.04	2000	N/A	N/A	N/A	4.0	1.0	T-1
P4567-04	WC-1	07	100.02	2000	N/A	N/A	N/A	7.2	1.5	T-1
P4567-08	WC-2	08	100.03	2000	N/A	N/A	N/A	7.6	1.0	T-1
P4567-12	WC-3	09	100.04	2000	N/A	N/A	N/A	7.6	1.5	T-1
P4574-03	GRAVEL-1	10	100.02	2000	N/A	N/A	N/A	4.5	1.0	T-1
P4574-06	GRAVEL-2	11	100.03	2000	N/A	N/A	N/A	3.5	1.5	T-2
P4574-08	CONCRETE-1	12	100.02	2000	N/A	N/A	N/A	10.5	1.0	T-2
P4574-10	CONCRETE-2	13	100.02	2000	N/A	N/A	N/A	10.5	1.5	T-2
P4574-12	CONCRETE-3	14	100.03	2000	N/A	N/A	N/A	11.0	1.0	T-2
P4578-06	WB-305-BOT	15	100.04	2000	N/A	N/A	N/A	5.8	1.5	T-2
P4578-08	WB-305-BOT-1	16	100.02	2000	N/A	N/A	N/A	6.0	1.5	T-2
PB164393TB	LEB393	17	N/A	2000	N/A	N/A	N/A	4.93	1.5	T-2

SampleID	ClientID	Sample Weight (g)	Filter Weight (g)	Filtrate (mL)	Filter + Solid (After 100°C)	% solids	% Dry Solids
P4452-03	ETGI-285	N/A	N/A	N/A	N/A	100	N/A
P4545-02	VNJ-215	N/A	N/A	N/A	N/A	100	N/A
P4547-04	BP-F-21	N/A	N/A	N/A	N/A	100	N/A
P4547-08	BP-F-20	N/A	N/A	N/A	N/A	100	N/A
P4561-04	BP-F-19	N/A	N/A	N/A	N/A	100	N/A
P4561-08	BP-F-18	N/A	N/A	N/A	N/A	100	N/A
P4567-04	WC-1	N/A	N/A	N/A	N/A	100	N/A
P4567-08	WC-2	N/A	N/A	N/A	N/A	100	N/A
P4567-12	WC-3	N/A	N/A	N/A	N/A	100	N/A
P4574-03	GRAVEL-1	N/A	N/A	N/A	N/A	100	N/A
P4574-06	GRAVEL-2	N/A	N/A	N/A	N/A	100	N/A
P4574-08	CONCRETE-1	N/A	N/A	N/A	N/A	100	N/A
P4574-10	CONCRETE-2	N/A	N/A	N/A	N/A	100	N/A
P4574-12	CONCRETE-3	N/A	N/A	N/A	N/A	100	N/A
P4578-06	WB-305-BOT	N/A	N/A	N/A	N/A	100	N/A
P4578-08	WB-305-BOT-1	N/A	N/A	N/A	N/A	100	N/A
PB164393TB	LEB393	N/A	N/A	N/A	N/A	N/A	N/A



TCLP Fluid Determination

PB164393

Hot Block ID : WC S-1 /WC S-2Thermometer ID : FLASHPOINT

SampleID	ClientID	Sample Weight (g)	Volume DI Water (mL)	pH after 5 min stir	pH after 10 min stir	Extraction Fluid 1 or 2	pH Extraction Fluid
P4452-03	ETGI-285	5.02	96.5	11.0	4.0	#1	4.93
P4545-02	VNJ-215	5.02	96.5	7.6	2.5	#1	4.93
P4547-04	BP-F-21	5.00	96.5	6.2	2.0	#1	4.93
P4547-08	BP-F-20	5.03	96.5	8.2	3.0	#1	4.93
P4561-04	BP-F-19	5.02	96.5	7.6	2.5	#1	4.93
P4561-08	BP-F-18	5.02	96.5	7.0	2.5	#1	4.93
P4567-04	WC-1	5.03	96.5	9.1	4.0	#1	4.93
P4567-08	WC-2	5.02	96.5	9.5	4.0	#1	4.93
P4567-12	WC-3	5.02	96.5	7.2	3.0	#1	4.93
P4574-03	GRAVEL-1	5.02	96.5	6.0	2.0	#1	4.93
P4574-06	GRAVEL-2	5.03	96.5	5.6	1.5	#1	4.93
P4574-08	CONCRETE-1	5.02	96.5	11.0	4.0	#1	4.93
P4574-10	CONCRETE-2	5.03	96.5	11.0	4.0	#1	4.93
P4574-12	CONCRETE-3	5.02	96.5	11.5	4.5	#1	4.93
P4578-06	WB-305-BOT	5.03	96.5	7.6	2.5	#1	4.93
P4578-08	WB-305-BOT-1	5.02	96.5	8.2	3.0	#1	4.93
PB164393TB	LEB393	N/A	N/A	N/A	N/A	#1	4.93

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tcp p4545

WorkList ID : 184760

Department : TCLP Extraction

Date : 10-25-2024 07:58:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4452-03	ETGI-285	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K51	10/18/2024	1311
P4545-02	VNJ-215	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K61	10/24/2024	1311
P4547-04	BP-F-21	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K51	10/24/2024	1311
P4547-08	BP-F-20	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K51	10/24/2024	1311
P4561-04	BP-F-19	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K51	10/24/2024	1311
P4561-08	BP-F-18	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K63	10/25/2024	1311
P4567-04	WC-1	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K63	10/25/2024	1311
P4567-08	WC-2	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K61	10/25/2024	1311
P4567-12	WC-3	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K61	10/25/2024	1311
P4574-03	GRAVEL-1	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K61	10/25/2024	1311
P4574-06	GRAVEL-2	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K63	10/25/2024	1311
P4574-08	CONCRETE-1	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K63	10/25/2024	1311
P4574-10	CONCRETE-2	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K63	10/25/2024	1311
P4574-12	CONCRETE-3	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K63	10/25/2024	1311
P4578-06	WB-305-BOT	Solid	TCLP Extraction	Cool 4 deg C	PORT06	K63	10/25/2024	1311
P4578-08	WB-305-BOT-1	Solid	TCLP Extraction	Cool 4 deg C	PORT06	K63	10/25/2024	1311

Date/Time 10/25/2024 14:25
 Raw Sample Received by: T.C. (w/c)
 Raw Sample Relinquished by: J.C. (w/c)

Date/Time 10/25/2024

Raw Sample Received by:

Raw Sample Relinquished by:

SOP ID:	M3541-ASE Extraction-14		
Clean Up SOP #:	N/A	Extraction Start Date :	10/26/2024
Matrix :	Water	Extraction Start Time :	12:10
Weigh By:	N/A	Extraction End Date :	10/26/2024
Balance check:	N/A	Extraction End Time :	17:10
Balance ID:	N/A	Concentration By:	EH
pH Strip Lot#:	N/A	Hood ID:	4,6,7
Supervisor By :	rajesh		
Extraction Method:	<input checked="" type="checkbox"/> Separatory Funnel <input type="checkbox"/> Continous Liquid/Liquid <input type="checkbox"/> Sonication <input type="checkbox"/> Waste Dilution <input type="checkbox"/> Soxhlet		

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	50/100 PPM	SP6630
Surrogate	1.0ML	100/150 PPM	SP6638
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3822
Baked Na ₂ SO ₄	N/A	EP2551
10N NaOH	N/A	EP2550
H ₂ SO ₄ 1:1	N/A	EP2548
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

1.5 ML Vial lot# 2210673. pH Adjusted<2 with 1:1 H₂SO₄ & >11with 10 N NaOH.

KD Bath ID: Water bath -01,02 Envap ID: NEVAP-02
KD Bath Temperature: 60 °C Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
10/26/24	RP (Eff - Lab)	Rcvd vrc
17:15	Preparation Group	Analysis Group

Analytical Method: M3541-ASE Extraction-14

Concentration Date: 10/26/2024

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB164393TB	PB164393TB	TCLP BNA	100	6	RUPESH	rajesh	1			SEP-01
PB164444BL	PB164444BL	TCLP BNA	1000	6	RUPESH	rajesh	1			2
PB164444BS	PB164444BS	TCLP BNA	1000	6	RUPESH	rajesh	1			3
P4545-02	VNJ-215	TCLP BNA	100	6	RUPESH	rajesh	1	A		4
P4545-02MS	VNJ-215MS	TCLP BNA	100	6	RUPESH	rajesh	1	A		5
P4545-02MS D	VNJ-215MSD	TCLP BNA	100	6	RUPESH	rajesh	1	A		6
P4547-04	BP-F-21	TCLP BNA	100	6	RUPESH	rajesh	1	A		7
P4547-08	BP-F-20	TCLP BNA	100	6	RUPESH	rajesh	1	A		8
P4561-04	BP-F-19	TCLP BNA	100	6	RUPESH	rajesh	1	A		9
P4561-08	BP-F-18	TCLP BNA	100	6	RUPESH	rajesh	1	A		10
P4567-04	WC-1	TCLP BNA	100	6	RUPESH	rajesh	1	A		11
P4567-08	WC-2	TCLP BNA	100	6	RUPESH	rajesh	1	A		12
P4567-12	WC-3	TCLP BNA	100	6	RUPESH	rajesh	1	A		13
P4574-03	GRAVEL-1	TCLP BNA	100	6	RUPESH	rajesh	1	A		14
P4574-06	GRAVEL-2	TCLP BNA	100	6	RUPESH	rajesh	1	A		15
P4578-06	WB-305-BOT	TCLP BNA	100	6	RUPESH	rajesh	1	A		16
P4578-08	WB-305-BOT-1	TCLP BNA	100	6	RUPESH	rajesh	1	A		17

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
P4452-03	ETGI-285	01	100.03	2000	N/A	N/A	N/A	11.0	1.5	T-1
P4545-02	VNJ-215	02	100.02	2000	N/A	N/A	N/A	5.8	1.5	T-1
P4547-04	BP-F-21	03	100.01	2000	N/A	N/A	N/A	4.5	1.0	T-1
P4547-08	BP-F-20	04	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-1
P4561-04	BP-F-19	05	100.03	2000	N/A	N/A	N/A	5.6	1.5	T-1
P4561-08	BP-F-18	06	100.04	2000	N/A	N/A	N/A	4.0	1.0	T-1
P4567-04	WC-1	07	100.02	2000	N/A	N/A	N/A	7.2	1.5	T-1
P4567-08	WC-2	08	100.03	2000	N/A	N/A	N/A	7.6	1.0	T-1
P4567-12	WC-3	09	100.04	2000	N/A	N/A	N/A	7.6	1.5	T-1
P4574-03	GRAVEL-1	10	100.02	2000	N/A	N/A	N/A	4.5	1.0	T-1
P4574-06	GRAVEL-2	11	100.03	2000	N/A	N/A	N/A	3.5	1.5	T-2
P4574-08	CONCRETE-1	12	100.02	2000	N/A	N/A	N/A	10.5	1.0	T-2
P4574-10	CONCRETE-2	13	100.02	2000	N/A	N/A	N/A	10.5	1.5	T-2
P4574-12	CONCRETE-3	14	100.03	2000	N/A	N/A	N/A	11.0	1.0	T-2
P4578-06	WB-305-BOT	15	100.04	2000	N/A	N/A	N/A	5.8	1.5	T-2
P4578-08	WB-305-BOT-1	16	100.02	2000	N/A	N/A	N/A	6.0	1.5	T-2
PB164393TB	LEB393	17	N/A	2000	N/A	N/A	N/A	4.93	1.5	T-2

10/26/2024
11:50



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789
8900, Fax : 908 789 8922

Prep Standard - Chemical Standard Summary

Order ID : P4578

Test : TCLP BNA

Prepbatch ID : PB164444,

Sequence ID/Qc Batch ID: BF102824,

Standard ID :

EP2548,EP2550,EP2551,SP6558,SP6559,SP6573,SP6620,SP6621,SP6622,SP6623,SP6624,SP6625,SP6626,SP6627,SP6628,SP6630,SP6638,

Chemical ID :

10ul/1000ul

sample,E3551,E3657,E3768,E3788,E3791,E3815,E3822,M5178,S10103,S10247,S10393,S10583,S10977,S10978,S10979,S10980,S11003,S11004,S11005,S11006,S11007,S11008,S11009,S11010,S11085,S11098,S11138,S11141,S11159,S11494,S11541,S11566,S11649,S11766,S11771,S11772,S11773,S11774,S11775,S12033,S12078,S12097,S12105,S12106,S12107,S12108,S12109,S12110,S12111,S12113,S12117,S12126,S12127,S12128,S12129,S12130,S12131,S12132,S12133,S12134,S12187,S12188,S12189,S12207,S12275,S12314,S12323,S12453,S12454,S12455,S12456,S12457,S12458,S12459,S12460,S12461,S12509,S12510,W3112,

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
314	1.1 H2SO4 SOLN	EP2548	10/16/2024	04/16/2025	Rajesh Parikh	None	None	RUPESHKUMAR SHAH 10/16/2024

FROM 1000.00000ml of W3112 + 4000.00000ml of M5178 = Final Quantity: 2000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1874	10 N SODIUM HYDROXIDE SOLN	EP2550	10/16/2024	04/16/2025	Rajesh Parikh	Extraction_SC ALE_2 (EX-SC-2)	None	RUPESHKUMAR SHAH 10/16/2024

FROM 1000.00000ml of W3112 + 400.00000gram of E3657 = Final Quantity: 1000.000 ml

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2551	10/18/2024	01/03/2025	Rajesh Parikh	Extraction_SC ALE_2 (EX-SC-2)	None	RUPESHKUMAR SHAH 10/18/2024

FROM 4000.0000gram of E3551 = Final Quantity: 4000.000 gram

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
18	Second Source Calibration Stock Standard, 100 PPM, (8270/625/CLP)	SP6558	07/09/2024	11/30/2024	Jagrut Upadhyay	None	None	mohammad ahmed 07/11/2024

FROM 0.04000ml of S10977 + 0.08000ml of S11003 + 0.10000ml of S11766 + 0.20000ml of S11566 + 0.20000ml of S12097 + 0.20000ml of S12117 + 1.18000ml of E3768 = Final Quantity: 2.000 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
416	40 ng BNA ICV, 40 PPM	SP6559	07/09/2024	11/30/2024	Jagrut Upadhyay	None	None	mohammad ahmed 07/11/2024

FROM 0.01000ml of S12033 + 0.60000ml of E3768 + 0.40000ml of SP6558 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3895	50 ug/ml DFTPP 8270E	SP6573	07/15/2024	01/08/2025	Rahul Chavli	None	None	Yogesh Patel 07/17/2024

FROM 1.00000ml of S10247 + 19.00000ml of E3768 = Final Quantity: 20.000 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3764	8270/625 Stock solution 100 ng	SP6620	09/11/2024	02/08/2025	Jagrut Upadhyay	None	None	mohammad ahmed 09/11/2024

FROM 0.26700ml of S10103 + 0.40000ml of S11494 + 0.50000ml of S12113 + 1.00000ml of S11085 + 1.00000ml of S11098 + 1.00000ml of S11159 + 1.00000ml of S12078 + 1.00000ml of S12275 + 3.83300ml of E3791 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
413	80 ng BNA ICC, 80 PPM	SP6621	09/11/2024	02/08/2025	Jagrut Upadhyay	None	None	mohammad ahmed 09/11/2024

FROM 0.01000ml of S12314 + 0.20000ml of E3791 + 0.80000ml of SP6620 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
412	60 ng BNA ICC, 60 PPM	SP6622	09/11/2024	02/08/2025	Jagrut Upadhyay	None	None	mohammad ahmed 09/11/2024

FROM 0.01000ml of S12314 + 0.40000ml of E3791 + 0.60000ml of SP6620 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
411	50 ng BNA ICC, 50 PPM	SP6623	09/11/2024	02/08/2025	Jagrut Upadhyay	None	None	mohammad ahmed 09/11/2024

FROM 0.01000ml of S12314 + 0.50000ml of E3791 + 0.50000ml of SP6620 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
410	40 ng BNA ICC, 40 PPM	SP6624	09/11/2024	02/08/2025	Jagrut Upadhyay	None	None	mohammad ahmed 09/11/2024

FROM 0.01000ml of S12314 + 0.60000ml of E3791 + 0.40000ml of SP6620 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3678	20 ng BNA ICC, 20 PPM	SP6625	09/11/2024	02/08/2025	Jagrut Upadhyay	None	None	mohammad ahmed 09/11/2024

FROM 0.01000ml of S12314 + 0.80000ml of E3791 + 0.20000ml of SP6620 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
408	10 ng BNA ICC, 10 PPM	SP6626	09/11/2024	02/08/2025	Jagrut Upadhyay	None	None	mohammad ahmed 09/11/2024

FROM 0.01000ml of S12314 + 0.90000ml of E3791 + 0.10000ml of SP6620 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
407	5 ng BNA ICC, 5 PPM	SP6627	09/11/2024	02/08/2025	Jagrut Upadhyay	None	None	mohammad ahmed 09/11/2024

FROM 0.01000ml of S12314 + 0.95000ml of E3791 + 0.05000ml of SP6620 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
175	2.5 ng BNA ICC, 2.5 PPM	SP6628	09/11/2024	02/08/2025	Jagrut Upadhyay	None	None	mohammad ahmed 09/11/2024

FROM 0.01000ml of S12314 + 0.50000ml of E3791 + 0.50000ml of SP6627 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
171	8270/625 Spike Solution, 50/100 PPM	SP6630	09/13/2024	02/12/2025	Rahul Chavli	None	None	Yogesh Patel 10/29/2024

FROM 0.10000ml of S12134 + 0.10000ml of S12461 + 0.20000ml of S12453 + 0.30000ml of S11775 + 0.30000ml of S12510 + 0.40000ml of S10393 + 0.40000ml of S10583 + 0.40000ml of S11138 + 0.40000ml of S11141 + 0.40000ml of S11649 + 0.70000ml of S11771 + 0.70000ml of S12105 + 0.80000ml of S12126 + 1.10000ml of S12110 + 1.10000ml of S12133 + 1.10000ml of S12460 + 1.20000ml of S12455 + 1.20000ml of S12456 + 1.20000ml of S12457 + 1.20000ml of S12458 + 1.20000ml of S12459 + 1.20000ml of S12509 + 1.30000ml of S11541 + 1.30000ml of S11772 + 1.30000ml of S11773 + 1.30000ml of S12106 + 1.30000ml of S12108 + 1.30000ml of S12111 + 1.30000ml of S12127 + 1.30000ml of S12129 + 1.30000ml of S12130 + 1.30000ml of S12131 + 1.30000ml of S12454 + 1.40000ml of S11774 + 1.40000ml of S12107 + 1.40000ml of S12109 + 1.40000ml of S12128 + 1.40000ml of S12132 + 163.00000ml of E3788 = Final Quantity: 200.000 ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	01/03/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551
PCI Scientific Supply, Inc.	PC19510-5 / Sodium Hydroxide Pellets 2.5 Kg, Pk of 4	23B1556310	12/31/2025	12/04/2023 / Rajesh	12/01/2023 / Rajesh	E3657
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24E2462004	01/08/2025	07/08/2024 / Rajesh	06/21/2024 / Rajesh	E3768
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	23H1462005	04/23/2025	08/13/2024 / Rajesh	08/13/2024 / Rajesh	E3788
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24G2362009	03/09/2025	09/09/2024 / Rajesh	09/03/2024 / Rajesh	E3791
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	04/04/2025	10/04/2024 / Rajesh	10/04/2024 / Rajesh	E3815

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24I2662006	04/23/2025	10/24/2024 / Rajesh	10/24/2024 / Rajesh	E3822
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	0000281827	03/29/2026	05/25/2022 / william	04/05/2022 / william	M5178
CPI International	Z-112090-04 / CLP Acid Surrogate Solution, 7500 mg/L, 1ml	440246	02/08/2025	08/08/2024 / Jagrut	12/09/2021 / Christian	S10103
Restek	31615 / SV Mixture, GC/MS Tuning Mixture, CH2Cl2, 1mL,	A0182667	01/15/2025	07/15/2024 / Rahul	03/18/2022 / Christian	S10247
Restek	555871 / Custom Standard, 4-nitrophenol Std [CS 5238-4]	A0185300	03/13/2025	09/13/2024 / yogesh	05/18/2022 / Christian	S10393
Restek	555868 / Custom Standard, Benzidine Std [CS 5328-1]	A0186373	02/12/2025	08/12/2024 / Rahul	07/05/2022 / Christian	S10583



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0188108	11/30/2024	05/31/2024 / Jagrut	12/28/2022 / Christian	S10977
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0188108	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S10978
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0188108	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S10979
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0188108	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S10980
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	11/30/2024	05/31/2024 / Jagrut	12/28/2022 / Christian	S11003
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S11004



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S11005
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S11006
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S11007
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S11008
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S11009
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S11010



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-010074-07 / 3,3'-Dichlorobenzidine Solution, 1,000 mg/L, 1 ml, (Maximum Expiration: 180 days)	406703	03/11/2025	09/11/2024 / Rahul	02/07/2023 / Christian	S11085
CPI International	z-110381-01 / 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1ml	495831	03/11/2025	09/11/2024 / Rahul	02/07/2023 / Christian	S11098
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555870 / Custom Standard, 2,4-dinitrophenol Std [CS 5328-3]	A0194698	02/12/2025	08/12/2024 / Rahul	02/20/2023 / Christian	S11138
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555869 / Custom Standard, hexachlorocyclopentadiene Std [CS 5328-2]	A0194702	02/12/2025	08/12/2024 / Rahul	02/20/2023 / Christian	S11141
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110817-01 / Custom 8270 Mix, 4-55, 1000 mg/L, 1 ml, (Maximum Expiration: 90 Days)	414125	03/11/2025	09/11/2024 / Rahul	03/06/2023 / Christian	S11159
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110094-02 / CLP Base/Neutral Surrogate Solution, 5000 mg/L, 1ml	506889	02/08/2025	08/08/2024 / Jagrut	08/11/2023 / Yogesh	S11494

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0201940	03/13/2025	09/13/2024 / yogesh	09/18/2023 / Kiran	S11541
[CS 4978-1]						
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0201940	12/05/2024	06/05/2024 / Rahul	09/18/2023 / Kiran	S11566
[CS 4978-1]						
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555872 / Custom Standard, pentachlorophenol Std [CS 5328-5]	A0201728	03/13/2025	09/13/2024 / anahy	11/09/2023 / Yogesh	S11649
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	12/14/2024	06/14/2024 / Rahul	11/21/2023 / Rahul	S11766
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	02/20/2025	08/20/2024 / Rahul	11/21/2023 / Rahul	S11771
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	03/13/2025	09/13/2024 / anahy	11/21/2023 / Rahul	S11772

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	03/13/2025	09/13/2024 / anahy	11/21/2023 / Rahul	S11773
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	03/13/2025	09/13/2024 / anahy	11/21/2023 / Rahul	S11774
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	03/13/2025	09/13/2024 / anahy	11/21/2023 / Rahul	S11775
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL	A0201320	01/01/2025	07/01/2024 / Rahul	12/21/2023 / Rahul	S12033
CPI International	Z-110816-01 / Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL, (Maximum Expiration: 180 Days)	414127	03/11/2025	09/11/2024 / Rahul	01/31/2024 / Rahul	S12078
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0207706	01/09/2025	07/09/2024 / Jagrut	02/05/2024 / Rahul	S12097

[CS 4978-2]



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0207706	02/12/2025	08/12/2024 / Rahul	02/05/2024 / Rahul	S12105

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0207706	03/13/2025	09/13/2024 / anahy	02/05/2024 / Rahul	S12106

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0207706	03/13/2025	09/13/2024 / anahy	02/05/2024 / Rahul	S12107

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0207706	03/13/2025	09/13/2024 / anahy	02/05/2024 / Rahul	S12108

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0207706	03/13/2025	09/13/2024 / anahy	02/05/2024 / Rahul	S12109

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0207706	03/13/2025	09/13/2024 / anahy	02/05/2024 / Rahul	S12110

[CS 4978-2]



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0207706	03/13/2025	09/13/2024 / anahy	02/05/2024 / Rahul	S12111
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-010223-01 / 1,4-Dioxane Solution, 2,000mg/L, 1ml	454157	02/09/2025	08/09/2024 / Jagrut	03/08/2024 / Rahul	S12113
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0203726	12/05/2024	06/05/2024 / Rahul	03/15/2024 / Rahul	S12117
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0203726	02/12/2025	08/12/2024 / Rahul	03/15/2024 / Rahul	S12126
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0203726	03/13/2025	09/13/2024 / anahy	03/15/2024 / Rahul	S12127
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0203726	03/13/2025	09/13/2024 / anahy	03/15/2024 / Rahul	S12128



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0203726	03/13/2025	09/13/2024 / anahy	03/15/2024 / Rahul	S12129
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0203726	03/13/2025	09/13/2024 / anahy	03/15/2024 / Rahul	S12130
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0203726	03/13/2025	09/13/2024 / anahy	03/15/2024 / Rahul	S12131
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0203726	03/13/2025	09/13/2024 / anahy	03/15/2024 / Rahul	S12132
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0203726	03/13/2025	09/13/2024 / anahy	03/15/2024 / Rahul	S12133
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0203726	03/13/2025	09/13/2024 / anahy	03/15/2024 / Rahul	S12134

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0206206	04/10/2025	10/10/2024 / anahy	03/15/2024 / Rahul	S12187
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0206206	04/10/2025	10/10/2024 / anahy	03/15/2024 / Rahul	S12188
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0206206	04/10/2025	10/10/2024 / anahy	03/15/2024 / Rahul	S12189
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0206381	04/10/2025	10/10/2024 / anahy	03/15/2024 / Rahul	S12207
CPI International	Z-010442-07 / Benzaldehyde Solution, 1000 mg/L, 1.3 ml, (Maximum Expiration: 90 Days)	495833	03/11/2025	09/11/2024 / Rahul	05/24/2024 / Rahul	S12275
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL	A0206540	03/04/2025	09/04/2024 / anahy	05/30/2024 / Rahul	S12314



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH ₂ Cl ₂ , 1mL	A0206540	04/23/2025	10/23/2024 / anahy	05/30/2024 / Rahul	S12323

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	02/12/2025	08/12/2024 / Rahul	07/23/2024 / RAHUL	S12453

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	03/13/2025	09/13/2024 / anahy	07/23/2024 / RAHUL	S12454

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	03/13/2025	09/13/2024 / anahy	07/23/2024 / RAHUL	S12455

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	03/13/2025	09/13/2024 / anahy	07/23/2024 / RAHUL	S12456

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	03/13/2025	09/13/2024 / anahy	07/23/2024 / RAHUL	S12457

[CS 4978-1]



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	03/13/2025	09/13/2024 / anahy	07/23/2024 / RAHUL	S12458

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	03/13/2025	09/13/2024 / anahy	07/23/2024 / RAHUL	S12459

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	03/13/2025	09/13/2024 / anahy	07/23/2024 / RAHUL	S12460

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	03/13/2025	09/13/2024 / anahy	07/23/2024 / RAHUL	S12461

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0214017	03/13/2025	09/13/2024 / anahy	07/23/2024 / RAHUL	S12509

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0214017	03/13/2025	09/13/2024 / anahy	07/23/2024 / RAHUL	S12510

[CS 4978-2]



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: Lot No.: Storage: Solvent: Exp. Date: Description:
Z-010074-07 406703 ≤ -10 °C Methylene Chloride 3/30/2025 3,3'-Dichlorobenzidine Solution, 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
3,3'-dichlorobenzidine	91-94-1	99.5	74.3.26P	989 ± 7.53

Received on
02/07/23
by
C6

S11084
to
S11088

*Not a certified value

Certified By:

A handwritten signature in black ink, appearing to read "Terry".

Jacob Mulloy
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: Lot No.: Storage: Solvent: Exp. Date: Description:
Z-110817-01 414125 ≤ -10 °C Methylene Chloride 6/21/2025 Custom 8270 Mix, 4-55, 1000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
acetophenone	98-86-2	99.2	85.8.1P	998 ± 11.5
benzoic acid	65-85-0	100	123.7.1P	1010 ± 5.88
biphenyl	92-52-4	99.9	366.29.1P	999 ± 5.82
1,2,4,5-tetrachlorobenzene	95-94-3	99.7	53.7.2P	993 ± 5.79

Received on

02/07/23

by

CG

S 11089

to

S 11093

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By:

Shane Overcash
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



5580 Skylane Blvd
Santa Rosa, CA 95403

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: Lot No.: Storage: Solvent: Exp. Date: Description:
Z-112090 440246 $\leq -10^{\circ}\text{C}$ Methylene Chloride 2/16/2026 CLP Acid Surrogate Solution, 7,500 mg/L, 1 mL
-04

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
2-chlorophenol-d ₄	93951-73-6	99.3	248.12.7P	7487 \pm 17.2
2-fluorophenol	367-12-4	99.8	10.7.3.3P	7513 \pm 17.26
phenol-d ₆	13127-88-3	99.9	949.120.8P	7481 \pm 17.19
2,4,6-tribromophenol	118-79-6	99.8	12.1.6P	7469 \pm 17.17

Received on

02/25/21

by
CG

S9236
+0

S9240

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA


All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certified By:

Erica Castiglione
Chemist



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Received on
02/07/23 by C6

SH067 S11096
to
S11099

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 4

Catalog No.: Lot No.: Storage: Solvent: Exp. Date: Description:
Z-110381-01 495831 ≤ -10 °C Methylene Chloride 10/30/2027 Method 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
acenaphthene	83-32-9	99.9	13.1.5P	1003 ± 17.27
acenaphthylene	208-96-8	97.6	14.290.1P	999.8 ± 17.22
aniline	62-53-3	99.9	64.7.1P	995 ± 17.13
anthracene	120-12-7	99.5	15.7.1P	1001 ± 17.24
azobenzene	103-33-3	98.1	252.7.2P	999.1 ± 17.21
benzo[a]anthracene	56-55-3	100	16.7.3P	1001 ± 17.24
benzo[b]fluoranthene	205-99-2	99.8	17.421.3P	1001 ± 19.91
benzo[k]fluoranthene	207-08-9	98.9	18.421.4P	1001 ± 17.92
benzo[ghi]perylene	191-24-2	93	19.286.4P	999.6 ± 19.88
benzo[a]pyrene	50-32-8	97	20.286.2P	999.1 ± 26.35
benzyl alcohol	100-51-6	99.9	65.18.1P	1001 ± 17.24
bis(2-chloroethoxy)methane	111-91-1	99.1	31.3.15P	999.7 ± 17.89
bis(2-chloroethyl)ether	111-44-4	99.8	32.7.1P	1001 ± 17.23
bis(2-chloro-1-methylethyl) ether	108-60-1	99.5	34.3.13P	999.5 ± 17.89
bis(2-ethylhexyl)adipate	103-23-1	99.5	874.7.1P	999.5 ± 17.21
bis(2-ethylhexyl)phthalate	117-81-7	99.4	33.29.1P	998.8 ± 19.86
4-bromophenyl phenyl ether	101-55-3	99.4	35.7.1P	999.1 ± 17.2
butyl benzyl phthalate	85-68-7	98.4	36.1.6P	984.7 ± 19.58
carbazole	86-74-8	99.4	239.7.2P	1000 ± 17.22

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By:

Briana Smith
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certificate of Analysis

Page 4 of 4

Catalog No.: Z-110381-01

Lot No.: 495831

Expiration Date: 10/30/2027

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2,4-trichlorobenzene	120-82-1	99.6	54.29.1P	1000 ± 17.22
2,4,5-trichlorophenol	95-95-4	96.5	121.7.1.1P	1000 ± 17.22
2,4,6-trichlorophenol	88-06-2	99.6	113.7.1P	1002 ± 17.25

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By:



Briana Smith
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

RESTEK® CERTIFIED REFERENCE MATERIAL

110 Benner Circle
 Bellefonte, PA 16823-8812
 Tel: (800)356-1688
 Fax: (814)353-1309

www.restek.com



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received on
 03/11/2022

b7
 CG

S10242
 to

S10247

Catalog No. : 31615

Lot No.: A0182667

Description : GC/MS Tuning Mixture

GC/MS Tuning Mixture 1,000 μ g/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2025

Storage: 10°C or colder

Handling: Contains carcinogen/reproductive toxin.

Ship: Ambient

C E R T I F I E D V A L U E S

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	Pentachlorophenol CAS # 87-86-5 Purity 99%	1,003.6 μ g/mL	+/- 5.8897 μ g/mL	+/- 45.7132 μ g/mL	+/- 66.0037 μ g/mL
2	DFTPP (Decafluorotriphenylphosphine) CAS # 5074-71-5 Purity 95%	1,006.6 μ g/mL	+/- 5.9074 μ g/mL	+/- 45.8508 μ g/mL	+/- 66.2023 μ g/mL
3	Benzidine CAS # 92-87-5 Purity 99%	1,008.4 μ g/mL	+/- 5.9179 μ g/mL	+/- 45.9318 μ g/mL	+/- 66.3193 μ g/mL
4	4,4'-DDT CAS # 50-29-3 Purity 99%	1,007.6 μ g/mL	+/- 5.9132 μ g/mL	+/- 45.8954 μ g/mL	+/- 66.2667 μ g/mL

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 08-Mar-2022 Balance: B345965662

Marilina Cowan - Operations Tech I

Date Passed: 10-Mar-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

RESTEK® CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com



Gravimetric Certificate

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. :	<u>555871</u>	Lot No.:	<u>A0185300</u>
Description :	Custom 4-Nitrophenol Standard		
	Custom 4-Nitrophenol Standard 25,000 μ g/mL, Methanol, 1mL/ampul		
Container Size :	<u>2 mL</u>	Pkg Amt:	<u>> 1 mL</u>
Expiration Date :	<u>May 31, 2025</u>	Storage:	<u>10°C or colder</u>
		Ship:	<u>Ambient</u>

Received by

CG on

05/18/22

\$10393

+0

\$10402

C E R T I F I E D V A L U E S

Component #	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	4-Nitrophenol CAS # 100-02-7 Purity 99%	25,060.0 μ g/mL	+/- 231.9100	μ g/mL	Gravimetric
	(Lot MKCN1089)		+/- 753.2622	μ g/mL	Unstressed
			+/- 905.6020	μ g/mL	Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Katelyn McGinn - Operations Tech I

Date Mixed: 16-May-2022 Balance: 1128342314

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined stressed}} = k \sqrt{U_{\text{gravimetric}}^2 + U_{\text{homogeneity}}^2 + U_{\text{storage stability}}^2 + U_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at [| Label Conditions | Standard Conditions | Non-Standard Conditions |
|---|---------------------|-------------------------|
| 25°C Nominal \(Room Temperature\) | < 60°C | ≥ 60°C up to 7 days |
| 10°C or colder \(Refrigerate\) | < 40°C | ≥ 40°C up to 7 days |
| 0°C or colder \(Freezer\)
-20°C or colder \(Deep Freezer\) | < 25°C | ≥ 25°C up to 7 days |](http://www.restek.com>Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.• Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.</div><div data-bbox=)

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at [### Manufacturing Notes:](http://www.restek.com>Contact-Us.• The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.</div><div data-bbox=)

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.

RESTEK® CERTIFIED REFERENCE MATERIAL

110 Benner Circle
 Bellefonte, PA 16823-8812
 Tel: (800)356-1688
 Fax: (814)353-1309

www.restek.com



Gravimetric Certificate



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555868

Lot No.: A0186373

Description : Custom Benzidine Standard

Custom Benzidine Standard 25,000 μ g/mL, Methanol, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : June 30, 2025

Storage: 10°C or colder

Handling: Contains carcinogen/reproductive toxin.

Ship: Ambient

Received by

CG

on

07/05/22

S 10583

to

S 10592

C E R T I F I E D V A L U E S

Component #	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	Benzidine	25,200.0 μ g/mL	+/- 233.2055	μ g/mL	Gravimetric
CAS #	92-87-5	(Lot 220511RSR)	+/- 351.6606	μ g/mL	Unstressed
Purity	99%		+/- 512.6054	μ g/mL	Stressed

Solvent: Methanol
 CAS # 67-56-1
 Purity 99%


 Tom Suckar - Mix Technician

Date Mixed: 16-Jun-2022 Balance: 1122030677

Manufactured under Restek's ISO 9001:2015
 Registered Quality System
 Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined stressed}} = k \sqrt{U_{\text{gravimetric}}^2 + U_{\text{homogeneity}}^2 + U_{\text{storage stability}}^2 + U_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.

RESTEK® CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087

Lot No.: A0188108

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000 μ g/mL, Methanol, 5mL/ampul

Container Size : 5 mL

Pkg Amt: > 5 mL

Expiration Date : August 31, 2030

Storage: 10°C or colder

Ship: Ambient

Received by
CG on
12/28/22
S10951
to
S10980

C E R T I F I E D V A L U E S

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	2-Fluorophenol CAS # 367-12-4 Purity 99%	10,088.5 μ g/mL	+/- 58.6554	μ g/mL	Gravimetric
	(Lot STBF3761V)		+/- 294.4162	μ g/mL	Unstressed
			+/- 357.2628	μ g/mL	Stressed
2	Phenol-d6 CAS # 13127-88-3 Purity 99%	10,043.3 μ g/mL	+/- 58.3923	μ g/mL	Gravimetric
	(Lot PR-31262)		+/- 293.0957	μ g/mL	Unstressed
			+/- 355.6603	μ g/mL	Stressed
3	2,4,6-Tribromophenol CAS # 118-79-6 Purity 99%	10,010.0 μ g/mL	+/- 58.1990	μ g/mL	Gravimetric
	(Lot MKCJ7664)		+/- 292.1253	μ g/mL	Unstressed
			+/- 354.4829	μ g/mL	Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.



Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022 Balance: 1127510105

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397


RESTEK® CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087

Lot No.: A0188108

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000 μ g/mL, Methanol, 5mL/ampul

Container Size : 5 mL

Pkg Amt: > 5 mL

Expiration Date : August 31, 2030

Storage: 10°C or colder

Ship: Ambient

Received by
CG on
12/28/22
S10951
to
S10980

C E R T I F I E D V A L U E S

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	2-Fluorophenol CAS # 367-12-4 Purity 99%	10,088.5 μ g/mL	+/- 58.6554	μ g/mL	Gravimetric
	(Lot STBF3761V)		+/- 294.4162	μ g/mL	Unstressed
			+/- 357.2628	μ g/mL	Stressed
2	Phenol-d6 CAS # 13127-88-3 Purity 99%	10,043.3 μ g/mL	+/- 58.3923	μ g/mL	Gravimetric
	(Lot PR-31262)		+/- 293.0957	μ g/mL	Unstressed
			+/- 355.6603	μ g/mL	Stressed
3	2,4,6-Tribromophenol CAS # 118-79-6 Purity 99%	10,010.0 μ g/mL	+/- 58.1990	μ g/mL	Gravimetric
	(Lot MKCJ7664)		+/- 292.1253	μ g/mL	Unstressed
			+/- 354.4829	μ g/mL	Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.



Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022 Balance: 1127510105

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397


RESTEK® CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087

Lot No.: A0188108

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000 μ g/mL, Methanol, 5mL/ampul

Container Size : 5 mL

Pkg Amt: > 5 mL

Expiration Date : August 31, 2030

Storage: 10°C or colder

Ship: Ambient

Received by
CG on
12/28/22
S10951
to
S10980

C E R T I F I E D V A L U E S

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	2-Fluorophenol CAS # 367-12-4 Purity 99%	10,088.5 μ g/mL	+/- 58.6554	μ g/mL	Gravimetric
	(Lot STBF3761V)		+/- 294.4162	μ g/mL	Unstressed
			+/- 357.2628	μ g/mL	Stressed
2	Phenol-d6 CAS # 13127-88-3 Purity 99%	10,043.3 μ g/mL	+/- 58.3923	μ g/mL	Gravimetric
	(Lot PR-31262)		+/- 293.0957	μ g/mL	Unstressed
			+/- 355.6603	μ g/mL	Stressed
3	2,4,6-Tribromophenol CAS # 118-79-6 Purity 99%	10,010.0 μ g/mL	+/- 58.1990	μ g/mL	Gravimetric
	(Lot MKCJ7664)		+/- 292.1253	μ g/mL	Unstressed
			+/- 354.4829	μ g/mL	Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.



Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022 Balance: 1127510105

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397


RESTEK® CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087

Lot No.: A0188108

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000 μ g/mL, Methanol, 5mL/ampul

Container Size : 5 mL

Pkg Amt: > 5 mL

Expiration Date : August 31, 2030

Storage: 10°C or colder

Ship: Ambient

Received by
CG on
12/28/22
S10951
to
S10980

C E R T I F I E D V A L U E S

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	2-Fluorophenol CAS # 367-12-4 Purity 99%	10,088.5 μ g/mL	+/- 58.6554	μ g/mL	Gravimetric
	(Lot STBF3761V)		+/- 294.4162	μ g/mL	Unstressed
			+/- 357.2628	μ g/mL	Stressed
2	Phenol-d6 CAS # 13127-88-3 Purity 99%	10,043.3 μ g/mL	+/- 58.3923	μ g/mL	Gravimetric
	(Lot PR-31262)		+/- 293.0957	μ g/mL	Unstressed
			+/- 355.6603	μ g/mL	Stressed
3	2,4,6-Tribromophenol CAS # 118-79-6 Purity 99%	10,010.0 μ g/mL	+/- 58.1990	μ g/mL	Gravimetric
	(Lot MKCJ7664)		+/- 292.1253	μ g/mL	Unstressed
			+/- 354.4829	μ g/mL	Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.



Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022 Balance: 1127510105

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 Lot No.: A0189418
 Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000 μ g/mL, Methylene Chloride, 5mL/ampul
 Container Size : 5 mL Pkg Amt: > 5 mL
 Expiration Date : August 31, 2028 Storage: 10°C or colder
 Handling: Sonicate prior to use. Ship: Ambient

Received by
CG on
12/28/22
Storage
to
Silo 10

C E R T I F I E D V A L U E S

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	Nitrobenzene-d5 CAS # 4165-60-0 Purity 99%	5,009.8 μ g/mL	+/- 29.1271 μ g/mL	+/- 225.6421 μ g/mL	+/- 250.3778 μ g/mL
2	2-Fluorobiphenyl CAS # 321-60-8 Purity 99%	5,026.6 μ g/mL	+/- 29.2250 μ g/mL	+/- 226.4003 μ g/mL	+/- 251.2191 μ g/mL
3	p-Terphenyl-d14 CAS # 1718-51-0 Purity 99%	5,027.3 μ g/mL	+/- 29.2289 μ g/mL	+/- 226.4304 μ g/mL	+/- 251.2524 μ g/mL

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)**Carrier Gas:**

hydrogen-constant pressure 10 psi.

Temp. Program:40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)**Inj. Temp:**


250°C

Det. Temp:


330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 Balance: 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 Lot No.: A0189418
 Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000 μ g/mL, Methylene Chloride, 5mL/ampul
 Container Size : 5 mL Pkg Amt: > 5 mL
 Expiration Date : August 31, 2028 Storage: 10°C or colder
 Handling: Sonicate prior to use. Ship: Ambient

Received by
CG on
12/28/22
Storage
to
Silo 10

C E R T I F I E D V A L U E S

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	Nitrobenzene-d5 CAS # 4165-60-0 Purity 99%	5,009.8 μ g/mL	+/- 29.1271 μ g/mL	+/- 225.6421 μ g/mL	+/- 250.3778 μ g/mL
2	2-Fluorobiphenyl CAS # 321-60-8 Purity 99%	5,026.6 μ g/mL	+/- 29.2250 μ g/mL	+/- 226.4003 μ g/mL	+/- 251.2191 μ g/mL
3	p-Terphenyl-d14 CAS # 1718-51-0 Purity 99%	5,027.3 μ g/mL	+/- 29.2289 μ g/mL	+/- 226.4304 μ g/mL	+/- 251.2524 μ g/mL

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)**Carrier Gas:**

hydrogen-constant pressure 10 psi.

Temp. Program:40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)**Inj. Temp:**


250°C

Det. Temp:


330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 Balance: 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 Lot No.: A0189418
 Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000 μ g/mL, Methylene Chloride, 5mL/ampul
 Container Size : 5 mL Pkg Amt: > 5 mL
 Expiration Date : August 31, 2028 Storage: 10°C or colder
 Handling: Sonicate prior to use. Ship: Ambient

Received by
CG on
12/28/22
Storage
to
Silo 10

C E R T I F I E D V A L U E S

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	Nitrobenzene-d5 CAS # 4165-60-0 Purity 99%	5,009.8 μ g/mL	+/- 29.1271 μ g/mL	+/- 225.6421 μ g/mL	+/- 250.3778 μ g/mL
2	2-Fluorobiphenyl CAS # 321-60-8 Purity 99%	5,026.6 μ g/mL	+/- 29.2250 μ g/mL	+/- 226.4003 μ g/mL	+/- 251.2191 μ g/mL
3	p-Terphenyl-d14 CAS # 1718-51-0 Purity 99%	5,027.3 μ g/mL	+/- 29.2289 μ g/mL	+/- 226.4304 μ g/mL	+/- 251.2524 μ g/mL

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)**Carrier Gas:**

hydrogen-constant pressure 10 psi.

Temp. Program:40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)**Inj. Temp:**


250°C

Det. Temp:


330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 Balance: 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 Lot No.: A0189418
 Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000 μ g/mL, Methylene Chloride, 5mL/ampul
 Container Size : 5 mL Pkg Amt: > 5 mL
 Expiration Date : August 31, 2028 Storage: 10°C or colder
 Handling: Sonicate prior to use. Ship: Ambient

Received by
CG on
12/28/22
Storage
to
Silo 10

C E R T I F I E D V A L U E S

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	Nitrobenzene-d5 CAS # 4165-60-0 Purity 99%	5,009.8 μ g/mL	+/- 29.1271 μ g/mL	+/- 225.6421 μ g/mL	+/- 250.3778 μ g/mL
2	2-Fluorobiphenyl CAS # 321-60-8 Purity 99%	5,026.6 μ g/mL	+/- 29.2250 μ g/mL	+/- 226.4003 μ g/mL	+/- 251.2191 μ g/mL
3	p-Terphenyl-d14 CAS # 1718-51-0 Purity 99%	5,027.3 μ g/mL	+/- 29.2289 μ g/mL	+/- 226.4304 μ g/mL	+/- 251.2524 μ g/mL

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)**Carrier Gas:**

hydrogen-constant pressure 10 psi.

Temp. Program:40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)**Inj. Temp:**


250°C

Det. Temp:


330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 Balance: 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 Lot No.: A0189418
 Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000 μ g/mL, Methylene Chloride, 5mL/ampul
 Container Size : 5 mL Pkg Amt: > 5 mL
 Expiration Date : August 31, 2028 Storage: 10°C or colder
 Handling: Sonicate prior to use. Ship: Ambient

Received by
CG on
12/28/22
Storage
to
Silo 10

C E R T I F I E D V A L U E S

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	Nitrobenzene-d5 CAS # 4165-60-0 Purity 99%	5,009.8 μ g/mL	+/- 29.1271 μ g/mL	+/- 225.6421 μ g/mL	+/- 250.3778 μ g/mL
2	2-Fluorobiphenyl CAS # 321-60-8 Purity 99%	5,026.6 μ g/mL	+/- 29.2250 μ g/mL	+/- 226.4003 μ g/mL	+/- 251.2191 μ g/mL
3	p-Terphenyl-d14 CAS # 1718-51-0 Purity 99%	5,027.3 μ g/mL	+/- 29.2289 μ g/mL	+/- 226.4304 μ g/mL	+/- 251.2524 μ g/mL

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)**Carrier Gas:**

hydrogen-constant pressure 10 psi.

Temp. Program:40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)**Inj. Temp:**


250°C

Det. Temp:


330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 Balance: 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 Lot No.: A0189418
 Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000 μ g/mL, Methylene Chloride, 5mL/ampul
 Container Size : 5 mL Pkg Amt: > 5 mL
 Expiration Date : August 31, 2028 Storage: 10°C or colder
 Handling: Sonicate prior to use. Ship: Ambient

Received by
CG on
12/28/22
Storage
to
Silo 10

C E R T I F I E D V A L U E S

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	Nitrobenzene-d5 CAS # 4165-60-0 Purity 99%	5,009.8 μ g/mL	+/- 29.1271 μ g/mL	+/- 225.6421 μ g/mL	+/- 250.3778 μ g/mL
2	2-Fluorobiphenyl CAS # 321-60-8 Purity 99%	5,026.6 μ g/mL	+/- 29.2250 μ g/mL	+/- 226.4003 μ g/mL	+/- 251.2191 μ g/mL
3	p-Terphenyl-d14 CAS # 1718-51-0 Purity 99%	5,027.3 μ g/mL	+/- 29.2289 μ g/mL	+/- 226.4304 μ g/mL	+/- 251.2524 μ g/mL

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)**Carrier Gas:**

hydrogen-constant pressure 10 psi.

Temp. Program:40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)**Inj. Temp:**


250°C

Det. Temp:


330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 Balance: 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 Lot No.: A0189418
 Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000 μ g/mL, Methylene Chloride, 5mL/ampul
 Container Size : 5 mL Pkg Amt: > 5 mL
 Expiration Date : August 31, 2028 Storage: 10°C or colder
 Handling: Sonicate prior to use. Ship: Ambient

Received by
CG on
12/28/22
Storage
to
Silo 10

C E R T I F I E D V A L U E S

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	Nitrobenzene-d5 CAS # 4165-60-0 Purity 99%	5,009.8 μ g/mL	+/- 29.1271 μ g/mL	+/- 225.6421 μ g/mL	+/- 250.3778 μ g/mL
2	2-Fluorobiphenyl CAS # 321-60-8 Purity 99%	5,026.6 μ g/mL	+/- 29.2250 μ g/mL	+/- 226.4003 μ g/mL	+/- 251.2191 μ g/mL
3	p-Terphenyl-d14 CAS # 1718-51-0 Purity 99%	5,027.3 μ g/mL	+/- 29.2289 μ g/mL	+/- 226.4304 μ g/mL	+/- 251.2524 μ g/mL

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

[Signature]
John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 Balance: 1128353505

[Signature]
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com



Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 Lot No.: A0189418
 Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000 μ g/mL, Methylene Chloride, 5mL/ampul
 Container Size : 5 mL Pkg Amt: > 5 mL
 Expiration Date : August 31, 2028 Storage: 10°C or colder
 Handling: Sonicate prior to use. Ship: Ambient

Received by
CG on
12/28/22
Storage
to
Silo 10

C E R T I F I E D V A L U E S

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	Nitrobenzene-d5 CAS # 4165-60-0 Purity 99%	5,009.8 μ g/mL	+/- 29.1271 μ g/mL	+/- 225.6421 μ g/mL	+/- 250.3778 μ g/mL
2	2-Fluorobiphenyl CAS # 321-60-8 Purity 99%	5,026.6 μ g/mL	+/- 29.2250 μ g/mL	+/- 226.4003 μ g/mL	+/- 251.2191 μ g/mL
3	p-Terphenyl-d14 CAS # 1718-51-0 Purity 99%	5,027.3 μ g/mL	+/- 29.2289 μ g/mL	+/- 226.4304 μ g/mL	+/- 251.2524 μ g/mL

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

[Signature]
John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 Balance: 1128353505

[Signature]
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555870 Lot No.: A0194698
Description : Custom 2,4-Dinitrophenol Standard
Custom 2,4-Dinitrophenol Standard 25,000 μ g/mL, Methanol, 1mL/ampul
Container Size : 2 mL Pkg Amt: > 1 mL
Expiration Date : February 28, 2026 Storage: 10°C or colder
Ship: Ambient

C E R T I F I C

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)
1	2,4-Dinitrophenol	51-28-5	DR221221RSR	99%	25,195.0 μ g/mL

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer - Operations Technician |

Date Mixed: 15-Feb-2023 Balance: B442140311

Manufactured under Restek
Registered Quality
Certificate #FM 8

Certified Reference Material Notes

Notes:

The date valid for unopened ampul stored in compliance with the recommended conditions. Purity, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, LC/MS, RI, and/or melting point.

Ampuls with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the compound in solution.

Isomeric compounds are reported as the sum of the isomers.

Values are rounded to the nearest whole number.

Uncertainty Value Notes:

Uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

A coverage factor of 2, which gives a level of confidence of approximately 95%.

Sampled amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Notes:

Preparation is based upon gravimetric preparation using either a balance whose calibration has been verified daily or traceable weights, and/or dilutions with Class A glassware.

The unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration date displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom service. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.

If dissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

REC

O

S

C

Catalog No. : 555869 **Lot No.:** A0194702
Description : Custom Hexachlorocyclopentadiene Standard
 Custom Hexachlorocyclopentadiene Standard 25,000 μ g/mL, Methanol,
 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : February 28, 2026 **Storage:** 10°C or colder
 Ship: Ambient

C E R T I F I E D

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)
1	Hexachlorocyclopentadiene	77-47-4	0012019	99%	25,008.0 μ g/mL

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer - Operations Technician I

Date Mixed: 15-Feb-2023 Balance: B442140311

Manufactured under Restek
Registered Quality
Certificate #FM 1

Certified Reference Material Notes

Notes:

The date valid for unopened ampul stored in compliance with the recommended conditions. The purity, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, LC/MS, RI, and/or melting point.

Ampuls with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the compound in solution.

Isomeric compounds are reported as the sum of the isomers.

Values are rounded to the nearest whole number.

Uncertainty Value Notes:

Uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

A coverage factor of 2, which gives a level of confidence of approximately 95%.

The stated amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Notes:

Preparation is based upon gravimetric preparation using either a balance whose calibration has been verified daily or traceable weights, and/or dilutions with Class A glassware.

:

The unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration date displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with small amounts packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, with complete instructions.

Unsolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreing matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

Recd. by R3 on 7/29/23 [E 3551]

RC-02-01, Ed. 3



Certificate of Analysis

Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:
Pellets

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025
Storage: Room Temperature

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Calcium	<= 0.005 %	<0.005 %	PASS
Chloride	<= 0.005 %	0.002 %	PASS
Heavy Metals	<= 0.002 %	<0.002 %	PASS
Iron	<= 0.001 %	<0.001 %	PASS
Magnesium	<= 0.002 %	<0.002 %	PASS
Mercury	<= 0.1 ppm	<0.1 ppm	PASS
Nickel	<= 0.001 %	<0.001 %	PASS
Nitrogen Compounds	<= 0.001 %	<0.001 %	PASS
Phosphate	<= 0.001 %	<0.001 %	PASS
Potassium	<= 0.02 %	<0.02 %	PASS
Purity	>= 97.0 %	99.2 %	PASS
Sodium Carbonate	<= 1.0 %	0.5 %	PASS
Sulfate	<= 0.003 %	<0.003 %	PASS

Internal ID #: 710

Signature

Additional Information

We certify that this batch conforms to the specifications listed.

Analysis may have been rounded to significant digits in specification limits.

This document has been electronically produced and is valid without a signature.

Product meets analytical specifications of the grades listed.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 24E2462004
Manufactured Date: 2024-04-10
Expiration Date: 2025-07-10
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	3
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	≥ 99.8 %	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Titrable Acid ($\mu\text{eq/g}$)	≤ 0.3	< 0.1
Chloride (Cl)	≤ 10 ppm	5 ppm
Water (by KF, coulometric)	≤ 0.02 %	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC
Manufacturer source batch: MG24D10725

E 3768

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087, U.S.A. Phone 610.386.1700
Page 1 of 1

Acetone

BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03
Batch No.: 23H1462005
Manufactured Date: 2023-07-26
Expiration Date: 2026-07-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	≥ 99.4 %	99.7 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titrable Acid (μeq/g)	≤ 0.3	0.1
Titrable Base (μeq/g)	≤ 0.6	< 0.1
Water (H ₂ O)	≤ 0.5 %	0.3 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	< 1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd. by RP on 8/13/24

E 3788

Ken Koehlein
Ken Koehlein
Sr. Manager, Quality Assurance

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 24G2362009
Manufactured Date: 2024-06-10
Expiration Date: 2025-09-09
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	2
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	≥ 99.8 %	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Titrable Acid ($\mu\text{eq/g}$)	≤ 0.3	< 0.1
Chloride (Cl)	≤ 10 ppm	< 5 ppm
Water (by KF, coulometric)	≤ 0.02 %	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC
Manufacturer source batch: MG24F10024

E3791

A handwritten signature in black ink that appears to read 'Jamie Croak'.

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone 610.386.1700
Page 1 of 1

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis



Material No.: 9254-03
Batch No.: 24H1462005
Manufactured Date: 2024-05-24
Expiration Date: 2027-05-24
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titrable Acid (μeq/g)	<= 0.3	0.2
Titrable Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3815

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087, U.S.A. Phone 610.386.1700

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4

Batch No.: 24I2662006

Manufactured Date: 2024-08-29

Expiration Date: 2025-11-28

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	2
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	3
Assay (CH ₂ Cl ₂) (by GC, exclusive of preservative, corrected for water)	>= 99.8 %	99.9 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Titrable Acid (μeq/g)	<= 0.3	<0.1
Chloride (Cl)	<= 10 ppm	<5 ppm
Water (by KF, coulometric)	<= 0.02 %	<0.01 %

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

E 3822

A handwritten signature in black ink that reads 'Jamie Croak'.

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Hydrochloric Acid, 36.5-38.0%
 BAKER INSTRUMENTS ANALYZED® Reagent
 For Trace Metal Analysis



Material No.: 9530-33
 Batch No.: 0000281827
 Manufactured Date: 2021/03/30
 Retest Date: 2026/03/29
 Revision No.: 1

Certificate of Analysis

Test	Specification	Result
ACS – Assay (as HCl) (by acid-base titrn)	36.5 – 38.0 %	37.6
ACS – Color (APHA)	<= 10	5
ACS – Residue after Ignition	<= 3 ppm	1
ACS – Specific Gravity at 60°/60°F	1.185 – 1.192	1.189
ACS – Bromide (Br)	<= 0.005 %	< 0.005
ACS – Extractable Organic Substances	<= 5 ppm	< 1
ACS – Free Chlorine (as Cl ₂)	<= 0.5 ppm	< 0.5
Phosphate (PO ₄)	<= 0.05 ppm	< 0.03
Sulfate (SO ₄)	<= 0.5 ppm	< 0.3
Sulfite (SO ₃)	<= 0.8 ppm	0.3
Ammonium (NH ₄)	<= 3 ppm	< 1
Trace Impurities – Arsenic (As)	<= 0.010 ppm	< 0.003
Trace Impurities – Aluminum (Al)	<= 10.0 ppb	0.5
Arsenic and Antimony (as As)	<= 5 ppb	< 3
Trace Impurities – Barium (Ba)	<= 1.0 ppb	< 0.2
Trace Impurities – Beryllium (Be)	<= 1.0 ppb	< 0.2
Trace Impurities – Bismuth (Bi)	<= 10.0 ppb	< 1.0
Trace Impurities – Boron (B)	<= 20.0 ppb	< 5.0
Trace Impurities – Cadmium (Cd)	<= 1.0 ppb	< 0.3
Trace Impurities – Calcium (Ca)	<= 50.0 ppb	15.0
Trace Impurities – Chromium (Cr)	<= 1.0 ppb	< 0.4
Trace Impurities – Cobalt (Co)	<= 1.0 ppb	< 0.3
Trace Impurities – Copper (Cu)	<= 1.0 ppb	< 0.1
Trace Impurities – Gallium (Ga)	<= 1.0 ppb	< 0.2

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC
 100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Test	Specification	Result
Trace Impurities – Germanium (Ge)	<= 3.0 ppb	< 2.0
Trace Impurities – Gold (Au)	<= 4.0 ppb	3.0
Heavy Metals (as Pb)	<= 100 ppb	< 50
Trace Impurities – Iron (Fe)	<= 15.0 ppb	1.0
Trace Impurities – Lead (Pb)	<= 1.0 ppb	< 0.5
Trace Impurities – Lithium (Li)	<= 1.0 ppb	< 0.2
Trace Impurities – Magnesium (Mg)	<= 10.0 ppb	< 0.4
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	< 0.4
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	0.2
Trace Impurities – Molybdenum (Mo)	<= 10.0 ppb	< 5.0
Trace Impurities – Nickel (Ni)	<= 4.0 ppb	< 0.3
Trace Impurities – Niobium (Nb)	<= 1.0 ppb	< 0.2
Trace Impurities – Potassium (K)	<= 9.0 ppb	< 2.0
Trace Impurities – Selenium (Se), For Information Only	ppb	1.0
Trace Impurities – Silicon (Si)	<= 100.0 ppb	18.0
Trace Impurities – Silver (Ag)	<= 1.0 ppb	< 0.3
Trace Impurities – Sodium (Na)	<= 100.0 ppb	< 5.0
Trace Impurities – Strontium (Sr)	<= 1.0 ppb	< 0.2
Trace Impurities – Tantalum (Ta)	<= 1.0 ppb	< 0.9
Trace Impurities – Thallium (Tl)	<= 5.0 ppb	< 2.0
Trace Impurities – Tin (Sn)	<= 5.0 ppb	< 0.8
Trace Impurities – Titanium (Ti)	<= 1.0 ppb	< 0.2
Trace Impurities – Vanadium (V)	<= 1.0 ppb	< 0.2
Trace Impurities – Zinc (Zn)	<= 5.0 ppb	0.4
Trace Impurities – Zirconium (Zr)	<= 1.0 ppb	< 0.1

For Laboratory, Research or Manufacturing Use

Product Information (not specifications):

Appearance (clear, fuming liquid)

Meets ACS Specifications

Country of Origin: US

Packaging Site: Phillipsburg Mfg Ctr & DC



Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700



5580 Skylene Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: Lot No.:	Storage:	Solvent:	Exp. Date:	Description:	
Compound		CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
Z-110094-02 506889	≤ -10 °C	Methylene Chloride	7/25/2028	CLP Base/Neutral Surrogate Solution, 5,000 mg/L, 1 ml	
1,2-dichlorobenzene-d ₄		2199-69-1	99.7	247.29.3P	5035 ± 28.02
2-fluorobiphenyl		321-60-8	99.69	8.286.1.1P	4999 ± 103.66
nitrobenzene-d ₅		4165-60-0	99.67	7.9.3P	4988 ± 27.32
p-terphenyl-d ₁₄		1718-51-0	99.3	9.120.8P	5005 ± 27.85

511494 Y.P.
↓ 08/11/2028
S11498

*Not a certified value

Mario Cadeau
Certified By:

Clint Tipton
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis *gravimetric*



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Expiration Date : September 30, 2025

Handling: This product is photosensitive.

Lot No.: A0201940

Pkg Amt: > 1 mL

Storage: 10°C or colder

Ship: Ambient

511539

↓
511568

Y.P.

{ 09/19/

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S230321RSR	99%	1,001.0 μ g/mL	+/- 22.9799
2	Atrazine	1912-24-9	5FYWL	99%	1,010.0 μ g/mL	+/- 23.1865
3	Benzidine	92-87-5	S221205RSR	99%	1,008.0 μ g/mL	+/- 23.1406
4	epsilon-Caprolactam	105-60-2	I16X016	99%	1,008.0 μ g/mL	+/- 23.1406
Solvent:	Methylene chloride					
	CAS #	75-09-2				
	Purity	99%				

Jennifer Pollino
Sam Moodler - Operations Tech I

Date Mixed: 13-Sep-2023 Balance: B345965662

REVIEWED
By Jennifer Pollino at 7:10 am, Sep 13, 2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/pECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis *gravimetric*



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Expiration Date : September 30, 2025

Handling: This product is photosensitive.

Lot No.: A0201940

Pkg Amt: > 1 mL

Storage: 10°C or colder

Ship: Ambient

511539

↓
511568

Y.P.

{ 09/19/

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S230321RSR	99%	1,001.0 μ g/mL	+/- 22.9799
2	Atrazine	1912-24-9	5FYWL	99%	1,010.0 μ g/mL	+/- 23.1865
3	Benzidine	92-87-5	S221205RSR	99%	1,008.0 μ g/mL	+/- 23.1406
4	epsilon-Caprolactam	105-60-2	I16X016	99%	1,008.0 μ g/mL	+/- 23.1406
Solvent:	Methylene chloride					
	CAS #	75-09-2				
	Purity	99%				

Jennifer Pollino
Sam Moodler - Operations Tech I

Date Mixed: 13-Sep-2023 Balance: B345965662

REVIEWED

By Jennifer Pollino at 7:10 am, Sep 13, 2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/pECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.



Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK

CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com**Certificate of Analysis****gravimetric**

ACCREDITED
ISO 17025 Accredited
Reference Materials Production
Certificate #4322-2.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #4322.202

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. :	555872	Lot No.:	A0201728
Description :	Custom Pentachlorophenol Standard		
	Custom Pentachlorophenol Standard 25,000 μ g/mL, Methanol, 1mL/ampul		
Container Size :	2 mL	Pkg Amt:	> 1 mL
Expiration Date :	September 30, 2026	Storage:	10°C or colder
		Ship:	Ambient

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	Pentachlorophenol	87-86-5	RP230530RSR	99%	25,000.0 μ g/mL	+/- 777.0837
Solvent:	Methanol					
	CAS #	67-56-1				
	Purity	99%				

Josh McCluskey - Operations Technician I

Date Mixed: 05-Sep-2023 Balance: B251644995

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853

Lot No.: A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000 μ g/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2028

Storage: 0°C or colder

Ship: Ambient

511749
↓ { RC /
511794 } 11/30/23

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 μ g/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 mL/min.

Inj. Vol

1 μ L



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Samuel Moodier
Sam Moodier - Operations Tech I

Date Mixed: 30-Mar-2023 Balance Serial #: B707717271

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/µECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853

Lot No.: A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000 μ g/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2028

Storage: 0°C or colder

Ship: Ambient

511749
↓ { RC /
511794 } 11/30/23

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 μ g/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 mL/min.

Inj. Vol

1 μ L



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Samuel Moodier
Sam Moodier - Operations Tech I

Date Mixed: 30-Mar-2023 Balance Serial #: B707717271

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/µECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853

Lot No.: A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000 μ g/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2028

Storage: 0°C or colder

Ship: Ambient

511749
↓ { RC /
511794 } 11/30/23

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 μ g/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 mL/min.

Inj. Vol

1 μ L



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Samuel Moodier
Sam Moodier - Operations Tech I

Date Mixed: 30-Mar-2023 Balance Serial #: B707717271

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/µECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853

Lot No.: A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000 μ g/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2028

Storage: 0°C or colder

Ship: Ambient

511749
↓ { RC /
511794 } 11/30/23

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 μ g/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 mL/min.

Inj. Vol

1 μ L



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodier
Sam Moodier - Operations Tech I

Date Mixed: 30-Mar-2023 Balance Serial #: B707717271

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/µECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853

Lot No.: A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000 μ g/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2028

Storage: 0°C or colder

Ship: Ambient

511749
↓ { RC /
511794 } 11/30/23

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 μ g/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 mL/min.

Inj. Vol

1 μ L



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Samuel Moodier
Sam Moodier - Operations Tech I

Date Mixed: 30-Mar-2023 Balance Serial #: B707717271

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/µECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853

Lot No.: A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000 μ g/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2028

Storage: 0°C or colder

Ship: Ambient

511749
↓ { RC /
511794 } 11/30/23

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 μ g/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 mL/min.

Inj. Vol

1 μ L



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodier
Sam Moodier - Operations Tech I

Date Mixed: 30-Mar-2023 Balance Serial #: B707717271

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/µECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



ILAC-MRA
ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ILAC-MRA
ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

Certificate of Analysis *chromatographic plus*

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31206

Lot No.: A0201320

Description : SV Internal Standard Mix 2mg/ml

SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride,
1mL/ampul

S12013 }
↓ } RC
S12042 } 12/26/23

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : July 31, 2029

Storage: 10°C or colder

Handling: Sonication required. Mix is
photosensitive.

Ship: Ambient

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,017.0 µg/mL	+/- 90.8469
2	Naphthalene-d8	1146-65-2	M-2180	99%	2,011.3 µg/mL	+/- 90.5917
3	Acenaphthene-d10	15067-26-2	PR-33507	99%	2,008.6 µg/mL	+/- 90.4685
4	Phenanthrene-d10	1517-22-2	PR-32303	99%	2,019.4 µg/mL	+/- 90.9550
5	Chrysene-d12	1719-03-5	PR-32210	99%	2,013.7 µg/mL	+/- 90.6968
6	Perylene-d12	1520-96-3	PR-33205	99%	2,012.7 µg/mL	+/- 90.6517

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1 μ l



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Peter Robbins - Operations Technician I

Date Mixed: 23-Aug-2023 Balance Serial #: B345965662


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 25-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: Lot No.:	Storage:	Solvent:	Exp. Date:	Description:	
Z-110816-01	414127	Methylene Chloride	6/21/2025	Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL	
Compound		CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
atrazine		1912-24-9	99.5	337.7.3P	997 ± 5.81
benzidine		92-87-5	99.9	124.18.6.2P	991.8 ± 5.77
caprolactam		105-60-2	99.9	271.1.6P	999 ± 5.82

512075 }
↓ } RC
512079 } 02/01/24

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certified By:

Shane Overcash
Chemist



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224

Lot No.: A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : February 28, 2026

Storage: 10°C or colder

Ship: Ambient

S12082
↓
S12111 } RC /
} 02/22/24

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 μ g/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 μ g/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 μ g/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 μ g/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 μ g/mL	+/- 29.571294

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224

Lot No.: A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : February 28, 2026

Storage: 10°C or colder

Ship: Ambient

S12082
↓
S12111 } RC /
} 02/22/24

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 μ g/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 μ g/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 μ g/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 μ g/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 μ g/mL	+/- 29.571294

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224

Lot No.: A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : February 28, 2026

Storage: 10°C or colder

Ship: Ambient

S12082
↓
S12111 } RC /
} 02/22/24

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 μ g/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 μ g/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 μ g/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 μ g/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 μ g/mL	+/- 29.571294

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224

Lot No.: A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : February 28, 2026

Storage: 10°C or colder

Ship: Ambient

S12082
↓
S12111 } RC /
} 02/22/24

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 μ g/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 μ g/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 μ g/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 μ g/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 μ g/mL	+/- 29.571294

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224

Lot No.: A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : February 28, 2026

Storage: 10°C or colder

Ship: Ambient

S12082
↓
S12111 } RC /
02/22/24

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 μ g/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 μ g/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 μ g/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 μ g/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 μ g/mL	+/- 29.571294

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224

Lot No.: A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : February 28, 2026

Storage: 10°C or colder

Ship: Ambient

S12082
↓
S12111 } RC /
02/22/24

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 μ g/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 μ g/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 μ g/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 μ g/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 μ g/mL	+/- 29.571294

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224

Lot No.: A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : February 28, 2026

Storage: 10°C or colder

Ship: Ambient

S12082
↓
S12111 } RC/
} 02/22/24

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 μ g/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 μ g/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 μ g/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 μ g/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 μ g/mL	+/- 29.571294

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224

Lot No.: A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : February 28, 2026

Storage: 10°C or colder

Ship: Ambient

S12082
↓
S12111 } RC /
} 02/22/24

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 μ g/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 μ g/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 μ g/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 μ g/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 μ g/mL	+/- 29.571294

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: Lot No.:	Storage:	Solvent:	Exp. Date:	Description:	
Z-020223-01	454157	≤ -10 °C	P/T Methanol	6/10/2026 1,4-Dioxane Solution, 2000 mg/L, 1 mL	
Compound		CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,4-dioxane		123-91-1	100	223.1.3P	1997 ± 57.08

512112 } RC /
↓
512116 } 03/08/24

*Not a certified value

Certified By:

Melissa Workoff
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850

Lot No.: A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : April 30, 2025

Storage: 0°C or colder

Handling: Sonication required. Mix is photosensitive.

Ship: Ambient

512117 } RC/
↓ } 03/18/24
512146

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/-	36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/-	36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/-	36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/-	36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/-	36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/-	36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/-	36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/-	36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/-	36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/-	36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/-	36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/-	36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/-	36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/-	36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/-	36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/-	36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/-	36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/-	36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/-	36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/-	36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/-	36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/-	36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/-	36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/-	36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/-	36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/-	36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/-	36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/-	36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/-	36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/-	36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/-	36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/-	36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/-	36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/-	36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/-	36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/-	36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850

Lot No.: A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : April 30, 2025

Storage: 0°C or colder

Handling: Sonication required. Mix is photosensitive.

Ship: Ambient

512117 } RC/
↓ } 03/18/24
512146

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/-	36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/-	36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/-	36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/-	36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/-	36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/-	36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/-	36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/-	36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/-	36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/-	36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/-	36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/-	36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/-	36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/-	36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/-	36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/-	36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/-	36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/-	36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/-	36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/-	36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/-	36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/-	36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/-	36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/-	36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/-	36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/-	36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/-	36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/-	36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/-	36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/-	36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/-	36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/-	36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/-	36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/-	36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/-	36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/-	36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850

Lot No.: A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : April 30, 2025

Storage: 0°C or colder

Handling: Sonication required. Mix is photosensitive.

Ship: Ambient

512117 } RC/
↓ } 03/18/24
512146

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/-	36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/-	36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/-	36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/-	36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/-	36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/-	36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/-	36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/-	36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/-	36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/-	36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/-	36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/-	36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/-	36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/-	36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/-	36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/-	36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/-	36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/-	36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/-	36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/-	36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/-	36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/-	36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/-	36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/-	36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/-	36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/-	36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/-	36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/-	36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/-	36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/-	36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/-	36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/-	36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/-	36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/-	36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/-	36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/-	36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850

Lot No.: A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : April 30, 2025

Storage: 0°C or colder

Handling: Sonication required. Mix is photosensitive.

Ship: Ambient

512117 } RC/
↓ } 03/18/24
512146

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/-	36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/-	36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/-	36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/-	36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/-	36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/-	36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/-	36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/-	36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/-	36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/-	36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/-	36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/-	36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/-	36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/-	36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/-	36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/-	36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/-	36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/-	36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/-	36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/-	36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/-	36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/-	36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/-	36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/-	36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/-	36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/-	36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/-	36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/-	36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/-	36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/-	36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/-	36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/-	36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/-	36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/-	36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/-	36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/-	36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850

Lot No.: A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : April 30, 2025

Storage: 0°C or colder

Handling: Sonication required. Mix is photosensitive.

Ship: Ambient

512117 } RC/
↓ } 03/18/24
512146

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/-	36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/-	36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/-	36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/-	36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/-	36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/-	36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/-	36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/-	36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/-	36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/-	36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/-	36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/-	36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/-	36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/-	36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/-	36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/-	36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/-	36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/-	36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/-	36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/-	36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/-	36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/-	36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/-	36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/-	36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/-	36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/-	36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/-	36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/-	36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/-	36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/-	36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/-	36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/-	36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/-	36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/-	36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/-	36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/-	36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850

Lot No.: A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : April 30, 2025

Storage: 0°C or colder

Handling: Sonication required. Mix is photosensitive.

Ship: Ambient

512117 } RC/
↓ } 03/18/24
512146

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/-	36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/-	36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/-	36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/-	36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/-	36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/-	36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/-	36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/-	36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/-	36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/-	36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/-	36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/-	36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/-	36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/-	36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/-	36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/-	36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/-	36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/-	36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/-	36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/-	36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/-	36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/-	36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/-	36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/-	36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/-	36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/-	36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/-	36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/-	36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/-	36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/-	36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/-	36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/-	36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/-	36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/-	36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/-	36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/-	36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850

Lot No.: A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : April 30, 2025

Storage: 0°C or colder

Handling: Sonication required. Mix is photosensitive.

Ship: Ambient

512117 } RC/
↓ } 03/18/24
512146

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/-	36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/-	36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/-	36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/-	36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/-	36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/-	36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/-	36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/-	36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/-	36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/-	36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/-	36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/-	36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/-	36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/-	36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/-	36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/-	36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/-	36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/-	36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/-	36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/-	36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/-	36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/-	36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/-	36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/-	36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/-	36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/-	36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/-	36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/-	36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/-	36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/-	36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/-	36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/-	36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/-	36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/-	36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/-	36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/-	36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850

Lot No.: A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : April 30, 2025

Storage: 0°C or colder

Handling: Sonication required. Mix is photosensitive.

Ship: Ambient

512117 } RC/
↓ } 03/18/24
512146

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/-	36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/-	36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/-	36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/-	36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/-	36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/-	36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/-	36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/-	36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/-	36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/-	36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/-	36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/-	36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/-	36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/-	36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/-	36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/-	36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/-	36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/-	36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/-	36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/-	36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/-	36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/-	36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/-	36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/-	36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/-	36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/-	36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/-	36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/-	36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/-	36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/-	36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/-	36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/-	36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/-	36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/-	36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/-	36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/-	36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850

Lot No.: A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : April 30, 2025

Storage: 0°C or colder

Handling: Sonication required. Mix is photosensitive.

Ship: Ambient

512117 } RC/
↓ } 03/18/24
512146

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/-	36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/-	36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/-	36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/-	36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/-	36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/-	36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/-	36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/-	36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/-	36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/-	36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/-	36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/-	36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/-	36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/-	36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/-	36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/-	36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/-	36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/-	36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/-	36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/-	36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/-	36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/-	36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/-	36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/-	36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/-	36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/-	36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/-	36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/-	36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/-	36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/-	36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/-	36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/-	36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/-	36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/-	36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/-	36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/-	36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850

Lot No.: A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : April 30, 2025

Storage: 0°C or colder

Handling: Sonication required. Mix is photosensitive.

Ship: Ambient

512117 } RC/
↓ } 03/18/24
512146

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/-	36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/-	36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/-	36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/-	36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/-	36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/-	36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/-	36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/-	36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/-	36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/-	36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/-	36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/-	36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/-	36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/-	36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/-	36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/-	36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/-	36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/-	36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/-	36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/-	36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/-	36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/-	36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/-	36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/-	36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/-	36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/-	36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/-	36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/-	36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/-	36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/-	36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/-	36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/-	36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/-	36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/-	36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/-	36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/-	36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



ILAC
ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ILAC
ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

Certificate of Analysis *chromatographic plus*

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087

Lot No.: A0206206

512187
↓
512206 } RC/
03/18/24

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000µg/mL, Methanol, 5mL/ampul

Container Size : 5 mL

Pkg Amt: > 5 mL

Expiration Date : January 31, 2032

Storage: 10°C or colder

Ship: Ambient

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Fluorophenol	367-12-4	STBK1705	99%	10,005.3 µg/mL	+/- 302.5390
2	Phenol-d6	13127-88-3	PR-33287A	99%	10,005.5 µg/mL	+/- 302.5475
3	2,4,6-Tribromophenol	118-79-6	RP230831RSR	99%	10,006.6 µg/mL	+/- 302.5783

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol

CAS # 67-56-1

Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

2 mL/min.

Inj. Vol

1 μ L



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Regin - Operations Tech |

Date Mixed: 04-Jan-2024 Balance Serial #: 1128360905

Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 08-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



ILAC
ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ILAC
ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

Certificate of Analysis *chromatographic plus*

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087

Lot No.: A0206206

512187 } RC/
↓ } 03/18/24
512206 }

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000 μ g/mL, Methanol, 5mL/ampul

Container Size : 5 mL

Pkg Amt: > 5 mL

Expiration Date : January 31, 2032

Storage: 10°C or colder

Ship: Ambient

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Fluorophenol	367-12-4	STBK1705	99%	10,005.3 μ g/mL	+/- 302.5390
2	Phenol-d6	13127-88-3	PR-33287A	99%	10,005.5 μ g/mL	+/- 302.5475
3	2,4,6-Tribromophenol	118-79-6	RP230831RSR	99%	10,006.6 μ g/mL	+/- 302.5783

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol

CAS # 67-56-1

Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

2 mL/min.

Inj. Vol

1 μ L



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Regin - Operations Tech |

Date Mixed: 04-Jan-2024 Balance Serial #: 1128360905

Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 08-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087

Lot No.: A0206206

512187 } RC/
↓ } 03/18/24
512206 }

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000 μ g/mL, Methanol, 5mL/ampul

Container Size : 5 mL

Pkg Amt: > 5 mL

Expiration Date : January 31, 2032

Storage: 10°C or colder

Ship: Ambient

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Fluorophenol	367-12-4	STBK1705	99%	10,005.3 μ g/mL	+/- 302.5390
2	Phenol-d6	13127-88-3	PR-33287A	99%	10,005.5 μ g/mL	+/- 302.5475
3	2,4,6-Tribromophenol	118-79-6	RP230831RSR	99%	10,006.6 μ g/mL	+/- 302.5783

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol

CAS # 67-56-1

Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

2 mL/min.

Inj. Vol

1 μ L



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Regin - Operations Tech |

Date Mixed: 04-Jan-2024 Balance Serial #: 1128360905

Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 08-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



Certificate of Analysis *chromatographic plus*

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 **Lot No.:** A0206381
Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000 μ g/mL, Methylene Chloride, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : December 31, 2029 **Storage:** 10°C or colder
Handling: Sonicate prior to use. **Ship:** Ambient

S12207 } RC /
↓ } 03/18/24
S12221 }

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Nitrobenzene-d5	4165-60-0	I-25158	99%	5,029.3 μ g/mL	+/- 226.5204
2	2-Fluorobiphenyl	321-60-8	00021384	99%	5,030.9 μ g/mL	+/- 226.5936
3	p-Terphenyl-d14	1718-51-0	PR-32599	99%	5,026.4 μ g/mL	+/- 226.3909

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

2 mL/min.

Inj. Vol

1 μ L



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Jess Hoy - Operations Tech I

Date Mixed: 09-Jan-2024 Balance Serial #: 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 11-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



5580 Skylan Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: Lot No.: Storage: Solvent: Exp. Date: Description:

Z-010442-07 495833 ≤ -10 °C Methylene Chloride 1/16/2028 Benzaldehyde Solution, 1000 mg/L, 1.3 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
benzaldehyde	100-52-7	98.3	442.421.1P	996.8 ± 11.49

S12275 } RC/
↓ } 05/24/24
S12279 }

*Not a certified value

Certified By: _____

Scott Hunter
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

Certificate of Analysis *chromatographic plus*

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 31206

Lot No.: A0206540

Description: SV Internal Standard Mix 2mg/ml

SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride,
1mL/ampul

Container Size: 2 mL

Pkg Amt: > 1 mL

Expiration Date: December 31, 2029

Storage: 10°C or colder

Handling: Sonication required. Mix is
photosensitive.

Ship: Ambient

512312 } RC/
↓ } 05/30/24
512331 }

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,007.1 µg/mL	+/- 90.4025
2	Naphthalene-d8	1146-65-2	M-2180	99%	2,005.9 µg/mL	+/- 90.3454
3	Acenaphthene-d10	15067-26-2	PR-33507	99%	2,007.9 µg/mL	+/- 90.4385
4	Phenanthrene-d10	1517-22-2	PR-32303	99%	2,006.7 µg/mL	+/- 90.3845
5	Chrysene-d12	1719-03-5	PR-32210	99%	2,015.5 µg/mL	+/- 90.7778
6	Perylene-d12	1520-96-3	PR-33205	99%	2,014.7 µg/mL	+/- 90.7448

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1 μ l



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Malina Homan
Malina Homan - Operations Technician |

Date Mixed: 12-Jan-2024 Balance Serial #: 1128360905

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 31206

Lot No.: A0206540

Description: SV Internal Standard Mix 2mg/ml
SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride,
1mL/ampul

Container Size: 2 mL

Expiration Date: December 31, 2029

Handling: Sonication required. Mix is photosensitive.

Pkg Amt: > 1 mL

Storage: 10°C or colder

Ship: Ambient

S12312 } RC /
↓ } 05/30/24
S12331 }

C E R T I F I E D V A L U E S

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,007.1 µg/mL	+/- 90.4025
2	Naphthalene-d8	1146-65-2	M-2180	99%	2,005.9 µg/mL	+/- 90.3454
3	Acenaphthene-d10	15067-26-2	PR-33507	99%	2,007.9 µg/mL	+/- 90.4385
4	Phenanthrene-d10	1517-22-2	PR-32303	99%	2,006.7 µg/mL	+/- 90.3845
5	Chrysene-d12	1719-03-5	PR-32210	99%	2,015.5 µg/mL	+/- 90.7778
6	Perylene-d12	1520-96-3	PR-33205	99%	2,014.7 µg/mL	+/- 90.7448

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25 μ m
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1 μ l



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Malina Homan
Malina Homan - Operations Technician |

Date Mixed: 12-Jan-2024 Balance Serial #: 1128360905

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 μ g/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 μ g/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 μ g/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 μ g/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓ } 7/24/24
S12508 }

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 μ g/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 μ g/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 μ g/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 μ g/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓ } 7/24/24
S12508 }

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 μ g/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 μ g/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 μ g/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 μ g/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓ } 7/24/24
S12508 }

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 μ g/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 μ g/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 μ g/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 μ g/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓ } 7/24/24
S12508 }

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 μ g/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 μ g/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 μ g/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 μ g/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓ } 7/24/24
S12508 }

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 μ g/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 μ g/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 μ g/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 μ g/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓ } 7/24/24
S12508 }

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 μ g/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 μ g/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 μ g/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 μ g/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓ } 7/24/24
S12508 }

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 μ g/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 μ g/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 μ g/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 μ g/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓ } 7/24/24
S12508 }

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



ILAC
ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

Certificate of Analysis *gravimetric*



ILAC
ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000 μ g/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 μ g/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 μ g/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 μ g/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 μ g/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓ } 7/24/24
S12508 }

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



ILAC
ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ILAC
ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

512509
↓
512568 } RC /
7/24/24

Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024 Balance: 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL



ILAC
ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01



ILAC
ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

C E R T I F I E D V A L U E S

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

512509
↓
512568 } RC / 7/24/24

Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024 Balance: 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{\text{combined uncertainty}} = k \sqrt{u_{\text{gravimetric}}^2 + u_{\text{homogeneity}}^2 + u_{\text{storage stability}}^2 + u_{\text{shipping stability}}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



SHIPPING DOCUMENTS

CLIENT INFORMATION

CLIENT PROJECT INFORMATION

CLIENT BILLING INFORMATION

REPORT TO BE SENT TO:

COMPANY: GANNETT FLEMING

ADDRESS: 1010 Adams Avenue

CITY Audobon STATE: PA ZIP: 19403

ATTENTION: Joe Kupansky

PHONE: 610-301-8342 FAX:

PROJECT NAME: Am Liab's top. of SB

PROJECT NO: 9500000518 LOCATION: Kearny, NJ

PROJECT MANAGER: Joe Kupansky

e-mail: DAAC@bennys.com

PHONE: 610-301-8342 FAX:

BILL TO: Chemtech

PO#:

ADDRESS: 284 Sheffield St.

CITY Mountainside STATE: NJ ZIP: 07042

ATTENTION: Samantha PHONE: 908-788-2981

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*

HARDCOPY (DATA PACKAGE): 10 DAYS*

EDD: 10 DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

- Level 1 (Results Only) Level 4 (QC + Full Raw Data)
 Level 2 (Results + QC) NJ Reduced US EPA CLP
 Level 3 (Results + QC) NYS ASP A NYS ASP B
+ Raw Data Other _____
 EDD FORMAT BEN EDD

1 VOC 10 PAH 1 PCB 1 SVOC 1 TAL Metals 1 EPH 1 TCAP - FULL
2 PCB 2 SVOC 2 TAL Metals 2 EPH 2 TCAP - FULL
3 SVOC 3 TAL Metals 3 EPH 3 TCAP - FULL
4 TAL Metals 4 EPH 4 TCAP - FULL
5 EPH 5 TCAP - FULL
6 TCAP - FULL
7 TCAP - FULL
8 TCAP - FULL
9 TCAP - FULL

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS			
			CMP	GRAB	DATE	TIME		A	1	2	3	4	5	6	7	8	9	← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER		
1.	NB-305-SW	GW	X		10/25	8:20	7	X	X	X	X	X	X							
2.	NB-305-SW-1	GW	X		10/25	8:30	7	X	X	X	X	X	X							
3.	NB-305-Top	S	X		10/25	9:20	11	X	X	X	X	X	X							
4.	NB-305-Top-1	S	X		10/25	9:30	11	X	X	X	X	X	X							
5.	NB-305-Bot	S	X		10/25	11:15	11	X	X	X	X	X	X	X	X	X	X	X	X	
6.	NB-305-Bot-1	S	X		10/25	11:20	11	X	X	X	X	X	X	X	X	X	X	X	X	
7.	TB-10252024	D/W			10/25		2	X												
8.																				
9.																				
10.																				

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:

1. *Vaunda Pujara*

DATE/TIME:

10/25/24 3:11

RECEIVED BY:

J. J.

RELINQUISHED BY SAMPLER:

2. *J. J.*

DATE/TIME:

10.25.2024

RECEIVED BY:

J. J.

RELINQUISHED BY SAMPLER:

3. *J. J.*

DATE/TIME:

10.25.2024

RECEIVED BY:

*J. J.*Conditions of bottles or coolers at receipt: COMPLIANT NON COMPLIANT COOLER TEMP

3. 16 °C

Comments:

Page ____ of ____

CLIENT: Hand Delivered Other _____
CHEMTECH: Picked Up Field Sampling

Shipment Complete

 YES NO



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (L-A-B)	L2219
Maine	2024021
Maryland	296
New Hampshire	255423
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488



284 Sheffield Street, Mountainside, New Jersey, 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID : P4578	PORT06	Order Date : 10/25/2024 3:58:00 PM	Project Mgr :
Client Name : Portal Partners Tri-Venture		Project Name : Amtrak Sawtooth Bridges 2	Report Type : NJ Reduced
Client Contact : Joseph Krupansky		Receive DateTime : 10/25/2024 4:04:00 PM	EDD Type : EXCEL NJCLEANUP
Invoice Name : Portal Partners Tri-Venture		Purchase Order :	Hard Copy Date :
Invoice Contact : Joseph Krupansky			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUe DATES
P4578-01	WB-305-SW	Water	10/25/2024	08:20	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
P4578-02	WB-305-SW-1	Water	10/25/2024	08:30	VOC-TCLVOA-10		8260-Low	10 Bus. Days	
P4578-03	WB-305-TOP	Solid	10/25/2024	09:20	VOC-TCLVOA-10		8260D	10 Bus. Days	
P4578-04	WB-305-TOP-1	Solid	10/25/2024	09:30	VOC-TCLVOA-10		8260D	10 Bus. Days	
P4578-05	WB-305-BOT	Solid	10/25/2024	11:15	VOC-TCLVOA-10		8260D	10 Bus. Days	
P4578-07	WB-305-BOT-1	Solid	10/25/2024	11:20	VOC-TCLVOA-10		8260D	10 Bus. Days	
P4578-09	TB-10252024	Water	10/25/2024	00:00	VOC-TCLVOA-10		8260-Low	10 Bus. Days	

LOGIN REPORT/SAMPLE TRANSFER

Order ID : P4578	PORT06	Order Date : 10/25/2024 3:58:00 PM	Project Mgr :
Client Name : Portal Partners Tri-Venture		Project Name : Amtrak Sawtooth Bridges 2	Report Type : NJ Reduced
Client Contact : Joseph Krupansky		Receive DateTime : 10/25/2024 4:04:00 PM	EDD Type : EXCEL NJCLEANUP
Invoice Name : Portal Partners Tri-Venture		Purchase Order :	Hard Copy Date :
Invoice Contact : Joseph Krupansky			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES

* J.C samples are in sm refrigerator &

Relinquished By :

J.C-

Date / Time : 10-25-2024 1710

Received By :

Sam

Date / Time :

10/25/2024 17:10

rg 6
rg 2
rg 5

Storage Area : VOA Refrigerator Room