284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Cover Page

Order ID : P4578

Project ID : Amtrak Sawtooth Bridges 2024

Client : Portal Partners Tri-Venture

Lab Sample Number	Client Sample Number
P4578-01	WB-305-SW
P4578-02	WB-305-SW-1
P4578-03	WB-305-TOP
P4578-04	WB-305-TOP-1
P4578-05	WB-305-BOT
P4578-06	WB-305-BOT
P4578-07	WB-305-BOT-1
P4578-08	WB-305-BOT-1
P4578-09	TB-10252024

I certify that the data package is in compliance with the terms and conditions of the contract,both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 10/30/2024

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Portal Partners Tri-Venture
Project Name: Amtrak Sawtooth Bridges 2024
Project # N/A
Chemtech Project # P4578
Test Name: TCLP BNA

A. Number of Samples and Date of Receipt:
6 Solid samples were received on 10/25/2024.
3 Water samples were received on 10/25/2024.

B. Parameters
According to the Chain of Custody document, the following analyses were requested: Corrosivity, EPH, EPH, Hexavalent Chromium, Ignitability, Mercury, Metals ICP-TAL, METALS-TAL, PCB, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SVOC-TCL BNA -20, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL, Trivalent Chromium, VOC-TCLVOA-10 and VOC-TCLVOA-10. This data package contains results for TCLP BNA.

C. Analytical Techniques:
The samples were analyzed on instrument BNA_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um df The analysis of TCLP BNA was based on method 8270E and extraction was done based on method 3510 and TCLP extraction method was 1311.

D. QA/ QC Samples:
The Holding Times were met for all analysis.
The Surrogate recoveries met the acceptable criteria.
The Internal Standards Areas met the acceptable requirements.
The Retention Times were acceptable for all samples.
The MS recoveries met the requirements for all compounds .
The MSD recoveries met the acceptable requirements .
The RPD met criteria .
The Blank Spike met requirements for all samples .
The Blank analysis did not indicate the presence of lab contamination.
The Initial Calibration met the requirements .
The Continuous Calibration met the requirements .
The Tuning criteria met requirements.

284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

E. Additional Comments:

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

284 Sheffield Street, Mountainside NJ 07092 Tel. 908-789-8900 Fax: 908-789-8922

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following " Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value**.** This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This is flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: **P4578**

	Completed
For thorough review, the report must have the following:	
GENERAL:	
Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)	✓
Check chain-of-custody for proper relinquish/return of samples	✓
Is the chain of custody signed and complete	✓
Check internal chain-of-custody for proper relinquish/return of samples /sample extracts	✓
Collect information for each project id from server. Were all requirements followed	✓
COVER PAGE:	
Do numbers of samples correspond to the number of samples in the Chain of Custody on login page	✓
Do lab numbers and client Ids on cover page agree with the Chain of Custody	✓
CHAIN OF CUSTODY:	
Do requested analyses on Chain of Custody agree with form I results	✓
Do requested analyses on Chain of Custody agree with the log-in page	✓
Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody	✓
Were the samples received within hold time	✓
Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle	✓
ANALYTICAL:	
Was method requirement followed?	✓
Was client requirement followed?	✓
Does the case narrative summarize all QC failure?	✓
All runlogs and manual integration are reviewed for requirements	✓
All manual calculations and /or hand notations verified	✓

QA Review Signature: **MAYUR DESAI** **Date:** **10/30/2024**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LAB CHRONICLE

OrderID:	P4578	**OrderDate:**	10/25/2024 3:58:00 PM
Client:	Portal Partners Tri-Venture	**Project:**	Amtrak Sawtooth Bridges 2024
Contact:	Joseph Krupansky	**Location:**	K63,VOA Ref. #2 Soil,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P4578-06	**WB-305-BOT**	**TCLP**			**10/25/24**			**10/25/24**
			TCLP BNA	8270E		10/26/24	10/28/24	
P4578-08	**WB-305-BOT-1**	**TCLP**			**10/25/24**			**10/25/24**
			TCLP BNA	8270E		10/26/24	10/28/24	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: P4578

Client: Portal Partners Tri-Venture

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :								
				0.000				
			Total Svoc :			**0.00**		
			Total Concentration:			**0.00**		



QC SUMMARY

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SW-846

SDG No.: P4578

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%) Low	Limits (%) High
P4545-02MS	VNJ-215MS	2-Fluorophenol	150	135	90		15 (10)	110 (139)
		Phenol-d6	150	129	86		15 (10)	110 (134)
		Nitrobenzene-d5	100	103	103		30 (49)	130 (133)
		2-Fluorobiphenyl	100	98.8	99		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	159	106		15 (44)	110 (137)
		Terphenyl-d14	100	105	105		30 (48)	130 (125)
P4545-02MSD	VNJ-215MSD	2-Fluorophenol	150	136	91		15 (10)	110 (139)
		Phenol-d6	150	132	88		15 (10)	110 (134)
		Nitrobenzene-d5	100	105	105		30 (49)	130 (133)
		2-Fluorobiphenyl	100	101	101		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	162	108		15 (44)	110 (137)
		Terphenyl-d14	100	105	105		30 (48)	130 (125)
P4578-06	WB-305-BOT	2-Fluorophenol	150	130	87		15 (10)	110 (139)
		Phenol-d6	150	120	80		15 (10)	110 (134)
		Nitrobenzene-d5	100	105	105		30 (49)	130 (133)
		2-Fluorobiphenyl	100	101	101		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	159	106		15 (44)	110 (137)
		Terphenyl-d14	100	111	111		30 (48)	130 (125)
P4578-08	WB-305-BOT-1	2-Fluorophenol	150	121	81		15 (10)	110 (139)
		Phenol-d6	150	112	75		15 (10)	110 (134)
		Nitrobenzene-d5	100	96.4	96		30 (49)	130 (133)
		2-Fluorobiphenyl	100	93.8	94		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	147	98		15 (44)	110 (137)
		Terphenyl-d14	100	102	102		30 (48)	130 (125)
PB164393TB	PB164393TB	2-Fluorophenol	150	122	82		15 (10)	110 (139)
		Phenol-d6	150	119	79		15 (10)	110 (134)
		Nitrobenzene-d5	100	92.2	92		30 (49)	130 (133)
		2-Fluorobiphenyl	100	87.9	88		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	137	91		15 (44)	110 (137)
		Terphenyl-d14	100	88.8	89		30 (48)	130 (125)
PB164444BL	PB164444BL	2-Fluorophenol	150	120	80		15 (10)	110 (139)
		Phenol-d6	150	114	76		15 (10)	110 (134)
		Nitrobenzene-d5	100	90.9	91		30 (49)	130 (133)
		2-Fluorobiphenyl	100	87.1	87		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	133	89		15 (44)	110 (137)
		Terphenyl-d14	100	89.1	89		30 (48)	130 (125)
PB164444BS	PB164444BS	2-Fluorophenol	150	120	80		15 (10)	110 (139)
		Phenol-d6	150	118	79		15 (10)	110 (134)
		Nitrobenzene-d5	100	92.6	93		30 (49)	130 (133)
		2-Fluorobiphenyl	100	87.2	87		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	141	94		15 (44)	110 (137)
		Terphenyl-d14	100	95.7	96		30 (48)	130 (125)

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4578

Client: Portal Partners Tri-Venture

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Limits Low	Limits High	RPD
Lab Sample ID: P4545-02MS		**Client Sample ID: VNJ-215MS**						**DataFile: BF140092.D**			
Pyridine	500	0	230	ug/L	46				20 (10)	160 (109)	
1,4-Dichlorobenzene	500	0	430	ug/L	86				70 (55)	130 (125)	
2-Methylphenol	500	0	490	ug/L	98				70 (37)	130 (126)	
3+4-Methylphenols	500	0	480	ug/L	96				20 (31)	160 (127)	
Hexachloroethane	500	0	420	ug/L	84				20 (49)	160 (110)	
Nitrobenzene	500	0	470	ug/L	94				70 (62)	130 (112)	
Hexachlorobutadiene	500	0	460	ug/L	92				70 (52)	130 (125)	
2,4,6-Trichlorophenol	500	0	550	ug/L	110				70 (78)	130 (112)	
2,4,5-Trichlorophenol	500	0	540	ug/L	108				70 (71)	130 (111)	
2,4-Dinitrotoluene	500	0	600	ug/L	120				70 (50)	130 (142)	
Hexachlorobenzene	500	0	540	ug/L	108				70 (72)	130 (115)	
Pentachlorophenol	1000	0	780	ug/L	78				20 (25)	160 (139)	

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4578

Client: Portal Partners Tri-Venture

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Limits Low	Limits High	RPD
Lab Sample ID: P4545-02MSD		**Client Sample ID: VNJ-215MSD**						**DataFile: BF140093.D**			
Pyridine	500	0	250	ug/L	50		8		20 (10)	160 (109)	20 (20)
1,4-Dichlorobenzene	500	0	440	ug/L	88		2		70 (55)	130 (125)	20 (20)
2-Methylphenol	500	0	500	ug/L	100		2		70 (37)	130 (126)	20 (20)
3+4-Methylphenols	500	0	500	ug/L	100		4		20 (31)	160 (127)	20 (20)
Hexachloroethane	500	0	430	ug/L	86		2		20 (49)	160 (110)	20 (20)
Nitrobenzene	500	0	470	ug/L	94		0		70 (62)	130 (112)	20 (20)
Hexachlorobutadiene	500	0	460	ug/L	92		0		70 (52)	130 (125)	20 (20)
2,4,6-Trichlorophenol	500	0	560	ug/L	112		2		70 (78)	130 (112)	20 (20)
2,4,5-Trichlorophenol	500	0	540	ug/L	108		0		70 (71)	130 (111)	20 (20)
2,4-Dinitrotoluene	500	0	610	ug/L	122		2		70 (50)	130 (142)	20 (20)
Hexachlorobenzene	500	0	560	ug/L	112		4		70 (72)	130 (115)	20 (20)
Pentachlorophenol	1000	0	850	ug/L	85		9		20 (25)	160 (139)	20 (20)

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4578

Client: Portal Partners Tri-Venture

Analytical Method: 8270E **DataFile:** BF140077.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD Qual	Limits Low	Limits High	Limits RPD
PB164444BS	Pyridine	50	38.6	ug/L	77				20 (29)	160 (97)	
	1,4-Dichlorobenzene	50	43.7	ug/L	87				70 (76)	130 (103)	
	2-Methylphenol	50	44.4	ug/L	89				70 (69)	130 (109)	
	3+4-Methylphenols	50	43.0	ug/L	86				20 (67)	160 (106)	
	Hexachloroethane	50	44.2	ug/L	88				20 (76)	160 (118)	
	Nitrobenzene	50	43.5	ug/L	87				70 (58)	130 (106)	
	Hexachlorobutadiene	50	44.9	ug/L	90				70 (69)	130 (101)	
	2,4,6-Trichlorophenol	50	45.9	ug/L	92				70 (61)	130 (110)	
	2,4,5-Trichlorophenol	50	44.6	ug/L	89				70 (70)	130 (106)	
	2,4-Dinitrotoluene	50	52.4	ug/L	105				70 (60)	130 (115)	
	Hexachlorobenzene	50	45.0	ug/L	90				70 (73)	130 (106)	
	Pentachlorophenol	100	89.7	ug/L	90				20 (47)	160 (114)	

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB164444BL

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

Lab File ID: BF140076.D Lab Sample ID: PB164444BL

Instrument ID: BNA_F Date Extracted: 10/26/2024

Matrix: (soil/water) water Date Analyzed: 10/28/2024

Level: (low/med) LOW Time Analyzed: 10:01

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB164393TB	PB164393TB	BF140078.D	10/28/2024
WB-305-BOT	P4578-06	BF140084.D	10/28/2024
WB-305-BOT-1	P4578-08	BF140085.D	10/28/2024
PB164444BS	PB164444BS	BF140077.D	10/28/2024
VNJ-215MS	P4545-02MS	BF140092.D	10/28/2024
VNJ-215MSD	P4545-02MSD	BF140093.D	10/28/2024

COMMENTS:

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM SAS No.: P4578 SDG NO.: P4578

Lab File ID: BF139843.D DFTPP Injection Date: 10/18/2024

Instrument ID: BNA_F DFTPP Injection Time: 09:22

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	43.8
68	Less than 2.0% of mass 69	0.7 (1.9) 1
69	Mass 69 relative abundance	37.8
70	Less than 2.0% of mass 69	0.3 (0.7) 1
127	10.0 - 80.0% of mass 198	46.9
197	Less than 2.0% of mass 198	0.8
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	26.9
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	14
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	17.9 (19.2) 2

1-Value is % mass 69 2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF139844.D	10/18/2024	10:27
SSTDICC005	SSTDICC005	BF139845.D	10/18/2024	10:55
SSTDICC010	SSTDICC010	BF139846.D	10/18/2024	11:23
SSTDICC020	SSTDICC020	BF139847.D	10/18/2024	11:52
SSTDICCC040	SSTDICCC040	BF139848.D	10/18/2024	12:20
SSTDICC050	SSTDICC050	BF139849.D	10/18/2024	12:49
SSTDICC060	SSTDICC060	BF139850.D	10/18/2024	13:17
SSTDICC080	SSTDICC080	BF139851.D	10/18/2024	13:46

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM SAS No.: P4578 SDG NO.: P4578

Lab File ID: BF140074.D DFTPP Injection Date: 10/28/2024

Instrument ID: BNA_F DFTPP Injection Time: 09:04

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	41.1
68	Less than 2.0% of mass 69	0.6 (1.6) 1
69	Mass 69 relative abundance	37.2
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	46.3
197	Less than 2.0% of mass 198	0.8
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.5
275	10.0 - 60.0% of mass 198	27.6
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	15.5
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.8 (18.8) 2

1-Value is % mass 69 2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF140075.D	10/28/2024	09:32
PB164444BL	PB164444BL	BF140076.D	10/28/2024	10:01
PB164444BS	PB164444BS	BF140077.D	10/28/2024	10:29
PB164393TB	PB164393TB	BF140078.D	10/28/2024	10:58
WB-305-BOT	P4578-06	BF140084.D	10/28/2024	13:51
WB-305-BOT-1	P4578-08	BF140085.D	10/28/2024	14:19
VNJ-215MS	P4545-02MS	BF140092.D	10/28/2024	17:35
VNJ-215MSD	P4545-02MSD	BF140093.D	10/28/2024	18:03

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/28/2024

Lab File ID: BF140075.D Time Analyzed: 09:32

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

		IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
	12 HOUR STD	132517	6.892	497360	8.18	274065	9.93
	UPPER LIMIT	265034	7.392	994720	8.675	548130	10.427
	LOWER LIMIT	66258.5	6.392	248680	7.675	137033	9.427
	EPA SAMPLE NO.						
01	PB164444BL	132330	6.89	518019	8.17	298114	9.92
02	PB164444BS	127868	6.89	495282	8.18	281323	9.93
03	PB164393TB	128211	6.89	508163	8.17	288600	9.92
04	VNJ-215MS	113612	6.89	441889	8.17	241647	9.93
05	VNJ-215MSD	111820	6.89	432287	8.17	236921	9.93
06	WB-305-BOT	112580	6.89	438516	8.17	246067	9.92
07	WB-305-BOT-1	111257	6.89	441878	8.17	245603	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4
IS2 (NPT) = Naphthalene-d8
IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/28/2024

Lab File ID: BF140075.D Time Analyzed: 09:32

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

		IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
	12 HOUR STD	494703	11.416	250901	14.051	263501	15.527
	UPPER LIMIT	989406	11.916	501802	14.551	527002	16.027
	LOWER LIMIT	247352	10.916	125451	13.551	131751	15.027
	EPA SAMPLE NO.						
01	PB164444BL	550598	11.41	341644	14.05	261224	15.53
02	PB164444BS	511374	11.42	262593	14.05	265684	15.53
03	PB164393TB	536321	11.41	334195	14.05	261975	15.53
04	VNJ-215MS	409717	11.41	203129	14.05	263019	15.53
05	VNJ-215MSD	398949	11.41	204155	14.05	256308	15.53
06	WB-305-BOT	437262	11.41	219498	14.05	258448	15.53
07	WB-305-BOT-1	437798	11.41	215751	14.05	230540	15.53

IS4 (PHN) = Phenanthrene-d10
IS5 (CRY) = Chrysene-d12
IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.



SAMPLE DATA

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	10/25/24
Project:	Amtrak Sawtooth Bridges 2024	Date Received:	10/25/24
Client Sample ID:	WB-305-BOT	SDG No.:	P4578
Lab Sample ID:	P4578-06	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140084.D	1	10/26/24 12:10	10/28/24 13:51	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	15.5	U	15.5	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	8.40	U	8.40	50.0	ug/L
95-48-7	2-Methylphenol	11.3	U	11.3	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.5	U	11.5	100	ug/L
67-72-1	Hexachloroethane	10.1	U	10.1	50.0	ug/L
98-95-3	Nitrobenzene	12.7	U	12.7	50.0	ug/L
87-68-3	Hexachlorobutadiene	12.7	U	12.7	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	8.90	U	8.90	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	10.1	U	10.1	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	15.2	U	15.2	50.0	ug/L
118-74-1	Hexachlorobenzene	11.4	U	11.4	50.0	ug/L
87-86-5	Pentachlorophenol	18.5	U	18.5	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	130		15 (10) - 110 (139)	87%	SPK: 150
13127-88-3	Phenol-d6	120		15 (10) - 110 (134)	80%	SPK: 150
4165-60-0	Nitrobenzene-d5	105		30 (49) - 130 (133)	105%	SPK: 100
321-60-8	2-Fluorobiphenyl	101		30 (52) - 130 (132)	101%	SPK: 100
118-79-6	2,4,6-Tribromophenol	159		15 (44) - 110 (137)	106%	SPK: 150
1718-51-0	Terphenyl-d14	111		30 (48) - 130 (125)	111%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	113000	6.887			
1146-65-2	Naphthalene-d8	439000	8.169			
15067-26-2	Acenaphthene-d10	246000	9.922			
1517-22-2	Phenanthrene-d10	437000	11.41			
1719-03-5	Chrysene-d12	219000	14.045			
1520-96-3	Perylene-d12	258000	15.527			

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	10/25/24
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	10/25/24
Client Sample ID:	WB-305-BOT			SDG No.:	P4578
Lab Sample ID:	P4578-06			Matrix:	TCLP
Analytical Method:	SW8270			% Solid:	0
Sample Wt/Vol:	100	Units:	mL	Final Vol:	1000 uL
Soil Aliquot Vol:			uL	Test:	TCLP BNA
Extraction Type :			Decanted : N	Level :	LOW
Injection Volume :			GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140084.D	1	10/26/24 12:10	10/28/24 13:51	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140084.D
Acq On : 28 Oct 2024 13:51
Operator : RC/JU
Sample : P4578-06
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WB-305-BOT

Quant Time: Oct 28 14:22:20 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.887	152	112580	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	438516	20.000	ng	0.00
39) Acenaphthene-d10	9.922	164	246067	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	437262	20.000	ng	0.00
76) Chrysene-d12	14.045	240	219498	20.000	ng	0.00
86) Perylene-d12	15.527	264	258448	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	938458	130.498	ng	0.02
7) Phenol-d6	6.510	99	1115782	119.796	ng	0.00
23) Nitrobenzene-d5	7.445	82	834614	105.486	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	367109	159.499	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1500601	100.787	ng	0.00
79) Terphenyl-d14	12.998	244	1489277	110.559	ng	0.00
Target Compounds						Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140084.D
Acq On : 28 Oct 2024 13:51
Operator : RC/JU
Sample : P4578-06
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WB-305-BOT

Quant Time: Oct 28 14:22:20 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.887 min Scan# 81
Delta R.T. -0.005 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51

Instrument :
BNA_F
ClientSampleId :
WB-305-BOT

Tgt Ion:152 Resp: 112580

Ion	Ratio	Lower	Upper
152	100		
150	159.0	130.2	195.2
115	65.7	51.4	77.2







#5
2-Fluorophenol
Concen: 130.498 ng
RT: 5.516 min Scan# 582
Delta R.T. 0.018 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51

Tgt Ion:112 Resp: 938458

Ion	Ratio	Lower	Upper
112	100		
64	65.4	53.0	79.6
63	33.9	27.0	40.4







#7
Phenol-d6
Concen: 119.796 ng
RT: 6.510 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51

Instrument :
BNA_F
ClientSampleId :
WB-305-BOT

Tgt Ion: 99 Resp: 1115782

Ion	Ratio	Lower	Upper
99	100		
42	19.6	16.7	25.1
71	35.6	27.7	41.5







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.169 min Scan# 1033
Delta R.T. -0.006 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51

Tgt Ion:136 Resp: 438516

Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.6	12.8
54	9.6	8.4	12.6
68	6.0	5.1	7.7







#23
Nitrobenzene-d5
Concen: 105.486 ng
RT: 7.445 min Scan# 91
Delta R.T. -0.006 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51

Instrument :
BNA_F
ClientSampleId :
WB-305-BOT

Tgt Ion: 82 Resp: 834614

Ion	Ratio	Lower	Upper
82	100		
128	41.9	33.4	50.0
54	56.8	47.8	71.8







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.922 min Scan# 1331
Delta R.T. -0.005 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51

Tgt Ion:164 Resp: 246067

Ion	Ratio	Lower	Upper
164	100		
162	101.7	81.0	121.4
160	44.9	35.4	53.0







#42
2,4,6-Tribromophenol
Concen: 159.499 ng
RT: 10.710 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51

Instrument :
BNA_F
ClientSampleId :
WB-305-BOT

Tgt Ion:330 Resp: 367109

Ion	Ratio	Lower	Upper
330	100		
332	96.4	78.1	117.1
141	29.4	26.6	39.8







#45
2-Fluorobiphenyl
Concen: 100.787 ng
RT: 9.245 min Scan# 1216
Delta R.T. -0.006 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51

Tgt Ion:172 Resp: 1500601

Ion	Ratio	Lower	Upper
172	100		
171	35.9	28.6	43.0
170	23.9	19.1	28.7







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51

Instrument :
BNA_F
ClientSampleId :
WB-305-BOT

Tgt Ion:188 Resp: 437262

Ion	Ratio	Lower	Upper
188	100		
94	9.5	7.9	11.9
80	10.3	9.0	13.4







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.045 min Scan# 2032
Delta R.T. -0.006 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51

Tgt Ion:240 Resp: 219498

Ion	Ratio	Lower	Upper
240	100		
120	11.6	9.4	14.2
236	25.4	20.9	31.3







#79
Terphenyl-d14
Concen: 110.559 ng
RT: 12.998 min Scan# 18
Delta R.T. 0.000 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51

Instrument :
BNA_F
ClientSampleId :
WB-305-BOT

Tgt Ion:244 Resp: 1489277

Ion	Ratio	Lower	Upper
244	100		
212	6.6	5.7	8.5
122	9.8	8.6	13.0






#86
Perylene-d12
Concen: 20.000 ng
RT: 15.527 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BF140084.D
Acq: 28 Oct 2024 13:51

Tgt Ion:264 Resp: 258448

Ion	Ratio	Lower	Upper
264	100		
260	24.0	19.4	29.2
265	21.8	17.4	26.0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	10/25/24
Project:	Amtrak Sawtooth Bridges 2024	Date Received:	10/25/24
Client Sample ID:	WB-305-BOT-1	SDG No.:	P4578
Lab Sample ID:	P4578-08	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140085.D	1	10/26/24 12:10	10/28/24 14:19	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	15.5	U	15.5	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	8.40	U	8.40	50.0	ug/L
95-48-7	2-Methylphenol	11.3	U	11.3	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.5	U	11.5	100	ug/L
67-72-1	Hexachloroethane	10.1	U	10.1	50.0	ug/L
98-95-3	Nitrobenzene	12.7	U	12.7	50.0	ug/L
87-68-3	Hexachlorobutadiene	12.7	U	12.7	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	8.90	U	8.90	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	10.1	U	10.1	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	15.2	U	15.2	50.0	ug/L
118-74-1	Hexachlorobenzene	11.4	U	11.4	50.0	ug/L
87-86-5	Pentachlorophenol	18.5	U	18.5	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	121		15 (10) - 110 (139)	81%	SPK: 150
13127-88-3	Phenol-d6	112		15 (10) - 110 (134)	75%	SPK: 150
4165-60-0	Nitrobenzene-d5	96.4		30 (49) - 130 (133)	96%	SPK: 100
321-60-8	2-Fluorobiphenyl	93.8		30 (52) - 130 (132)	94%	SPK: 100
118-79-6	2,4,6-Tribromophenol	147		15 (44) - 110 (137)	98%	SPK: 150
1718-51-0	Terphenyl-d14	102		30 (48) - 130 (125)	102%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	111000	6.887			
1146-65-2	Naphthalene-d8	442000	8.169			
15067-26-2	Acenaphthene-d10	246000	9.922			
1517-22-2	Phenanthrene-d10	438000	11.41			
1719-03-5	Chrysene-d12	216000	14.045			
1520-96-3	Perylene-d12	231000	15.527			

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	10/25/24
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	10/25/24
Client Sample ID:	WB-305-BOT-1			SDG No.:	P4578
Lab Sample ID:	P4578-08			Matrix:	TCLP
Analytical Method:	SW8270			% Solid:	0
Sample Wt/Vol:	100	Units:	mL	Final Vol:	1000 uL
Soil Aliquot Vol:			uL	Test:	TCLP BNA
Extraction Type :			Decanted : N	Level :	LOW
Injection Volume :			GPC Factor : 1.0	GPC Cleanup : N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140085.D	1	10/26/24 12:10	10/28/24 14:19	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140085.D
Acq On : 28 Oct 2024 14:19
Operator : RC/JU
Sample : P4578-08
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WB-305-BOT-1

Quant Time: Oct 28 14:45:06 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.887	152	111257	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	441878	20.000	ng	0.00
39) Acenaphthene-d10	9.922	164	245603	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	437798	20.000	ng	0.00
76) Chrysene-d12	14.045	240	215751	20.000	ng	0.00
86) Perylene-d12	15.527	264	230540	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	863113	121.448	ng	0.02
7) Phenol-d6	6.510	99	1033565	112.289	ng	0.00
23) Nitrobenzene-d5	7.445	82	768626	96.407	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	338552	147.370	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1393721	93.785	ng	0.00
79) Terphenyl-d14	12.998	244	1353675	102.238	ng	0.00
Target Compounds						Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140085.D
Acq On : 28 Oct 2024 14:19
Operator : RC/JU
Sample : P4578-08
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WB-305-BOT-1

Quant Time: Oct 28 14:45:06 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.887 min Scan# 81
Delta R.T. -0.005 min
Lab File: BF140085.D
Acq: 28 Oct 2024 14:19

Instrument :
BNA_F
ClientSampleId :
WB-305-BOT-1

Tgt Ion:152 Resp: 111257

Ion	Ratio	Lower	Upper
152	100		
150	157.9	130.2	195.2
115	64.3	51.4	77.2







#5
2-Fluorophenol
Concen: 121.448 ng
RT: 5.516 min Scan# 582
Delta R.T. 0.018 min
Lab File: BF140085.D
Acq: 28 Oct 2024 14:19

Tgt Ion:112 Resp: 863113

Ion	Ratio	Lower	Upper
112	100		
64	64.4	53.0	79.6
63	33.1	27.0	40.4







#7
Phenol-d6
Concen: 112.289 ng
RT: 6.510 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF140085.D
Acq: 28 Oct 2024 14:19

Instrument :
BNA_F
ClientSampleId :
WB-305-BOT-1

Tgt Ion: 99 Resp: 1033565

Ion	Ratio	Lower	Upper
99	100		
42	19.3	16.7	25.1
71	34.9	27.7	41.5







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.169 min Scan# 1033
Delta R.T. -0.006 min
Lab File: BF140085.D
Acq: 28 Oct 2024 14:19

Tgt Ion:136 Resp: 441878

Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.6	12.8
54	9.5	8.4	12.6
68	6.3	5.1	7.7







#23
Nitrobenzene-d5
Concen: 96.407 ng
RT: 7.445 min Scan# 91
Delta R.T. -0.006 min
Lab File: BF140085.D
Acq: 28 Oct 2024 14:19

Instrument :
BNA_F
ClientSampleId :
WB-305-BOT-1

Tgt Ion: 82 Resp: 768626

Ion	Ratio	Lower	Upper
82	100		
128	42.3	33.4	50.0
54	56.8	47.8	71.8







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.922 min Scan# 1331
Delta R.T. -0.005 min
Lab File: BF140085.D
Acq: 28 Oct 2024 14:19

Tgt Ion:164 Resp: 245603

Ion	Ratio	Lower	Upper
164	100		
162	100.9	81.0	121.4
160	44.5	35.4	53.0







#42
2,4,6-Tribromophenol
Concen: 147.370 ng
RT: 10.710 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF140085.D
Acq: 28 Oct 2024 14:19

Instrument :
BNA_F
ClientSampleId :
WB-305-BOT-1

Tgt Ion:330 Resp: 338552

Ion	Ratio	Lower	Upper
330	100		
332	94.9	78.1	117.1
141	29.4	26.6	39.8







#45
2-Fluorobiphenyl
Concen: 93.785 ng
RT: 9.245 min Scan# 1216
Delta R.T. -0.006 min
Lab File: BF140085.D
Acq: 28 Oct 2024 14:19

Tgt Ion:172 Resp: 1393721

Ion	Ratio	Lower	Upper
172	100		
171	35.6	28.6	43.0
170	23.7	19.1	28.7







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140085.D
Acq: 28 Oct 2024 14:19

Instrument :
BNA_F
ClientSampleId :
WB-305-BOT-1

Tgt Ion:188 Resp: 437798

Ion	Ratio	Lower	Upper
188	100		
94	9.8	7.9	11.9
80	10.4	9.0	13.4







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.045 min Scan# 2032
Delta R.T. -0.006 min
Lab File: BF140085.D
Acq: 28 Oct 2024 14:19

Tgt Ion:240 Resp: 215751

Ion	Ratio	Lower	Upper
240	100		
120	11.4	9.4	14.2
236	25.6	20.9	31.3







#79
Terphenyl-d14
Concen: 102.238 ng
RT: 12.998 min Scan# 18
Delta R.T. 0.000 min
Lab File: BF140085.D
Acq: 28 Oct 2024 14:19

Instrument :
BNA_F
ClientSampleId :
WB-305-BOT-1

Tgt Ion:244 Resp: 1353675

Ion	Ratio	Lower	Upper
244	100		
212	6.8	5.7	8.5
122	9.8	8.6	13.0






#86
Perylene-d12
Concen: 20.000 ng
RT: 15.527 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BF140085.D
Acq: 28 Oct 2024 14:19

Tgt Ion:264 Resp: 230540

Ion	Ratio	Lower	Upper
264	100		
260	24.3	19.4	29.2
265	21.5	17.4	26.0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	10/26/24
Project:	Amtrak Sawtooth Bridges 2024	Date Received:	10/26/24
Client Sample ID:	PB164393TB	SDG No.:	P4578
Lab Sample ID:	PB164393TB	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140078.D	1	10/26/24 12:10	10/28/24 10:58	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	15.5	U	15.5	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	8.40	U	8.40	50.0	ug/L
95-48-7	2-Methylphenol	11.3	U	11.3	50.0	ug/L
65794-96-9	3+4-Methylphenols	11.5	U	11.5	100	ug/L
67-72-1	Hexachloroethane	10.1	U	10.1	50.0	ug/L
98-95-3	Nitrobenzene	12.7	U	12.7	50.0	ug/L
87-68-3	Hexachlorobutadiene	12.7	U	12.7	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	8.90	U	8.90	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	10.1	U	10.1	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	15.2	U	15.2	50.0	ug/L
118-74-1	Hexachlorobenzene	11.4	U	11.4	50.0	ug/L
87-86-5	Pentachlorophenol	18.5	U	18.5	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	122		15 (10) - 110 (139)	82%	SPK: 150
13127-88-3	Phenol-d6	119		15 (10) - 110 (134)	79%	SPK: 150
4165-60-0	Nitrobenzene-d5	92.2		30 (49) - 130 (133)	92%	SPK: 100
321-60-8	2-Fluorobiphenyl	87.9		30 (52) - 130 (132)	88%	SPK: 100
118-79-6	2,4,6-Tribromophenol	137		15 (44) - 110 (137)	91%	SPK: 150
1718-51-0	Terphenyl-d14	88.8		30 (48) - 130 (125)	89%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	128000	6.887			
1146-65-2	Naphthalene-d8	508000	8.169			
15067-26-2	Acenaphthene-d10	289000	9.922			
1517-22-2	Phenanthrene-d10	536000	11.41			
1719-03-5	Chrysene-d12	334000	14.051			
1520-96-3	Perylene-d12	262000	15.527			

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	10/26/24
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	10/26/24
Client Sample ID:	PB164393TB			SDG No.:	P4578
Lab Sample ID:	PB164393TB			Matrix:	TCLP
Analytical Method:	SW8270			% Solid:	0
Sample Wt/Vol:	100	Units:	mL	Final Vol:	1000 uL
Soil Aliquot Vol:			uL	Test:	TCLP BNA
Extraction Type :			Decanted : N	Level :	LOW
Injection Volume :			GPC Factor : 1.0	GPC Cleanup : N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140078.D	1	10/26/24 12:10	10/28/24 10:58	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140078.D
Acq On : 28 Oct 2024 10:58
Operator : RC/JU
Sample : PB164393TB
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB164393TB

Quant Time: Oct 28 11:28:25 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.887	152	128211	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	508163	20.000	ng	0.00
39) Acenaphthene-d10	9.922	164	288600	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	536321	20.000	ng	0.00
76) Chrysene-d12	14.051	240	334195	20.000	ng	0.00
86) Perylene-d12	15.527	264	261975	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	1001524	122.289	ng	0.02
7) Phenol-d6	6.510	99	1264624	119.224	ng	0.00
23) Nitrobenzene-d5	7.446	82	845367	92.201	ng	0.00
42) 2,4,6-Tribromophenol	10.716	330	368703	136.583	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1534211	87.858	ng	0.00
79) Terphenyl-d14	12.998	244	1821703	88.823	ng	0.00
Target Compounds						Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140078.D
Acq On : 28 Oct 2024 10:58
Operator : RC/JU
Sample : PB164393TB
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB164393TB

Quant Time: Oct 28 11:28:25 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.887 min Scan# 81
Delta R.T. -0.005 min
Lab File: BF140078.D
Acq: 28 Oct 2024 10:58

Instrument :
BNA_F
ClientSampleId :
PB164393TB

Tgt Ion:152 Resp: 128211

Ion	Ratio	Lower	Upper
152	100		
150	156.8	130.2	195.2
115	64.7	51.4	77.2







#5
2-Fluorophenol
Concen: 122.289 ng
RT: 5.516 min Scan# 582
Delta R.T. 0.018 min
Lab File: BF140078.D
Acq: 28 Oct 2024 10:58

Tgt Ion:112 Resp: 1001524

Ion	Ratio	Lower	Upper
112	100		
64	64.3	53.0	79.6
63	33.5	27.0	40.4







#7
Phenol-d6
Concen: 119.224 ng
RT: 6.510 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF140078.D
Acq: 28 Oct 2024 10:58

Instrument :
BNA_F
ClientSampleId :
PB164393TB

Tgt Ion: 99 Resp: 1264624

Ion	Ratio	Lower	Upper
99	100		
42	19.9	16.7	25.1
71	36.0	27.7	41.5







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.169 min Scan# 1033
Delta R.T. -0.006 min
Lab File: BF140078.D
Acq: 28 Oct 2024 10:58

Tgt Ion:136 Resp: 508163

Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.6	12.8
54	10.1	8.4	12.6
68	6.0	5.1	7.7







#23
Nitrobenzene-d5
Concen: 92.201 ng
RT: 7.446 min Scan# 91
Delta R.T. -0.006 min
Lab File: BF140078.D
Acq: 28 Oct 2024 10:58

Instrument :
BNA_F
ClientSampleId :
PB164393TB

Tgt Ion: 82 Resp: 845367

Ion	Ratio	Lower	Upper
82	100		
128	42.3	33.4	50.0
54	56.9	47.8	71.8







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.922 min Scan# 1331
Delta R.T. -0.005 min
Lab File: BF140078.D
Acq: 28 Oct 2024 10:58

Tgt Ion:164 Resp: 288600

Ion	Ratio	Lower	Upper
164	100		
162	101.6	81.0	121.4
160	45.2	35.4	53.0







#42
2,4,6-Tribromophenol
Concen: 136.583 ng
RT: 10.716 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF140078.D
Acq: 28 Oct 2024 10:58

Instrument :
BNA_F
ClientSampleId :
PB164393TB

Tgt Ion:330 Resp: 368703

Ion	Ratio	Lower	Upper
330	100		
332	97.6	78.1	117.1
141	29.6	26.6	39.8







#45
2-Fluorobiphenyl
Concen: 87.858 ng
RT: 9.245 min Scan# 1216
Delta R.T. -0.006 min
Lab File: BF140078.D
Acq: 28 Oct 2024 10:58

Tgt Ion:172 Resp: 1534211

Ion	Ratio	Lower	Upper
172	100		
171	35.4	28.6	43.0
170	23.8	19.1	28.7







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140078.D
Acq: 28 Oct 2024 10:58

Instrument :
BNA_F
ClientSampleId :
PB164393TB

Tgt Ion:188 Resp: 536321

Ion	Ratio	Lower	Upper
188	100		
94	9.4	7.9	11.9
80	10.6	9.0	13.4







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.051 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BF140078.D
Acq: 28 Oct 2024 10:58

Tgt Ion:240 Resp: 334195

Ion	Ratio	Lower	Upper
240	100		
120	10.5	9.4	14.2
236	26.0	20.9	31.3







#79
Terphenyl-d14
Concen: 88.823 ng
RT: 12.998 min Scan# 18
Delta R.T. 0.000 min
Lab File: BF140078.D
Acq: 28 Oct 2024 10:58

Instrument :
BNA_F
ClientSampleId :
PB164393TB

Tgt Ion:244 Resp: 1821703

Ion	Ratio	Lower	Upper
244	100		
212	6.8	5.7	8.5
122	9.9	8.6	13.0






#86
Perylene-d12
Concen: 20.000 ng
RT: 15.527 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BF140078.D
Acq: 28 Oct 2024 10:58

Tgt Ion:264 Resp: 261975

Ion	Ratio	Lower	Upper
264	100		
260	23.4	19.4	29.2
265	21.3	17.4	26.0



CALIBRATION SUMMARY

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG No.: P4578

Instrument ID: BNA_F Calibration Date(s): 10/18/2024 10/18/2024

Calibration Time(s): 10:27 13:46

LAB FILE ID: RRF2.5 = BF139844.D RRF005 = BF139845.D RRF010 = BF139846.D
RRF020 = BF139847.D RRF040 = BF139848.D RRF050 = BF139849.D

COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Pyridine		1.622	1.570	1.522	1.504	1.386	1.476	7.3
2-Fluorophenol		1.465	1.398	1.331	1.278	1.179	1.278	10.1
Phenol-d6		1.900	1.818	1.709	1.647	1.538	1.655	10.1
1,4-Dichlorobenzene		1.723	1.641	1.558	1.487	1.392	1.498	10.2
2-Methylphenol		1.264	1.164	1.134	1.114	1.053	1.114	7.7
3+4-Methylphenols		1.678	1.573	1.520	1.418	1.309	1.424	12.4
Nitrobenzene-d5		0.365	0.365	0.372	0.371	0.354	0.361	2.9
Hexachloroethane		0.583	0.571	0.549	0.541	0.507	0.534	7.1
Nitrobenzene		0.417	0.402	0.408	0.403	0.383	0.396	4.1
Hexachlorobutadiene		0.224	0.206	0.203	0.197	0.185	0.197	8.0
2,4,6-Trichlorophenol		0.405	0.382	0.400	0.387	0.363	0.382	4.8
2-Fluorobiphenyl		1.512	1.396	1.280	1.162	1.088	1.210	16.0
2,4,5-Trichlorophenol		0.426	0.425	0.398	0.401	0.381	0.394	6.6
2,4-Dinitrotoluene		0.280	0.314	0.337	0.356	0.344	0.332	7.9
2,4,6-Tribromophenol		0.201	0.196	0.193	0.188	0.179	0.187	5.5
Hexachlorobenzene		0.258	0.245	0.237	0.231	0.217	0.232	7.2
Pentachlorophenol		0.118	0.137	0.148	0.152	0.145	0.141	8.0
Terphenyl-d14		1.381	1.335	1.340	1.244	1.154	1.227	11.0

All other compounds must meet a minimum RRF of 0.010.

Response Factor Report BNA_F

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF101824.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Oct 18 15:07:50 2024
Response Via : Initial Calibration

Calibration Files
2.5 =BF139844.D 5 =BF139845.D 10 =BF139846.D 20 =BF139847.D 40 =BF139848.D 50 =BF139849.D 60 =BF139850.D 80 =BF1398
51.D

		Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1)	I	1,4-Dichlorobenzen...	----------ISTD----------									
2)		1,4-Dioxane		0.651	0.625	0.603	0.591	0.571	0.581	0.548	0.596	5.74
3)		Pyridine		1.622	1.570	1.522	1.504	1.386	1.406	1.326	1.476	7.26
4)		n-Nitrosodimet...		0.815	0.800	0.799	0.806	0.782	0.794	0.757	0.793	2.39
5)	S	2-Fluorophenol		1.465	1.398	1.331	1.278	1.179	1.186	1.106	1.278	10.10
6)		Aniline		1.673	1.649	1.618	1.582	1.471	1.479	1.324	1.542	8.05
7)	S	Phenol-d6		1.900	1.818	1.709	1.647	1.538	1.539	1.432	1.655	10.06
8)		2-Chlorophenol		1.503	1.422	1.358	1.310	1.212	1.221	1.116	1.306	10.24
9)		Benzaldehyde			1.137	1.042	0.940	0.873	0.853	0.740	0.931	15.22
10)	C	Phenol		1.952	1.832	1.760	1.712	1.583	1.601	1.502	1.706	9.19
11)		bis(2-Chloroet...		1.470	1.423	1.359	1.294	1.251	1.249	1.183	1.319	7.82
12)		1,3-Dichlorobe...		1.718	1.656	1.544	1.499	1.392	1.391	1.294	1.499	10.19
13)	C	1,4-Dichlorobe...		1.723	1.641	1.558	1.487	1.392	1.391	1.291	1.498	10.19
14)		1,2-Dichlorobe...		1.660	1.579	1.478	1.379	1.273	1.267	1.149	1.398	13.15
15)		Benzyl Alcohol		1.355	1.299	1.257	1.213	1.154	1.146	1.071	1.214	8.07
16)		2,2'-oxybis(1-...		2.524	2.409	2.353	2.255	2.117	2.115	1.964	2.248	8.69
17)		2-Methylphenol		1.264	1.164	1.134	1.114	1.053	1.063	1.004	1.114	7.69
18)		Hexachloroethane		0.583	0.571	0.549	0.541	0.507	0.511	0.477	0.534	7.06
19)	P	n-Nitroso-di-n...	1.105	1.141	1.066	1.025	0.974	0.921	0.927	0.869	1.004	9.63
20)		3+4-Methylphenols		1.678	1.573	1.520	1.418	1.309	1.300	1.173	1.424	12.41
21)	I	Naphthalene-d8	----------ISTD----------									
22)		Acetophenone		0.578	0.533	0.516	0.481	0.452	0.454	0.414	0.490	11.43
23)	S	Nitrobenzene-d5		0.365	0.365	0.372	0.371	0.354	0.357	0.342	0.361	2.92
24)		Nitrobenzene		0.417	0.402	0.408	0.403	0.383	0.388	0.370	0.396	4.10
25)		Isophorone		0.755	0.709	0.702	0.679	0.652	0.660	0.637	0.685	5.90
26)	C	2-Nitrophenol		0.124	0.137	0.150	0.157	0.158	0.161	0.157	0.149	9.22
27)		2,4-Dimethylph...		0.285	0.259	0.256	0.248	0.237	0.234	0.223	0.249	8.07
28)		bis(2-Chloroet...		0.468	0.440	0.432	0.412	0.394	0.390	0.369	0.415	8.22
29)	C	2,4-Dichloroph...		0.308	0.297	0.293	0.283	0.272	0.274	0.257	0.283	6.17
30)		1,2,4-Trichlor...		0.351	0.331	0.325	0.315	0.298	0.298	0.279	0.314	7.68
31)		Naphthalene		1.209	1.121	1.091	1.023	0.958	0.944	0.875	1.031	11.23
32)		Benzoic acid			0.175	0.207	0.220	0.231	0.235	0.235	0.217	10.69
33)		4-Chloroaniline		0.397	0.382	0.368	0.351	0.332	0.326	0.306	0.352	9.26
34)	C	Hexachlorobuta...		0.224	0.206	0.203	0.197	0.185	0.187	0.177	0.197	7.96
35)		Caprolactam		0.092	0.092	0.092	0.092	0.088	0.088	0.086	0.090	2.85
36)	C	4-Chloro-3-met...		0.341	0.328	0.324	0.315	0.299	0.303	0.287	0.314	6.03
37)		2-Methylnaphth...		0.740	0.689	0.666	0.621	0.587	0.583	0.537	0.632	11.11
38)		1-Methylnaphth...		0.726	0.683	0.655	0.612	0.569	0.567	0.526	0.620	11.55


Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF101824.M

39)	I	Acenaphthene-d10	----------ISTD----------								
40)		1,2,4,5-Tetrac...	0.609	0.579	0.562	0.537	0.512	0.506	0.472	0.540	8.72
41)	P	Hexachlorocycl...	0.185	0.193	0.198	0.198	0.186	0.185	0.169	0.188	5.28
42)	S	2,4,6-Tribromo...	0.201	0.196	0.193	0.188	0.179	0.180	0.173	0.187	5.47
43)	C	2,4,6-Trichlor...	0.405	0.382	0.400	0.387	0.363	0.383	0.354	0.382	4.80
44)		2,4,5-Trichlor...	0.426	0.425	0.398	0.401	0.381	0.369	0.359	0.394	6.58
45)	S	2-Fluorobiphenyl	1.512	1.396	1.280	1.162	1.088	1.060	0.973	1.210	16.05
46)		1,1'-Biphenyl	1.640	1.536	1.483	1.379	1.285	1.262	1.161	1.392	12.18
47)		2-Chloronaphth...	1.288	1.225	1.164	1.111	1.048	1.040	0.973	1.121	9.95
48)		2-Nitroaniline	0.299	0.318	0.353	0.365	0.354	0.362	0.349	0.343	7.20
49)		Acenaphthylene	1.854	1.756	1.717	1.615	1.507	1.505	1.394	1.621	10.06
50)		Dimethylphthalate	1.410	1.344	1.283	1.237	1.172	1.163	1.110	1.246	8.60
51)		2,6-Dinitrotol...	0.256	0.266	0.278	0.280	0.273	0.274	0.260	0.270	3.41
52)	C	Acenaphthene	1.200	1.136	1.089	1.044	0.977	0.983	0.913	1.049	9.55
53)		3-Nitroaniline	0.270	0.275	0.296	0.292	0.276	0.282	0.256	0.278	4.84
54)	P	2,4-Dinitrophenol		0.056	0.077	0.101	0.101	0.113	0.112	0.093	23.89
55)		Dibenzofuran	1.767	1.659	1.579	1.486	1.389	1.375	1.278	1.505	11.52
56)	P	4-Nitrophenol	0.187	0.207	0.225	0.232	0.219	0.220	0.208	0.214	6.84
57)		2,4-Dinitrotol...	0.280	0.314	0.337	0.356	0.344	0.351	0.338	0.332	7.94
58)		Fluorene	1.409	1.309	1.201	1.110	1.029	1.021	0.944	1.146	14.68
59)		2,3,4,6-Tetrac...	0.334	0.322	0.326	0.310	0.296	0.293	0.281	0.309	6.33
60)		Diethylphthalate	1.366	1.308	1.267	1.214	1.165	1.161	1.080	1.223	7.99
61)		4-Chlorophenyl...	0.698	0.638	0.617	0.569	0.526	0.520	0.484	0.579	13.14
62)		4-Nitroaniline	0.249	0.259	0.270	0.275	0.263	0.269	0.254	0.263	3.65
63)		Azobenzene	1.459	1.385	1.355	1.306	1.216	1.212	1.143	1.297	8.62
64)	I	Phenanthrene-d10	----------ISTD----------								
65)		4,6-Dinitro-2-...		0.055	0.072	0.087	0.090	0.092	0.093	0.082	18.60
66)	c	n-Nitrosodiphe...	0.672	0.641	0.621	0.595	0.568	0.561	0.533	0.599	8.14
67)		4-Bromophenyl-...	0.230	0.217	0.211	0.202	0.197	0.196	0.189	0.206	6.98
68)		Hexachlorobenzene	0.258	0.245	0.237	0.231	0.217	0.221	0.212	0.232	7.20
69)		Atrazine	0.189	0.175	0.156	0.175	0.135	0.153	0.151	0.162	11.36
70)	C	Pentachlorophenol	0.118	0.137	0.148	0.152	0.145	0.144	0.140	0.141	8.04
71)		Phenanthrene	1.121	1.035	0.994	0.933	0.863	0.860	0.807	0.945	11.79
72)		Anthracene	1.082	1.014	0.972	0.913	0.851	0.833	0.787	0.922	11.53
73)		Carbazole	1.002	0.964	0.923	0.846	0.776	0.772	0.717	0.857	12.64
74)		Di-n-butylphth...	1.104	1.071	1.062	1.014	0.923	0.909	0.850	0.990	9.77
75)	C	Fluoranthene	1.149	1.108	1.036	0.943	0.842	0.835	0.772	0.955	15.34
76)	I	Chrysene-d12	----------ISTD----------								
77)		Benzidine	0.457	0.461	0.293	0.366	0.276	0.203	0.246	0.329	30.86
78)		Pyrene	1.892	1.828	1.900	1.805	1.685	1.649	1.496	1.751	8.43
79)	S	Terphenyl-d14	1.381	1.335	1.340	1.244	1.154	1.121	1.018	1.227	10.96
80)		Butylbenzylpht...	0.490	0.513	0.536	0.555	0.535	0.531	0.514	0.525	3.99
81)		Benzo(a)anthra...	1.425	1.355	1.329	1.331	1.267	1.237	1.169	1.302	6.48
82)		3,3'-Dichlorob...	0.368	0.372	0.390	0.387	0.374	0.382	0.384	0.380	2.15
83)		Chrysene	1.325	1.244	1.234	1.167	1.134	1.151	1.101	1.194	6.52
84)		Bis(2-ethylhex...	0.520	0.539	0.577	0.626	0.620	0.626	0.614	0.589	7.48
85)	c	Di-n-octyl pht...		0.786	0.930	1.135	1.182	1.207	1.186	1.071	16.14

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF101824.M

86)	I	Perylene-d12	----------ISTD----------							
87)		Indeno(1,2,3-c...	1.209	1.261	1.307	1.352	1.299	1.326	1.253	1.287	3.77
88)		Benzo(b)fluora...	1.317	1.240	1.319	1.196	1.105	1.239	1.111	1.218	7.15
89)		Benzo(k)fluora...	1.213	1.177	0.992	1.066	1.030	0.929	0.947	1.051	10.42
90)	C	Benzo(a)pyrene	1.030	1.024	1.018	1.025	0.974	0.999	0.947	1.002	3.12
91)		Dibenzo(a,h)an...	1.021	1.064	1.103	1.120	1.083	1.085	1.036	1.073	3.31
92)		Benzo(g,h,i)pe...	1.030	1.046	1.090	1.128	1.081	1.095	1.035	1.072	3.37

(#) = Out of Range



Response = 1.220e-001 * Amt - 3.983e-002
Coef of Det (r^2) = 0.997175 Curve Fit: Linear
Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF101824.M
Calibration Table Last Updated: Fri Oct 18 15:07:50 2024

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139844.D
Acq On : 18 Oct 2024 10:27
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Quant Time: Oct 18 14:15:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.887	152	190003	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	748539	20.000	ng	0.00
39) Acenaphthene-d10	9.928	164	423999	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	760140	20.000	ng	0.00
76) Chrysene-d12	14.051	240	519381	20.000	ng	0.00
86) Perylene-d12	15.527	264	410408	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						Qvalue
19) n-Nitroso-di-n-propyla...	7.287	70	26252	2.754	ng	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139844.D
Acq On : 18 Oct 2024 10:27
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Quant Time: Oct 18 14:15:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.887 min Scan# 8
Delta R.T. -0.005 min
Lab File: BF139844.D
Acq: 18 Oct 2024 10:27

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:152 Resp: 190003

Ion	Ratio	Lower	Upper
152	100		
150	158.9	130.2	195.2
115	64.4	51.4	77.2







#5
2-Fluorophenol
Concen: 0.000 ng
Expected RT: 5.50 min

Lab File: BF139844.D
Acq: 18 Oct 2024 10:27

Tgt Ion: 112

Sig	Exp Ratio
112	100
64	66.3
63	33.7






#7
Phenol-d6
Concen: 0.000 ng
Expected RT: 6.51 min


Lab File: BF139844.D
Acq: 18 Oct 2024 10:27

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion: 99

Sig	Exp Ratio
99	100
42	20.9
71	34.6








#19
n-Nitroso-di-n-propylamine
Concen: 2.754 ng
RT: 7.287 min Scan# 883
Delta R.T. -0.017 min
Lab File: BF139844.D
Acq: 18 Oct 2024 10:27


Tgt Ion: 70 Resp: 26252

Ion	Ratio	Lower	Upper
70	100		
42	70.4	51.5	77.3
101	10.6	7.5	11.3
130	18.9	15.5	23.3











#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.169 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139844.D
Acq: 18 Oct 2024 10:27

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:136 Resp: 748539

Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.6	12.8
54	10.6	8.4	12.6
68	6.5	5.1	7.7







#23
Nitrobenzene-d5
Concen: 0.000 ng
Expected RT: 7.45 min

Lab File: BF139844.D
Acq: 18 Oct 2024 10:27

Tgt Ion: 82

Sig	Exp Ratio
82	100
128	41.7
54	59.8







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.928 min Scan# 1
Delta R.T. 0.001 min
Lab File: BF139844.D
Acq: 18 Oct 2024 10:27

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:164 Resp: 423999

Ion	Ratio	Lower	Upper
164	100		
162	99.5	81.0	121.4
160	43.8	35.4	53.0






#42
2,4,6-Tribromophenol
Concen: 0.000 ng
Expected RT: 10.72 min

Lab File: BF139844.D
Acq: 18 Oct 2024 10:27

Tgt Ion: 330

Sig	Exp Ratio
330	100
332	97.6
141	33.2




#45
2-Fluorobiphenyl
Concen: 0.000 ng
Expected RT: 9.25 min

Lab File: BF139844.D
Acq: 18 Oct 2024 10:27

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion: 172


Sig	Exp Ratio
172	100
171	35.8
170	23.9




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1584
Delta R.T. -0.006 min
Lab File: BF139844.D
Acq: 18 Oct 2024 10:27

Tgt Ion:188 Resp: 760140

Ion	Ratio	Lower	Upper
188	100		
94	10.4	7.9	11.9
80	11.6	9.0	13.4







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.051 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139844.D
Acq: 18 Oct 2024 10:27

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:240 Resp: 519381

Ion	Ratio	Lower	Upper
240	100		
120	11.9	9.4	14.2
236	26.4	20.9	31.3







#79
Terphenyl-d14
Concen: 0.000 ng
Expected RT: 13.00 min

Lab File: BF139844.D
Acq: 18 Oct 2024 10:27

Tgt Ion: 244

Sig	Exp Ratio
244	100
212	7.1
122	10.8






#86
Perylene-d12
Concen: 20.000 ng
RT: 15.527 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139844.D
Acq: 18 Oct 2024 10:27

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Tgt Ion:264 Resp: 410408

Ion	Ratio	Lower	Upper
264	100		
260	23.9	19.4	29.2
265	21.5	17.4	26.0



Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139845.D
Acq On : 18 Oct 2024 10:55
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Quant Time: Oct 18 14:16:29 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.887	152	189112	20.000	ng		0.00
21) Naphthalene-d8	8.169	136	737062	20.000	ng		0.00
39) Acenaphthene-d10	9.928	164	414324	20.000	ng		0.00
64) Phenanthrene-d10	11.410	188	741430	20.000	ng		0.00
76) Chrysene-d12	14.051	240	477223	20.000	ng		0.00
86) Perylene-d12	15.527	264	377685	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.493	112	138495	11.465	ng		0.00
7) Phenol-d6	6.493	99	179648	11.482	ng		-0.02
23) Nitrobenzene-d5	7.445	82	134439	10.109	ng		0.00
42) 2,4,6-Tribromophenol	10.710	330	41682	10.755	ng		0.00
45) 2-Fluorobiphenyl	9.245	172	313171	12.492	ng		0.00
79) Terphenyl-d14	12.992	244	329424	11.248	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.687	88	30759	5.461	ng		97
3) Pyridine	3.452	79	76703	5.494	ng		98
4) n-Nitrosodimethylamine	3.375	42	38552	5.140	ng	#	96
6) Aniline	6.545	93	79083	5.424	ng		98
8) 2-Chlorophenol	6.669	128	71065	5.754	ng		98
10) Phenol	6.510	94	92287	5.620	ng		95
11) bis(2-Chloroethyl)ether	6.616	93	69504	5.575	ng		98
12) 1,3-Dichlorobenzene	6.828	146	81239	5.731	ng		99
13) 1,4-Dichlorobenzene	6.904	146	81442	5.750	ng		99
14) 1,2-Dichlorobenzene	7.057	146	78458	5.936	ng		97
15) Benzyl Alcohol	7.016	79	64060	5.582	ng		99
16) 2,2'-oxybis(1-Chloropr...	7.163	45	119350	5.614	ng		100
17) 2-Methylphenol	7.122	107	59770	5.676	ng		96
18) Hexachloroethane	7.404	117	27566	5.458	ng		96
19) n-Nitroso-di-n-propyla...	7.287	70	53958	5.686	ng		97
20) 3+4-Methylphenols	7.275	107	79329	5.890	ng		92
22) Acetophenone	7.287	105	106522	5.900	ng		98
24) Nitrobenzene	7.463	77	76831	5.269	ng		96
25) Isophorone	7.698	82	139084	5.510	ng		100
26) 2-Nitrophenol	7.781	139	22867	4.157	ng		95
27) 2,4-Dimethylphenol	7.810	122	52430	5.716	ng		97
28) bis(2-Chloroethoxy)met...	7.916	93	86298	5.643	ng		99
29) 2,4-Dichlorophenol	8.016	162	56772	5.434	ng		99
30) 1,2,4-Trichlorobenzene	8.110	180	64640	5.592	ng		99
31) Naphthalene	8.192	128	222725	5.859	ng		99
33) 4-Chloroaniline	8.234	127	73196	5.647	ng		99
34) Hexachlorobutadiene	8.310	225	41186	5.670	ng		97
35) Caprolactam	8.557	113	16989	5.114	ng		98
36) 4-Chloro-3-methylphenol	8.704	107	62903	5.438	ng		99
37) 2-Methylnaphthalene	8.881	142	136292	5.854	ng		99
38) 1-Methylnaphthalene	8.981	142	133747	5.856	ng		100
40) 1,2,4,5-Tetrachloroben...	9.045	216	63046	5.640	ng		100
41) Hexachlorocyclopentadiene	9.039	237	19121	4.916	ng		97
43) 2,4,6-Trichlorophenol	9.151	196	41997	5.307	ng		98
44) 2,4,5-Trichlorophenol	9.187	196	44086	5.399	ng		99
46) 1,1'-Biphenyl	9.345	154	169920	5.891	ng		99

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139845.D
Acq On : 18 Oct 2024 10:55
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Quant Time: Oct 18 14:16:29 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
47) 2-Chloronaphthalene	9.369	162	133456	5.745	ng		99
48) 2-Nitroaniline	9.457	65	30999	4.363	ng		99
49) Acenaphthylene	9.786	152	192014	5.717	ng		99
50) Dimethylphthalate	9.639	163	146096	5.661	ng		100
51) 2,6-Dinitrotoluene	9.698	165	26490	4.741	ng		96
52) Acenaphthene	9.957	154	124288	5.721	ng		100
53) 3-Nitroaniline	9.863	138	27995	4.858	ng	#	94
55) Dibenzofuran	10.128	168	182985	5.870	ng		98
56) 4-Nitrophenol	10.004	139	19402	4.376	ng		89
57) 2,4-Dinitrotoluene	10.098	165	29038	4.226	ng		92
58) Fluorene	10.469	166	145913	6.146	ng		99
59) 2,3,4,6-Tetrachlorophenol	10.239	232	34636	5.411	ng		97
60) Diethylphthalate	10.339	149	141466	5.583	ng		99
61) 4-Chlorophenyl-phenyle...	10.469	204	72331	6.033	ng		96
62) 4-Nitroaniline	10.469	138	25747	4.731	ng		96
63) Azobenzene	10.622	77	151158	5.627	ng		99
66) n-Nitrosodiphenylamine	10.581	169	124472	5.609	ng		100
67) 4-Bromophenyl-phenylether	10.957	248	42707	5.591	ng		94
68) Hexachlorobenzene	11.016	284	47895	5.575	ng		98
69) Atrazine	11.104	200	35075	5.835	ng		99
70) Pentachlorophenol	11.210	266	21781	4.178	ng		93
71) Phenanthrene	11.433	178	207716	5.931	ng		98
72) Anthracene	11.486	178	200621	5.871	ng		99
73) Carbazole	11.639	167	185743	5.845	ng		99
74) Di-n-butylphthalate	11.975	149	204647	5.574	ng		99
75) Fluoranthene	12.622	202	213038	6.017	ng		100
77) Benzidine	12.745	184	54501	7.022	ng		99
78) Pyrene	12.851	202	225742	5.404	ng		99
80) Butylbenzylphthalate	13.474	149	58502	4.671	ng		100
81) Benzo(a)anthracene	14.039	228	170026	5.473	ng		99
82) 3,3'-Dichlorobenzidine	14.004	252	43937	4.851	ng		100
83) Chrysene	14.074	228	158121	5.551	ng		98
84) Bis(2-ethylhexyl)phtha...	14.033	149	62085	4.419	ng		97
87) Indeno(1,2,3-cd)pyrene	17.010	276	114185	4.699	ng		98
88) Benzo(b)fluoranthene	15.092	252	124353	5.405	ng		100
89) Benzo(k)fluoranthene	15.121	252	114525	5.772	ng		99
90) Benzo(a)pyrene	15.463	252	97243	5.137	ng		100
91) Dibenzo(a,h)anthracene	17.033	278	96416	4.757	ng		99
92) Benzo(g,h,i)perylene	17.457	276	97292	4.805	ng		99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139845.D
Acq On : 18 Oct 2024 10:55
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Quant Time: Oct 18 14:16:29 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.887 min Scan# 8
Delta R.T. -0.005 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:152 Resp: 189112

Ion	Ratio	Lower	Upper
152	100		
150	157.4	130.2	195.2
115	65.7	51.4	77.2







#2
1,4-Dioxane
Concen: 5.461 ng
RT: 2.687 min Scan# 101
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion: 88 Resp: 30759

Ion	Ratio	Lower	Upper
88	100		
58	89.3	73.3	109.9
43	35.6	27.1	40.7







#3
Pyridine
Concen: 5.494 ng
RT: 3.452 min Scan# 21
Delta R.T. 0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion: 79 Resp: 76703

Ion	Ratio	Lower	Upper
79	100		
52	86.7	68.5	102.7
51	42.2	32.3	48.5







#4
n-Nitrosodimethylamine
Concen: 5.140 ng
RT: 3.375 min Scan# 218
Delta R.T. -0.023 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion: 42 Resp: 38552

Ion	Ratio	Lower	Upper
42	100		
74	115.9	96.2	144.4
44	11.0	6.2	9.2#







#5
2-Fluorophenol
Concen: 11.465 ng
RT: 5.493 min Scan# 5
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:112 Resp: 138495

Ion	Ratio	Lower	Upper
112	100		
64	65.5	53.0	79.6
63	33.2	27.0	40.4







#6
Aniline
Concen: 5.424 ng
RT: 6.545 min Scan# 757
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion: 93 Resp: 79083

Ion	Ratio	Lower	Upper
93	100		
66	41.4	34.0	51.0
65	21.1	17.4	26.0







#7
Phenol-d6
Concen: 11.482 ng
RT: 6.493 min Scan# 74
Delta R.T. -0.017 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion: 99 Resp: 179648

Ion	Ratio	Lower	Upper
99	100		
42	20.7	16.7	25.1
71	34.3	27.7	41.5







#8
2-Chlorophenol
Concen: 5.754 ng
RT: 6.669 min Scan# 778
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:128 Resp: 71065

Ion	Ratio	Lower	Upper
128	100		
130	33.3	12.3	52.3
64	52.8	34.6	74.6







#10
Phenol
Concen: 5.620 ng
RT: 6.510 min Scan# 7
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion: 94 Resp: 92287

Ion	Ratio	Lower	Upper
94	100		
65	29.4	11.1	51.1
66	38.6	22.7	62.7







#11
bis(2-Chloroethyl)ether
Concen: 5.575 ng
RT: 6.616 min Scan# 769
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion: 93 Resp: 69504

Ion	Ratio	Lower	Upper
93	100		
63	87.5	65.0	105.0
95	32.6	12.4	52.4







#12
1,3-Dichlorobenzene
Concen: 5.731 ng
RT: 6.828 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:146 Resp: 81239

Ion	Ratio	Lower	Upper
146	100		
148	64.8	51.0	76.4
75	33.2	26.4	39.6







#13
1,4-Dichlorobenzene
Concen: 5.750 ng
RT: 6.904 min Scan# 818
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:146 Resp: 81442

Ion	Ratio	Lower	Upper
146	100		
148	63.7	51.8	77.8
111	43.5	34.6	51.8







#14
1,2-Dichlorobenzene
Concen: 5.936 ng
RT: 7.057 min Scan# 84
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:146 Resp: 78458

Ion	Ratio	Lower	Upper
146	100		
148	63.0	51.4	77.2
111	43.3	36.7	55.1







#15
Benzyl Alcohol
Concen: 5.582 ng
RT: 7.016 min Scan# 837
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion: 79 Resp: 64060

Ion	Ratio	Lower	Upper
79	100		
108	75.6	61.4	92.0
77	62.3	49.9	74.9







#16
2,2'-oxybis(1-Chloropropane)
Concen: 5.614 ng
RT: 7.163 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion: 45 Resp: 119350

Ion	Ratio	Lower	Upper
45	100		
77	11.8	0.0	31.5
79	9.1	0.0	29.1







#17
2-Methylphenol
Concen: 5.676 ng
RT: 7.122 min Scan# 855
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:107 Resp: 59770

Ion	Ratio	Lower	Upper
107	100		
108	111.2	91.3	136.9
77	43.6	37.1	55.7
79	42.5	37.0	55.6







#18
Hexachloroethane
Concen: 5.458 ng
RT: 7.404 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:117 Resp: 27566

Ion	Ratio	Lower	Upper
117	100		
119	99.1	77.0	115.4
201	101.0	76.2	114.4







#19
n-Nitroso-di-n-propylamine
Concen: 5.686 ng
RT: 7.287 min Scan# 883
Delta R.T. -0.017 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion: 70 Resp: 53958

Ion	Ratio	Lower	Upper
70	100		
42	68.0	51.5	77.3
101	9.2	7.5	11.3
130	19.0	15.5	23.3







#20
3+4-Methylphenols
Concen: 5.890 ng
RT: 7.275 min Scan# 8
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:107 Resp: 79329

Ion	Ratio	Lower	Upper
107	100		
108	89.7	69.8	109.8
77	54.7	50.5	90.5
79	28.9	10.2	50.2







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.169 min Scan# 1033
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:136 Resp: 737062

Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.6	12.8
54	10.7	8.4	12.6
68	6.9	5.1	7.7







#22
Acetophenone
Concen: 5.900 ng
RT: 7.287 min Scan# 8
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:105 Resp: 106522

Ion	Ratio	Lower	Upper
105	100		
71	8.6	5.9	8.9
51	36.4	29.4	44.2
120	20.4	17.5	26.3







#23
Nitrobenzene-d5
Concen: 10.109 ng
RT: 7.445 min Scan# 910
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion: 82 Resp: 134439

Ion	Ratio	Lower	Upper
82	100		
128	42.0	33.4	50.0
54	60.5	47.8	71.8







#24
Nitrobenzene
Concen: 5.269 ng
RT: 7.463 min Scan# 9
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion: 77 Resp: 76831

Ion	Ratio	Lower	Upper
77	100		
123	41.4	35.7	53.5
65	14.0	11.5	17.3







#25
Isophorone
Concen: 5.510 ng
RT: 7.698 min Scan# 953
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion: 82 Resp: 139084

Ion	Ratio	Lower	Upper
82	100		
95	6.8	5.5	8.3
138	16.5	13.2	19.8







#26
2-Nitrophenol
Concen: 4.157 ng
RT: 7.781 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:139 Resp: 22867

Ion	Ratio	Lower	Upper
139	100		
109	33.3	25.1	37.7
65	63.8	47.5	71.3







#27
2,4-Dimethylphenol
Concen: 5.716 ng
RT: 7.810 min Scan# 972
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:122 Resp: 52430

Ion	Ratio	Lower	Upper
122	100		
107	112.5	91.9	137.9
121	54.7	46.2	69.4







#28
bis(2-Chloroethoxy)methane
Concen: 5.643 ng
RT: 7.916 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion: 93 Resp: 86298

Ion	Ratio	Lower	Upper
93	100		
95	32.5	26.0	39.0
123	10.6	9.8	14.8







#29
2,4-Dichlorophenol
Concen: 5.434 ng
RT: 8.016 min Scan# 1007
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:162 Resp: 56772

Ion	Ratio	Lower	Upper
162	100		
164	63.6	43.9	83.9
98	39.2	19.9	59.9







#30
1,2,4-Trichlorobenzene
Concen: 5.592 ng
RT: 8.110 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:180 Resp: 64640

Ion	Ratio	Lower	Upper
180	100		
182	98.8	80.5	120.7
145	31.8	25.4	38.0







#31
Naphthalene
Concen: 5.859 ng
RT: 8.192 min Scan# 1037
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:128 Resp: 222725

Ion	Ratio	Lower	Upper
128	100		
129	10.7	8.9	13.3
127	12.9	10.6	16.0







#33
4-Chloroaniline
Concen: 5.647 ng
RT: 8.234 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:127 Resp: 73196

Ion	Ratio	Lower	Upper
127	100		
129	32.9	25.4	38.2
65	34.5	28.4	42.6
92	20.6	16.6	25.0







#34
Hexachlorobutadiene
Concen: 5.670 ng
RT: 8.310 min Scan# 1057
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:225 Resp: 41186

Ion	Ratio	Lower	Upper
225	100		
223	62.1	50.6	76.0
227	61.9	51.8	77.8







#35
Caprolactam
Concen: 5.114 ng
RT: 8.557 min Scan# 1
Delta R.T. -0.053 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:113 Resp: 16989

Ion	Ratio	Lower	Upper
113	100		
55	212.2	189.2	229.2
56	156.7	132.5	172.5







#36
4-Chloro-3-methylphenol
Concen: 5.438 ng
RT: 8.704 min Scan# 1124
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:107 Resp: 62903

Ion	Ratio	Lower	Upper
107	100		
144	24.3	19.1	28.7
142	72.0	57.2	85.8







#37
2-Methylnaphthalene
Concen: 5.854 ng
RT: 8.881 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:142 Resp: 136292

Ion	Ratio	Lower	Upper
142	100		
141	87.8	70.8	106.2







#38
1-Methylnaphthalene
Concen: 5.856 ng
RT: 8.981 min Scan# 1171
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:142 Resp: 133747

Ion	Ratio	Lower	Upper
142	100		
141	91.7	73.5	110.3







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.928 min Scan# 1
Delta R.T. 0.001 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:164 Resp: 414324

Ion	Ratio	Lower	Upper
164	100		
162	101.3	81.0	121.4
160	44.1	35.4	53.0







#40
1,2,4,5-Tetrachlorobenzene
Concen: 5.640 ng
RT: 9.045 min Scan# 1182
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:216 Resp: 63046

Ion	Ratio	Lower	Upper
216	100		
214	79.3	63.4	95.0
179	21.4	17.2	25.8
108	18.7	15.6	23.4







#41
Hexachlorocyclopentadiene
Concen: 4.916 ng
RT: 9.039 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:237 Resp: 19121

Ion	Ratio	Lower	Upper
237	100		
235	65.2	42.9	82.9
272	12.4	0.0	32.3







#42
2,4,6-Tribromophenol
Concen: 10.755 ng
RT: 10.710 min Scan# 1465
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:330 Resp: 41682

Ion	Ratio	Lower	Upper
330	100		
332	96.2	78.1	117.1
141	32.3	26.6	39.8







#43
2,4,6-Trichlorophenol
Concen: 5.307 ng
RT: 9.151 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:196 Resp: 41997

Ion	Ratio	Lower	Upper
196	100		
198	92.8	76.2	114.2
200	28.8	24.0	36.0







#44
2,4,5-Trichlorophenol
Concen: 5.399 ng
RT: 9.187 min Scan# 1206
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:196 Resp: 44086

Ion	Ratio	Lower	Upper
196	100		
198	96.0	75.8	113.6
97	51.2	41.8	62.6
132	29.1	22.9	34.3







#45
2-Fluorobiphenyl
Concen: 12.492 ng
RT: 9.245 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:172 Resp: 313171

Ion	Ratio	Lower	Upper
172	100		
171	35.6	28.6	43.0
170	23.5	19.1	28.7







#46
1,1'-Biphenyl
Concen: 5.891 ng
RT: 9.345 min Scan# 1233
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:154 Resp: 169920

Ion	Ratio	Lower	Upper
154	100		
153	39.7	20.3	60.3
76	14.1	0.0	33.9







#47
2-Chloronaphthalene
Concen: 5.745 ng
RT: 9.369 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:162 Resp: 133456

Ion	Ratio	Lower	Upper
162	100		
127	41.4	33.8	50.6
164	32.7	26.6	39.8







#48
2-Nitroaniline
Concen: 4.363 ng
RT: 9.457 min Scan# 1252
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion: 65 Resp: 30999

Ion	Ratio	Lower	Upper
65	100		
92	65.6	50.9	76.3
138	91.1	73.4	110.0







#49
Acenaphthylene
Concen: 5.717 ng
RT: 9.786 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:152 Resp: 192014

Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.9	23.9
153	13.4	10.7	16.1







#50
Dimethylphthalate
Concen: 5.661 ng
RT: 9.639 min Scan# 1283
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:163 Resp: 146096

Ion	Ratio	Lower	Upper
163	100		
194	4.3	3.4	5.0
164	10.1	8.2	12.2







#51
2,6-Dinitrotoluene
Concen: 4.741 ng
RT: 9.698 min Scan# 1
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:165 Resp: 26490

Ion	Ratio	Lower	Upper
165	100		
63	66.6	50.6	75.8
89	61.8	47.5	71.3







#52
Acenaphthene
Concen: 5.721 ng
RT: 9.957 min Scan# 1337
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:154 Resp: 124288

Ion	Ratio	Lower	Upper
154	100		
153	112.9	89.8	134.6
152	53.0	42.4	63.6







#53
3-Nitroaniline
Concen: 4.858 ng
RT: 9.863 min Scan# 1
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:138 Resp: 27995

Ion	Ratio	Lower	Upper
138	100		
108	17.3	10.2	15.2#
92	128.5	98.3	147.5







#55
Dibenzofuran
Concen: 5.870 ng
RT: 10.128 min Scan# 1366
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:168 Resp: 182985

Ion	Ratio	Lower	Upper
168	100		
139	38.5	31.8	47.6
169	14.3	11.6	17.4







#56
4-Nitrophenol
Concen: 4.376 ng
RT: 10.004 min Scan# 1
Delta R.T. -0.017 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:139 Resp: 19402

Ion	Ratio	Lower	Upper
139	100		
109	75.6	53.6	93.6
65	124.8	87.8	127.8







#57
2,4-Dinitrotoluene
Concen: 4.226 ng
RT: 10.098 min Scan# 1361
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:165 Resp: 29038

Ion	Ratio	Lower	Upper
165	100		
63	55.0	40.1	60.1
89	83.3	60.4	90.6
182	2.5	2.0	3.0







#58
Fluorene
Concen: 6.146 ng
RT: 10.469 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:166 Resp: 145913

Ion	Ratio	Lower	Upper
166	100		
165	97.5	78.5	117.7
167	13.2	10.9	16.3







#59
2,3,4,6-Tetrachlorophenol
Concen: 5.411 ng
RT: 10.239 min Scan# 1385
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:232 Resp: 34636

Ion	Ratio	Lower	Upper
232	100		
131	43.7	37.0	55.6
130	2.3	1.9	2.9
166	29.5	24.2	36.2







#60
Diethylphthalate
Concen: 5.583 ng
RT: 10.339 min Scan# 14
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:149 Resp: 141466

Ion	Ratio	Lower	Upper
149	100		
177	22.0	17.9	26.9
150	12.0	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 6.033 ng
RT: 10.469 min Scan# 1424
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:204 Resp: 72331

Ion	Ratio	Lower	Upper
204	100		
206	33.3	26.3	39.5
141	56.2	48.6	73.0







#62
4-Nitroaniline
Concen: 4.731 ng
RT: 10.469 min Scan# 14
Delta R.T. -0.017 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:138 Resp: 25747

Ion	Ratio	Lower	Upper
138	100		
92	57.8	35.7	75.7
108	81.3	57.5	97.5







#63
Azobenzene
Concen: 5.627 ng
RT: 10.622 min Scan# 1450
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion: 77 Resp: 151158

Ion	Ratio	Lower	Upper
77	100		
182	23.2	3.3	43.3
105	21.5	1.2	41.2
51	32.4	13.3	53.3







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:188 Resp: 741430

Ion	Ratio	Lower	Upper
188	100		
94	10.2	7.9	11.9
80	11.4	9.0	13.4







#66
n-Nitrosodiphenylamine
Concen: 5.609 ng
RT: 10.581 min Scan# 1443
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:169 Resp: 124472

Ion	Ratio	Lower	Upper
169	100		
168	61.9	49.8	74.6
167	35.0	28.0	42.0







#67
4-Bromophenyl-phenylether
Concen: 5.591 ng
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:248 Resp: 42707

Ion	Ratio	Lower	Upper
248	100		
250	93.5	79.0	118.6
141	71.1	61.7	92.5







#68
Hexachlorobenzene
Concen: 5.575 ng
RT: 11.016 min Scan# 1517
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:284 Resp: 47895

Ion	Ratio	Lower	Upper
284	100		
142	35.4	26.9	40.3
249	31.8	25.3	37.9







#69
Atrazine
Concen: 5.835 ng
RT: 11.104 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:200 Resp: 35075

Ion	Ratio	Lower	Upper
200	100		
173	26.4	7.7	47.7
215	48.2	27.7	67.7







#70
Pentachlorophenol
Concen: 4.178 ng
RT: 11.210 min Scan# 1550
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:266 Resp: 21781

Ion	Ratio	Lower	Upper
266	100		
268	68.6	50.4	75.6
264	68.9	50.6	75.8







#71
Phenanthrene
Concen: 5.931 ng
RT: 11.433 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:178 Resp: 207716

Ion	Ratio	Lower	Upper
178	100		
176	18.6	15.8	23.6
179	14.9	12.6	18.8







#72
Anthracene
Concen: 5.871 ng
RT: 11.486 min Scan# 1597
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:178 Resp: 200621

Ion	Ratio	Lower	Upper
178	100		
176	18.4	15.3	22.9
179	15.0	12.4	18.6







#73
Carbazole
Concen: 5.845 ng
RT: 11.639 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:167 Resp: 185743

Ion	Ratio	Lower	Upper
167	100		
166	21.4	17.0	25.6
139	12.7	10.8	16.2







#74
Di-n-butylphthalate
Concen: 5.574 ng
RT: 11.975 min Scan# 1680
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:149 Resp: 204647

Ion	Ratio	Lower	Upper
149	100		
150	8.9	7.5	11.3
104	5.0	4.1	6.1







#75
Fluoranthene
Concen: 6.017 ng
RT: 12.622 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:202 Resp: 213038

Ion	Ratio	Lower	Upper
202	100		
101	12.2	0.0	31.9
203	17.5	0.0	37.5







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.051 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:240 Resp: 477223

Ion	Ratio	Lower	Upper
240	100		
120	11.1	9.4	14.2
236	26.2	20.9	31.3







#77
Benzidine
Concen: 7.022 ng
RT: 12.745 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:184 Resp: 54501

Ion	Ratio	Lower	Upper
184	100		
185	15.5	12.0	18.0
183	11.4	9.4	14.2







#78
Pyrene
Concen: 5.404 ng
RT: 12.851 min Scan# 1829
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:202 Resp: 225742

Ion	Ratio	Lower	Upper
202	100		
200	21.3	17.2	25.8
203	17.4	14.2	21.4







#79
Terphenyl-d14
Concen: 11.248 ng
RT: 12.992 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:244 Resp: 329424

Ion	Ratio	Lower	Upper
244	100		
212	7.0	5.7	8.5
122	11.6	8.6	13.0







#80
Butylbenzylphthalate
Concen: 4.671 ng
RT: 13.474 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:149 Resp: 58502

Ion	Ratio	Lower	Upper
149	100		
91	74.4	59.4	89.2
206	20.3	16.4	24.6







#81
Benzo(a)anthracene
Concen: 5.473 ng
RT: 14.039 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:228 Resp: 170026

Ion	Ratio	Lower	Upper
228	100		
226	26.6	21.6	32.4
229	19.9	16.1	24.1







#82
3,3'-Dichlorobenzidine
Concen: 4.851 ng
RT: 14.004 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:252 Resp: 43937

Ion	Ratio	Lower	Upper
252	100		
254	63.3	50.8	76.2
126	11.1	8.9	13.3







#83
Chrysene
Concen: 5.551 ng
RT: 14.074 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:228 Resp: 158121

Ion	Ratio	Lower	Upper
228	100		
226	29.3	24.1	36.1
229	18.6	15.8	23.6







#84
Bis(2-ethylhexyl)phthalate
Concen: 4.419 ng
RT: 14.033 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:149 Resp: 62085

Ion	Ratio	Lower	Upper
149	100		
167	24.8	21.4	32.0
279	3.7	3.0	4.4







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.527 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:264 Resp: 377685

Ion	Ratio	Lower	Upper
264	100		
260	24.5	19.4	29.2
265	21.5	17.4	26.0







#87
Indeno(1,2,3-cd)pyrene
Concen: 4.699 ng
RT: 17.010 min Scan# 2536
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:276 Resp: 114185

Ion	Ratio	Lower	Upper
276	100		
138	27.1	20.4	30.6
277	25.3	20.2	30.4







#88
Benzo(b)fluoranthene
Concen: 5.405 ng
RT: 15.092 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:252 Resp: 124353

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.6	26.4
125	9.7	7.9	11.9







#89
Benzo(k)fluoranthene
Concen: 5.772 ng
RT: 15.121 min Scan# 2215
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:252 Resp: 114525

Ion	Ratio	Lower	Upper
252	100		
253	21.4	17.6	26.4
125	10.0	7.9	11.9







#90
Benzo(a)pyrene
Concen: 5.137 ng
RT: 15.463 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:252 Resp: 97243

Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.3	25.9
125	10.9	8.7	13.1







#91
Dibenzo(a,h)anthracene
Concen: 4.757 ng
RT: 17.033 min Scan# 2540
Delta R.T. -0.012 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Tgt Ion:278 Resp: 96416

Ion	Ratio	Lower	Upper
278	100		
139	16.6	13.1	19.7
279	24.4	19.0	28.4






#92
Benzo(g,h,i)perylene
Concen: 4.805 ng
RT: 17.457 min Scan# 2
Delta R.T. -0.023 min
Lab File: BF139845.D
Acq: 18 Oct 2024 10:55

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:276 Resp: 97292

Ion	Ratio	Lower	Upper
276	100		
277	23.1	19.0	28.6
138	21.3	17.4	26.0



Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139846.D
Acq On : 18 Oct 2024 11:23
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Quant Time: Oct 18 14:17:15 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.887	152	188831	20.000	ng		0.00
21) Naphthalene-d8	8.169	136	733621	20.000	ng		0.00
39) Acenaphthene-d10	9.928	164	411556	20.000	ng		0.00
64) Phenanthrene-d10	11.410	188	732661	20.000	ng		0.00
76) Chrysene-d12	14.051	240	452443	20.000	ng		0.00
86) Perylene-d12	15.527	264	375863	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.493	112	264054	21.891	ng		0.00
7) Phenol-d6	6.498	99	343334	21.977	ng		-0.01
23) Nitrobenzene-d5	7.445	82	268128	20.257	ng		0.00
42) 2,4,6-Tribromophenol	10.710	330	80655	20.952	ng		0.00
45) 2-Fluorobiphenyl	9.245	172	574563	23.073	ng		0.00
79) Terphenyl-d14	12.998	244	603787	21.745	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.681	88	58969	10.485	ng		97
3) Pyridine	3.440	79	148260	10.636	ng		97
4) n-Nitrosodimethylamine	3.375	42	75504	10.081	ng		96
6) Aniline	6.546	93	155680	10.692	ng		98
8) 2-Chlorophenol	6.669	128	134290	10.890	ng		99
9) Benzaldehyde	6.440	77	107330	12.213	ng		98
10) Phenol	6.510	94	172965	10.549	ng		98
11) bis(2-Chloroethyl)ether	6.616	93	134367	10.794	ng		100
12) 1,3-Dichlorobenzene	6.828	146	156338	11.045	ng		99
13) 1,4-Dichlorobenzene	6.904	146	154957	10.957	ng		99
14) 1,2-Dichlorobenzene	7.057	146	149129	11.299	ng		98
15) Benzyl Alcohol	7.022	79	122669	10.704	ng		99
16) 2,2'-oxybis(1-Chloropr...	7.163	45	227443	10.715	ng		99
17) 2-Methylphenol	7.128	107	109901	10.452	ng		98
18) Hexachloroethane	7.404	117	53892	10.686	ng		99
19) n-Nitroso-di-n-propyla...	7.293	70	100635	10.621	ng		98
20) 3+4-Methylphenols	7.275	107	148497	11.041	ng	#	90
22) Acetophenone	7.293	105	195595	10.885	ng		96
24) Nitrobenzene	7.463	77	147356	10.154	ng		98
25) Isophorone	7.698	82	260247	10.359	ng		98
26) 2-Nitrophenol	7.781	139	50159	9.162	ng		100
27) 2,4-Dimethylphenol	7.810	122	94947	10.400	ng		99
28) bis(2-Chloroethoxy)met...	7.916	93	161423	10.605	ng		99
29) 2,4-Dichlorophenol	8.016	162	109102	10.492	ng		100
30) 1,2,4-Trichlorobenzene	8.110	180	121232	10.537	ng		99
31) Naphthalene	8.192	128	411053	10.864	ng		99
32) Benzoic acid	7.869	122	64294	8.075	ng		99
33) 4-Chloroaniline	8.234	127	140021	10.853	ng		100
34) Hexachlorobutadiene	8.310	225	75740	10.477	ng		99
35) Caprolactam	8.569	113	33908	10.256	ng		99
36) 4-Chloro-3-methylphenol	8.704	107	120339	10.452	ng		98
37) 2-Methylnaphthalene	8.881	142	252554	10.899	ng		99
38) 1-Methylnaphthalene	8.981	142	250609	11.023	ng		99
40) 1,2,4,5-Tetrachloroben...	9.045	216	119227	10.738	ng		99
41) Hexachlorocyclopentadiene	9.039	237	39754	10.290	ng		98
43) 2,4,6-Trichlorophenol	9.151	196	78578	9.996	ng		99

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139846.D
Acq On : 18 Oct 2024 11:23
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Quant Time: Oct 18 14:17:15 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.187	196	87482	10.787	ng		100
46) 1,1'-Biphenyl	9.345	154	316146	11.035	ng		99
47) 2-Chloronaphthalene	9.369	162	252140	10.927	ng		98
48) 2-Nitroaniline	9.457	65	65425	9.270	ng		98
49) Acenaphthylene	9.787	152	361398	10.833	ng		99
50) Dimethylphthalate	9.639	163	276645	10.792	ng		99
51) 2,6-Dinitrotoluene	9.698	165	54824	9.877	ng		95
52) Acenaphthene	9.957	154	233749	10.832	ng		98
53) 3-Nitroaniline	9.869	138	56574	9.884	ng	#	98
54) 2,4-Dinitrophenol	9.969	184	11577	11.138	ng	#	1
55) Dibenzofuran	10.128	168	341446	11.028	ng		98
56) 4-Nitrophenol	10.010	139	42688	9.694	ng		98
57) 2,4-Dinitrotoluene	10.098	165	64666	9.474	ng		94
58) Fluorene	10.475	166	269323	11.420	ng		99
59) 2,3,4,6-Tetrachlorophenol	10.239	232	66170	10.408	ng		97
60) Diethylphthalate	10.345	149	269178	10.694	ng		100
61) 4-Chlorophenyl-phenyle...	10.469	204	131191	11.016	ng		97
62) 4-Nitroaniline	10.475	138	53288	9.857	ng		98
63) Azobenzene	10.622	77	284960	10.679	ng		100
65) 4,6-Dinitro-2-methylph...	10.504	198	20120	6.732	ng		97
66) n-Nitrosodiphenylamine	10.581	169	234741	10.705	ng		99
67) 4-Bromophenyl-phenylether	10.957	248	79420	10.522	ng		98
68) Hexachlorobenzene	11.016	284	89927	10.594	ng		98
69) Atrazine	11.104	200	64267	10.819	ng		99
70) Pentachlorophenol	11.210	266	50361	9.775	ng		100
71) Phenanthrene	11.433	178	379233	10.958	ng		99
72) Anthracene	11.486	178	371391	10.999	ng		99
73) Carbazole	11.639	167	353253	11.249	ng		99
74) Di-n-butylphthalate	11.975	149	392482	10.818	ng		99
75) Fluoranthene	12.622	202	405919	11.602	ng		99
77) Benzidine	12.745	184	104253	14.168	ng		99
78) Pyrene	12.851	202	413606	10.444	ng		99
80) Butylbenzylphthalate	13.474	149	116053	9.774	ng		100
81) Benzo(a)anthracene	14.039	228	306594	10.410	ng		99
82) 3,3'-Dichlorobenzidine	14.004	252	84205	9.806	ng		99
83) Chrysene	14.074	228	281467	10.423	ng		99
84) Bis(2-ethylhexyl)phtha...	14.033	149	121932	9.153	ng		99
85) Di-n-octyl phthalate	14.651	149	177824	7.339	ng		99
87) Indeno(1,2,3-cd)pyrene	17.010	276	236914	9.798	ng		99
88) Benzo(b)fluoranthene	15.092	252	232988	10.176	ng		99
89) Benzo(k)fluoranthene	15.121	252	221219	11.203	ng		99
90) Benzo(a)pyrene	15.463	252	192493	10.217	ng		100
91) Dibenzo(a,h)anthracene	17.033	278	199946	9.912	ng		99
92) Benzo(g,h,i)perylene	17.457	276	196609	9.758	ng		99


(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139846.D
Acq On : 18 Oct 2024 11:23
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Quant Time: Oct 18 14:17:15 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.887 min Scan# 8
Delta R.T. -0.005 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:152 Resp: 188831

Ion	Ratio	Lower	Upper
152	100		
150	157.0	130.2	195.2
115	64.1	51.4	77.2







#2
1,4-Dioxane
Concen: 10.485 ng
RT: 2.681 min Scan# 100
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion: 88 Resp: 58969

Ion	Ratio	Lower	Upper
88	100		
58	87.5	73.3	109.9
43	34.3	27.1	40.7







#3
Pyridine
Concen: 10.636 ng
RT: 3.440 min Scan# 21
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 79 Resp: 148260

Ion	Ratio	Lower	Upper
79	100		
52	88.5	68.5	102.7
51	41.6	32.3	48.5







#4
n-Nitrosodimethylamine
Concen: 10.081 ng
RT: 3.375 min Scan# 218
Delta R.T. -0.023 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion: 42 Resp: 75504

Ion	Ratio	Lower	Upper
42	100		
74	115.8	96.2	144.4
44	8.2	6.2	9.2







#5
2-Fluorophenol
Concen: 21.891 ng
RT: 5.493 min Scan# 5
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:112 Resp: 264054

Ion	Ratio	Lower	Upper
112	100		
64	65.6	53.0	79.6
63	34.0	27.0	40.4







#6
Aniline
Concen: 10.692 ng
RT: 6.546 min Scan# 757
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion: 93 Resp: 155680

Ion	Ratio	Lower	Upper
93	100		
66	41.1	34.0	51.0
65	21.2	17.4	26.0







#7
Phenol-d6
Concen: 21.977 ng
RT: 6.498 min Scan# 7
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 99 Resp: 343334

Ion	Ratio	Lower	Upper
99	100		
42	20.3	16.7	25.1
71	32.7	27.7	41.5







#8
2-Chlorophenol
Concen: 10.890 ng
RT: 6.669 min Scan# 778
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:128 Resp: 134290

Ion	Ratio	Lower	Upper
128	100		
130	32.4	12.3	52.3
64	54.1	34.6	74.6







#9
Benzaldehyde
Concen: 12.213 ng
RT: 6.440 min Scan# 7
Delta R.T. 0.001 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 77 Resp: 107330

Ion	Ratio	Lower	Upper
77	100		
105	94.7	75.5	115.5
106	94.6	71.7	111.7







#10
Phenol
Concen: 10.549 ng
RT: 6.510 min Scan# 751
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion: 94 Resp: 172965

Ion	Ratio	Lower	Upper
94	100		
65	30.1	11.1	51.1
66	41.0	22.7	62.7







#11
bis(2-Chloroethyl)ether
Concen: 10.794 ng
RT: 6.616 min Scan# 7
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 93 Resp: 134367

Ion	Ratio	Lower	Upper
93	100		
63	85.2	65.0	105.0
95	31.8	12.4	52.4







#12
1,3-Dichlorobenzene
Concen: 11.045 ng
RT: 6.828 min Scan# 805
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:146 Resp: 156338

Ion	Ratio	Lower	Upper
146	100		
148	62.9	51.0	76.4
75	31.8	26.4	39.6







#13
1,4-Dichlorobenzene
Concen: 10.957 ng
RT: 6.904 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:146 Resp: 154957

Ion	Ratio	Lower	Upper
146	100		
148	63.7	51.8	77.8
111	43.7	34.6	51.8







#14
1,2-Dichlorobenzene
Concen: 11.299 ng
RT: 7.057 min Scan# 844
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:146 Resp: 149129

Ion	Ratio	Lower	Upper
146	100		
148	63.1	51.4	77.2
111	44.0	36.7	55.1







#15
Benzyl Alcohol
Concen: 10.704 ng
RT: 7.022 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 79 Resp: 122669

Ion	Ratio	Lower	Upper
79	100		
108	77.7	61.4	92.0
77	62.8	49.9	74.9







#16
2,2'-oxybis(1-Chloropropane)
Concen: 10.715 ng
RT: 7.163 min Scan# 862
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion: 45 Resp: 227443

Ion	Ratio	Lower	Upper
45	100		
77	11.2	0.0	31.5
79	8.7	0.0	29.1







#17
2-Methylphenol
Concen: 10.452 ng
RT: 7.128 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:107 Resp: 109901

Ion	Ratio	Lower	Upper
107	100		
108	116.1	91.3	136.9
77	44.4	37.1	55.7
79	45.6	37.0	55.6







#18
Hexachloroethane
Concen: 10.686 ng
RT: 7.404 min Scan# 903
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:117 Resp: 53892

Ion	Ratio	Lower	Upper
117	100		
119	94.7	77.0	115.4
201	96.4	76.2	114.4







#19
n-Nitroso-di-n-propylamine
Concen: 10.621 ng
RT: 7.293 min Scan# 8
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 70 Resp: 100635

Ion	Ratio	Lower	Upper
70	100		
42	66.4	51.5	77.3
101	9.6	7.5	11.3
130	19.4	15.5	23.3







#20
3+4-Methylphenols
Concen: 11.041 ng
RT: 7.275 min Scan# 881
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:107 Resp: 148497

Ion	Ratio	Lower	Upper
107	100		
108	90.1	69.8	109.8
77	49.6	50.5	90.5#
79	28.8	10.2	50.2







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.169 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:136 Resp: 733621

Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.6	12.8
54	10.9	8.4	12.6
68	6.6	5.1	7.7







#22
Acetophenone
Concen: 10.885 ng
RT: 7.293 min Scan# 884
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:105 Resp: 195595

Ion	Ratio	Lower	Upper
105	100		
71	8.5	5.9	8.9
51	33.6	29.4	44.2
120	22.2	17.5	26.3







#23
Nitrobenzene-d5
Concen: 20.257 ng
RT: 7.445 min Scan# 91
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 82 Resp: 268128

Ion	Ratio	Lower	Upper
82	100		
128	41.3	33.4	50.0
54	59.5	47.8	71.8







#24
Nitrobenzene
Concen: 10.154 ng
RT: 7.463 min Scan# 913
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion: 77 Resp: 147356

Ion	Ratio	Lower	Upper
77	100		
123	43.0	35.7	53.5
65	14.6	11.5	17.3







#25
Isophorone
Concen: 10.359 ng
RT: 7.698 min Scan# 9
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 82 Resp: 260247

Ion	Ratio	Lower	Upper
82	100		
95	6.6	5.5	8.3
138	15.7	13.2	19.8







#26
2-Nitrophenol
Concen: 9.162 ng
RT: 7.781 min Scan# 967
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:139 Resp: 50159

Ion	Ratio	Lower	Upper
139	100		
109	30.9	25.1	37.7
65	59.2	47.5	71.3







#27
2,4-Dimethylphenol
Concen: 10.400 ng
RT: 7.810 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:122 Resp: 94947

Ion	Ratio	Lower	Upper
122	100		
107	114.1	91.9	137.9
121	58.0	46.2	69.4







#28
bis(2-Chloroethoxy)methane
Concen: 10.605 ng
RT: 7.916 min Scan# 990
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion: 93 Resp: 161423

Ion	Ratio	Lower	Upper
93	100		
95	33.0	26.0	39.0
123	11.2	9.8	14.8







#29
2,4-Dichlorophenol
Concen: 10.492 ng
RT: 8.016 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:162 Resp: 109102

Ion	Ratio	Lower	Upper
162	100		
164	64.1	43.9	83.9
98	39.8	19.9	59.9







#30
1,2,4-Trichlorobenzene
Concen: 10.537 ng
RT: 8.110 min Scan# 1023
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:180 Resp: 121232

Ion	Ratio	Lower	Upper
180	100		
182	101.5	80.5	120.7
145	32.7	25.4	38.0







#31
Naphthalene
Concen: 10.864 ng
RT: 8.192 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:128 Resp: 411053

Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.9	13.3
127	13.0	10.6	16.0







#32
Benzoic acid
Concen: 8.075 ng
RT: 7.869 min Scan# 982
Delta R.T. -0.053 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:122 Resp: 64294

Ion	Ratio	Lower	Upper
122	100		
105	126.9	107.6	147.6
77	99.2	80.2	120.2







#33
4-Chloroaniline
Concen: 10.853 ng
RT: 8.234 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:127 Resp: 140021

Ion	Ratio	Lower	Upper
127	100		
129	31.8	25.4	38.2
65	35.4	28.4	42.6
92	20.4	16.6	25.0







#34
Hexachlorobutadiene
Concen: 10.477 ng
RT: 8.310 min Scan# 1057
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:225 Resp: 75740

Ion	Ratio	Lower	Upper
225	100		
223	63.0	50.6	76.0
227	64.3	51.8	77.8







#35
Caprolactam
Concen: 10.256 ng
RT: 8.569 min Scan# 11
Delta R.T. -0.041 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:113 Resp: 33908

Ion	Ratio	Lower	Upper
113	100		
55	208.3	189.2	229.2
56	151.4	132.5	172.5







#36
4-Chloro-3-methylphenol
Concen: 10.452 ng
RT: 8.704 min Scan# 1124
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:107 Resp: 120339

Ion	Ratio	Lower	Upper
107	100		
144	25.2	19.1	28.7
142	72.5	57.2	85.8







#37
2-Methylnaphthalene
Concen: 10.899 ng
RT: 8.881 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:142 Resp: 252554

Ion	Ratio	Lower	Upper
142	100		
141	89.7	70.8	106.2







#38
1-Methylnaphthalene
Concen: 11.023 ng
RT: 8.981 min Scan# 1171
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:142 Resp: 250609

Ion	Ratio	Lower	Upper
142	100		
141	91.3	73.5	110.3







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.928 min Scan# 1
Delta R.T. 0.001 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:164 Resp: 411556

Ion	Ratio	Lower	Upper
164	100		
162	100.2	81.0	121.4
160	43.4	35.4	53.0







#40
1,2,4,5-Tetrachlorobenzene
Concen: 10.738 ng
RT: 9.045 min Scan# 1182
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:216 Resp: 119227

Ion	Ratio	Lower	Upper
216	100		
214	78.1	63.4	95.0
179	21.2	17.2	25.8
108	18.7	15.6	23.4







#41
Hexachlorocyclopentadiene
Concen: 10.290 ng
RT: 9.039 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:237 Resp: 39754

Ion	Ratio	Lower	Upper
237	100		
235	61.3	42.9	82.9
272	12.3	0.0	32.3







#42
2,4,6-Tribromophenol
Concen: 20.952 ng
RT: 10.710 min Scan# 1465
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:330 Resp: 80655

Ion	Ratio	Lower	Upper
330	100		
332	95.9	78.1	117.1
141	34.2	26.6	39.8







#43
2,4,6-Trichlorophenol
Concen: 9.996 ng
RT: 9.151 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:196 Resp: 78578

Ion	Ratio	Lower	Upper
196	100		
198	95.7	76.2	114.2
200	29.8	24.0	36.0







#44
2,4,5-Trichlorophenol
Concen: 10.787 ng
RT: 9.187 min Scan# 1206
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:196 Resp: 87482

Ion	Ratio	Lower	Upper
196	100		
198	94.9	75.8	113.6
97	52.0	41.8	62.6
132	28.0	22.9	34.3







#45
2-Fluorobiphenyl
Concen: 23.073 ng
RT: 9.245 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:172 Resp: 574563

Ion	Ratio	Lower	Upper
172	100		
171	35.2	28.6	43.0
170	23.5	19.1	28.7







#46
1,1'-Biphenyl
Concen: 11.035 ng
RT: 9.345 min Scan# 1233
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:154 Resp: 316146

Ion	Ratio	Lower	Upper
154	100		
153	39.4	20.3	60.3
76	13.9	0.0	33.9







#47
2-Chloronaphthalene
Concen: 10.927 ng
RT: 9.369 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:162 Resp: 252140

Ion	Ratio	Lower	Upper
162	100		
127	40.9	33.8	50.6
164	31.8	26.6	39.8







#48
2-Nitroaniline
Concen: 9.270 ng
RT: 9.457 min Scan# 1252
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion: 65 Resp: 65425

Ion	Ratio	Lower	Upper
65	100		
92	65.5	50.9	76.3
138	92.7	73.4	110.0







#49
Acenaphthylene
Concen: 10.833 ng
RT: 9.787 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:152 Resp: 361398

Ion	Ratio	Lower	Upper
152	100		
151	19.9	15.9	23.9
153	12.7	10.7	16.1







#50
Dimethylphthalate
Concen: 10.792 ng
RT: 9.639 min Scan# 1283
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:163 Resp: 276645

Ion	Ratio	Lower	Upper
163	100		
194	3.7	3.4	5.0
164	10.0	8.2	12.2







#51
2,6-Dinitrotoluene
Concen: 9.877 ng
RT: 9.698 min Scan# 1
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:165 Resp: 54824

Ion	Ratio	Lower	Upper
165	100		
63	66.0	50.6	75.8
89	64.1	47.5	71.3







#52
Acenaphthene
Concen: 10.832 ng
RT: 9.957 min Scan# 1337
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:154 Resp: 233749

Ion	Ratio	Lower	Upper
154	100		
153	109.8	89.8	134.6
152	52.9	42.4	63.6







#53
3-Nitroaniline
Concen: 9.884 ng
RT: 9.869 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:138 Resp: 56574

Ion	Ratio	Lower	Upper
138	100		
108	15.7	10.2	15.2#
92	124.7	98.3	147.5







#54
2,4-Dinitrophenol
Concen: 11.138 ng
RT: 9.969 min Scan# 1339
Delta R.T. -0.005 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:184 Resp: 11577

Ion	Ratio	Lower	Upper
184	100		
63	142.4	84.7	127.1#
154	870.2	338.2	507.2#







#55
Dibenzofuran
Concen: 11.028 ng
RT: 10.128 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:168 Resp: 341446

Ion	Ratio	Lower	Upper
168	100		
139	38.2	31.8	47.6
169	13.9	11.6	17.4







#56
4-Nitrophenol
Concen: 9.694 ng
RT: 10.010 min Scan# 1346
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:139 Resp: 42688

Ion	Ratio	Lower	Upper
139	100		
109	74.3	53.6	93.6
65	111.6	87.8	127.8







#57
2,4-Dinitrotoluene
Concen: 9.474 ng
RT: 10.098 min Scan# 1
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:165 Resp: 64666

Ion	Ratio	Lower	Upper
165	100		
63	53.7	40.1	60.1
89	81.9	60.4	90.6
182	2.9	2.0	3.0







#58
Fluorene
Concen: 11.420 ng
RT: 10.475 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:166 Resp: 269323

Ion	Ratio	Lower	Upper
166	100		
165	97.4	78.5	117.7
167	13.4	10.9	16.3







#59
2,3,4,6-Tetrachlorophenol
Concen: 10.408 ng
RT: 10.239 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:232 Resp: 66170

Ion	Ratio	Lower	Upper
232	100		
131	44.4	37.0	55.6
130	2.3	1.9	2.9
166	29.0	24.2	36.2







#60
Diethylphthalate
Concen: 10.694 ng
RT: 10.345 min Scan# 1403
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:149 Resp: 269178

Ion	Ratio	Lower	Upper
149	100		
177	22.4	17.9	26.9
150	12.3	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 11.016 ng
RT: 10.469 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:204 Resp: 131191

Ion	Ratio	Lower	Upper
204	100		
206	32.5	26.3	39.5
141	58.0	48.6	73.0







#62
4-Nitroaniline
Concen: 9.857 ng
RT: 10.475 min Scan# 1425
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:138 Resp: 53288

Ion	Ratio	Lower	Upper
138	100		
92	54.6	35.7	75.7
108	75.6	57.5	97.5







#63
Azobenzene
Concen: 10.679 ng
RT: 10.622 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 77 Resp: 284960

Ion	Ratio	Lower	Upper
77	100		
182	23.3	3.3	43.3
105	21.3	1.2	41.2
51	33.8	13.3	53.3







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1584
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:188 Resp: 732661

Ion	Ratio	Lower	Upper
188	100		
94	10.5	7.9	11.9
80	11.5	9.0	13.4







#65
4,6-Dinitro-2-methylphenol
Concen: 6.732 ng
RT: 10.504 min Scan# 14
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:198 Resp: 20120

Ion	Ratio	Lower	Upper
198	100		
51	61.0	38.2	78.2
105	49.9	27.9	67.9







#66
n-Nitrosodiphenylamine
Concen: 10.705 ng
RT: 10.581 min Scan# 1443
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:169 Resp: 234741

Ion	Ratio	Lower	Upper
169	100		
168	63.0	49.8	74.6
167	34.3	28.0	42.0







#67
4-Bromophenyl-phenylether
Concen: 10.522 ng
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:248 Resp: 79420

Ion	Ratio	Lower	Upper
248	100		
250	96.8	79.0	118.6
141	74.5	61.7	92.5







#68
Hexachlorobenzene
Concen: 10.594 ng
RT: 11.016 min Scan# 1517
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:284 Resp: 89927

Ion	Ratio	Lower	Upper
284	100		
142	35.1	26.9	40.3
249	32.2	25.3	37.9







#69
Atrazine
Concen: 10.819 ng
RT: 11.104 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:200 Resp: 64267

Ion	Ratio	Lower	Upper
200	100		
173	27.4	7.7	47.7
215	46.8	27.7	67.7







#70
Pentachlorophenol
Concen: 9.775 ng
RT: 11.210 min Scan# 1550
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:266 Resp: 50361

Ion	Ratio	Lower	Upper
266	100		
268	63.2	50.4	75.6
264	62.8	50.6	75.8







#71
Phenanthrene
Concen: 10.958 ng
RT: 11.433 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:178 Resp: 379233

Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.8	23.6
179	15.4	12.6	18.8







#72
Anthracene
Concen: 10.999 ng
RT: 11.486 min Scan# 1597
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:178 Resp: 371391

Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.3	22.9
179	15.0	12.4	18.6







#73
Carbazole
Concen: 11.249 ng
RT: 11.639 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:167 Resp: 353253

Ion	Ratio	Lower	Upper
167	100		
166	21.0	17.0	25.6
139	12.8	10.8	16.2







#74
Di-n-butylphthalate
Concen: 10.818 ng
RT: 11.975 min Scan# 1680
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:149 Resp: 392482

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.5	11.3
104	4.9	4.1	6.1







#75
Fluoranthene
Concen: 11.602 ng
RT: 12.622 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:202 Resp: 405919

Ion	Ratio	Lower	Upper
202	100		
101	12.5	0.0	31.9
203	17.2	0.0	37.5







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.051 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:240 Resp: 452443

Ion	Ratio	Lower	Upper
240	100		
120	11.4	9.4	14.2
236	26.0	20.9	31.3







#77
Benzidine
Concen: 14.168 ng
RT: 12.745 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:184 Resp: 104253

Ion	Ratio	Lower	Upper
184	100		
185	15.0	12.0	18.0
183	11.4	9.4	14.2







#78
Pyrene
Concen: 10.444 ng
RT: 12.851 min Scan# 1829
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:202 Resp: 413606

Ion	Ratio	Lower	Upper
202	100		
200	21.0	17.2	25.8
203	17.4	14.2	21.4







#79
Terphenyl-d14
Concen: 21.745 ng
RT: 12.998 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:244 Resp: 603787

Ion	Ratio	Lower	Upper
244	100		
212	6.8	5.7	8.5
122	10.6	8.6	13.0







#80
Butylbenzylphthalate
Concen: 9.774 ng
RT: 13.474 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:149 Resp: 116053

Ion	Ratio	Lower	Upper
149	100		
91	74.4	59.4	89.2
206	20.1	16.4	24.6







#81
Benzo(a)anthracene
Concen: 10.410 ng
RT: 14.039 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:228 Resp: 306594

Ion	Ratio	Lower	Upper
228	100		
226	26.7	21.6	32.4
229	19.5	16.1	24.1







#82
3,3'-Dichlorobenzidine
Concen: 9.806 ng
RT: 14.004 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:252 Resp: 84205

Ion	Ratio	Lower	Upper
252	100		
254	64.7	50.8	76.2
126	11.3	8.9	13.3







#83
Chrysene
Concen: 10.423 ng
RT: 14.074 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:228 Resp: 281467

Ion	Ratio	Lower	Upper
228	100		
226	29.4	24.1	36.1
229	19.6	15.8	23.6







#84
Bis(2-ethylhexyl)phthalate
Concen: 9.153 ng
RT: 14.033 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:149 Resp: 121932

Ion	Ratio	Lower	Upper
149	100		
167	26.0	21.4	32.0
279	3.8	3.0	4.4







#85
Di-n-octyl phthalate
Concen: 7.339 ng
RT: 14.651 min Scan# 21
Delta R.T. -0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:149 Resp: 177824

Ion	Ratio	Lower	Upper
149	100		
167	1.4	1.2	1.8
43	11.8	9.8	14.6







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.527 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:264 Resp: 375863

Ion	Ratio	Lower	Upper
264	100		
260	23.9	19.4	29.2
265	21.3	17.4	26.0







#87
Indeno(1,2,3-cd)pyrene
Concen: 9.798 ng
RT: 17.010 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:276 Resp: 236914

Ion	Ratio	Lower	Upper
276	100		
138	25.8	20.4	30.6
277	25.0	20.2	30.4







#88
Benzo(b)fluoranthene
Concen: 10.176 ng
RT: 15.092 min Scan# 2210
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:252 Resp: 232988

Ion	Ratio	Lower	Upper
252	100		
253	21.4	17.6	26.4
125	10.3	7.9	11.9







#89
Benzo(k)fluoranthene
Concen: 11.203 ng
RT: 15.121 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:252 Resp: 221219

Ion	Ratio	Lower	Upper
252	100		
253	21.5	17.6	26.4
125	10.2	7.9	11.9







#90
Benzo(a)pyrene
Concen: 10.217 ng
RT: 15.463 min Scan# 2273
Delta R.T. -0.006 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:252 Resp: 192493

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.3	25.9
125	11.3	8.7	13.1







#91
Dibenzo(a,h)anthracene
Concen: 9.912 ng
RT: 17.033 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:278 Resp: 199946

Ion	Ratio	Lower	Upper
278	100		
139	16.0	13.1	19.7
279	23.2	19.0	28.4






#92
Benzo(g,h,i)perylene
Concen: 9.758 ng
RT: 17.457 min Scan# 2612
Delta R.T. -0.023 min
Lab File: BF139846.D
Acq: 18 Oct 2024 11:23

Tgt Ion:276 Resp: 196609

Ion	Ratio	Lower	Upper
276	100		
277	23.3	19.0	28.6
138	21.4	17.4	26.0



Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139847.D
Acq On : 18 Oct 2024 11:52
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Quant Time: Oct 18 14:18:00 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.893	152	192582	20.000	ng	0.00
21) Naphthalene-d8	8.175	136	735124	20.000	ng	0.00
39) Acenaphthene-d10	9.928	164	413023	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	724973	20.000	ng	0.00
76) Chrysene-d12	14.051	240	402251	20.000	ng	0.00
86) Perylene-d12	15.527	264	371811	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.498	112	512648	41.673	ng	0.00
7) Phenol-d6	6.504	99	658086	41.304	ng	0.00
23) Nitrobenzene-d5	7.451	82	546550	41.206	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	159165	41.199	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1057523	42.316	ng	0.00
79) Terphenyl-d14	12.998	244	1078101	43.673	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.681	88	116188	20.257	ng	98
3) Pyridine	3.440	79	293019	20.611	ng	98
4) n-Nitrosodimethylamine	3.381	42	153824	20.138	ng	99
6) Aniline	6.551	93	311561	20.982	ng	99
8) 2-Chlorophenol	6.669	128	261502	20.794	ng	98
9) Benzaldehyde	6.440	77	200586	22.379	ng	98
10) Phenol	6.516	94	338933	20.269	ng	98
11) bis(2-Chloroethyl)ether	6.622	93	261690	20.612	ng	99
12) 1,3-Dichlorobenzene	6.828	146	297432	20.603	ng	99
13) 1,4-Dichlorobenzene	6.910	146	300107	20.808	ng	98
14) 1,2-Dichlorobenzene	7.063	146	284645	21.146	ng	99
15) Benzyl Alcohol	7.022	79	242106	20.715	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.163	45	453174	20.933	ng	99
17) 2-Methylphenol	7.128	107	218455	20.371	ng	99
18) Hexachloroethane	7.404	117	105775	20.566	ng	98
19) n-Nitroso-di-n-propyla...	7.293	70	197390	20.427	ng	97
20) 3+4-Methylphenols	7.281	107	292795	21.346	ng	96
22) Acetophenone	7.293	105	379127	21.056	ng	99
24) Nitrobenzene	7.469	77	299625	20.604	ng	99
25) Isophorone	7.704	82	516119	20.502	ng	100
26) 2-Nitrophenol	7.787	139	110497	20.141	ng	97
27) 2,4-Dimethylphenol	7.816	122	187930	20.544	ng	98
28) bis(2-Chloroethoxy)met...	7.916	93	317497	20.816	ng	99
29) 2,4-Dichlorophenol	8.016	162	215409	20.673	ng	98
30) 1,2,4-Trichlorobenzene	8.110	180	238559	20.691	ng	100
31) Naphthalene	8.192	128	801971	21.153	ng	99
32) Benzoic acid	7.892	122	152028	19.054	ng	99
33) 4-Chloroaniline	8.239	127	270435	20.919	ng	98
34) Hexachlorobutadiene	8.310	225	149527	20.641	ng	100
35) Caprolactam	8.587	113	67624	20.411	ng	100
36) 4-Chloro-3-methylphenol	8.704	107	238461	20.668	ng	100
37) 2-Methylnaphthalene	8.881	142	489343	21.075	ng	100
38) 1-Methylnaphthalene	8.981	142	481797	21.149	ng	100
40) 1,2,4,5-Tetrachloroben...	9.045	216	232025	20.823	ng	99
41) Hexachlorocyclopentadiene	9.039	237	81579	21.041	ng	99
43) 2,4,6-Trichlorophenol	9.157	196	165112	20.930	ng	99

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139847.D
Acq On : 18 Oct 2024 11:52
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Quant Time: Oct 18 14:18:00 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.192	196	164517	20.213	ng		98
46) 1,1'-Biphenyl	9.345	154	612342	21.297	ng		99
47) 2-Chloronaphthalene	9.375	162	480916	20.767	ng		99
48) 2-Nitroaniline	9.463	65	146002	20.614	ng		98
49) Acenaphthylene	9.786	152	709161	21.182	ng		99
50) Dimethylphthalate	9.645	163	529745	20.592	ng		100
51) 2,6-Dinitrotoluene	9.704	165	114726	20.596	ng		98
52) Acenaphthene	9.963	154	449770	20.768	ng		99
53) 3-Nitroaniline	9.869	138	122186	21.271	ng		99
54) 2,4-Dinitrophenol	9.975	184	31801	19.148	ng	#	67
55) Dibenzofuran	10.134	168	652249	20.991	ng		98
56) 4-Nitrophenol	10.010	139	93060	21.057	ng		95
57) 2,4-Dinitrotoluene	10.104	165	139381	20.348	ng		98
58) Fluorene	10.475	166	496012	20.958	ng		100
59) 2,3,4,6-Tetrachlorophenol	10.245	232	134842	21.133	ng		98
60) Diethylphthalate	10.345	149	523458	20.723	ng		100
61) 4-Chlorophenyl-phenyle...	10.469	204	254950	21.331	ng		97
62) 4-Nitroaniline	10.481	138	111683	20.586	ng		98
63) Azobenzene	10.628	77	559837	20.906	ng		99
65) 4,6-Dinitro-2-methylph...	10.510	198	52180	17.645	ng		100
66) n-Nitrosodiphenylamine	10.581	169	449925	20.736	ng		99
67) 4-Bromophenyl-phenylether	10.957	248	153196	20.512	ng		98
68) Hexachlorobenzene	11.022	284	172091	20.488	ng		99
69) Atrazine	11.104	200	113089	19.240	ng		99
70) Pentachlorophenol	11.210	266	107595	21.106	ng		99
71) Phenanthrene	11.439	178	720363	21.036	ng		100
72) Anthracene	11.486	178	704417	21.083	ng		99
73) Carbazole	11.639	167	669328	21.540	ng		100
74) Di-n-butylphthalate	11.975	149	769779	21.442	ng		99
75) Fluoranthene	12.622	202	751337	21.703	ng		99
77) Benzidine	12.745	184	118035	18.042	ng		100
78) Pyrene	12.851	202	764077	21.700	ng		100
80) Butylbenzylphthalate	13.474	149	215683	20.432	ng		98
81) Benzo(a)anthracene	14.039	228	534525	20.413	ng		100
82) 3,3'-Dichlorobenzidine	14.004	252	156770	20.534	ng		100
83) Chrysene	14.074	228	496447	20.678	ng		99
84) Bis(2-ethylhexyl)phtha...	14.033	149	232200	19.606	ng		99
85) Di-n-octyl phthalate	14.651	149	373952	17.359	ng		100
87) Indeno(1,2,3-cd)pyrene	17.015	276	485858	20.312	ng		100
88) Benzo(b)fluoranthene	15.092	252	490566	21.659	ng		100
89) Benzo(k)fluoranthene	15.121	252	368835	18.883	ng		99
90) Benzo(a)pyrene	15.463	252	378617	20.316	ng		100
91) Dibenzo(a,h)anthracene	17.039	278	410285	20.561	ng		99
92) Benzo(g,h,i)perylene	17.468	276	405140	20.326	ng		99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139847.D
Acq On : 18 Oct 2024 11:52
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Quant Time: Oct 18 14:18:00 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.893 min Scan# 8
Delta R.T. 0.001 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:152 Resp: 192582

Ion	Ratio	Lower	Upper
152	100		
150	156.0	130.2	195.2
115	62.6	51.4	77.2







#2
1,4-Dioxane
Concen: 20.257 ng
RT: 2.681 min Scan# 100
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion: 88 Resp: 116188

Ion	Ratio	Lower	Upper
88	100		
58	88.9	73.3	109.9
43	33.6	27.1	40.7







#3
Pyridine
Concen: 20.611 ng
RT: 3.440 min Scan# 21
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 79 Resp: 293019

Ion	Ratio	Lower	Upper
79	100		
52	87.3	68.5	102.7
51	41.6	32.3	48.5







#4
n-Nitrosodimethylamine
Concen: 20.138 ng
RT: 3.381 min Scan# 219
Delta R.T. -0.017 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion: 42 Resp: 153824

Ion	Ratio	Lower	Upper
42	100		
74	118.8	96.2	144.4
44	8.0	6.2	9.2







#5
2-Fluorophenol
Concen: 41.673 ng
RT: 5.498 min Scan# 5
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:112 Resp: 512648

Ion	Ratio	Lower	Upper
112	100		
64	65.1	53.0	79.6
63	33.1	27.0	40.4







#6
Aniline
Concen: 20.982 ng
RT: 6.551 min Scan# 758
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion: 93 Resp: 311561

Ion	Ratio	Lower	Upper
93	100		
66	42.0	34.0	51.0
65	21.2	17.4	26.0







#7
Phenol-d6
Concen: 41.304 ng
RT: 6.504 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 99 Resp: 658086

Ion	Ratio	Lower	Upper
99	100		
42	20.5	16.7	25.1
71	33.8	27.7	41.5







#8
2-Chlorophenol
Concen: 20.794 ng
RT: 6.669 min Scan# 778
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:128 Resp: 261502

Ion	Ratio	Lower	Upper
128	100		
130	32.8	12.3	52.3
64	56.8	34.6	74.6







#9
Benzaldehyde
Concen: 22.379 ng
RT: 6.440 min Scan# 7
Delta R.T. 0.001 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 77 Resp: 200586

Ion	Ratio	Lower	Upper
77	100		
105	97.1	75.5	115.5
106	94.0	71.7	111.7







#10
Phenol
Concen: 20.269 ng
RT: 6.516 min Scan# 752
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion: 94 Resp: 338933

Ion	Ratio	Lower	Upper
94	100		
65	30.3	11.1	51.1
66	40.8	22.7	62.7







#11
bis(2-Chloroethyl)ether
Concen: 20.612 ng
RT: 6.622 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 93 Resp: 261690

Ion	Ratio	Lower	Upper
93	100		
63	86.3	65.0	105.0
95	32.9	12.4	52.4







#12
1,3-Dichlorobenzene
Concen: 20.603 ng
RT: 6.828 min Scan# 805
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:146 Resp: 297432

Ion	Ratio	Lower	Upper
146	100		
148	63.8	51.0	76.4
75	33.9	26.4	39.6







#13
1,4-Dichlorobenzene
Concen: 20.808 ng
RT: 6.910 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:146 Resp: 300107

Ion	Ratio	Lower	Upper
146	100		
148	63.1	51.8	77.8
111	42.1	34.6	51.8







#14
1,2-Dichlorobenzene
Concen: 21.146 ng
RT: 7.063 min Scan# 845
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:146 Resp: 284645

Ion	Ratio	Lower	Upper
146	100		
148	64.2	51.4	77.2
111	44.2	36.7	55.1







#15
Benzyl Alcohol
Concen: 20.715 ng
RT: 7.022 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 79 Resp: 242106

Ion	Ratio	Lower	Upper
79	100		
108	77.2	61.4	92.0
77	62.8	49.9	74.9







#16
2,2'-oxybis(1-Chloropropane)
Concen: 20.933 ng
RT: 7.163 min Scan# 862
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion: 45 Resp: 453174

Ion	Ratio	Lower	Upper
45	100		
77	11.1	0.0	31.5
79	8.5	0.0	29.1







#17
2-Methylphenol
Concen: 20.371 ng
RT: 7.128 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:107 Resp: 218455

Ion	Ratio	Lower	Upper
107	100		
108	115.0	91.3	136.9
77	46.6	37.1	55.7
79	45.5	37.0	55.6







#18
Hexachloroethane
Concen: 20.566 ng
RT: 7.404 min Scan# 903
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:117 Resp: 105775

Ion	Ratio	Lower	Upper
117	100		
119	97.3	77.0	115.4
201	98.4	76.2	114.4







#19
n-Nitroso-di-n-propylamine
Concen: 20.427 ng
RT: 7.293 min Scan# 8
Delta R.T. -0.012 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 70 Resp: 197390

Ion	Ratio	Lower	Upper
70	100		
42	67.2	51.5	77.3
101	9.2	7.5	11.3
130	19.1	15.5	23.3







#20
3+4-Methylphenols
Concen: 21.346 ng
RT: 7.281 min Scan# 882
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:107 Resp: 292795

Ion	Ratio	Lower	Upper
107	100		
108	89.6	69.8	109.8
77	61.4	50.5	90.5
79	29.4	10.2	50.2







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.175 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:136 Resp: 735124

Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.6	12.8
54	10.3	8.4	12.6
68	6.5	5.1	7.7







#22
Acetophenone
Concen: 21.056 ng
RT: 7.293 min Scan# 884
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:105 Resp: 379127

Ion	Ratio	Lower	Upper
105	100		
71	8.3	5.9	8.9
51	36.5	29.4	44.2
120	21.3	17.5	26.3







#23
Nitrobenzene-d5
Concen: 41.206 ng
RT: 7.451 min Scan# 91
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 82 Resp: 546550

Ion	Ratio	Lower	Upper
82	100		
128	42.3	33.4	50.0
54	59.2	47.8	71.8







#24
Nitrobenzene
Concen: 20.604 ng
RT: 7.469 min Scan# 914
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion: 77 Resp: 299625

Ion	Ratio	Lower	Upper
77	100		
123	44.2	35.7	53.5
65	14.7	11.5	17.3







#25
Isophorone
Concen: 20.502 ng
RT: 7.704 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 82 Resp: 516119

Ion	Ratio	Lower	Upper
82	100		
95	7.1	5.5	8.3
138	16.3	13.2	19.8







#26
2-Nitrophenol
Concen: 20.141 ng
RT: 7.787 min Scan# 968
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:139 Resp: 110497

Ion	Ratio	Lower	Upper
139	100		
109	30.0	25.1	37.7
65	56.5	47.5	71.3







#27
2,4-Dimethylphenol
Concen: 20.544 ng
RT: 7.816 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:122 Resp: 187930

Ion	Ratio	Lower	Upper
122	100		
107	112.6	91.9	137.9
121	57.1	46.2	69.4







#28
bis(2-Chloroethoxy)methane
Concen: 20.816 ng
RT: 7.916 min Scan# 990
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion: 93 Resp: 317497

Ion	Ratio	Lower	Upper
93	100		
95	32.9	26.0	39.0
123	11.1	9.8	14.8







#29
2,4-Dichlorophenol
Concen: 20.673 ng
RT: 8.016 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:162 Resp: 215409

Ion	Ratio	Lower	Upper
162	100		
164	62.7	43.9	83.9
98	41.1	19.9	59.9







#30
1,2,4-Trichlorobenzene
Concen: 20.691 ng
RT: 8.110 min Scan# 1023
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:180 Resp: 238559

Ion	Ratio	Lower	Upper
180	100		
182	100.6	80.5	120.7
145	32.5	25.4	38.0







#31
Naphthalene
Concen: 21.153 ng
RT: 8.192 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:128 Resp: 801971

Ion	Ratio	Lower	Upper
128	100		
129	11.3	8.9	13.3
127	13.1	10.6	16.0







#32
Benzoic acid
Concen: 19.054 ng
RT: 7.892 min Scan# 986
Delta R.T. -0.030 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:122 Resp: 152028

Ion	Ratio	Lower	Upper
122	100		
105	127.9	107.6	147.6
77	98.2	80.2	120.2







#33
4-Chloroaniline
Concen: 20.919 ng
RT: 8.239 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:127 Resp: 270435

Ion	Ratio	Lower	Upper
127	100		
129	33.0	25.4	38.2
65	34.3	28.4	42.6
92	20.5	16.6	25.0







#34
Hexachlorobutadiene
Concen: 20.641 ng
RT: 8.310 min Scan# 1057
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:225 Resp: 149527

Ion	Ratio	Lower	Upper
225	100		
223	63.4	50.6	76.0
227	64.6	51.8	77.8







#35
Caprolactam
Concen: 20.411 ng
RT: 8.587 min Scan# 11
Delta R.T. -0.023 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:113 Resp: 67624

Ion	Ratio	Lower	Upper
113	100		
55	209.6	189.2	229.2
56	152.6	132.5	172.5







#36
4-Chloro-3-methylphenol
Concen: 20.668 ng
RT: 8.704 min Scan# 1124
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:107 Resp: 238461

Ion	Ratio	Lower	Upper
107	100		
144	23.9	19.1	28.7
142	71.9	57.2	85.8







#37
2-Methylnaphthalene
Concen: 21.075 ng
RT: 8.881 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:142 Resp: 489343

Ion	Ratio	Lower	Upper
142	100		
141	88.4	70.8	106.2







#38
1-Methylnaphthalene
Concen: 21.149 ng
RT: 8.981 min Scan# 1171
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:142 Resp: 481797

Ion	Ratio	Lower	Upper
142	100		
141	91.7	73.5	110.3







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.928 min Scan# 1
Delta R.T. 0.001 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:164 Resp: 413023

Ion	Ratio	Lower	Upper
164	100		
162	99.7	81.0	121.4
160	43.8	35.4	53.0







#40
1,2,4,5-Tetrachlorobenzene
Concen: 20.823 ng
RT: 9.045 min Scan# 1182
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:216 Resp: 232025

Ion	Ratio	Lower	Upper
216	100		
214	78.2	63.4	95.0
179	20.8	17.2	25.8
108	18.7	15.6	23.4







#41
Hexachlorocyclopentadiene
Concen: 21.041 ng
RT: 9.039 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:237 Resp: 81579

Ion	Ratio	Lower	Upper
237	100		
235	62.3	42.9	82.9
272	12.7	0.0	32.3







#42
2,4,6-Tribromophenol
Concen: 41.199 ng
RT: 10.710 min Scan# 1465
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:330 Resp: 159165

Ion	Ratio	Lower	Upper
330	100		
332	96.4	78.1	117.1
141	32.9	26.6	39.8







#43
2,4,6-Trichlorophenol
Concen: 20.930 ng
RT: 9.157 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:196 Resp: 165112

Ion	Ratio	Lower	Upper
196	100		
198	95.4	76.2	114.2
200	30.8	24.0	36.0







#44
2,4,5-Trichlorophenol
Concen: 20.213 ng
RT: 9.192 min Scan# 1207
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:196 Resp: 164517

Ion	Ratio	Lower	Upper
196	100		
198	96.2	75.8	113.6
97	49.9	41.8	62.6
132	28.4	22.9	34.3







#45
2-Fluorobiphenyl
Concen: 42.316 ng
RT: 9.245 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:172 Resp: 1057523

Ion	Ratio	Lower	Upper
172	100		
171	35.0	28.6	43.0
170	23.6	19.1	28.7







#46
1,1'-Biphenyl
Concen: 21.297 ng
RT: 9.345 min Scan# 1233
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:154 Resp: 612342

Ion	Ratio	Lower	Upper
154	100		
153	39.6	20.3	60.3
76	14.4	0.0	33.9







#47
2-Chloronaphthalene
Concen: 20.767 ng
RT: 9.375 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:162 Resp: 480916

Ion	Ratio	Lower	Upper
162	100		
127	41.7	33.8	50.6
164	32.8	26.6	39.8







#48
2-Nitroaniline
Concen: 20.614 ng
RT: 9.463 min Scan# 1253
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion: 65 Resp: 146002

Ion	Ratio	Lower	Upper
65	100		
92	64.0	50.9	76.3
138	94.4	73.4	110.0







#49
Acenaphthylene
Concen: 21.182 ng
RT: 9.786 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:152 Resp: 709161

Ion	Ratio	Lower	Upper
152	100		
151	20.0	15.9	23.9
153	13.0	10.7	16.1







#50
Dimethylphthalate
Concen: 20.592 ng
RT: 9.645 min Scan# 1284
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:163 Resp: 529745

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.4	5.0
164	10.1	8.2	12.2







#51
2,6-Dinitrotoluene
Concen: 20.596 ng
RT: 9.704 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:165 Resp: 114726

Ion	Ratio	Lower	Upper
165	100		
63	64.6	50.6	75.8
89	58.4	47.5	71.3







#52
Acenaphthene
Concen: 20.768 ng
RT: 9.963 min Scan# 1338
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:154 Resp: 449770

Ion	Ratio	Lower	Upper
154	100		
153	110.4	89.8	134.6
152	52.5	42.4	63.6







#53
3-Nitroaniline
Concen: 21.271 ng
RT: 9.869 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:138 Resp: 122186

Ion	Ratio	Lower	Upper
138	100		
108	13.5	10.2	15.2
92	121.9	98.3	147.5







#54
2,4-Dinitrophenol
Concen: 19.148 ng
RT: 9.975 min Scan# 1340
Delta R.T. 0.001 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:184 Resp: 31801

Ion	Ratio	Lower	Upper
184	100		
63	91.3	84.7	127.1
154	331.4	338.2	507.2#







#55
Dibenzofuran
Concen: 20.991 ng
RT: 10.134 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:168 Resp: 652249

Ion	Ratio	Lower	Upper
168	100		
139	38.0	31.8	47.6
169	14.2	11.6	17.4







#56
4-Nitrophenol
Concen: 21.057 ng
RT: 10.010 min Scan# 1346
Delta R.T. -0.012 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:139 Resp: 93060

Ion	Ratio	Lower	Upper
139	100		
109	74.2	53.6	93.6
65	115.7	87.8	127.8







#57
2,4-Dinitrotoluene
Concen: 20.348 ng
RT: 10.104 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:165 Resp: 139381

Ion	Ratio	Lower	Upper
165	100		
63	50.6	40.1	60.1
89	77.9	60.4	90.6
182	2.8	2.0	3.0







#58
Fluorene
Concen: 20.958 ng
RT: 10.475 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:166 Resp: 496012

Ion	Ratio	Lower	Upper
166	100		
165	98.3	78.5	117.7
167	13.5	10.9	16.3







#59
2,3,4,6-Tetrachlorophenol
Concen: 21.133 ng
RT: 10.245 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:232 Resp: 134842

Ion	Ratio	Lower	Upper
232	100		
131	45.1	37.0	55.6
130	2.5	1.9	2.9
166	29.2	24.2	36.2







#60
Diethylphthalate
Concen: 20.723 ng
RT: 10.345 min Scan# 1403
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:149 Resp: 523458

Ion	Ratio	Lower	Upper
149	100		
177	22.5	17.9	26.9
150	12.7	10.1	15.1







#61
4-Chlorophenyl-phenylether
Concen: 21.331 ng
RT: 10.469 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:204 Resp: 254950

Ion	Ratio	Lower	Upper
204	100		
206	32.3	26.3	39.5
141	58.3	48.6	73.0







#62
4-Nitroaniline
Concen: 20.586 ng
RT: 10.481 min Scan# 1426
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:138 Resp: 111683

Ion	Ratio	Lower	Upper
138	100		
92	53.4	35.7	75.7
108	76.5	57.5	97.5







#63
Azobenzene
Concen: 20.906 ng
RT: 10.628 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 77 Resp: 559837

Ion	Ratio	Lower	Upper
77	100		
182	24.4	3.3	43.3
105	21.3	1.2	41.2
51	33.0	13.3	53.3







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1584
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:188 Resp: 724973

Ion	Ratio	Lower	Upper
188	100		
94	10.5	7.9	11.9
80	11.7	9.0	13.4







#65
4,6-Dinitro-2-methylphenol
Concen: 17.645 ng
RT: 10.510 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:198 Resp: 52180

Ion	Ratio	Lower	Upper
198	100		
51	58.5	38.2	78.2
105	48.3	27.9	67.9







#66
n-Nitrosodiphenylamine
Concen: 20.736 ng
RT: 10.581 min Scan# 1443
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:169 Resp: 449925

Ion	Ratio	Lower	Upper
169	100		
168	63.1	49.8	74.6
167	34.4	28.0	42.0







#67
4-Bromophenyl-phenylether
Concen: 20.512 ng
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:248 Resp: 153196

Ion	Ratio	Lower	Upper
248	100		
250	97.4	79.0	118.6
141	74.3	61.7	92.5







#68
Hexachlorobenzene
Concen: 20.488 ng
RT: 11.022 min Scan# 1518
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:284 Resp: 172091

Ion	Ratio	Lower	Upper
284	100		
142	32.7	26.9	40.3
249	31.5	25.3	37.9







#69
Atrazine
Concen: 19.240 ng
RT: 11.104 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:200 Resp: 113089

Ion	Ratio	Lower	Upper
200	100		
173	27.2	7.7	47.7
215	46.9	27.7	67.7







#70
Pentachlorophenol
Concen: 21.106 ng
RT: 11.210 min Scan# 1550
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:266 Resp: 107595

Ion	Ratio	Lower	Upper
266	100		
268	63.4	50.4	75.6
264	62.8	50.6	75.8







#71
Phenanthrene
Concen: 21.036 ng
RT: 11.439 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:178 Resp: 720363

Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.8	23.6
179	15.5	12.6	18.8







#72
Anthracene
Concen: 21.083 ng
RT: 11.486 min Scan# 1597
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:178 Resp: 704417

Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.3	22.9
179	15.3	12.4	18.6







#73
Carbazole
Concen: 21.540 ng
RT: 11.639 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:167 Resp: 669328

Ion	Ratio	Lower	Upper
167	100		
166	21.3	17.0	25.6
139	13.1	10.8	16.2







#74
Di-n-butylphthalate
Concen: 21.442 ng
RT: 11.975 min Scan# 1680
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:149 Resp: 769779

Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.5	11.3
104	4.8	4.1	6.1







#75
Fluoranthene
Concen: 21.703 ng
RT: 12.622 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:202 Resp: 751337

Ion	Ratio	Lower	Upper
202	100		
101	12.6	0.0	31.9
203	17.7	0.0	37.5







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.051 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:240 Resp: 402251

Ion	Ratio	Lower	Upper
240	100		
120	11.3	9.4	14.2
236	26.2	20.9	31.3







#77
Benzidine
Concen: 18.042 ng
RT: 12.745 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:184 Resp: 118035

Ion	Ratio	Lower	Upper
184	100		
185	15.0	12.0	18.0
183	11.8	9.4	14.2







#78
Pyrene
Concen: 21.700 ng
RT: 12.851 min Scan# 1829
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:202 Resp: 764077

Ion	Ratio	Lower	Upper
202	100		
200	21.5	17.2	25.8
203	17.4	14.2	21.4







#79
Terphenyl-d14
Concen: 43.673 ng
RT: 12.998 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:244 Resp: 1078101

Ion	Ratio	Lower	Upper
244	100		
212	6.8	5.7	8.5
122	10.7	8.6	13.0







#80
Butylbenzylphthalate
Concen: 20.432 ng
RT: 13.474 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:149 Resp: 215683

Ion	Ratio	Lower	Upper
149	100		
91	76.6	59.4	89.2
206	20.9	16.4	24.6







#81
Benzo(a)anthracene
Concen: 20.413 ng
RT: 14.039 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:228 Resp: 534525

Ion	Ratio	Lower	Upper
228	100		
226	26.8	21.6	32.4
229	19.8	16.1	24.1







#82
3,3'-Dichlorobenzidine
Concen: 20.534 ng
RT: 14.004 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:252 Resp: 156770

Ion	Ratio	Lower	Upper
252	100		
254	63.2	50.8	76.2
126	11.3	8.9	13.3







#83
Chrysene
Concen: 20.678 ng
RT: 14.074 min Scan# 2(
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:228 Resp: 496447

Ion	Ratio	Lower	Upper
228	100		
226	29.9	24.1	36.1
229	20.1	15.8	23.6







#84
Bis(2-ethylhexyl)phthalate
Concen: 19.606 ng
RT: 14.033 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:149 Resp: 232200

Ion	Ratio	Lower	Upper
149	100		
167	26.3	21.4	32.0
279	4.0	3.0	4.4







#85
Di-n-octyl phthalate
Concen: 17.359 ng
RT: 14.651 min Scan# 21
Delta R.T. -0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:149 Resp: 373952

Ion	Ratio	Lower	Upper
149	100		
167	1.4	1.2	1.8
43	12.2	9.8	14.6







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.527 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:264 Resp: 371811

Ion	Ratio	Lower	Upper
264	100		
260	23.6	19.4	29.2
265	21.1	17.4	26.0







#87
Indeno(1,2,3-cd)pyrene
Concen: 20.312 ng
RT: 17.015 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:276 Resp: 485858

Ion	Ratio	Lower	Upper
276	100		
138	25.9	20.4	30.6
277	25.2	20.2	30.4







#88
Benzo(b)fluoranthene
Concen: 21.659 ng
RT: 15.092 min Scan# 2210
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:252 Resp: 490566

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.6	26.4
125	10.2	7.9	11.9







#89
Benzo(k)fluoranthene
Concen: 18.883 ng
RT: 15.121 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:252 Resp: 368835

Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.6	26.4
125	10.2	7.9	11.9







#90
Benzo(a)pyrene
Concen: 20.316 ng
RT: 15.463 min Scan# 2273
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:252 Resp: 378617

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.3	25.9
125	11.2	8.7	13.1







#91
Dibenzo(a,h)anthracene
Concen: 20.561 ng
RT: 17.039 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:278 Resp: 410285

Ion	Ratio	Lower	Upper
278	100		
139	15.4	13.1	19.7
279	23.8	19.0	28.4






#92
Benzo(g,h,i)perylene
Concen: 20.326 ng
RT: 17.468 min Scan# 2614
Delta R.T. -0.012 min
Lab File: BF139847.D
Acq: 18 Oct 2024 11:52

Tgt Ion:276 Resp: 405140

Ion	Ratio	Lower	Upper
276	100		
277	23.9	19.0	28.6
138	21.2	17.4	26.0



Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139848.D
Acq On : 18 Oct 2024 12:20
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Oct 18 14:18:47 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.892	152	181899	20.000	ng	0.00
21) Naphthalene-d8	8.175	136	688880	20.000	ng	0.00
39) Acenaphthene-d10	9.927	164	379753	20.000	ng	0.00
64) Phenanthrene-d10	11.416	188	663459	20.000	ng	0.00
76) Chrysene-d12	14.051	240	345800	20.000	ng	0.00
86) Perylene-d12	15.527	264	365076	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.498	112	929569	80.002	ng	0.00
7) Phenol-d6	6.510	99	1197998	79.607	ng	0.00
23) Nitrobenzene-d5	7.451	82	1021808	82.209	ng	0.00
42) 2,4,6-Tribromophenol	10.716	330	285205	80.292	ng	0.00
45) 2-Fluorobiphenyl	9.251	172	1764495	76.791	ng	0.00
79) Terphenyl-d14	12.998	244	1721125	81.103	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.687	88	214958	39.679	ng	100
3) Pyridine	3.446	79	547145	40.746	ng	100
4) n-Nitrosodimethylamine	3.398	42	293138	40.631	ng	100
6) Aniline	6.551	93	575402	41.026	ng	100
8) 2-Chlorophenol	6.675	128	476614	40.124	ng	100
9) Benzaldehyde	6.439	77	341959	40.393	ng	100
10) Phenol	6.522	94	622748m	39.429	ng	
11) bis(2-Chloroethyl)ether	6.628	93	470765	39.258	ng	100
12) 1,3-Dichlorobenzene	6.834	146	545252	39.988	ng	100
13) 1,4-Dichlorobenzene	6.910	146	541068	39.718	ng	100
14) 1,2-Dichlorobenzene	7.063	146	501857	39.473	ng	100
15) Benzyl Alcohol	7.028	79	441461	39.991	ng	100
16) 2,2'-oxybis(1-Chloropr...	7.169	45	820216	40.113	ng	100
17) 2-Methylphenol	7.133	107	405109	39.996	ng	100
18) Hexachloroethane	7.404	117	196637	40.478	ng	100
19) n-Nitroso-di-n-propyla...	7.304	70	354454	38.835	ng	100
20) 3+4-Methylphenols	7.286	107	515857	39.818	ng	100
22) Acetophenone	7.298	105	663304	39.312	ng	100
24) Nitrobenzene	7.475	77	555115	40.735	ng	100
25) Isophorone	7.710	82	935400	39.651	ng	100
26) 2-Nitrophenol	7.786	139	216916	42.193	ng	100
27) 2,4-Dimethylphenol	7.816	122	341985	39.894	ng	100
28) bis(2-Chloroethoxy)met...	7.922	93	567188	39.683	ng	100
29) 2,4-Dichlorophenol	8.022	162	390330	39.976	ng	100
30) 1,2,4-Trichlorobenzene	8.116	180	434274	40.195	ng	100
31) Naphthalene	8.198	128	1409796	39.681	ng	100
32) Benzoic acid	7.922	122	302443	40.451	ng	100
33) 4-Chloroaniline	8.239	127	483982	39.950	ng	100
34) Hexachlorobutadiene	8.316	225	271315	39.967	ng	100
35) Caprolactam	8.610	113	126585	40.772	ng	100
36) 4-Chloro-3-methylphenol	8.710	107	433727	40.116	ng	100
37) 2-Methylnaphthalene	8.886	142	856111	39.345	ng	100
38) 1-Methylnaphthalene	8.986	142	842638	39.471	ng	100
40) 1,2,4,5-Tetrachloroben...	9.051	216	407511	39.776	ng	100
41) Hexachlorocyclopentadiene	9.039	237	150540	42.229	ng	100
43) 2,4,6-Trichlorophenol	9.157	196	293577	40.474	ng	100

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139848.D
Acq On : 18 Oct 2024 12:20
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Oct 18 14:18:47 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.192	196	304213	40.651	ng	100
46) 1,1'-Biphenyl	9.351	154	1047094	39.608	ng	100
47) 2-Chloronaphthalene	9.375	162	843452	39.613	ng	100
48) 2-Nitroaniline	9.463	65	277165	42.562	ng	100
49) Acenaphthylene	9.792	152	1226273	39.837	ng	100
50) Dimethylphthalate	9.651	163	939670	39.726	ng	100
51) 2,6-Dinitrotoluene	9.710	165	212956	41.580	ng	100
52) Acenaphthene	9.963	154	792756	39.811	ng	100
53) 3-Nitroaniline	9.874	138	222098	42.051	ng	100
54) 2,4-Dinitrophenol	9.974	184	76948	39.739	ng	100
55) Dibenzofuran	10.133	168	1128670	39.505	ng	100
56) 4-Nitrophenol	10.022	139	175913	43.292	ng	100
57) 2,4-Dinitrotoluene	10.110	165	270604	42.965	ng	100
58) Fluorene	10.474	166	843048	38.742	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.245	232	235175	40.087	ng	100
60) Diethylphthalate	10.351	149	922384	39.715	ng	100
61) 4-Chlorophenyl-phenyle...	10.469	204	432393	39.347	ng	100
62) 4-Nitroaniline	10.486	138	208926	41.884	ng	100
63) Azobenzene	10.627	77	991989	40.289	ng	100
65) 4,6-Dinitro-2-methylph...	10.516	198	115824	42.799	ng	100
66) n-Nitrosodiphenylamine	10.586	169	789753	39.772	ng	100
67) 4-Bromophenyl-phenylether	10.957	248	268509	39.285	ng	100
68) Hexachlorobenzene	11.021	284	306919	39.927	ng	100
69) Atrazine	11.110	200	232078	43.145	ng	100
70) Pentachlorophenol	11.210	266	201559	43.204	ng	100
71) Phenanthrene	11.439	178	1238155	39.508	ng	100
72) Anthracene	11.492	178	1211718	39.628	ng	100
73) Carbazole	11.639	167	1122566	39.475	ng	100
74) Di-n-butylphthalate	11.974	149	1345878	40.965	ng	100
75) Fluoranthene	12.627	202	1251202	39.493	ng	100
77) Benzidine	12.745	184	252893	44.967	ng	100
78) Pyrene	12.857	202	1248309	41.241	ng	100
80) Butylbenzylphthalate	13.474	149	383831	42.297	ng	100
81) Benzo(a)anthracene	14.039	228	920698	40.901	ng	100
82) 3,3'-Dichlorobenzidine	14.004	252	267769	40.798	ng	100
83) Chrysene	14.080	228	806950	39.098	ng	100
84) Bis(2-ethylhexyl)phtha...	14.033	149	432750	42.504	ng	100
85) Di-n-octyl phthalate	14.651	149	785179	42.399	ng	100
87) Indeno(1,2,3-cd)pyrene	17.021	276	987345	42.039	ng	100
88) Benzo(b)fluoranthene	15.098	252	873168	39.263	ng	100
89) Benzo(k)fluoranthene	15.127	252	778470	40.589	ng	100
90) Benzo(a)pyrene	15.468	252	748507	40.905	ng	100
91) Dibenzo(a,h)anthracene	17.045	278	818044	41.752	ng	100
92) Benzo(g,h,i)perylene	17.480	276	823727	42.090	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139848.D
Acq On : 18 Oct 2024 12:20
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1


Quant Time: Oct 18 14:18:47 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.892 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion:152 Resp: 18189


Ion	Ratio	Lower	Upper
152	100		
150	162.7	130.2	195.2
115	64.3	51.4	77.2

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#2
1,4-Dioxane
Concen: 39.679 ng
RT: 2.687 min Scan# 101
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion: 88 Resp: 214958

Ion	Ratio	Lower	Upper
88	100		
58	91.6	73.3	109.9
43	33.9	27.1	40.7






#3
Pyridine
Concen: 40.746 ng
RT: 3.446 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion: 79 Resp: 54714


Ion	Ratio	Lower	Upper
79	100		
52	85.6	68.5	102.7
51	40.4	32.3	48.5



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#4
n-Nitrosodimethylamine
Concen: 40.631 ng
RT: 3.398 min Scan# 222
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion: 42 Resp: 293138

Ion	Ratio	Lower	Upper
42	100		
74	120.3	96.2	144.4
44	7.7	6.2	9.2





#5
2-Fluorophenol
Concen: 80.002 ng
RT: 5.498 min Scan# 5
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion:112 Resp: 92956


Ion	Ratio	Lower	Upper
112	100		
64	66.3	53.0	79.6
63	33.7	27.0	40.4

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#6
Aniline
Concen: 41.026 ng
RT: 6.551 min Scan# 758
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion: 93 Resp: 575402

Ion	Ratio	Lower	Upper
93	100		
66	42.5	34.0	51.0
65	21.7	17.4	26.0





#7
Phenol-d6
Concen: 79.607 ng
RT: 6.510 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion: 99 Resp: 119799


Ion	Ratio	Lower	Upper
99	100		
42	20.9	16.7	25.1
71	34.6	27.7	41.5

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#8
2-Chlorophenol
Concen: 40.124 ng
RT: 6.675 min Scan# 779
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:128 Resp: 476614

Ion	Ratio	Lower	Upper
128	100		
130	32.3	12.3	52.3
64	54.6	34.6	74.6






#9
Benzaldehyde
Concen: 40.393 ng
RT: 6.439 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion: 77 Resp: 34195


Ion	Ratio	Lower	Upper
77	100		
105	95.5	75.5	115.5
106	91.7	71.7	111.7



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#10
Phenol
Concen: 39.429 ng m
RT: 6.522 min Scan# 753
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion: 94 Resp: 622748

Ion	Ratio	Lower	Upper
94	100		
65	31.1	11.1	51.1
66	42.7	22.7	62.7





#11
bis(2-Chloroethyl)ether
Concen: 39.258 ng
RT: 6.628 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion: 93 Resp: 47076


Ion	Ratio	Lower	Upper
93	100		
63	85.0	65.0	105.0
95	32.4	12.4	52.4

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#12
1,3-Dichlorobenzene
Concen: 39.988 ng
RT: 6.834 min Scan# 806
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:146 Resp: 545252

Ion	Ratio	Lower	Upper
146	100		
148	63.7	51.0	76.4
75	33.0	26.4	39.6






#13
1,4-Dichlorobenzene
Concen: 39.718 ng
RT: 6.910 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:146 Resp: 54106


Ion	Ratio	Lower	Upper
146	100		
148	64.8	51.8	77.8
111	43.2	34.6	51.8



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#14
1,2-Dichlorobenzene
Concen: 39.473 ng
RT: 7.063 min Scan# 845
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:146 Resp: 501857

Ion	Ratio	Lower	Upper
146	100		
148	64.3	51.4	77.2
111	45.9	36.7	55.1





#15
Benzyl Alcohol
Concen: 39.991 ng
RT: 7.028 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion: 79 Resp: 44146


Ion	Ratio	Lower	Upper
79	100		
108	76.7	61.4	92.0
77	62.4	49.9	74.9

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#16
2,2'-oxybis(1-Chloropropane)
Concen: 40.113 ng
RT: 7.169 min Scan# 863
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion: 45 Resp: 820216

Ion	Ratio	Lower	Upper
45	100		
77	11.5	0.0	31.5
79	9.1	0.0	29.1






#17
2-Methylphenol
Concen: 39.996 ng
RT: 7.133 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:107 Resp: 40510


Ion	Ratio	Lower	Upper
107	100		
108	114.1	91.3	136.9
77	46.4	37.1	55.7
79	46.3	37.0	55.6



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#18
Hexachloroethane
Concen: 40.478 ng
RT: 7.404 min Scan# 903
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:117 Resp: 196637

Ion	Ratio	Lower	Upper
117	100		
119	96.2	77.0	115.4
201	95.3	76.2	114.4






#19
n-Nitroso-di-n-propylamine
Concen: 38.835 ng
RT: 7.304 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion: 70 Resp: 35445


Ion	Ratio	Lower	Upper
70	100		
42	64.4	51.5	77.3
101	9.4	7.5	11.3
130	19.4	15.5	23.3



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#20
3+4-Methylphenols
Concen: 39.818 ng
RT: 7.286 min Scan# 883
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:107 Resp: 515857

Ion	Ratio	Lower	Upper
107	100		
108	89.8	69.8	109.8
77	70.5	50.5	90.5
79	30.2	10.2	50.2






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.175 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:136 Resp: 68888


Ion	Ratio	Lower	Upper
136	100		
137	10.7	8.6	12.8
54	10.5	8.4	12.6
68	6.4	5.1	7.7



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#22
Acetophenone
Concen: 39.312 ng
RT: 7.298 min Scan# 885
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:105 Resp: 663304

Ion	Ratio	Lower	Upper
105	100		
71	7.4	5.9	8.9
51	36.8	29.4	44.2
120	21.9	17.5	26.3





#23
Nitrobenzene-d5
Concen: 82.209 ng
RT: 7.451 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion: 82 Resp: 102180


Ion	Ratio	Lower	Upper
82	100		
128	41.7	33.4	50.0
54	59.8	47.8	71.8

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#24
Nitrobenzene
Concen: 40.735 ng
RT: 7.475 min Scan# 915
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion: 77 Resp: 555115

Ion	Ratio	Lower	Upper
77	100		
123	44.6	35.7	53.5
65	14.4	11.5	17.3





#25
Isophorone
Concen: 39.651 ng
RT: 7.710 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion: 82 Resp: 93540


Ion	Ratio	Lower	Upper
82	100		
95	6.9	5.5	8.3
138	16.5	13.2	19.8

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#26
2-Nitrophenol
Concen: 42.193 ng
RT: 7.786 min Scan# 968
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:139 Resp: 216916

Ion	Ratio	Lower	Upper
139	100		
109	31.4	25.1	37.7
65	59.4	47.5	71.3






#27
2,4-Dimethylphenol
Concen: 39.894 ng
RT: 7.816 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:122 Resp: 34198


Ion	Ratio	Lower	Upper
122	100		
107	114.9	91.9	137.9
121	57.8	46.2	69.4



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#28
bis(2-Chloroethoxy)methane
Concen: 39.683 ng
RT: 7.922 min Scan# 991
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion: 93 Resp: 567188

Ion	Ratio	Lower	Upper
93	100		
95	32.5	26.0	39.0
123	12.3	9.8	14.8





#29
2,4-Dichlorophenol
Concen: 39.976 ng
RT: 8.022 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion:162 Resp: 39033


Ion	Ratio	Lower	Upper
162	100		
164	63.9	43.9	83.9
98	39.9	19.9	59.9

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#30
1,2,4-Trichlorobenzene
Concen: 40.195 ng
RT: 8.116 min Scan# 1024
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:180 Resp: 434274

Ion	Ratio	Lower	Upper
180	100		
182	100.6	80.5	120.7
145	31.7	25.4	38.0






#31
Naphthalene
Concen: 39.681 ng
RT: 8.198 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:128 Resp: 140979


Ion	Ratio	Lower	Upper
128	100		
129	11.1	8.9	13.3
127	13.3	10.6	16.0



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#32
Benzoic acid
Concen: 40.451 ng
RT: 7.922 min Scan# 991
Delta R.T. -0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:122 Resp: 302443

Ion	Ratio	Lower	Upper
122	100		
105	127.6	107.6	147.6
77	100.2	80.2	120.2






#33
4-Chloroaniline
Concen: 39.950 ng
RT: 8.239 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:127 Resp: 48398


Ion	Ratio	Lower	Upper
127	100		
129	31.8	25.4	38.2
65	35.5	28.4	42.6
92	20.8	16.6	25.0



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#34
Hexachlorobutadiene
Concen: 39.967 ng
RT: 8.316 min Scan# 1058
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:225 Resp: 271315

Ion	Ratio	Lower	Upper
225	100		
223	63.3	50.6	76.0
227	64.8	51.8	77.8





#35
Caprolactam
Concen: 40.772 ng
RT: 8.610 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion:113 Resp: 12658


Ion	Ratio	Lower	Upper
113	100		
55	209.2	189.2	229.2
56	152.5	132.5	172.5

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#36
4-Chloro-3-methylphenol
Concen: 40.116 ng
RT: 8.710 min Scan# 1125
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:107 Resp: 433727

Ion	Ratio	Lower	Upper
107	100		
144	23.9	19.1	28.7
142	71.5	57.2	85.8





#37
2-Methylnaphthalene
Concen: 39.345 ng
RT: 8.886 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion:142 Resp: 85611


Ion	Ratio	Lower	Upper
142	100		
141	88.5	70.8	106.2

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#38
1-Methylnaphthalene
Concen: 39.471 ng
RT: 8.986 min Scan# 1172
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:142 Resp: 842638

Ion	Ratio	Lower	Upper
142	100		
141	91.9	73.5	110.3





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.927 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion:164 Resp: 37975


Ion	Ratio	Lower	Upper
164	100		
162	101.2	81.0	121.4
160	44.2	35.4	53.0

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#40
1,2,4,5-Tetrachlorobenzene
Concen: 39.776 ng
RT: 9.051 min Scan# 1183
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:216 Resp: 407511

Ion	Ratio	Lower	Upper
216	100		
214	79.2	63.4	95.0
179	21.5	17.2	25.8
108	19.5	15.6	23.4





#41
Hexachlorocyclopentadiene
Concen: 42.229 ng
RT: 9.039 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion:237 Resp: 15054


Ion	Ratio	Lower	Upper
237	100		
235	62.9	42.9	82.9
272	12.3	0.0	32.3

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#42
2,4,6-Tribromophenol
Concen: 80.292 ng
RT: 10.716 min Scan# 1466
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:330 Resp: 285205

Ion	Ratio	Lower	Upper
330	100		
332	97.6	78.1	117.1
141	33.2	26.6	39.8






#43
2,4,6-Trichlorophenol
Concen: 40.474 ng
RT: 9.157 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:196 Resp: 29357


Ion	Ratio	Lower	Upper
196	100		
198	95.2	76.2	114.2
200	30.0	24.0	36.0



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#44
2,4,5-Trichlorophenol
Concen: 40.651 ng
RT: 9.192 min Scan# 1207
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:196 Resp: 304213

Ion	Ratio	Lower	Upper
196	100		
198	94.7	75.8	113.6
97	52.2	41.8	62.6
132	28.6	22.9	34.3






#45
2-Fluorobiphenyl
Concen: 76.791 ng
RT: 9.251 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:172 Resp: 176449


Ion	Ratio	Lower	Upper
172	100		
171	35.8	28.6	43.0
170	23.9	19.1	28.7



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#46
1,1'-Biphenyl
Concen: 39.608 ng
RT: 9.351 min Scan# 1234
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:154 Resp: 1047094

Ion	Ratio	Lower	Upper
154	100		
153	40.3	20.3	60.3
76	13.9	0.0	33.9






#47
2-Chloronaphthalene
Concen: 39.613 ng
RT: 9.375 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:162 Resp: 84345


Ion	Ratio	Lower	Upper
162	100		
127	42.2	33.8	50.6
164	33.2	26.6	39.8



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#48
2-Nitroaniline
Concen: 42.562 ng
RT: 9.463 min Scan# 1253
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion: 65 Resp: 277165

Ion	Ratio	Lower	Upper
65	100		
92	63.6	50.9	76.3
138	91.7	73.4	110.0





#49
Acenaphthylene
Concen: 39.837 ng
RT: 9.792 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion:152 Resp: 122627


Ion	Ratio	Lower	Upper
152	100		
151	19.9	15.9	23.9
153	13.4	10.7	16.1

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#50
Dimethylphthalate
Concen: 39.726 ng
RT: 9.651 min Scan# 1285
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:163 Resp: 939670

Ion	Ratio	Lower	Upper
163	100		
194	4.2	3.4	5.0
164	10.2	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 41.580 ng
RT: 9.710 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion:165 Resp: 21295


Ion	Ratio	Lower	Upper
165	100		
63	63.2	50.6	75.8
89	59.4	47.5	71.3

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#52
Acenaphthene
Concen: 39.811 ng
RT: 9.963 min Scan# 1338
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:154 Resp: 792756

Ion	Ratio	Lower	Upper
154	100		
153	112.2	89.8	134.6
152	53.0	42.4	63.6





#53
3-Nitroaniline
Concen: 42.051 ng
RT: 9.874 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion:138 Resp: 22209


Ion	Ratio	Lower	Upper
138	100		
108	12.7	10.2	15.2
92	122.9	98.3	147.5

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#54
2,4-Dinitrophenol
Concen: 39.739 ng
RT: 9.974 min Scan# 1340
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:184 Resp: 76948

Ion	Ratio	Lower	Upper
184	100		
63	105.9	84.7	127.1
154	422.7	338.2	507.2






#55
Dibenzofuran
Concen: 39.505 ng
RT: 10.133 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:168 Resp: 112867


Ion	Ratio	Lower	Upper
168	100		
139	39.7	31.8	47.6
169	14.5	11.6	17.4



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#56
4-Nitrophenol
Concen: 43.292 ng
RT: 10.022 min Scan# 1348
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:139 Resp: 175913

Ion	Ratio	Lower	Upper
139	100		
109	73.6	53.6	93.6
65	107.8	87.8	127.8






#57
2,4-Dinitrotoluene
Concen: 42.965 ng
RT: 10.110 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:165 Resp: 27060


Ion	Ratio	Lower	Upper
165	100		
63	50.1	40.1	60.1
89	75.5	60.4	90.6
182	2.5	2.0	3.0



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#58
Fluorene
Concen: 38.742 ng
RT: 10.474 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:166 Resp: 843048

Ion	Ratio	Lower	Upper
166	100		
165	98.1	78.5	117.7
167	13.6	10.9	16.3






#59
2,3,4,6-Tetrachlorophenol
Concen: 40.087 ng
RT: 10.245 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:232 Resp: 23517


Ion	Ratio	Lower	Upper
232	100		
131	46.3	37.0	55.6
130	2.4	1.9	2.9
166	30.2	24.2	36.2



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#60
Diethylphthalate
Concen: 39.715 ng
RT: 10.351 min Scan# 1404
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:149 Resp: 922384

Ion	Ratio	Lower	Upper
149	100		
177	22.4	17.9	26.9
150	12.6	10.1	15.1





#61
4-Chlorophenyl-phenylether
Concen: 39.347 ng
RT: 10.469 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion:204 Resp: 43239


Ion	Ratio	Lower	Upper
204	100		
206	32.9	26.3	39.5
141	60.8	48.6	73.0

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#62
4-Nitroaniline
Concen: 41.884 ng
RT: 10.486 min Scan# 1427
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:138 Resp: 208926

Ion	Ratio	Lower	Upper
138	100		
92	55.7	35.7	75.7
108	77.5	57.5	97.5






#63
Azobenzene
Concen: 40.289 ng
RT: 10.627 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion: 77 Resp: 99198


Ion	Ratio	Lower	Upper
77	100		
182	23.3	3.3	43.3
105	21.2	1.2	41.2
51	33.3	13.3	53.3



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.416 min Scan# 1585
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:188 Resp: 663459

Ion	Ratio	Lower	Upper
188	100		
94	9.9	7.9	11.9
80	11.2	9.0	13.4





#65
4,6-Dinitro-2-methylphenol
Concen: 42.799 ng
RT: 10.516 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion:198 Resp: 11582


Ion	Ratio	Lower	Upper
198	100		
51	58.2	38.2	78.2
105	47.9	27.9	67.9

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#66
n-Nitrosodiphenylamine
Concen: 39.772 ng
RT: 10.586 min Scan# 1444
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:169 Resp: 789753

Ion	Ratio	Lower	Upper
169	100		
168	62.2	49.8	74.6
167	35.0	28.0	42.0






#67
4-Bromophenyl-phenylether
Concen: 39.285 ng
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:248 Resp: 26850


Ion	Ratio	Lower	Upper
248	100		
250	98.8	79.0	118.6
141	77.1	61.7	92.5



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#68
Hexachlorobenzene
Concen: 39.927 ng
RT: 11.021 min Scan# 1518
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:284 Resp: 306919

Ion	Ratio	Lower	Upper
284	100		
142	33.6	26.9	40.3
249	31.6	25.3	37.9





#69
Atrazine
Concen: 43.145 ng
RT: 11.110 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Instrument :
BNA_F
ClientSampleId :
SSTDICCC040


Tgt Ion:200 Resp: 23207

Ion	Ratio	Lower	Upper
200	100		
173	27.7	7.7	47.7
215	47.7	27.7	67.7

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#70
Pentachlorophenol
Concen: 43.204 ng
RT: 11.210 min Scan# 1550
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:266 Resp: 201559

Ion	Ratio	Lower	Upper
266	100		
268	63.0	50.4	75.6
264	63.2	50.6	75.8






#71
Phenanthrene
Concen: 39.508 ng
RT: 11.439 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:178 Resp: 123815


Ion	Ratio	Lower	Upper
178	100		
176	19.7	15.8	23.6
179	15.7	12.6	18.8



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#72
Anthracene
Concen: 39.628 ng
RT: 11.492 min Scan# 1598
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:178 Resp: 1211718

Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.3	22.9
179	15.5	12.4	18.6






#73
Carbazole
Concen: 39.475 ng
RT: 11.639 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:167 Resp: 112256


Ion	Ratio	Lower	Upper
167	100		
166	21.3	17.0	25.6
139	13.5	10.8	16.2



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#74
Di-n-butylphthalate
Concen: 40.965 ng
RT: 11.974 min Scan# 1680
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:149 Resp: 1345878

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.5	11.3
104	5.1	4.1	6.1





#75
Fluoranthene
Concen: 39.493 ng
RT: 12.627 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion:202 Resp: 125120


Ion	Ratio	Lower	Upper
202	100		
101	11.9	0.0	31.9
203	17.5	0.0	37.5

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.051 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:240 Resp: 345800

Ion	Ratio	Lower	Upper
240	100		
120	11.8	9.4	14.2
236	26.1	20.9	31.3





#77
Benzidine
Concen: 44.967 ng
RT: 12.745 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion:184 Resp: 25289


Ion	Ratio	Lower	Upper
184	100		
185	15.0	12.0	18.0
183	11.8	9.4	14.2

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#78
Pyrene
Concen: 41.241 ng
RT: 12.857 min Scan# 1830
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:202 Resp: 1248309

Ion	Ratio	Lower	Upper
202	100		
200	21.5	17.2	25.8
203	17.8	14.2	21.4






#79
Terphenyl-d14
Concen: 81.103 ng
RT: 12.998 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:244 Resp: 172112


Ion	Ratio	Lower	Upper
244	100		
212	7.1	5.7	8.5
122	10.8	8.6	13.0



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#80
Butylbenzylphthalate
Concen: 42.297 ng
RT: 13.474 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:149 Resp: 383831

Ion	Ratio	Lower	Upper
149	100		
91	74.3	59.4	89.2
206	20.5	16.4	24.6






#81
Benzo(a)anthracene
Concen: 40.901 ng
RT: 14.039 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:228 Resp: 92069


Ion	Ratio	Lower	Upper
228	100		
226	27.0	21.6	32.4
229	20.1	16.1	24.1



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#82
3,3'-Dichlorobenzidine
Concen: 40.798 ng
RT: 14.004 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:252 Resp: 267769

Ion	Ratio	Lower	Upper
252	100		
254	63.5	50.8	76.2
126	11.1	8.9	13.3





#83
Chrysene
Concen: 39.098 ng
RT: 14.080 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion:228 Resp: 80695


Ion	Ratio	Lower	Upper
228	100		
226	30.1	24.1	36.1
229	19.7	15.8	23.6

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#84
Bis(2-ethylhexyl)phthalate
Concen: 42.504 ng
RT: 14.033 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:149 Resp: 432750

Ion	Ratio	Lower	Upper
149	100		
167	26.7	21.4	32.0
279	3.7	3.0	4.4






#85
Di-n-octyl phthalate
Concen: 42.399 ng
RT: 14.651 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:149 Resp: 78517


Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.2	1.8
43	12.2	9.8	14.6



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#86
Perylene-d12
Concen: 20.000 ng
RT: 15.527 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:264 Resp: 365076

Ion	Ratio	Lower	Upper
264	100		
260	24.3	19.4	29.2
265	21.7	17.4	26.0






#87
Indeno(1,2,3-cd)pyrene
Concen: 42.039 ng
RT: 17.021 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:276 Resp: 98734


Ion	Ratio	Lower	Upper
276	100		
138	25.5	20.4	30.6
277	25.3	20.2	30.4



Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#88
Benzo(b)fluoranthene
Concen: 39.263 ng
RT: 15.098 min Scan# 2211
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:252 Resp: 873168

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.6	26.4
125	9.9	7.9	11.9





#89
Benzo(k)fluoranthene
Concen: 40.589 ng
RT: 15.127 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Tgt Ion:252 Resp: 77847


Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.6	26.4
125	9.9	7.9	11.9

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#90
Benzo(a)pyrene
Concen: 40.905 ng
RT: 15.468 min Scan# 2274
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:252 Resp: 748507

Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.3	25.9
125	10.9	8.7	13.1





#91
Dibenzo(a,h)anthracene
Concen: 41.752 ng
RT: 17.045 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20


Instrument :
BNA_F
ClientSampleId :
SSTDICCC040


Tgt Ion:278 Resp: 81804

Ion	Ratio	Lower	Upper
278	100		
139	16.4	13.1	19.7
279	23.7	19.0	28.4

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024






#92
Benzo(g,h,i)perylene
Concen: 42.090 ng
RT: 17.480 min Scan# 2616
Delta R.T. 0.000 min
Lab File: BF139848.D
Acq: 18 Oct 2024 12:20

Tgt Ion:276 Resp: 823727

Ion	Ratio	Lower	Upper
276	100		
277	23.8	19.0	28.6
138	21.7	17.4	26.0



Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139849.D
Acq On : 18 Oct 2024 12:49
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Oct 18 14:19:35 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED
Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.892	152	197235	20.000	ng		0.00
21) Naphthalene-d8	8.175	136	739665	20.000	ng		0.00
39) Acenaphthene-d10	9.927	164	405194	20.000	ng		0.00
64) Phenanthrene-d10	11.416	188	694873	20.000	ng		0.00
76) Chrysene-d12	14.051	240	349496	20.000	ng		0.00
86) Perylene-d12	15.527	264	390757	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.498	112	1162820	92.295	ng		0.00
7) Phenol-d6	6.510	99	1517079	92.972	ng		0.00
23) Nitrobenzene-d5	7.457	82	1310450	98.193	ng		0.00
42) 2,4,6-Tribromophenol	10.716	330	362330	95.600	ng		0.00
45) 2-Fluorobiphenyl	9.251	172	2204460	89.915	ng		0.00
79) Terphenyl-d14	12.998	244	2015874	93.988	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.687	88	281572	47.933	ng		99
3) Pyridine	3.446	79	683175	46.920	ng		99
4) n-Nitrosodimethylamine	3.404	42	385387	49.264	ng		97
6) Aniline	6.557	93	725204	47.686	ng		98
8) 2-Chlorophenol	6.675	128	597810	46.414	ng		98
9) Benzaldehyde	6.439	77	430623	46.911	ng		100
10) Phenol	6.528	94	780406m	45.569	ng		
11) bis(2-Chloroethyl)ether	6.628	93	617014	47.453	ng		99
12) 1,3-Dichlorobenzene	6.834	146	686600	46.439	ng		99
13) 1,4-Dichlorobenzene	6.910	146	686616	46.483	ng		99
14) 1,2-Dichlorobenzene	7.063	146	627851	45.543	ng		99
15) Benzyl Alcohol	7.028	79	569012	47.538	ng		99
16) 2,2'-oxybis(1-Chloropr...	7.169	45	1043921	47.084	ng		100
17) 2-Methylphenol	7.134	107	519035	47.259	ng		99
18) Hexachloroethane	7.404	117	250100	47.480	ng		100
19) n-Nitroso-di-n-propyla...	7.310	70	454173	45.892	ng		99
20) 3+4-Methylphenols	7.292	107	645650	45.961	ng	#	88
22) Acetophenone	7.304	105	835908	46.140	ng		97
24) Nitrobenzene	7.475	77	707970	48.385	ng		100
25) Isophorone	7.710	82	1205277	47.583	ng		99
26) 2-Nitrophenol	7.786	139	292634	53.013	ng		99
27) 2,4-Dimethylphenol	7.822	122	439040	47.699	ng		98
28) bis(2-Chloroethoxy)met...	7.922	93	728968	47.500	ng		100
29) 2,4-Dichlorophenol	8.028	162	502930	47.971	ng		98
30) 1,2,4-Trichlorobenzene	8.116	180	550757	47.476	ng		99
31) Naphthalene	8.198	128	1770616	46.415	ng		100
32) Benzoic acid	7.933	122	426532	53.131	ng		100
33) 4-Chloroaniline	8.245	127	613822	47.189	ng		99
34) Hexachlorobutadiene	8.316	225	343002	47.058	ng		100
35) Caprolactam	8.616	113	162260	48.675	ng		98
36) 4-Chloro-3-methylphenol	8.716	107	553487	47.678	ng		99
37) 2-Methylnaphthalene	8.886	142	1085491	46.462	ng		99
38) 1-Methylnaphthalene	8.986	142	1052879	45.933	ng		100
40) 1,2,4,5-Tetrachloroben...	9.051	216	518960	47.474	ng		99
41) Hexachlorocyclopentadiene	9.039	237	188234	49.488	ng		99
43) 2,4,6-Trichlorophenol	9.157	196	367373	47.468	ng		99

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139849.D
Acq On : 18 Oct 2024 12:49
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Oct 18 14:19:35 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.192	196	385963	48.336	ng		98
46) 1,1'-Biphenyl	9.351	154	1301562	46.143	ng		99
47) 2-Chloronaphthalene	9.375	162	1061417	46.720	ng		99
48) 2-Nitroaniline	9.469	65	358283	51.564	ng		97
49) Acenaphthylene	9.792	152	1526583	46.479	ng		99
50) Dimethylphthalate	9.651	163	1187472	47.050	ng		99
51) 2,6-Dinitrotoluene	9.710	165	277023	50.693	ng		99
52) Acenaphthene	9.963	154	989603	46.576	ng		100
53) 3-Nitroaniline	9.880	138	279467	49.591	ng		100
54) 2,4-Dinitrophenol	9.980	184	102786	48.105	ng	#	18
55) Dibenzofuran	10.133	168	1407211	46.162	ng		99
56) 4-Nitrophenol	10.022	139	221410	51.068	ng		97
57) 2,4-Dinitrotoluene	10.110	165	348628	51.878	ng		97
58) Fluorene	10.480	166	1042268	44.889	ng		100
59) 2,3,4,6-Tetrachlorophenol	10.251	232	300306	47.975	ng		99
60) Diethylphthalate	10.351	149	1180411	47.634	ng		99
61) 4-Chlorophenyl-phenyle...	10.469	204	532654	45.427	ng		98
62) 4-Nitroaniline	10.492	138	266320	50.037	ng		99
63) Azobenzene	10.633	77	1231644	46.881	ng		98
65) 4,6-Dinitro-2-methylph...	10.522	198	156461	55.202	ng		94
66) n-Nitrosodiphenylamine	10.586	169	987295	47.472	ng		100
67) 4-Bromophenyl-phenylether	10.963	248	341790	47.746	ng		97
68) Hexachlorobenzene	11.027	284	376968	46.823	ng		98
69) Atrazine	11.110	200	234998	41.713	ng		99
70) Pentachlorophenol	11.216	266	252064	51.587	ng		98
71) Phenanthrene	11.439	178	1500010	45.700	ng		99
72) Anthracene	11.492	178	1478655	46.172	ng		99
73) Carbazole	11.645	167	1348098	45.263	ng		99
74) Di-n-butylphthalate	11.974	149	1602928	46.583	ng		100
75) Fluoranthene	12.627	202	1462936	44.089	ng		99
77) Benzidine	12.745	184	240940	42.388	ng		99
78) Pyrene	12.857	202	1471893	48.113	ng		100
80) Butylbenzylphthalate	13.474	149	467242	50.944	ng		100
81) Benzo(a)anthracene	14.045	228	1107235	48.667	ng		99
82) 3,3'-Dichlorobenzidine	14.004	252	326706	49.251	ng		98
83) Chrysene	14.080	228	990517	47.484	ng		99
84) Bis(2-ethylhexyl)phtha...	14.033	149	541828	52.655	ng		100
85) Di-n-octyl phthalate	14.651	149	1033054	55.195	ng		100
87) Indeno(1,2,3-cd)pyrene	17.027	276	1268552	50.463	ng		100
88) Benzo(b)fluoranthene	15.098	252	1079928	45.369	ng		100
89) Benzo(k)fluoranthene	15.127	252	1006425	49.026	ng		99
90) Benzo(a)pyrene	15.468	252	951346	48.573	ng		99
91) Dibenzo(a,h)anthracene	17.045	278	1058119	50.456	ng		100
92) Benzo(g,h,i)perylene	17.480	276	1055618	50.394	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139849.D
Acq On : 18 Oct 2024 12:49
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1


Quant Time: Oct 18 14:19:35 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.892 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:152 Resp: 19723


Ion	Ratio	Lower	Upper
152	100		
150	160.0	130.2	195.2
115	63.6	51.4	77.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#2
1,4-Dioxane
Concen: 47.933 ng
RT: 2.687 min Scan# 101
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion: 88 Resp: 281572

Ion	Ratio	Lower	Upper
88	100		
58	91.1	73.3	109.9
43	33.2	27.1	40.7






#3
Pyridine
Concen: 46.920 ng
RT: 3.446 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion: 79 Resp: 68317


Ion	Ratio	Lower	Upper
79	100		
52	86.6	68.5	102.7
51	41.3	32.3	48.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#4
n-Nitrosodimethylamine
Concen: 49.264 ng
RT: 3.404 min Scan# 223
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion: 42 Resp: 385387

Ion	Ratio	Lower	Upper
42	100		
74	116.3	96.2	144.4
44	7.2	6.2	9.2





#5
2-Fluorophenol
Concen: 92.295 ng
RT: 5.498 min Scan# 5
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49


Tgt Ion:112 Resp: 116282


Ion	Ratio	Lower	Upper
112	100		
64	67.1	53.0	79.6
63	34.7	27.0	40.4

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#6
Aniline
Concen: 47.686 ng
RT: 6.557 min Scan# 759
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion: 93 Resp: 725204

Ion	Ratio	Lower	Upper
93	100		
66	41.3	34.0	51.0
65	21.3	17.4	26.0





#7
Phenol-d6
Concen: 92.972 ng
RT: 6.510 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49


Tgt Ion: 99 Resp: 151707


Ion	Ratio	Lower	Upper
99	100		
42	21.0	16.7	25.1
71	35.3	27.7	41.5

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#8
2-Chlorophenol
Concen: 46.414 ng
RT: 6.675 min Scan# 779
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:128 Resp: 597810

Ion	Ratio	Lower	Upper
128	100		
130	32.6	12.3	52.3
64	57.0	34.6	74.6






#9
Benzaldehyde
Concen: 46.911 ng
RT: 6.439 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion: 77 Resp: 43062


Ion	Ratio	Lower	Upper
77	100		
105	95.2	75.5	115.5
106	92.0	71.7	111.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#10
Phenol
Concen: 45.569 ng m
RT: 6.528 min Scan# 754
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion: 94 Resp: 780406

Ion	Ratio	Lower	Upper
94	100		
65	31.6	11.1	51.1
66	43.3	22.7	62.7





#11
bis(2-Chloroethyl)ether
Concen: 47.453 ng
RT: 6.628 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49


Tgt Ion: 93 Resp: 61701


Ion	Ratio	Lower	Upper
93	100		
63	85.8	65.0	105.0
95	32.2	12.4	52.4

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#12
1,3-Dichlorobenzene
Concen: 46.439 ng
RT: 6.834 min Scan# 806
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:146 Resp: 686600

Ion	Ratio	Lower	Upper
146	100		
148	64.7	51.0	76.4
75	33.3	26.4	39.6






#13
1,4-Dichlorobenzene
Concen: 46.483 ng
RT: 6.910 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:146 Resp: 68661


Ion	Ratio	Lower	Upper
146	100		
148	64.1	51.8	77.8
111	43.4	34.6	51.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#14
1,2-Dichlorobenzene
Concen: 45.543 ng
RT: 7.063 min Scan# 845
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:146 Resp: 627851

Ion	Ratio	Lower	Upper
146	100		
148	63.7	51.4	77.2
111	46.3	36.7	55.1





#15
Benzyl Alcohol
Concen: 47.538 ng
RT: 7.028 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49


Tgt Ion: 79 Resp: 56901


Ion	Ratio	Lower	Upper
79	100		
108	75.3	61.4	92.0
77	62.7	49.9	74.9

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#16
2,2'-oxybis(1-Chloropropane)
Concen: 47.084 ng
RT: 7.169 min Scan# 863
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion: 45 Resp: 1043921

Ion	Ratio	Lower	Upper
45	100		
77	11.6	0.0	31.5
79	9.0	0.0	29.1






#17
2-Methylphenol
Concen: 47.259 ng
RT: 7.134 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:107 Resp: 51903


Ion	Ratio	Lower	Upper
107	100		
108	113.7	91.3	136.9
77	46.7	37.1	55.7
79	46.9	37.0	55.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#18
Hexachloroethane
Concen: 47.480 ng
RT: 7.404 min Scan# 903
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:117 Resp: 250100

Ion	Ratio	Lower	Upper
117	100		
119	96.0	77.0	115.4
201	95.3	76.2	114.4






#19
n-Nitroso-di-n-propylamine
Concen: 45.892 ng
RT: 7.310 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion: 70 Resp: 45417


Ion	Ratio	Lower	Upper
70	100		
42	63.4	51.5	77.3
101	9.3	7.5	11.3
130	20.2	15.5	23.3



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#20
3+4-Methylphenols
Concen: 45.961 ng
RT: 7.292 min Scan# 884
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:107 Resp: 645650

Ion	Ratio	Lower	Upper
107	100		
108	90.8	69.8	109.8
77	95.4	50.5	90.5#
79	31.2	10.2	50.2






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.175 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:136 Resp: 73966


Ion	Ratio	Lower	Upper
136	100		
137	11.2	8.6	12.8
54	10.6	8.4	12.6
68	6.7	5.1	7.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#22
Acetophenone
Concen: 46.140 ng
RT: 7.304 min Scan# 886
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:105 Resp: 835908

Ion	Ratio	Lower	Upper
105	100		
71	7.9	5.9	8.9
51	33.7	29.4	44.2
120	22.0	17.5	26.3





#23
Nitrobenzene-d5
Concen: 98.193 ng
RT: 7.457 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49


Tgt Ion: 82 Resp: 131045


Ion	Ratio	Lower	Upper
82	100		
128	42.2	33.4	50.0
54	59.3	47.8	71.8

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#24
Nitrobenzene
Concen: 48.385 ng
RT: 7.475 min Scan# 915
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion: 77 Resp: 707970

Ion	Ratio	Lower	Upper
77	100		
123	44.2	35.7	53.5
65	14.5	11.5	17.3





#25
Isophorone
Concen: 47.583 ng
RT: 7.710 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49


Tgt Ion: 82 Resp: 120527


Ion	Ratio	Lower	Upper
82	100		
95	7.2	5.5	8.3
138	16.1	13.2	19.8

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#26
2-Nitrophenol
Concen: 53.013 ng
RT: 7.786 min Scan# 968
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:139 Resp: 292634

Ion	Ratio	Lower	Upper
139	100		
109	31.4	25.1	37.7
65	58.4	47.5	71.3






#27
2,4-Dimethylphenol
Concen: 47.699 ng
RT: 7.822 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:122 Resp: 43904


Ion	Ratio	Lower	Upper
122	100		
107	113.0	91.9	137.9
121	56.6	46.2	69.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#28
bis(2-Chloroethoxy)methane
Concen: 47.500 ng
RT: 7.922 min Scan# 991
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion: 93 Resp: 728968

Ion	Ratio	Lower	Upper
93	100		
95	32.4	26.0	39.0
123	12.0	9.8	14.8






#29
2,4-Dichlorophenol
Concen: 47.971 ng
RT: 8.028 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:162 Resp: 50293


Ion	Ratio	Lower	Upper
162	100		
164	65.3	43.9	83.9
98	38.6	19.9	59.9



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#30
1,2,4-Trichlorobenzene
Concen: 47.476 ng
RT: 8.116 min Scan# 1024
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:180 Resp: 550757

Ion	Ratio	Lower	Upper
180	100		
182	99.7	80.5	120.7
145	32.4	25.4	38.0





#31
Naphthalene
Concen: 46.415 ng
RT: 8.198 min Scan# 1(
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49


Tgt Ion:128 Resp: 177061(


Ion	Ratio	Lower	Upper
128	100		
129	11.2	8.9	13.3
127	13.3	10.6	16.0

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#32
Benzoic acid
Concen: 53.131 ng
RT: 7.933 min Scan# 993
Delta R.T. 0.011 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:122 Resp: 426532

Ion	Ratio	Lower	Upper
122	100		
105	127.2	107.6	147.6
77	100.6	80.2	120.2






#33
4-Chloroaniline
Concen: 47.189 ng
RT: 8.245 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:127 Resp: 61382


Ion	Ratio	Lower	Upper
127	100		
129	32.7	25.4	38.2
65	34.6	28.4	42.6
92	20.5	16.6	25.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#34
Hexachlorobutadiene
Concen: 47.058 ng
RT: 8.316 min Scan# 1058
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:225 Resp: 343002

Ion	Ratio	Lower	Upper
225	100		
223	63.7	50.6	76.0
227	65.1	51.8	77.8





#35
Caprolactam
Concen: 48.675 ng
RT: 8.616 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49


Tgt Ion:113 Resp: 16226


Ion	Ratio	Lower	Upper
113	100		
55	208.2	189.2	229.2
56	157.0	132.5	172.5

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#36
4-Chloro-3-methylphenol
Concen: 47.678 ng
RT: 8.716 min Scan# 1126
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:107 Resp: 553487

Ion	Ratio	Lower	Upper
107	100		
144	24.3	19.1	28.7
142	72.6	57.2	85.8






#37
2-Methylnaphthalene
Concen: 46.462 ng
RT: 8.886 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:142 Resp: 108549


Ion	Ratio	Lower	Upper
142	100		
141	88.0	70.8	106.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#38
1-Methylnaphthalene
Concen: 45.933 ng
RT: 8.986 min Scan# 1172
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:142 Resp: 1052879

Ion	Ratio	Lower	Upper
142	100		
141	91.8	73.5	110.3






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.927 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:164 Resp: 40519


Ion	Ratio	Lower	Upper
164	100		
162	100.7	81.0	121.4
160	44.4	35.4	53.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#40
1,2,4,5-Tetrachlorobenzene
Concen: 47.474 ng
RT: 9.051 min Scan# 1183
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:216 Resp: 518960

Ion	Ratio	Lower	Upper
216	100		
214	78.3	63.4	95.0
179	21.5	17.2	25.8
108	19.4	15.6	23.4





#41
Hexachlorocyclopentadiene
Concen: 49.488 ng
RT: 9.039 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49


Tgt Ion:237 Resp: 18823


Ion	Ratio	Lower	Upper
237	100		
235	63.8	42.9	82.9
272	12.0	0.0	32.3

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#42
2,4,6-Tribromophenol
Concen: 95.600 ng
RT: 10.716 min Scan# 1466
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:330 Resp: 362330

Ion	Ratio	Lower	Upper
330	100		
332	98.1	78.1	117.1
141	32.6	26.6	39.8






#43
2,4,6-Trichlorophenol
Concen: 47.468 ng
RT: 9.157 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:196 Resp: 36737


Ion	Ratio	Lower	Upper
196	100		
198	94.5	76.2	114.2
200	30.2	24.0	36.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#44
2,4,5-Trichlorophenol
Concen: 48.336 ng
RT: 9.192 min Scan# 1207
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:196 Resp: 385963

Ion	Ratio	Lower	Upper
196	100		
198	96.2	75.8	113.6
97	54.3	41.8	62.6
132	28.5	22.9	34.3






#45
2-Fluorobiphenyl
Concen: 89.915 ng
RT: 9.251 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:172 Resp: 220446


Ion	Ratio	Lower	Upper
172	100		
171	35.9	28.6	43.0
170	24.2	19.1	28.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#46
1,1'-Biphenyl
Concen: 46.143 ng
RT: 9.351 min Scan# 1234
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:154 Resp: 1301562

Ion	Ratio	Lower	Upper
154	100		
153	40.7	20.3	60.3
76	14.1	0.0	33.9






#47
2-Chloronaphthalene
Concen: 46.720 ng
RT: 9.375 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:162 Resp: 106141


Ion	Ratio	Lower	Upper
162	100		
127	43.3	33.8	50.6
164	32.7	26.6	39.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#48
2-Nitroaniline
Concen: 51.564 ng
RT: 9.469 min Scan# 1254
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion: 65 Resp: 358283

Ion	Ratio	Lower	Upper
65	100		
92	64.6	50.9	76.3
138	96.2	73.4	110.0






#49
Acenaphthylene
Concen: 46.479 ng
RT: 9.792 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:152 Resp: 152658


Ion	Ratio	Lower	Upper
152	100		
151	20.4	15.9	23.9
153	13.4	10.7	16.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#50
Dimethylphthalate
Concen: 47.050 ng
RT: 9.651 min Scan# 1285
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:163 Resp: 1187472

Ion	Ratio	Lower	Upper
163	100		
194	3.9	3.4	5.0
164	10.4	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 50.693 ng
RT: 9.710 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49


Tgt Ion:165 Resp: 27702


Ion	Ratio	Lower	Upper
165	100		
63	65.3	50.6	75.8
89	59.5	47.5	71.3

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#52
Acenaphthene
Concen: 46.576 ng
RT: 9.963 min Scan# 1338
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:154 Resp: 989603

Ion	Ratio	Lower	Upper
154	100		
153	112.4	89.8	134.6
152	53.4	42.4	63.6





#53
3-Nitroaniline
Concen: 49.591 ng
RT: 9.880 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49


Tgt Ion:138 Resp: 27946


Ion	Ratio	Lower	Upper
138	100		
108	12.2	10.2	15.2
92	123.1	98.3	147.5

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#54
2,4-Dinitrophenol
Concen: 48.105 ng
RT: 9.980 min Scan# 1341
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:184 Resp: 102786

Ion	Ratio	Lower	Upper
184	100		
63	84.8	84.7	127.1
154	182.9	338.2	507.2#






#55
Dibenzofuran
Concen: 46.162 ng
RT: 10.133 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:168 Resp: 140721


Ion	Ratio	Lower	Upper
168	100		
139	40.1	31.8	47.6
169	14.7	11.6	17.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#56
4-Nitrophenol
Concen: 51.068 ng
RT: 10.022 min Scan# 1348
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:139 Resp: 221410

Ion	Ratio	Lower	Upper
139	100		
109	74.0	53.6	93.6
65	112.1	87.8	127.8






#57
2,4-Dinitrotoluene
Concen: 51.878 ng
RT: 10.110 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:165 Resp: 34862


Ion	Ratio	Lower	Upper
165	100		
63	52.2	40.1	60.1
89	78.4	60.4	90.6
182	2.6	2.0	3.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#58
Fluorene
Concen: 44.889 ng
RT: 10.480 min Scan# 1426
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:166 Resp: 1042268

Ion	Ratio	Lower	Upper
166	100		
165	98.6	78.5	117.7
167	13.5	10.9	16.3






#59
2,3,4,6-Tetrachlorophenol
Concen: 47.975 ng
RT: 10.251 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:232 Resp: 30030


Ion	Ratio	Lower	Upper
232	100		
131	45.6	37.0	55.6
130	2.4	1.9	2.9
166	29.3	24.2	36.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#60
Diethylphthalate
Concen: 47.634 ng
RT: 10.351 min Scan# 1404
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:149 Resp: 1180411

Ion	Ratio	Lower	Upper
149	100		
177	21.9	17.9	26.9
150	12.5	10.1	15.1





#61
4-Chlorophenyl-phenylether
Concen: 45.427 ng
RT: 10.469 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49


Tgt Ion:204 Resp: 53265


Ion	Ratio	Lower	Upper
204	100		
206	32.9	26.3	39.5
141	62.9	48.6	73.0

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#62
4-Nitroaniline
Concen: 50.037 ng
RT: 10.492 min Scan# 1428
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:138 Resp: 266320

Ion	Ratio	Lower	Upper
138	100		
92	55.4	35.7	75.7
108	78.6	57.5	97.5






#63
Azobenzene
Concen: 46.881 ng
RT: 10.633 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion: 77 Resp: 123164


Ion	Ratio	Lower	Upper
77	100		
182	25.4	3.3	43.3
105	21.5	1.2	41.2
51	33.2	13.3	53.3



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.416 min Scan# 1585
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:188 Resp: 694873

Ion	Ratio	Lower	Upper
188	100		
94	10.2	7.9	11.9
80	11.1	9.0	13.4





#65
4,6-Dinitro-2-methylphenol
Concen: 55.202 ng
RT: 10.522 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49


Tgt Ion:198 Resp: 15646


Ion	Ratio	Lower	Upper
198	100		
51	54.1	38.2	78.2
105	43.9	27.9	67.9

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#66
n-Nitrosodiphenylamine
Concen: 47.472 ng
RT: 10.586 min Scan# 1444
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:169 Resp: 987295

Ion	Ratio	Lower	Upper
169	100		
168	62.4	49.8	74.6
167	34.6	28.0	42.0






#67
4-Bromophenyl-phenylether
Concen: 47.746 ng
RT: 10.963 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:248 Resp: 34179


Ion	Ratio	Lower	Upper
248	100		
250	98.1	79.0	118.6
141	72.9	61.7	92.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#68
Hexachlorobenzene
Concen: 46.823 ng
RT: 11.027 min Scan# 1519
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:284 Resp: 376968

Ion	Ratio	Lower	Upper
284	100		
142	32.2	26.9	40.3
249	31.3	25.3	37.9






#69
Atrazine
Concen: 41.713 ng
RT: 11.110 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:200 Resp: 23499


Ion	Ratio	Lower	Upper
200	100		
173	27.4	7.7	47.7
215	46.8	27.7	67.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#70
Pentachlorophenol
Concen: 51.587 ng
RT: 11.216 min Scan# 1551
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:266 Resp: 252064

Ion	Ratio	Lower	Upper
266	100		
268	64.2	50.4	75.6
264	61.7	50.6	75.8






#71
Phenanthrene
Concen: 45.700 ng
RT: 11.439 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:178 Resp: 150001


Ion	Ratio	Lower	Upper
178	100		
176	20.1	15.8	23.6
179	16.0	12.6	18.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#72
Anthracene
Concen: 46.172 ng
RT: 11.492 min Scan# 1598
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:178 Resp: 1478655

Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.3	22.9
179	15.8	12.4	18.6






#73
Carbazole
Concen: 45.263 ng
RT: 11.645 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:167 Resp: 134809


Ion	Ratio	Lower	Upper
167	100		
166	21.9	17.0	25.6
139	13.5	10.8	16.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#74
Di-n-butylphthalate
Concen: 46.583 ng
RT: 11.974 min Scan# 1680
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:149 Resp: 1602928

Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.5	11.3
104	5.2	4.1	6.1






#75
Fluoranthene
Concen: 44.089 ng
RT: 12.627 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:202 Resp: 146293


Ion	Ratio	Lower	Upper
202	100		
101	12.2	0.0	31.9
203	17.7	0.0	37.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.051 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:240 Resp: 349496

Ion	Ratio	Lower	Upper
240	100		
120	12.0	9.4	14.2
236	26.0	20.9	31.3






#77
Benzidine
Concen: 42.388 ng
RT: 12.745 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:184 Resp: 24094


Ion	Ratio	Lower	Upper
184	100		
185	15.3	12.0	18.0
183	11.5	9.4	14.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#78
Pyrene
Concen: 48.113 ng
RT: 12.857 min Scan# 1830
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:202 Resp: 1471893

Ion	Ratio	Lower	Upper
202	100		
200	21.8	17.2	25.8
203	17.8	14.2	21.4






#79
Terphenyl-d14
Concen: 93.988 ng
RT: 12.998 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:244 Resp: 201587


Ion	Ratio	Lower	Upper
244	100		
212	7.1	5.7	8.5
122	11.0	8.6	13.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#80
Butylbenzylphthalate
Concen: 50.944 ng
RT: 13.474 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:149 Resp: 467242

Ion	Ratio	Lower	Upper
149	100		
91	74.5	59.4	89.2
206	20.2	16.4	24.6





#81
Benzo(a)anthracene
Concen: 48.667 ng
RT: 14.045 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49


Tgt Ion:228 Resp: 110723


Ion	Ratio	Lower	Upper
228	100		
226	27.1	21.6	32.4
229	19.7	16.1	24.1

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#82
3,3'-Dichlorobenzidine
Concen: 49.251 ng
RT: 14.004 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:252 Resp: 326706

Ion	Ratio	Lower	Upper
252	100		
254	65.6	50.8	76.2
126	11.0	8.9	13.3






#83
Chrysene
Concen: 47.484 ng
RT: 14.080 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:228 Resp: 99051


Ion	Ratio	Lower	Upper
228	100		
226	30.5	24.1	36.1
229	20.1	15.8	23.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#84
Bis(2-ethylhexyl)phthalate
Concen: 52.655 ng
RT: 14.033 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:149 Resp: 541828

Ion	Ratio	Lower	Upper
149	100		
167	26.7	21.4	32.0
279	4.1	3.0	4.4






#85
Di-n-octyl phthalate
Concen: 55.195 ng
RT: 14.651 min Scan# 21
Delta R.T. -0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:149 Resp: 103305


Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.2	1.8
43	12.4	9.8	14.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#86
Perylene-d12
Concen: 20.000 ng
RT: 15.527 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:264 Resp: 390757

Ion	Ratio	Lower	Upper
264	100		
260	23.7	19.4	29.2
265	21.7	17.4	26.0






#87
Indeno(1,2,3-cd)pyrene
Concen: 50.463 ng
RT: 17.027 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:276 Resp: 126855


Ion	Ratio	Lower	Upper
276	100		
138	25.2	20.4	30.6
277	25.4	20.2	30.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#88
Benzo(b)fluoranthene
Concen: 45.369 ng
RT: 15.098 min Scan# 2211
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:252 Resp: 1079928

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.6	26.4
125	10.0	7.9	11.9





#89
Benzo(k)fluoranthene
Concen: 49.026 ng
RT: 15.127 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49


Tgt Ion:252 Resp: 100642


Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.6	26.4
125	10.0	7.9	11.9

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#90
Benzo(a)pyrene
Concen: 48.573 ng
RT: 15.468 min Scan# 2274
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:252 Resp: 951346

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.3	25.9
125	11.1	8.7	13.1






#91
Dibenzo(a,h)anthracene
Concen: 50.456 ng
RT: 17.045 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:278 Resp: 105811


Ion	Ratio	Lower	Upper
278	100		
139	16.2	13.1	19.7
279	23.6	19.0	28.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#92
Benzo(g,h,i)perylene
Concen: 50.394 ng
RT: 17.480 min Scan# 2616
Delta R.T. 0.000 min
Lab File: BF139849.D
Acq: 18 Oct 2024 12:49

Tgt Ion:276 Resp: 1055618

Ion	Ratio	Lower	Upper
276	100		
277	23.5	19.0	28.6
138	22.4	17.4	26.0



Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139850.D
Acq On : 18 Oct 2024 13:17
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Oct 18 14:20:24 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.893	152	189346	20.000	ng	0.00
21) Naphthalene-d8	8.175	136	712033	20.000	ng	0.00
39) Acenaphthene-d10	9.928	164	389299	20.000	ng	0.00
64) Phenanthrene-d10	11.416	188	667372	20.000	ng	0.00
76) Chrysene-d12	14.057	240	337032	20.000	ng	0.00
86) Perylene-d12	15.527	264	377203	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.499	112	1347774	111.432	ng	0.00
7) Phenol-d6	6.516	99	1748088	111.592	ng	0.00
23) Nitrobenzene-d5	7.457	82	1523975	118.624	ng	0.00
42) 2,4,6-Tribromophenol	10.722	330	420891	115.586	ng	0.00
45) 2-Fluorobiphenyl	9.251	172	2475821	105.106	ng	0.00
79) Terphenyl-d14	13.004	244	2266127	109.563	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.687	88	330127	58.541	ng	98
3) Pyridine	3.446	79	798529	57.127	ng	99
4) n-Nitrosodimethylamine	3.410	42	451209	60.081	ng	98
6) Aniline	6.551	93	840113	57.544	ng	100
8) 2-Chlorophenol	6.675	128	693549	56.091	ng	96
9) Benzaldehyde	6.440	77	484510	54.980	ng	99
10) Phenol	6.528	94	909456m	55.317	ng	
11) bis(2-Chloroethyl)ether	6.628	93	709494	56.838	ng	99
12) 1,3-Dichlorobenzene	6.834	146	790296	55.679	ng	98
13) 1,4-Dichlorobenzene	6.910	146	790397	55.738	ng	99
14) 1,2-Dichlorobenzene	7.063	146	719470	54.363	ng	99
15) Benzyl Alcohol	7.028	79	650971	56.651	ng	98
16) 2,2'-oxybis(1-Chloropr...	7.169	45	1201584	56.453	ng	99
17) 2-Methylphenol	7.134	107	603882	57.276	ng	99
18) Hexachloroethane	7.404	117	290250	57.398	ng	100
19) n-Nitroso-di-n-propyla...	7.310	70	526516	55.418	ng	99
20) 3+4-Methylphenols	7.293	107	738542	54.764	ng	93
22) Acetophenone	7.304	105	969975	55.618	ng	99
24) Nitrobenzene	7.475	77	828015	58.785	ng	98
25) Isophorone	7.716	82	1409786	57.816	ng	100
26) 2-Nitrophenol	7.787	139	344300	64.793	ng	100
27) 2,4-Dimethylphenol	7.822	122	500332	56.468	ng	99
28) bis(2-Chloroethoxy)met...	7.922	93	832755	56.369	ng	99
29) 2,4-Dichlorophenol	8.028	162	584871	57.952	ng	99
30) 1,2,4-Trichlorobenzene	8.116	180	635946	56.947	ng	99
31) Naphthalene	8.198	128	2016330	54.907	ng	99
32) Benzoic acid	7.945	122	502169	64.980	ng	98
33) 4-Chloroaniline	8.245	127	695566	55.548	ng	99
34) Hexachlorobutadiene	8.316	225	399483	56.934	ng	99
35) Caprolactam	8.622	113	188948	58.880	ng	97
36) 4-Chloro-3-methylphenol	8.716	107	646732	57.872	ng	100
37) 2-Methylnaphthalene	8.887	142	1245977	55.401	ng	99
38) 1-Methylnaphthalene	8.987	142	1211981	54.926	ng	100
40) 1,2,4,5-Tetrachloroben...	9.051	216	591461	56.315	ng	100
41) Hexachlorocyclopentadiene	9.039	237	216567	59.262	ng	100
43) 2,4,6-Trichlorophenol	9.157	196	447702	60.209	ng	99

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139850.D
Acq On : 18 Oct 2024 13:17
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Oct 18 14:20:24 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.198	196	431080	56.191	ng		99
46) 1,1'-Biphenyl	9.351	154	1474193	54.396	ng		99
47) 2-Chloronaphthalene	9.375	162	1214856	55.657	ng		98
48) 2-Nitroaniline	9.469	65	423080	63.375	ng		97
49) Acenaphthylene	9.792	152	1758248	55.718	ng		99
50) Dimethylphthalate	9.657	163	1358812	56.037	ng		100
51) 2,6-Dinitrotoluene	9.710	165	320076	60.963	ng		97
52) Acenaphthene	9.969	154	1148095	56.242	ng		98
53) 3-Nitroaniline	9.881	138	328790	60.726	ng		98
54) 2,4-Dinitrophenol	9.981	184	131611	61.938	ng	#	29
55) Dibenzofuran	10.139	168	1605572	54.819	ng		100
56) 4-Nitrophenol	10.028	139	256728	61.631	ng		100
57) 2,4-Dinitrotoluene	10.116	165	410361	63.558	ng		99
58) Fluorene	10.481	166	1192018	53.435	ng		100
59) 2,3,4,6-Tetrachlorophenol	10.251	232	342475	56.946	ng		99
60) Diethylphthalate	10.351	149	1355992	56.954	ng		99
61) 4-Chlorophenyl-phenyle...	10.469	204	606730	53.858	ng		98
62) 4-Nitroaniline	10.498	138	314525	61.507	ng		98
63) Azobenzene	10.633	77	1415762	56.090	ng		99
65) 4,6-Dinitro-2-methylph...	10.522	198	184974	67.951	ng		99
66) n-Nitrosodiphenylamine	10.592	169	1122831	56.214	ng		100
67) 4-Bromophenyl-phenylether	10.963	248	392570	57.100	ng		98
68) Hexachlorobenzene	11.028	284	442187	57.187	ng		98
69) Atrazine	11.116	200	307149	56.767	ng		99
70) Pentachlorophenol	11.216	266	289014	61.587	ng		99
71) Phenanthrene	11.439	178	1721763	54.618	ng		99
72) Anthracene	11.492	178	1667221	54.206	ng		99
73) Carbazole	11.645	167	1545085	54.015	ng		99
74) Di-n-butylphthalate	11.975	149	1819929	55.069	ng		100
75) Fluoranthene	12.627	202	1671405	52.448	ng		99
77) Benzidine	12.745	184	205548	37.499	ng		99
78) Pyrene	12.857	202	1667489	56.522	ng		99
80) Butylbenzylphthalate	13.474	149	537039	60.719	ng		98
81) Benzo(a)anthracene	14.045	228	1250382	56.992	ng		100
82) 3,3'-Dichlorobenzidine	14.004	252	386072	60.353	ng		99
83) Chrysene	14.080	228	1163808	57.855	ng		100
84) Bis(2-ethylhexyl)phtha...	14.033	149	632594	63.749	ng		100
85) Di-n-octyl phthalate	14.651	149	1220737	67.634	ng		100
87) Indeno(1,2,3-cd)pyrene	17.027	276	1500084	61.817	ng		99
88) Benzo(b)fluoranthene	15.098	252	1402395	61.033	ng		100
89) Benzo(k)fluoranthene	15.127	252	1051303	53.053	ng		100
90) Benzo(a)pyrene	15.468	252	1129942	59.764	ng		100
91) Dibenzo(a,h)anthracene	17.051	278	1228124	60.667	ng		99
92) Benzo(g,h,i)perylene	17.486	276	1238955	61.272	ng		100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139850.D
Acq On : 18 Oct 2024 13:17
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1


Quant Time: Oct 18 14:20:24 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.893 min Scan# 8
Delta R.T. 0.001 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:152 Resp: 18934


Ion	Ratio	Lower	Upper
152	100		
150	162.7	130.2	195.2
115	63.2	51.4	77.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#2
1,4-Dioxane
Concen: 58.541 ng
RT: 2.687 min Scan# 101
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion: 88 Resp: 330127

Ion	Ratio	Lower	Upper
88	100		
58	89.7	73.3	109.9
43	33.0	27.1	40.7






#3
Pyridine
Concen: 57.127 ng
RT: 3.446 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion: 79 Resp: 79852


Ion	Ratio	Lower	Upper
79	100		
52	86.8	68.5	102.7
51	41.4	32.3	48.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#4
n-Nitrosodimethylamine
Concen: 60.081 ng
RT: 3.410 min Scan# 224
Delta R.T. 0.012 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion: 42 Resp: 451209

Ion	Ratio	Lower	Upper
42	100		
74	118.0	96.2	144.4
44	7.1	6.2	9.2





#5
2-Fluorophenol
Concen: 111.432 ng
RT: 5.499 min Scan# 5
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17


Tgt Ion:112 Resp: 134777


Ion	Ratio	Lower	Upper
112	100		
64	67.5	53.0	79.6
63	34.5	27.0	40.4

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#6
Aniline
Concen: 57.544 ng
RT: 6.551 min Scan# 758
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion: 93 Resp: 840113

Ion	Ratio	Lower	Upper
93	100		
66	42.2	34.0	51.0
65	21.7	17.4	26.0






#7
Phenol-d6
Concen: 111.592 ng
RT: 6.516 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion: 99 Resp: 174808


Ion	Ratio	Lower	Upper
99	100		
42	20.9	16.7	25.1
71	35.5	27.7	41.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#8
2-Chlorophenol
Concen: 56.091 ng
RT: 6.675 min Scan# 779
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:128 Resp: 693549

Ion	Ratio	Lower	Upper
128	100		
130	32.5	12.3	52.3
64	58.5	34.6	74.6






#9
Benzaldehyde
Concen: 54.980 ng
RT: 6.440 min Scan# 7
Delta R.T. 0.001 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion: 77 Resp: 48451


Ion	Ratio	Lower	Upper
77	100		
105	95.6	75.5	115.5
106	92.7	71.7	111.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#10
Phenol
Concen: 55.317 ng m
RT: 6.528 min Scan# 754
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion: 94 Resp: 909456

Ion	Ratio	Lower	Upper
94	100		
65	32.2	11.1	51.1
66	44.9	22.7	62.7






#11
bis(2-Chloroethyl)ether
Concen: 56.838 ng
RT: 6.628 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion: 93 Resp: 70949


Ion	Ratio	Lower	Upper
93	100		
63	86.7	65.0	105.0
95	32.4	12.4	52.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#12
1,3-Dichlorobenzene
Concen: 55.679 ng
RT: 6.834 min Scan# 806
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:146 Resp: 790296

Ion	Ratio	Lower	Upper
146	100		
148	65.1	51.0	76.4
75	34.3	26.4	39.6






#13
1,4-Dichlorobenzene
Concen: 55.738 ng
RT: 6.910 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:146 Resp: 79039


Ion	Ratio	Lower	Upper
146	100		
148	64.4	51.8	77.8
111	43.7	34.6	51.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#14
1,2-Dichlorobenzene
Concen: 54.363 ng
RT: 7.063 min Scan# 845
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:146 Resp: 719470

Ion	Ratio	Lower	Upper
146	100		
148	63.8	51.4	77.2
111	46.8	36.7	55.1






#15
Benzyl Alcohol
Concen: 56.651 ng
RT: 7.028 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion: 79 Resp: 65097


Ion	Ratio	Lower	Upper
79	100		
108	74.3	61.4	92.0
77	61.5	49.9	74.9



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#16
2,2'-oxybis(1-Chloropropane)
Concen: 56.453 ng
RT: 7.169 min Scan# 863
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion: 45 Resp: 1201584

Ion	Ratio	Lower	Upper
45	100		
77	11.9	0.0	31.5
79	9.3	0.0	29.1






#17
2-Methylphenol
Concen: 57.276 ng
RT: 7.134 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:107 Resp: 60388


Ion	Ratio	Lower	Upper
107	100		
108	113.7	91.3	136.9
77	48.0	37.1	55.7
79	47.6	37.0	55.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#18
Hexachloroethane
Concen: 57.398 ng
RT: 7.404 min Scan# 903
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:117 Resp: 290250

Ion	Ratio	Lower	Upper
117	100		
119	96.0	77.0	115.4
201	94.9	76.2	114.4






#19
n-Nitroso-di-n-propylamine
Concen: 55.418 ng
RT: 7.310 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion: 70 Resp: 52651


Ion	Ratio	Lower	Upper
70	100		
42	63.7	51.5	77.3
101	9.7	7.5	11.3
130	19.4	15.5	23.3



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#20
3+4-Methylphenols
Concen: 54.764 ng
RT: 7.293 min Scan# 884
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:107 Resp: 738542

Ion	Ratio	Lower	Upper
107	100		
108	89.8	69.8	109.8
77	86.6	50.5	90.5
79	30.9	10.2	50.2






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.175 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:136 Resp: 71203


Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.6	12.8
54	10.6	8.4	12.6
68	6.7	5.1	7.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#22
Acetophenone
Concen: 55.618 ng
RT: 7.304 min Scan# 886
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:105 Resp: 969975

Ion	Ratio	Lower	Upper
105	100		
71	7.0	5.9	8.9
51	35.8	29.4	44.2
120	21.5	17.5	26.3





#23
Nitrobenzene-d5
Concen: 118.624 ng
RT: 7.457 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17


Tgt Ion: 82 Resp: 152397


Ion	Ratio	Lower	Upper
82	100		
128	42.1	33.4	50.0
54	59.3	47.8	71.8

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#24
Nitrobenzene
Concen: 58.785 ng
RT: 7.475 min Scan# 915
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion: 77 Resp: 828015

Ion	Ratio	Lower	Upper
77	100		
123	43.4	35.7	53.5
65	14.9	11.5	17.3





#25
Isophorone
Concen: 57.816 ng
RT: 7.716 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17


Tgt Ion: 82 Resp: 140978


Ion	Ratio	Lower	Upper
82	100		
95	7.1	5.5	8.3
138	16.7	13.2	19.8

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#26
2-Nitrophenol
Concen: 64.793 ng
RT: 7.787 min Scan# 968
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:139 Resp: 344300

Ion	Ratio	Lower	Upper
139	100		
109	31.6	25.1	37.7
65	59.3	47.5	71.3






#27
2,4-Dimethylphenol
Concen: 56.468 ng
RT: 7.822 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:122 Resp: 50033


Ion	Ratio	Lower	Upper
122	100		
107	115.6	91.9	137.9
121	58.3	46.2	69.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#28
bis(2-Chloroethoxy)methane
Concen: 56.369 ng
RT: 7.922 min Scan# 991
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion: 93 Resp: 832755

Ion	Ratio	Lower	Upper
93	100		
95	32.9	26.0	39.0
123	11.8	9.8	14.8






#29
2,4-Dichlorophenol
Concen: 57.952 ng
RT: 8.028 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:162 Resp: 58487


Ion	Ratio	Lower	Upper
162	100		
164	63.9	43.9	83.9
98	38.9	19.9	59.9



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#30
1,2,4-Trichlorobenzene
Concen: 56.947 ng
RT: 8.116 min Scan# 1024
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:180 Resp: 635946

Ion	Ratio	Lower	Upper
180	100		
182	99.7	80.5	120.7
145	32.6	25.4	38.0





#31
Naphthalene
Concen: 54.907 ng
RT: 8.198 min Scan# 1(
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17


Tgt Ion:128 Resp: 201633(


Ion	Ratio	Lower	Upper
128	100		
129	11.5	8.9	13.3
127	13.7	10.6	16.0

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#32
Benzoic acid
Concen: 64.980 ng
RT: 7.945 min Scan# 995
Delta R.T. 0.023 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:122 Resp: 502169

Ion	Ratio	Lower	Upper
122	100		
105	127.1	107.6	147.6
77	96.2	80.2	120.2






#33
4-Chloroaniline
Concen: 55.548 ng
RT: 8.245 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:127 Resp: 69556


Ion	Ratio	Lower	Upper
127	100		
129	32.3	25.4	38.2
65	35.7	28.4	42.6
92	20.7	16.6	25.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#34
Hexachlorobutadiene
Concen: 56.934 ng
RT: 8.316 min Scan# 1058
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:225 Resp: 399483

Ion	Ratio	Lower	Upper
225	100		
223	62.8	50.6	76.0
227	63.6	51.8	77.8





#35
Caprolactam
Concen: 58.880 ng
RT: 8.622 min Scan# 11
Delta R.T. 0.012 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17


Tgt Ion:113 Resp: 18894


Ion	Ratio	Lower	Upper
113	100		
55	213.0	189.2	229.2
56	156.6	132.5	172.5

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#36
4-Chloro-3-methylphenol
Concen: 57.872 ng
RT: 8.716 min Scan# 1126
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:107 Resp: 646732

Ion	Ratio	Lower	Upper
107	100		
144	24.3	19.1	28.7
142	71.2	57.2	85.8





#37
2-Methylnaphthalene
Concen: 55.401 ng
RT: 8.887 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17


Tgt Ion:142 Resp: 124597


Ion	Ratio	Lower	Upper
142	100		
141	87.7	70.8	106.2

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#38
1-Methylnaphthalene
Concen: 54.926 ng
RT: 8.987 min Scan# 1172
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:142 Resp: 1211981

Ion	Ratio	Lower	Upper
142	100		
141	91.6	73.5	110.3






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.928 min Scan# 1
Delta R.T. 0.001 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:164 Resp: 38929


Ion	Ratio	Lower	Upper
164	100		
162	100.1	81.0	121.4
160	43.3	35.4	53.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#40
1,2,4,5-Tetrachlorobenzene
Concen: 56.315 ng
RT: 9.051 min Scan# 1183
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:216 Resp: 591461

Ion	Ratio	Lower	Upper
216	100		
214	79.6	63.4	95.0
179	21.5	17.2	25.8
108	19.6	15.6	23.4






#41
Hexachlorocyclopentadiene
Concen: 59.262 ng
RT: 9.039 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:237 Resp: 21656


Ion	Ratio	Lower	Upper
237	100		
235	62.9	42.9	82.9
272	12.5	0.0	32.3



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#42
2,4,6-Tribromophenol
Concen: 115.586 ng
RT: 10.722 min Scan# 1467
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:330 Resp: 420891

Ion	Ratio	Lower	Upper
330	100		
332	95.9	78.1	117.1
141	32.0	26.6	39.8






#43
2,4,6-Trichlorophenol
Concen: 60.209 ng
RT: 9.157 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:196 Resp: 44770


Ion	Ratio	Lower	Upper
196	100		
198	94.1	76.2	114.2
200	29.6	24.0	36.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#44
2,4,5-Trichlorophenol
Concen: 56.191 ng
RT: 9.198 min Scan# 1208
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:196 Resp: 431080

Ion	Ratio	Lower	Upper
196	100		
198	94.5	75.8	113.6
97	50.5	41.8	62.6
132	28.0	22.9	34.3






#45
2-Fluorobiphenyl
Concen: 105.106 ng
RT: 9.251 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:172 Resp: 247582


Ion	Ratio	Lower	Upper
172	100		
171	36.2	28.6	43.0
170	24.3	19.1	28.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#46
1,1'-Biphenyl
Concen: 54.396 ng
RT: 9.351 min Scan# 1234
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:154 Resp: 1474193

Ion	Ratio	Lower	Upper
154	100		
153	40.9	20.3	60.3
76	14.5	0.0	33.9






#47
2-Chloronaphthalene
Concen: 55.657 ng
RT: 9.375 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:162 Resp: 121485


Ion	Ratio	Lower	Upper
162	100		
127	43.7	33.8	50.6
164	32.9	26.6	39.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#48
2-Nitroaniline
Concen: 63.375 ng
RT: 9.469 min Scan# 1254
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion: 65 Resp: 423080

Ion	Ratio	Lower	Upper
65	100		
92	64.6	50.9	76.3
138	95.8	73.4	110.0






#49
Acenaphthylene
Concen: 55.718 ng
RT: 9.792 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:152 Resp: 175824


Ion	Ratio	Lower	Upper
152	100		
151	20.5	15.9	23.9
153	13.4	10.7	16.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#50
Dimethylphthalate
Concen: 56.037 ng
RT: 9.657 min Scan# 1286
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:163 Resp: 1358812

Ion	Ratio	Lower	Upper
163	100		
194	4.0	3.4	5.0
164	10.1	8.2	12.2






#51
2,6-Dinitrotoluene
Concen: 60.963 ng
RT: 9.710 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:165 Resp: 32007


Ion	Ratio	Lower	Upper
165	100		
63	66.3	50.6	75.8
89	61.0	47.5	71.3



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#52
Acenaphthene
Concen: 56.242 ng
RT: 9.969 min Scan# 1339
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:154 Resp: 1148095

Ion	Ratio	Lower	Upper
154	100		
153	109.6	89.8	134.6
152	52.5	42.4	63.6





#53
3-Nitroaniline
Concen: 60.726 ng
RT: 9.881 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17


Tgt Ion:138 Resp: 32879


Ion	Ratio	Lower	Upper
138	100		
108	12.0	10.2	15.2
92	120.8	98.3	147.5

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#54
2,4-Dinitrophenol
Concen: 61.938 ng
RT: 9.981 min Scan# 1341
Delta R.T. 0.007 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:184 Resp: 131611

Ion	Ratio	Lower	Upper
184	100		
63	88.4	84.7	127.1
154	215.7	338.2	507.2#






#55
Dibenzofuran
Concen: 54.819 ng
RT: 10.139 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:168 Resp: 160557


Ion	Ratio	Lower	Upper
168	100		
139	39.7	31.8	47.6
169	14.4	11.6	17.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#56
4-Nitrophenol
Concen: 61.631 ng
RT: 10.028 min Scan# 1349
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:139 Resp: 256728

Ion	Ratio	Lower	Upper
139	100		
109	74.0	53.6	93.6
65	108.1	87.8	127.8






#57
2,4-Dinitrotoluene
Concen: 63.558 ng
RT: 10.116 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:165 Resp: 41036


Ion	Ratio	Lower	Upper
165	100		
63	50.6	40.1	60.1
89	74.9	60.4	90.6
182	2.7	2.0	3.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#58
Fluorene
Concen: 53.435 ng
RT: 10.481 min Scan# 1426
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:166 Resp: 1192018

Ion	Ratio	Lower	Upper
166	100		
165	98.0	78.5	117.7
167	13.5	10.9	16.3






#59
2,3,4,6-Tetrachlorophenol
Concen: 56.946 ng
RT: 10.251 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:232 Resp: 34247


Ion	Ratio	Lower	Upper
232	100		
131	46.7	37.0	55.6
130	2.4	1.9	2.9
166	29.2	24.2	36.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#60
Diethylphthalate
Concen: 56.954 ng
RT: 10.351 min Scan# 1404
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:149 Resp: 1355992

Ion	Ratio	Lower	Upper
149	100		
177	22.0	17.9	26.9
150	12.8	10.1	15.1





#61
4-Chlorophenyl-phenylether
Concen: 53.858 ng
RT: 10.469 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17


Tgt Ion:204 Resp: 60673


Ion	Ratio	Lower	Upper
204	100		
206	32.9	26.3	39.5
141	63.3	48.6	73.0

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#62
4-Nitroaniline
Concen: 61.507 ng
RT: 10.498 min Scan# 1429
Delta R.T. 0.012 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:138 Resp: 314525

Ion	Ratio	Lower	Upper
138	100		
92	53.4	35.7	75.7
108	78.2	57.5	97.5






#63
Azobenzene
Concen: 56.090 ng
RT: 10.633 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion: 77 Resp: 141576


Ion	Ratio	Lower	Upper
77	100		
182	24.9	3.3	43.3
105	21.6	1.2	41.2
51	33.4	13.3	53.3



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.416 min Scan# 1585
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:188 Resp: 667372

Ion	Ratio	Lower	Upper
188	100		
94	10.0	7.9	11.9
80	10.9	9.0	13.4






#65
4,6-Dinitro-2-methylphenol
Concen: 67.951 ng
RT: 10.522 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:198 Resp: 18497


Ion	Ratio	Lower	Upper
198	100		
51	57.4	38.2	78.2
105	47.0	27.9	67.9



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#66
n-Nitrosodiphenylamine
Concen: 56.214 ng
RT: 10.592 min Scan# 1445
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:169 Resp: 1122831

Ion	Ratio	Lower	Upper
169	100		
168	62.5	49.8	74.6
167	34.8	28.0	42.0






#67
4-Bromophenyl-phenylether
Concen: 57.100 ng
RT: 10.963 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:248 Resp: 39257


Ion	Ratio	Lower	Upper
248	100		
250	98.7	79.0	118.6
141	72.9	61.7	92.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#68
Hexachlorobenzene
Concen: 57.187 ng
RT: 11.028 min Scan# 1519
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:284 Resp: 442187

Ion	Ratio	Lower	Upper
284	100		
142	31.2	26.9	40.3
249	31.3	25.3	37.9






#69
Atrazine
Concen: 56.767 ng
RT: 11.116 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:200 Resp: 30714


Ion	Ratio	Lower	Upper
200	100		
173	27.1	7.7	47.7
215	47.5	27.7	67.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#70
Pentachlorophenol
Concen: 61.587 ng
RT: 11.216 min Scan# 1551
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:266 Resp: 289014

Ion	Ratio	Lower	Upper
266	100		
268	63.7	50.4	75.6
264	63.0	50.6	75.8





#71
Phenanthrene
Concen: 54.618 ng
RT: 11.439 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17


Tgt Ion:178 Resp: 172176


Ion	Ratio	Lower	Upper
178	100		
176	20.1	15.8	23.6
179	15.9	12.6	18.8

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#72
Anthracene
Concen: 54.206 ng
RT: 11.492 min Scan# 1598
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:178 Resp: 1667221

Ion	Ratio	Lower	Upper
178	100		
176	19.6	15.3	22.9
179	15.9	12.4	18.6






#73
Carbazole
Concen: 54.015 ng
RT: 11.645 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:167 Resp: 154508


Ion	Ratio	Lower	Upper
167	100		
166	22.1	17.0	25.6
139	13.7	10.8	16.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#74
Di-n-butylphthalate
Concen: 55.069 ng
RT: 11.975 min Scan# 1680
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:149 Resp: 1819929

Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.5	11.3
104	5.3	4.1	6.1






#75
Fluoranthene
Concen: 52.448 ng
RT: 12.627 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:202 Resp: 167140


Ion	Ratio	Lower	Upper
202	100		
101	12.1	0.0	31.9
203	17.8	0.0	37.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.057 min Scan# 2034
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:240 Resp: 337032

Ion	Ratio	Lower	Upper
240	100		
120	11.0	9.4	14.2
236	26.7	20.9	31.3






#77
Benzidine
Concen: 37.499 ng
RT: 12.745 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:184 Resp: 20554


Ion	Ratio	Lower	Upper
184	100		
185	15.5	12.0	18.0
183	11.5	9.4	14.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#78
Pyrene
Concen: 56.522 ng
RT: 12.857 min Scan# 1830
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:202 Resp: 1667489

Ion	Ratio	Lower	Upper
202	100		
200	21.9	17.2	25.8
203	18.0	14.2	21.4






#79
Terphenyl-d14
Concen: 109.563 ng
RT: 13.004 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:244 Resp: 226612


Ion	Ratio	Lower	Upper
244	100		
212	7.0	5.7	8.5
122	9.7	8.6	13.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#80
Butylbenzylphthalate
Concen: 60.719 ng
RT: 13.474 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:149 Resp: 537039

Ion	Ratio	Lower	Upper
149	100		
91	76.1	59.4	89.2
206	20.7	16.4	24.6






#81
Benzo(a)anthracene
Concen: 56.992 ng
RT: 14.045 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:228 Resp: 125038


Ion	Ratio	Lower	Upper
228	100		
226	27.3	21.6	32.4
229	20.0	16.1	24.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#82
3,3'-Dichlorobenzidine
Concen: 60.353 ng
RT: 14.004 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:252 Resp: 386072

Ion	Ratio	Lower	Upper
252	100		
254	64.3	50.8	76.2
126	11.2	8.9	13.3






#83
Chrysene
Concen: 57.855 ng
RT: 14.080 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:228 Resp: 116380


Ion	Ratio	Lower	Upper
228	100		
226	30.3	24.1	36.1
229	19.9	15.8	23.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#84
Bis(2-ethylhexyl)phthalate
Concen: 63.749 ng
RT: 14.033 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:149 Resp: 632594

Ion	Ratio	Lower	Upper
149	100		
167	26.7	21.4	32.0
279	3.9	3.0	4.4






#85
Di-n-octyl phthalate
Concen: 67.634 ng
RT: 14.651 min Scan# 21
Delta R.T. -0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:149 Resp: 122073


Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.2	1.8
43	12.3	9.8	14.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#86
Perylene-d12
Concen: 20.000 ng
RT: 15.527 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:264 Resp: 377203

Ion	Ratio	Lower	Upper
264	100		
260	24.1	19.4	29.2
265	21.5	17.4	26.0





#87
Indeno(1,2,3-cd)pyrene
Concen: 61.817 ng
RT: 17.027 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17


Instrument :
BNA_F
ClientSampleId :
SSTDICC060


Tgt Ion:276 Resp: 150008

Ion	Ratio	Lower	Upper
276	100		
138	25.0	20.4	30.6
277	25.4	20.2	30.4

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#88
Benzo(b)fluoranthene
Concen: 61.033 ng
RT: 15.098 min Scan# 2211
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:252 Resp: 1402395

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.6	26.4
125	10.1	7.9	11.9





#89
Benzo(k)fluoranthene
Concen: 53.053 ng
RT: 15.127 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17


Tgt Ion:252 Resp: 105130


Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.6	26.4
125	10.2	7.9	11.9

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#90
Benzo(a)pyrene
Concen: 59.764 ng
RT: 15.468 min Scan# 2274
Delta R.T. 0.000 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:252 Resp: 1129942

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.3	25.9
125	11.0	8.7	13.1






#91
Dibenzo(a,h)anthracene
Concen: 60.667 ng
RT: 17.051 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:278 Resp: 122812


Ion	Ratio	Lower	Upper
278	100		
139	16.0	13.1	19.7
279	24.1	19.0	28.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#92
Benzo(g,h,i)perylene
Concen: 61.272 ng
RT: 17.486 min Scan# 2617
Delta R.T. 0.006 min
Lab File: BF139850.D
Acq: 18 Oct 2024 13:17

Tgt Ion:276 Resp: 1238955

Ion	Ratio	Lower	Upper
276	100		
277	23.6	19.0	28.6
138	21.4	17.4	26.0



Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139851.D
Acq On : 18 Oct 2024 13:46
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Quant Time: Oct 18 14:21:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED
Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.892	152	187399	20.000	ng		0.00
21) Naphthalene-d8	8.175	136	693285	20.000	ng		0.00
39) Acenaphthene-d10	9.933	164	377723	20.000	ng		0.00
64) Phenanthrene-d10	11.416	188	633721	20.000	ng		0.00
76) Chrysene-d12	14.057	240	323513	20.000	ng		0.00
86) Perylene-d12	15.533	264	374491	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.504	112	1657822	138.491	ng		0.00
7) Phenol-d6	6.522	99	2147173	138.492	ng		0.01
23) Nitrobenzene-d5	7.463	82	1897374	151.683	ng		0.01
42) 2,4,6-Tribromophenol	10.722	330	522350	147.845	ng		0.00
45) 2-Fluorobiphenyl	9.251	172	2941186	128.689	ng		0.00
79) Terphenyl-d14	13.004	244	2634676	132.704	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.693	88	410806	73.604	ng		99
3) Pyridine	3.452	79	993777	71.834	ng		99
4) n-Nitrosodimethylamine	3.422	42	567647	76.370	ng		98
6) Aniline	6.598	93	992375m	68.680	ng		
8) 2-Chlorophenol	6.681	128	836233	68.333	ng		98
9) Benzaldehyde	6.445	77	554983	63.632	ng		99
10) Phenol	6.534	94	1125558	69.172	ng		95
11) bis(2-Chloroethyl)ether	6.634	93	886786	71.780	ng		99
12) 1,3-Dichlorobenzene	6.834	146	969627	69.024	ng		98
13) 1,4-Dichlorobenzene	6.910	146	968088	68.978	ng		99
14) 1,2-Dichlorobenzene	7.069	146	861333	65.758	ng		99
15) Benzyl Alcohol	7.034	79	803101	70.616	ng		98
16) 2,2'-oxybis(1-Chloropr...	7.169	45	1472196	69.886	ng		99
17) 2-Methylphenol	7.140	107	752459	72.109	ng		99
18) Hexachloroethane	7.410	117	357637	71.459	ng		97
19) n-Nitroso-di-n-propyla...	7.316	70	651048	69.238	ng		99
20) 3+4-Methylphenols	7.298	107	878987	65.856	ng	#	85
22) Acetophenone	7.310	105	1149288	67.682	ng		98
24) Nitrobenzene	7.481	77	1025571	74.779	ng		99
25) Isophorone	7.722	82	1767234	74.436	ng		100
26) 2-Nitrophenol	7.792	139	434755	84.029	ng		99
27) 2,4-Dimethylphenol	7.822	122	619136	71.765	ng		99
28) bis(2-Chloroethoxy)met...	7.928	93	1022457	71.081	ng		100
29) 2,4-Dichlorophenol	8.028	162	711892	72.445	ng		99
30) 1,2,4-Trichlorobenzene	8.116	180	774108	71.194	ng		99
31) Naphthalene	8.198	128	2427578	67.894	ng		99
32) Benzoic acid	7.963	122	651887	86.634	ng		99
33) 4-Chloroaniline	8.251	127	849601	69.684	ng		99
34) Hexachlorobutadiene	8.316	225	490203	71.753	ng		99
35) Caprolactam	8.634	113	239184	76.550	ng		99
36) 4-Chloro-3-methylphenol	8.722	107	794640	73.031	ng		99
37) 2-Methylnaphthalene	8.886	142	1487841	67.944	ng		100
38) 1-Methylnaphthalene	8.986	142	1457890	67.858	ng		99
40) 1,2,4,5-Tetrachloroben...	9.051	216	712936	69.962	ng		100
41) Hexachlorocyclopentadiene	9.039	237	256007	72.201	ng		100
43) 2,4,6-Trichlorophenol	9.163	196	535533	74.228	ng		99

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139851.D
Acq On : 18 Oct 2024 13:46
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Quant Time: Oct 18 14:21:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.198	196	542663	72.904	ng		99
46) 1,1'-Biphenyl	9.357	154	1753952	66.703	ng		99
47) 2-Chloronaphthalene	9.381	162	1470012	69.411	ng		99
48) 2-Nitroaniline	9.469	65	527552	81.447	ng		99
49) Acenaphthylene	9.792	152	2106357	68.796	ng		99
50) Dimethylphthalate	9.657	163	1676969	71.277	ng		99
51) 2,6-Dinitrotoluene	9.716	165	393234	77.193	ng		99
52) Acenaphthene	9.969	154	1378724	69.610	ng		99
53) 3-Nitroaniline	9.886	138	387157	73.697	ng		96
54) 2,4-Dinitrophenol	9.986	184	169167	79.932	ng	#	1
55) Dibenzofuran	10.139	168	1930402	67.930	ng		99
56) 4-Nitrophenol	10.033	139	314170	77.732	ng		99
57) 2,4-Dinitrotoluene	10.116	165	510703	81.523	ng		97
58) Fluorene	10.480	166	1426794	65.920	ng		100
59) 2,3,4,6-Tetrachlorophenol	10.251	232	424598	72.764	ng		99
60) Diethylphthalate	10.357	149	1631842	70.641	ng		100
61) 4-Chlorophenyl-phenyle...	10.475	204	730565	66.837	ng		100
62) 4-Nitroaniline	10.504	138	383383	77.270	ng		98
63) Azobenzene	10.633	77	1727533	70.539	ng		99
65) 4,6-Dinitro-2-methylph...	10.528	198	235323	91.037	ng		98
66) n-Nitrosodiphenylamine	10.592	169	1350839	71.220	ng		99
67) 4-Bromophenyl-phenylether	10.963	248	477995	73.216	ng		99
68) Hexachlorobenzene	11.028	284	536511	73.070	ng		98
69) Atrazine	11.116	200	382294	74.406	ng		99
70) Pentachlorophenol	11.216	266	354105	79.464	ng		99
71) Phenanthrene	11.445	178	2045577	68.336	ng		100
72) Anthracene	11.492	178	1995750	68.333	ng		99
73) Carbazole	11.645	167	1818340	66.943	ng		99
74) Di-n-butylphthalate	11.980	149	2153517	68.623	ng		100
75) Fluoranthene	12.627	202	1955847	64.632	ng		99
77) Benzidine	12.745	184	318678	60.567	ng		99
78) Pyrene	12.857	202	1935829	68.360	ng		99
80) Butylbenzylphthalate	13.474	149	664532	78.274	ng		98
81) Benzo(a)anthracene	14.045	228	1512920	71.840	ng		99
82) 3,3'-Dichlorobenzidine	14.010	252	496976	80.937	ng		98
83) Chrysene	14.080	228	1424522	73.774	ng		99
84) Bis(2-ethylhexyl)phtha...	14.033	149	794387	83.399	ng		99
85) Di-n-octyl phthalate	14.651	149	1534331	88.561	ng		100
87) Indeno(1,2,3-cd)pyrene	17.033	276	1877616	77.935	ng		99
88) Benzo(b)fluoranthene	15.098	252	1664856	72.980	ng		99
89) Benzo(k)fluoranthene	15.133	252	1419158	72.134	ng		99
90) Benzo(a)pyrene	15.474	252	1419026	75.598	ng		100
91) Dibenzo(a,h)anthracene	17.057	278	1552091	77.226	ng		99
92) Benzo(g,h,i)perylene	17.492	276	1550622	77.240	ng		100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139851.D
Acq On : 18 Oct 2024 13:46
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1


Quant Time: Oct 18 14:21:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 14:11:20 2024
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.892 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:152 Resp: 18739


Ion	Ratio	Lower	Upper
152	100		
150	161.2	130.2	195.2
115	64.4	51.4	77.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#2
1,4-Dioxane
Concen: 73.604 ng
RT: 2.693 min Scan# 102
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion: 88 Resp: 410806

Ion	Ratio	Lower	Upper
88	100		
58	91.2	73.3	109.9
43	33.5	27.1	40.7






#3
Pyridine
Concen: 71.834 ng
RT: 3.452 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion: 79 Resp: 99377


Ion	Ratio	Lower	Upper
79	100		
52	86.1	68.5	102.7
51	41.5	32.3	48.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#4
n-Nitrosodimethylamine
Concen: 76.370 ng
RT: 3.422 min Scan# 226
Delta R.T. 0.024 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion: 42 Resp: 567647

Ion	Ratio	Lower	Upper
42	100		
74	117.8	96.2	144.4
44	7.1	6.2	9.2






#5
2-Fluorophenol
Concen: 138.491 ng
RT: 5.504 min Scan# 5
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:112 Resp: 165782


Ion	Ratio	Lower	Upper
112	100		
64	67.0	53.0	79.6
63	35.0	27.0	40.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#6
Aniline
Concen: 68.680 ng m
RT: 6.598 min Scan# 766
Delta R.T. 0.047 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion: 93 Resp: 992375

Ion	Ratio	Lower	Upper
93	100		
66	41.8	34.0	51.0
65	21.1	17.4	26.0






#7
Phenol-d6
Concen: 138.492 ng
RT: 6.522 min Scan# 7
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion: 99 Resp: 214717


Ion	Ratio	Lower	Upper
99	100		
42	20.8	16.7	25.1
71	36.1	27.7	41.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#8
2-Chlorophenol
Concen: 68.333 ng
RT: 6.681 min Scan# 780
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:128 Resp: 836233

Ion	Ratio	Lower	Upper
128	100		
130	32.8	12.3	52.3
64	56.2	34.6	74.6






#9
Benzaldehyde
Concen: 63.632 ng
RT: 6.445 min Scan# 74
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion: 77 Resp: 55498


Ion	Ratio	Lower	Upper
77	100		
105	96.4	75.5	115.5
106	92.6	71.7	111.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#10
Phenol
Concen: 69.172 ng
RT: 6.534 min Scan# 755
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion: 94 Resp: 1125558

Ion	Ratio	Lower	Upper
94	100		
65	33.1	11.1	51.1
66	47.1	22.7	62.7






#11
bis(2-Chloroethyl)ether
Concen: 71.780 ng
RT: 6.634 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion: 93 Resp: 88678


Ion	Ratio	Lower	Upper
93	100		
63	85.6	65.0	105.0
95	32.2	12.4	52.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#12
1,3-Dichlorobenzene
Concen: 69.024 ng
RT: 6.834 min Scan# 806
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:146 Resp: 969627

Ion	Ratio	Lower	Upper
146	100		
148	64.1	51.0	76.4
75	35.0	26.4	39.6






#13
1,4-Dichlorobenzene
Concen: 68.978 ng
RT: 6.910 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:146 Resp: 96808


Ion	Ratio	Lower	Upper
146	100		
148	64.0	51.8	77.8
111	44.3	34.6	51.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#14
1,2-Dichlorobenzene
Concen: 65.758 ng
RT: 7.069 min Scan# 846
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:146 Resp: 861333

Ion	Ratio	Lower	Upper
146	100		
148	63.5	51.4	77.2
111	46.5	36.7	55.1





#15
Benzyl Alcohol
Concen: 70.616 ng
RT: 7.034 min Scan# 84
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46


Tgt Ion: 79 Resp: 80310


Ion	Ratio	Lower	Upper
79	100		
108	74.0	61.4	92.0
77	62.3	49.9	74.9

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#16
2,2'-oxybis(1-Chloropropane)
Concen: 69.886 ng
RT: 7.169 min Scan# 863
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion: 45 Resp: 1472196

Ion	Ratio	Lower	Upper
45	100		
77	12.2	0.0	31.5
79	9.2	0.0	29.1






#17
2-Methylphenol
Concen: 72.109 ng
RT: 7.140 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:107 Resp: 75245


Ion	Ratio	Lower	Upper
107	100		
108	114.3	91.3	136.9
77	47.6	37.1	55.7
79	47.6	37.0	55.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#18
Hexachloroethane
Concen: 71.459 ng
RT: 7.410 min Scan# 904
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:117 Resp: 357637

Ion	Ratio	Lower	Upper
117	100		
119	95.6	77.0	115.4
201	99.9	76.2	114.4






#19
n-Nitroso-di-n-propylamine
Concen: 69.238 ng
RT: 7.316 min Scan# 8
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion: 70 Resp: 6510


Ion	Ratio	Lower	Upper
70	100		
42	63.4	51.5	77.3
101	9.4	7.5	11.3
130	19.5	15.5	23.3



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#20
3+4-Methylphenols
Concen: 65.856 ng
RT: 7.298 min Scan# 885
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:107 Resp: 878987

Ion	Ratio	Lower	Upper
107	100		
108	91.7	69.8	109.8
77	100.5	50.5	90.5#
79	31.7	10.2	50.2






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.175 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:136 Resp: 69328


Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.6	12.8
54	10.6	8.4	12.6
68	6.5	5.1	7.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#22
Acetophenone
Concen: 67.682 ng
RT: 7.310 min Scan# 887
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:105 Resp: 1149288

Ion	Ratio	Lower	Upper
105	100		
71	7.6	5.9	8.9
51	35.3	29.4	44.2
120	21.8	17.5	26.3





#23
Nitrobenzene-d5
Concen: 151.683 ng
RT: 7.463 min Scan# 9
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46


Tgt Ion: 82 Resp: 189737


Ion	Ratio	Lower	Upper
82	100		
128	42.6	33.4	50.0
54	58.8	47.8	71.8

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#24
Nitrobenzene
Concen: 74.779 ng
RT: 7.481 min Scan# 916
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion: 77 Resp: 1025571

Ion	Ratio	Lower	Upper
77	100		
123	44.0	35.7	53.5
65	14.6	11.5	17.3





#25
Isophorone
Concen: 74.436 ng
RT: 7.722 min Scan# 9
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46


Tgt Ion: 82 Resp: 176723


Ion	Ratio	Lower	Upper
82	100		
95	7.2	5.5	8.3
138	16.5	13.2	19.8

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#26
2-Nitrophenol
Concen: 84.029 ng
RT: 7.792 min Scan# 969
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:139 Resp: 434755

Ion	Ratio	Lower	Upper
139	100		
109	31.6	25.1	37.7
65	57.8	47.5	71.3






#27
2,4-Dimethylphenol
Concen: 71.765 ng
RT: 7.822 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:122 Resp: 61913


Ion	Ratio	Lower	Upper
122	100		
107	116.2	91.9	137.9
121	58.6	46.2	69.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#28
bis(2-Chloroethoxy)methane
Concen: 71.081 ng
RT: 7.928 min Scan# 992
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion: 93 Resp: 1022457

Ion	Ratio	Lower	Upper
93	100		
95	32.7	26.0	39.0
123	12.2	9.8	14.8






#29
2,4-Dichlorophenol
Concen: 72.445 ng
RT: 8.028 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:162 Resp: 71189


Ion	Ratio	Lower	Upper
162	100		
164	64.5	43.9	83.9
98	40.8	19.9	59.9



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#30
1,2,4-Trichlorobenzene
Concen: 71.194 ng
RT: 8.116 min Scan# 1024
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:180 Resp: 774108

Ion	Ratio	Lower	Upper
180	100		
182	99.8	80.5	120.7
145	33.0	25.4	38.0






#31
Naphthalene
Concen: 67.894 ng
RT: 8.198 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:128 Resp: 242757


Ion	Ratio	Lower	Upper
128	100		
129	11.7	8.9	13.3
127	13.6	10.6	16.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#32
Benzoic acid
Concen: 86.634 ng
RT: 7.963 min Scan# 998
Delta R.T. 0.041 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:122 Resp: 651887

Ion	Ratio	Lower	Upper
122	100		
105	126.8	107.6	147.6
77	98.1	80.2	120.2






#33
4-Chloroaniline
Concen: 69.684 ng
RT: 8.251 min Scan# 1
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:127 Resp: 84960


Ion	Ratio	Lower	Upper
127	100		
129	32.5	25.4	38.2
65	35.3	28.4	42.6
92	20.7	16.6	25.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#34
Hexachlorobutadiene
Concen: 71.753 ng
RT: 8.316 min Scan# 1058
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:225 Resp: 490203

Ion	Ratio	Lower	Upper
225	100		
223	63.2	50.6	76.0
227	64.0	51.8	77.8





#35
Caprolactam
Concen: 76.550 ng
RT: 8.634 min Scan# 11
Delta R.T. 0.024 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46


Tgt Ion:113 Resp: 23918


Ion	Ratio	Lower	Upper
113	100		
55	206.7	189.2	229.2
56	152.7	132.5	172.5

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#36
4-Chloro-3-methylphenol
Concen: 73.031 ng
RT: 8.722 min Scan# 1127
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:107 Resp: 794640

Ion	Ratio	Lower	Upper
107	100		
144	24.4	19.1	28.7
142	72.3	57.2	85.8





#37
2-Methylnaphthalene
Concen: 67.944 ng
RT: 8.886 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46


Tgt Ion:142 Resp: 148784


Ion	Ratio	Lower	Upper
142	100		
141	88.7	70.8	106.2

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#38
1-Methylnaphthalene
Concen: 67.858 ng
RT: 8.986 min Scan# 1172
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:142 Resp: 1457890

Ion	Ratio	Lower	Upper
142	100		
141	91.4	73.5	110.3






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.933 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:164 Resp: 37772


Ion	Ratio	Lower	Upper
164	100		
162	100.7	81.0	121.4
160	44.3	35.4	53.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#40
1,2,4,5-Tetrachlorobenzene
Concen: 69.962 ng
RT: 9.051 min Scan# 1183
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:216 Resp: 712936

Ion	Ratio	Lower	Upper
216	100		
214	79.7	63.4	95.0
179	21.7	17.2	25.8
108	19.5	15.6	23.4






#41
Hexachlorocyclopentadiene
Concen: 72.201 ng
RT: 9.039 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:237 Resp: 25600


Ion	Ratio	Lower	Upper
237	100		
235	62.9	42.9	82.9
272	12.5	0.0	32.3



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#42
2,4,6-Tribromophenol
Concen: 147.845 ng
RT: 10.722 min Scan# 1467
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:330 Resp: 522350

Ion	Ratio	Lower	Upper
330	100		
332	96.0	78.1	117.1
141	31.2	26.6	39.8






#43
2,4,6-Trichlorophenol
Concen: 74.228 ng
RT: 9.163 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:196 Resp: 53553


Ion	Ratio	Lower	Upper
196	100		
198	96.0	76.2	114.2
200	30.7	24.0	36.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#44
2,4,5-Trichlorophenol
Concen: 72.904 ng
RT: 9.198 min Scan# 1208
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:196 Resp: 542663

Ion	Ratio	Lower	Upper
196	100		
198	95.2	75.8	113.6
97	54.3	41.8	62.6
132	29.0	22.9	34.3






#45
2-Fluorobiphenyl
Concen: 128.689 ng
RT: 9.251 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:172 Resp: 294118


Ion	Ratio	Lower	Upper
172	100		
171	36.6	28.6	43.0
170	24.7	19.1	28.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#46
1,1'-Biphenyl
Concen: 66.703 ng
RT: 9.357 min Scan# 1235
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:154 Resp: 1753952

Ion	Ratio	Lower	Upper
154	100		
153	41.3	20.3	60.3
76	13.8	0.0	33.9






#47
2-Chloronaphthalene
Concen: 69.411 ng
RT: 9.381 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:162 Resp: 147001


Ion	Ratio	Lower	Upper
162	100		
127	43.0	33.8	50.6
164	33.3	26.6	39.8



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#48
2-Nitroaniline
Concen: 81.447 ng
RT: 9.469 min Scan# 1254
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion: 65 Resp: 527552

Ion	Ratio	Lower	Upper
65	100		
92	64.1	50.9	76.3
138	93.3	73.4	110.0






#49
Acenaphthylene
Concen: 68.796 ng
RT: 9.792 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:152 Resp: 210635


Ion	Ratio	Lower	Upper
152	100		
151	20.5	15.9	23.9
153	13.7	10.7	16.1



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#50
Dimethylphthalate
Concen: 71.277 ng
RT: 9.657 min Scan# 1286
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:163 Resp: 1676969

Ion	Ratio	Lower	Upper
163	100		
194	4.2	3.4	5.0
164	10.5	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 77.193 ng
RT: 9.716 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46


Tgt Ion:165 Resp: 39323


Ion	Ratio	Lower	Upper
165	100		
63	64.7	50.6	75.8
89	59.4	47.5	71.3

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#52
Acenaphthene
Concen: 69.610 ng
RT: 9.969 min Scan# 1339
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:154 Resp: 1378724

Ion	Ratio	Lower	Upper
154	100		
153	110.4	89.8	134.6
152	53.4	42.4	63.6





#53
3-Nitroaniline
Concen: 73.697 ng
RT: 9.886 min Scan# 1
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46


Tgt Ion:138 Resp: 38715


Ion	Ratio	Lower	Upper
138	100		
108	12.1	10.2	15.2
92	118.4	98.3	147.5

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#54
2,4-Dinitrophenol
Concen: 79.932 ng
RT: 9.986 min Scan# 1342
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:184 Resp: 169167

Ion	Ratio	Lower	Upper
184	100		
63	77.7	84.7	127.1#
154	102.0	338.2	507.2#






#55
Dibenzofuran
Concen: 67.930 ng
RT: 10.139 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:168 Resp: 193040


Ion	Ratio	Lower	Upper
168	100		
139	40.6	31.8	47.6
169	14.6	11.6	17.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#56
4-Nitrophenol
Concen: 77.732 ng
RT: 10.033 min Scan# 1350
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:139 Resp: 314170

Ion	Ratio	Lower	Upper
139	100		
109	73.7	53.6	93.6
65	109.6	87.8	127.8






#57
2,4-Dinitrotoluene
Concen: 81.523 ng
RT: 10.116 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:165 Resp: 51070


Ion	Ratio	Lower	Upper
165	100		
63	51.9	40.1	60.1
89	78.4	60.4	90.6
182	2.6	2.0	3.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#58
Fluorene
Concen: 65.920 ng
RT: 10.480 min Scan# 1426
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:166 Resp: 1426794

Ion	Ratio	Lower	Upper
166	100		
165	98.6	78.5	117.7
167	13.5	10.9	16.3






#59
2,3,4,6-Tetrachlorophenol
Concen: 72.764 ng
RT: 10.251 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:232 Resp: 42459


Ion	Ratio	Lower	Upper
232	100		
131	45.6	37.0	55.6
130	2.4	1.9	2.9
166	29.6	24.2	36.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#60
Diethylphthalate
Concen: 70.641 ng
RT: 10.357 min Scan# 1405
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:149 Resp: 1631842

Ion	Ratio	Lower	Upper
149	100		
177	22.2	17.9	26.9
150	12.8	10.1	15.1





#61
4-Chlorophenyl-phenylether
Concen: 66.837 ng
RT: 10.475 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46


Tgt Ion:204 Resp: 73056


Ion	Ratio	Lower	Upper
204	100		
206	32.7	26.3	39.5
141	60.6	48.6	73.0

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#62
4-Nitroaniline
Concen: 77.270 ng
RT: 10.504 min Scan# 1430
Delta R.T. 0.018 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:138 Resp: 383383

Ion	Ratio	Lower	Upper
138	100		
92	53.7	35.7	75.7
108	78.3	57.5	97.5






#63
Azobenzene
Concen: 70.539 ng
RT: 10.633 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion: 77 Resp: 172753


Ion	Ratio	Lower	Upper
77	100		
182	24.6	3.3	43.3
105	21.6	1.2	41.2
51	33.6	13.3	53.3



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.416 min Scan# 1585
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:188 Resp: 633721

Ion	Ratio	Lower	Upper
188	100		
94	10.7	7.9	11.9
80	11.1	9.0	13.4





#65
4,6-Dinitro-2-methylphenol
Concen: 91.037 ng
RT: 10.528 min Scan# 14
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46


Tgt Ion:198 Resp: 23532


Ion	Ratio	Lower	Upper
198	100		
51	56.5	38.2	78.2
105	46.4	27.9	67.9

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#66
n-Nitrosodiphenylamine
Concen: 71.220 ng
RT: 10.592 min Scan# 1445
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:169 Resp: 1350839

Ion	Ratio	Lower	Upper
169	100		
168	63.4	49.8	74.6
167	35.5	28.0	42.0






#67
4-Bromophenyl-phenylether
Concen: 73.216 ng
RT: 10.963 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:248 Resp: 47799


Ion	Ratio	Lower	Upper
248	100		
250	97.5	79.0	118.6
141	75.7	61.7	92.5



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#68
Hexachlorobenzene
Concen: 73.070 ng
RT: 11.028 min Scan# 1519
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:284 Resp: 536511

Ion	Ratio	Lower	Upper
284	100		
142	31.9	26.9	40.3
249	31.3	25.3	37.9






#69
Atrazine
Concen: 74.406 ng
RT: 11.116 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:200 Resp: 38229


Ion	Ratio	Lower	Upper
200	100		
173	26.9	7.7	47.7
215	47.9	27.7	67.7



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#70
Pentachlorophenol
Concen: 79.464 ng
RT: 11.216 min Scan# 1551
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:266 Resp: 354105

Ion	Ratio	Lower	Upper
266	100		
268	63.1	50.4	75.6
264	62.5	50.6	75.8





#71
Phenanthrene
Concen: 68.336 ng
RT: 11.445 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46


Tgt Ion:178 Resp: 204557


Ion	Ratio	Lower	Upper
178	100		
176	20.0	15.8	23.6
179	15.7	12.6	18.8

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#72
Anthracene
Concen: 68.333 ng
RT: 11.492 min Scan# 1598
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:178 Resp: 1995750

Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.3	22.9
179	16.0	12.4	18.6






#73
Carbazole
Concen: 66.943 ng
RT: 11.645 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:167 Resp: 181834


Ion	Ratio	Lower	Upper
167	100		
166	21.9	17.0	25.6
139	13.6	10.8	16.2



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#74
Di-n-butylphthalate
Concen: 68.623 ng
RT: 11.980 min Scan# 1681
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:149 Resp: 2153517

Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.5	11.3
104	5.2	4.1	6.1





#75
Fluoranthene
Concen: 64.632 ng
RT: 12.627 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46


Tgt Ion:202 Resp: 195584


Ion	Ratio	Lower	Upper
202	100		
101	12.4	0.0	31.9
203	17.7	0.0	37.5

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.057 min Scan# 2034
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:240 Resp: 323513

Ion	Ratio	Lower	Upper
240	100		
120	10.6	9.4	14.2
236	26.1	20.9	31.3





#77
Benzidine
Concen: 60.567 ng
RT: 12.745 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46


Tgt Ion:184 Resp: 31867


Ion	Ratio	Lower	Upper
184	100		
185	15.1	12.0	18.0
183	11.4	9.4	14.2

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#78
Pyrene
Concen: 68.360 ng
RT: 12.857 min Scan# 1830
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:202 Resp: 1935829

Ion	Ratio	Lower	Upper
202	100		
200	21.7	17.2	25.8
203	18.2	14.2	21.4






#79
Terphenyl-d14
Concen: 132.704 ng
RT: 13.004 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:244 Resp: 263467


Ion	Ratio	Lower	Upper
244	100		
212	7.1	5.7	8.5
122	9.8	8.6	13.0



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#80
Butylbenzylphthalate
Concen: 78.274 ng
RT: 13.474 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:149 Resp: 664532

Ion	Ratio	Lower	Upper
149	100		
91	76.5	59.4	89.2
206	20.5	16.4	24.6





#81
Benzo(a)anthracene
Concen: 71.840 ng
RT: 14.045 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46


Tgt Ion:228 Resp: 151292


Ion	Ratio	Lower	Upper
228	100		
226	27.4	21.6	32.4
229	20.0	16.1	24.1

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#82
3,3'-Dichlorobenzidine
Concen: 80.937 ng
RT: 14.010 min Scan# 2026
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:252 Resp: 496976

Ion	Ratio	Lower	Upper
252	100		
254	65.7	50.8	76.2
126	11.0	8.9	13.3






#83
Chrysene
Concen: 73.774 ng
RT: 14.080 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:228 Resp: 142452


Ion	Ratio	Lower	Upper
228	100		
226	30.4	24.1	36.1
229	20.3	15.8	23.6



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#84
Bis(2-ethylhexyl)phthalate
Concen: 83.399 ng
RT: 14.033 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:149 Resp: 794387

Ion	Ratio	Lower	Upper
149	100		
167	26.3	21.4	32.0
279	4.0	3.0	4.4





#85
Di-n-octyl phthalate
Concen: 88.561 ng
RT: 14.651 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46


Tgt Ion:149 Resp: 153433


Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.2	1.8
43	12.4	9.8	14.6

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.533 min Scan# 2285
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:264 Resp: 374491

Ion	Ratio	Lower	Upper
264	100		
260	24.1	19.4	29.2
265	21.4	17.4	26.0






#87
Indeno(1,2,3-cd)pyrene
Concen: 77.935 ng
RT: 17.033 min Scan# 2
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:276 Resp: 187761


Ion	Ratio	Lower	Upper
276	100		
138	25.1	20.4	30.6
277	25.4	20.2	30.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024




#88
Benzo(b)fluoranthene
Concen: 72.980 ng
RT: 15.098 min Scan# 2211
Delta R.T. 0.000 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:252 Resp: 1664856

Ion	Ratio	Lower	Upper
252	100		
253	22.4	17.6	26.4
125	10.4	7.9	11.9





#89
Benzo(k)fluoranthene
Concen: 72.134 ng
RT: 15.133 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46


Tgt Ion:252 Resp: 141915


Ion	Ratio	Lower	Upper
252	100		
253	22.7	17.6	26.4
125	9.9	7.9	11.9

Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024







#90
Benzo(a)pyrene
Concen: 75.598 ng
RT: 15.474 min Scan# 2275
Delta R.T. 0.006 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:252 Resp: 1419026

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.3	25.9
125	10.7	8.7	13.1






#91
Dibenzo(a,h)anthracene
Concen: 77.226 ng
RT: 17.057 min Scan# 2
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:278 Resp: 155209


Ion	Ratio	Lower	Upper
278	100		
139	15.6	13.1	19.7
279	23.7	19.0	28.4



Instrument :
BNA_F
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/21/2024
Supervised By :mohammad ahmed 10/21/2024



#92
Benzo(g,h,i)perylene
Concen: 77.240 ng
RT: 17.492 min Scan# 2618
Delta R.T. 0.012 min
Lab File: BF139851.D
Acq: 18 Oct 2024 13:46

Tgt Ion:276 Resp: 1550622

Ion	Ratio	Lower	Upper
276	100		
277	23.9	19.0	28.6
138	21.6	17.4	26.0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG No.: P4578

Instrument ID: BNA_F Calibration Date/Time: 10/28/2024 09:32

Lab File ID: BF140075.D Init. Calib. Date(s): 10/18/2024 10/18/2024

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:27 13:46

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.476	1.412		-4.3	
2-Fluorophenol	1.278	1.198		-6.3	
Phenol-d6	1.655	1.537		-7.1	
1,4-Dichlorobenzene	1.498	1.486		-0.8	20.0
2-Methylphenol	1.114	1.050		-5.7	
3+4-Methylphenols	1.424	1.320		-7.3	
Nitrobenzene-d5	0.361	0.377		4.4	
Hexachloroethane	0.534	0.526		-1.5	
Nitrobenzene	0.396	0.397		0.3	
Hexachlorobutadiene	0.197	0.200		1.5	20.0
2,4,6-Trichlorophenol	0.382	0.375		-1.8	20.0
2-Fluorobiphenyl	1.210	1.195		-1.2	
2,4,5-Trichlorophenol	0.394	0.392		-0.5	
2,4-Dinitrotoluene	0.332	0.382		15.1	
2,4,6-Tribromophenol	0.187	0.197		5.3	
Hexachlorobenzene	0.232	0.233		0.4	
Pentachlorophenol	0.141	0.144		2.1	20.0
Terphenyl-d14	1.227	1.250		1.9	

All other compounds must meet a minimum RRF of 0.010.

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140075.D
Acq On : 28 Oct 2024 09:32
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Oct 28 09:55:45 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.892	152	132517	20.000	ng		0.00
21) Naphthalene-d8	8.175	136	497360	20.000	ng		0.00
39) Acenaphthene-d10	9.927	164	274065	20.000	ng		0.00
64) Phenanthrene-d10	11.416	188	494703	20.000	ng		0.00
76) Chrysene-d12	14.051	240	250901	20.000	ng		0.00
86) Perylene-d12	15.527	264	263501	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.504	112	635250	75.045	ng		0.00
7) Phenol-d6	6.516	99	814577	74.300	ng		0.00
23) Nitrobenzene-d5	7.451	82	750063	83.584	ng		0.00
42) 2,4,6-Tribromophenol	10.716	330	216066	84.285	ng		0.00
45) 2-Fluorobiphenyl	9.245	172	1310048	79.000	ng		0.00
79) Terphenyl-d14	12.998	244	1254654	81.484	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.669	88	148154	37.538	ng		96
3) Pyridine	3.440	79	374133	38.244	ng		98
4) n-Nitrosodimethylamine	3.387	42	197677	37.610	ng		96
6) Aniline	6.551	93	388187	37.992	ng		99
8) 2-Chlorophenol	6.675	128	328283	37.936	ng		100
9) Benzaldehyde	6.439	77	234409	38.007	ng		99
10) Phenol	6.528	94	427262m	37.802	ng		
11) bis(2-Chloroethyl)ether	6.628	93	324860	37.186	ng		100
12) 1,3-Dichlorobenzene	6.833	146	395383	39.802	ng		99
13) 1,4-Dichlorobenzene	6.910	146	393761	39.676	ng		99
14) 1,2-Dichlorobenzene	7.063	146	364568	39.360	ng		99
15) Benzyl Alcohol	7.028	79	307563	38.244	ng		98
16) 2,2'-oxybis(1-Chloropr...	7.163	45	538199	36.130	ng		98
17) 2-Methylphenol	7.133	107	278276	37.712	ng		99
18) Hexachloroethane	7.404	117	139330	39.369	ng		98
19) n-Nitroso-di-n-propyla...	7.304	70	243835	36.671	ng		98
20) 3+4-Methylphenols	7.292	107	349881	37.070	ng	#	83
22) Acetophenone	7.298	105	466951	38.331	ng		99
24) Nitrobenzene	7.475	77	394888	40.136	ng		100
25) Isophorone	7.710	82	654467	38.425	ng		99
26) 2-Nitrophenol	7.786	139	151084	40.704	ng		98
27) 2,4-Dimethylphenol	7.816	122	220819	35.678	ng		97
28) bis(2-Chloroethoxy)met...	7.922	93	392131	38.000	ng		100
29) 2,4-Dichlorophenol	8.028	162	277268	39.331	ng		99
30) 1,2,4-Trichlorobenzene	8.116	180	313301	40.164	ng		100
31) Naphthalene	8.198	128	1010972	39.413	ng		100
32) Benzoic acid	7.928	122	184222	34.127	ng		99
33) 4-Chloroaniline	8.239	127	341100	38.998	ng		99
34) Hexachlorobutadiene	8.310	225	198499	40.501	ng		100
35) Caprolactam	8.616	113	88634	39.542	ng		92
36) 4-Chloro-3-methylphenol	8.716	107	305431	39.128	ng		98
37) 2-Methylnaphthalene	8.886	142	616924	39.271	ng		99
38) 1-Methylnaphthalene	8.986	142	607270	39.400	ng		100
40) 1,2,4,5-Tetrachloroben...	9.051	216	303883	41.099	ng		99
41) Hexachlorocyclopentadiene	9.039	237	99773	38.781	ng		100
43) 2,4,6-Trichlorophenol	9.157	196	205507	39.258	ng		99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140075.D
Acq On : 28 Oct 2024 09:32
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

Quant Time: Oct 28 09:55:45 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.198	196	214705	39.754	ng		98
46) 1,1'-Biphenyl	9.351	154	756187	39.635	ng		100
47) 2-Chloronaphthalene	9.375	162	620778	40.398	ng		99
48) 2-Nitroaniline	9.469	65	201129	42.796	ng		94
49) Acenaphthylene	9.792	152	891367	40.124	ng		99
50) Dimethylphthalate	9.651	163	682031	39.953	ng		99
51) 2,6-Dinitrotoluene	9.710	165	161527	43.701	ng		98
52) Acenaphthene	9.963	154	579769	40.343	ng		99
53) 3-Nitroaniline	9.874	138	160653	42.148	ng		98
54) 2,4-Dinitrophenol	9.980	184	64847	45.309	ng	#	1
55) Dibenzofuran	10.133	168	813269	39.443	ng		99
56) 4-Nitrophenol	10.027	139	120058	40.940	ng		97
57) 2,4-Dinitrotoluene	10.110	165	209655	46.125	ng		99
58) Fluorene	10.474	166	633488	40.338	ng		100
59) 2,3,4,6-Tetrachlorophenol	10.251	232	170444	40.257	ng		98
60) Diethylphthalate	10.345	149	675530	40.303	ng		100
61) 4-Chlorophenyl-phenyle...	10.469	204	324430	40.907	ng		99
62) 4-Nitroaniline	10.492	138	156727	43.535	ng		98
63) Azobenzene	10.627	77	685497	38.577	ng		98
65) 4,6-Dinitro-2-methylph...	10.522	198	96169	47.659	ng		92
66) n-Nitrosodiphenylamine	10.586	169	569092	38.436	ng		100
67) 4-Bromophenyl-phenylether	10.957	248	205221	40.268	ng		98
68) Hexachlorobenzene	11.021	284	230272	40.175	ng		99
69) Atrazine	11.110	200	155137	38.680	ng		99
70) Pentachlorophenol	11.216	266	142153	40.865	ng		98
71) Phenanthrene	11.439	178	899645	38.500	ng		99
72) Anthracene	11.492	178	889039	38.994	ng		100
73) Carbazole	11.645	167	803004	37.870	ng		99
74) Di-n-butylphthalate	11.974	149	957414	39.082	ng		99
75) Fluoranthene	12.627	202	901406	38.158	ng		99
77) Benzidine	12.745	184	186877	45.296	ng		99
78) Pyrene	12.857	202	894517	40.730	ng		99
80) Butylbenzylphthalate	13.468	149	271701	41.265	ng		99
81) Benzo(a)anthracene	14.039	228	665269	40.732	ng		99
82) 3,3'-Dichlorobenzidine	14.004	252	211560	44.426	ng		99
83) Chrysene	14.080	228	587960	39.262	ng		100
84) Bis(2-ethylhexyl)phtha...	14.027	149	334928	45.339	ng		99
85) Di-n-octyl phthalate	14.639	149	547586	40.754	ng		98
87) Indeno(1,2,3-cd)pyrene	17.033	276	684787	40.396	ng		98
88) Benzo(b)fluoranthene	15.098	252	615487	38.345	ng		99
89) Benzo(k)fluoranthene	15.127	252	563702	40.721	ng		99
90) Benzo(a)pyrene	15.468	252	533309	40.379	ng		99
91) Dibenzo(a,h)anthracene	17.051	278	563984	39.881	ng		98
92) Benzo(g,h,i)perylene	17.486	276	583161	41.284	ng		98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140075.D
Acq On : 28 Oct 2024 09:32
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1


Quant Time: Oct 28 09:55:45 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.892 min Scan# 81
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:152 Resp: 13251


Ion	Ratio	Lower	Upper
152	100		
150	159.1	130.2	195.2
115	63.1	51.4	77.2



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#2
1,4-Dioxane
Concen: 37.538 ng
RT: 2.669 min Scan# 98
Delta R.T. -0.018 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion: 88 Resp: 148154

Ion	Ratio	Lower	Upper
88	100		
58	87.9	73.3	109.9
43	32.5	27.1	40.7






#3
Pyridine
Concen: 38.244 ng
RT: 3.440 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion: 79 Resp: 37413


Ion	Ratio	Lower	Upper
79	100		
52	83.2	68.5	102.7
51	40.7	32.3	48.5



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#4
n-Nitrosodimethylamine
Concen: 37.610 ng
RT: 3.387 min Scan# 220
Delta R.T. -0.012 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion: 42 Resp: 197677

Ion	Ratio	Lower	Upper
42	100		
74	115.6	96.2	144.4
44	8.1	6.2	9.2






#5
2-Fluorophenol
Concen: 75.045 ng
RT: 5.504 min Scan# 58
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:112 Resp: 635250


Ion	Ratio	Lower	Upper
112	100		
64	64.3	53.0	79.6
63	33.6	27.0	40.4



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#6
Aniline
Concen: 37.992 ng
RT: 6.551 min Scan# 758
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion: 93 Resp: 388187

Ion	Ratio	Lower	Upper
93	100		
66	41.5	34.0	51.0
65	21.6	17.4	26.0





#7
Phenol-d6
Concen: 74.300 ng
RT: 6.516 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32


Tgt Ion: 99 Resp: 81457


Ion	Ratio	Lower	Upper
99	100		
42	20.2	16.7	25.1
71	35.1	27.7	41.5

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#8
2-Chlorophenol
Concen: 37.936 ng
RT: 6.675 min Scan# 779
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:128 Resp: 328283

Ion	Ratio	Lower	Upper
128	100		
130	32.5	12.3	52.3
64	54.8	34.6	74.6






#9
Benzaldehyde
Concen: 38.007 ng
RT: 6.439 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion: 77 Resp: 23440


Ion	Ratio	Lower	Upper
77	100		
105	94.6	75.5	115.5
106	91.6	71.7	111.7



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#10
Phenol
Concen: 37.802 ng m
RT: 6.528 min Scan# 754
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion: 94 Resp: 427262

Ion	Ratio	Lower	Upper
94	100		
65	31.7	11.1	51.1
66	43.6	22.7	62.7





#11
bis(2-Chloroethyl)ether
Concen: 37.186 ng
RT: 6.628 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32


Tgt Ion: 93 Resp: 32486


Ion	Ratio	Lower	Upper
93	100		
63	85.0	65.0	105.0
95	32.6	12.4	52.4

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#12
1,3-Dichlorobenzene
Concen: 39.802 ng
RT: 6.833 min Scan# 806
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:146 Resp: 395383

Ion	Ratio	Lower	Upper
146	100		
148	64.0	51.0	76.4
75	32.4	26.4	39.6






#13
1,4-Dichlorobenzene
Concen: 39.676 ng
RT: 6.910 min Scan# 81
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:146 Resp: 39376


Ion	Ratio	Lower	Upper
146	100		
148	64.2	51.8	77.8
111	42.7	34.6	51.8



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#14
1,2-Dichlorobenzene
Concen: 39.360 ng
RT: 7.063 min Scan# 845
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:146 Resp: 364568

Ion	Ratio	Lower	Upper
146	100		
148	63.2	51.4	77.2
111	44.9	36.7	55.1





#15
Benzyl Alcohol
Concen: 38.244 ng
RT: 7.028 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32


Tgt Ion: 79 Resp: 30756


Ion	Ratio	Lower	Upper
79	100		
108	73.7	61.4	92.0
77	62.8	49.9	74.9

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#16
2,2'-oxybis(1-Chloropropane)
Concen: 36.130 ng
RT: 7.163 min Scan# 862
Delta R.T. -0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion: 45 Resp: 538199

Ion	Ratio	Lower	Upper
45	100		
77	12.5	0.0	31.5
79	9.4	0.0	29.1






#17
2-Methylphenol
Concen: 37.712 ng
RT: 7.133 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:107 Resp: 27827


Ion	Ratio	Lower	Upper
107	100		
108	113.0	91.3	136.9
77	48.0	37.1	55.7
79	47.2	37.0	55.6



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#18
Hexachloroethane
Concen: 39.369 ng
RT: 7.404 min Scan# 903
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:117 Resp: 139330

Ion	Ratio	Lower	Upper
117	100		
119	96.9	77.0	115.4
201	99.5	76.2	114.4






#19
n-Nitroso-di-n-propylamine
Concen: 36.671 ng
RT: 7.304 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion: 70 Resp: 24383


Ion	Ratio	Lower	Upper
70	100		
42	62.6	51.5	77.3
101	9.8	7.5	11.3
130	20.0	15.5	23.3



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#20
3+4-Methylphenols
Concen: 37.070 ng
RT: 7.292 min Scan# 884
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:107 Resp: 349881

Ion	Ratio	Lower	Upper
107	100		
108	90.4	69.8	109.8
77	107.5	50.5	90.5#
79	31.3	10.2	50.2






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.175 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:136 Resp: 49736


Ion	Ratio	Lower	Upper
136	100		
137	10.7	8.6	12.8
54	9.7	8.4	12.6
68	6.1	5.1	7.7



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#22
Acetophenone
Concen: 38.331 ng
RT: 7.298 min Scan# 885
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:105 Resp: 466951

Ion	Ratio	Lower	Upper
105	100		
71	7.7	5.9	8.9
51	37.2	29.4	44.2
120	21.4	17.5	26.3





#23
Nitrobenzene-d5
Concen: 83.584 ng
RT: 7.451 min Scan# 91
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32


Tgt Ion: 82 Resp: 75006


Ion	Ratio	Lower	Upper
82	100		
128	41.7	33.4	50.0
54	57.8	47.8	71.8

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#24
Nitrobenzene
Concen: 40.136 ng
RT: 7.475 min Scan# 915
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion: 77 Resp: 394888

Ion	Ratio	Lower	Upper
77	100		
123	44.3	35.7	53.5
65	14.3	11.5	17.3





#25
Isophorone
Concen: 38.425 ng
RT: 7.710 min Scan# 9
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32


Tgt Ion: 82 Resp: 65446


Ion	Ratio	Lower	Upper
82	100		
95	7.2	5.5	8.3
138	16.7	13.2	19.8

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#26
2-Nitrophenol
Concen: 40.704 ng
RT: 7.786 min Scan# 968
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:139 Resp: 151084

Ion	Ratio	Lower	Upper
139	100		
109	31.1	25.1	37.7
65	57.4	47.5	71.3






#27
2,4-Dimethylphenol
Concen: 35.678 ng
RT: 7.816 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:122 Resp: 220819


Ion	Ratio	Lower	Upper
122	100		
107	118.6	91.9	137.9
121	59.5	46.2	69.4



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#28
bis(2-Chloroethoxy)methane
Concen: 38.000 ng
RT: 7.922 min Scan# 991
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion: 93 Resp: 392131

Ion	Ratio	Lower	Upper
93	100		
95	32.7	26.0	39.0
123	12.0	9.8	14.8






#29
2,4-Dichlorophenol
Concen: 39.331 ng
RT: 8.028 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:162 Resp: 27726


Ion	Ratio	Lower	Upper
162	100		
164	63.6	43.9	83.9
98	37.9	19.9	59.9



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#30
1,2,4-Trichlorobenzene
Concen: 40.164 ng
RT: 8.116 min Scan# 1024
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:180 Resp: 313301

Ion	Ratio	Lower	Upper
180	100		
182	100.6	80.5	120.7
145	32.0	25.4	38.0






#31
Naphthalene
Concen: 39.413 ng
RT: 8.198 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:128 Resp: 101097


Ion	Ratio	Lower	Upper
128	100		
129	11.1	8.9	13.3
127	13.3	10.6	16.0



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#32
Benzoic acid
Concen: 34.127 ng
RT: 7.928 min Scan# 992
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:122 Resp: 184222

Ion	Ratio	Lower	Upper
122	100		
105	129.3	107.6	147.6
77	98.6	80.2	120.2






#33
4-Chloroaniline
Concen: 38.998 ng
RT: 8.239 min Scan# 10
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:127 Resp: 34110


Ion	Ratio	Lower	Upper
127	100		
129	32.3	25.4	38.2
65	34.5	28.4	42.6
92	20.6	16.6	25.0



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#34
Hexachlorobutadiene
Concen: 40.501 ng
RT: 8.310 min Scan# 1057
Delta R.T. -0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:225 Resp: 198499

Ion	Ratio	Lower	Upper
225	100		
223	63.6	50.6	76.0
227	64.9	51.8	77.8






#35
Caprolactam
Concen: 39.542 ng
RT: 8.616 min Scan# 11
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:113 Resp: 8863


Ion	Ratio	Lower	Upper
113	100		
55	193.0	189.2	229.2
56	146.0	132.5	172.5



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#36
4-Chloro-3-methylphenol
Concen: 39.128 ng
RT: 8.716 min Scan# 1126
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:107 Resp: 305431

Ion	Ratio	Lower	Upper
107	100		
144	24.4	19.1	28.7
142	73.2	57.2	85.8





#37
2-Methylnaphthalene
Concen: 39.271 ng
RT: 8.886 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32


Tgt Ion:142 Resp: 61692


Ion	Ratio	Lower	Upper
142	100		
141	89.2	70.8	106.2

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#38
1-Methylnaphthalene
Concen: 39.400 ng
RT: 8.986 min Scan# 1172
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:142 Resp: 607270

Ion	Ratio	Lower	Upper
142	100		
141	92.0	73.5	110.3






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.927 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:164 Resp: 27406


Ion	Ratio	Lower	Upper
164	100		
162	102.4	81.0	121.4
160	45.1	35.4	53.0



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#40
1,2,4,5-Tetrachlorobenzene
Concen: 41.099 ng
RT: 9.051 min Scan# 1183
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:216 Resp: 303883

Ion	Ratio	Lower	Upper
216	100		
214	80.2	63.4	95.0
179	20.9	17.2	25.8
108	18.9	15.6	23.4






#41
Hexachlorocyclopentadiene
Concen: 38.781 ng
RT: 9.039 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:237 Resp: 9977


Ion	Ratio	Lower	Upper
237	100		
235	62.8	42.9	82.9
272	12.8	0.0	32.3



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#42
2,4,6-Tribromophenol
Concen: 84.285 ng
RT: 10.716 min Scan# 1466
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:330 Resp: 216066

Ion	Ratio	Lower	Upper
330	100		
332	96.6	78.1	117.1
141	31.7	26.6	39.8






#43
2,4,6-Trichlorophenol
Concen: 39.258 ng
RT: 9.157 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:196 Resp: 20550


Ion	Ratio	Lower	Upper
196	100		
198	96.0	76.2	114.2
200	30.5	24.0	36.0



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#44
2,4,5-Trichlorophenol
Concen: 39.754 ng
RT: 9.198 min Scan# 1208
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:196 Resp: 214705

Ion	Ratio	Lower	Upper
196	100		
198	95.8	75.8	113.6
97	50.4	41.8	62.6
132	27.9	22.9	34.3





#45
2-Fluorobiphenyl
Concen: 79.000 ng
RT: 9.245 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32


Tgt Ion:172 Resp: 131004


Ion	Ratio	Lower	Upper
172	100		
171	35.9	28.6	43.0
170	23.8	19.1	28.7

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#46
1,1'-Biphenyl
Concen: 39.635 ng
RT: 9.351 min Scan# 1234
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:154 Resp: 756187

Ion	Ratio	Lower	Upper
154	100		
153	40.4	20.3	60.3
76	13.3	0.0	33.9






#47
2-Chloronaphthalene
Concen: 40.398 ng
RT: 9.375 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:162 Resp: 62077


Ion	Ratio	Lower	Upper
162	100		
127	42.1	33.8	50.6
164	32.6	26.6	39.8



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#48
2-Nitroaniline
Concen: 42.796 ng
RT: 9.469 min Scan# 1254
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion: 65 Resp: 201129

Ion	Ratio	Lower	Upper
65	100		
92	65.6	50.9	76.3
138	99.2	73.4	110.0






#49
Acenaphthylene
Concen: 40.124 ng
RT: 9.792 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:152 Resp: 89136


Ion	Ratio	Lower	Upper
152	100		
151	20.1	15.9	23.9
153	13.1	10.7	16.1



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#50
Dimethylphthalate
Concen: 39.953 ng
RT: 9.651 min Scan# 1285
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:163 Resp: 682031

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.4	5.0
164	9.8	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 43.701 ng
RT: 9.710 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32


Tgt Ion:165 Resp: 16152


Ion	Ratio	Lower	Upper
165	100		
63	61.6	50.6	75.8
89	58.0	47.5	71.3

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#52
Acenaphthene
Concen: 40.343 ng
RT: 9.963 min Scan# 1338
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:154 Resp: 579769

Ion	Ratio	Lower	Upper
154	100		
153	112.4	89.8	134.6
152	54.4	42.4	63.6





#53
3-Nitroaniline
Concen: 42.148 ng
RT: 9.874 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32


Tgt Ion:138 Resp: 16065


Ion	Ratio	Lower	Upper
138	100		
108	11.0	10.2	15.2
92	125.0	98.3	147.5

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#54
2,4-Dinitrophenol
Concen: 45.309 ng
RT: 9.980 min Scan# 1341
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:184 Resp: 64847

Ion	Ratio	Lower	Upper
184	100		
63	78.4	84.7	127.1#
154	125.1	338.2	507.2#






#55
Dibenzofuran
Concen: 39.443 ng
RT: 10.133 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:168 Resp: 81326


Ion	Ratio	Lower	Upper
168	100		
139	40.0	31.8	47.6
169	14.2	11.6	17.4



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#56
4-Nitrophenol
Concen: 40.940 ng
RT: 10.027 min Scan# 1349
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:139 Resp: 120058

Ion	Ratio	Lower	Upper
139	100		
109	76.0	53.6	93.6
65	111.5	87.8	127.8






#57
2,4-Dinitrotoluene
Concen: 46.125 ng
RT: 10.110 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:165 Resp: 20965


Ion	Ratio	Lower	Upper
165	100		
63	49.7	40.1	60.1
89	76.1	60.4	90.6
182	2.7	2.0	3.0



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#58
Fluorene
Concen: 40.338 ng
RT: 10.474 min Scan# 1425
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:166 Resp: 633488

Ion	Ratio	Lower	Upper
166	100		
165	97.7	78.5	117.7
167	13.6	10.9	16.3






#59
2,3,4,6-Tetrachlorophenol
Concen: 40.257 ng
RT: 10.251 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:232 Resp: 170444


Ion	Ratio	Lower	Upper
232	100		
131	45.6	37.0	55.6
130	2.1	1.9	2.9
166	28.5	24.2	36.2



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#60
Diethylphthalate
Concen: 40.303 ng
RT: 10.345 min Scan# 1403
Delta R.T. -0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:149 Resp: 675530

Ion	Ratio	Lower	Upper
149	100		
177	22.2	17.9	26.9
150	12.6	10.1	15.1





#61
4-Chlorophenyl-phenylether
Concen: 40.907 ng
RT: 10.469 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32


Tgt Ion:204 Resp: 32443


Ion	Ratio	Lower	Upper
204	100		
206	32.5	26.3	39.5
141	59.6	48.6	73.0

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#62
4-Nitroaniline
Concen: 43.535 ng
RT: 10.492 min Scan# 1428
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:138 Resp: 156727

Ion	Ratio	Lower	Upper
138	100		
92	52.9	35.7	75.7
108	78.6	57.5	97.5






#63
Azobenzene
Concen: 38.577 ng
RT: 10.627 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion: 77 Resp: 68549


Ion	Ratio	Lower	Upper
77	100		
182	24.7	3.3	43.3
105	21.5	1.2	41.2
51	32.6	13.3	53.3



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.416 min Scan# 1585
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:188 Resp: 494703

Ion	Ratio	Lower	Upper
188	100		
94	9.0	7.9	11.9
80	9.9	9.0	13.4





#65
4,6-Dinitro-2-methylphenol
Concen: 47.659 ng
RT: 10.522 min Scan# 14
Delta R.T. 0.005 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32


Tgt Ion:198 Resp: 9616


Ion	Ratio	Lower	Upper
198	100		
51	51.1	38.2	78.2
105	43.0	27.9	67.9

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#66
n-Nitrosodiphenylamine
Concen: 38.436 ng
RT: 10.586 min Scan# 1444
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:169 Resp: 569092

Ion	Ratio	Lower	Upper
169	100		
168	62.5	49.8	74.6
167	35.2	28.0	42.0






#67
4-Bromophenyl-phenylether
Concen: 40.268 ng
RT: 10.957 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:248 Resp: 20522


Ion	Ratio	Lower	Upper
248	100		
250	98.0	79.0	118.6
141	74.2	61.7	92.5



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#68
Hexachlorobenzene
Concen: 40.175 ng
RT: 11.021 min Scan# 1518
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:284 Resp: 230272

Ion	Ratio	Lower	Upper
284	100		
142	32.7	26.9	40.3
249	31.8	25.3	37.9






#69
Atrazine
Concen: 38.680 ng
RT: 11.110 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:200 Resp: 15513


Ion	Ratio	Lower	Upper
200	100		
173	26.7	7.7	47.7
215	47.6	27.7	67.7



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#70
Pentachlorophenol
Concen: 40.865 ng
RT: 11.216 min Scan# 1551
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:266 Resp: 142153

Ion	Ratio	Lower	Upper
266	100		
268	65.1	50.4	75.6
264	63.6	50.6	75.8






#71
Phenanthrene
Concen: 38.500 ng
RT: 11.439 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:178 Resp: 89964


Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.8	23.6
179	15.4	12.6	18.8



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#72
Anthracene
Concen: 38.994 ng
RT: 11.492 min Scan# 1598
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:178 Resp: 889039

Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.3	22.9
179	15.4	12.4	18.6






#73
Carbazole
Concen: 37.870 ng
RT: 11.645 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:167 Resp: 80300


Ion	Ratio	Lower	Upper
167	100		
166	21.6	17.0	25.6
139	13.3	10.8	16.2



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#74
Di-n-butylphthalate
Concen: 39.082 ng
RT: 11.974 min Scan# 1680
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:149 Resp: 957414

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.5	11.3
104	4.9	4.1	6.1






#75
Fluoranthene
Concen: 38.158 ng
RT: 12.627 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:202 Resp: 90140


Ion	Ratio	Lower	Upper
202	100		
101	11.3	0.0	31.9
203	17.8	0.0	37.5



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.051 min Scan# 2033
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:240 Resp: 250901

Ion	Ratio	Lower	Upper
240	100		
120	11.0	9.4	14.2
236	25.6	20.9	31.3





#77
Benzidine
Concen: 45.296 ng
RT: 12.745 min Scan# 18
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32


Tgt Ion:184 Resp: 18687


Ion	Ratio	Lower	Upper
184	100		
185	14.2	12.0	18.0
183	11.5	9.4	14.2

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#78
Pyrene
Concen: 40.730 ng
RT: 12.857 min Scan# 1830
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:202 Resp: 894517

Ion	Ratio	Lower	Upper
202	100		
200	21.3	17.2	25.8
203	17.5	14.2	21.4






#79
Terphenyl-d14
Concen: 81.484 ng
RT: 12.998 min Scan# 18
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:244 Resp: 125465


Ion	Ratio	Lower	Upper
244	100		
212	6.8	5.7	8.5
122	10.0	8.6	13.0



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#80
Butylbenzylphthalate
Concen: 41.265 ng
RT: 13.468 min Scan# 1934
Delta R.T. -0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:149 Resp: 271701

Ion	Ratio	Lower	Upper
149	100		
91	75.1	59.4	89.2
206	20.0	16.4	24.6






#81
Benzo(a)anthracene
Concen: 40.732 ng
RT: 14.039 min Scan# 20
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:228 Resp: 665269


Ion	Ratio	Lower	Upper
228	100		
226	26.6	21.6	32.4
229	19.6	16.1	24.1



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#82
3,3'-Dichlorobenzidine
Concen: 44.426 ng
RT: 14.004 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:252 Resp: 211560

Ion	Ratio	Lower	Upper
252	100		
254	64.1	50.8	76.2
126	10.5	8.9	13.3






#83
Chrysene
Concen: 39.262 ng
RT: 14.080 min Scan# 20
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:228 Resp: 587960


Ion	Ratio	Lower	Upper
228	100		
226	29.7	24.1	36.1
229	19.6	15.8	23.6



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#84
Bis(2-ethylhexyl)phthalate
Concen: 45.339 ng
RT: 14.027 min Scan# 2029
Delta R.T. -0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:149 Resp: 334928

Ion	Ratio	Lower	Upper
149	100		
167	26.2	21.4	32.0
279	4.0	3.0	4.4






#85
Di-n-octyl phthalate
Concen: 40.754 ng
RT: 14.639 min Scan# 21
Delta R.T. -0.012 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:149 Resp: 54758


Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	11.5	9.8	14.6



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#86
Perylene-d12
Concen: 20.000 ng
RT: 15.527 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:264 Resp: 263501

Ion	Ratio	Lower	Upper
264	100		
260	24.1	19.4	29.2
265	22.0	17.4	26.0






#87
Indeno(1,2,3-cd)pyrene
Concen: 40.396 ng
RT: 17.033 min Scan# 2
Delta R.T. 0.012 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:276 Resp: 68478


Ion	Ratio	Lower	Upper
276	100		
138	23.3	20.4	30.6
277	25.3	20.2	30.4



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#88
Benzo(b)fluoranthene
Concen: 38.345 ng
RT: 15.098 min Scan# 2211
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:252 Resp: 615487

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.6	26.4
125	9.0	7.9	11.9






#89
Benzo(k)fluoranthene
Concen: 40.721 ng
RT: 15.127 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:252 Resp: 56370


Ion	Ratio	Lower	Upper
252	100		
253	22.4	17.6	26.4
125	9.3	7.9	11.9



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#90
Benzo(a)pyrene
Concen: 40.379 ng
RT: 15.468 min Scan# 2274
Delta R.T. 0.000 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:252 Resp: 533309

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.3	25.9
125	10.4	8.7	13.1






#91
Dibenzo(a,h)anthracene
Concen: 39.881 ng
RT: 17.051 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:278 Resp: 56398


Ion	Ratio	Lower	Upper
278	100		
139	14.5	13.1	19.7
279	23.9	19.0	28.4



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#92
Benzo(g,h,i)perylene
Concen: 41.284 ng
RT: 17.486 min Scan# 2617
Delta R.T. 0.006 min
Lab File: BF140075.D
Acq: 28 Oct 2024 09:32

Tgt Ion:276 Resp: 583161

Ion	Ratio	Lower	Upper
276	100		
277	23.6	19.0	28.6
138	20.4	17.4	26.0



Evaluate Continuing Calibration Report

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140075.D
Acq On : 28 Oct 2024 09:32
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 28 09:55:45 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	73	0.00
2		1,4-Dioxane	40.000	37.538	6.2	69	-0.02
3		Pyridine	40.000	38.244	4.4	68	0.00
4		n-Nitrosodimethylamine	40.000	37.610	6.0	67	-0.01
5	S	2-Fluorophenol	80.000	75.045	6.2	68	0.00
6		Aniline	40.000	37.992	5.0	67	0.00
7	S	Phenol-d6	80.000	74.300	7.1	68	0.00
8		2-Chlorophenol	40.000	37.936	5.2	69	0.00
9		Benzaldehyde	40.000	38.007	5.0	69	0.00
10	C	Phenol	40.000	37.802	5.5	69	0.00
11		bis(2-Chloroethyl)ether	40.000	37.186	7.0	69	0.00
12		1,3-Dichlorobenzene	40.000	39.802	0.5	73	0.00
13	C	1,4-Dichlorobenzene	40.000	39.676	0.8	73	0.00
14		1,2-Dichlorobenzene	40.000	39.360	1.6	73	0.00
15		Benzyl Alcohol	40.000	38.244	4.4	70	0.00
16		2,2'-oxybis(1-Chloropropane	40.000	36.130	9.7	66	0.00
17		2-Methylphenol	40.000	37.712	5.7	69	0.00
18		Hexachloroethane	40.000	39.369	1.6	71	0.00
19	P	n-Nitroso-di-n-propylamine	40.000	36.671	8.3	69	0.00
20		3+4-Methylphenols	40.000	37.070	7.3	68	0.00
21	I	Naphthalene-d8	20.000	20.000	0.0	72	0.00
22		Acetophenone	40.000	38.331	4.2	70	0.00
23	S	Nitrobenzene-d5	80.000	83.584	-4.5	73	0.00
24		Nitrobenzene	40.000	40.136	-0.3	71	0.00
25		Isophorone	40.000	38.425	3.9	70	0.00
26	C	2-Nitrophenol	40.000	40.704	-1.8	70	0.00
27		2,4-Dimethylphenol	40.000	35.678	10.8	65	0.00
28		bis(2-Chloroethoxy)methane	40.000	38.000	5.0	69	0.00
29	C	2,4-Dichlorophenol	40.000	39.331	1.7	71	0.00
30		1,2,4-Trichlorobenzene	40.000	40.164	-0.4	72	0.00
31		Naphthalene	40.000	39.413	1.5	72	0.00
32		Benzoic acid	40.000	34.127	14.7	61	0.00
33		4-Chloroaniline	40.000	38.998	2.5	70	0.00
34	C	Hexachlorobutadiene	40.000	40.501	-1.3	73	0.00
35		Caprolactam	40.000	39.542	1.1	70	0.00
36	C	4-Chloro-3-methylphenol	40.000	39.128	2.2	70	0.00
37		2-Methylnaphthalene	40.000	39.271	1.8	72	0.00
38		1-Methylnaphthalene	40.000	39.400	1.5	72	0.00
39	I	Acenaphthene-d10	20.000	20.000	0.0	72	0.00
40		1,2,4,5-Tetrachlorobenzene	40.000	41.099	-2.7	75	0.00
41	P	Hexachlorocyclopentadiene	40.000	38.781	3.0	66	0.00
42	S	2,4,6-Tribromophenol	80.000	84.285	-5.4	76	0.00
43	C	2,4,6-Trichlorophenol	40.000	39.258	1.9	70	0.00
44		2,4,5-Trichlorophenol	40.000	39.754	0.6	71	0.00
45	S	2-Fluorobiphenyl	80.000	79.000	1.3	74	0.00
46		1,1'-Biphenyl	40.000	39.635	0.9	72	0.00
47		2-Chloronaphthalene	40.000	40.398	-1.0	74	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140075.D
Acq On : 28 Oct 2024 09:32
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 28 09:55:45 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	42.796	-7.0	73	0.00
49		Acenaphthylene	40.000	40.124	-0.3	73	0.00
50		Dimethylphthalate	40.000	39.953	0.1	73	0.00
51		2,6-Dinitrotoluene	40.000	43.701	-9.3	76	0.00
52	C	Acenaphthene	40.000	40.343	-0.9	73	0.00
53		3-Nitroaniline	40.000	42.148	-5.4	72	0.00
54	P	2,4-Dinitrophenol	40.000	45.309	-13.3	84	0.00
55		Dibenzofuran	40.000	39.443	1.4	72	0.00
56	P	4-Nitrophenol	40.000	40.940	-2.3	68	0.00
57		2,4-Dinitrotoluene	40.000	46.125	-15.3	77	0.00
58		Fluorene	40.000	40.338	-0.8	75	0.00
59		2,3,4,6-Tetrachlorophenol	40.000	40.257	-0.6	72	0.00
60		Diethylphthalate	40.000	40.303	-0.8	73	0.00
61		4-Chlorophenyl-phenylether	40.000	40.907	-2.3	75	0.00
62		4-Nitroaniline	40.000	43.535	-8.8	75	0.00
63		Azobenzene	40.000	38.577	3.6	69	0.00
64	I	Phenanthrene-d10	20.000	20.000	0.0	75	0.00
65		4,6-Dinitro-2-methylphenol	40.000	47.659	-19.1	83	0.00
66	c	n-Nitrosodiphenylamine	40.000	38.436	3.9	72	0.00
67		4-Bromophenyl-phenylether	40.000	40.268	-0.7	76	0.00
68		Hexachlorobenzene	40.000	40.175	-0.4	75	0.00
69		Atrazine	40.000	38.680	3.3	67	0.00
70	C	Pentachlorophenol	40.000	40.865	-2.2	71	0.00
71		Phenanthrene	40.000	38.500	3.8	73	0.00
72		Anthracene	40.000	38.994	2.5	73	0.00
73		Carbazole	40.000	37.870	5.3	72	0.00
74		Di-n-butylphthalate	40.000	39.082	2.3	71	0.00
75	C	Fluoranthene	40.000	38.158	4.6	72	0.00
76	I	Chrysene-d12	20.000	20.000	0.0	73	0.00
77		Benzidine	40.000	45.296	-13.2	74	0.00
78		Pyrene	40.000	40.730	-1.8	72	0.00
79	S	Terphenyl-d14	80.000	81.484	-1.9	73	0.00
80		Butylbenzylphthalate	40.000	41.265	-3.2	71	0.00
81		Benzo(a)anthracene	40.000	40.732	-1.8	72	0.00
82		3,3'-Dichlorobenzidine	40.000	44.426	-11.1	79	0.00
83		Chrysene	40.000	39.262	1.8	73	0.00
84		Bis(2-ethylhexyl)phthalate	40.000	45.339	-13.3	77	0.00
85	c	Di-n-octyl phthalate	40.000	40.754	-1.9	70	-0.01
86	I	Perylene-d12	20.000	20.000	0.0	72	0.00
87		Indeno(1,2,3-cd)pyrene	40.000	40.396	-1.0	69	0.01
88		Benzo(b)fluoranthene	40.000	38.345	4.1	70	0.00
89		Benzo(k)fluoranthene	40.000	40.721	-1.8	72	0.00
90	C	Benzo(a)pyrene	40.000	40.379	-0.9	71	0.00
91		Dibenzo(a,h)anthracene	40.000	39.881	0.3	69	0.00
92		Benzo(g,h,i)perylene	40.000	41.284	-3.2	71	0.00

Evaluate Continuing Calibration Report

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140075.D
Acq On : 28 Oct 2024 09:32
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 28 09:55:45 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Evaluate Continuing Calibration Report

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140075.D
Acq On : 28 Oct 2024 09:32
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 28 09:55:45 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	73	0.00
2		1,4-Dioxane	40.000	37.538	6.2	69	-0.02
3		Pyridine	40.000	38.244	4.4	68	0.00
4		n-Nitrosodimethylamine	40.000	37.610	6.0	67	-0.01
5	S	2-Fluorophenol	80.000	75.045	6.2	68	0.00
6		Aniline	40.000	37.992	5.0	67	0.00
7	S	Phenol-d6	80.000	74.300	7.1	68	0.00
8		2-Chlorophenol	40.000	37.936	5.2	69	0.00
9		Benzaldehyde	40.000	38.007	5.0	69	0.00
10	C	Phenol	40.000	37.802	5.5	69	0.00
11		bis(2-Chloroethyl)ether	40.000	37.186	7.0	69	0.00
12		1,3-Dichlorobenzene	40.000	39.802	0.5	73	0.00
13	C	1,4-Dichlorobenzene	40.000	39.676	0.8	73	0.00
14		1,2-Dichlorobenzene	40.000	39.360	1.6	73	0.00
15		Benzyl Alcohol	40.000	38.244	4.4	70	0.00
16		2,2'-oxybis(1-Chloropropane	40.000	36.130	9.7	66	0.00
17		2-Methylphenol	40.000	37.712	5.7	69	0.00
18		Hexachloroethane	40.000	39.369	1.6	71	0.00
19	P	n-Nitroso-di-n-propylamine	40.000	36.671	8.3	69	0.00
20		3+4-Methylphenols	40.000	37.070	7.3	68	0.00
21	I	Naphthalene-d8	20.000	20.000	0.0	72	0.00
22		Acetophenone	40.000	38.331	4.2	70	0.00
23	S	Nitrobenzene-d5	80.000	83.584	-4.5	73	0.00
24		Nitrobenzene	40.000	40.136	-0.3	71	0.00
25		Isophorone	40.000	38.425	3.9	70	0.00
26	C	2-Nitrophenol	40.000	40.704	-1.8	70	0.00
27		2,4-Dimethylphenol	40.000	35.678	10.8	65	0.00
28		bis(2-Chloroethoxy)methane	40.000	38.000	5.0	69	0.00
29	C	2,4-Dichlorophenol	40.000	39.331	1.7	71	0.00
30		1,2,4-Trichlorobenzene	40.000	40.164	-0.4	72	0.00
31		Naphthalene	40.000	39.413	1.5	72	0.00
32		Benzoic acid	40.000	34.127	14.7	61	0.00
33		4-Chloroaniline	40.000	38.998	2.5	70	0.00
34	C	Hexachlorobutadiene	40.000	40.501	-1.3	73	0.00
35		Caprolactam	40.000	39.542	1.1	70	0.00
36	C	4-Chloro-3-methylphenol	40.000	39.128	2.2	70	0.00
37		2-Methylnaphthalene	40.000	39.271	1.8	72	0.00
38		1-Methylnaphthalene	40.000	39.400	1.5	72	0.00
39	I	Acenaphthene-d10	20.000	20.000	0.0	72	0.00
40		1,2,4,5-Tetrachlorobenzene	40.000	41.099	-2.7	75	0.00
41	P	Hexachlorocyclopentadiene	40.000	38.781	3.0	66	0.00
42	S	2,4,6-Tribromophenol	80.000	84.285	-5.4	76	0.00
43	C	2,4,6-Trichlorophenol	40.000	39.258	1.9	70	0.00
44		2,4,5-Trichlorophenol	40.000	39.754	0.6	71	0.00
45	S	2-Fluorobiphenyl	80.000	79.000	1.3	74	0.00
46		1,1'-Biphenyl	40.000	39.635	0.9	72	0.00
47		2-Chloronaphthalene	40.000	40.398	-1.0	74	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140075.D
Acq On : 28 Oct 2024 09:32
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 28 09:55:45 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	42.796	-7.0	73	0.00
49		Acenaphthylene	40.000	40.124	-0.3	73	0.00
50		Dimethylphthalate	40.000	39.953	0.1	73	0.00
51		2,6-Dinitrotoluene	40.000	43.701	-9.3	76	0.00
52	C	Acenaphthene	40.000	40.343	-0.9	73	0.00
53		3-Nitroaniline	40.000	42.148	-5.4	72	0.00
54	P	2,4-Dinitrophenol	40.000	45.309	-13.3	84	0.00
55		Dibenzofuran	40.000	39.443	1.4	72	0.00
56	P	4-Nitrophenol	40.000	40.940	-2.3	68	0.00
57		2,4-Dinitrotoluene	40.000	46.125	-15.3	77	0.00
58		Fluorene	40.000	40.338	-0.8	75	0.00
59		2,3,4,6-Tetrachlorophenol	40.000	40.257	-0.6	72	0.00
60		Diethylphthalate	40.000	40.303	-0.8	73	0.00
61		4-Chlorophenyl-phenylether	40.000	40.907	-2.3	75	0.00
62		4-Nitroaniline	40.000	43.535	-8.8	75	0.00
63		Azobenzene	40.000	38.577	3.6	69	0.00
64	I	Phenanthrene-d10	20.000	20.000	0.0	75	0.00
65		4,6-Dinitro-2-methylphenol	40.000	47.659	-19.1	83	0.00
66	c	n-Nitrosodiphenylamine	40.000	38.436	3.9	72	0.00
67		4-Bromophenyl-phenylether	40.000	40.268	-0.7	76	0.00
68		Hexachlorobenzene	40.000	40.175	-0.4	75	0.00
69		Atrazine	40.000	38.680	3.3	67	0.00
70	C	Pentachlorophenol	40.000	40.865	-2.2	71	0.00
71		Phenanthrene	40.000	38.500	3.8	73	0.00
72		Anthracene	40.000	38.994	2.5	73	0.00
73		Carbazole	40.000	37.870	5.3	72	0.00
74		Di-n-butylphthalate	40.000	39.082	2.3	71	0.00
75	C	Fluoranthene	40.000	38.158	4.6	72	0.00
76	I	Chrysene-d12	20.000	20.000	0.0	73	0.00
77		Benzidine	40.000	45.296	-13.2	74	0.00
78		Pyrene	40.000	40.730	-1.8	72	0.00
79	S	Terphenyl-d14	80.000	81.484	-1.9	73	0.00
80		Butylbenzylphthalate	40.000	41.265	-3.2	71	0.00
81		Benzo(a)anthracene	40.000	40.732	-1.8	72	0.00
82		3,3'-Dichlorobenzidine	40.000	44.426	-11.1	79	0.00
83		Chrysene	40.000	39.262	1.8	73	0.00
84		Bis(2-ethylhexyl)phthalate	40.000	45.339	-13.3	77	0.00
85	c	Di-n-octyl phthalate	40.000	40.754	-1.9	70	-0.01
86	I	Perylene-d12	20.000	20.000	0.0	72	0.00
87		Indeno(1,2,3-cd)pyrene	40.000	40.396	-1.0	69	0.01
88		Benzo(b)fluoranthene	40.000	38.345	4.1	70	0.00
89		Benzo(k)fluoranthene	40.000	40.721	-1.8	72	0.00
90	C	Benzo(a)pyrene	40.000	40.379	-0.9	71	0.00
91		Dibenzo(a,h)anthracene	40.000	39.881	0.3	69	0.00
92		Benzo(g,h,i)perylene	40.000	41.284	-3.2	71	0.00

Evaluate Continuing Calibration Report

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140075.D
Acq On : 28 Oct 2024 09:32
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 28 09:55:45 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

QC SAMPLE DATA


DFTPP

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139843.D
Acq On : 18 Oct 2024 09:22
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Oct 18 15:07:50 2024



AutoFind: Scans 1627, 1628, 1629; Background Corrected with Scan 1621

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.8	92701	PASS
68	69	0.00	2	1.9	1486	PASS
69	198	0.00	100	37.8	79959	PASS
70	69	0.00	2	0.7	529	PASS
127	198	10	80	46.9	99176	PASS
197	198	0.00	2	0.8	1650	PASS
198	198	100	100	100.0	211520	PASS
199	198	5	9	6.7	14078	PASS
275	198	10	60	26.9	56867	PASS
365	198	1	100	3.4	7148	PASS
441	198	0.01	100	14.0	29512	PASS
442	442	50	100	100.0	196979	PASS
443	442	15	24	19.2	37896	PASS

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/18/2024	BNA_F	BF139843.D
Compound Name	**Response**	**Retention Time**
DDT	585942	13.574
DDD	33558	13.274
DDE	421	12.939
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
33979	619921	5.48


Instrument :
BNA_F
ClientSampleId :
DFTPP


Quantitation Report (Qedit)


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139843.D
Acq On : 18 Oct 2024 09:22
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Oct 18 11:03:23 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 11:03:04 2024
Response via : Initial Calibration







TIC: BF139843.D\data.ms

(70) Pentachlorophenol (C)

11.210min (-0.006) 60102.24 ng

response 277191


Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.00	63.60
264.00	62.50	63.30
0.00	0.00	0.00


Quantitation Report (Qedit)


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF101824\
Data File : BF139843.D
Acq On : 18 Oct 2024 09:22
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Oct 18 11:03:23 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 11:03:04 2024
Response via : Initial Calibration







TIC: BF139843.D\data.ms

(77) Benzidine

12.751min (+ 0.006) 86231.01 ng

response 1303015

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	14.39
183.00	11.70	11.82
0.00	0.00	0.00


DFTPP

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140074.D
Acq On : 28 Oct 2024 09:04
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Oct 18 15:07:50 2024



AutoFind: Scans 1626, 1627, 1628; Background Corrected with Scan 1620


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.7	78035	PASS
68	69	0.00	2	1.6	1158	PASS
69	198	0.00	100	38.6	70576	PASS
70	69	0.00	2	0.5	343	PASS
127	198	10	80	48.1	87888	PASS
197	198	0.00	2	0.8	1489	PASS
198	198	100	100	100.0	182720	PASS
199	198	5	9	6.8	12368	PASS
275	198	10	60	28.6	52275	PASS
365	198	1	100	3.5	6329	PASS
441	198	0.01	100	16.1	29440	PASS
442	442	50	100	100.0	189696	PASS
443	442	15	24	18.8	35661	PASS


Quantitation Report (Qedit)


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140074.D
Acq On : 28 Oct 2024 09:04
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Oct 28 11:24:41 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration







TIC: BF140074.D\data.ms

(70) Pentachlorophenol (C)

11.216min (+ 0.006) 128927.16 ng

response 209420


Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.00	64.64
264.00	63.20	63.11
0.00	0.00	0.00


Quantitation Report (Qedit)


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140074.D
Acq On : 28 Oct 2024 09:04
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Oct 28 11:24:41 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration







TIC: BF140074.D\data.ms

(77) Benzidine

12.751min (+ 0.006) 156036.69 ng

response 1051967

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.00	14.13
183.00	11.80	11.80
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/28/2024	BNA_F	BF140074.D
Compound Name	**Response**	**Retention Time**
DDT	521702	13.574
DDD	8598	13.304
DDE	1230	12.933
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
9828	531530	1.85

Instrument :
BNA_F
ClientSampleId :
DFTPP

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	
Client Sample ID:	PB164444BL			SDG No.:	P4578
Lab Sample ID:	PB164444BL			Matrix:	TCLP
Analytical Method:	SW8270			% Solid:	0
Sample Wt/Vol:	1000	Units: mL		Final Vol:	1000 uL
Soil Aliquot Vol:		uL		Test:	TCLP BNA
Extraction Type :		Decanted : N		Level :	LOW
Injection Volume :		GPC Factor : 1.0		GPC Cleanup : N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140076.D	1	10/26/24 12:10	10/28/24 10:01	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	1.60	U	1.60	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.84	U	0.84	5.00	ug/L
95-48-7	2-Methylphenol	1.10	U	1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	1.20	U	1.20	10.0	ug/L
67-72-1	Hexachloroethane	1.00	U	1.00	5.00	ug/L
98-95-3	Nitrobenzene	1.30	U	1.30	5.00	ug/L
87-68-3	Hexachlorobutadiene	1.30	U	1.30	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	0.89	U	0.89	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	1.00	U	1.00	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	1.50	U	1.50	5.00	ug/L
118-74-1	Hexachlorobenzene	1.10	U	1.10	5.00	ug/L
87-86-5	Pentachlorophenol	1.90	U	1.90	10.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	120		15 (10) - 110 (139)	80%	SPK: 150
13127-88-3	Phenol-d6	114		15 (10) - 110 (134)	76%	SPK: 150
4165-60-0	Nitrobenzene-d5	90.9		30 (49) - 130 (133)	91%	SPK: 100
321-60-8	2-Fluorobiphenyl	87.1		30 (52) - 130 (132)	87%	SPK: 100
118-79-6	2,4,6-Tribromophenol	133		15 (44) - 110 (137)	89%	SPK: 150
1718-51-0	Terphenyl-d14	89.1		30 (48) - 130 (125)	89%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	132000	6.886			
1146-65-2	Naphthalene-d8	518000	8.169			
15067-26-2	Acenaphthene-d10	298000	9.922			
1517-22-2	Phenanthrene-d10	551000	11.41			
1719-03-5	Chrysene-d12	342000	14.045			
1520-96-3	Perylene-d12	261000	15.527			

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	
Client Sample ID:	PB164444BL			SDG No.:	P4578
Lab Sample ID:	PB164444BL			Matrix:	TCLP
Analytical Method:	SW8270			% Solid:	0
Sample Wt/Vol:	1000	Units:	mL	Final Vol:	1000 uL
Soil Aliquot Vol:			uL	Test:	TCLP BNA
Extraction Type :			Decanted : N	Level :	LOW
Injection Volume :			GPC Factor : 1.0	GPC Cleanup : N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140076.D	1	10/26/24 12:10	10/28/24 10:01	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Quantitation Report (QT/LSC Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140076.D
Acq On : 28 Oct 2024 10:01
Operator : RC/JU
Sample : PB164444BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB164444BL

Quant Time: Oct 28 11:22:30 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.886	152	132330	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	518019	20.000	ng	0.00
39) Acenaphthene-d10	9.922	164	298114	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	550598	20.000	ng	0.00
76) Chrysene-d12	14.045	240	341644	20.000	ng	0.00
86) Perylene-d12	15.527	264	261224	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	1015752	120.165	ng	0.02
7) Phenol-d6	6.510	99	1244979	113.718	ng	0.00
23) Nitrobenzene-d5	7.445	82	849954	90.938	ng	0.00
42) 2,4,6-Tribromophenol	10.716	330	370207	132.764	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1571487	87.121	ng	0.00
79) Terphenyl-d14	12.998	244	1868766	89.131	ng	0.00
Target Compounds						Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140076.D
Acq On : 28 Oct 2024 10:01
Operator : RC/JU
Sample : PB164444BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB164444BL

Quant Time: Oct 28 11:22:30 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.886 min Scan# 81
Delta R.T. -0.006 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01

Instrument :
BNA_F
ClientSampleId :
PB164444BL

Tgt Ion:152 Resp: 132330

Ion	Ratio	Lower	Upper
152	100		
150	156.9	130.2	195.2
115	63.9	51.4	77.2







#5
2-Fluorophenol
Concen: 120.165 ng
RT: 5.516 min Scan# 582
Delta R.T. 0.018 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01

Tgt Ion:112 Resp: 1015752

Ion	Ratio	Lower	Upper
112	100		
64	65.1	53.0	79.6
63	34.3	27.0	40.4







#7
Phenol-d6
Concen: 113.718 ng
RT: 6.510 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01

Instrument :
BNA_F
ClientSampleId :
PB164444BL

Tgt Ion: 99 Resp: 1244979

Ion	Ratio	Lower	Upper
99	100		
42	20.0	16.7	25.1
71	36.5	27.7	41.5







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.169 min Scan# 1033
Delta R.T. -0.006 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01

Tgt Ion:136 Resp: 518019

Ion	Ratio	Lower	Upper
136	100		
137	10.7	8.6	12.8
54	10.0	8.4	12.6
68	6.1	5.1	7.7







#23
Nitrobenzene-d5
Concen: 90.938 ng
RT: 7.445 min Scan# 91
Delta R.T. -0.006 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01

Instrument :
BNA_F
ClientSampleId :
PB164444BL

Tgt Ion: 82 Resp: 849954

Ion	Ratio	Lower	Upper
82	100		
128	41.5	33.4	50.0
54	57.6	47.8	71.8







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.922 min Scan# 1331
Delta R.T. -0.005 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01

Tgt Ion:164 Resp: 298114

Ion	Ratio	Lower	Upper
164	100		
162	100.2	81.0	121.4
160	43.8	35.4	53.0







#42
2,4,6-Tribromophenol
Concen: 132.764 ng
RT: 10.716 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01

Instrument :
BNA_F
ClientSampleId :
PB164444BL

Tgt Ion:330 Resp: 370207

Ion	Ratio	Lower	Upper
330	100		
332	95.2	78.1	117.1
141	30.0	26.6	39.8







#45
2-Fluorobiphenyl
Concen: 87.121 ng
RT: 9.245 min Scan# 1216
Delta R.T. -0.006 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01

Tgt Ion:172 Resp: 1571487

Ion	Ratio	Lower	Upper
172	100		
171	35.9	28.6	43.0
170	23.9	19.1	28.7







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01

Instrument :
BNA_F
ClientSampleId :
PB164444BL

Tgt Ion:188 Resp: 550598

Ion	Ratio	Lower	Upper
188	100		
94	9.8	7.9	11.9
80	10.6	9.0	13.4







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.045 min Scan# 2032
Delta R.T. -0.006 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01

Tgt Ion:240 Resp: 341644

Ion	Ratio	Lower	Upper
240	100		
120	11.6	9.4	14.2
236	26.2	20.9	31.3







#79
Terphenyl-d14
Concen: 89.131 ng
RT: 12.998 min Scan# 18
Delta R.T. 0.000 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01

Instrument :
BNA_F
ClientSampleId :
PB164444BL

Tgt Ion:244 Resp: 1868766

Ion	Ratio	Lower	Upper
244	100		
212	6.9	5.7	8.5
122	10.0	8.6	13.0






#86
Perylene-d12
Concen: 20.000 ng
RT: 15.527 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BF140076.D
Acq: 28 Oct 2024 10:01

Tgt Ion:264 Resp: 261224

Ion	Ratio	Lower	Upper
264	100		
260	23.3	19.4	29.2
265	21.3	17.4	26.0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	
Client Sample ID:	PB164444BS			SDG No.:	P4578
Lab Sample ID:	PB164444BS			Matrix:	TCLP
Analytical Method:	SW8270			% Solid:	0
Sample Wt/Vol:	1000	Units: mL		Final Vol:	1000 uL
Soil Aliquot Vol:		uL		Test:	TCLP BNA
Extraction Type :		Decanted : N		Level :	LOW
Injection Volume :		GPC Factor : 1.0		GPC Cleanup : N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140077.D	1	10/26/24 12:10	10/28/24 10:29	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	38.6		1.60	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	43.7		0.84	5.00	ug/L
95-48-7	2-Methylphenol	44.4		1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	43.0		1.20	10.0	ug/L
67-72-1	Hexachloroethane	44.2		1.00	5.00	ug/L
98-95-3	Nitrobenzene	43.5		1.30	5.00	ug/L
87-68-3	Hexachlorobutadiene	44.9		1.30	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	45.9		0.89	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	44.6		1.00	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	52.4		1.50	5.00	ug/L
118-74-1	Hexachlorobenzene	45.0		1.10	5.00	ug/L
87-86-5	Pentachlorophenol	89.7	E	1.90	10.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	120		15 (10) - 110 (139)	80%	SPK: 150
13127-88-3	Phenol-d6	118		15 (10) - 110 (134)	79%	SPK: 150
4165-60-0	Nitrobenzene-d5	92.6		30 (49) - 130 (133)	93%	SPK: 100
321-60-8	2-Fluorobiphenyl	87.2		30 (52) - 130 (132)	87%	SPK: 100
118-79-6	2,4,6-Tribromophenol	141		15 (44) - 110 (137)	94%	SPK: 150
1718-51-0	Terphenyl-d14	95.7		30 (48) - 130 (125)	96%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	128000	6.887			
1146-65-2	Naphthalene-d8	495000	8.175			
15067-26-2	Acenaphthene-d10	281000	9.928			
1517-22-2	Phenanthrene-d10	511000	11.416			
1719-03-5	Chrysene-d12	263000	14.051			
1520-96-3	Perylene-d12	266000	15.527			

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	
Client Sample ID:	PB164444BS			SDG No.:	P4578
Lab Sample ID:	PB164444BS			Matrix:	TCLP
Analytical Method:	SW8270			% Solid:	0
Sample Wt/Vol:	1000	Units:	mL	Final Vol:	1000 uL
Soil Aliquot Vol:			uL	Test:	TCLP BNA
Extraction Type :			Decanted : N	Level :	LOW
Injection Volume :			GPC Factor : 1.0	GPC Cleanup : N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140077.D	1	10/26/24 12:10	10/28/24 10:29	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140077.D
Acq On : 28 Oct 2024 10:29
Operator : RC/JU
Sample : PB164444BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Oct 28 11:23:23 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.887	152	127868	20.000	ng		0.00
21) Naphthalene-d8	8.175	136	495282	20.000	ng		0.00
39) Acenaphthene-d10	9.928	164	281323	20.000	ng		0.00
64) Phenanthrene-d10	11.416	188	511374	20.000	ng		0.00
76) Chrysene-d12	14.051	240	262593	20.000	ng		0.00
86) Perylene-d12	15.527	264	265684	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.516	112	983498	120.410	ng		0.02
7) Phenol-d6	6.516	99	1248557	118.025	ng		0.00
23) Nitrobenzene-d5	7.451	82	827399	92.589	ng		0.00
42) 2,4,6-Tribromophenol	10.722	330	372077	141.398	ng		0.00
45) 2-Fluorobiphenyl	9.245	172	1485083	87.245	ng		0.00
79) Terphenyl-d14	12.998	244	1542921	95.744	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	2.781	88	135793	35.657	ng		97
3) Pyridine	3.534	79	364598	38.624	ng		98
4) n-Nitrosodimethylamine	3.481	42	210444	41.494	ng	#	99
6) Aniline	6.551	93	381952	38.741	ng		100
8) 2-Chlorophenol	6.675	128	380511	45.570	ng		98
9) Benzaldehyde	6.440	77	75047	12.611	ng		96
10) Phenol	6.528	94	480470m	44.055	ng		
11) bis(2-Chloroethyl)ether	6.622	93	369586	43.843	ng		99
12) 1,3-Dichlorobenzene	6.834	146	409351	42.706	ng		99
13) 1,4-Dichlorobenzene	6.910	146	418487	43.700	ng		99
14) 1,2-Dichlorobenzene	7.057	146	399902	44.744	ng		98
15) Benzyl Alcohol	7.028	79	358350	46.179	ng		99
16) 2,2'-oxybis(1-Chloropr...	7.163	45	620620	43.177	ng		98
17) 2-Methylphenol	7.139	107	315983	44.379	ng		98
18) Hexachloroethane	7.404	117	150834	44.169	ng		96
19) n-Nitroso-di-n-propyla...	7.304	70	281684	43.903	ng		97
20) 3+4-Methylphenols	7.292	107	391158	42.951	ng	#	77
22) Acetophenone	7.298	105	532350	43.883	ng		99
24) Nitrobenzene	7.469	77	425814	43.461	ng		99
25) Isophorone	7.710	82	773146	45.583	ng		99
26) 2-Nitrophenol	7.786	139	188880	51.101	ng		98
27) 2,4-Dimethylphenol	7.822	122	320582	52.015	ng		99
28) bis(2-Chloroethoxy)met...	7.916	93	456094	44.384	ng		99
29) 2,4-Dichlorophenol	8.028	162	318680	45.395	ng		98
30) 1,2,4-Trichlorobenzene	8.110	180	336007	43.256	ng		98
31) Naphthalene	8.192	128	1127006	44.121	ng		99
32) Benzoic acid	7.939	122	216873	40.344	ng		99
33) 4-Chloroaniline	8.239	127	159322	18.292	ng		98
34) Hexachlorobutadiene	8.310	225	219251	44.923	ng		98
35) Caprolactam	8.610	113	103223m	46.244	ng		
36) 4-Chloro-3-methylphenol	8.716	107	357353	45.972	ng		100
37) 2-Methylnaphthalene	8.886	142	713865	45.632	ng		99
38) 1-Methylnaphthalene	8.986	142	669536	43.622	ng		100
40) 1,2,4,5-Tetrachloroben...	9.051	216	338977	44.663	ng		99
41) Hexachlorocyclopentadiene	9.039	237	437259	165.576	ng		99
43) 2,4,6-Trichlorophenol	9.163	196	246720	45.915	ng		99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140077.D
Acq On : 28 Oct 2024 10:29
Operator : RC/JU
Sample : PB164444BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Oct 28 11:23:23 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations APPROVED
Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.204	196	247409	44.627	ng		98
46) 1,1'-Biphenyl	9.351	154	859069	43.865	ng		99
47) 2-Chloronaphthalene	9.375	162	677574	42.957	ng		100
48) 2-Nitroaniline	9.469	65	236689	49.063	ng		95
49) Acenaphthylene	9.792	152	1088844	47.749	ng		100
50) Dimethylphthalate	9.651	163	804892	45.933	ng		100
51) 2,6-Dinitrotoluene	9.710	165	179884	47.412	ng		99
52) Acenaphthene	9.963	154	735816	49.881	ng		99
53) 3-Nitroaniline	9.875	138	109911	28.091	ng		98
54) 2,4-Dinitrophenol	9.986	184	177437	109.904	ng	#	1
55) Dibenzofuran	10.133	168	947573	44.771	ng		100
56) 4-Nitrophenol	10.033	139	286527	95.186	ng		96
57) 2,4-Dinitrotoluene	10.116	165	244414	52.385	ng		99
58) Fluorene	10.480	166	727905	45.154	ng		99
59) 2,3,4,6-Tetrachlorophenol	10.251	232	212306	48.851	ng		97
60) Diethylphthalate	10.351	149	791021	45.976	ng		100
61) 4-Chlorophenyl-phenyle...	10.469	204	369072	45.336	ng		99
62) 4-Nitroaniline	10.498	138	180917	48.958	ng		97
63) Azobenzene	10.627	77	806460	44.214	ng		98
65) 4,6-Dinitro-2-methylph...	10.522	198	126328	60.564	ng		95
66) n-Nitrosodiphenylamine	10.586	169	680576	44.467	ng		100
67) 4-Bromophenyl-phenylether	10.957	248	237977	45.173	ng		98
68) Hexachlorobenzene	11.027	284	266322	44.950	ng		96
69) Atrazine	11.116	200	224215	54.080	ng		99
70) Pentachlorophenol	11.216	266	322659	89.731	ng		99
71) Phenanthrene	11.439	178	1074904	44.500	ng		100
72) Anthracene	11.492	178	1107771	47.004	ng		100
73) Carbazole	11.645	167	952915	43.475	ng		99
74) Di-n-butylphthalate	11.974	149	1167048	46.086	ng		99
75) Fluoranthene	12.627	202	1073772	43.973	ng		99
77) Benzidine	12.745	184	237729	55.056	ng		99
78) Pyrene	12.857	202	1070018	46.552	ng		100
80) Butylbenzylphthalate	13.468	149	358899	52.081	ng		99
81) Benzo(a)anthracene	14.045	228	824539	48.236	ng		100
82) 3,3'-Dichlorobenzidine	14.004	252	161964	32.497	ng		99
83) Chrysene	14.080	228	734278	46.850	ng		100
84) Bis(2-ethylhexyl)phtha...	14.027	149	433668	56.091	ng		100
85) Di-n-octyl phthalate	14.639	149	699522	49.743	ng		97
87) Indeno(1,2,3-cd)pyrene	17.033	276	839150	49.096	ng		98
88) Benzo(b)fluoranthene	15.098	252	705155	43.570	ng		99
89) Benzo(k)fluoranthene	15.127	252	711830	50.999	ng		99
90) Benzo(a)pyrene	15.468	252	674445	50.645	ng		99
91) Dibenzo(a,h)anthracene	17.051	278	679796	47.676	ng		98
92) Benzo(g,h,i)perylene	17.486	276	644416	45.246	ng		98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140077.D
Acq On : 28 Oct 2024 10:29
Operator : RC/JU
Sample : PB164444BS
Misc :
ALS Vial : 4 Sample Multiplier: 1


Quant Time: Oct 28 11:23:23 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.887 min Scan# 81
Delta R.T. -0.005 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:152 Resp: 12786


Ion	Ratio	Lower	Upper
152	100		
150	159.6	130.2	195.2
115	65.7	51.4	77.2



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#2
1,4-Dioxane
Concen: 35.657 ng
RT: 2.781 min Scan# 117
Delta R.T. 0.094 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion: 88 Resp: 135793

Ion	Ratio	Lower	Upper
88	100		
58	88.2	73.3	109.9
43	32.6	27.1	40.7





#3
Pyridine
Concen: 38.624 ng
RT: 3.534 min Scan# 24
Delta R.T. 0.088 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29


Tgt Ion: 79 Resp: 364598


Ion	Ratio	Lower	Upper
79	100		
52	82.9	68.5	102.7
51	40.0	32.3	48.5

Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#4
n-Nitrosodimethylamine
Concen: 41.494 ng
RT: 3.481 min Scan# 236
Delta R.T. 0.082 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion: 42 Resp: 210444

Ion	Ratio	Lower	Upper
42	100		
74	119.2	96.2	144.4
44	9.6	6.2	9.2#






#5
2-Fluorophenol
Concen: 120.410 ng
RT: 5.516 min Scan# 58
Delta R.T. 0.018 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:112 Resp: 98349


Ion	Ratio	Lower	Upper
112	100		
64	65.8	53.0	79.6
63	34.5	27.0	40.4



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#6
Aniline
Concen: 38.741 ng
RT: 6.551 min Scan# 758
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion: 93 Resp: 381952

Ion	Ratio	Lower	Upper
93	100		
66	42.5	34.0	51.0
65	21.6	17.4	26.0






#7
Phenol-d6
Concen: 118.025 ng
RT: 6.516 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion: 99 Resp: 124855


Ion	Ratio	Lower	Upper
99	100		
42	20.5	16.7	25.1
71	36.3	27.7	41.5



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#8
2-Chlorophenol
Concen: 45.570 ng
RT: 6.675 min Scan# 779
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:128 Resp: 380511

Ion	Ratio	Lower	Upper
128	100		
130	32.8	12.3	52.3
64	52.7	34.6	74.6






#9
Benzaldehyde
Concen: 12.611 ng
RT: 6.440 min Scan# 7
Delta R.T. 0.001 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion: 77 Resp: 7504


Ion	Ratio	Lower	Upper
77	100		
105	99.9	75.5	115.5
106	94.6	71.7	111.7



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#10
Phenol
Concen: 44.055 ng m
RT: 6.528 min Scan# 754
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion: 94 Resp: 480470

Ion	Ratio	Lower	Upper
94	100		
65	32.0	11.1	51.1
66	47.1	22.7	62.7





#11
bis(2-Chloroethyl)ether
Concen: 43.843 ng
RT: 6.622 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29


Tgt Ion: 93 Resp: 36958


Ion	Ratio	Lower	Upper
93	100		
63	86.2	65.0	105.0
95	32.7	12.4	52.4

Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#12
1,3-Dichlorobenzene
Concen: 42.706 ng
RT: 6.834 min Scan# 806
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:146 Resp: 409351

Ion	Ratio	Lower	Upper
146	100		
148	63.5	51.0	76.4
75	32.1	26.4	39.6






#13
1,4-Dichlorobenzene
Concen: 43.700 ng
RT: 6.910 min Scan# 81
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:146 Resp: 41848


Ion	Ratio	Lower	Upper
146	100		
148	64.9	51.8	77.8
111	42.4	34.6	51.8



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#14
1,2-Dichlorobenzene
Concen: 44.744 ng
RT: 7.057 min Scan# 844
Delta R.T. -0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:146 Resp: 399902

Ion	Ratio	Lower	Upper
146	100		
148	62.6	51.4	77.2
111	45.4	36.7	55.1





#15
Benzyl Alcohol
Concen: 46.179 ng
RT: 7.028 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29


Tgt Ion: 79 Resp: 35835


Ion	Ratio	Lower	Upper
79	100		
108	75.1	61.4	92.0
77	62.3	49.9	74.9

Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#16
2,2'-oxybis(1-Chloropropane)
Concen: 43.177 ng
RT: 7.163 min Scan# 862
Delta R.T. -0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion: 45 Resp: 620620

Ion	Ratio	Lower	Upper
45	100		
77	12.6	0.0	31.5
79	9.7	0.0	29.1






#17
2-Methylphenol
Concen: 44.379 ng
RT: 7.139 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:107 Resp: 31598


Ion	Ratio	Lower	Upper
107	100		
108	113.5	91.3	136.9
77	49.4	37.1	55.7
79	48.3	37.0	55.6



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#18
Hexachloroethane
Concen: 44.169 ng
RT: 7.404 min Scan# 903
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:117 Resp: 150834

Ion	Ratio	Lower	Upper
117	100		
119	95.0	77.0	115.4
201	102.8	76.2	114.4






#19
n-Nitroso-di-n-propylamine
Concen: 43.903 ng
RT: 7.304 min Scan# 88
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion: 70 Resp: 28168


Ion	Ratio	Lower	Upper
70	100		
42	61.2	51.5	77.3
101	9.6	7.5	11.3
130	20.3	15.5	23.3



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#20
3+4-Methylphenols
Concen: 42.951 ng
RT: 7.292 min Scan# 884
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:107 Resp: 391158

Ion	Ratio	Lower	Upper
107	100		
108	91.5	69.8	109.8
77	118.6	50.5	90.5#
79	31.5	10.2	50.2






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.175 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:136 Resp: 49528


Ion	Ratio	Lower	Upper
136	100		
137	10.5	8.6	12.8
54	9.6	8.4	12.6
68	6.1	5.1	7.7



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#22
Acetophenone
Concen: 43.883 ng
RT: 7.298 min Scan# 885
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:105 Resp: 532350

Ion	Ratio	Lower	Upper
105	100		
71	7.9	5.9	8.9
51	36.6	29.4	44.2
120	21.5	17.5	26.3





#23
Nitrobenzene-d5
Concen: 92.589 ng
RT: 7.451 min Scan# 91
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29


Tgt Ion: 82 Resp: 827399


Ion	Ratio	Lower	Upper
82	100		
128	42.3	33.4	50.0
54	57.7	47.8	71.8

Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#24
Nitrobenzene
Concen: 43.461 ng
RT: 7.469 min Scan# 914
Delta R.T. -0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion: 77 Resp: 425814

Ion	Ratio	Lower	Upper
77	100		
123	44.2	35.7	53.5
65	14.5	11.5	17.3






#25
Isophorone
Concen: 45.583 ng
RT: 7.710 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion: 82 Resp: 77314


Ion	Ratio	Lower	Upper
82	100		
95	7.1	5.5	8.3
138	16.3	13.2	19.8



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#26
2-Nitrophenol
Concen: 51.101 ng
RT: 7.786 min Scan# 968
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:139 Resp: 188880

Ion	Ratio	Lower	Upper
139	100		
109	31.2	25.1	37.7
65	57.5	47.5	71.3






#27
2,4-Dimethylphenol
Concen: 52.015 ng
RT: 7.822 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:122 Resp: 32058


Ion	Ratio	Lower	Upper
122	100		
107	115.7	91.9	137.9
121	57.3	46.2	69.4



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#28
bis(2-Chloroethoxy)methane
Concen: 44.384 ng
RT: 7.916 min Scan# 990
Delta R.T. -0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion: 93 Resp: 456094

Ion	Ratio	Lower	Upper
93	100		
95	32.7	26.0	39.0
123	11.9	9.8	14.8






#29
2,4-Dichlorophenol
Concen: 45.395 ng
RT: 8.028 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:162 Resp: 31868


Ion	Ratio	Lower	Upper
162	100		
164	65.2	43.9	83.9
98	38.3	19.9	59.9



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#30
1,2,4-Trichlorobenzene
Concen: 43.256 ng
RT: 8.110 min Scan# 1023
Delta R.T. -0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:180 Resp: 336007

Ion	Ratio	Lower	Upper
180	100		
182	98.4	80.5	120.7
145	32.4	25.4	38.0






#31
Naphthalene
Concen: 44.121 ng
RT: 8.192 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:128 Resp: 112700


Ion	Ratio	Lower	Upper
128	100		
129	11.3	8.9	13.3
127	13.0	10.6	16.0



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#32
Benzoic acid
Concen: 40.344 ng
RT: 7.939 min Scan# 994
Delta R.T. 0.017 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:122 Resp: 216873

Ion	Ratio	Lower	Upper
122	100		
105	126.8	107.6	147.6
77	99.3	80.2	120.2






#33
4-Chloroaniline
Concen: 18.292 ng
RT: 8.239 min Scan# 10
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:127 Resp: 15932


Ion	Ratio	Lower	Upper
127	100		
129	32.5	25.4	38.2
65	32.9	28.4	42.6
92	20.3	16.6	25.0



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#34
Hexachlorobutadiene
Concen: 44.923 ng
RT: 8.310 min Scan# 1057
Delta R.T. -0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:225 Resp: 219251

Ion	Ratio	Lower	Upper
225	100		
223	62.2	50.6	76.0
227	63.2	51.8	77.8






#35
Caprolactam
Concen: 46.244 ng m
RT: 8.610 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:113 Resp: 10322


Ion	Ratio	Lower	Upper
113	100		
55	204.6	189.2	229.2
56	154.7	132.5	172.5



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#36
4-Chloro-3-methylphenol
Concen: 45.972 ng
RT: 8.716 min Scan# 1126
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:107 Resp: 357353

Ion	Ratio	Lower	Upper
107	100		
144	24.0	19.1	28.7
142	71.2	57.2	85.8





#37
2-Methylnaphthalene
Concen: 45.632 ng
RT: 8.886 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29


Tgt Ion:142 Resp: 71386


Ion	Ratio	Lower	Upper
142	100		
141	89.1	70.8	106.2

Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#38
1-Methylnaphthalene
Concen: 43.622 ng
RT: 8.986 min Scan# 1172
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:142 Resp: 669536

Ion	Ratio	Lower	Upper
142	100		
141	91.4	73.5	110.3






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.928 min Scan# 1
Delta R.T. 0.001 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:164 Resp: 28132


Ion	Ratio	Lower	Upper
164	100		
162	100.9	81.0	121.4
160	44.3	35.4	53.0



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#40
1,2,4,5-Tetrachlorobenzene
Concen: 44.663 ng
RT: 9.051 min Scan# 1183
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:216 Resp: 338977

Ion	Ratio	Lower	Upper
216	100		
214	79.1	63.4	95.0
179	21.1	17.2	25.8
108	21.4	15.6	23.4





#41
Hexachlorocyclopentadiene
Concen: 165.576 ng
RT: 9.039 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29


Tgt Ion:237 Resp: 437259


Ion	Ratio	Lower	Upper
237	100		
235	62.5	42.9	82.9
272	13.0	0.0	32.3

Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#42
2,4,6-Tribromophenol
Concen: 141.398 ng
RT: 10.722 min Scan# 1467
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:330 Resp: 372077

Ion	Ratio	Lower	Upper
330	100		
332	96.2	78.1	117.1
141	30.1	26.6	39.8






#43
2,4,6-Trichlorophenol
Concen: 45.915 ng
RT: 9.163 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:196 Resp: 24672


Ion	Ratio	Lower	Upper
196	100		
198	96.8	76.2	114.2
200	30.1	24.0	36.0



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#44
2,4,5-Trichlorophenol
Concen: 44.627 ng
RT: 9.204 min Scan# 1209
Delta R.T. 0.012 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:196 Resp: 247409

Ion	Ratio	Lower	Upper
196	100		
198	96.3	75.8	113.6
97	50.5	41.8	62.6
132	27.3	22.9	34.3





#45
2-Fluorobiphenyl
Concen: 87.245 ng
RT: 9.245 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29


Tgt Ion:172 Resp: 148508


Ion	Ratio	Lower	Upper
172	100		
171	36.4	28.6	43.0
170	23.9	19.1	28.7

Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#46
1,1'-Biphenyl
Concen: 43.865 ng
RT: 9.351 min Scan# 1234
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:154 Resp: 859069

Ion	Ratio	Lower	Upper
154	100		
153	40.9	20.3	60.3
76	13.1	0.0	33.9






#47
2-Chloronaphthalene
Concen: 42.957 ng
RT: 9.375 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:162 Resp: 67757


Ion	Ratio	Lower	Upper
162	100		
127	42.6	33.8	50.6
164	33.2	26.6	39.8



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#48
2-Nitroaniline
Concen: 49.063 ng
RT: 9.469 min Scan# 1254
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion: 65 Resp: 236689

Ion	Ratio	Lower	Upper
65	100		
92	66.1	50.9	76.3
138	97.9	73.4	110.0






#49
Acenaphthylene
Concen: 47.749 ng
RT: 9.792 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:152 Resp: 108884


Ion	Ratio	Lower	Upper
152	100		
151	20.0	15.9	23.9
153	13.2	10.7	16.1



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#50
Dimethylphthalate
Concen: 45.933 ng
RT: 9.651 min Scan# 1285
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:163 Resp: 804892

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.4	5.0
164	10.1	8.2	12.2






#51
2,6-Dinitrotoluene
Concen: 47.412 ng
RT: 9.710 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:165 Resp: 17988


Ion	Ratio	Lower	Upper
165	100		
63	62.9	50.6	75.8
89	58.5	47.5	71.3



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#52
Acenaphthene
Concen: 49.881 ng
RT: 9.963 min Scan# 1338
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:154 Resp: 735816

Ion	Ratio	Lower	Upper
154	100		
153	111.6	89.8	134.6
152	54.0	42.4	63.6





#53
3-Nitroaniline
Concen: 28.091 ng
RT: 9.875 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29


Tgt Ion:138 Resp: 10991


Ion	Ratio	Lower	Upper
138	100		
108	11.2	10.2	15.2
92	121.4	98.3	147.5

Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#54
2,4-Dinitrophenol
Concen: 109.904 ng
RT: 9.986 min Scan# 1342
Delta R.T. 0.012 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:184 Resp: 177437

Ion	Ratio	Lower	Upper
184	100		
63	73.3	84.7	127.1#
154	64.0	338.2	507.2#






#55
Dibenzofuran
Concen: 44.771 ng
RT: 10.133 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:168 Resp: 94757


Ion	Ratio	Lower	Upper
168	100		
139	40.0	31.8	47.6
169	14.3	11.6	17.4



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#56
4-Nitrophenol
Concen: 95.186 ng
RT: 10.033 min Scan# 1350
Delta R.T. 0.012 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:139 Resp: 286527

Ion	Ratio	Lower	Upper
139	100		
109	76.9	53.6	93.6
65	112.6	87.8	127.8






#57
2,4-Dinitrotoluene
Concen: 52.385 ng
RT: 10.116 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:165 Resp: 24441


Ion	Ratio	Lower	Upper
165	100		
63	49.8	40.1	60.1
89	73.8	60.4	90.6
182	2.8	2.0	3.0



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#58
Fluorene
Concen: 45.154 ng
RT: 10.480 min Scan# 1426
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:166 Resp: 727905

Ion	Ratio	Lower	Upper
166	100		
165	97.4	78.5	117.7
167	13.2	10.9	16.3






#59
2,3,4,6-Tetrachlorophenol
Concen: 48.851 ng
RT: 10.251 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:232 Resp: 212306


Ion	Ratio	Lower	Upper
232	100		
131	44.5	37.0	55.6
130	2.3	1.9	2.9
166	28.9	24.2	36.2



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#60
Diethylphthalate
Concen: 45.976 ng
RT: 10.351 min Scan# 1404
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:149 Resp: 791021

Ion	Ratio	Lower	Upper
149	100		
177	22.2	17.9	26.9
150	12.7	10.1	15.1






#61
4-Chlorophenyl-phenylether
Concen: 45.336 ng
RT: 10.469 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:204 Resp: 369071


Ion	Ratio	Lower	Upper
204	100		
206	32.5	26.3	39.5
141	60.0	48.6	73.0



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#62
4-Nitroaniline
Concen: 48.958 ng
RT: 10.498 min Scan# 1429
Delta R.T. 0.012 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:138 Resp: 180917

Ion	Ratio	Lower	Upper
138	100		
92	51.3	35.7	75.7
108	77.4	57.5	97.5






#63
Azobenzene
Concen: 44.214 ng
RT: 10.627 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion: 77 Resp: 80646


Ion	Ratio	Lower	Upper
77	100		
182	24.9	3.3	43.3
105	21.4	1.2	41.2
51	32.5	13.3	53.3



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.416 min Scan# 1585
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:188 Resp: 511374

Ion	Ratio	Lower	Upper
188	100		
94	9.4	7.9	11.9
80	10.3	9.0	13.4





#65
4,6-Dinitro-2-methylphenol
Concen: 60.564 ng
RT: 10.522 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29


Tgt Ion:198 Resp: 12632


Ion	Ratio	Lower	Upper
198	100		
51	52.4	38.2	78.2
105	46.1	27.9	67.9

Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#66
n-Nitrosodiphenylamine
Concen: 44.467 ng
RT: 10.586 min Scan# 1444
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:169 Resp: 680576

Ion	Ratio	Lower	Upper
169	100		
168	62.5	49.8	74.6
167	34.8	28.0	42.0





#67
4-Bromophenyl-phenylether
Concen: 45.173 ng
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29


Instrument :
BNA_F
ClientSampleId :
PB164444BS


Tgt Ion:248 Resp: 23797

Ion	Ratio	Lower	Upper
248	100		
250	97.6	79.0	118.6
141	73.9	61.7	92.5

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#68
Hexachlorobenzene
Concen: 44.950 ng
RT: 11.027 min Scan# 1519
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:284 Resp: 266322

Ion	Ratio	Lower	Upper
284	100		
142	29.5	26.9	40.3
249	31.2	25.3	37.9






#69
Atrazine
Concen: 54.080 ng
RT: 11.116 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:200 Resp: 22421


Ion	Ratio	Lower	Upper
200	100		
173	26.7	7.7	47.7
215	47.3	27.7	67.7



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#70
Pentachlorophenol
Concen: 89.731 ng
RT: 11.216 min Scan# 1551
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:266 Resp: 322659

Ion	Ratio	Lower	Upper
266	100		
268	62.7	50.4	75.6
264	62.1	50.6	75.8





#71
Phenanthrene
Concen: 44.500 ng
RT: 11.439 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29


Tgt Ion:178 Resp: 107490


Ion	Ratio	Lower	Upper
178	100		
176	19.9	15.8	23.6
179	15.8	12.6	18.8

Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#72
Anthracene
Concen: 47.004 ng
RT: 11.492 min Scan# 1598
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:178 Resp: 1107771

Ion	Ratio	Lower	Upper
178	100		
176	19.2	15.3	22.9
179	15.6	12.4	18.6






#73
Carbazole
Concen: 43.475 ng
RT: 11.645 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:167 Resp: 95291


Ion	Ratio	Lower	Upper
167	100		
166	21.6	17.0	25.6
139	13.2	10.8	16.2



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#74
Di-n-butylphthalate
Concen: 46.086 ng
RT: 11.974 min Scan# 1680
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:149 Resp: 1167048

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.5	11.3
104	5.0	4.1	6.1






#75
Fluoranthene
Concen: 43.973 ng
RT: 12.627 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:202 Resp: 107377


Ion	Ratio	Lower	Upper
202	100		
101	11.3	0.0	31.9
203	17.4	0.0	37.5



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.051 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:240 Resp: 262593

Ion	Ratio	Lower	Upper
240	100		
120	11.1	9.4	14.2
236	26.5	20.9	31.3





#77
Benzidine
Concen: 55.056 ng
RT: 12.745 min Scan# 18
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29


Tgt Ion:184 Resp: 237729


Ion	Ratio	Lower	Upper
184	100		
185	14.9	12.0	18.0
183	11.4	9.4	14.2

Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#78
Pyrene
Concen: 46.552 ng
RT: 12.857 min Scan# 1830
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:202 Resp: 1070018

Ion	Ratio	Lower	Upper
202	100		
200	21.8	17.2	25.8
203	17.7	14.2	21.4






#79
Terphenyl-d14
Concen: 95.744 ng
RT: 12.998 min Scan# 18
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:244 Resp: 154292


Ion	Ratio	Lower	Upper
244	100		
212	6.8	5.7	8.5
122	10.1	8.6	13.0



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#80
Butylbenzylphthalate
Concen: 52.081 ng
RT: 13.468 min Scan# 1934
Delta R.T. -0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:149 Resp: 358899

Ion	Ratio	Lower	Upper
149	100		
91	74.9	59.4	89.2
206	20.0	16.4	24.6






#81
Benzo(a)anthracene
Concen: 48.236 ng
RT: 14.045 min Scan# 20
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:228 Resp: 824539


Ion	Ratio	Lower	Upper
228	100		
226	26.8	21.6	32.4
229	19.8	16.1	24.1



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#82
3,3'-Dichlorobenzidine
Concen: 32.497 ng
RT: 14.004 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:252 Resp: 161964

Ion	Ratio	Lower	Upper
252	100		
254	64.2	50.8	76.2
126	10.5	8.9	13.3






#83
Chrysene
Concen: 46.850 ng
RT: 14.080 min Scan# 20
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:228 Resp: 734278


Ion	Ratio	Lower	Upper
228	100		
226	29.9	24.1	36.1
229	19.8	15.8	23.6



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#84
Bis(2-ethylhexyl)phthalate
Concen: 56.091 ng
RT: 14.027 min Scan# 2029
Delta R.T. -0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:149 Resp: 433668

Ion	Ratio	Lower	Upper
149	100		
167	26.6	21.4	32.0
279	4.0	3.0	4.4





#85
Di-n-octyl phthalate
Concen: 49.743 ng
RT: 14.639 min Scan# 21
Delta R.T. -0.012 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29


Tgt Ion:149 Resp: 69952


Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.2	1.8
43	11.0	9.8	14.6

Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.527 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:264 Resp: 265684

Ion	Ratio	Lower	Upper
264	100		
260	24.0	19.4	29.2
265	21.8	17.4	26.0






#87
Indeno(1,2,3-cd)pyrene
Concen: 49.096 ng
RT: 17.033 min Scan# 2
Delta R.T. 0.012 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:276 Resp: 83915


Ion	Ratio	Lower	Upper
276	100		
138	23.3	20.4	30.6
277	25.3	20.2	30.4



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#88
Benzo(b)fluoranthene
Concen: 43.570 ng
RT: 15.098 min Scan# 2211
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:252 Resp: 705155

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.6	26.4
125	9.4	7.9	11.9





#89
Benzo(k)fluoranthene
Concen: 50.999 ng
RT: 15.127 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29


Tgt Ion:252 Resp: 71183


Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.6	26.4
125	9.1	7.9	11.9

Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#90
Benzo(a)pyrene
Concen: 50.645 ng
RT: 15.468 min Scan# 2274
Delta R.T. 0.000 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:252 Resp: 674445

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.3	25.9
125	10.7	8.7	13.1






#91
Dibenzo(a,h)anthracene
Concen: 47.676 ng
RT: 17.051 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:278 Resp: 67979


Ion	Ratio	Lower	Upper
278	100		
139	14.9	13.1	19.7
279	24.0	19.0	28.4



Instrument :
BNA_F
ClientSampleId :
PB164444BS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#92
Benzo(g,h,i)perylene
Concen: 45.246 ng
RT: 17.486 min Scan# 2617
Delta R.T. 0.006 min
Lab File: BF140077.D
Acq: 28 Oct 2024 10:29

Tgt Ion:276 Resp: 644416

Ion	Ratio	Lower	Upper
276	100		
277	23.7	19.0	28.6
138	20.0	17.4	26.0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	10/24/24
Project:	Amtrak Sawtooth Bridges 2024	Date Received:	10/24/24
Client Sample ID:	VNJ-215MS	SDG No.:	P4578
Lab Sample ID:	P4545-02MS	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140092.D	1	10/26/24 12:10	10/28/24 17:35	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	230		15.5	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	430		8.40	50.0	ug/L
95-48-7	2-Methylphenol	490		11.3	50.0	ug/L
65794-96-9	3+4-Methylphenols	480		11.5	100	ug/L
67-72-1	Hexachloroethane	420		10.1	50.0	ug/L
98-95-3	Nitrobenzene	470		12.7	50.0	ug/L
87-68-3	Hexachlorobutadiene	460		12.7	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	550		8.90	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	540		10.1	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	600		15.2	50.0	ug/L
118-74-1	Hexachlorobenzene	540		11.4	50.0	ug/L
87-86-5	Pentachlorophenol	780		18.5	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	135		15 (10) - 110 (139)	90%	SPK: 150
13127-88-3	Phenol-d6	129		15 (10) - 110 (134)	86%	SPK: 150
4165-60-0	Nitrobenzene-d5	103		30 (49) - 130 (133)	103%	SPK: 100
321-60-8	2-Fluorobiphenyl	98.8		30 (52) - 130 (132)	99%	SPK: 100
118-79-6	2,4,6-Tribromophenol	159		15 (44) - 110 (137)	106%	SPK: 150
1718-51-0	Terphenyl-d14	105		30 (48) - 130 (125)	105%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	114000	6.892			
1146-65-2	Naphthalene-d8	442000	8.169			
15067-26-2	Acenaphthene-d10	242000	9.928			
1517-22-2	Phenanthrene-d10	410000	11.41			
1719-03-5	Chrysene-d12	203000	14.051			
1520-96-3	Perylene-d12	263000	15.527			

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture			Date Collected:	10/24/24
Project:	Amtrak Sawtooth Bridges 2024			Date Received:	10/24/24
Client Sample ID:	VNJ-215MS			SDG No.:	P4578
Lab Sample ID:	P4545-02MS			Matrix:	TCLP
Analytical Method:	SW8270			% Solid:	0
Sample Wt/Vol:	100	Units:	mL	Final Vol:	1000 uL
Soil Aliquot Vol:			uL	Test:	TCLP BNA
Extraction Type :			Decanted : N	Level :	LOW
Injection Volume :			GPC Factor : 1.0	GPC Cleanup : N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140092.D	1	10/26/24 12:10	10/28/24 17:35	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140092.D
Acq On : 28 Oct 2024 17:35
Operator : RC/JU
Sample : P4545-02MS
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations APPROVED
Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

Quant Time: Oct 28 18:05:02 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.892	152	113612	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	441889	20.000	ng	0.00
39) Acenaphthene-d10	9.928	164	241647	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	409717	20.000	ng	0.00
76) Chrysene-d12	14.051	240	203129	20.000	ng	0.00
86) Perylene-d12	15.527	264	263019	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.522	112	977769	134.729	ng	0.02
7) Phenol-d6	6.516	99	1208635	128.587	ng	0.00
23) Nitrobenzene-d5	7.451	82	821761	103.069	ng	0.00
42) 2,4,6-Tribromophenol	10.716	330	358712	158.702	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1444688	98.807	ng	0.00
79) Terphenyl-d14	12.998	244	1306207	104.783	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.881	88	126476	37.378	ng	# 82
3) Pyridine	3.587	79	189602	22.606	ng	98
4) n-Nitrosodimethylamine	3.516	42	187502	41.610	ng	# 97
6) Aniline	6.557	93	172468	19.688	ng	99
8) 2-Chlorophenol	6.675	128	370207	49.899	ng	99
9) Benzaldehyde	6.440	77	47671	9.016	ng	99
10) Phenol	6.528	94	445105m	45.933	ng	
11) bis(2-Chloroethyl)ether	6.628	93	351685	46.955	ng	99
12) 1,3-Dichlorobenzene	6.834	146	361312	42.425	ng	99
13) 1,4-Dichlorobenzene	6.910	146	367844	43.232	ng	99
14) 1,2-Dichlorobenzene	7.057	146	355662	44.788	ng	99
15) Benzyl Alcohol	7.028	79	347790	50.442	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.163	45	584140	45.739	ng	97
17) 2-Methylphenol	7.134	107	310465	49.075	ng	99
18) Hexachloroethane	7.404	117	128123	42.227	ng	96
19) n-Nitroso-di-n-propyla...	7.304	70	276089	48.431	ng	95
20) 3+4-Methylphenols	7.287	107	391991	48.443	ng	94
22) Acetophenone	7.298	105	508513	46.983	ng	98
24) Nitrobenzene	7.469	77	408928	46.780	ng	100
25) Isophorone	7.710	82	765851	50.609	ng	99
26) 2-Nitrophenol	7.787	139	189690	57.521	ng	97
27) 2,4-Dimethylphenol	7.816	122	314864	57.260	ng	98
28) bis(2-Chloroethoxy)met...	7.916	93	447113	48.767	ng	99
29) 2,4-Dichlorophenol	8.022	162	320220	51.126	ng	100
30) 1,2,4-Trichlorobenzene	8.110	180	316171	45.620	ng	100
31) Naphthalene	8.192	128	1066174	46.783	ng	100
32) Benzoic acid	7.898	122	84400	17.598	ng	95
33) 4-Chloroaniline	8.234	127	133599	17.192	ng	98
34) Hexachlorobutadiene	8.310	225	199456	45.805	ng	99
35) Caprolactam	8.604	113	83595m	41.975	ng	
36) 4-Chloro-3-methylphenol	8.716	107	363875	52.467	ng	99
37) 2-Methylnaphthalene	8.881	142	696991	49.937	ng	100
38) 1-Methylnaphthalene	8.981	142	647449	47.280	ng	99
40) 1,2,4,5-Tetrachloroben...	9.051	216	331367	50.829	ng	99
41) Hexachlorocyclopentadiene	9.034	237	362304	159.719	ng	100
43) 2,4,6-Trichlorophenol	9.157	196	253954	55.021	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140092.D
Acq On : 28 Oct 2024 17:35
Operator : RC/JU
Sample : P4545-02MS
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations APPROVED
Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

Quant Time: Oct 28 18:05:02 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.204	196	258939	54.376	ng		98
46) 1,1'-Biphenyl	9.345	154	856147	50.894	ng		100
47) 2-Chloronaphthalene	9.375	162	675337	49.845	ng		100
48) 2-Nitroaniline	9.463	65	241309	58.234	ng		97
49) Acenaphthylene	9.786	152	1080746	55.175	ng		99
50) Dimethylphthalate	9.645	163	822973	54.677	ng		99
51) 2,6-Dinitrotoluene	9.710	165	178715	54.837	ng		96
52) Acenaphthene	9.963	154	661593	52.213	ng		99
53) 3-Nitroaniline	9.875	138	131503	39.128	ng		98
54) 2,4-Dinitrophenol	9.980	184	60419	47.508	ng	#	1
55) Dibenzofuran	10.133	168	947629	52.125	ng		100
56) 4-Nitrophenol	10.033	139	254819	98.551	ng		98
57) 2,4-Dinitrotoluene	10.110	165	238568	59.527	ng		99
58) Fluorene	10.475	166	716535	51.747	ng		99
59) 2,3,4,6-Tetrachlorophenol	10.245	232	201940	54.095	ng		98
60) Diethylphthalate	10.345	149	792061	53.595	ng		99
61) 4-Chlorophenyl-phenyle...	10.469	204	368778	52.737	ng		98
62) 4-Nitroaniline	10.492	138	167633	52.812	ng		98
63) Azobenzene	10.628	77	807556	51.543	ng		97
65) 4,6-Dinitro-2-methylph...	10.516	198	60965	36.479	ng		94
66) n-Nitrosodiphenylamine	10.586	169	670666	54.692	ng		100
67) 4-Bromophenyl-phenylether	10.957	248	233236	55.258	ng		97
68) Hexachlorobenzene	11.022	284	255815	53.889	ng		98
69) Atrazine	11.110	200	212035	63.831	ng		99
70) Pentachlorophenol	11.216	266	224019	77.757	ng		99
71) Phenanthrene	11.439	178	1031591	53.303	ng		100
72) Anthracene	11.492	178	1041472	55.155	ng		100
73) Carbazole	11.639	167	882304	50.241	ng		100
74) Di-n-butylphthalate	11.969	149	1119955	55.199	ng		100
75) Fluoranthene	12.627	202	912791	46.655	ng		99
77) Benzidine	12.745	184	17706	5.301	ng		98
78) Pyrene	12.851	202	903442	50.811	ng		100
80) Butylbenzylphthalate	13.469	149	345144	64.747	ng		99
81) Benzo(a)anthracene	14.039	228	724089	54.760	ng		100
82) 3,3'-Dichlorobenzidine	14.004	252	185837	48.202	ng		100
83) Chrysene	14.080	228	701255	57.841	ng		99
84) Bis(2-ethylhexyl)phtha...	14.027	149	456891	76.394	ng		100
85) Di-n-octyl phthalate	14.639	149	793362	72.931	ng		98
87) Indeno(1,2,3-cd)pyrene	17.027	276	775966	45.859	ng		98
88) Benzo(b)fluoranthene	15.098	252	831851	51.919	ng		99
89) Benzo(k)fluoranthene	15.127	252	782283	56.615	ng		99
90) Benzo(a)pyrene	15.468	252	780503	59.203	ng		99
91) Dibenzo(a,h)anthracene	17.045	278	651067	46.124	ng		99
92) Benzo(g,h,i)perylene	17.474	276	529084	37.525	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140092.D
Acq On : 28 Oct 2024 17:35
Operator : RC/JU
Sample : P4545-02MS
Misc :
ALS Vial : 19 Sample Multiplier: 1


Quant Time: Oct 28 18:05:02 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.892 min Scan# 81
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35


Tgt Ion:152 Resp: 11361


Ion	Ratio	Lower	Upper
152	100		
150	164.2	130.2	195.2
115	63.0	51.4	77.2

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#2
1,4-Dioxane
Concen: 37.378 ng
RT: 2.881 min Scan# 134
Delta R.T. 0.194 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion: 88 Resp: 126476

Ion	Ratio	Lower	Upper
88	100		
58	89.6	73.3	109.9
43	0.0	27.1	40.7#






#3
Pyridine
Concen: 22.606 ng
RT: 3.587 min Scan# 2
Delta R.T. 0.141 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion: 79 Resp: 18960


Ion	Ratio	Lower	Upper
79	100		
52	83.4	68.5	102.7
51	40.2	32.3	48.5



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#4
n-Nitrosodimethylamine
Concen: 41.610 ng
RT: 3.516 min Scan# 242
Delta R.T. 0.118 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion: 42 Resp: 187502

Ion	Ratio	Lower	Upper
42	100		
74	123.8	96.2	144.4
44	11.0	6.2	9.2#





#5
2-Fluorophenol
Concen: 134.729 ng
RT: 5.522 min Scan# 58
Delta R.T. 0.024 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35


Tgt Ion:112 Resp: 977769


Ion	Ratio	Lower	Upper
112	100		
64	65.0	53.0	79.6
63	33.4	27.0	40.4

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#6
Aniline
Concen: 19.688 ng
RT: 6.557 min Scan# 759
Delta R.T. 0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion: 93 Resp: 172468

Ion	Ratio	Lower	Upper
93	100		
66	42.0	34.0	51.0
65	21.4	17.4	26.0






#7
Phenol-d6
Concen: 128.587 ng
RT: 6.516 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion: 99 Resp: 120863


Ion	Ratio	Lower	Upper
99	100		
42	19.9	16.7	25.1
71	35.8	27.7	41.5



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#8
2-Chlorophenol
Concen: 49.899 ng
RT: 6.675 min Scan# 779
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:128 Resp: 370207

Ion	Ratio	Lower	Upper
128	100		
130	32.8	12.3	52.3
64	53.5	34.6	74.6






#9
Benzaldehyde
Concen: 9.016 ng
RT: 6.440 min Scan# 7
Delta R.T. 0.001 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion: 77 Resp: 4767


Ion	Ratio	Lower	Upper
77	100		
105	95.5	75.5	115.5
106	92.9	71.7	111.7



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#10
Phenol
Concen: 45.933 ng m
RT: 6.528 min Scan# 754
Delta R.T. 0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion: 94 Resp: 445105

Ion	Ratio	Lower	Upper
94	100		
65	32.3	11.1	51.1
66	45.7	22.7	62.7





#11
bis(2-Chloroethyl)ether
Concen: 46.955 ng
RT: 6.628 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35


Tgt Ion: 93 Resp: 35168


Ion	Ratio	Lower	Upper
93	100		
63	84.3	65.0	105.0
95	32.5	12.4	52.4

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#12
1,3-Dichlorobenzene
Concen: 42.425 ng
RT: 6.834 min Scan# 806
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:146 Resp: 361312

Ion	Ratio	Lower	Upper
146	100		
148	64.1	51.0	76.4
75	31.4	26.4	39.6






#13
1,4-Dichlorobenzene
Concen: 43.232 ng
RT: 6.910 min Scan# 81
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:146 Resp: 36784


Ion	Ratio	Lower	Upper
146	100		
148	64.2	51.8	77.8
111	42.6	34.6	51.8



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#14
1,2-Dichlorobenzene
Concen: 44.788 ng
RT: 7.057 min Scan# 844
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:146 Resp: 355662

Ion	Ratio	Lower	Upper
146	100		
148	64.4	51.4	77.2
111	45.0	36.7	55.1





#15
Benzyl Alcohol
Concen: 50.442 ng
RT: 7.028 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35


Tgt Ion: 79 Resp: 34779


Ion	Ratio	Lower	Upper
79	100		
108	76.0	61.4	92.0
77	63.1	49.9	74.9

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#16
2,2'-oxybis(1-Chloropropane)
Concen: 45.739 ng
RT: 7.163 min Scan# 862
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion: 45 Resp: 584140

Ion	Ratio	Lower	Upper
45	100		
77	13.0	0.0	31.5
79	9.6	0.0	29.1






#17
2-Methylphenol
Concen: 49.075 ng
RT: 7.134 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:107 Resp: 31046


Ion	Ratio	Lower	Upper
107	100		
108	113.6	91.3	136.9
77	46.7	37.1	55.7
79	46.9	37.0	55.6



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#18
Hexachloroethane
Concen: 42.227 ng
RT: 7.404 min Scan# 903
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:117 Resp: 128123

Ion	Ratio	Lower	Upper
117	100		
119	96.3	77.0	115.4
201	103.7	76.2	114.4






#19
n-Nitroso-di-n-propylamine
Concen: 48.431 ng
RT: 7.304 min Scan# 88
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion: 70 Resp: 276089


Ion	Ratio	Lower	Upper
70	100		
42	59.4	51.5	77.3
101	9.7	7.5	11.3
130	20.3	15.5	23.3



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#20
3+4-Methylphenols
Concen: 48.443 ng
RT: 7.287 min Scan# 883
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:107 Resp: 391991

Ion	Ratio	Lower	Upper
107	100		
108	89.9	69.8	109.8
77	83.7	50.5	90.5
79	29.5	10.2	50.2






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.169 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:136 Resp: 44188


Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.6	12.8
54	9.9	8.4	12.6
68	6.4	5.1	7.7



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#22
Acetophenone
Concen: 46.983 ng
RT: 7.298 min Scan# 885
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:105 Resp: 508513

Ion	Ratio	Lower	Upper
105	100		
71	8.1	5.9	8.9
51	34.9	29.4	44.2
120	21.7	17.5	26.3





#23
Nitrobenzene-d5
Concen: 103.069 ng
RT: 7.451 min Scan# 91
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35


Tgt Ion: 82 Resp: 82176


Ion	Ratio	Lower	Upper
82	100		
128	42.9	33.4	50.0
54	56.6	47.8	71.8

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#24
Nitrobenzene
Concen: 46.780 ng
RT: 7.469 min Scan# 914
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion: 77 Resp: 408928

Ion	Ratio	Lower	Upper
77	100		
123	44.8	35.7	53.5
65	14.4	11.5	17.3





#25
Isophorone
Concen: 50.609 ng
RT: 7.710 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35


Tgt Ion: 82 Resp: 765851


Ion	Ratio	Lower	Upper
82	100		
95	7.2	5.5	8.3
138	16.8	13.2	19.8

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#26
2-Nitrophenol
Concen: 57.521 ng
RT: 7.787 min Scan# 968
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:139 Resp: 189690

Ion	Ratio	Lower	Upper
139	100		
109	30.8	25.1	37.7
65	56.2	47.5	71.3






#27
2,4-Dimethylphenol
Concen: 57.260 ng
RT: 7.816 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:122 Resp: 31486


Ion	Ratio	Lower	Upper
122	100		
107	117.9	91.9	137.9
121	58.7	46.2	69.4



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#28
bis(2-Chloroethoxy)methane
Concen: 48.767 ng
RT: 7.916 min Scan# 990
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion: 93 Resp: 447113

Ion	Ratio	Lower	Upper
93	100		
95	32.8	26.0	39.0
123	11.2	9.8	14.8






#29
2,4-Dichlorophenol
Concen: 51.126 ng
RT: 8.022 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:162 Resp: 32022


Ion	Ratio	Lower	Upper
162	100		
164	63.7	43.9	83.9
98	40.2	19.9	59.9



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#30
1,2,4-Trichlorobenzene
Concen: 45.620 ng
RT: 8.110 min Scan# 1023
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:180 Resp: 316171

Ion	Ratio	Lower	Upper
180	100		
182	100.3	80.5	120.7
145	32.3	25.4	38.0






#31
Naphthalene
Concen: 46.783 ng
RT: 8.192 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:128 Resp: 106617


Ion	Ratio	Lower	Upper
128	100		
129	11.3	8.9	13.3
127	13.2	10.6	16.0



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#32
Benzoic acid
Concen: 17.598 ng
RT: 7.898 min Scan# 987
Delta R.T. -0.024 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:122 Resp: 84400

Ion	Ratio	Lower	Upper
122	100		
105	122.6	107.6	147.6
77	94.4	80.2	120.2






#33
4-Chloroaniline
Concen: 17.192 ng
RT: 8.234 min Scan# 10
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:127 Resp: 133599


Ion	Ratio	Lower	Upper
127	100		
129	32.0	25.4	38.2
65	33.9	28.4	42.6
92	19.7	16.6	25.0



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#34
Hexachlorobutadiene
Concen: 45.805 ng
RT: 8.310 min Scan# 1057
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:225 Resp: 199456

Ion	Ratio	Lower	Upper
225	100		
223	62.7	50.6	76.0
227	63.7	51.8	77.8





#35
Caprolactam
Concen: 41.975 ng m
RT: 8.604 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35


Tgt Ion:113 Resp: 8359


Ion	Ratio	Lower	Upper
113	100		
55	199.8	189.2	229.2
56	147.9	132.5	172.5

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#36
4-Chloro-3-methylphenol
Concen: 52.467 ng
RT: 8.716 min Scan# 1126
Delta R.T. 0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:107 Resp: 363875

Ion	Ratio	Lower	Upper
107	100		
144	24.4	19.1	28.7
142	72.4	57.2	85.8





#37
2-Methylnaphthalene
Concen: 49.937 ng
RT: 8.881 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35


Tgt Ion:142 Resp: 69699


Ion	Ratio	Lower	Upper
142	100		
141	88.2	70.8	106.2

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#38
1-Methylnaphthalene
Concen: 47.280 ng
RT: 8.981 min Scan# 1171
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:142 Resp: 647449

Ion	Ratio	Lower	Upper
142	100		
141	92.6	73.5	110.3






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.928 min Scan# 1
Delta R.T. 0.001 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:164 Resp: 24164


Ion	Ratio	Lower	Upper
164	100		
162	101.1	81.0	121.4
160	44.2	35.4	53.0



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#40
1,2,4,5-Tetrachlorobenzene
Concen: 50.829 ng
RT: 9.051 min Scan# 1183
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:216 Resp: 331367

Ion	Ratio	Lower	Upper
216	100		
214	78.7	63.4	95.0
179	21.2	17.2	25.8
108	21.1	15.6	23.4






#41
Hexachlorocyclopentadiene
Concen: 159.719 ng
RT: 9.034 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:237 Resp: 362304


Ion	Ratio	Lower	Upper
237	100		
235	62.8	42.9	82.9
272	12.8	0.0	32.3



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#42
2,4,6-Tribromophenol
Concen: 158.702 ng
RT: 10.716 min Scan# 1466
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:330 Resp: 358712

Ion	Ratio	Lower	Upper
330	100		
332	96.9	78.1	117.1
141	29.1	26.6	39.8






#43
2,4,6-Trichlorophenol
Concen: 55.021 ng
RT: 9.157 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:196 Resp: 25395


Ion	Ratio	Lower	Upper
196	100		
198	93.7	76.2	114.2
200	29.8	24.0	36.0



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#44
2,4,5-Trichlorophenol
Concen: 54.376 ng
RT: 9.204 min Scan# 1209
Delta R.T. 0.012 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:196 Resp: 258939

Ion	Ratio	Lower	Upper
196	100		
198	94.2	75.8	113.6
97	49.0	41.8	62.6
132	26.5	22.9	34.3






#45
2-Fluorobiphenyl
Concen: 98.807 ng
RT: 9.245 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:172 Resp: 144468


Ion	Ratio	Lower	Upper
172	100		
171	35.9	28.6	43.0
170	23.9	19.1	28.7



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#46
1,1'-Biphenyl
Concen: 50.894 ng
RT: 9.345 min Scan# 1233
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:154 Resp: 856147

Ion	Ratio	Lower	Upper
154	100		
153	40.5	20.3	60.3
76	13.7	0.0	33.9






#47
2-Chloronaphthalene
Concen: 49.845 ng
RT: 9.375 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:162 Resp: 67533


Ion	Ratio	Lower	Upper
162	100		
127	42.3	33.8	50.6
164	33.1	26.6	39.8



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#48
2-Nitroaniline
Concen: 58.234 ng
RT: 9.463 min Scan# 1253
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion: 65 Resp: 241309

Ion	Ratio	Lower	Upper
65	100		
92	65.7	50.9	76.3
138	94.2	73.4	110.0






#49
Acenaphthylene
Concen: 55.175 ng
RT: 9.786 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:152 Resp: 108074


Ion	Ratio	Lower	Upper
152	100		
151	20.2	15.9	23.9
153	13.3	10.7	16.1



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#50
Dimethylphthalate
Concen: 54.677 ng
RT: 9.645 min Scan# 1284
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:163 Resp: 822973

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.4	5.0
164	10.5	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 54.837 ng
RT: 9.710 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35


Tgt Ion:165 Resp: 17871


Ion	Ratio	Lower	Upper
165	100		
63	60.0	50.6	75.8
89	56.2	47.5	71.3

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#52
Acenaphthene
Concen: 52.213 ng
RT: 9.963 min Scan# 1338
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:154 Resp: 661593

Ion	Ratio	Lower	Upper
154	100		
153	113.2	89.8	134.6
152	53.7	42.4	63.6





#53
3-Nitroaniline
Concen: 39.128 ng
RT: 9.875 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35


Tgt Ion:138 Resp: 13150


Ion	Ratio	Lower	Upper
138	100		
108	11.0	10.2	15.2
92	120.6	98.3	147.5

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#54
2,4-Dinitrophenol
Concen: 47.508 ng
RT: 9.980 min Scan# 1341
Delta R.T. 0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:184 Resp: 60419

Ion	Ratio	Lower	Upper
184	100		
63	71.1	84.7	127.1#
154	103.2	338.2	507.2#






#55
Dibenzofuran
Concen: 52.125 ng
RT: 10.133 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:168 Resp: 94762


Ion	Ratio	Lower	Upper
168	100		
139	39.4	31.8	47.6
169	14.4	11.6	17.4



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#56
4-Nitrophenol
Concen: 98.551 ng
RT: 10.033 min Scan# 1350
Delta R.T. 0.012 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:139 Resp: 254819

Ion	Ratio	Lower	Upper
139	100		
109	73.1	53.6	93.6
65	104.9	87.8	127.8






#57
2,4-Dinitrotoluene
Concen: 59.527 ng
RT: 10.110 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:165 Resp: 23856


Ion	Ratio	Lower	Upper
165	100		
63	49.4	40.1	60.1
89	74.2	60.4	90.6
182	2.8	2.0	3.0



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#58
Fluorene
Concen: 51.747 ng
RT: 10.475 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:166 Resp: 716535

Ion	Ratio	Lower	Upper
166	100		
165	98.9	78.5	117.7
167	13.5	10.9	16.3






#59
2,3,4,6-Tetrachlorophenol
Concen: 54.095 ng
RT: 10.245 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:232 Resp: 201940


Ion	Ratio	Lower	Upper
232	100		
131	44.9	37.0	55.6
130	2.1	1.9	2.9
166	29.1	24.2	36.2



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#60
Diethylphthalate
Concen: 53.595 ng
RT: 10.345 min Scan# 1403
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:149 Resp: 792061

Ion	Ratio	Lower	Upper
149	100		
177	22.1	17.9	26.9
150	12.8	10.1	15.1





#61
4-Chlorophenyl-phenylether
Concen: 52.737 ng
RT: 10.469 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35


Tgt Ion:204 Resp: 36877


Ion	Ratio	Lower	Upper
204	100		
206	33.1	26.3	39.5
141	58.5	48.6	73.0

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#62
4-Nitroaniline
Concen: 52.812 ng
RT: 10.492 min Scan# 1428
Delta R.T. 0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:138 Resp: 167633

Ion	Ratio	Lower	Upper
138	100		
92	52.5	35.7	75.7
108	77.6	57.5	97.5






#63
Azobenzene
Concen: 51.543 ng
RT: 10.628 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion: 77 Resp: 80755


Ion	Ratio	Lower	Upper
77	100		
182	25.4	3.3	43.3
105	21.5	1.2	41.2
51	31.4	13.3	53.3



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1584
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:188 Resp: 409717

Ion	Ratio	Lower	Upper
188	100		
94	10.2	7.9	11.9
80	10.9	9.0	13.4





#65
4,6-Dinitro-2-methylphenol
Concen: 36.479 ng
RT: 10.516 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35


Tgt Ion:198 Resp: 6096


Ion	Ratio	Lower	Upper
198	100		
51	52.8	38.2	78.2
105	44.2	27.9	67.9

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#66
n-Nitrosodiphenylamine
Concen: 54.692 ng
RT: 10.586 min Scan# 1444
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:169 Resp: 670666

Ion	Ratio	Lower	Upper
169	100		
168	62.2	49.8	74.6
167	34.5	28.0	42.0






#67
4-Bromophenyl-phenylether
Concen: 55.258 ng
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:248 Resp: 23323


Ion	Ratio	Lower	Upper
248	100		
250	98.0	79.0	118.6
141	71.2	61.7	92.5



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#68
Hexachlorobenzene
Concen: 53.889 ng
RT: 11.022 min Scan# 1518
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:284 Resp: 255815

Ion	Ratio	Lower	Upper
284	100		
142	32.4	26.9	40.3
249	32.2	25.3	37.9





#69
Atrazine
Concen: 63.831 ng
RT: 11.110 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35


Instrument :
BNA_F
ClientSampleId :
VNJ-215MS


Tgt Ion:200 Resp: 21203

Ion	Ratio	Lower	Upper
200	100		
173	27.0	7.7	47.7
215	48.1	27.7	67.7

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#70
Pentachlorophenol
Concen: 77.757 ng
RT: 11.216 min Scan# 1551
Delta R.T. 0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:266 Resp: 224019

Ion	Ratio	Lower	Upper
266	100		
268	64.2	50.4	75.6
264	62.9	50.6	75.8






#71
Phenanthrene
Concen: 53.303 ng
RT: 11.439 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:178 Resp: 103159


Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.8	23.6
179	15.7	12.6	18.8



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#72
Anthracene
Concen: 55.155 ng
RT: 11.492 min Scan# 1598
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:178 Resp: 1041472

Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.3	22.9
179	15.4	12.4	18.6






#73
Carbazole
Concen: 50.241 ng
RT: 11.639 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:167 Resp: 88230


Ion	Ratio	Lower	Upper
167	100		
166	21.5	17.0	25.6
139	13.4	10.8	16.2



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#74
Di-n-butylphthalate
Concen: 55.199 ng
RT: 11.969 min Scan# 1679
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:149 Resp: 1119955

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.5	11.3
104	5.1	4.1	6.1





#75
Fluoranthene
Concen: 46.655 ng
RT: 12.627 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35


Tgt Ion:202 Resp: 91279


Ion	Ratio	Lower	Upper
202	100		
101	11.2	0.0	31.9
203	17.5	0.0	37.5

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.051 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:240 Resp: 203129

Ion	Ratio	Lower	Upper
240	100		
120	10.6	9.4	14.2
236	25.9	20.9	31.3






#77
Benzidine
Concen: 5.301 ng
RT: 12.745 min Scan# 18
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:184 Resp: 1770


Ion	Ratio	Lower	Upper
184	100		
185	16.3	12.0	18.0
183	11.7	9.4	14.2



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#78
Pyrene
Concen: 50.811 ng
RT: 12.851 min Scan# 1829
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:202 Resp: 903442

Ion	Ratio	Lower	Upper
202	100		
200	21.4	17.2	25.8
203	17.6	14.2	21.4






#79
Terphenyl-d14
Concen: 104.783 ng
RT: 12.998 min Scan# 18
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:244 Resp: 130620


Ion	Ratio	Lower	Upper
244	100		
212	6.8	5.7	8.5
122	9.9	8.6	13.0



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#80
Butylbenzylphthalate
Concen: 64.747 ng
RT: 13.469 min Scan# 1934
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:149 Resp: 345144

Ion	Ratio	Lower	Upper
149	100		
91	75.1	59.4	89.2
206	20.6	16.4	24.6





#81
Benzo(a)anthracene
Concen: 54.760 ng
RT: 14.039 min Scan# 20
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35


Tgt Ion:228 Resp: 724089


Ion	Ratio	Lower	Upper
228	100		
226	26.8	21.6	32.4
229	19.8	16.1	24.1

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#82
3,3'-Dichlorobenzidine
Concen: 48.202 ng
RT: 14.004 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:252 Resp: 185837

Ion	Ratio	Lower	Upper
252	100		
254	63.6	50.8	76.2
126	10.7	8.9	13.3






#83
Chrysene
Concen: 57.841 ng
RT: 14.080 min Scan# 20
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:228 Resp: 70125


Ion	Ratio	Lower	Upper
228	100		
226	29.7	24.1	36.1
229	19.9	15.8	23.6



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#84
Bis(2-ethylhexyl)phthalate
Concen: 76.394 ng
RT: 14.027 min Scan# 2029
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:149 Resp: 456891

Ion	Ratio	Lower	Upper
149	100		
167	26.8	21.4	32.0
279	4.1	3.0	4.4





#85
Di-n-octyl phthalate
Concen: 72.931 ng
RT: 14.639 min Scan# 21
Delta R.T. -0.012 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35


Tgt Ion:149 Resp: 79336


Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	11.2	9.8	14.6

Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.527 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:264 Resp: 263019

Ion	Ratio	Lower	Upper
264	100		
260	23.4	19.4	29.2
265	21.7	17.4	26.0






#87
Indeno(1,2,3-cd)pyrene
Concen: 45.859 ng
RT: 17.027 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:276 Resp: 77596


Ion	Ratio	Lower	Upper
276	100		
138	24.1	20.4	30.6
277	25.1	20.2	30.4



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#88
Benzo(b)fluoranthene
Concen: 51.919 ng
RT: 15.098 min Scan# 2211
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:252 Resp: 831851

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.6	26.4
125	9.3	7.9	11.9






#89
Benzo(k)fluoranthene
Concen: 56.615 ng
RT: 15.127 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:252 Resp: 78228


Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.6	26.4
125	9.3	7.9	11.9



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#90
Benzo(a)pyrene
Concen: 59.203 ng
RT: 15.468 min Scan# 2274
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:252 Resp: 780503

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.3	25.9
125	10.4	8.7	13.1






#91
Dibenzo(a,h)anthracene
Concen: 46.124 ng
RT: 17.045 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:278 Resp: 65106


Ion	Ratio	Lower	Upper
278	100		
139	15.2	13.1	19.7
279	23.7	19.0	28.4



Instrument :
BNA_F
ClientSampleId :
VNJ-215MS

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#92
Benzo(g,h,i)perylene
Concen: 37.525 ng
RT: 17.474 min Scan# 2615
Delta R.T. -0.006 min
Lab File: BF140092.D
Acq: 28 Oct 2024 17:35

Tgt Ion:276 Resp: 529084

Ion	Ratio	Lower	Upper
276	100		
277	23.3	19.0	28.6
138	21.2	17.4	26.0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	10/24/24
Project:	Amtrak Sawtooth Bridges 2024	Date Received:	10/24/24
Client Sample ID:	VNJ-215MSD	SDG No.:	P4578
Lab Sample ID:	P4545-02MSD	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140093.D	1	10/26/24 12:10	10/28/24 18:03	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	250		15.5	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	440		8.40	50.0	ug/L
95-48-7	2-Methylphenol	500		11.3	50.0	ug/L
65794-96-9	3+4-Methylphenols	500		11.5	100	ug/L
67-72-1	Hexachloroethane	430		10.1	50.0	ug/L
98-95-3	Nitrobenzene	470		12.7	50.0	ug/L
87-68-3	Hexachlorobutadiene	460		12.7	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	560		8.90	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	540		10.1	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	610		15.2	50.0	ug/L
118-74-1	Hexachlorobenzene	560		11.4	50.0	ug/L
87-86-5	Pentachlorophenol	850	E	18.5	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	136		15 (10) - 110 (139)	91%	SPK: 150
13127-88-3	Phenol-d6	132		15 (10) - 110 (134)	88%	SPK: 150
4165-60-0	Nitrobenzene-d5	105		30 (49) - 130 (133)	105%	SPK: 100
321-60-8	2-Fluorobiphenyl	101		30 (52) - 130 (132)	101%	SPK: 100
118-79-6	2,4,6-Tribromophenol	162		15 (44) - 110 (137)	108%	SPK: 150
1718-51-0	Terphenyl-d14	105		30 (48) - 130 (125)	105%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	112000	6.886			
1146-65-2	Naphthalene-d8	432000	8.169			
15067-26-2	Acenaphthene-d10	237000	9.927			
1517-22-2	Phenanthrene-d10	399000	11.41			
1719-03-5	Chrysene-d12	204000	14.051			
1520-96-3	Perylene-d12	256000	15.527			

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Portal Partners Tri-Venture	Date Collected:	10/24/24
Project:	Amtrak Sawtooth Bridges 2024	Date Received:	10/24/24
Client Sample ID:	VNJ-215MSD	SDG No.:	P4578
Lab Sample ID:	P4545-02MSD	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140093.D	1	10/26/24 12:10	10/28/24 18:03	PB164444

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140093.D
Acq On : 28 Oct 2024 18:03
Operator : RC/JU
Sample : P4545-02MSD
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

Quant Time: Oct 28 18:37:04 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.886	152	111820	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	432287	20.000	ng	0.00
39) Acenaphthene-d10	9.927	164	236921	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	398949	20.000	ng	0.00
76) Chrysene-d12	14.051	240	204155	20.000	ng	0.00
86) Perylene-d12	15.527	264	256308	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.522	112	974076	136.371	ng	0.02
7) Phenol-d6	6.516	99	1219104	131.779	ng	0.00
23) Nitrobenzene-d5	7.451	82	820015	105.134	ng	0.00
42) 2,4,6-Tribromophenol	10.716	330	358251	161.659	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1443612	100.702	ng	0.00
79) Terphenyl-d14	12.998	244	1318745	105.257	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.881	88	128884	38.700	ng	# 81
3) Pyridine	3.587	79	206806	25.053	ng	98
4) n-Nitrosodimethylamine	3.516	42	189912	42.820	ng	# 98
6) Aniline	6.551	93	182449	21.161	ng	99
8) 2-Chlorophenol	6.675	128	372961	51.076	ng	97
9) Benzaldehyde	6.439	77	44949	8.637	ng	95
10) Phenol	6.528	94	448413m	47.016	ng	
11) bis(2-Chloroethyl)ether	6.628	93	356154	48.313	ng	99
12) 1,3-Dichlorobenzene	6.834	146	359207	42.853	ng	99
13) 1,4-Dichlorobenzene	6.904	146	367348	43.866	ng	99
14) 1,2-Dichlorobenzene	7.057	146	355848	45.529	ng	99
15) Benzyl Alcohol	7.028	79	349195	51.458	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.157	45	577398	45.935	ng	96
17) 2-Methylphenol	7.134	107	313004	50.269	ng	99
18) Hexachloroethane	7.404	117	129514	43.369	ng	95
19) n-Nitroso-di-n-propyla...	7.298	70	273178	48.688	ng	98
20) 3+4-Methylphenols	7.286	107	394617	49.549	ng	92
22) Acetophenone	7.298	105	510185	48.185	ng	97
24) Nitrobenzene	7.469	77	405962	47.472	ng	100
25) Isophorone	7.710	82	761382	51.431	ng	99
26) 2-Nitrophenol	7.786	139	190721	59.118	ng	96
27) 2,4-Dimethylphenol	7.816	122	316975	58.924	ng	99
28) bis(2-Chloroethoxy)met...	7.916	93	451618	50.352	ng	99
29) 2,4-Dichlorophenol	8.022	162	320207	52.260	ng	100
30) 1,2,4-Trichlorobenzene	8.110	180	314131	46.333	ng	100
31) Naphthalene	8.192	128	1064132	47.730	ng	100
32) Benzoic acid	7.904	122	123617	26.347	ng	98
33) 4-Chloroaniline	8.233	127	150937	19.854	ng	99
34) Hexachlorobutadiene	8.310	225	196299	46.081	ng	99
35) Caprolactam	8.604	113	85333m	43.800	ng	
36) 4-Chloro-3-methylphenol	8.716	107	365021	53.801	ng	98
37) 2-Methylnaphthalene	8.880	142	701854	51.402	ng	99
38) 1-Methylnaphthalene	8.980	142	652303	48.693	ng	100
40) 1,2,4,5-Tetrachloroben...	9.045	216	331879	51.923	ng	99
41) Hexachlorocyclopentadiene	9.033	237	348918	156.886	ng	99
43) 2,4,6-Trichlorophenol	9.157	196	253091	55.928	ng	99

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140093.D
Acq On : 28 Oct 2024 18:03
Operator : RC/JU
Sample : P4545-02MSD
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations APPROVED
Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024

Quant Time: Oct 28 18:37:04 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	9.204	196	253581	54.313	ng		98
46) 1,1'-Biphenyl	9.345	154	857737	52.006	ng		100
47) 2-Chloronaphthalene	9.375	162	680498	51.228	ng		99
48) 2-Nitroaniline	9.463	65	244814	60.258	ng		99
49) Acenaphthylene	9.786	152	1071538	55.796	ng		99
50) Dimethylphthalate	9.645	163	828465	56.139	ng		100
51) 2,6-Dinitrotoluene	9.710	165	177848	55.660	ng		95
52) Acenaphthene	9.963	154	668819	53.836	ng		99
53) 3-Nitroaniline	9.875	138	128514	39.002	ng		97
54) 2,4-Dinitrophenol	9.980	184	66396	52.460	ng	#	1
55) Dibenzofuran	10.133	168	950618	53.332	ng		99
56) 4-Nitrophenol	10.033	139	257304	101.497	ng		99
57) 2,4-Dinitrotoluene	10.110	165	238589	60.720	ng		99
58) Fluorene	10.475	166	727096	53.557	ng		99
59) 2,3,4,6-Tetrachlorophenol	10.245	232	204109	55.766	ng		99
60) Diethylphthalate	10.345	149	792166	54.672	ng		99
61) 4-Chlorophenyl-phenyle...	10.469	204	374542	54.630	ng		97
62) 4-Nitroaniline	10.492	138	166610	53.537	ng		98
63) Azobenzene	10.627	77	802948	52.271	ng		97
65) 4,6-Dinitro-2-methylph...	10.516	198	66402	40.805	ng		95
66) n-Nitrosodiphenylamine	10.586	169	673533	56.408	ng		99
67) 4-Bromophenyl-phenylether	10.957	248	231386	56.299	ng		98
68) Hexachlorobenzene	11.022	284	257174	55.638	ng		98
69) Atrazine	11.110	200	211444	65.372	ng		99
70) Pentachlorophenol	11.216	266	237118	84.525	ng		99
71) Phenanthrene	11.439	178	1027335	54.516	ng		100
72) Anthracene	11.492	178	1041878	56.666	ng		100
73) Carbazole	11.639	167	877488	51.316	ng		100
74) Di-n-butylphthalate	11.969	149	1110467	56.209	ng		100
75) Fluoranthene	12.621	202	920478	48.318	ng		100
77) Benzidine	12.745	184	27530	8.201	ng		98
78) Pyrene	12.851	202	902116	50.481	ng		99
80) Butylbenzylphthalate	13.468	149	344594	64.319	ng		100
81) Benzo(a)anthracene	14.039	228	746125	56.143	ng		99
82) 3,3'-Dichlorobenzidine	14.004	252	188100	48.544	ng		99
83) Chrysene	14.074	228	707350	58.050	ng		100
84) Bis(2-ethylhexyl)phtha...	14.027	149	458667	76.306	ng		100
85) Di-n-octyl phthalate	14.639	149	799157	73.095	ng		97
87) Indeno(1,2,3-cd)pyrene	17.021	276	748376	45.387	ng		99
88) Benzo(b)fluoranthene	15.092	252	856677	54.869	ng		100
89) Benzo(k)fluoranthene	15.127	252	777981	57.778	ng		99
90) Benzo(a)pyrene	15.468	252	787766	61.319	ng		100
91) Dibenzo(a,h)anthracene	17.045	278	626364	45.536	ng		98
92) Benzo(g,h,i)perylene	17.474	276	506321	36.850	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102824\
Data File : BF140093.D
Acq On : 28 Oct 2024 18:03
Operator : RC/JU
Sample : P4545-02MSD
Misc :
ALS Vial : 20 Sample Multiplier: 1


Quant Time: Oct 28 18:37:04 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.886 min Scan# 81
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:152 Resp: 11182


Ion	Ratio	Lower	Upper
152	100		
150	158.5	130.2	195.2
115	63.3	51.4	77.2



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#2
1,4-Dioxane
Concen: 38.700 ng
RT: 2.881 min Scan# 134
Delta R.T. 0.194 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion: 88 Resp: 128884

Ion	Ratio	Lower	Upper
88	100		
58	87.7	73.3	109.9
43	0.0	27.1	40.7#






#3
Pyridine
Concen: 25.053 ng
RT: 3.587 min Scan# 2
Delta R.T. 0.141 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion: 79 Resp: 20680


Ion	Ratio	Lower	Upper
79	100		
52	83.6	68.5	102.7
51	39.7	32.3	48.5



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#4
n-Nitrosodimethylamine
Concen: 42.820 ng
RT: 3.516 min Scan# 242
Delta R.T. 0.118 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion: 42 Resp: 189912

Ion	Ratio	Lower	Upper
42	100		
74	121.6	96.2	144.4
44	10.3	6.2	9.2#






#5
2-Fluorophenol
Concen: 136.371 ng
RT: 5.522 min Scan# 58
Delta R.T. 0.024 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:112 Resp: 97407


Ion	Ratio	Lower	Upper
112	100		
64	65.0	53.0	79.6
63	33.5	27.0	40.4



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#6
Aniline
Concen: 21.161 ng
RT: 6.551 min Scan# 758
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion: 93 Resp: 182449

Ion	Ratio	Lower	Upper
93	100		
66	43.1	34.0	51.0
65	22.1	17.4	26.0






#7
Phenol-d6
Concen: 131.779 ng
RT: 6.516 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion: 99 Resp: 121910


Ion	Ratio	Lower	Upper
99	100		
42	19.7	16.7	25.1
71	35.2	27.7	41.5



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#8
2-Chlorophenol
Concen: 51.076 ng
RT: 6.675 min Scan# 779
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:128 Resp: 372961

Ion	Ratio	Lower	Upper
128	100		
130	32.7	12.3	52.3
64	51.8	34.6	74.6






#9
Benzaldehyde
Concen: 8.637 ng
RT: 6.439 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion: 77 Resp: 4494


Ion	Ratio	Lower	Upper
77	100		
105	98.6	75.5	115.5
106	97.6	71.7	111.7



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#10
Phenol
Concen: 47.016 ng m
RT: 6.528 min Scan# 754
Delta R.T. 0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion: 94 Resp: 448413

Ion	Ratio	Lower	Upper
94	100		
65	31.6	11.1	51.1
66	45.3	22.7	62.7





#11
bis(2-Chloroethyl)ether
Concen: 48.313 ng
RT: 6.628 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03


Tgt Ion: 93 Resp: 35615


Ion	Ratio	Lower	Upper
93	100		
63	83.5	65.0	105.0
95	32.4	12.4	52.4

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#12
1,3-Dichlorobenzene
Concen: 42.853 ng
RT: 6.834 min Scan# 806
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:146 Resp: 359207

Ion	Ratio	Lower	Upper
146	100		
148	63.9	51.0	76.4
75	31.6	26.4	39.6






#13
1,4-Dichlorobenzene
Concen: 43.866 ng
RT: 6.904 min Scan# 81
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:146 Resp: 36734


Ion	Ratio	Lower	Upper
146	100		
148	63.6	51.8	77.8
111	43.3	34.6	51.8



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#14
1,2-Dichlorobenzene
Concen: 45.529 ng
RT: 7.057 min Scan# 844
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:146 Resp: 355848

Ion	Ratio	Lower	Upper
146	100		
148	63.9	51.4	77.2
111	45.0	36.7	55.1





#15
Benzyl Alcohol
Concen: 51.458 ng
RT: 7.028 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03


Tgt Ion: 79 Resp: 34919


Ion	Ratio	Lower	Upper
79	100		
108	76.0	61.4	92.0
77	62.6	49.9	74.9

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#16
2,2'-oxybis(1-Chloropropane)
Concen: 45.935 ng
RT: 7.157 min Scan# 861
Delta R.T. -0.012 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion: 45 Resp: 577398

Ion	Ratio	Lower	Upper
45	100		
77	13.2	0.0	31.5
79	10.0	0.0	29.1






#17
2-Methylphenol
Concen: 50.269 ng
RT: 7.134 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:107 Resp: 313004


Ion	Ratio	Lower	Upper
107	100		
108	113.3	91.3	136.9
77	47.3	37.1	55.7
79	45.7	37.0	55.6



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#18
Hexachloroethane
Concen: 43.369 ng
RT: 7.404 min Scan# 903
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:117 Resp: 129514

Ion	Ratio	Lower	Upper
117	100		
119	94.7	77.0	115.4
201	104.2	76.2	114.4






#19
n-Nitroso-di-n-propylamine
Concen: 48.688 ng
RT: 7.298 min Scan# 88
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion: 70 Resp: 273178


Ion	Ratio	Lower	Upper
70	100		
42	61.7	51.5	77.3
101	9.7	7.5	11.3
130	19.5	15.5	23.3



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#20
3+4-Methylphenols
Concen: 49.549 ng
RT: 7.286 min Scan# 883
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:107 Resp: 394617

Ion	Ratio	Lower	Upper
107	100		
108	89.6	69.8	109.8
77	88.5	50.5	90.5
79	29.4	10.2	50.2






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.169 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:136 Resp: 43228


Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.6	12.8
54	10.2	8.4	12.6
68	6.0	5.1	7.7



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#22
Acetophenone
Concen: 48.185 ng
RT: 7.298 min Scan# 885
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:105 Resp: 510185

Ion	Ratio	Lower	Upper
105	100		
71	8.7	5.9	8.9
51	34.6	29.4	44.2
120	21.4	17.5	26.3






#23
Nitrobenzene-d5
Concen: 105.134 ng
RT: 7.451 min Scan# 91
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion: 82 Resp: 82001


Ion	Ratio	Lower	Upper
82	100		
128	43.2	33.4	50.0
54	56.8	47.8	71.8



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#24
Nitrobenzene
Concen: 47.472 ng
RT: 7.469 min Scan# 914
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion: 77 Resp: 405962

Ion	Ratio	Lower	Upper
77	100		
123	44.7	35.7	53.5
65	14.5	11.5	17.3






#25
Isophorone
Concen: 51.431 ng
RT: 7.710 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion: 82 Resp: 76138


Ion	Ratio	Lower	Upper
82	100		
95	7.3	5.5	8.3
138	16.8	13.2	19.8



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#26
2-Nitrophenol
Concen: 59.118 ng
RT: 7.786 min Scan# 968
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:139 Resp: 190721

Ion	Ratio	Lower	Upper
139	100		
109	30.4	25.1	37.7
65	55.5	47.5	71.3






#27
2,4-Dimethylphenol
Concen: 58.924 ng
RT: 7.816 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:122 Resp: 31697


Ion	Ratio	Lower	Upper
122	100		
107	116.1	91.9	137.9
121	58.5	46.2	69.4



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#28
bis(2-Chloroethoxy)methane
Concen: 50.352 ng
RT: 7.916 min Scan# 990
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion: 93 Resp: 451618

Ion	Ratio	Lower	Upper
93	100		
95	32.6	26.0	39.0
123	11.5	9.8	14.8






#29
2,4-Dichlorophenol
Concen: 52.260 ng
RT: 8.022 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:162 Resp: 32020


Ion	Ratio	Lower	Upper
162	100		
164	63.9	43.9	83.9
98	39.8	19.9	59.9



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#30
1,2,4-Trichlorobenzene
Concen: 46.333 ng
RT: 8.110 min Scan# 1023
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:180 Resp: 314131

Ion	Ratio	Lower	Upper
180	100		
182	100.4	80.5	120.7
145	31.8	25.4	38.0






#31
Naphthalene
Concen: 47.730 ng
RT: 8.192 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:128 Resp: 106413


Ion	Ratio	Lower	Upper
128	100		
129	11.1	8.9	13.3
127	13.2	10.6	16.0



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#32
Benzoic acid
Concen: 26.347 ng
RT: 7.904 min Scan# 988
Delta R.T. -0.018 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:122 Resp: 123617

Ion	Ratio	Lower	Upper
122	100		
105	125.8	107.6	147.6
77	97.8	80.2	120.2






#33
4-Chloroaniline
Concen: 19.854 ng
RT: 8.233 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:127 Resp: 15093


Ion	Ratio	Lower	Upper
127	100		
129	31.5	25.4	38.2
65	33.9	28.4	42.6
92	20.5	16.6	25.0



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#34
Hexachlorobutadiene
Concen: 46.081 ng
RT: 8.310 min Scan# 1057
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:225 Resp: 196299

Ion	Ratio	Lower	Upper
225	100		
223	63.5	50.6	76.0
227	64.1	51.8	77.8





#35
Caprolactam
Concen: 43.800 ng m
RT: 8.604 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03


Tgt Ion:113 Resp: 8533


Ion	Ratio	Lower	Upper
113	100		
55	201.5	189.2	229.2
56	148.0	132.5	172.5

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#36
4-Chloro-3-methylphenol
Concen: 53.801 ng
RT: 8.716 min Scan# 1126
Delta R.T. 0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:107 Resp: 365021

Ion	Ratio	Lower	Upper
107	100		
144	24.6	19.1	28.7
142	72.9	57.2	85.8





#37
2-Methylnaphthalene
Concen: 51.402 ng
RT: 8.880 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03


Tgt Ion:142 Resp: 70185


Ion	Ratio	Lower	Upper
142	100		
141	87.9	70.8	106.2

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#38
1-Methylnaphthalene
Concen: 48.693 ng
RT: 8.980 min Scan# 1171
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:142 Resp: 652303

Ion	Ratio	Lower	Upper
142	100		
141	91.8	73.5	110.3





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.927 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03


Tgt Ion:164 Resp: 23692


Ion	Ratio	Lower	Upper
164	100		
162	100.8	81.0	121.4
160	44.0	35.4	53.0

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#40
1,2,4,5-Tetrachlorobenzene
Concen: 51.923 ng
RT: 9.045 min Scan# 1182
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:216 Resp: 331879

Ion	Ratio	Lower	Upper
216	100		
214	79.1	63.4	95.0
179	21.1	17.2	25.8
108	21.1	15.6	23.4





#41
Hexachlorocyclopentadiene
Concen: 156.886 ng
RT: 9.033 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03


Tgt Ion:237 Resp: 348918


Ion	Ratio	Lower	Upper
237	100		
235	62.2	42.9	82.9
272	12.5	0.0	32.3

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#42
2,4,6-Tribromophenol
Concen: 161.659 ng
RT: 10.716 min Scan# 1466
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:330 Resp: 358251

Ion	Ratio	Lower	Upper
330	100		
332	95.9	78.1	117.1
141	29.3	26.6	39.8






#43
2,4,6-Trichlorophenol
Concen: 55.928 ng
RT: 9.157 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:196 Resp: 25309


Ion	Ratio	Lower	Upper
196	100		
198	94.7	76.2	114.2
200	30.2	24.0	36.0



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#44
2,4,5-Trichlorophenol
Concen: 54.313 ng
RT: 9.204 min Scan# 1209
Delta R.T. 0.012 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:196 Resp: 253581

Ion	Ratio	Lower	Upper
196	100		
198	96.3	75.8	113.6
97	50.3	41.8	62.6
132	26.9	22.9	34.3






#45
2-Fluorobiphenyl
Concen: 100.702 ng
RT: 9.245 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:172 Resp: 144361


Ion	Ratio	Lower	Upper
172	100		
171	35.6	28.6	43.0
170	24.0	19.1	28.7



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#46
1,1'-Biphenyl
Concen: 52.006 ng
RT: 9.345 min Scan# 1233
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:154 Resp: 857737

Ion	Ratio	Lower	Upper
154	100		
153	40.3	20.3	60.3
76	13.5	0.0	33.9






#47
2-Chloronaphthalene
Concen: 51.228 ng
RT: 9.375 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:162 Resp: 68049


Ion	Ratio	Lower	Upper
162	100		
127	41.8	33.8	50.6
164	32.8	26.6	39.8



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#48
2-Nitroaniline
Concen: 60.258 ng
RT: 9.463 min Scan# 1253
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion: 65 Resp: 244814

Ion	Ratio	Lower	Upper
65	100		
92	63.9	50.9	76.3
138	93.8	73.4	110.0






#49
Acenaphthylene
Concen: 55.796 ng
RT: 9.786 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:152 Resp: 107153


Ion	Ratio	Lower	Upper
152	100		
151	20.3	15.9	23.9
153	13.5	10.7	16.1



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#50
Dimethylphthalate
Concen: 56.139 ng
RT: 9.645 min Scan# 1284
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:163 Resp: 828465

Ion	Ratio	Lower	Upper
163	100		
194	4.0	3.4	5.0
164	10.2	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 55.660 ng
RT: 9.710 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03


Tgt Ion:165 Resp: 17784


Ion	Ratio	Lower	Upper
165	100		
63	58.8	50.6	75.8
89	55.8	47.5	71.3

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#52
Acenaphthene
Concen: 53.836 ng
RT: 9.963 min Scan# 1338
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:154 Resp: 668819

Ion	Ratio	Lower	Upper
154	100		
153	110.8	89.8	134.6
152	53.8	42.4	63.6





#53
3-Nitroaniline
Concen: 39.002 ng
RT: 9.875 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03


Tgt Ion:138 Resp: 12851


Ion	Ratio	Lower	Upper
138	100		
108	11.3	10.2	15.2
92	119.0	98.3	147.5

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#54
2,4-Dinitrophenol
Concen: 52.460 ng
RT: 9.980 min Scan# 1341
Delta R.T. 0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:184 Resp: 66396

Ion	Ratio	Lower	Upper
184	100		
63	70.8	84.7	127.1#
154	98.9	338.2	507.2#






#55
Dibenzofuran
Concen: 53.332 ng
RT: 10.133 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:168 Resp: 95061


Ion	Ratio	Lower	Upper
168	100		
139	39.2	31.8	47.6
169	14.4	11.6	17.4



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#56
4-Nitrophenol
Concen: 101.497 ng
RT: 10.033 min Scan# 1350
Delta R.T. 0.012 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:139 Resp: 257304

Ion	Ratio	Lower	Upper
139	100		
109	74.2	53.6	93.6
65	105.7	87.8	127.8






#57
2,4-Dinitrotoluene
Concen: 60.720 ng
RT: 10.110 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:165 Resp: 23858


Ion	Ratio	Lower	Upper
165	100		
63	49.3	40.1	60.1
89	74.4	60.4	90.6
182	2.8	2.0	3.0



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#58
Fluorene
Concen: 53.557 ng
RT: 10.475 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:166 Resp: 727096

Ion	Ratio	Lower	Upper
166	100		
165	97.3	78.5	117.7
167	13.4	10.9	16.3






#59
2,3,4,6-Tetrachlorophenol
Concen: 55.766 ng
RT: 10.245 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:232 Resp: 204109


Ion	Ratio	Lower	Upper
232	100		
131	45.6	37.0	55.6
130	2.3	1.9	2.9
166	29.1	24.2	36.2



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#60
Diethylphthalate
Concen: 54.672 ng
RT: 10.345 min Scan# 1403
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:149 Resp: 792166

Ion	Ratio	Lower	Upper
149	100		
177	21.8	17.9	26.9
150	12.6	10.1	15.1





#61
4-Chlorophenyl-phenylether
Concen: 54.630 ng
RT: 10.469 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03


Tgt Ion:204 Resp: 37454


Ion	Ratio	Lower	Upper
204	100		
206	32.5	26.3	39.5
141	58.0	48.6	73.0

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#62
4-Nitroaniline
Concen: 53.537 ng
RT: 10.492 min Scan# 1428
Delta R.T. 0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:138 Resp: 166610

Ion	Ratio	Lower	Upper
138	100		
92	51.9	35.7	75.7
108	77.9	57.5	97.5






#63
Azobenzene
Concen: 52.271 ng
RT: 10.627 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion: 77 Resp: 80294


Ion	Ratio	Lower	Upper
77	100		
182	25.1	3.3	43.3
105	21.6	1.2	41.2
51	31.8	13.3	53.3



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1584
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:188 Resp: 398949

Ion	Ratio	Lower	Upper
188	100		
94	10.1	7.9	11.9
80	10.6	9.0	13.4





#65
4,6-Dinitro-2-methylphenol
Concen: 40.805 ng
RT: 10.516 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03


Tgt Ion:198 Resp: 6640


Ion	Ratio	Lower	Upper
198	100		
51	52.3	38.2	78.2
105	46.4	27.9	67.9

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#66
n-Nitrosodiphenylamine
Concen: 56.408 ng
RT: 10.586 min Scan# 1444
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:169 Resp: 673533

Ion	Ratio	Lower	Upper
169	100		
168	61.7	49.8	74.6
167	34.4	28.0	42.0






#67
4-Bromophenyl-phenylether
Concen: 56.299 ng
RT: 10.957 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:248 Resp: 23138


Ion	Ratio	Lower	Upper
248	100		
250	98.9	79.0	118.6
141	72.4	61.7	92.5



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#68
Hexachlorobenzene
Concen: 55.638 ng
RT: 11.022 min Scan# 1518
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:284 Resp: 257174

Ion	Ratio	Lower	Upper
284	100		
142	32.0	26.9	40.3
249	31.6	25.3	37.9






#69
Atrazine
Concen: 65.372 ng
RT: 11.110 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:200 Resp: 21144


Ion	Ratio	Lower	Upper
200	100		
173	27.1	7.7	47.7
215	48.3	27.7	67.7



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#70
Pentachlorophenol
Concen: 84.525 ng
RT: 11.216 min Scan# 1551
Delta R.T. 0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:266 Resp: 237118

Ion	Ratio	Lower	Upper
266	100		
268	62.5	50.4	75.6
264	61.7	50.6	75.8






#71
Phenanthrene
Concen: 54.516 ng
RT: 11.439 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:178 Resp: 102733


Ion	Ratio	Lower	Upper
178	100		
176	19.8	15.8	23.6
179	15.6	12.6	18.8



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#72
Anthracene
Concen: 56.666 ng
RT: 11.492 min Scan# 1598
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:178 Resp: 1041878

Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.3	22.9
179	15.6	12.4	18.6






#73
Carbazole
Concen: 51.316 ng
RT: 11.639 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:167 Resp: 87748


Ion	Ratio	Lower	Upper
167	100		
166	21.4	17.0	25.6
139	13.5	10.8	16.2



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#74
Di-n-butylphthalate
Concen: 56.209 ng
RT: 11.969 min Scan# 1679
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:149 Resp: 1110467

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.5	11.3
104	5.1	4.1	6.1






#75
Fluoranthene
Concen: 48.318 ng
RT: 12.621 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:202 Resp: 92047


Ion	Ratio	Lower	Upper
202	100		
101	12.0	0.0	31.9
203	17.7	0.0	37.5



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.051 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:240 Resp: 204155

Ion	Ratio	Lower	Upper
240	100		
120	10.4	9.4	14.2
236	26.1	20.9	31.3





#77
Benzidine
Concen: 8.201 ng
RT: 12.745 min Scan# 18
Delta R.T. -0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03


Tgt Ion:184 Resp: 2753


Ion	Ratio	Lower	Upper
184	100		
185	16.3	12.0	18.0
183	11.6	9.4	14.2

Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024







#78
Pyrene
Concen: 50.481 ng
RT: 12.851 min Scan# 1829
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:202 Resp: 902116

Ion	Ratio	Lower	Upper
202	100		
200	21.6	17.2	25.8
203	17.4	14.2	21.4






#79
Terphenyl-d14
Concen: 105.257 ng
RT: 12.998 min Scan# 18
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:244 Resp: 131874


Ion	Ratio	Lower	Upper
244	100		
212	6.6	5.7	8.5
122	9.7	8.6	13.0



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#80
Butylbenzylphthalate
Concen: 64.319 ng
RT: 13.468 min Scan# 1934
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:149 Resp: 344594

Ion	Ratio	Lower	Upper
149	100		
91	74.3	59.4	89.2
206	20.6	16.4	24.6






#81
Benzo(a)anthracene
Concen: 56.143 ng
RT: 14.039 min Scan# 20
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:228 Resp: 74612


Ion	Ratio	Lower	Upper
228	100		
226	26.7	21.6	32.4
229	19.6	16.1	24.1



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#82
3,3'-Dichlorobenzidine
Concen: 48.544 ng
RT: 14.004 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:252 Resp: 188100

Ion	Ratio	Lower	Upper
252	100		
254	64.2	50.8	76.2
126	10.6	8.9	13.3






#83
Chrysene
Concen: 58.050 ng
RT: 14.074 min Scan# 20
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:228 Resp: 70735


Ion	Ratio	Lower	Upper
228	100		
226	30.2	24.1	36.1
229	19.9	15.8	23.6



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#84
Bis(2-ethylhexyl)phthalate
Concen: 76.306 ng
RT: 14.027 min Scan# 2029
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:149 Resp: 458667

Ion	Ratio	Lower	Upper
149	100		
167	26.7	21.4	32.0
279	4.1	3.0	4.4






#85
Di-n-octyl phthalate
Concen: 73.095 ng
RT: 14.639 min Scan# 21
Delta R.T. -0.012 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:149 Resp: 79915


Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	10.9	9.8	14.6



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#86
Perylene-d12
Concen: 20.000 ng
RT: 15.527 min Scan# 2284
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:264 Resp: 256308

Ion	Ratio	Lower	Upper
264	100		
260	24.0	19.4	29.2
265	21.0	17.4	26.0






#87
Indeno(1,2,3-cd)pyrene
Concen: 45.387 ng
RT: 17.021 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:276 Resp: 74837


Ion	Ratio	Lower	Upper
276	100		
138	24.2	20.4	30.6
277	25.2	20.2	30.4



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#88
Benzo(b)fluoranthene
Concen: 54.869 ng
RT: 15.092 min Scan# 2210
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:252 Resp: 856677

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.6	26.4
125	10.2	7.9	11.9






#89
Benzo(k)fluoranthene
Concen: 57.778 ng
RT: 15.127 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:252 Resp: 77798


Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.6	26.4
125	9.1	7.9	11.9



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024




#90
Benzo(a)pyrene
Concen: 61.319 ng
RT: 15.468 min Scan# 2274
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:252 Resp: 787766

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.3	25.9
125	10.6	8.7	13.1






#91
Dibenzo(a,h)anthracene
Concen: 45.536 ng
RT: 17.045 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:278 Resp: 626364


Ion	Ratio	Lower	Upper
278	100		
139	14.9	13.1	19.7
279	24.0	19.0	28.4



Instrument :
BNA_F
ClientSampleId :
VNJ-215MSD

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/29/2024
Supervised By :mohammad ahmed 10/30/2024



#92
Benzo(g,h,i)perylene
Concen: 36.850 ng
RT: 17.474 min Scan# 2615
Delta R.T. -0.006 min
Lab File: BF140093.D
Acq: 28 Oct 2024 18:03

Tgt Ion:276 Resp: 506321

Ion	Ratio	Lower	Upper
276	100		
277	23.8	19.0	28.6
138	21.2	17.4	26.0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BF101824	Instrument	BNA_f

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDICCC040	BF139848.D	Phenol	yogesh	10/21/2024 6:33:45 AM	mohammad	10/21/2024 6:38:35 AM	Peak Integrated by Software
SSTDICC050	BF139849.D	Phenol	yogesh	10/21/2024 6:33:46 AM	mohammad	10/21/2024 6:38:35 AM	Peak Integrated by Software
SSTDICC060	BF139850.D	Phenol	yogesh	10/21/2024 6:33:47 AM	mohammad	10/21/2024 6:38:35 AM	Peak Integrated by Software
SSTDICC080	BF139851.D	Aniline	yogesh	10/21/2024 6:33:49 AM	mohammad	10/21/2024 6:38:35 AM	Peak Integrated by Software
SSTDICV040	BF139852.D	Phenol	yogesh	10/21/2024 6:33:50 AM	mohammad	10/21/2024 6:38:35 AM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BF102824	Instrument	BNA_f

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDCCC040	BF140075.D	Phenol	yogesh	10/29/2024 1:35:05 AM	mohammad	10/29/2024 1:38:39 AM	Peak Integrated by Software
PB164444BS	BF140077.D	Caprolactam	yogesh	10/29/2024 1:35:06 AM	mohammad	10/29/2024 1:38:39 AM	Peak Integrated by Software
PB164444BS	BF140077.D	Phenol	yogesh	10/29/2024 1:35:06 AM	mohammad	10/29/2024 1:38:39 AM	Peak Integrated by Software
P4545-02MS	BF140092.D	Caprolactam	yogesh	10/29/2024 4:52:56 AM	mohammad	10/29/2024 1:38:39 AM	Peak Integrated by Software
P4545-02MS	BF140092.D	Phenol	yogesh	10/29/2024 4:52:56 AM	mohammad	10/29/2024 1:38:39 AM	Peak Integrated by Software
P4545-02MSD	BF140093.D	Caprolactam	yogesh	10/29/2024 1:35:09 AM	mohammad	10/29/2024 1:38:39 AM	Peak Integrated by Software
P4545-02MSD	BF140093.D	Phenol	yogesh	10/29/2024 1:35:09 AM	mohammad	10/29/2024 1:38:39 AM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF101824

Review By	yogesh	Review On	10/21/2024 6:34:01 AM				
Supervise By	mohammad	Supervise On	10/21/2024 6:38:35 AM				
SubDirectory	BF101824	HP Acquire Method	BNA_F		HP Processing Method	bf101824	

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12322,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF139843.D	18 Oct 2024 09:22	RC/JU	Ok
2	SSTDICC2.5	BF139844.D	18 Oct 2024 10:27	RC/JU	Ok
3	SSTDICC005	BF139845.D	18 Oct 2024 10:55	RC/JU	Ok
4	SSTDICC010	BF139846.D	18 Oct 2024 11:23	RC/JU	Ok
5	SSTDICC020	BF139847.D	18 Oct 2024 11:52	RC/JU	Ok
6	SSTDICCC040	BF139848.D	18 Oct 2024 12:20	RC/JU	Ok,M
7	SSTDICC050	BF139849.D	18 Oct 2024 12:49	RC/JU	Ok,M
8	SSTDICC060	BF139850.D	18 Oct 2024 13:17	RC/JU	Ok,M
9	SSTDICC080	BF139851.D	18 Oct 2024 13:46	RC/JU	Ok,M
10	SSTDICV040	BF139852.D	18 Oct 2024 14:19	RC/JU	Ok,M
11	PB164211BL	BF139853.D	18 Oct 2024 14:48	RC/JU	Ok
12	P4405-01	BF139854.D	18 Oct 2024 15:21	RC/JU	Ok
13	P4431-01	BF139855.D	18 Oct 2024 15:50	RC/JU	Ok,M
14	P4421-01	BF139856.D	18 Oct 2024 16:18	RC/JU	Ok,M
15	P4422-01	BF139857.D	18 Oct 2024 16:46	RC/JU	Ok
16	P4425-01	BF139858.D	18 Oct 2024 17:15	RC/JU	Ok
17	P4425-03	BF139859.D	18 Oct 2024 17:44	RC/JU	Ok
18	P4425-05	BF139860.D	18 Oct 2024 18:12	RC/JU	Ok
19	P4425-07	BF139861.D	18 Oct 2024 18:41	RC/JU	ReRun
20	P4425-09	BF139862.D	18 Oct 2024 19:09	RC/JU	ReRun
21	P4426-03	BF139863.D	18 Oct 2024 19:37	RC/JU	ReRun

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF101824

Review By	yogesh	Review On	10/21/2024 6:34:01 AM		
Supervise By	mohammad	Supervise On	10/21/2024 6:38:35 AM		
SubDirectory	BF101824	HP Acquire Method	BNA_F	HP Processing Method	bf101824

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12322,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	P4426-07	BF139864.D	18 Oct 2024 20:05	RC/JU	ReRun
23	P4426-17	BF139865.D	18 Oct 2024 20:34	RC/JU	ReRun
24	P4426-11	BF139866.D	18 Oct 2024 21:02	RC/JU	ReRun

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF102824

Review By	yogesh	Review On	10/29/2024 1:35:27 AM				
Supervise By	mohammad	Supervise On	10/29/2024 1:38:39 AM				
SubDirectory	BF102824	HP Acquire Method	BNA_F		HP Processing Method	bf101824	

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12323,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF140074.D	28 Oct 2024 09:04	RC/JU	Ok
2	SSTDCCC040	BF140075.D	28 Oct 2024 09:32	RC/JU	Ok,M
3	PB164444BL	BF140076.D	28 Oct 2024 10:01	RC/JU	Ok
4	PB164444BS	BF140077.D	28 Oct 2024 10:29	RC/JU	Ok,M
5	PB164393TB	BF140078.D	28 Oct 2024 10:58	RC/JU	Ok
6	PB164444BL	BF140079.D	28 Oct 2024 11:31	RC/JU	Not Ok
7	P4547-04	BF140080.D	28 Oct 2024 11:59	RC/JU	Ok
8	P4547-08	BF140081.D	28 Oct 2024 12:27	RC/JU	Ok
9	P4561-04	BF140082.D	28 Oct 2024 12:55	RC/JU	Ok
10	P4561-08	BF140083.D	28 Oct 2024 13:23	RC/JU	Ok
11	P4578-06	BF140084.D	28 Oct 2024 13:51	RC/JU	Ok
12	P4578-08	BF140085.D	28 Oct 2024 14:19	RC/JU	Ok
13	P4567-04	BF140086.D	28 Oct 2024 14:47	RC/JU	Ok
14	P4567-08	BF140087.D	28 Oct 2024 15:15	RC/JU	Ok
15	P4567-12	BF140088.D	28 Oct 2024 15:43	RC/JU	Ok
16	P4574-03	BF140089.D	28 Oct 2024 16:11	RC/JU	Ok
17	P4574-06	BF140090.D	28 Oct 2024 16:39	RC/JU	Ok
18	P4545-02	BF140091.D	28 Oct 2024 17:07	RC/JU	Ok
19	P4545-02MS	BF140092.D	28 Oct 2024 17:35	RC/JU	Ok,M
20	P4545-02MSD	BF140093.D	28 Oct 2024 18:03	RC/JU	Ok,M
21	P4567-01	BF140094.D	28 Oct 2024 18:31	RC/JU	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF102824

Review By	yogesh	Review On	10/29/2024 1:35:27 AM		
Supervise By	mohammad	Supervise On	10/29/2024 1:38:39 AM		
SubDirectory	BF102824	HP Acquire Method	BNA_F	HP Processing Method	bf101824

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12323,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	P4574-04	BF140095.D	28 Oct 2024 18:59	RC/JU	Ok
23	P4577-01	BF140096.D	28 Oct 2024 19:27	RC/JU	Ok,NS
24	P4566-01	BF140097.D	28 Oct 2024 19:55	RC/JU	Ok,NS
25	P4575-01	BF140098.D	28 Oct 2024 20:23	RC/JU	Ok,NS

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF101824

Review By	yogesh	Review On	10/21/2024 6:34:01 AM		
Supervise By	mohammad	Supervise On	10/21/2024 6:38:35 AM		
SubDirectory	BF101824	HP Acquire Method	BNA_F	HP Processing Method	bf101824

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12322,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF139843.D	18 Oct 2024 09:22		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF139844.D	18 Oct 2024 10:27		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF139845.D	18 Oct 2024 10:55	Compound #9,32,54,65,85 removed from 5 ppm	RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF139846.D	18 Oct 2024 11:23		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF139847.D	18 Oct 2024 11:52	Compound #54 Kept on LR	RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF139848.D	18 Oct 2024 12:20	The Calibration is Good For 8270 DOD Except com#77 and good for 625.1 Method	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BF139849.D	18 Oct 2024 12:49	Com#77(Benzidine) Failed in the calibration for both DOD and NON-DOD	RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BF139850.D	18 Oct 2024 13:17		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BF139851.D	18 Oct 2024 13:46		RC/JU	Ok,M
10	SSTDICV040	SSTDICV040	BF139852.D	18 Oct 2024 14:19		RC/JU	Ok,M
11	PB164211BL	PB164211BL	BF139853.D	18 Oct 2024 14:48		RC/JU	Ok
12	P4405-01	MH-121	BF139854.D	18 Oct 2024 15:21		RC/JU	Ok
13	P4431-01	72-11934	BF139855.D	18 Oct 2024 15:50		RC/JU	Ok,M
14	P4421-01	EO-02-101624	BF139856.D	18 Oct 2024 16:18		RC/JU	Ok,M
15	P4422-01	EO-01-101624	BF139857.D	18 Oct 2024 16:46		RC/JU	Ok
16	P4425-01	TP-1	BF139858.D	18 Oct 2024 17:15		RC/JU	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF101824

Review By	yogesh	Review On	10/21/2024 6:34:01 AM		
Supervise By	mohammad	Supervise On	10/21/2024 6:38:35 AM		
SubDirectory	BF101824	HP Acquire Method	BNA_F	HP Processing Method	bf101824

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12322,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

17	P4425-03	TP-2	BF139859.D	18 Oct 2024 17:44		RC/JU	Ok
18	P4425-05	TP-3	BF139860.D	18 Oct 2024 18:12		RC/JU	Ok
19	P4425-07	TP-4	BF139861.D	18 Oct 2024 18:41	Internal Standrad Fail	RC/JU	ReRun
20	P4425-09	TP-5	BF139862.D	18 Oct 2024 19:09	Internal Standrad Fail	RC/JU	ReRun
21	P4426-03	PAD-2	BF139863.D	18 Oct 2024 19:37	Internal Standrad Fail	RC/JU	ReRun
22	P4426-07	PAD-4	BF139864.D	18 Oct 2024 20:05	Internal Standrad Fail	RC/JU	ReRun
23	P4426-17	PAD-9	BF139865.D	18 Oct 2024 20:34	Internal Standard Fail	RC/JU	ReRun
24	P4426-11	PAD-6	BF139866.D	18 Oct 2024 21:02	Internal Standard Fail	RC/JU	ReRun

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF102824

Review By	yogesh	Review On	10/29/2024 1:35:27 AM		
Supervise By	mohammad	Supervise On	10/29/2024 1:38:39 AM		
SubDirectory	BF102824	HP Acquire Method	BNA_F	HP Processing Method	bf101824

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12323,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140074.D	28 Oct 2024 09:04		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140075.D	28 Oct 2024 09:32		RC/JU	Ok,M
3	PB164444BL	PB164444BL	BF140076.D	28 Oct 2024 10:01		RC/JU	Ok
4	PB164444BS	PB164444BS	BF140077.D	28 Oct 2024 10:29		RC/JU	Ok,M
5	PB164393TB	PB164393TB	BF140078.D	28 Oct 2024 10:58		RC/JU	Ok
6	PB164444BL	PB164444BL	BF140079.D	28 Oct 2024 11:31	Already analysed with ok status	RC/JU	Not Ok
7	P4547-04	BP-F-21	BF140080.D	28 Oct 2024 11:59		RC/JU	Ok
8	P4547-08	BP-F-20	BF140081.D	28 Oct 2024 12:27		RC/JU	Ok
9	P4561-04	BP-F-19	BF140082.D	28 Oct 2024 12:55		RC/JU	Ok
10	P4561-08	BP-F-18	BF140083.D	28 Oct 2024 13:23		RC/JU	Ok
11	P4578-06	WB-305-BOT	BF140084.D	28 Oct 2024 13:51		RC/JU	Ok
12	P4578-08	WB-305-BOT-1	BF140085.D	28 Oct 2024 14:19		RC/JU	Ok
13	P4567-04	WC-1	BF140086.D	28 Oct 2024 14:47		RC/JU	Ok
14	P4567-08	WC-2	BF140087.D	28 Oct 2024 15:15		RC/JU	Ok
15	P4567-12	WC-3	BF140088.D	28 Oct 2024 15:43		RC/JU	Ok
16	P4574-03	GRAVEL-1	BF140089.D	28 Oct 2024 16:11		RC/JU	Ok
17	P4574-06	GRAVEL-2	BF140090.D	28 Oct 2024 16:39		RC/JU	Ok
18	P4545-02	VNJ-215	BF140091.D	28 Oct 2024 17:07		RC/JU	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QCBatch ID # BF102824

Review By	yogesh	Review On	10/29/2024 1:35:27 AM		
Supervise By	mohammad	Supervise On	10/29/2024 1:38:39 AM		
SubDirectory	BF102824	HP Acquire Method	BNA_F	HP Processing Method	bf101824

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12323,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	P4545-02MS	VNJ-215MS	BF140092.D	28 Oct 2024 17:35		RC/JU	Ok,M
20	P4545-02MSD	VNJ-215MSD	BF140093.D	28 Oct 2024 18:03		RC/JU	Ok,M
21	P4567-01	WC-1	BF140094.D	28 Oct 2024 18:31		RC/JU	Ok
22	P4574-04	GRAVEL-2	BF140095.D	28 Oct 2024 18:59		RC/JU	Ok
23	P4577-01	TR-05-102524	BF140096.D	28 Oct 2024 19:27		RC/JU	Ok,NS
24	P4566-01	HD-01-102524	BF140097.D	28 Oct 2024 19:55		RC/JU	Ok,NS
25	P4575-01	PL-02-102424	BF140098.D	28 Oct 2024 20:23		RC/JU	Ok,NS

M : Manual Integration

Alliance TECHNICAL GROUP

TCLP EXTRACTION LOGPAGE

PB164393

SOP ID : M1311-TCLP-15

SDG No : N/A

Weigh By : JP

Balance ID : WC SC-4

pH Meter ID : WC PH METER-1

Extraction By : JP

Filter By : JP

Pippete ID : WC

Tumbler ID : T-1 / T-2

TCLP Filter ID : 114771

Start Prep Date : 10/25/2024 **Time :** 17:40

End Prep Date : 10/26/2024 **Time :** 10:37

Combination Ratio : 20

ZHE Cleaning Batch : N/A

Initial Room Temperature: 23 °C

Final Room Temperature: 22 °C

TCLP Technician Signature : JP

Supervisor By : 12

Standared Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
TCLP-FLUID-1	N/A	WP108622
HCL-TCLP,1N	N/A	WP108584
HNO3-TCLP,1N	N/A	WP108585
pH Strips	N/A	W1931,W1934,W2350,W2755
pH Strips	N/A	W1937,W1938,W1939,W1940,W1941,W1942
1 Liter Amber	N/A	23091
120ml Plastic bottle	N/A	21029
1:1 HNO3	MP81119	N/A

Extraction Conformance/Non-Conformance Comments:

Matrix spikes are added after fiftration and before preservation. Tumbler T-1 / T-2 CHECKED,30 RPM. Particle size reduction is not required. p4578-08 is used for MS-MSD.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
10/26/24 11:45	JP TCLP Room	JP [illegible]
	Preparation Group	Analysis Group

Alliance TECHNICAL GROUP

TCLP EXTRACTION LOGPAGE

PB164393

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
P4452-03	ETGI-285	01	100.03	2000	N/A	N/A	N/A	11.0	1.5	T-1
P4545-02	VNJ-215	02	100.02	2000	N/A	N/A	N/A	5.8	1.5	T-1
P4547-04	BP-F-21	03	100.01	2000	N/A	N/A	N/A	4.5	1.0	T-1
P4547-08	BP-F-20	04	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-1
P4561-04	BP-F-19	05	100.03	2000	N/A	N/A	N/A	5.6	1.5	T-1
P4561-08	BP-F-18	06	100.04	2000	N/A	N/A	N/A	4.0	1.0	T-1
P4567-04	WC-1	07	100.02	2000	N/A	N/A	N/A	7.2	1.5	T-1
P4567-08	WC-2	08	100.03	2000	N/A	N/A	N/A	7.6	1.0	T-1
P4567-12	WC-3	09	100.04	2000	N/A	N/A	N/A	7.6	1.5	T-1
P4574-03	GRAVEL-1	10	100.02	2000	N/A	N/A	N/A	4.5	1.0	T-1
P4574-06	GRAVEL-2	11	100.03	2000	N/A	N/A	N/A	3.5	1.5	T-2
P4574-08	CONCRETE-1	12	100.02	2000	N/A	N/A	N/A	10.5	1.0	T-2
P4574-10	CONCRETE-2	13	100.02	2000	N/A	N/A	N/A	10.5	1.5	T-2
P4574-12	CONCRETE-3	14	100.03	2000	N/A	N/A	N/A	11.0	1.0	T-2
P4578-06	WB-305-BOT	15	100.04	2000	N/A	N/A	N/A	5.8	1.5	T-2
P4578-08	WB-305-BOT-1	16	100.02	2000	N/A	N/A	N/A	6.0	1.5	T-2
PB164393TB	LEB393	17	N/A	2000	N/A	N/A	N/A	4.93	1.5	T-2

SampleID	ClientID	Sample Weight (g)	Filter Weight (g)	Filtrate (mL)	Filter + Solid (After 100°C)	% solids	% Dry Solids
P4452-03	ETGI-285	N/A	N/A	N/A	N/A	100	N/A
P4545-02	VNJ-215	N/A	N/A	N/A	N/A	100	N/A
P4547-04	BP-F-21	N/A	N/A	N/A	N/A	100	N/A
P4547-08	BP-F-20	N/A	N/A	N/A	N/A	100	N/A
P4561-04	BP-F-19	N/A	N/A	N/A	N/A	100	N/A
P4561-08	BP-F-18	N/A	N/A	N/A	N/A	100	N/A
P4567-04	WC-1	N/A	N/A	N/A	N/A	100	N/A
P4567-08	WC-2	N/A	N/A	N/A	N/A	100	N/A
P4567-12	WC-3	N/A	N/A	N/A	N/A	100	N/A
P4574-03	GRAVEL-1	N/A	N/A	N/A	N/A	100	N/A
P4574-06	GRAVEL-2	N/A	N/A	N/A	N/A	100	N/A
P4574-08	CONCRETE-1	N/A	N/A	N/A	N/A	100	N/A
P4574-10	CONCRETE-2	N/A	N/A	N/A	N/A	100	N/A
P4574-12	CONCRETE-3	N/A	N/A	N/A	N/A	100	N/A
P4578-06	WB-305-BOT	N/A	N/A	N/A	N/A	100	N/A
P4578-08	WB-305-BOT-1	N/A	N/A	N/A	N/A	100	N/A
PB164393TB	LEB393	N/A	N/A	N/A	N/A	N/A	N/A

Alliance TECHNICAL GROUP

TCLP Fluid Determination

PB164393

Hot Block ID : WC S-1 /WC S-2

Thermometer ID : FLASHPOINT

SampleID	ClientID	Sample Weight (g)	Volume DI Water (mL)	PH after 5 min stir	PH after 10 min stir	Extraction Fluid 1 or 2	pH Extraction Fluid
P4452-03	ETGI-285	5.02	96.5	11.0	4.0	#1	4.93
P4545-02	VNJ-215	5.02	96.5	7.6	2.5	#1	4.93
P4547-04	BP-F-21	5.00	96.5	6.2	2.0	#1	4.93
P4547-08	BP-F-20	5.03	96.5	8.2	3.0	#1	4.93
P4561-04	BP-F-19	5.02	96.5	7.6	2.5	#1	4.93
P4561-08	BP-F-18	5.02	96.5	7.0	2.5	#1	4.93
P4567-04	WC-1	5.03	96.5	9.1	4.0	#1	4.93
P4567-08	WC-2	5.02	96.5	9.5	4.0	#1	4.93
P4567-12	WC-3	5.02	96.5	7.2	3.0	#1	4.93
P4574-03	GRAVEL-1	5.02	96.5	6.0	2.0	#1	4.93
P4574-06	GRAVEL-2	5.03	96.5	5.6	1.5	#1	4.93
P4574-08	CONCRETE-1	5.02	96.5	11.0	4.0	#1	4.93
P4574-10	CONCRETE-2	5.03	96.5	11.0	4.0	#1	4.93
P4574-12	CONCRETE-3	5.02	96.5	11.5	4.5	#1	4.93
P4578-06	WB-305-BOT	5.03	96.5	7.6	2.5	#1	4.93
P4578-08	WB-305-BOT-1	5.02	96.5	8.2	3.0	#1	4.93
PB164393TB	LEB393	N/A	N/A	N/A	N/A	#1	4.93

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tclp p4545 **WorkList ID :** 184760 **Department :** TCLP Extraction **Date :** 10-25-2024 07:58:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4452-03	ETGI-285	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K51	10/18/2024	1311
P4545-02	VNJ-215	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K61	10/24/2024	1311
P4547-04	BP-F-21	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K51	10/24/2024	1311
P4547-08	BP-F-20	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K51	10/24/2024	1311
P4561-04	BP-F-19	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K63	10/25/2024	1311
P4561-08	BP-F-18	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K63	10/25/2024	1311
P4567-04	WC-1	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K61	10/25/2024	1311
P4567-08	WC-2	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K61	10/25/2024	1311
P4567-12	WC-3	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K61	10/25/2024	1311
P4574-03	GRAVEL-1	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K63	10/25/2024	1311
P4574-06	GRAVEL-2	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K63	10/25/2024	1311
P4574-08	CONCRETE-1	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K63	10/25/2024	1311
P4574-10	CONCRETE-2	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K63	10/25/2024	1311
P4574-12	CONCRETE-3	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	K63	10/25/2024	1311
P4578-06	WB-305-BOT	Solid	TCLP Extraction	Cool 4 deg C	PORT06	K63	10/25/2024	1311
P4578-08	WB-305-BOT-1	Solid	TCLP Extraction	Cool 4 deg C	PORT06	K63	10/25/2024	1311

Date/Time 10/25/24 17:25
Raw Sample Received by: JD (WOC)
Raw Sample Relinquished by: J.C (SM)

Date/Time 10/25/24 19:30
Raw Sample Received by: J.C (SM)
Raw Sample Relinquished by: JD WOC

Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PB164444

SOP ID: M3541-ASE Extraction-14

Clean Up SOP #: N/A

Matrix : Water

Weigh By: N/A

Balance check: N/A

Balance ID: N/A

pH Strip Lot#: N/A

Extraction By: RS

Filter By: RS

pH Meter ID: N/A

Hood ID: 4,6,7

Extraction Start Date : 10/26/2024

Extraction Start Time : 12:10

Extraction End Date : 10/26/2024

Extraction End Time : 17:10

Concentration By: EH

Supervisor By : rajesh

Extraction Method: [√] Seperatory Funnel [ ] Continious Liquid/Liquid [ ] Sonication [ ] Waste Dilution [ ] Soxhlet

Standared Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	50/100 PPM	SP6630
Surrogate	1.0ML	100/150 PPM	SP6638
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3822
Baked Na2SO4	N/A	EP2551
10N NaoH	N/A	EP2550
H2SO4 1:1	N/A	EP2548
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

1.5 ML Vial lot# 2210673. pH Adjusted<2 with 1:1 H2SO4 & >11with 10 N NaOH.

KD Bath ID: Water bath -01,02

KD Bath Temperature: 60 °C

Envap ID: NEVAP-02

Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
10|26|24	RP (Ext-Lab)	RC/SVOC
17:15	Preparation Group	Analysis Group

Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PrepBatch ID : PB164444

Analytical Method: M3541-ASE Extraction-14 **Concentration Date:** 10/26/2024

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB164393TB	PB164393TB	TCLP BNA	100	6	RUPESH	rajesh	1			SEP-01
PB164444BL	PB164444BL	TCLP BNA	1000	6	RUPESH	rajesh	1			2
PB164444BS	PB164444BS	TCLP BNA	1000	6	RUPESH	rajesh	1			3
P4545-02	VNJ-215	TCLP BNA	100	6	RUPESH	rajesh	1	A		4
P4545-02MS	VNJ-215MS	TCLP BNA	100	6	RUPESH	rajesh	1	A		5
P4545-02MSD	VNJ-215MSD	TCLP BNA	100	6	RUPESH	rajesh	1	A		6
P4547-04	BP-F-21	TCLP BNA	100	6	RUPESH	rajesh	1	A		7
P4547-08	BP-F-20	TCLP BNA	100	6	RUPESH	rajesh	1	A		8
P4561-04	BP-F-19	TCLP BNA	100	6	RUPESH	rajesh	1	A		9
P4561-08	BP-F-18	TCLP BNA	100	6	RUPESH	rajesh	1	A		10
P4567-04	WC-1	TCLP BNA	100	6	RUPESH	rajesh	1	A		11
P4567-08	WC-2	TCLP BNA	100	6	RUPESH	rajesh	1	A		12
P4567-12	WC-3	TCLP BNA	100	6	RUPESH	rajesh	1	A		13
P4574-03	GRAVEL-1	TCLP BNA	100	6	RUPESH	rajesh	1	A		14
P4574-06	GRAVEL-2	TCLP BNA	100	6	RUPESH	rajesh	1	A		15
P4578-06	WB-305-BOT	TCLP BNA	100	6	RUPESH	rajesh	1	A		16
P4578-08	WB-305-BOT-1	TCLP BNA	100	6	RUPESH	rajesh	1	A		17

* **Extracts relinguished on the same date as received.**

10/26/24

Alliance TECHNICAL GROUP

TCLP EXTRACTION LOGPAGE

PB164393

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
P4452-03	ETGI-285	01	100.03	2000	N/A	N/A	N/A	11.0	1.5	T-1
P4545-02	VNJ-215	02	100.02	2000	N/A	N/A	N/A	5.8	1.5	T-1
P4547-04	BP-F-21	03	100.01	2000	N/A	N/A	N/A	4.5	1.0	T-1
P4547-08	BP-F-20	04	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-1
P4561-04	BP-F-19	05	100.03	2000	N/A	N/A	N/A	5.6	1.5	T-1
P4561-08	BP-F-18	06	100.04	2000	N/A	N/A	N/A	4.0	1.0	T-1
P4567-04	WC-1	07	100.02	2000	N/A	N/A	N/A	7.2	1.5	T-1
P4567-08	WC-2	08	100.03	2000	N/A	N/A	N/A	7.6	1.0	T-1
P4567-12	WC-3	09	100.04	2000	N/A	N/A	N/A	7.6	1.5	T-1
P4574-03	GRAVEL-1	10	100.02	2000	N/A	N/A	N/A	4.5	1.0	T-1
P4574-06	GRAVEL-2	11	100.03	2000	N/A	N/A	N/A	3.5	1.5	T-2
P4574-08	CONCRETE-1	12	100.02	2000	N/A	N/A	N/A	10.5	1.0	T-2
P4574-10	CONCRETE-2	13	100.02	2000	N/A	N/A	N/A	10.5	1.5	T-2
P4574-12	CONCRETE-3	14	100.03	2000	N/A	N/A	N/A	11.0	1.0	T-2
P4578-06	WB-305-BOT	15	100.04	2000	N/A	N/A	N/A	5.8	1.5	T-2
P4578-08	WB-305-BOT-1	16	100.02	2000	N/A	N/A	N/A	6.0	1.5	T-2
PB164393TB	LEB393	17	N/A	2000	N/A	N/A	N/A	4.93	1.5	T-2

10/26/2024
11:50

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Prep Standard - Chemical Standard Summary

Order ID : P4578

Test : TCLP BNA

Prepbatch ID : PB164444,

Sequence ID/Qc Batch ID: BF102824,

Standard ID :
EP2548,EP2550,EP2551,SP6558,SP6559,SP6573,SP6620,SP6621,SP6622,SP6623,SP6624,SP6625,SP6626,SP6627,SP6628,SP6630,SP6638,

Chemical ID :
10ul/1000ul sample,E3551,E3657,E3768,E3788,E3791,E3815,E3822,M5178,S10103,S10247,S10393,S10583,S10977,S10978,S10979,S10980,S11003,S11004,S11005,S11006,S11007,S11008,S11009,S11010,S11085,S11098,S11138,S11141,S11159,S11494,S11541,S11566,S11649,S11766,S11771,S11772,S11773,S11774,S11775,S12033,S12078,S12097,S12105,S12106,S12107,S12108,S12109,S12110,S12111,S12113,S12117,S12126,S12127,S12128,S12129,S12130,S12131,S12132,S12133,S12134,S12187,S12188,S12189,S12207,S12275,S12314,S12323,S12453,S12454,S12455,S12456,S12457,S12458,S12459,S12460,S12461,S12509,S12510,W3112,

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Extractions STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
314	1.1 H2SO4 SOLN	EP2548	10/16/2024	04/16/2025	Rajesh Parikh	None	None	RUPESHKUMAR SHAH 10/16/2024

FROM 1000.00000ml of W3112 + 4000.00000ml of M5178 = Final Quantity: 2000.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1874	10 N SODIUM HYDROXIDE SOLN	EP2550	10/16/2024	04/16/2025	Rajesh Parikh	Extraction_SCALE_2 (EX-SC-2)	None	RUPESHKUMAR SHAH 10/16/2024

FROM 1000.00000ml of W3112 + 400.00000gram of E3657 = Final Quantity: 1000.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Extractions STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3923	Baked Sodium Sulfate	EP2551	10/18/2024	01/03/2025	Rajesh Parikh	Extraction_SCALE_2 (EX-SC-2)	None	RUPESHKUMAR SHAH 10/18/2024

FROM 4000.00000gram of E3551 = Final Quantity: 4000.000 gram

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
18	Second Source Calibration Stock Standard, 100 PPM, (8270/625/CLP)	SP6558	07/09/2024	11/30/2024	Jagrut Upadhyay	None	None	mohammad ahmed 07/11/2024

FROM 0.04000ml of S10977 + 0.08000ml of S11003 + 0.10000ml of S11766 + 0.20000ml of S11566 + 0.20000ml of S12097 + 0.20000ml of S12117 + 1.18000ml of E3768 = Final Quantity: 2.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
416	40 ng BNA ICV, 40 PPM	SP6559	07/09/2024	11/30/2024	Jagrut Upadhyay	None	None	mohammad ahmed 07/11/2024

FROM 0.01000ml of S12033 + 0.60000ml of E3768 + 0.40000ml of SP6558 = Final Quantity: 1.010 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3895	50 ug/ml DFTPP 8270E	SP6573	07/15/2024	01/08/2025	Rahul Chavli	None	None	Yogesh Patel 07/17/2024

FROM 1.00000ml of S10247 + 19.00000ml of E3768 = Final Quantity: 20.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3764	8270/625 Stock solution 100 ng	SP6620	09/11/2024	02/08/2025	Jagrut Upadhyay	None	None	mohammad ahmed 09/11/2024

FROM 0.26700ml of S10103 + 0.40000ml of S11494 + 0.50000ml of S12113 + 1.00000ml of S11085 + 1.00000ml of S11098 + 1.00000ml of S11159 + 1.00000ml of S12078 + 1.00000ml of S12275 + 3.83300ml of E3791 = Final Quantity: 10.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
413	80 ng BNA ICC, 80 PPM	SP6621	09/11/2024	02/08/2025	Jagrut Upadhyay	None	None	mohammad ahmed 09/11/2024

FROM 0.01000ml of S12314 + 0.20000ml of E3791 + 0.80000ml of SP6620 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
412	60 ng BNA ICC, 60 PPM	SP6622	09/11/2024	02/08/2025	Jagrut Upadhyay	None	None	mohammad ahmed 09/11/2024

FROM 0.01000ml of S12314 + 0.40000ml of E3791 + 0.60000ml of SP6620 = Final Quantity: 1.010 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
411	50 ng BNA ICC, 50 PPM	SP6623	09/11/2024	02/08/2025	Jagrut Upadhyay	None	None	mohammad ahmed 09/11/2024

FROM 0.01000ml of S12314 + 0.50000ml of E3791 + 0.50000ml of SP6620 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
410	40 ng BNA ICC, 40 PPM	SP6624	09/11/2024	02/08/2025	Jagrut Upadhyay	None	None	mohammad ahmed 09/11/2024

FROM 0.01000ml of S12314 + 0.60000ml of E3791 + 0.40000ml of SP6620 = Final Quantity: 1.010 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3678	20 ng BNA ICC, 20 PPM	SP6625	09/11/2024	02/08/2025	Jagrut Upadhyay	None	None	mohammad ahmed 09/11/2024

FROM 0.01000ml of S12314 + 0.80000ml of E3791 + 0.20000ml of SP6620 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
408	10 ng BNA ICC, 10 PPM	SP6626	09/11/2024	02/08/2025	Jagrut Upadhyay	None	None	mohammad ahmed 09/11/2024

FROM 0.01000ml of S12314 + 0.90000ml of E3791 + 0.10000ml of SP6620 = Final Quantity: 1.010 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
407	5 ng BNA ICC, 5 PPM	SP6627	09/11/2024	02/08/2025	Jagrut Upadhyay	None	None	mohammad ahmed 09/11/2024

FROM 0.01000ml of S12314 + 0.95000ml of E3791 + 0.05000ml of SP6620 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
175	2.5 ng BNA ICC, 2.5 PPM	SP6628	09/11/2024	02/08/2025	Jagrut Upadhyay	None	None	mohammad ahmed 09/11/2024

FROM 0.01000ml of S12314 + 0.50000ml of E3791 + 0.50000ml of SP6627 = Final Quantity: 1.010 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
171	8270/625 Spike Solution, 50/100 PPM	SP6630	09/13/2024	02/12/2025	Rahul Chavli	None	None	Yogesh Patel 10/29/2024

FROM 0.10000ml of S12134 + 0.10000ml of S12461 + 0.20000ml of S12453 + 0.30000ml of S11775 + 0.30000ml of S12510 + 0.40000ml of S10393 + 0.40000ml of S10583 + 0.40000ml of S11138 + 0.40000ml of S11141 + 0.40000ml of S11649 + 0.70000ml of S11771 + 0.70000ml of S12105 + 0.80000ml of S12126 + 1.10000ml of S12110 + 1.10000ml of S12133 + 1.10000ml of S12460 + 1.20000ml of S12455 + 1.20000ml of S12456 + 1.20000ml of S12457 + 1.20000ml of S12458 + 1.20000ml of S12459 + 1.20000ml of S12509 + 1.30000ml of S11541 + 1.30000ml of S11772 + 1.30000ml of S11773 + 1.30000ml of S12106 + 1.30000ml of S12108 + 1.30000ml of S12111 + 1.30000ml of S12127 + 1.30000ml of S12129 + 1.30000ml of S12130 + 1.30000ml of S12131 + 1.30000ml of S12454 + 1.40000ml of S11774 + 1.40000ml of S12107 + 1.40000ml of S12109 + 1.40000ml of S12128 + 1.40000ml of S12132 + 163.00000ml of E3788 = Final Quantity: 200.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
19	8270/CLP Surrogate Solution, 100 PPM BN/150 PPM ACID	SP6638	10/10/2024	04/04/2025	Jagrut Upadhyay	None	None	mohammad ahmed 10/18/2024

FROM 1930.00000ml of E3815 + 5.00000ml of S10978 + 5.00000ml of S10979 + 5.00000ml of S10980 + 5.00000ml of S11004 + 5.00000ml of S11005 + 5.00000ml of S11006 + 5.00000ml of S11007 + 5.00000ml of S11008 + 5.00000ml of S11009 + 5.00000ml of S11010 + 5.00000ml of S12187 + 5.00000ml of S12188 + 5.00000ml of S12189 + 5.00000ml of S12207 = Final Quantity: 2000.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	01/03/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19510-5 / Sodium Hydroxide Pellets 2.5 Kg, Pk of 4	23B1556310	12/31/2025	12/04/2023 / Rajesh	12/01/2023 / Rajesh	E3657

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24E2462004	01/08/2025	07/08/2024 / Rajesh	06/21/2024 / Rajesh	E3768

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	23H1462005	04/23/2025	08/13/2024 / Rajesh	08/13/2024 / Rajesh	E3788

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24G2362009	03/09/2025	09/09/2024 / Rajesh	09/03/2024 / Rajesh	E3791

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	04/04/2025	10/04/2024 / Rajesh	10/04/2024 / Rajesh	E3815

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24I2662006	04/23/2025	10/24/2024 / Rajesh	10/24/2024 / Rajesh	E3822

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	0000281827	03/29/2026	05/25/2022 / william	04/05/2022 / william	M5178

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-112090-04 / CLP Acid Surrogate Solution, 7500 mg/L, 1ml	440246	02/08/2025	08/08/2024 / Jagrut	12/09/2021 / Christian	S10103

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31615 / SV Mixture, GC/MS Tuning Mixture, CH2Cl2, 1mL,	A0182667	01/15/2025	07/15/2024 / Rahul	03/18/2022 / Christian	S10247

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555871 / Custom Standard, 4-nitrophenol Std [CS 5238-4]	A0185300	03/13/2025	09/13/2024 / yogesh	05/18/2022 / Christian	S10393

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555868 / Custom Standard, Benzidine Std [CS 5328-1]	A0186373	02/12/2025	08/12/2024 / Rahul	07/05/2022 / Christian	S10583

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0188108	11/30/2024	05/31/2024 / Jagrut	12/28/2022 / Christian	S10977

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0188108	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S10978

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0188108	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S10979

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0188108	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S10980

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	11/30/2024	05/31/2024 / Jagrut	12/28/2022 / Christian	S11003

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S11004

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S11005

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S11006

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S11007

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S11008

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S11009

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0189418	04/10/2025	10/10/2024 / anahy	12/28/2022 / Christian	S11010

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-010074-07 / 3,3'-Dichlorobenzidine Solution, 1,000 mg/L, 1 ml, (Maximum Expiration: 180 days)	406703	03/11/2025	09/11/2024 / Rahul	02/07/2023 / Christian	S11085

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-110381-01 / 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1ml	495831	03/11/2025	09/11/2024 / Rahul	02/07/2023 / Christian	S11098

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555870 / Custom Standard, 2,4-dinitrophenol Std [CS 5328-3]	A0194698	02/12/2025	08/12/2024 / Rahul	02/20/2023 / Christian	S11138

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555869 / Custom Standard, hexachlorocyclopentadiene Std [CS 5328-2]	A0194702	02/12/2025	08/12/2024 / Rahul	02/20/2023 / Christian	S11141

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110817-01 / Custom 8270 Mix, 4-55, 1000 mg/L, 1 ml, (Maximum Expiration: 90 Days)	414125	03/11/2025	09/11/2024 / Rahul	03/06/2023 / Christian	S11159

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110094-02 / CLP Base/Neutral Surrogate Solution, 5000 mg/L, 1ml	506889	02/08/2025	08/08/2024 / Jagrut	08/11/2023 / Yogesh	S11494

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0201940	03/13/2025	09/13/2024 / yogesh	09/18/2023 / Kiran	S11541

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0201940	12/05/2024	06/05/2024 / Rahul	09/18/2023 / Kiran	S11566

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555872 / Custom Standard, pentachlorophenol Std [CS 5328-5]	A0201728	03/13/2025	09/13/2024 / anahy	11/09/2023 / Yogesh	S11649

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	12/14/2024	06/14/2024 / Rahul	11/21/2023 / Rahul	S11766

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	02/20/2025	08/20/2024 / Rahul	11/21/2023 / Rahul	S11771

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	03/13/2025	09/13/2024 / anahy	11/21/2023 / Rahul	S11772

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	03/13/2025	09/13/2024 / anahy	11/21/2023 / Rahul	S11773

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	03/13/2025	09/13/2024 / anahy	11/21/2023 / Rahul	S11774

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	03/13/2025	09/13/2024 / anahy	11/21/2023 / Rahul	S11775

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL	A0201320	01/01/2025	07/01/2024 / Rahul	12/21/2023 / Rahul	S12033

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110816-01 / Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL, (Maximum Expiration: 180 Days)	414127	03/11/2025	09/11/2024 / Rahul	01/31/2024 / Rahul	S12078

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0207706	01/09/2025	07/09/2024 / Jagrut	02/05/2024 / Rahul	S12097

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0207706	02/12/2025	08/12/2024 / Rahul	02/05/2024 / Rahul	S12105

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0207706	03/13/2025	09/13/2024 / anahy	02/05/2024 / Rahul	S12106

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0207706	03/13/2025	09/13/2024 / anahy	02/05/2024 / Rahul	S12107

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0207706	03/13/2025	09/13/2024 / anahy	02/05/2024 / Rahul	S12108

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0207706	03/13/2025	09/13/2024 / anahy	02/05/2024 / Rahul	S12109

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0207706	03/13/2025	09/13/2024 / anahy	02/05/2024 / Rahul	S12110

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0207706	03/13/2025	09/13/2024 / anahy	02/05/2024 / Rahul	S12111

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-010223-01 / 1,4-Dioxane Solution, 2,000mg/L, 1ml	454157	02/09/2025	08/09/2024 / Jagrut	03/08/2024 / Rahul	S12113

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0203726	12/05/2024	06/05/2024 / Rahul	03/15/2024 / Rahul	S12117

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0203726	02/12/2025	08/12/2024 / Rahul	03/15/2024 / Rahul	S12126

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0203726	03/13/2025	09/13/2024 / anahy	03/15/2024 / Rahul	S12127

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0203726	03/13/2025	09/13/2024 / anahy	03/15/2024 / Rahul	S12128

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0203726	03/13/2025	09/13/2024 / anahy	03/15/2024 / Rahul	S12129

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0203726	03/13/2025	09/13/2024 / anahy	03/15/2024 / Rahul	S12130

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0203726	03/13/2025	09/13/2024 / anahy	03/15/2024 / Rahul	S12131

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0203726	03/13/2025	09/13/2024 / anahy	03/15/2024 / Rahul	S12132

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0203726	03/13/2025	09/13/2024 / anahy	03/15/2024 / Rahul	S12133

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0203726	03/13/2025	09/13/2024 / anahy	03/15/2024 / Rahul	S12134

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0206206	04/10/2025	10/10/2024 / anahy	03/15/2024 / Rahul	S12187

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0206206	04/10/2025	10/10/2024 / anahy	03/15/2024 / Rahul	S12188

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0206206	04/10/2025	10/10/2024 / anahy	03/15/2024 / Rahul	S12189

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0206381	04/10/2025	10/10/2024 / anahy	03/15/2024 / Rahul	S12207

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-010442-07 / Benzaldehyde Solution, 1000 mg/L, 1.3 ml, (Maximum Expiration: 90 Days)	495833	03/11/2025	09/11/2024 / Rahul	05/24/2024 / Rahul	S12275

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL	A0206540	03/04/2025	09/04/2024 / anahy	05/30/2024 / Rahul	S12314

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL	A0206540	04/23/2025	10/23/2024 / anahy	05/30/2024 / Rahul	S12323

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	02/12/2025	08/12/2024 / Rahul	07/23/2024 / RAHUL	S12453

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	03/13/2025	09/13/2024 / anahy	07/23/2024 / RAHUL	S12454

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	03/13/2025	09/13/2024 / anahy	07/23/2024 / RAHUL	S12455

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	03/13/2025	09/13/2024 / anahy	07/23/2024 / RAHUL	S12456

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	03/13/2025	09/13/2024 / anahy	07/23/2024 / RAHUL	S12457

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	03/13/2025	09/13/2024 / anahy	07/23/2024 / RAHUL	S12458

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	03/13/2025	09/13/2024 / anahy	07/23/2024 / RAHUL	S12459

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	03/13/2025	09/13/2024 / anahy	07/23/2024 / RAHUL	S12460

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	03/13/2025	09/13/2024 / anahy	07/23/2024 / RAHUL	S12461

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	03/13/2025	09/13/2024 / anahy	07/23/2024 / RAHUL	S12509

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	03/13/2025	09/13/2024 / anahy	07/23/2024 / RAHUL	S12510

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-010074-07	406703	≤ -10 °C	Methylene Chloride	3/30/2025	3,3'-Dichlorobenzidine Solution, 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
3,3'-dichlorobenzidine	91-94-1	99.5	74.3.26P	989 ± 7.53

Received on
02/07/23
by
CG

S11084
to
S11088

*Not a certified value

Certified By: _____
Jacob Mulloy
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricllly.

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110817-01	414125	≤ -10 °C	Methylene Chloride	6/21/2025	Custom 8270 Mix, 4-55, 1000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
acetophenone	98-86-2	99.2	85.8.1P	998 ± 11.5
benzoic acid	65-85-0	100	123.7.1P	1010 ± 5.88
biphenyl	92-52-4	99.9	366.29.1P	999 ± 5.82
1,2,4,5-tetrachlorobenzene	95-94-3	99.7	53.7.2P	993 ± 5.79

Received on
02/07/23
by
CG
S11089
to
S11093

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: _____
Shane Overcash
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-112090-04	440246	≤ -10 °C	Methylene Chloride	2/16/2026	CLP Acid Surrogate Solution, 7,500 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
2-chlorophenol-d_4	93951-73-6	99.3	248.12.7P	7487 ± 17.2
2-fluorophenol	367-12-4	99.8	10.7.3.3P	7513 ± 17.26
phenol-d6	13127-88-3	99.9	949.120.8P	7481 ± 17.19
2,4,6-tribromophenol	118-79-6	99.8	12.1.6P	7469 ± 17.17

Received on
02/25/21
by
CG

S9236
to
S9240

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: Erica Castiglione

Erica Castiglione
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Received on 02/07/23 by CG

~~S11097~~ S11096 to S11099

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110381-01	495831	≤ -10 °C	Methylene Chloride	10/30/2027	Method 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
acenaphthene	83-32-9	99.9	13.1.5P	1003 ± 17.27
acenaphthylene	208-96-8	97.6	14.290.1P	999.8 ± 17.22
aniline	62-53-3	99.9	64.7.1P	995 ± 17.13
anthracene	120-12-7	99.5	15.7.1P	1001 ± 17.24
azobenzene	103-33-3	98.1	252.7.2P	999.1 ± 17.21
benzo[a]anthracene	56-55-3	100	16.7.3P	1001 ± 17.24
benzo[b]fluoranthene	205-99-2	99.8	17.421.3P	1001 ± 19.91
benzo[k]fluoranthene	207-08-9	98.9	18.421.4P	1001 ± 17.92
benzo[ghi]perylene	191-24-2	93	19.286.4P	999.6 ± 19.88
benzo[a]pyrene	50-32-8	97	20.286.2P	999.1 ± 26.35
benzyl alcohol	100-51-6	99.9	65.18.1P	1001 ± 17.24
bis(2-chloroethoxy)methane	111-91-1	99.1	31.3.15P	999.7 ± 17.89
bis(2-chloroethyl)ether	111-44-4	99.8	32.7.1P	1001 ± 17.23
bis(2-chloro-1-methylethyl) ether	108-60-1	99.5	34.3.13P	999.5 ± 17.89
bis(2-ethylhexyl)adipate	103-23-1	99.5	874.7.1P	999.5 ± 17.21
bis(2-ethylhexyl)phthalate	117-81-7	99.4	33.29.1P	998.8 ± 19.86
4-bromophenyl phenyl ether	101-55-3	99.4	35.7.1P	999.1 ± 17.2
butyl benzyl phthalate	85-68-7	98.4	36.1.6P	984.7 ± 19.58
carbazole	86-74-8	99.4	239.7.2P	1000 ± 17.22

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: Briana Smith

Briana Smith
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

Certificate of Analysis

Catalog No.: Z-110381-01 **Lot No.:** 495831 **Expiration Date:** 10/30/2027

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2,4-trichlorobenzene	120-82-1	99.6	54.29.1P	1000 ± 17.22
2,4,5-trichlorophenol	95-95-4	96.5	121.7.1.1P	1000 ± 17.22
2,4,6-trichlorophenol	88-06-2	99.6	113.7.1P	1002 ± 17.25

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: Briana Smith

Briana Smith
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

RESTEK®
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received on 03/08/22 by CG S10242 to S10247

Catalog No. : 31615 **Lot No.:** A0182667

Description : GC/MS Tuning Mixture

GC/MS Tuning Mixture 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2025 **Storage:** 10°C or colder

Handling: Contains carcinogen/reproductive toxin. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Pentachlorophenol **CAS #** 87-86-5 (Lot 211229RSR) **Purity** 99%	1,003.6 µg/mL	+/- 5.8897 µg/mL +/- 45.7132 µg/mL +/- 66.0037 µg/mL	Gravimetric Unstressed Stressed
2	DFTPP (Decafluorotriphenylphosphine) **CAS #** 5074-71-5 (Lot Q117-147) **Purity** 95%	1,006.6 µg/mL	+/- 5.9074 µg/mL +/- 45.8508 µg/mL +/- 66.2023 µg/mL	Gravimetric Unstressed Stressed
3	Benzidine **CAS #** 92-87-5 (Lot 211228JLM) **Purity** 99%	1,008.4 µg/mL	+/- 5.9179 µg/mL +/- 45.9318 µg/mL +/- 66.3193 µg/mL	Gravimetric Unstressed Stressed
4	4,4'-DDT **CAS #** 50-29-3 (Lot 210916JLM) **Purity** 99%	1,007.6 µg/mL	+/- 5.9132 µg/mL +/- 45.8954 µg/mL +/- 66.2667 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 08-Mar-2022 **Balance:** B345965662

Merlina Cowan - Operations Tech I

Date Passed: 10-Mar-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK®
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Gravimetric Certificate

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 05/19/22

S10393 to S10402

Catalog No. : 555871 **Lot No.:** A0185300

Description : Custom 4-Nitrophenol Standard

Custom 4-Nitrophenol Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2025 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	4-Nitrophenol	25,060.0 µg/mL	+/- 231.9100 µg/mL	Gravimetric
	CAS # 100-02-7 (Lot MKCN1089)		+/- 753.2622 µg/mL	Unstressed
	Purity 99%		+/- 905.6020 µg/mL	Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Katelyn McGinnis - Operations Tech I

Date Mixed: 16-May-2022 **Balance:** 1128342314

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k\sqrt{U^2_{gravimetric} + U^2_{homogeneity} + U^2_{storage\ stability} + U^2_{shipping\ stability}}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.

RESTEK®
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Gravimetric Certificate

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by
CG
on
07/05/22
S 10583
to
S 10592

Catalog No. : 555868 **Lot No.:** A0186373

Description : Custom Benzidine Standard

Custom Benzidine Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : June 30, 2025 **Storage:** 10°C or colder

Handling: Contains carcinogen/reproductive toxin. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Benzidine	25,200.0 µg/mL	+/- 233.2055 µg/mL	Gravimetric
	CAS # 92-87-5 (Lot 220511RSR)		+/- 351.6606 µg/mL	Unstressed
	Purity 99%		+/- 512.6054 µg/mL	Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Tom Suckar - Mix Technician **Date Mixed:** 16-Jun-2022 **Balance:** 1122030677

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k\sqrt{U^2_{gravimetric} + U^2_{homogeneity} + U^2_{storage\ stability} + U^2_{shipping\ stability}}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 S10951 to S10980

Catalog No. : 31087 **Lot No.:** A0188108

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000µg/mL, Methanol, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2030 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound		Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	2-Fluorophenol **CAS #** 367-12-4 **Purity** 99%	(Lot STBF3761V)	10,088.5 µg/mL	+/- 58.6554 µg/mL +/- 294.4162 µg/mL +/- 357.2628 µg/mL		Gravimetric Unstressed Stressed
2	Phenol-d6 **CAS #** 13127-88-3 **Purity** 99%	(Lot PR-31262)	10,043.3 µg/mL	+/- 58.3923 µg/mL +/- 293.0957 µg/mL +/- 355.6603 µg/mL		Gravimetric Unstressed Stressed
3	2,4,6-Tribromophenol **CAS #** 118-79-6 **Purity** 99%	(Lot MKCJ7664)	10,010.0 µg/mL	+/- 58.1990 µg/mL +/- 292.1253 µg/mL +/- 354.4829 µg/mL		Gravimetric Unstressed Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022 **Balance:** 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 S10951 to S10980

Catalog No. : 31087 **Lot No.:** A0188108

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000µg/mL, Methanol, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2030 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound		Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	2-Fluorophenol **CAS #** 367-12-4 **Purity** 99%	(Lot STBF3761V)	10,088.5 µg/mL	+/- 58.6554 µg/mL +/- 294.4162 µg/mL +/- 357.2628 µg/mL		Gravimetric Unstressed Stressed
2	Phenol-d6 **CAS #** 13127-88-3 **Purity** 99%	(Lot PR-31262)	10,043.3 µg/mL	+/- 58.3923 µg/mL +/- 293.0957 µg/mL +/- 355.6603 µg/mL		Gravimetric Unstressed Stressed
3	2,4,6-Tribromophenol **CAS #** 118-79-6 **Purity** 99%	(Lot MKCJ7664)	10,010.0 µg/mL	+/- 58.1990 µg/mL +/- 292.1253 µg/mL +/- 354.4829 µg/mL		Gravimetric Unstressed Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022 **Balance:** 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 S10951 to S10980

Catalog No. : 31087 **Lot No.:** A0188108

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000µg/mL, Methanol, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2030 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound		Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	2-Fluorophenol **CAS #** 367-12-4 **Purity** 99%	(Lot STBF3761V)	10,088.5 µg/mL	+/- 58.6554 µg/mL +/- 294.4162 µg/mL +/- 357.2628 µg/mL		Gravimetric Unstressed Stressed
2	Phenol-d6 **CAS #** 13127-88-3 **Purity** 99%	(Lot PR-31262)	10,043.3 µg/mL	+/- 58.3923 µg/mL +/- 293.0957 µg/mL +/- 355.6603 µg/mL		Gravimetric Unstressed Stressed
3	2,4,6-Tribromophenol **CAS #** 118-79-6 **Purity** 99%	(Lot MKCJ7664)	10,010.0 µg/mL	+/- 58.1990 µg/mL +/- 292.1253 µg/mL +/- 354.4829 µg/mL		Gravimetric Unstressed Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022 **Balance:** 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 S10951 to S10980

Catalog No. : 31087 **Lot No.:** A0188108

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10,000µg/mL, Methanol, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2030 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound		Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	2-Fluorophenol **CAS #** 367-12-4 **Purity** 99%	(Lot STBF3761V)	10,088.5 µg/mL	+/- 58.6554 µg/mL +/- 294.4162 µg/mL +/- 357.2628 µg/mL		Gravimetric Unstressed Stressed
2	Phenol-d6 **CAS #** 13127-88-3 **Purity** 99%	(Lot PR-31262)	10,043.3 µg/mL	+/- 58.3923 µg/mL +/- 293.0957 µg/mL +/- 355.6603 µg/mL		Gravimetric Unstressed Stressed
3	2,4,6-Tribromophenol **CAS #** 118-79-6 **Purity** 99%	(Lot MKCJ7664)	10,010.0 µg/mL	+/- 58.1990 µg/mL +/- 292.1253 µg/mL +/- 354.4829 µg/mL		Gravimetric Unstressed Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022 **Balance:** 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 $10981 to $11010

Catalog No. : 31086 **Lot No.:** A0189418

Description : B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2028 **Storage:** 10°C or colder

Handling: Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Nitrobenzene-d5 **CAS #** 4165-60-0 (Lot PR-29940A) **Purity** 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL +/- 225.6421 µg/mL +/- 250.3778 µg/mL	Gravimetric Unstressed Stressed
2	2-Fluorobiphenyl **CAS #** 321-60-8 (Lot 00021384) **Purity** 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL +/- 226.4003 µg/mL +/- 251.2191 µg/mL	Gravimetric Unstressed Stressed
3	p-Terphenyl-d14 **CAS #** 1718-51-0 (Lot PR-30504) **Purity** 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL +/- 226.4304 µg/mL +/- 251.2524 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 **Balance:** 1128353505

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 $10981 to $11010

Catalog No. : 31086 **Lot No.:** A0189418

Description : B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2028 **Storage:** 10°C or colder

Handling: Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Nitrobenzene-d5 **CAS #** 4165-60-0 (Lot PR-29940A) **Purity** 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL +/- 225.6421 µg/mL +/- 250.3778 µg/mL	Gravimetric Unstressed Stressed
2	2-Fluorobiphenyl **CAS #** 321-60-8 (Lot 00021384) **Purity** 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL +/- 226.4003 µg/mL +/- 251.2191 µg/mL	Gravimetric Unstressed Stressed
3	p-Terphenyl-d14 **CAS #** 1718-51-0 (Lot PR-30504) **Purity** 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL +/- 226.4304 µg/mL +/- 251.2524 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 **Balance:** 1128353505

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 $10981 to $11010

Catalog No. : 31086 **Lot No.:** A0189418

Description : B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2028 **Storage:** 10°C or colder

Handling: Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Nitrobenzene-d5 **CAS #** 4165-60-0 (Lot PR-29940A) **Purity** 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL +/- 225.6421 µg/mL +/- 250.3778 µg/mL	Gravimetric Unstressed Stressed
2	2-Fluorobiphenyl **CAS #** 321-60-8 (Lot 00021384) **Purity** 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL +/- 226.4003 µg/mL +/- 251.2191 µg/mL	Gravimetric Unstressed Stressed
3	p-Terphenyl-d14 **CAS #** 1718-51-0 (Lot PR-30504) **Purity** 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL +/- 226.4304 µg/mL +/- 251.2524 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 **Balance:** 1128353505

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 $10981 to $11010

Catalog No. : 31086 **Lot No.:** A0189418

Description : B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2028 **Storage:** 10°C or colder

Handling: Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Nitrobenzene-d5 **CAS #** 4165-60-0 (Lot PR-29940A) **Purity** 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL +/- 225.6421 µg/mL +/- 250.3778 µg/mL	Gravimetric Unstressed Stressed
2	2-Fluorobiphenyl **CAS #** 321-60-8 (Lot 00021384) **Purity** 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL +/- 226.4003 µg/mL +/- 251.2191 µg/mL	Gravimetric Unstressed Stressed
3	p-Terphenyl-d14 **CAS #** 1718-51-0 (Lot PR-30504) **Purity** 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL +/- 226.4304 µg/mL +/- 251.2524 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 **Balance:** 1128353505

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 $10981 to $11010

Catalog No. : 31086 **Lot No.:** A0189418

Description : B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2028 **Storage:** 10°C or colder

Handling: Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Nitrobenzene-d5 **CAS #** 4165-60-0 (Lot PR-29940A) **Purity** 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL +/- 225.6421 µg/mL +/- 250.3778 µg/mL	Gravimetric Unstressed Stressed
2	2-Fluorobiphenyl **CAS #** 321-60-8 (Lot 00021384) **Purity** 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL +/- 226.4003 µg/mL +/- 251.2191 µg/mL	Gravimetric Unstressed Stressed
3	p-Terphenyl-d14 **CAS #** 1718-51-0 (Lot PR-30504) **Purity** 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL +/- 226.4304 µg/mL +/- 251.2524 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 **Balance:** 1128353505

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 $10981 to $11010

Catalog No. : 31086 **Lot No.:** A0189418

Description : B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2028 **Storage:** 10°C or colder

Handling: Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Nitrobenzene-d5 **CAS #** 4165-60-0 (Lot PR-29940A) **Purity** 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL +/- 225.6421 µg/mL +/- 250.3778 µg/mL	Gravimetric Unstressed Stressed
2	2-Fluorobiphenyl **CAS #** 321-60-8 (Lot 00021384) **Purity** 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL +/- 226.4003 µg/mL +/- 251.2191 µg/mL	Gravimetric Unstressed Stressed
3	p-Terphenyl-d14 **CAS #** 1718-51-0 (Lot PR-30504) **Purity** 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL +/- 226.4304 µg/mL +/- 251.2524 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 **Balance:** 1128353505

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 $10981 to $11010

Catalog No. : 31086 **Lot No.:** A0189418

Description : B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2028 **Storage:** 10°C or colder

Handling: Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Nitrobenzene-d5 **CAS #** 4165-60-0 (Lot PR-29940A) **Purity** 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL +/- 225.6421 µg/mL +/- 250.3778 µg/mL	Gravimetric Unstressed Stressed
2	2-Fluorobiphenyl **CAS #** 321-60-8 (Lot 00021384) **Purity** 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL +/- 226.4003 µg/mL +/- 251.2191 µg/mL	Gravimetric Unstressed Stressed
3	p-Terphenyl-d14 **CAS #** 1718-51-0 (Lot PR-30504) **Purity** 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL +/- 226.4304 µg/mL +/- 251.2524 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 **Balance:** 1128353505

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 12/28/22 $10981 to $11010

Catalog No. : 31086 **Lot No.:** A0189418

Description : B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : August 31, 2028 **Storage:** 10°C or colder

Handling: Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Nitrobenzene-d5 **CAS #** 4165-60-0 (Lot PR-29940A) **Purity** 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL +/- 225.6421 µg/mL +/- 250.3778 µg/mL	Gravimetric Unstressed Stressed
2	2-Fluorobiphenyl **CAS #** 321-60-8 (Lot 00021384) **Purity** 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL +/- 226.4003 µg/mL +/- 251.2191 µg/mL	Gravimetric Unstressed Stressed
3	p-Terphenyl-d14 **CAS #** 1718-51-0 (Lot PR-30504) **Purity** 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL +/- 226.4304 µg/mL +/- 251.2524 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022 **Balance:** 1128353505

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555870 **Lot No.:** A0194698

Description : Custom 2,4-Dinitrophenol Standard

Custom 2,4-Dinitrophenol Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)
1	2,4-Dinitrophenol	51-28-5	DR221221RSR	99%	25,195.0 µg/mL

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer - Operations Technician I

Date Mixed: 15-Feb-2023 **Balance:** B442140311

Manufactured under Restek's Registered Quality Certificate #FM 8

tified Reference Material Notes

:es:

n date valid for unopened ampul stored in compliance with the recommended conditions.

ıty, concentration, and expiration of the CRM are based on the unopened product being stored according to the ɜnded condition found in the storage field.

d/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, LC/MS, RI, and/or melting point.

ıds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A n factor is used to calculate the amount of compound necessary to achieve the desired concentration of the ɔmpound in solution.

isomeric compounds is reported as the sum of the isomers.

lues are rounded to the nearest whole number.

rtainty Value Notes:

ɜrtainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded ıty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability ıty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

ɜrage factor of 2, which gives a level of confidence of approximately 95%.

ıged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure inimum packaged amount can be sufficiently transferred.

Notes:

tion is based upon gravimetric preparation using either a balance whose calibration has been verified daily Γ traceable weights, and/or dilutions with Class A glassware.

:

the unopened product, when stored in compliance with the recommended conditions, is guaranteed through ion displayed on the product label and certificate. Contact Restek for additional opened product stability ɳ, with the knowledge/understanding that open product stability is subject to the specific handling and ntal conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with ɟards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ɜm. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, ıdes complete instructions.

ssolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555869 **Lot No.:** A0194702

Description : Custom Hexachlorocyclopentadiene Standard

Custom Hexachlorocyclopentadiene Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)
1	Hexachlorocyclopentadiene	77-47-4	0012019	99%	25,008.0 µg/mL

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer - Operations Technician I

Date Mixed: 15-Feb-2023 **Balance:** B442140311

Manufactured under Restek Registered Quality Certificate #FM

tified Reference Material Notes

:es:

n date valid for unopened ampul stored in compliance with the recommended conditions.

nty, concentration, and expiration of the CRM are based on the unopened product being stored according to the ended condition found in the storage field.

d/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, LC/MS, RI, and/or melting point.

nds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A n factor is used to calculate the amount of compound necessary to achieve the desired concentration of the mpound in solution.

isomeric compounds is reported as the sum of the isomers.

lues are rounded to the nearest whole number.

rtainty Value Notes:

rtainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded ty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability ty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

erage factor of 2, which gives a level of confidence of approximately 95%.

ged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure nimum packaged amount can be sufficiently transferred.

Notes:

tion is based upon gravimetric preparation using either a balance whose calibration has been verified daily traceable weights, and/or dilutions with Class A glassware.

:

the unopened product, when stored in compliance with the recommended conditions, is guaranteed through ion displayed on the product label and certificate. Contact Restek for additional opened product stability ı, with the knowledge/understanding that open product stability is subject to the specific handling and ntal conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with ards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom m. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, des complete instructions.

ssolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely

PRODUCTOS QUIMICOS MONTERREY, S.A. DE C.V.

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na_2SO_4
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na_2SO_4)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO_4)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreing matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

RG-02-01, Ed. 3

Recd. by RB on 7/24/23 E3551

VWR®

Certificate of Analysis

Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:

Pellets

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025

Storage: Room Temperature

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Calcium	<= 0.005 %	<0.005 %	PASS
Chloride	<= 0.005 %	0.002 %	PASS
Heavy Metals	<= 0.002 %	<0.002 %	PASS
Iron	<= 0.001 %	<0.001 %	PASS
Magnesium	<= 0.002 %	<0.002 %	PASS
Mercury	<= 0.1 ppm	<0.1 ppm	PASS
Nickel	<= 0.001 %	<0.001 %	PASS
Nitrogen Compounds	<= 0.001 %	<0.001 %	PASS
Phosphate	<= 0.001 %	<0.001 %	PASS
Potassium	<= 0.02 %	<0.02 %	PASS
Purity	>= 97.0 %	99.2 %	PASS
Sodium Carbonate	<= 1.0 %	0.5 %	PASS
Sulfate	<= 0.003 %	<0.003 %	PASS

Internal ID #: 710

Signature

We certify that this batch conforms to the specifications listed.

This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.

E 3657	E 3659
E 3654	E 3660

PO: PO1-8886 PRODUCT CODE: SHIP DATE: 6/21/2024

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor™ | J.T.Baker

Material No.: 9266-A4
Batch No.: 24E2462004
Manufactured Date: 2024-04-10
Expiration Date: 2025-07-10
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	3
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	≥ 99.8 %	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Titrable Acid (µeq/g)	≤ 0.3	< 0.1
Chloride (Cl)	≤ 10 ppm	5 ppm
Water (by KF, coulometric)	≤ 0.02 %	< 0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC
Manufacturer source batch: MG24D10725

E 3768

JCroak
Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone 610.386.1700

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™ | J.T.Baker®

Material No.: 9254-03
Batch No.: 23H1462005
Manufactured Date: 2023-07-26
Expiration Date: 2026-07-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ($(CH_3)_2CO$) (by GC, corrected for water)	≥ 99.4 %	99.7 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titrable Acid (µeq/g)	≤ 0.3	0.1
Titrable Base (µeq/g)	≤ 0.6	< 0.1
Water (H_2O)	≤ 0.5 %	0.3 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	< 1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd. by RP on 8/13/24

E 3788

Ken Koehnlein
Sr. Manager, Quality Assurance

PO: PO2-76 PRODUCT CODE: SHIP DATE: 9/3/2024

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor™ | J.T.Baker®

Material No.: 9266-A4
Batch No.: 24G2362009
Manufactured Date: 2024-06-10
Expiration Date: 2025-09-09
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	2
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	≥ 99.8 %	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Titrable Acid (µeq/g)	≤ 0.3	< 0.1
Chloride (Cl)	≤ 10 ppm	< 5 ppm
Water (by KF, coulometric)	≤ 0.02 %	< 0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC
Manufacturer source batch: MG24F10024

E 3791

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone 610.386.1700

PO: PO2-329 PRODUCT CODE: SHIP DATE: 9/30/2024

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor | J.T.Baker

Material No.: 9254-03
Batch No.: 24H1462005
Manufactured Date: 2024-05-24
Expiration Date:2027-05-24
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ($(CH_3)_2CO$) (by GC, corrected forwater)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titrable Acid (µeq/g)	<= 0.3	0.2
Titrable Base (µeq/g)	<= 0.6	<0.1
Water (H_2O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3815

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials,LLC
100 Matsonford Rd, Suite 200,Radnor,PA,19087.U.S.A.Phone 610.386.1700

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor™ | J.T.Baker

Material No.: 9266-A4
Batch No.: 2412662006
Manufactured Date: 2024-08-29
Expiration Date:2025-11-28
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	2
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	3
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	>= 99.8 %	99.9 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Titrable Acid (µeq/g)	<= 0.3	<0.1
Chloride (Cl)	<= 10 ppm	<5 ppm
Water (by KF, coulometric)	<= 0.02 %	<0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3822

JCroak
Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Hydrochloric Acid, 36.5-38.0%
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis

Material No.: 9530-33
Batch No.: 0000281827
Manufactured Date: 2021/03/30
Retest Date: 2026/03/29
Revision No: 1

Certificate of Analysis

Test	Specification	Result
ACS - Assay (as HCl) (by acid-base titrn)	36.5 - 38.0 %	37.6
ACS - Color (APHA)	<= 10	5
ACS - Residue after Ignition	<= 3 ppm	1
ACS - Specific Gravity at 60°/60°F	1.185 - 1.192	1.189
ACS - Bromide (Br)	<= 0.005 %	< 0.005
ACS - Extractable Organic Substances	<= 5 ppm	< 1
ACS - Free Chlorine (as Cl_2)	<= 0.5 ppm	< 0.5
Phosphate (PO_4)	<= 0.05 ppm	< 0.03
Sulfate (SO_4)	<= 0.5 ppm	< 0.3
Sulfite (SO_3)	<= 0.8 ppm	0.3
Ammonium (NH_4)	<= 3 ppm	< 1
Trace Impurities - Arsenic (As)	<= 0.010 ppm	< 0.003
Trace Impurities - Aluminum (Al)	<= 10.0 ppb	0.5
Arsenic and Antimony (as As)	<= 5 ppb	< 3
Trace Impurities - Barium (Ba)	<= 1.0 ppb	< 0.2
Trace Impurities - Beryllium (Be)	<= 1.0 ppb	< 0.2
Trace Impurities - Bismuth (Bi)	<= 10.0 ppb	< 1.0
Trace Impurities - Boron (B)	<= 20.0 ppb	< 5.0
Trace Impurities - Cadmium (Cd)	<= 1.0 ppb	< 0.3
Trace Impurities - Calcium (Ca)	<= 50.0 ppb	15.0
Trace Impurities - Chromium (Cr)	<= 1.0 ppb	< 0.4
Trace Impurities - Cobalt (Co)	<= 1.0 ppb	< 0.3
Trace Impurities - Copper (Cu)	<= 1.0 ppb	< 0.1
Trace Impurities - Gallium (Ga)	<= 1.0 ppb	< 0.2

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Material No.: 9530-33
Batch No.: 0000281827

Test	Specification	Result
Trace Impurities - Germanium (Ge)	<= 3.0 ppb	< 2.0
Trace Impurities - Gold (Au)	<= 4.0 ppb	3.0
Heavy Metals (as Pb)	<= 100 ppb	< 50
Trace Impurities - Iron (Fe)	<= 15.0 ppb	1.0
Trace Impurities - Lead (Pb)	<= 1.0 ppb	< 0.5
Trace Impurities - Lithium (Li)	<= 1.0 ppb	< 0.2
Trace Impurities - Magnesium (Mg)	<= 10.0 ppb	< 0.4
Trace Impurities - Manganese (Mn)	<= 1.0 ppb	< 0.4
Trace Impurities - Mercury (Hg)	<= 0.5 ppb	0.2
Trace Impurities - Molybdenum (Mo)	<= 10.0 ppb	< 5.0
Trace Impurities - Nickel (Ni)	<= 4.0 ppb	< 0.3
Trace Impurities - Niobium (Nb)	<= 1.0 ppb	< 0.2
Trace Impurities - Potassium (K)	<= 9.0 ppb	< 2.0
Trace Impurities - Selenium (Se), For Information Only	ppb	1.0
Trace Impurities - Silicon (Si)	<= 100.0 ppb	18.0
Trace Impurities - Silver (Ag)	<= 1.0 ppb	< 0.3
Trace Impurities - Sodium (Na)	<= 100.0 ppb	< 5.0
Trace Impurities - Strontium (Sr)	<= 1.0 ppb	< 0.2
Trace Impurities - Tantalum (Ta)	<= 1.0 ppb	< 0.9
Trace Impurities - Thallium (Tl)	<= 5.0 ppb	< 2.0
Trace Impurities - Tin (Sn)	<= 5.0 ppb	< 0.8
Trace Impurities - Titanium (Ti)	<= 1.0 ppb	< 0.2
Trace Impurities - Vanadium (V)	<= 1.0 ppb	< 0.2
Trace Impurities - Zinc (Zn)	<= 5.0 ppb	0.4
Trace Impurities - Zirconium (Zr)	<= 1.0 ppb	< 0.1

For Laboratory, Research or Manufacturing Use
Product Information (not specifications):
Appearance (clear, fuming liquid)
Meets ACS Specifications

Country of Origin: US
Packaging Site: Phillipsburg Mfg Ctr & DC



For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110094-02	506889	≤ -10 °C	Methylene Chloride	7/25/2028	CLP Base/Neutral Surrogate Solution, 5,000 mg/L, 1 ml

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2-dichlorobenzene-d_4	2199-69-1	99.7	247.29.3P	5035 ± 28.02
2-fluorobiphenyl	321-60-8	99.69	8.286.1.1P	4999 ± 103.66
nitrobenzene-d5	4165-60-0	99.67	7.9.3P	4988 ± 27.32
p-terphenyl-d14	1718-51-0	99.3	9.120.8P	5005 ± 27.85

511494 to 511498 } Y.P. 08/11/2023

*Not a certified value

Certified By: _____

Clint Tipton
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01
ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0201940

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

511539 ↓ 511568 Y.P. 091 191

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S230321RSR	99%	1,001.0 µg/mL	+/- 22.9799
2	Atrazine	1912-24-9	5FYWL	99%	1,010.0 µg/mL	+/- 23.1865
3	Benzidine	92-87-5	S221205RSR	99%	1,008.0 µg/mL	+/- 23.1406
4	epsilon-Caprolactam	105-60-2	I16X016	99%	1,008.0 µg/mL	+/- 23.1406

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Sam Moodler - Operations Tech I

REVIEWED
By Jennifer Pollino at 7:10 am, Sep 13, 2023

Date Mixed: 13-Sep-2023 **Balance:** B345965662

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01
ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0201940

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

511539 ↓ 511568 Y.P. 091 191

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S230321RSR	99%	1,001.0 µg/mL	+/- 22.9799
2	Atrazine	1912-24-9	5FYWL	99%	1,010.0 µg/mL	+/- 23.1865
3	Benzidine	92-87-5	S221205RSR	99%	1,008.0 µg/mL	+/- 23.1406
4	epsilon-Caprolactam	105-60-2	I16X016	99%	1,008.0 µg/mL	+/- 23.1406

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Sam Moodler - Operations Tech I

REVIEWED
By Jennifer Pollino at 7:10 am, Sep 13, 2023

Date Mixed: 13-Sep-2023 **Balance:** B345965662

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555872 **Lot No.:** A0201728

Description : Custom Pentachlorophenol Standard

Custom Pentachlorophenol Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2026 **Storage:** 10°C or colder

Ship: Ambient

511649 ↓ 511658 Y.P. 11/13/23

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pentachlorophenol	87-86-5	RP230530RSR	99%	25,000.0 µg/mL	+/- 777.0837

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Josh McCloskey - Operations Technician I

Date Mixed: 05-Sep-2023 **Balance:** B251644995

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2028 **Storage:** 0°C or colder

Ship: Ambient

S11749 ↓ S11794 RC 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant flow 1.8 mL/min.

Temp. Program:
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)


Inj. Temp:
250°C

Det. Temp:
340°C

Det. Type:
FID

Split Vent:
100 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2028 **Storage:** 0°C or colder

Ship: Ambient

S11749 ↓ S11794 RC 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant flow 1.8 mL/min.

Temp. Program:
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)


Inj. Temp:
250°C

Det. Temp:
340°C

Det. Type:
FID

Split Vent:
100 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2028 **Storage:** 0°C or colder

Ship: Ambient

S11749 ↓ S11794 RC 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant flow 1.8 mL/min.

Temp. Program:
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)


Inj. Temp:
250°C

Det. Temp:
340°C

Det. Type:
FID

Split Vent:
100 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2028 **Storage:** 0°C or colder

Ship: Ambient

S11749 ↓ S11794 RC 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant flow 1.8 mL/min.

Temp. Program:
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)


Inj. Temp:
250°C

Det. Temp:
340°C

Det. Type:
FID

Split Vent:
100 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2028 **Storage:** 0°C or colder

Ship: Ambient

S11749 ↓ S11794 RC 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant flow 1.8 mL/min.

Temp. Program:
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)


Inj. Temp:
250°C

Det. Temp:
340°C

Det. Type:
FID

Split Vent:
100 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2028 **Storage:** 0°C or colder

Ship: Ambient

S11749 ↓ S11794 RC 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant flow 1.8 mL/min.

Temp. Program:
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)


Inj. Temp:
250°C

Det. Temp:
340°C

Det. Type:
FID

Split Vent:
100 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31206 **Lot No.:** A0201320

S12013 ↓ S12042 RC 12/26/23

Description : SV Internal Standard Mix 2mg/ml

SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2029 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,017.0 µg/mL	+/- 90.8469
2	Naphthalene-d8	1146-65-2	M-2180	99%	2,011.3 µg/mL	+/- 90.5917
3	Acenaphthene-d10	15067-26-2	PR-33507	99%	2,008.6 µg/mL	+/- 90.4685
4	Phenanthrene-d10	1517-22-2	PR-32303	99%	2,019.4 µg/mL	+/- 90.9550
5	Chrysene-d12	1719-03-5	PR-32210	99%	2,013.7 µg/mL	+/- 90.6968
6	Perylene-d12	1520-96-3	PR-33205	99%	2,012.7 µg/mL	+/- 90.6517

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Peter Robbins - Operations Technician I

Date Mixed: 23-Aug-2023 **Balance Serial #** B345965662

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 25-Aug-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110816-01	414127	≤ -10 °C	Methylene Chloride	6/21/2025	Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
atrazine	1912-24-9	99.5	337.7.3P	997 ± 5.81
benzidine	92-87-5	99.9	124.18.6.2P	991.8 ± 5.77
caprolactam	105-60-2	99.9	271.1.6P	999 ± 5.82

S12075
↓
S12079
RC
02/01/24

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: _____
Shane Overcash
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

S12082 ↓ S12111 RC/ 02/22/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 µg/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 µg/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 µg/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 µg/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 µg/mL	+/- 29.571294

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024 **Balance:** B345965662

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

S12082 ↓ S12111 RC/ 02/22/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 µg/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 µg/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 µg/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 µg/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 µg/mL	+/- 29.571294

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024 **Balance:** B345965662

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

S12082 ↓ S12111 RC/ 02/22/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 µg/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 µg/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 µg/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 µg/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 µg/mL	+/- 29.571294

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024 **Balance:** B345965662

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

S12082 ↓ S12111 RC/ 02/22/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 µg/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 µg/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 µg/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 µg/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 µg/mL	+/- 29.571294

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024 **Balance:** B345965662

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

S12082 ↓ S12111 RC/ 02/22/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 µg/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 µg/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 µg/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 µg/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 µg/mL	+/- 29.571294

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024 **Balance:** B345965662

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

S12082 ↓ S12111 RC/ 02/22/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 µg/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 µg/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 µg/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 µg/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 µg/mL	+/- 29.571294

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024 **Balance:** B345965662

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

S12082 ↓ S12111 RC/ 02/22/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 µg/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 µg/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 µg/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 µg/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 µg/mL	+/- 29.571294

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024 **Balance:** B345965662

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

S12082 ↓ S12111 RC/ 02/22/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 µg/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 µg/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 µg/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 µg/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 µg/mL	+/- 29.571294

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024 **Balance:** B345965662

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-020223-01	454157	≤ -10 °C	P/T Methanol	6/10/2026	1,4-Dioxane Solution, 2000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,4-dioxane	123-91-1	100	223.1.3P	1997 ± 57.08

S12112 ↓ S12116 } RC/ 03/08/24

*Not a certified value

Certified By: Melissa Workoff

Melissa Workoff
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

S12117 ↓ S12146 RC/ 03/18/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

S12117 ↓ S12146 RC/ 03/18/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

S12117 ↓ S12146 RC/ 03/18/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

S12117 ↓ S12146 RC/ 03/18/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

S12117 ↓ S12146 RC/ 03/18/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

S12117 ↓ S12146 RC/ 03/18/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

S12117 ↓ S12146 RC/ 03/18/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

S12117 ↓ S12146 RC/ 03/18/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

S12117 ↓ S12146 RC/ 03/18/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

S12117 ↓ S12146 RC/ 03/18/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCCK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087 **Lot No.:** A0206206

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10, 000µg/mL, Methanol, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : January 31, 2032 **Storage:** 10°C or colder

Ship: Ambient

S12187 ↓ S12206 } RC/ 03/18/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Fluorophenol	367-12-4	STBK1705	99%	10,005.3 µg/mL	+/- 302.5390
2	Phenol-d6	13127-88-3	PR-33287A	99%	10,005.5 µg/mL	+/- 302.5475
3	2,4,6-Tribromophenol	118-79-6	RP230831RSR	99%	10,006.6 µg/mL	+/- 302.5783

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID

Split Vent:
2 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

Date Mixed: 04-Jan-2024 **Balance Serial #** 1128360905

Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 08-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087 **Lot No.:** A0206206

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10, 000µg/mL, Methanol, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : January 31, 2032 **Storage:** 10°C or colder

Ship: Ambient

S12187 ↓ S12206 } RC/ 03/18/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Fluorophenol	367-12-4	STBK1705	99%	10,005.3 µg/mL	+/- 302.5390
2	Phenol-d6	13127-88-3	PR-33287A	99%	10,005.5 µg/mL	+/- 302.5475
3	2,4,6-Tribromophenol	118-79-6	RP230831RSR	99%	10,006.6 µg/mL	+/- 302.5783

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID

Split Vent:
2 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

Date Mixed: 04-Jan-2024 **Balance Serial #** 1128360905

Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 08-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087 **Lot No.:** A0206206

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10, 000µg/mL, Methanol, 5mL/ampul

S12187 ↓ S12206 } RC/ 03/18/24

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : January 31, 2032 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Fluorophenol	367-12-4	STBK1705	99%	10,005.3 µg/mL	+/- 302.5390
2	Phenol-d6	13127-88-3	PR-33287A	99%	10,005.5 µg/mL	+/- 302.5475
3	2,4,6-Tribromophenol	118-79-6	RP230831RSR	99%	10,006.6 µg/mL	+/- 302.5783

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID

Split Vent:
2 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

Date Mixed: 04-Jan-2024 **Balance Serial #** 1128360905

Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 08-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 **Lot No.:** A0206381

Description : B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : December 31, 2029 **Storage:** 10°C or colder

Handling: Sonicate prior to use. **Ship:** Ambient

S12207 ↓ S12221 RC/ 03/18/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Nitrobenzene-d5	4165-60-0	I-25158	99%	5,029.3 µg/mL	+/- 226.5204
2	2-Fluorobiphenyl	321-60-8	00021384	99%	5,030.9 µg/mL	+/- 226.5936
3	p-Terphenyl-d14	1718-51-0	PR-32599	99%	5,026.4 µg/mL	+/- 226.3909

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID

Split Vent:
2 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Jess Hoy - Operations Tech I

Date Mixed: 09-Jan-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 11-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-010442-07	495833	≤ -10 °C	Methylene Chloride	1/16/2028	Benzaldehyde Solution, 1000 mg/L, 1.3 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
benzaldehyde	100-52-7	98.3	442.421.1P	996.8 ± 11.49

S12275 ↓ S12279 } RC/ 05/24/24

*Not a certified value

Certified By: S. Hunter

Scott Hunter
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

S12312 ↓ S12331 } RC/ 05/30/24

Catalog No. : 31206 **Lot No.:** A0206540

Description : SV Internal Standard Mix 2mg/ml

SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2029 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,007.1 µg/mL	+/- 90.4025
2	Naphthalene-d8	1146-65-2	M-2180	99%	2,005.9 µg/mL	+/- 90.3454
3	Acenaphthene-d10	15067-26-2	PR-33507	99%	2,007.9 µg/mL	+/- 90.4385
4	Phenanthrene-d10	1517-22-2	PR-32303	99%	2,006.7 µg/mL	+/- 90.3845
5	Chrysene-d12	1719-03-5	PR-32210	99%	2,015.5 µg/mL	+/- 90.7778
6	Perylene-d12	1520-96-3	PR-33205	99%	2,014.7 µg/mL	+/- 90.7448

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Malina Homan - Operations Technician I

Date Mixed: 12-Jan-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

S12312 ↓ S12331 } RC/ 05/30/24

Catalog No. : 31206 **Lot No.:** A0206540

Description : SV Internal Standard Mix 2mg/ml

SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2029 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,007.1 µg/mL	+/- 90.4025
2	Naphthalene-d8	1146-65-2	M-2180	99%	2,005.9 µg/mL	+/- 90.3454
3	Acenaphthene-d10	15067-26-2	PR-33507	99%	2,007.9 µg/mL	+/- 90.4385
4	Phenanthrene-d10	1517-22-2	PR-32303	99%	2,006.7 µg/mL	+/- 90.3845
5	Chrysene-d12	1719-03-5	PR-32210	99%	2,015.5 µg/mL	+/- 90.7778
6	Perylene-d12	1520-96-3	PR-33205	99%	2,014.7 µg/mL	+/- 90.7448

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Malina Homan - Operations Technician I

Date Mixed: 12-Jan-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 ↓ S12508 } RC 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 ↓ S12508 } RC 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 ↓ S12508 } RC 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 ↓ S12508 } RC 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 ↓ S12508 } RC 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 ↓ S12508 } RC 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 ↓ S12508 } RC 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 ↓ S12508 } RC 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 ↓ S12508 } RC 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



SHIPPING DOCUMENTS

CHEMTECH
CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 • Fax (908) 789-8922
www.chemtech.net

CHEMTECH PROJECT NO. P4578/79
QUOTE NO.
COC Number 2042038

CLIENT INFORMATION

REPORT TO BE SENT TO:
COMPANY: GANNETT FLEMING
ADDRESS: 1010 Adams Avenue
CITY Audubon STATE: PA ZIP: 19403
ATTENTION: Joe Krupansky
PHONE: 610-391-8342 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: Amtrak's Mop. of SB
PROJECT NO.: 95000005B LOCATION: Kearny, NJ
PROJECT MANAGER: Joe Krupansky
e-mail: aaac@bemgs.com
PHONE: 610-391-8342 FAX:

CLIENT BILLING INFORMATION

BILL TO: Chemtech PO#:
ADDRESS: 284 Sheffield St.
CITY Mountainside STATE: NJ ZIP: 07092
ATTENTION: Samantha PHONE: 908-788-398

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): 10 DAYS*
EDD: 10 DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

- ☐ Level 1 (Results Only)
- ☐ Level 2 (Results + QC)
- ☐ Level 3 (Results + QC + Raw Data)
- ☐ EDD FORMAT BEM EDD
- ☐ Level 4 (QC + Full Raw Data)
- ☒ NJ Reduced
- ☐ US EPA CLP
- ☐ NYS ASP A
- ☐ NYS ASP B
- ☐ Other _____

ANALYSIS

1 TCL VOC +10 | 2 TCL PAH | 3 PCB's | 4 Cr(III), Cr(VI) | 5 TAL Metals | 6 EPH | 7 TCLP - Full | 8 RCRA Parameters | 9 TDX

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE COMP	SAMPLE TYPE GRAB	SAMPLE COLLECTION DATE	SAMPLE COLLECTION TIME	# OF BOTTLES	PRESERVATIVES 1	2	3	4	5	6	7	8	9	COMMENTS
								A			F	B					← Specify Preservatives A-HCl D-NaOH B-HN03 E-ICE C-H2SO4 F-OTHER
1.	WB-305-SW	GW	X		10/25	8:20	7	X	X	X	X	X	X				
2.	WB-305-SW-1	GW	X		10/25	8:30	7	X	X	X	X	X	X				
3.	WB-305-TOP	S	X		10/25	9:20	11	X	X	X	X	X	X				
4.	WB-305-TOP-1	S	X		10/25	9:30	11	X	X	X	X	X	X				
5.	WB-305-Bot	S	X		10/25	11:15	11	X	X	X	X	X	X	X	X	X	
6.	WB-305-Bot-1	S	X		10/25	11:20	11	X	X	X	X	X	X	X	X	X	
7.	FB-10252024	D.I.W.			10/25		2	X									
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:
1. Vaunda Pujara	10/25/24-3:11	1.
2.		2.
3.	10.25.2024 1604	3.

Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☒ COOLER TEMP 3.1 °C
Comments:

Page ____ of ____

CLIENT: ☐ Hand Delivered ☐ Other _____
CHEMTECH: ☐ Picked Up ☐ Field Sampling

Shipment Complete ☐ YES ☐ NO

Copyright © 2023

WHITE - CHEMTECH COPY FOR RETURN TO CLIENT YELLOW - CHEMTECH COPY PINK - SAMPLER COPY

284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (L-A-B)	L2219
Maine	2024021
Maryland	296
New Hampshire	255423
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID : P4578 PORT06
Client Name : Portal Partners Tri-Venture
Client Contact : Joseph Krupansky
Invoice Name : Portal Partners Tri-Venture
Invoice Contact : Joseph Krupansky

Order Date : 10/25/2024 3:58:00 PM
Project Name : Amtrak Sawtooth Bridges 2
Receive DateTime : 10/25/2024 4:04:00 PM
Purchase Order :

Project Mgr :
Report Type : NJ Reduced
EDD Type : EXCEL NJCLEANUP
Hard Copy Date :
Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
P4578-01	WB-305-SW	Water	10/25/2024	08:20					
					VOC-TCLVOA-10		8260-Low		10 Bus. Days
P4578-02	WB-305-SW-1	Water	10/25/2024	08:30					
					VOC-TCLVOA-10		8260-Low		10 Bus. Days
P4578-03	WB-305-TOP	Solid	10/25/2024	09:20					
					VOC-TCLVOA-10		8260D		10 Bus. Days
P4578-04	WB-305-TOP-1	Solid	10/25/2024	09:30					
					VOC-TCLVOA-10		8260D		10 Bus. Days
P4578-05	WB-305-BOT	Solid	10/25/2024	11:15					
					VOC-TCLVOA-10		8260D		10 Bus. Days
P4578-07	WB-305-BOT-1	Solid	10/25/2024	11:20					
					VOC-TCLVOA-10		8260D		10 Bus. Days
P4578-09	TB-10252024	Water	10/25/2024	00:00					
					VOC-TCLVOA-10		8260-Low		10 Bus. Days

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID :	P4578 PORT06	**Order Date :**	10/25/2024 3:58:00 PM	**Project Mgr :**	
Client Name :	Portal Partners Tri-Venture	**Project Name :**	Amtrak Sawtooth Bridges 2	**Report Type :**	NJ Reduced
Client Contact :	Joseph Krupansky	**Receive DateTime :**	10/25/2024 4:04:00 PM	**EDD Type :**	EXCEL NJCLEANUP
Invoice Name :	Portal Partners Tri-Venture	**Purchase Order :**		**Hard Copy Date :**	
Invoice Contact :	Joseph Krupansky			**Date Signoff :**	

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES

* J.C samples are in SM Refrigerator *

Relinguished By : J. C
Date / Time : 10-25-2024 1710

Received By : Sam 10/25/2024 1710
Date / Time : Sam.

RJ 6 +2 2 RJ 5

Storage Area : VOA Refridgerator Room