

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4578

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

DataFile: BF140077.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		
								Qual	Low	High	RPD
PB164444BS	Pyridine	50	38.6	ug/L	77				20 (29)	160 (97)	
	1,4-Dichlorobenzene	50	43.7	ug/L	87				70 (76)	130 (103)	
	2-Methylphenol	50	44.4	ug/L	89				70 (69)	130 (109)	
	3+4-Methylphenols	50	43.0	ug/L	86				20 (67)	160 (106)	
	Hexachloroethane	50	44.2	ug/L	88				20 (76)	160 (118)	
	Nitrobenzene	50	43.5	ug/L	87				70 (58)	130 (106)	
	Hexachlorobutadiene	50	44.9	ug/L	90				70 (69)	130 (101)	
	2,4,6-Trichlorophenol	50	45.9	ug/L	92				70 (61)	130 (110)	
	2,4,5-Trichlorophenol	50	44.6	ug/L	89				70 (70)	130 (106)	
	2,4-Dinitrotoluene	50	52.4	ug/L	105				70 (60)	130 (115)	
	Hexachlorobenzene	50	45.0	ug/L	90				70 (73)	130 (106)	
	Pentachlorophenol	100	89.7	ug/L	90				20 (47)	160 (114)	