

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4578

Client: Portal Partners Tri-Venture

Analytical Method: SW8260D

Datafile : VY020051.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1029SBS01	Dichlorodifluoromethane	20	14.9	ug/Kg	75			40 (64)	160 (136)	
	Chloromethane	20	15.7	ug/Kg	79			40 (70)	160 (130)	
	Vinyl chloride	20	17.8	ug/Kg	89			70 (72)	130 (129)	
	Bromomethane	20	18.5	ug/Kg	93			40 (58)	160 (141)	
	Chloroethane	20	19.7	ug/Kg	99			40 (69)	160 (130)	
	Trichlorofluoromethane	20	19.3	ug/Kg	97			40 (69)	160 (134)	
	1,1,2-Trichlorotrifluoroethane	20	20.3	ug/Kg	102			70 (81)	130 (123)	
	1,1-Dichloroethene	20	17.4	ug/Kg	87			70 (79)	130 (121)	
	Acetone	100	140	ug/Kg	140			40 (60)	160 (131)	
	Carbon disulfide	20	10.2	ug/Kg	51			40 (45)	160 (154)	
	Methyl tert-butyl Ether	20	21.2	ug/Kg	106			70 (77)	130 (129)	
	Methyl Acetate	20	21.6	ug/Kg	108			70 (69)	130 (149)	
	Methylene Chloride	20	21.8	ug/Kg	109			70 (56)	130 (174)	
	trans-1,2-Dichloroethene	20	17.6	ug/Kg	88			70 (80)	130 (123)	
	1,1-Dichloroethane	20	22.3	ug/Kg	112			70 (82)	130 (123)	
	Cyclohexane	20	16.9	ug/Kg	85			70 (76)	130 (122)	
	2-Butanone	100	130	ug/Kg	130			40 (69)	160 (131)	
	Carbon Tetrachloride	20	20.0	ug/Kg	100			70 (76)	130 (129)	
	cis-1,2-Dichloroethene	20	20.5	ug/Kg	103			70 (82)	130 (123)	
	Bromochloromethane	20	23.4	ug/Kg	117			70 (80)	130 (127)	
	Chloroform	20	23.3	ug/Kg	117			70 (82)	130 (125)	
	1,1,1-Trichloroethane	20	21.7	ug/Kg	109			70 (80)	130 (126)	
	Methylcyclohexane	20	16.1	ug/Kg	81			70 (77)	130 (123)	
	Benzene	20	19.9	ug/Kg	100			70 (84)	130 (121)	
	1,2-Dichloroethane	20	21.0	ug/Kg	105			70 (81)	130 (126)	
	Trichloroethene	20	18.6	ug/Kg	93			70 (83)	130 (122)	
	1,2-Dichloropropane	20	22.5	ug/Kg	113			70 (83)	130 (122)	
	Bromodichloromethane	20	22.5	ug/Kg	113			70 (82)	130 (123)	
	4-Methyl-2-Pentanone	100	120	ug/Kg	120			40 (70)	160 (135)	
	Toluene	20	20.4	ug/Kg	102			70 (83)	130 (122)	
	t-1,3-Dichloropropene	20	20.4	ug/Kg	102			70 (78)	130 (124)	
	cis-1,3-Dichloropropene	20	20.6	ug/Kg	103			70 (81)	130 (122)	
	1,1,2-Trichloroethane	20	22.8	ug/Kg	114			70 (82)	130 (125)	
	2-Hexanone	100	130	ug/Kg	130			40 (66)	160 (138)	
	Dibromochloromethane	20	21.7	ug/Kg	109			70 (79)	130 (125)	
	1,2-Dibromoethane	20	20.9	ug/Kg	104			70 (80)	130 (125)	
	Tetrachloroethene	20	17.8	ug/Kg	89			70 (83)	130 (125)	
	Chlorobenzene	20	20.3	ug/Kg	102			70 (84)	130 (122)	
	Ethyl Benzene	20	20.6	ug/Kg	103			70 (82)	130 (124)	
	m/p-Xylenes	40	40.8	ug/Kg	102			70 (83)	130 (124)	
	o-Xylene	20	20.2	ug/Kg	101			70 (83)	130 (123)	
	Styrene	20	21.2	ug/Kg	106			70 (82)	130 (124)	
	Bromoform	20	21.7	ug/Kg	109			70 (75)	130 (127)	
	Isopropylbenzene	20	21.4	ug/Kg	107			70 (82)	130 (124)	
	1,1,2,2-Tetrachloroethane	20	23.9	ug/Kg	119			70 (77)	130 (127)	
	1,3-Dichlorobenzene	20	21.2	ug/Kg	106			70 (83)	130 (122)	
	1,4-Dichlorobenzene	20	21.4	ug/Kg	107			70 (84)	130 (121)	
	1,2-Dichlorobenzene	20	21.6	ug/Kg	108			70 (83)	130 (124)	

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4578

Client: Portal Partners Tri-Venture

Analytical Method: SW8260D

Datafile : VY020051.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VY1029SBS01	1,2-Dibromo-3-Chloropropane	20	22.3	ug/Kg	112			40 (66)	160 (134)	
	1,2,4-Trichlorobenzene	20	19.1	ug/Kg	96			70 (78)	130 (127)	
	1,2,3-Trichlorobenzene	20	19.0	ug/Kg	95			70 (70)	130 (137)	