

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578
 Lab File ID: VX043583.D Date Analyzed: 10/29/2024
 Instrument ID: MSVOA_X Time Analyzed: 09:29
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	136145	12.018				
UPPER LIMIT	272290	12.518				
LOWER LIMIT	68072.5	11.518				
EPA SAMPLE NO.						
WB-305-SW	100896	12.02				
WB-305-SW-1	107787	12.02				
TB-10252024	105429	12.02				
VX1029WBL01	128354	12.02				
VX1029WBS01	142936	12.02				
VX1029WBSD01	131226	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.