

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578
 Lab File ID: VY020029.D Date Analyzed: 10/28/2024
 Instrument ID: MSVOA_Y Time Analyzed: 09:29
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	147527	13.352				
UPPER LIMIT	295054	13.852				
LOWER LIMIT	73763.5	12.852				
EPA SAMPLE NO.						
WB-305-TOP	8668 *	13.36				
WB-305-TOP-1	133013	13.35				
WB-305-BOT-1	145301	13.35				
VY1028SBL01	125121	13.35				
VY1028SBS01	143516	13.35				
VY1028SBSD01	146119	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.