

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: P4578 SAS No.: P4578 SDG NO.: P4578
 Lab File ID: VY020049.D Date Analyzed: 10/29/2024
 Instrument ID: MSVOA_Y Time Analyzed: 09:31
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	148124	13.352				
UPPER LIMIT	296248	13.852				
LOWER LIMIT	74062	12.852				
EPA SAMPLE NO.						
WB-305-BOT	127440	13.35				
VY1029SBL01	118867	13.35				
VY1029SBS01	134900	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.