

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51716 MA No.: SDG No.: MC0PI1
Instrument ID: CV1 Start Date: 10/31/2024
Analytical Method: CVAA Run Batch: LB133241
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.0	0	0.2	0.17	-16	2.5	2.5	-2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51716 MA No.: _____ SDG No.: MC0PI1
Instrument ID: CV1 Start Date: 10/31/2024
Analytical Method: CVAA Run Batch: LB133241
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	5.1	1	7.5	7.8	4	10	9.7	-3

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51716 MA No.: _____ SDG No.: MC0PI1
Instrument ID: CV1 Start Date: 11/15/2024
Analytical Method: CVAA Run Batch: LB133469
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.0	0	0.2	0.17	-14	2.5	2.5	2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51716 MA No.: _____ SDG No.: MC0PI1
Instrument ID: CV1 Start Date: 11/15/2024
Analytical Method: CVAA Run Batch: LB133469
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	5.1	2	7.5	7.4	-1	10	10	0