


LB133244

SFAM01.1

INSTRUMENT M1CV2



A= 0.0000e+000
B= 2.4300e-004 *slope*
C= -2.2893e-002 *intercept*
Rho= 0.9989934
Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.030	-0.030	-31	0.000	-31					=
0.05	0.050					442	0				-16
0.20	0.200	0.169	-0.031	788	0.0 %	788					-2
2.50	2.500	2.462	-0.038	10225	0.0 %	10225					1
5.00	5.000	5.064	0.064	20932	0.0 %	20932					4
7.50	7.500	7.798	0.298	32183	0.0 %	32183					-3
10.0	10.000	9.739	-0.261	40170	0.0 %	40170					

LB133241

INSTRUMENT ID: CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-31	-	0	SFAM01.1	PPB	10/31/2024 18:33	Std
	0.2 S01	788	-	0.2	SFAM01.1	PPB	10/31/2024 18:35	Std
	2.5 S02	10225	-	2.5	SFAM01.1	PPB	10/31/2024 18:37	Std
	5 S03	20932	-	5	SFAM01.1	PPB	10/31/2024 18:40	Std
	7.5 S04	32183	-	7.5	SFAM01.1	PPB	10/31/2024 18:44	Std
	10 S05	40170	-	10	SFAM01.1	PPB	10/31/2024 18:49	Std
ICV064	ICV064	16004	3.8662 -		SFAM01.1	PPB	10/31/2024 18:52	SMPL
ICB064	ICB064	-277	-0.0902 -		SFAM01.1	PPB	10/31/2024 18:55	SMPL
CCV011	CCV011	21041	5.0902 -		SFAM01.1	PPB	10/31/2024 18:57	SMPL
CCB011	CCB011	-223	-0.0771 -		SFAM01.1	PPB	10/31/2024 18:59	SMPL
PB164581BL	PBW581	-51	-0.0353 -		SFAM01.1	PPB	10/31/2024 19:01	SMPL
P4496-01	MBHCX5	-123	-0.0528 -		SFAM01.1	PPB	10/31/2024 19:04	SMPL
P4496-02	MBHCY6	603	0.1236 -		SFAM01.1	PPB	10/31/2024 19:09	SMPL
P4496-03	MBHCY7	-106	-0.0487 -		SFAM01.1	PPB	10/31/2024 19:11	SMPL
P4584-01	MC0PI1	-89	-0.0445 -		SFAM01.1	PPB	10/31/2024 19:13	SMPL
P4584-02	MC0PI5	-94	-0.0457 -		SFAM01.1	PPB	10/31/2024 19:16	SMPL
CCV012	CCV012	20269	4.9026 -		SFAM01.1	PPB	10/31/2024 19:18	SMPL
CCB012	CCB012	-246	-0.0827 -		SFAM01.1	PPB	10/31/2024 19:20	SMPL