

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133479

Review By	Sarabjit Jaswal	Review On	11/18/2024 7:24:59 AM
Supervise By	Mohan Bera	Supervise On	11/18/2024 7:28:13 AM

STD. NAME	STD REF.#
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136
ICV Standard	MP83143
CCV Standard	MP83146
ICSA Standard	MP83144,MP83145
CRI Standard	
LCS Standard	MP83141
Chk Standard	MP83148,MP83149

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/17/24 13:02		Kareem	OK
2	S01	S01	CAL2	11/17/24 13:07		Kareem	OK
3	S02	S02	CAL3	11/17/24 13:11		Kareem	OK
4	S03	S03	CAL4	11/17/24 13:16		Kareem	OK
5	S04	S04	CAL5	11/17/24 13:20		Kareem	OK
6	S05	S05	CAL6	11/17/24 13:24		Kareem	OK
7	S06	S06	CAL7	11/17/24 13:29		Kareem	OK
8	ICV002	ICV002	ICV	11/17/24 13:42	Inst. paused for calibration review	Kareem	OK
9	ICB002	ICB002	ICB	11/17/24 13:47		Kareem	OK
10	ICSA002	ICSA002	ICSA	11/17/24 13:51		Kareem	OK
11	ICSAB002	ICSAB002	ICSAB	11/17/24 13:56		Kareem	OK
12	CCV009	CCV009	CCV	11/17/24 14:00		Kareem	OK
13	CCB009	CCB009	CCB	11/17/24 14:05		Kareem	OK
14	PB164661BL	PBS661	MB	11/17/24 14:09		Kareem	OK
15	PB164661BS	LCS661	LCS	11/17/24 14:14		Kareem	OK
16	PB164661BSD	LCSD661	LCSD	11/17/24 14:18		Kareem	OK
17	PB164697BL	PBS697	MB	11/17/24 14:23		Kareem	OK
18	PB164697BS	LCS697	LCS	11/17/24 14:27		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133479

Review By	Sarabjit Jaswal	Review On	11/18/2024 7:24:59 AM
Supervise By	Mohan Bera	Supervise On	11/18/2024 7:28:13 AM

STD. NAME	STD REF.#
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136
ICV Standard	MP83143
CCV Standard	MP83146
ICSA Standard	MP83144,MP83145
CRI Standard	
LCS Standard	MP83141
Chk Standard	MP83148,MP83149

19	PB164697BSD	LCSD697	LCSD	11/17/24 14:32		Kareem	OK
20	PB164701BL	PBS701	MB	11/17/24 14:36		Kareem	OK
21	PB164701BS	LCS701	LCS	11/17/24 14:41		Kareem	OK
22	PB164701BSD	LCSD701	LCSD	11/17/24 14:45		Kareem	OK
23	PB164800BL	PBS800	MB	11/17/24 14:50		Kareem	OK
24	PB164800BS	LCS800	LCS	11/17/24 14:54		Kareem	OK
25	PB164801BL	PBS801	MB	11/17/24 14:59		Kareem	OK
26	PB164801BS	LCS801	LCS	11/17/24 15:03		Kareem	OK
27	PB164802BL	PBS802	MB	11/17/24 15:08		Kareem	OK
28	PB164802BS	LCS802	LCS	11/17/24 15:12		Kareem	OK
29	P4587-01DL	MYE5G4	SAM	11/17/24 15:17		Kareem	OK
30	P4587-02DL	MYE5G5	SAM	11/17/24 15:21		Kareem	OK
31	P4587-03DL	MYE5G6	SAM	11/17/24 15:26		Kareem	OK
32	P4587-04DL	MYE5G7	SAM	11/17/24 15:30		Kareem	OK
33	P4587-05DL	MYE5G8	SAM	11/17/24 15:35		Kareem	OK
34	CCV010	CCV010	CCV	11/17/24 15:39		Kareem	OK
35	CCB010	CCB010	CCB	11/17/24 15:43		Kareem	OK
36	P4587-06DLRE	MYE5G9	SAM	11/17/24 15:48	Not Use	Kareem	Not Ok
37	P4587-07DLRE	MYE5G9D	DUP	11/17/24 15:52	Not Use	Kareem	Not Ok
38	P4587-06LDLRE	MYE5G9L	SD	11/17/24 15:57	Not Use	Kareem	Not Ok

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133479

Review By	Sarabjit Jaswal	Review On	11/18/2024 7:24:59 AM
Supervise By	Mohan Bera	Supervise On	11/18/2024 7:28:13 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard	MP83141		
Chk Standard	MP83148,MP83149		

39	P4587-09DL	MYE5H0	SAM	11/17/24 16:06		Kareem	OK
40	P4587-10DL	MYE5H1	SAM	11/17/24 16:10		Kareem	OK
41	P4587-11DL	MYE5H2	SAM	11/17/24 16:15		Kareem	OK
42	P4587-12DL	MYE5H3	SAM	11/17/24 16:19		Kareem	OK
43	P4587-13DL	MYE5H4	SAM	11/17/24 16:23		Kareem	OK
44	P4587-14DL	MYE5H5	SAM	11/17/24 16:28		Kareem	OK
45	P4587-15DL	MYE5H6	SAM	11/17/24 16:32		Kareem	OK
46	P4587-16DL	MYE5H7	SAM	11/17/24 16:37		Kareem	OK
47	P4587-17DL	MYE5H8	SAM	11/17/24 16:41		Kareem	OK
48	P4587-18DL	MYE5H9	SAM	11/17/24 16:46		Kareem	OK
49	P4587-19DL	MYE5J0	SAM	11/17/24 16:50		Kareem	OK
50	P4587-20DL	MYE5J1	SAM	11/17/24 16:54		Kareem	OK
51	P4588-01DL	MYE5J2	SAM	11/17/24 16:59		Kareem	OK
52	P4588-02DL	MYE5J3	SAM	11/17/24 17:03		Kareem	OK
53	P4588-03DL	MYE5J4	SAM	11/17/24 17:08		Kareem	OK
54	P4588-04DL	MYE5J5	SAM	11/17/24 17:12		Kareem	OK
55	CCV011	CCV011	CCV	11/17/24 17:17		Kareem	OK
56	CCB011	CCB011	CCB	11/17/24 17:21		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133481

Review By	Sarabjit Jaswal	Review On	11/18/2024 7:27:18 AM
Supervise By	Mohan Bera	Supervise On	11/18/2024 7:28:03 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard			
Chk Standard	MP83148,MP83149		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/17/24 19:56		Kareem	OK
2	S1	S01	CAL2	11/17/24 20:01		Kareem	OK
3	S2	S02	CAL3	11/17/24 20:06		Kareem	OK
4	S3	S03	CAL4	11/17/24 20:10		Kareem	OK
5	S4	S04	CAL5	11/17/24 20:14		Kareem	OK
6	S5	S05	CAL6	11/17/24 20:19		Kareem	OK
7	S6	S06	CAL7	11/17/24 20:23		Kareem	OK
8	ICV003	ICV003	ICV	11/17/24 20:35	Inst. paused for calibration review	Kareem	OK
9	ICB003	ICB003	ICB	11/17/24 20:39		Kareem	OK
10	ICSA003	ICSA003	ICSA	11/17/24 20:44		Kareem	OK
11	ICSAB003	ICSAB003	ICSAB	11/17/24 20:48		Kareem	OK
12	CCV012	CCV012	CCV	11/17/24 20:53		Kareem	OK
13	CCB012	CCB012	CCB	11/17/24 20:57		Kareem	OK
14	P4588-05DL	MYE5J6	SAM	11/17/24 21:02		Kareem	OK
15	P4588-06DL	MYE5J7	SAM	11/17/24 21:06		Kareem	OK
16	P4588-07DL	MYE5J8	SAM	11/17/24 21:10		Kareem	OK
17	P4588-08DL	MYE5J9	SAM	11/17/24 21:15		Kareem	OK
18	P4588-09DL	MYE5K0	SAM	11/17/24 21:19		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133481

Review By	Sarabjit Jaswal	Review On	11/18/2024 7:27:18 AM
Supervise By	Mohan Bera	Supervise On	11/18/2024 7:28:03 AM

STD. NAME	STD REF.#
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136
ICV Standard	MP83143
CCV Standard	MP83146
ICSA Standard	MP83144,MP83145
CRI Standard	
LCS Standard	
Chk Standard	MP83148,MP83149

19	P4588-10DL	MYE5K1	SAM	11/17/24 21:24		Kareem	OK
20	P4588-11DL	MYE5K2	SAM	11/17/24 21:28		Kareem	OK
21	P4588-12DL	MYE5K3	SAM	11/17/24 21:33		Kareem	OK
22	P4588-13DL	MYE5K4	SAM	11/17/24 21:37		Kareem	OK
23	P4588-14DL	MYE5K5	SAM	11/17/24 21:42		Kareem	OK
24	P4588-15DL	MYE5K5D	DUP	11/17/24 21:46		Kareem	OK
25	P4588-14LDL	MYE5K5L	SD	11/17/24 21:50		Kareem	OK
26	P4588-17DL	MYE5K6	SAM	11/17/24 21:59		Kareem	OK
27	P4588-18DL	MYE5K7	SAM	11/17/24 22:04		Kareem	OK
28	P4588-19DL	MYE5K8	SAM	11/17/24 22:08		Kareem	OK
29	P4588-20DL	MYE5K9	SAM	11/17/24 22:12		Kareem	OK
30	P4589-01DL	MYE5L0	SAM	11/17/24 22:17		Kareem	OK
31	P4589-02DL	MYE5L1	SAM	11/17/24 22:21		Kareem	OK
32	P4589-03DL	MYE5L2	SAM	11/17/24 22:26		Kareem	OK
33	CCV013	CCV013	CCV	11/17/24 22:30		Kareem	OK
34	CCB013	CCB013	CCB	11/17/24 22:35		Kareem	OK
35	P4589-04DL	MYE5M3	SAM	11/17/24 22:39		Kareem	OK
36	P4589-05DL	MYE5M4	SAM	11/17/24 22:44		Kareem	OK
37	P4589-06DL	MYE5M5	SAM	11/17/24 22:48		Kareem	OK
38	P4589-07DL	MYE5M6	SAM	11/17/24 22:52		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133481

Review By	Sarabjit Jaswal	Review On	11/18/2024 7:27:18 AM
Supervise By	Mohan Bera	Supervise On	11/18/2024 7:28:03 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard			
Chk Standard	MP83148,MP83149		

39	P4589-08DL	MYE5M7	SAM	11/17/24 22:57		Kareem	OK
40	P4589-09DL	MYE5M8	SAM	11/17/24 23:01		Kareem	OK
41	P4589-10DL	MYE5M9	SAM	11/17/24 23:06		Kareem	OK
42	P4589-11DL	MYE5N0	SAM	11/17/24 23:10		Kareem	OK
43	P4589-12DL	MYE5N1	SAM	11/17/24 23:14		Kareem	OK
44	P4589-13DL	MYE5N2	SAM	11/17/24 23:19		Kareem	OK
45	P4589-14DL	MYE5N2D	DUP	11/17/24 23:23		Kareem	OK
46	P4589-13LDL	MYE5N2L	SD	11/17/24 23:28		Kareem	OK
47	P4584-03	MC0PI0	SAM	11/17/24 23:37		Kareem	OK
48	P4584-04	MCC0Q4	SAM	11/17/24 23:41		Kareem	OK
49	P4584-05	MCC0Q5	SAM	11/17/24 23:45		Kareem	OK
50	P4584-06	MCC0Q6	SAM	11/17/24 23:50		Kareem	OK
51	P4584-07	MCC0Q7	SAM	11/17/24 23:54		Kareem	OK
52	P4584-08	MCC0Q8	SAM	11/17/24 23:59		Kareem	OK
53	P4584-09	MCC0Q9	SAM	11/18/24 00:03		Kareem	OK
54	CCV014	CCV014	CCV	11/18/24 00:07		Kareem	OK
55	CCB014	CCB014	CCB	11/18/24 00:12		Kareem	OK
56	P4584-10	MCC0R0	SAM	11/18/24 00:16		Kareem	OK
57	P4584-11	MCC0R1	SAM	11/18/24 00:21		Kareem	OK
58	P4584-12	MCC0R2	SAM	11/18/24 00:25		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133481

Review By	Sarabjit Jaswal	Review On	11/18/2024 7:27:18 AM
Supervise By	Mohan Bera	Supervise On	11/18/2024 7:28:03 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard			
Chk Standard	MP83148,MP83149		

59	P4584-13	MCC0R3	SAM	11/18/24 00:30		Kareem	OK
60	P4584-14	MCC0R4	SAM	11/18/24 00:34		Kareem	OK
61	P4584-15	MCC0R5	SAM	11/18/24 00:39		Kareem	OK
62	P4584-16	MCC0R6	SAM	11/18/24 00:43		Kareem	OK
63	P4584-17	MCC0R6D	DUP	11/18/24 00:47		Kareem	OK
64	P4584-16L	MCC0R6L	SD	11/18/24 00:52		Kareem	OK
65	P4584-18	MCC0R6S	MS	11/18/24 00:56	MS fail for Pb,Se	Kareem	OK
66	P4584-19	MCC0R8	SAM	11/18/24 01:01		Kareem	OK
67	P4607-01	MC0PI2	SAM	11/18/24 01:05		Kareem	OK
68	P4607-02	MC0PI6	SAM	11/18/24 01:10		Kareem	OK
69	P4607-03	MC0PI7	SAM	11/18/24 01:14		Kareem	OK
70	P4607-04	MC0PI8	SAM	11/18/24 01:19		Kareem	OK
71	P4607-05	MC0PI9	SAM	11/18/24 01:23		Kareem	OK
72	P4607-06	MCC0P0	SAM	11/18/24 01:27		Kareem	OK
73	P4607-07	MCC0P1	SAM	11/18/24 01:32		Kareem	OK
74	P4607-08	MCC0P2	SAM	11/18/24 01:36		Kareem	OK
75	P4607-09	MCC0P3	SAM	11/18/24 01:41		Kareem	OK
76	CCV015	CCV015	CCV	11/18/24 01:45		Kareem	OK
77	CCB015	CCB015	CCB	11/18/24 01:50		Kareem	OK
78	P4588-16DL	MYE5K5S	MS	11/18/24 01:54	MS fail for Se	Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133481

Review By	Sarabjit Jaswal	Review On	11/18/2024 7:27:18 AM
Supervise By	Mohan Bera	Supervise On	11/18/2024 7:28:03 AM

STD. NAME	STD REF.#
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136
ICV Standard	MP83143
CCV Standard	MP83146
ICSA Standard	MP83144,MP83145
CRI Standard	
LCS Standard	
Chk Standard	MP83148,MP83149

79	P4589-15DL	MYE5N2S	MS	11/18/24 01:59	MS fail for As,Pb,Zn	Kareem	OK
80	P4587-06DL	MYE5G9	SAM	11/18/24 02:03		Kareem	OK
81	P4587-07DL	MYE5G9D	DUP	11/18/24 02:07		Kareem	OK
82	P4587-06LDL	MYE5G9L	SD	11/18/24 02:12		Kareem	OK
83	P4587-08DL	MYE5G9S	MS	11/18/24 02:16		Kareem	OK
84	CCV016	CCV016	CCV	11/18/24 02:30		Kareem	OK
85	CCB016	CCB016	CCB	11/18/24 02:34		Kareem	OK
86	P4607-10	MCC0P4	SAM	11/18/24 02:39		Kareem	OK
87	P4607-11	MCC0P5	SAM	11/18/24 02:43		Kareem	OK
88	P4607-12	MCC0P6	SAM	11/18/24 02:48	Pb high	Kareem	Dilution
89	P4607-13	MCC0P7	SAM	11/18/24 02:52		Kareem	OK
90	P4607-14	MCC0P8	SAM	11/18/24 02:56		Kareem	OK
91	P4607-15	MCC0P9	SAM	11/18/24 03:01	Fe high	Kareem	Dilution
92	P4607-16	MCC0Q0	SAM	11/18/24 03:05		Kareem	OK
93	P4607-17	MCC0Q1	SAM	11/18/24 03:10		Kareem	OK
94	P4607-18	MCC0Q2	SAM	11/18/24 03:14		Kareem	OK
95	P4607-19	MCC0Q3	SAM	11/18/24 03:18		Kareem	OK
96	P4607-20	MCC0R7	SAM	11/18/24 03:23		Kareem	OK
97	P4607-21	MCC0R7D	DUP	11/18/24 03:27		Kareem	OK
98	P4607-20L	MCC0R7L	SD	11/18/24 03:32		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133481

Review By	Sarabjit Jaswal	Review On	11/18/2024 7:27:18 AM
Supervise By	Mohan Bera	Supervise On	11/18/2024 7:28:03 AM

STD. NAME	STD REF.#
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136
ICV Standard	MP83143
CCV Standard	MP83146
ICSA Standard	MP83144,MP83145
CRI Standard	
LCS Standard	
Chk Standard	MP83148,MP83149

99	P4607-22	MCC0R7S	MS	11/18/24 03:36	MS fail for Pb,Mn,Se,Ag,Tl,Zn	Kareem	OK
100	P4608-01	MCC0P0	SAM	11/18/24 03:40		Kareem	OK
101	P4608-02	MCC0P0D	DUP	11/18/24 03:45		Kareem	OK
102	P4608-01L	MCC0P0L	SD	11/18/24 03:49		Kareem	OK
103	P4608-03	MCC0P0S	MS	11/18/24 03:54	MS fail for As,Se,Zn	Kareem	OK
104	P4584-16A	MCC0R6A	PS	11/18/24 03:58	PS for Pb - 0.024ml (MP83156), Se - 0.07ml (MP83152) - 10 ml sample	Kareem	OK
105	P4588-14ADL	MYE5K5A	PS	11/18/24 04:03	PS for Se - 0.07ml (MP83152) - 10 ml sample	Kareem	OK
106	CCV017	CCV017	CCV	11/18/24 04:07		Kareem	OK
107	CCB017	CCB017	CCB	11/18/24 04:12		Kareem	OK
108	P4589-13ADL	MYE5N2A	PS	11/18/24 04:16	PS for As - 0.035 ml (MP83228), Pb- 0.011ml (MP83156), Zn - 0.0052ml (M5819) - 10ml sample	Kareem	OK
109	P4607-20A	MCC0R7A	PS	11/18/24 04:22	PS for Pb,Mn,Se,Tl,Zn	Kareem	OK
110	P4608-01A	MCC0P0A	PS	11/18/24 04:26	PS for As,Se,Zn	Kareem	OK
111	CCV018	CCV018	CCV	11/18/24 04:36		Kareem	OK
112	CCB018	CCB018	CCB	11/18/24 04:40		Kareem	OK