

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: SDG No.: MYE5J2
 Instrument ID: P7 Start Date : 11/12/2024
 Run Batch: LB133416 End Date : 11/13/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element	Q	Element	Q	Element	Q	Element	Q	Element	Q
		6Li		45Sc		89Y		103Rh		159Tb	
S0	1238	100		100		100		100		100	
S02	1245	104		103		102		103		102	
S03	1248	110		108		108		107		108	
S04	1251	110		107		107		106		108	
S05	1254	108		105		106		103		107	
S06	1256	106		103		106		102		107	
S07	1259	103		104		104		100		108	
S08	1302	90		94		94		84		94	
ICV001	1343	103		108		105		104		101	
ICB001	1346	103		109		106		106		103	
ICSA001	1349	95		106		107		100		107	
ICSAB001	1352	90		109		113		104		110	
CCV001	1359	93		111		110		101		108	
CCB001	1402	98		111		111		110		107	
ZZZZZZ	1407										
ZZZZZZ	1410										
ZZZZZZ	1413										
ZZZZZZ	1417										
ZZZZZZ	1420										
ZZZZZZ	1423										
ZZZZZZ	1426										
ZZZZZZ	1430										
ZZZZZZ	1433										
ZZZZZZ	1436										
ZZZZZZ	1439										
ZZZZZZ	1442										
ZZZZZZ	1445										
ZZZZZZ	1448										
ZZZZZZ	1451										
ZZZZZZ	1454										
ZZZZZZ	1457										
ZZZZZZ	1500										
ZZZZZZ	1504										
ZZZZZZ	1507										
ZZZZZZ	1510										
ZZZZZZ	1513										
ZZZZZZ	1516										
ZZZZZZ	1519										

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: SDG No.: MYE5J2
 Instrument ID: P7 Start Date : 11/12/2024
 Run Batch: LB133416 End Date : 11/13/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
ZZZZZZ	1521										
ZZZZZZ	1524										
ZZZZZZ	1528										
ZZZZZZ	1530										
ZZZZZZ	1533										
ZZZZZZ	1536										
ZZZZZZ	1539										
CCV002	1542	89		104		108		99		110	
CCB002	1545	95		105		107		108		108	
ZZZZZZ	1548										
ZZZZZZ	1551										
ZZZZZZ	1554										
ZZZZZZ	1557										
ZZZZZZ	1600										
ZZZZZZ	1603										
ZZZZZZ	1606										
ZZZZZZ	1610										
ZZZZZZ	1613										
ZZZZZZ	1616										
ZZZZZZ	1619										
ZZZZZZ	1622										
ZZZZZZ	1625										
ZZZZZZ	1628										
ZZZZZZ	1631										
ZZZZZZ	1634										
ZZZZZZ	1637										
ZZZZZZ	1640										
ZZZZZZ	1643										
ZZZZZZ	1646										
ZZZZZZ	1649										
ZZZZZZ	1652										
ZZZZZZ	1655										
ZZZZZZ	1658										
ZZZZZZ	1701										
ZZZZZZ	1704										
CCV003	1707	86		101		106		98		110	
CCB003	1710	91		103		107		108		110	
ZZZZZZ	1713										

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: SDG No.: MYE5J2
 Instrument ID: P7 Start Date : 11/12/2024
 Run Batch: LB133416 End Date : 11/13/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
ZZZZZZ	1716										
ZZZZZZ	1719										
ZZZZZZ	1722										
ZZZZZZ	1725										
ZZZZZZ	1728										
ZZZZZZ	1731										
ZZZZZZ	1734										
ZZZZZZ	1737										
ZZZZZZ	1740										
ZZZZZZ	1743										
ZZZZZZ	1746										
ZZZZZZ	1749										
ZZZZZZ	1752										
ZZZZZZ	1755										
ZZZZZZ	1758										
ZZZZZZ	1801										
ZZZZZZ	1805										
ZZZZZZ	1808										
ZZZZZZ	1811										
ZZZZZZ	1814										
ZZZZZZ	1817										
ZZZZZZ	1820										
ZZZZZZ	1823										
ZZZZZZ	1826										
ZZZZZZ	1829										
ZZZZZZ	1833										
ZZZZZZ	1836										
CCV004	1839	90		100		103		96		107	
CCB004	1842	98		106		109		109		111	
ZZZZZZ	1845										
ZZZZZZ	1848										
ZZZZZZ	1851										
ZZZZZZ	1854										
ZZZZZZ	1857										
ZZZZZZ	1900										
ZZZZZZ	1912										
ZZZZZZ	1915										
ZZZZZZ	1919										

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: SDG No.: MYE5J2
 Instrument ID: P7 Start Date : 11/12/2024
 Run Batch: LB133416 End Date : 11/13/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element		Element		Element		Element		Element	
		6Li	Q	45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q
ZZZZZZ	1922										
ZZZZZZ	1925										
ZZZZZZ	1928										
ZZZZZZ	1931										
ZZZZZZ	1934										
ZZZZZZ	1937										
ZZZZZZ	1941										
ZZZZZZ	1944										
ZZZZZZ	1948										
ZZZZZZ	1951										
ZZZZZZ	1954										
ZZZZZZ	1957										
ZZZZZZ	2000										
ZZZZZZ	2003										
ZZZZZZ	2007										
ZZZZZZ	2010										
ZZZZZZ	2013										
ZZZZZZ	2016										
CCV005	2019	91		104		105		97		106	
CCB005	2022	99		106		108		107		108	
ZZZZZZ	2025										
ZZZZZZ	2028										
ZZZZZZ	2031										
ZZZZZZ	2039										
ZZZZZZ	2042										
ZZZZZZ	2045										
ZZZZZZ	2048										
ZZZZZZ	2052										
ZZZZZZ	2055										
ZZZZZZ	2058										
ZZZZZZ	2101										
ZZZZZZ	2104										
ZZZZZZ	2107										
ZZZZZZ	2110										
ZZZZZZ	2113										
ZZZZZZ	2116										
ZZZZZZ	2119										
ZZZZZZ	2122										

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: SDG No.: MYE5J2
 Instrument ID: P7 Start Date : 11/12/2024
 Run Batch: LB133416 End Date : 11/13/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
ZZZZZZ	2125										
ZZZZZZ	2128										
ZZZZZZ	2131										
ZZZZZZ	2134										
ZZZZZZ	2138										
ZZZZZZ	2141										
ZZZZZZ	2144										
ZZZZZZ	2147										
ZZZZZZ	2150										
ZZZZZZ	2153										
CCV006	2158	87		102		104		97		107	
CCB006	2203	96		107		109		109		108	
ZZZZZZ	2206										
ZZZZZZ	2209										
ZZZZZZ	2212										
ZZZZZZ	2215										
ZZZZZZ	2218										
ZZZZZZ	2221										
ZZZZZZ	2224										
ZZZZZZ	2227										
PBS696	2230	90		105		110		110		112	
LCS696	2233	93		105		108		108		109	
LCSD696	2237	95		105		108		109		110	
MYE5J2	2240	91		108		114		107		112	
MYE5J3	2243	88		107		113		108		113	
MYE5J4	2246	85		106		112		108		113	
MYE5J5	2249	81		104		111		105		111	
MYE5J6	2252	80		104		113		105		112	
MYE5J7	2255	83		105		114		108		114	
MYE5J8	2259	83		106		114		108		115	
MYE5J9	2302	84		108		117		110		117	
MYE5K0	2305	84		106		114		107		115	
MYE5K1	2308	84		105		115		108		115	
MYE5K2	2311	81		106		115		107		114	
MYE5K3	2314	81		106		115		108		115	
MYE5K4	2317	81		105		115		109		115	
MYE5K5	2320	79		107		116		106		113	
MYE5K5D	2323	77		107		118		105		113	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: _____ SDG No.: MYE5J2
 Instrument ID: P7 Start Date : 11/12/2024
 Run Batch: LB133416 End Date : 11/13/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element	Q	Element	Q	Element	Q	Element	Q	Element	Q
		6Li		45Sc		89Y		103Rh		159Tb	
MYE5K5L	2326	84		104		110		108		111	
MYE5K5S	2329	78		107		118		105		114	
CCV007	2332	81		99		105		97		109	
CCB007	2335	88		102		108		108		110	
MYE5K5SRE	2338	80		110		119		105		112	
MYE5K6	2341	76		106		114		107		112	
MYE5K7	2344	77		106		115		108		113	
MYE5K8	2347	76		107		116		111		116	
MYE5K9	2350	77		109		117		110		115	
ZZZZZZ	2353										
ZZZZZZ	2357										
ZZZZZZ	0000										
ZZZZZZ	0003										
ZZZZZZ	0006										
ZZZZZZ	0009										
ZZZZZZ	0012										
ZZZZZZ	0015										
ZZZZZZ	0019										
ZZZZZZ	0022										
ZZZZZZ	0025										
ZZZZZZ	0028										
ZZZZZZ	0031										
ZZZZZZ	0034										
ZZZZZZ	0038										
MYE5K5	0047	91		108		112		107		109	
MYE5K5D	0050	86		108		111		108		110	
MYE5K5L	0053	89		107		111		110		110	
MYE5K5S	0057	91		109		113		109		112	
MYE5K5SRE	0100	86		102		107		103		105	
CCV008	0103	89		102		106		97		108	
CCB008	0105	96		105		108		108		110	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: SDG No.: MYE5J2
 Instrument ID: P7 Start Date : 11/12/2024
 Run Batch: LB133416 End Date : 11/13/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
S0	1238	100		100		100		100		100	
S02	1245	104		102		102		103		102	
S03	1248	106		106		107		107		106	
S04	1251	106		106		105		108		106	
S05	1254	102		103		100		106		105	
S06	1256	104		102		100		107		105	
S07	1259	104		104		99		107		107	
S08	1302	102		102		89		99		99	
ICV001	1343	110		109		107		105		104	
ICB001	1346	111		109		108		106		104	
ICSA001	1349	111		110		104		109		107	
ICSAB001	1352	114		113		107		111		109	
CCV001	1359	113		112		101		108		107	
CCB001	1402	112		112		111		108		107	
ZZZZZZ	1407										
ZZZZZZ	1410										
ZZZZZZ	1413										
ZZZZZZ	1417										
ZZZZZZ	1420										
ZZZZZZ	1423										
ZZZZZZ	1426										
ZZZZZZ	1430										
ZZZZZZ	1433										
ZZZZZZ	1436										
ZZZZZZ	1439										
ZZZZZZ	1442										
ZZZZZZ	1445										
ZZZZZZ	1448										
ZZZZZZ	1451										
ZZZZZZ	1454										
ZZZZZZ	1457										
ZZZZZZ	1500										
ZZZZZZ	1504										
ZZZZZZ	1507										
ZZZZZZ	1510										
ZZZZZZ	1513										
ZZZZZZ	1516										
ZZZZZZ	1519										

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: SDG No.: MYE5J2
 Instrument ID: P7 Start Date : 11/12/2024
 Run Batch: LB133416 End Date : 11/13/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
ZZZZZZ	1521										
ZZZZZZ	1524										
ZZZZZZ	1528										
ZZZZZZ	1530										
ZZZZZZ	1533										
ZZZZZZ	1536										
ZZZZZZ	1539										
CCV002	1542	104		105		98		106		104	
CCB002	1545	107		108		109		108		107	
ZZZZZZ	1548										
ZZZZZZ	1551										
ZZZZZZ	1554										
ZZZZZZ	1557										
ZZZZZZ	1600										
ZZZZZZ	1603										
ZZZZZZ	1606										
ZZZZZZ	1610										
ZZZZZZ	1613										
ZZZZZZ	1616										
ZZZZZZ	1619										
ZZZZZZ	1622										
ZZZZZZ	1625										
ZZZZZZ	1628										
ZZZZZZ	1631										
ZZZZZZ	1634										
ZZZZZZ	1637										
ZZZZZZ	1640										
ZZZZZZ	1643										
ZZZZZZ	1646										
ZZZZZZ	1649										
ZZZZZZ	1652										
ZZZZZZ	1655										
ZZZZZZ	1658										
ZZZZZZ	1701										
ZZZZZZ	1704										
CCV003	1707	103		104		98		107		106	
CCB003	1710	103		107		109		109		106	
ZZZZZZ	1713										

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: SDG No.: MYE5J2
 Instrument ID: P7 Start Date : 11/12/2024
 Run Batch: LB133416 End Date : 11/13/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
ZZZZZZ	1716										
ZZZZZZ	1719										
ZZZZZZ	1722										
ZZZZZZ	1725										
ZZZZZZ	1728										
ZZZZZZ	1731										
ZZZZZZ	1734										
ZZZZZZ	1737										
ZZZZZZ	1740										
ZZZZZZ	1743										
ZZZZZZ	1746										
ZZZZZZ	1749										
ZZZZZZ	1752										
ZZZZZZ	1755										
ZZZZZZ	1758										
ZZZZZZ	1801										
ZZZZZZ	1805										
ZZZZZZ	1808										
ZZZZZZ	1811										
ZZZZZZ	1814										
ZZZZZZ	1817										
ZZZZZZ	1820										
ZZZZZZ	1823										
ZZZZZZ	1826										
ZZZZZZ	1829										
ZZZZZZ	1833										
ZZZZZZ	1836										
CCV004	1839	101		101		95		104		102	
CCB004	1842	104		106		107		108		106	
ZZZZZZ	1845										
ZZZZZZ	1848										
ZZZZZZ	1851										
ZZZZZZ	1854										
ZZZZZZ	1857										
ZZZZZZ	1900										
ZZZZZZ	1912										
ZZZZZZ	1915										
ZZZZZZ	1919										

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: SDG No.: MYE5J2
 Instrument ID: P7 Start Date : 11/12/2024
 Run Batch: LB133416 End Date : 11/13/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
ZZZZZZ	1922										
ZZZZZZ	1925										
ZZZZZZ	1928										
ZZZZZZ	1931										
ZZZZZZ	1934										
ZZZZZZ	1937										
ZZZZZZ	1941										
ZZZZZZ	1944										
ZZZZZZ	1948										
ZZZZZZ	1951										
ZZZZZZ	1954										
ZZZZZZ	1957										
ZZZZZZ	2000										
ZZZZZZ	2003										
ZZZZZZ	2007										
ZZZZZZ	2010										
ZZZZZZ	2013										
ZZZZZZ	2016										
CCV005	2019	102		103		95		103		102	
CCB005	2022	107		109		110		108		108	
ZZZZZZ	2025										
ZZZZZZ	2028										
ZZZZZZ	2031										
ZZZZZZ	2039										
ZZZZZZ	2042										
ZZZZZZ	2045										
ZZZZZZ	2048										
ZZZZZZ	2052										
ZZZZZZ	2055										
ZZZZZZ	2058										
ZZZZZZ	2101										
ZZZZZZ	2104										
ZZZZZZ	2107										
ZZZZZZ	2110										
ZZZZZZ	2113										
ZZZZZZ	2116										
ZZZZZZ	2119										
ZZZZZZ	2122										

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: SDG No.: MYE5J2
 Instrument ID: P7 Start Date : 11/12/2024
 Run Batch: LB133416 End Date : 11/13/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
ZZZZZZ	2125										
ZZZZZZ	2128										
ZZZZZZ	2131										
ZZZZZZ	2134										
ZZZZZZ	2138										
ZZZZZZ	2141										
ZZZZZZ	2144										
ZZZZZZ	2147										
ZZZZZZ	2150										
ZZZZZZ	2153										
CCV006	2158	102		102		95		104		102	
CCB006	2203	105		106		107		108		105	
ZZZZZZ	2206										
ZZZZZZ	2209										
ZZZZZZ	2212										
ZZZZZZ	2215										
ZZZZZZ	2218										
ZZZZZZ	2221										
ZZZZZZ	2224										
ZZZZZZ	2227										
PBS696	2230	104		105		108		107		105	
LCS696	2233	104		105		107		106		104	
LCSD696	2237	105		105		107		106		104	
MYE5J2	2240	106		108		106		107		107	
MYE5J3	2243	105		108		106		108		106	
MYE5J4	2246	106		107		107		108		106	
MYE5J5	2249	104		106		104		107		105	
MYE5J6	2252	106		109		106		108		106	
MYE5J7	2255	105		108		107		110		107	
MYE5J8	2259	106		108		106		109		107	
MYE5J9	2302	105		110		108		109		107	
MYE5K0	2305	105		109		107		109		107	
MYE5K1	2308	107		109		107		108		107	
MYE5K2	2311	106		109		107		109		107	
MYE5K3	2314	107		109		108		110		107	
MYE5K4	2317	105		108		106		108		107	
MYE5K5	2320	109		112		104		107		105	
MYE5K5D	2323	109		113		103		107		105	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: _____ SDG No.: MYE5J2
 Instrument ID: P7 Start Date : 11/12/2024
 Run Batch: LB133416 End Date : 11/13/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
MYE5K5L	2326	104		105		106		108		104	
MYE5K5S	2329	106		112		103		107		105	
CCV007	2332	99		101		96		104		103	
CCB007	2335	99		100		102		103		101	
MYE5K5SRE	2338	108		113		105		108		106	
MYE5K6	2341	107		111		107		109		107	
MYE5K7	2344	107		111		109		111		108	
MYE5K8	2347	107		111		109		110		107	
MYE5K9	2350	107		111		109		110		108	
ZZZZZZ	2353										
ZZZZZZ	2357										
ZZZZZZ	0000										
ZZZZZZ	0003										
ZZZZZZ	0006										
ZZZZZZ	0009										
ZZZZZZ	0012										
ZZZZZZ	0015										
ZZZZZZ	0019										
ZZZZZZ	0022										
ZZZZZZ	0025										
ZZZZZZ	0028										
ZZZZZZ	0031										
ZZZZZZ	0034										
ZZZZZZ	0038										
MYE5K5	0047	107		108		107		106		104	
MYE5K5D	0050	106		108		107		107		105	
MYE5K5L	0053	105		107		109		107		106	
MYE5K5S	0057	108		110		109		108		106	
MYE5K5SRE	0100	107		107		106		107		105	
CCV008	0103	103		103		97		105		104	
CCB008	0105	105		106		108		108		105	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: SDG No.: MYE5J2
 Instrument ID: P7 Start Date : 11/13/2024
 Run Batch: LB133439 End Date : 11/13/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element	Q	Element	Q	Element	Q	Element	Q	Element	Q
		6Li		45Sc		89Y		103Rh		159Tb	
S0	1625	100		100		100		100		100	
S02	1632	102		101		103		102		102	
S03	1635	108		100		101		101		102	
S04	1638	110		98		99		99		103	
S05	1641	107		94		96		95		101	
S06	1643	108		93		97		94		102	
S07	1646	101		92		95		92		102	
S08	1649	92		85		86		78		91	
ICV007	1718	107		99		98		98		100	
ICB007	1723	104		96		98		98		98	
ICSA007	1749	94		93		95		90		100	
ICSAB007	1801	91		94		97		93		102	
CCV031	1815	91		93		96		89		101	
CCB031	1821	97		95		97		97		100	
ZZZZZZ	1825										
ZZZZZZ	1828										
ZZZZZZ	1831										
ZZZZZZ	1834										
ZZZZZZ	1837										
ZZZZZZ	1840										
ZZZZZZ	1843										
ZZZZZZ	1846										
ZZZZZZ	1849										
ZZZZZZ	1852										
ZZZZZZ	1855										
ZZZZZZ	1858										
ZZZZZZ	1901										
ZZZZZZ	1904										
ZZZZZZ	1907										
ZZZZZZ	1911										
ZZZZZZ	1914										
ZZZZZZ	1917										
ZZZZZZ	1920										
ZZZZZZ	1923										
ZZZZZZ	1926										
ZZZZZZ	1929										
ZZZZZZ	1932										
ZZZZZZ	1935										

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: _____ SDG No.: MYE5J2
 Instrument ID: P7 Start Date : 11/13/2024
 Run Batch: LB133439 End Date : 11/13/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element	Q	Element	Q	Element	Q	Element	Q	Element	Q
		6Li		45Sc		89Y		103Rh		159Tb	
CCV032	1946	94		93		96		90		102	
CCB032	1950	101		94		99		99		104	
MYE5K5	1954	93		102		109		98		106	
MYE5K5D	1957	84		101		108		98		104	
MYE5K5L	2000	92		98		104		102		107	
MYE5K5S	2003	89		102		110		100		107	
MYE5K5SRE	2009	93		103		109		99		108	
MYE5K5	2013	99		101		106		104		110	
MYE5K5D	2016	95		97		102		99		105	
MYE5K5L	2019	93		95		100		99		103	
MYE5K5S	2023	93		98		103		101		105	
MYE5K5SRE	2026	97		100		105		103		108	
ZZZZZZ	2029										
ZZZZZZ	2033										
ZZZZZZ	2036										
ZZZZZZ	2040										
ZZZZZZ	2043										
ZZZZZZ	2046										
ZZZZZZ	2049										
CCV033	2052	95		94		98		91		107	
CCB033	2055	104		93		98		99		105	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: SDG No.: MYE5J2
 Instrument ID: P7 Start Date : 11/13/2024
 Run Batch: LB133439 End Date : 11/13/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
S0	1625	100		100		100		100		100	
S02	1632	100		102		102		103		102	
S03	1635	100		102		102		104		103	
S04	1638	98		98		97		101		101	
S05	1641	96		95		94		100		99	
S06	1643	94		94		92		99		100	
S07	1646	95		94		91		99		99	
S08	1649	88		88		79		91		91	
ICV007	1718	100		99		99		100		100	
ICB007	1723	99		98		98		99		99	
ICSA007	1749	96		96		92		101		101	
ICSAB007	1801	95		97		93		100		101	
CCV031	1815	94		96		90		99		99	
CCB031	1821	96		98		99		100		100	
ZZZZZZ	1825										
ZZZZZZ	1828										
ZZZZZZ	1831										
ZZZZZZ	1834										
ZZZZZZ	1837										
ZZZZZZ	1840										
ZZZZZZ	1843										
ZZZZZZ	1846										
ZZZZZZ	1849										
ZZZZZZ	1852										
ZZZZZZ	1855										
ZZZZZZ	1858										
ZZZZZZ	1901										
ZZZZZZ	1904										
ZZZZZZ	1907										
ZZZZZZ	1911										
ZZZZZZ	1914										
ZZZZZZ	1917										
ZZZZZZ	1920										
ZZZZZZ	1923										
ZZZZZZ	1926										
ZZZZZZ	1929										
ZZZZZZ	1932										
ZZZZZZ	1935										

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: SDG No.: MYE5J2
 Instrument ID: P7 Start Date : 11/13/2024
 Run Batch: LB133439 End Date : 11/13/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
CCV032	1946	94		94		89		98		99	
CCB032	1950	95		96		98		100		100	
MYE5K5	1954	104		107		99		103		103	
MYE5K5D	1957	104		108		100		103		102	
MYE5K5L	2000	100		102		101		104		103	
MYE5K5S	2003	102		106		100		104		104	
MYE5K5SRE	2009	101		105		98		103		103	
MYE5K5	2013	99		99		99		102		102	
MYE5K5D	2016	97		99		99		102		101	
MYE5K5L	2019	97		98		100		101		102	
MYE5K5S	2023	100		102		101		104		103	
MYE5K5SRE	2026	100		102		102		104		105	
ZZZZZZ	2029										
ZZZZZZ	2033										
ZZZZZZ	2036										
ZZZZZZ	2040										
ZZZZZZ	2043										
ZZZZZZ	2046										
ZZZZZZ	2049										
CCV033	2052	93		95		89		99		99	
CCB033	2055	95		97		99		101		101	