

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No. : SDG No.: MYE5L0
 Initial Calibration Verification Source : MP83041
 Continuing Calibration Verification Source : MP83042
 Run Batch: LB133416 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV001				ID: CCV001				ID: CCV002		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony	200	210	105	1	500	520	104	1	510	101	2
Arsenic	200	210	103	1	500	510	103	0	520	103	1
Barium	100	110	106	1	2500	2500	102	2	2500	99	2
Beryllium	100	99	99	2	500	530	105	4	520	103	3
Cadmium	100	110	110	1	500	530	106	1	510	102	2
Chromium	100	100	104	1	500	490	99	1	510	102	2
Cobalt	100	110	106	2	500	520	103	2	530	106	1
Copper	100	100	100	1	5000	5000	99	1	5200	103	1
Lead	200	200	101	1	2500	2500	102	0	2600	102	2
Nickel	110	110	96	1	500	510	102	2	520	105	2
Selenium	200	200	102	0	500	500	100	1	490	98	1
Silver	50.0	54	109	0	500	520	104	2	500	100	2
Thallium	210	200	95	2	500	510	102	0	510	102	2
Vanadium	100	99	99	1	500	500	100	1	520	103	2
Zinc	200	200	99	2	5000	5000	99	0	5100	103	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No. : SDG No.: MYE5L0
 Initial Calibration Verification Source : MP83041
 Continuing Calibration Verification Source : MP83042
 Run Batch: LB133416 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV003				ID: CCV004		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony					500	500	100	1	490	97	1
Arsenic					500	510	102	0	510	101	1
Barium					2500	2500	99	1	2500	99	1
Beryllium					500	510	102	3	500	101	1
Cadmium					500	510	102	1	500	100	1
Chromium					500	500	100	1	500	100	2
Cobalt					500	520	104	2	520	104	2
Copper					5000	5100	102	2	5100	103	2
Lead					2500	2500	102	1	2500	101	0
Nickel					500	520	103	1	520	103	1
Selenium					500	500	99	1	490	99	1
Silver					500	500	99	1	500	99	0
Thallium					500	510	102	1	510	102	1
Vanadium					500	510	101	2	500	101	2
Zinc					5000	5000	100	2	5000	101	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No. : SDG No.: MYE5L0
 Initial Calibration Verification Source : MP83041
 Continuing Calibration Verification Source : MP83042
 Run Batch: LB133416 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV005				ID: CCV006		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony					500	480	97	0	490	99	2
Arsenic					500	500	101	1	510	102	0
Barium					2500	2400	98	1	2500	100	1
Beryllium					500	490	98	1	510	102	2
Cadmium					500	490	98	1	500	100	1
Chromium					500	500	100	0	500	100	0
Cobalt					500	510	103	1	510	103	1
Copper					5000	5100	101	2	5100	102	1
Lead					2500	2500	99	1	2500	101	1
Nickel					500	510	101	1	510	103	1
Selenium					500	480	96	1	490	99	0
Silver					500	490	98	1	500	99	1
Thallium					500	500	100	0	520	104	1
Vanadium					500	500	99	1	500	100	0
Zinc					5000	5000	100	2	5100	102	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No. : SDG No.: MYE5L0
 Initial Calibration Verification Source : MP83041
 Continuing Calibration Verification Source : MP83042
 Run Batch: LB133416 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV007				ID: CCV008		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony					500	490	98	1	490	99	1
Arsenic					500	510	102	1	510	101	1
Barium					2500	2400	97	1	2400	98	0
Beryllium					500	480	96	2	500	100	1
Cadmium					500	500	99	0	500	99	1
Chromium					500	500	100	1	490	98	1
Cobalt					500	520	104	1	520	103	1
Copper					5000	5100	102	0	5100	101	1
Lead					2500	2500	100	1	2500	99	1
Nickel					500	520	104	1	510	103	2
Selenium					500	500	99	1	490	99	1
Silver					500	490	98	0	490	97	0
Thallium					500	500	100	2	500	100	2
Vanadium					500	500	100	1	490	98	2
Zinc					5000	5100	103	0	5100	101	1