

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51817 MA No. : SDG No.: MYE5N3Initial Calibration Verification Source : MP83143Continuing Calibration Verification Source : MP83146Run Batch: LB133461 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV012				ID: CCV087				ID: CCV088		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2500	99	2	400000	400000	100	3	400000	101	0
Antimony	1000	920	92	1	5000	4800	95	0	4700	95	0
Arsenic	1000	1100	105	1	5000	4900	99	0	4900	98	0
Barium	520	510	98	2	10000	10000	101	3	10000	102	0
Beryllium	510	530	104	1	500	500	100	3	510	101	0
Cadmium	510	510	99	0	2500	2500	99	0	2500	98	0
Calcium	10000	10000	101	2	400000	410000	102	3	410000	103	0
Chromium	520	560	108	0	15500	15000	99	0	15000	98	0
Cobalt	520	500	96	0	2500	2500	99	0	2500	99	0
Copper	510	480	95	1	15000	15000	99	3	15000	100	0
Iron	10000	11000	109	2	400000	410000	103	3	420000	104	0
Lead	1000	980	98	1	25000	25000	98	0	24000	97	0
Magnesium	6000	5900	98	3	400000	400000	101	3	410000	102	0
Manganese	520	530	102	2	15000	15000	99	4	15000	100	0
Nickel	530	510	96	0	2500	2500	99	0	2500	98	0
Potassium	9900	11000	107	2	150000	150000	101	3	150000	102	0
Selenium	1000	990	99	1	5000	4800	95	0	4700	94	0
Silver	250	240	97	0	1250	1300	100	1	1200	100	0
Sodium	10000	10000	101	2	400000	410000	102	3	410000	103	0
Thallium	1000	1100	107	0	5000	4800	96	1	4800	95	1
Vanadium	500	530	106	3	2500	2500	100	4	2500	101	0
Zinc	1000	970	97	0	15000	15000	97	0	14000	96	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51817 MA No. : SDG No.: MYE5N3Initial Calibration Verification Source : MP83143Continuing Calibration Verification Source : MP83146Run Batch: LB133461 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV089				ID: CCV090		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	101	1	410000	101	0
Antimony					5000	4800	96	0	4800	97	1
Arsenic					5000	5000	100	0	5000	99	1
Barium					10000	10000	103	1	10000	103	0
Beryllium					500	510	102	1	510	102	0
Cadmium					2500	2500	100	1	2500	99	0
Calcium					400000	410000	102	1	410000	102	0
Chromium					15500	16000	100	1	15000	100	0
Cobalt					2500	2500	100	1	2500	100	0
Copper					15000	15000	101	0	15000	101	0
Iron					400000	410000	103	1	410000	104	0
Lead					25000	25000	99	1	25000	99	0
Magnesium					400000	410000	102	1	410000	102	0
Manganese					15000	15000	100	1	15000	100	0
Nickel					2500	2500	100	1	2500	100	0
Potassium					150000	160000	104	1	160000	104	0
Selenium					5000	4800	96	1	4800	97	0
Silver					1250	1300	101	1	1300	102	0
Sodium					400000	410000	103	1	420000	104	0
Thallium					5000	4900	97	1	4800	96	0
Vanadium					2500	2500	101	1	2500	102	0
Zinc					15000	15000	98	2	15000	98	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51817 MA No. : SDG No.: MYE5N3Initial Calibration Verification Source : MP83143Continuing Calibration Verification Source : MP83146Run Batch: LB133461 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV091				ID: CCV092		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	101	0	400000	101	0
Antimony					5000	4700	95	1	4800	96	0
Arsenic					5000	4900	98	2	5000	99	0
Barium					10000	10000	103	0	10000	102	0
Beryllium					500	510	101	1	510	101	0
Cadmium					2500	2500	98	1	2500	99	0
Calcium					400000	410000	102	0	400000	101	0
Chromium					15500	15000	99	0	15000	100	0
Cobalt					2500	2500	99	1	2500	100	0
Copper					15000	15000	101	0	15000	101	0
Iron					400000	410000	102	0	410000	103	0
Lead					25000	24000	97	1	24000	98	0
Magnesium					400000	400000	101	0	400000	101	0
Manganese					15000	15000	99	0	15000	99	0
Nickel					2500	2500	99	1	2500	100	0
Potassium					150000	150000	103	0	160000	104	0
Selenium					5000	4700	95	1	4900	97	0
Silver					1250	1300	100	0	1300	101	0
Sodium					400000	410000	103	0	410000	103	0
Thallium					5000	4800	96	1	4800	96	0
Vanadium					2500	2500	100	0	2500	100	0
Zinc					15000	15000	97	0	15000	97	1