

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133461

Review By	Sarabjit Jaswal	Review On	11/15/2024 4:33:08 AM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard	MP83141		
Chk Standard	MP83148,MP83149		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/14/24 15:50		Kareem	OK
2	S1	S01	CAL2	11/14/24 15:54		Kareem	OK
3	S2	S02	CAL3	11/14/24 15:59		Kareem	OK
4	S3	S03	CAL4	11/14/24 16:03		Kareem	OK
5	S4	S04	CAL5	11/14/24 16:08		Kareem	OK
6	S5	S05	CAL6	11/14/24 16:12		Kareem	OK
7	S6	S06	CAL7	11/14/24 16:17		Kareem	OK
8	ICV012	ICV012	ICV	11/14/24 16:21		Kareem	OK
9	ICB012	ICB012	ICB	11/14/24 16:38	Inst. paused for review	Kareem	OK
10	ICSA012	ICSA012	ICSA	11/14/24 16:43		Kareem	OK
11	ICSAB012	ICSAB012	ICSAB	11/14/24 16:47		Kareem	OK
12	CCV087	CCV087	CCV	11/14/24 16:52		Kareem	OK
13	CCB087	CCB087	CCB	11/14/24 16:56		Kareem	OK
14	PB164636BL	PBS636	MB	11/14/24 17:00		Kareem	OK
15	PB164636BS	LCS636	LCS	11/14/24 17:05		Kareem	OK
16	PB164636BSD	LCSD636	LCSD	11/14/24 17:09		Kareem	OK
17	PB164656BL	PBS656	MB	11/14/24 17:14		Kareem	OK
18	PB164656BS	LCS656	LCS	11/14/24 17:19		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133461

Review By	Sarabjit Jaswal	Review On	11/15/2024 4:33:08 AM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard	MP83141		
Chk Standard	MP83148,MP83149		

19	PB164656BSD	LCSD656	LCSD	11/14/24 17:23		Kareem	OK
20	PB164658BL	PBS658	MB	11/14/24 17:28		Kareem	OK
21	PB164658BS	LCS658	LCS	11/14/24 17:32		Kareem	OK
22	PB164658BSD	LCSD658	LCSD	11/14/24 17:37		Kareem	OK
23	PB164655BL	PBS655	MB	11/14/24 17:41	NOT USE	Kareem	Not Ok
24	PB164655BS	LCS655	LCS	11/14/24 17:46	NOT USE	Kareem	Not Ok
25	PB164655BSD	LCSD655	LCSD	11/14/24 17:50	NOT USE	Kareem	Not Ok
26	PB164715BL	PBS715	MB	11/14/24 17:55		Kareem	OK
27	PB164715BS	LCS715	LCS	11/14/24 17:59		Kareem	OK
28	PB164715BSD	LCSD715	LCSD	11/14/24 18:04		Kareem	OK
29	P4572-11DL	MYE586	SAM	11/14/24 18:08		Kareem	OK
30	P4572-12DL	MYE586D	DUP	11/14/24 18:13		Kareem	OK
31	P4572-11LDL	MYE586L	SD	11/14/24 18:17		Kareem	OK
32	P4572-13DL	MYE586S	MS	11/14/24 18:22	MS fail for Cu,Zn	Kareem	OK
33	P4573-07DL	MYE5A2	SAM	11/14/24 18:26		Kareem	OK
34	CCV088	CCV088	CCV	11/14/24 18:30		Kareem	OK
35	CCB088	CCB088	CCB	11/14/24 18:35		Kareem	OK
36	P4573-08DL	MYE5A2D	DUP	11/14/24 18:39		Kareem	OK
37	P4573-07LDL	MYE5A2L	SD	11/14/24 18:44		Kareem	OK
38	P4573-09DL	MYE5A2S	MS	11/14/24 18:48	MS fail for As	Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133461

Review By	Sarabjit Jaswal	Review On	11/15/2024 4:33:08 AM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard	MP83141		
Chk Standard	MP83148,MP83149		

39	P4590-01DL	MYE5N3	SAM	11/14/24 18:52		Kareem	OK
40	P4590-02DL	MYE5N3D	DUP	11/14/24 18:57		Kareem	OK
41	P4590-01LDL	MYE5N3L	SD	11/14/24 19:01		Kareem	OK
42	P4590-03DL	MYE5N3S	MS	11/14/24 19:06	MS fail for As	Kareem	OK
43	P4591-01DL	MYE5N4	SAM	11/14/24 19:10		Kareem	OK
44	P4591-02DL	MYE5N4D	DUP	11/14/24 19:15		Kareem	OK
45	P4591-01LDL	MYE5N4L	SD	11/14/24 19:19		Kareem	OK
46	P4591-03DL	MYE5N4S	MS	11/14/24 19:23		Kareem	OK
47	P4572-14DL	MYE587	SAM	11/14/24 19:28		Kareem	OK
48	P4572-15DL	MYE588	SAM	11/14/24 19:32		Kareem	OK
49	P4572-16DL	MYE589	SAM	11/14/24 19:37		Kareem	OK
50	P4572-17DL	MYE590	SAM	11/14/24 19:41		Kareem	OK
51	P4572-18DL	MYE591	SAM	11/14/24 19:46		Kareem	OK
52	P4572-19DL	MYE592	SAM	11/14/24 19:50		Kareem	OK
53	P4572-20DL	MYE593	SAM	11/14/24 19:54		Kareem	OK
54	P4572-21DL	MYE594	SAM	11/14/24 19:59		Kareem	OK
55	P4572-22DL	MYE595	SAM	11/14/24 20:03		Kareem	OK
56	CCV089	CCV089	CCV	11/14/24 20:08		Kareem	OK
57	CCB089	CCB089	CCB	11/14/24 20:12		Kareem	OK
58	P4591-04DL	MYE5N5	SAM	11/14/24 20:16		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133461

Review By	Sarabjit Jaswal	Review On	11/15/2024 4:33:08 AM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard	MP83141		
Chk Standard	MP83148,MP83149		

59	P4571-01ADL	MYE543A	PS	11/14/24 20:21	PS for As,Pb	Kareem	OK
60	P4573-01DL	MYE596	SAM	11/14/24 20:25		Kareem	OK
61	P4573-02DL	MYE597	SAM	11/14/24 20:30		Kareem	OK
62	P4573-03DL	MYE598	SAM	11/14/24 20:34		Kareem	OK
63	P4573-04DL	MYE599	SAM	11/14/24 20:39		Kareem	OK
64	P4573-05DL	MYE5A0	SAM	11/14/24 20:43		Kareem	OK
65	P4573-06DL	MYE5A1	SAM	11/14/24 20:47		Kareem	OK
66	P4573-07REDL	MYE5A2	SAM	11/14/24 20:52	NOT USE	Kareem	Not Ok
67	P4573-08REDL	MYE5A2D	DUP	11/14/24 20:56	NOT USE	Kareem	Not Ok
68	P4573-07LREDL	MYE5A2L	SD	11/14/24 21:01	NOT USE	Kareem	Not Ok
69	P4573-09REDL	MYE5A2S	MS	11/14/24 21:05	NOT USE	Kareem	Not Ok
70	P4573-10DL	MYE5A3	SAM	11/14/24 21:10		Kareem	OK
71	P4573-11DL	MYE5A4	SAM	11/14/24 21:14		Kareem	OK
72	P4573-12DL	MYE5A5	SAM	11/14/24 21:18		Kareem	OK
73	P4573-13DL	MYE5A6	SAM	11/14/24 21:23		Kareem	OK
74	P4573-14DL	MYE5A7	SAM	11/14/24 21:27		Kareem	OK
75	P4573-15DL	MYE5A8	SAM	11/14/24 21:32		Kareem	OK
76	P4573-16DL	MYE5A9	SAM	11/14/24 21:36		Kareem	OK
77	P4573-17DL	MYE5B0	SAM	11/14/24 21:41		Kareem	OK
78	CCV090	CCV090	CCV	11/14/24 21:45		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133461

Review By	Sarabjit Jaswal	Review On	11/15/2024 4:33:08 AM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136
ICV Standard	MP83143
CCV Standard	MP83146
ICSA Standard	MP83144,MP83145
CRI Standard	
LCS Standard	MP83141
Chk Standard	MP83148,MP83149

79	CCB090	CCB090	CCB	11/14/24 21:49		Kareem	OK
80	P4573-18DL	MYE5B1	SAM	11/14/24 21:54		Kareem	OK
81	P4573-19DL	MYE5B2	SAM	11/14/24 22:07		Kareem	OK
82	CCV091	CCV091	CCV	11/14/24 22:25		Kareem	OK
83	CCB091	CCB091	CCB	11/14/24 22:29		Kareem	OK
84	P4572-11ADL	MYE586A	PS	11/14/24 23:06	PS for Cu-0.009ml (M5751), Zn-0.0012ml (M5819) - 10 ml sample	Kareem	OK
85	P4573-07ADL	MYE5A2A	PS	11/14/24 23:11	PS for As-0.003ml (M5801) - 10 ml sample	Kareem	OK
86	P4590-01ADL	MYE5N3A	PS	11/14/24 23:15	PS for As-0.0023ml (M5801) - 10 ml sample	Kareem	OK
87	CCV092	CCV092	CCV	11/14/24 23:20		Kareem	OK
88	CCB092	CCB092	CCB	11/14/24 23:24		Kareem	OK