

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: _____ SDG No.: MYE5N4
 Instrument ID: P5 Start Date: 11/14/2024
 Analytical Method: ICP-AES Run Batch: LB133461
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	0	0.010	0	200	210	5	38750	37000	-5
Antimony	0	0.020	0	60	53	-12	1250	1100	-12
Arsenic	0	0.0	0	10	9.3	-7	1250	1100	-12
Barium	0	-0.010	0	200	210	5	2500	2500	0
Beryllium	0	-0	0	5	5.7	14	125	130	4
Cadmium	0	-0	0	5	5.5	10	625	620	-1
Calcium	0	-0.83	0	5000	5600	12	62500	64000	2
Chromium	0	-0	0	10	9.9	-1	3875	4000	3
Cobalt	0	0.0	0	50	50	0	625	610	-2
Copper	0	0.010	0	25	19	-24	3750	3800	1
Iron	0	-0.040	0	100	120	20	37500	40000	7
Lead	0	0.0	0	10	7.2	-28	6250	6100	-2
Magnesium	0	-0.19	0	5000	5400	8	62500	61000	-2
Manganese	0	-0	0	15	16	7	3750	3900	4
Nickel	0	-0	0	40	42	5	625	620	-1
Potassium	0	0.43	0	5000	5000	0	37500	36000	-4
Selenium	0	0.010	0	35	31	-11	1250	1100	-12
Silver	0	-0	0	10	13	30	312.5	310	-1
Sodium	0	-0.24	0	5000	5300	6	37500	37000	-1
Thallium	0	-0	0	25	25	0	1250	1300	4
Vanadium	0	-0	0	50	53	6	625	630	1
Zinc	0	0.010	0	60	57	-5	3750	3400	-9

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: _____ SDG No.: MYE5N4
 Instrument ID: P5 Start Date: 11/14/2024
 Analytical Method: ICP-AES Run Batch: LB133461
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	77500	75000	-3	155000	150000	-3	310000	310000	0
Antimony	2500	2200	-12	5000	4700	-6	10000	11000	10
Arsenic	2500	2400	-4	5000	4800	-4	10000	10000	0
Barium	5000	5000	0	10000	9900	-1	20000	20000	0
Beryllium	250	250	0	500	490	-2	1000	1000	0
Cadmium	1250	1300	4	2500	2500	0	5000	5000	0
Calcium	125000	130000	4	250000	250000	0	500000	510000	2
Chromium	7750	7900	2	15500	15000	-3	31000	31000	0
Cobalt	1250	1200	-4	2500	2500	0	5000	5100	2
Copper	7500	7500	0	15000	15000	0	30000	30000	0
Iron	75000	79000	5	150000	150000	0	300000	310000	3
Lead	12500	12000	-4	25000	25000	0	50000	51000	2
Magnesium	125000	120000	-4	250000	250000	0	500000	510000	2
Manganese	7500	7700	3	15000	15000	0	30000	30000	0
Nickel	1250	1300	4	2500	2500	0	5000	5000	0
Potassium	75000	72000	-4	150000	150000	0	300000	310000	3
Selenium	2500	2300	-8	5000	4700	-6	10000	11000	10
Silver	625	610	-2	1250	1200	-4	2500	2600	4
Sodium	75000	74000	-1	150000	150000	0	300000	310000	3
Thallium	2500	2600	4	5000	5000	0	10000	9900	-1
Vanadium	1250	1200	-4	2500	2500	0	5000	5000	0
Zinc	7500	7000	-7	15000	14000	-7	30000	32000	7

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51817 MA No.: SDG No.: MYE5N4
Instrument ID: P5 Start Date: 11/14/2024
Analytical Method: ICP-AES Run Batch: LB133461
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	800000	810000	1						
Calcium	800000	780000	-3						
Iron	800000	780000	-3						
Magnesium	800000	800000	0						
Sodium	800000	800000	0						