



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Tully Construction Co., Inc.

SDG No.: P4595

Project: Pitkin Yard

RunNo.: LB133178

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.02	7	100	90-110	10/28/2024
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	10/28/2024
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	10/28/2024
Sample ID: Corrosivity	CCV3	pH	2.01	2.00	101	90-110	10/28/2024

Initial and Continuing Calibration Verification

Client: Tully Construction Co., Inc.

SDG No.: P4595

Project: Pitkin Yard

RunNo.: LB133190

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.099	0.099	100	85-115	10/29/2024
Sample ID: CCV1 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	10/29/2024
Sample ID: CCV2 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	10/29/2024
Sample ID: CCV3 Reactive Cyanide	mg/L	0.26	0.25	104	90-110	10/29/2024