

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc. **SDG No.:** P4595
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P4595 **SAS No.:** P4595
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV57	Mercury	4.16	4.0	104	90 - 110	CV	10/29/2024	10:32	LB133185

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc. **SDG No.:** P4595
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P4595 **SAS No.:** P4595
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV86	Mercury	5.13	5.0	103	90 - 110	CV	10/29/2024	10:36	LB133185
CCV87	Mercury	5.18	5.0	104	90 - 110	CV	10/29/2024	11:04	LB133185
CCV88	Mercury	5.28	5.0	106	90 - 110	CV	10/29/2024	11:32	LB133185
CCV89	Mercury	5.26	5.0	105	90 - 110	CV	10/29/2024	11:57	LB133185

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc. **SDG No.:** P4595

Contract: TULL02 **Lab Code:** CHEM **Case No.:** P4595 **SAS No.:** P4595

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	2430	2500	97	90 - 110	P	10/30/2024	16:59	LB133235
	Antimony	1020	1000	102	90 - 110	P	10/30/2024	16:59	LB133235
	Arsenic	963	1000	96	90 - 110	P	10/30/2024	16:59	LB133235
	Barium	477	520	92	90 - 110	P	10/30/2024	16:59	LB133235
	Beryllium	486	510	95	90 - 110	P	10/30/2024	16:59	LB133235
	Cadmium	482	510	94	90 - 110	P	10/30/2024	16:59	LB133235
	Calcium	9500	10000	95	90 - 110	P	10/30/2024	16:59	LB133235
	Chromium	526	520	101	90 - 110	P	10/30/2024	16:59	LB133235
	Cobalt	517	520	99	90 - 110	P	10/30/2024	16:59	LB133235
	Copper	534	510	105	90 - 110	P	10/30/2024	16:59	LB133235
	Iron	9960	10000	100	90 - 110	P	10/30/2024	16:59	LB133235
	Lead	972	1000	97	90 - 110	P	10/30/2024	16:59	LB133235
	Magnesium	5480	6000	91	90 - 110	P	10/30/2024	16:59	LB133235
	Manganese	486	520	94	90 - 110	P	10/30/2024	16:59	LB133235
	Nickel	521	530	98	90 - 110	P	10/30/2024	16:59	LB133235
	Potassium	9810	9900	99	90 - 110	P	10/30/2024	16:59	LB133235
	Selenium	1000	1000	100	90 - 110	P	10/30/2024	16:59	LB133235
	Silver	260	250	104	90 - 110	P	10/30/2024	16:59	LB133235
	Sodium	9440	10000	94	90 - 110	P	10/30/2024	16:59	LB133235
	Thallium	1010	1000	101	90 - 110	P	10/30/2024	16:59	LB133235
	Vanadium	467	500	93	90 - 110	P	10/30/2024	16:59	LB133235
	Zinc	1030	1000	103	90 - 110	P	10/30/2024	16:59	LB133235

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc. **SDG No.:** P4595

Contract: TULL02 **Lab Code:** CHEM **Case No.:** P4595 **SAS No.:** P4595

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	99.5	100	100	80 - 120	P	10/30/2024	17:03	LB133235
	Antimony	47.8	50.0	96	80 - 120	P	10/30/2024	17:03	LB133235
	Arsenic	19.3	20.0	96	80 - 120	P	10/30/2024	17:03	LB133235
	Barium	89.9	100	90	80 - 120	P	10/30/2024	17:03	LB133235
	Beryllium	5.36	6.0	89	80 - 120	P	10/30/2024	17:03	LB133235
	Cadmium	5.74	6.0	96	80 - 120	P	10/30/2024	17:03	LB133235
	Calcium	1830	2000	91	80 - 120	P	10/30/2024	17:03	LB133235
	Chromium	9.65	10.0	96	80 - 120	P	10/30/2024	17:03	LB133235
	Cobalt	28.1	30.0	94	80 - 120	P	10/30/2024	17:03	LB133235
	Copper	21.4	20.0	107	80 - 120	P	10/30/2024	17:03	LB133235
	Iron	95.8	100	96	80 - 120	P	10/30/2024	17:03	LB133235
	Lead	11.8	12.0	99	80 - 120	P	10/30/2024	17:03	LB133235
	Magnesium	1780	2000	89	80 - 120	P	10/30/2024	17:03	LB133235
	Manganese	18.3	20.0	92	80 - 120	P	10/30/2024	17:03	LB133235
	Nickel	37.8	40.0	94	80 - 120	P	10/30/2024	17:03	LB133235
	Potassium	1770	2000	88	80 - 120	P	10/30/2024	17:03	LB133235
	Selenium	18.3	20.0	91	80 - 120	P	10/30/2024	17:03	LB133235
	Silver	10.2	10.0	102	80 - 120	P	10/30/2024	17:03	LB133235
	Sodium	1670	2000	84	80 - 120	P	10/30/2024	17:03	LB133235
	Thallium	38.1	40.0	95	80 - 120	P	10/30/2024	17:03	LB133235
	Vanadium	33.4	40.0	84	80 - 120	P	10/30/2024	17:03	LB133235
	Zinc	41.3	40.0	103	80 - 120	P	10/30/2024	17:03	LB133235

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc. **SDG No.:** P4595

Contract: TULL02 **Lab Code:** CHEM **Case No.:** P4595 **SAS No.:** P4595

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9480	10000	95	90 - 110	P	10/30/2024	17:33	LB133235
	Antimony	4880	5000	98	90 - 110	P	10/30/2024	17:33	LB133235
	Arsenic	4830	5000	97	90 - 110	P	10/30/2024	17:33	LB133235
	Barium	9200	10000	92	90 - 110	P	10/30/2024	17:33	LB133235
	Beryllium	231	250	92	90 - 110	P	10/30/2024	17:33	LB133235
	Cadmium	2370	2500	95	90 - 110	P	10/30/2024	17:33	LB133235
	Calcium	23200	25000	93	90 - 110	P	10/30/2024	17:33	LB133235
	Chromium	920	1000	92	90 - 110	P	10/30/2024	17:33	LB133235
	Cobalt	2380	2500	95	90 - 110	P	10/30/2024	17:33	LB133235
	Copper	1220	1250	98	90 - 110	P	10/30/2024	17:33	LB133235
	Iron	4680	5000	94	90 - 110	P	10/30/2024	17:33	LB133235
	Lead	4770	5000	96	90 - 110	P	10/30/2024	17:33	LB133235
	Magnesium	26000	25000	104	90 - 110	P	10/30/2024	17:33	LB133235
	Manganese	2260	2500	91	90 - 110	P	10/30/2024	17:33	LB133235
	Nickel	2400	2500	96	90 - 110	P	10/30/2024	17:33	LB133235
	Potassium	23800	25000	95	90 - 110	P	10/30/2024	17:33	LB133235
	Selenium	4910	5000	98	90 - 110	P	10/30/2024	17:33	LB133235
	Silver	1200	1250	96	90 - 110	P	10/30/2024	17:33	LB133235
	Sodium	23200	25000	93	90 - 110	P	10/30/2024	17:33	LB133235
	Thallium	4840	5000	97	90 - 110	P	10/30/2024	17:33	LB133235
	Vanadium	2700	2500	108	90 - 110	P	10/30/2024	17:33	LB133235
	Zinc	2440	2500	97	90 - 110	P	10/30/2024	17:33	LB133235
CCV02	Aluminum	9440	10000	94	90 - 110	P	10/30/2024	18:29	LB133235
	Antimony	4890	5000	98	90 - 110	P	10/30/2024	18:29	LB133235
	Arsenic	4850	5000	97	90 - 110	P	10/30/2024	18:29	LB133235
	Barium	9140	10000	91	90 - 110	P	10/30/2024	18:29	LB133235
	Beryllium	226	250	90	90 - 110	P	10/30/2024	18:29	LB133235
	Cadmium	2370	2500	95	90 - 110	P	10/30/2024	18:29	LB133235
	Calcium	22800	25000	91	90 - 110	P	10/30/2024	18:29	LB133235
	Chromium	912	1000	91	90 - 110	P	10/30/2024	18:29	LB133235
	Cobalt	2380	2500	95	90 - 110	P	10/30/2024	18:29	LB133235
	Copper	1230	1250	98	90 - 110	P	10/30/2024	18:29	LB133235
	Iron	4670	5000	93	90 - 110	P	10/30/2024	18:29	LB133235
	Lead	4760	5000	95	90 - 110	P	10/30/2024	18:29	LB133235

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc. **SDG No.:** P4595

Contract: TULL02 **Lab Code:** CHEM **Case No.:** P4595 **SAS No.:** P4595

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	25300	25000	101	90 - 110	P	10/30/2024	18:29	LB133235
	Manganese	2730	2500	109	90 - 110	P	10/30/2024	18:29	LB133235
	Nickel	2400	2500	96	90 - 110	P	10/30/2024	18:29	LB133235
	Potassium	24400	25000	97	90 - 110	P	10/30/2024	18:29	LB133235
	Selenium	4940	5000	99	90 - 110	P	10/30/2024	18:29	LB133235
	Silver	1200	1250	96	90 - 110	P	10/30/2024	18:29	LB133235
	Sodium	23400	25000	94	90 - 110	P	10/30/2024	18:29	LB133235
	Thallium	4870	5000	97	90 - 110	P	10/30/2024	18:29	LB133235
	Vanadium	2740	2500	110	90 - 110	P	10/30/2024	18:29	LB133235
	Zinc	2440	2500	98	90 - 110	P	10/30/2024	18:29	LB133235
CCV03	Aluminum	9550	10000	96	90 - 110	P	10/30/2024	19:23	LB133235
	Antimony	4930	5000	99	90 - 110	P	10/30/2024	19:23	LB133235
	Arsenic	4900	5000	98	90 - 110	P	10/30/2024	19:23	LB133235
	Barium	9020	10000	90	90 - 110	P	10/30/2024	19:23	LB133235
	Beryllium	230	250	92	90 - 110	P	10/30/2024	19:23	LB133235
	Cadmium	2390	2500	96	90 - 110	P	10/30/2024	19:23	LB133235
	Calcium	23000	25000	92	90 - 110	P	10/30/2024	19:23	LB133235
	Chromium	922	1000	92	90 - 110	P	10/30/2024	19:23	LB133235
	Cobalt	2410	2500	96	90 - 110	P	10/30/2024	19:23	LB133235
	Copper	1240	1250	99	90 - 110	P	10/30/2024	19:23	LB133235
	Iron	4720	5000	94	90 - 110	P	10/30/2024	19:23	LB133235
	Lead	4820	5000	96	90 - 110	P	10/30/2024	19:23	LB133235
	Magnesium	25400	25000	102	90 - 110	P	10/30/2024	19:23	LB133235
	Manganese	2480	2500	99	90 - 110	P	10/30/2024	19:23	LB133235
	Nickel	2420	2500	97	90 - 110	P	10/30/2024	19:23	LB133235
	Potassium	24600	25000	98	90 - 110	P	10/30/2024	19:23	LB133235
	Selenium	5000	5000	100	90 - 110	P	10/30/2024	19:23	LB133235
	Silver	1220	1250	98	90 - 110	P	10/30/2024	19:23	LB133235
	Sodium	23600	25000	95	90 - 110	P	10/30/2024	19:23	LB133235
	Thallium	4950	5000	99	90 - 110	P	10/30/2024	19:23	LB133235
	Vanadium	2640	2500	106	90 - 110	P	10/30/2024	19:23	LB133235
	Zinc	2480	2500	99	90 - 110	P	10/30/2024	19:23	LB133235
CCV04	Aluminum	9590	10000	96	90 - 110	P	10/30/2024	20:18	LB133235
	Antimony	4960	5000	99	90 - 110	P	10/30/2024	20:18	LB133235

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc. **SDG No.:** P4595

Contract: TULL02 **Lab Code:** CHEM **Case No.:** P4595 **SAS No.:** P4595

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	4920	5000	98	90 - 110	P	10/30/2024	20:18	LB133235
	Barium	9130	10000	91	90 - 110	P	10/30/2024	20:18	LB133235
	Beryllium	229	250	91	90 - 110	P	10/30/2024	20:18	LB133235
	Cadmium	2390	2500	96	90 - 110	P	10/30/2024	20:18	LB133235
	Calcium	23100	25000	92	90 - 110	P	10/30/2024	20:18	LB133235
	Chromium	924	1000	92	90 - 110	P	10/30/2024	20:18	LB133235
	Cobalt	2400	2500	96	90 - 110	P	10/30/2024	20:18	LB133235
	Copper	1240	1250	99	90 - 110	P	10/30/2024	20:18	LB133235
	Iron	4700	5000	94	90 - 110	P	10/30/2024	20:18	LB133235
	Lead	4800	5000	96	90 - 110	P	10/30/2024	20:18	LB133235
	Magnesium	25600	25000	102	90 - 110	P	10/30/2024	20:18	LB133235
	Manganese	2500	2500	100	90 - 110	P	10/30/2024	20:18	LB133235
	Nickel	2420	2500	97	90 - 110	P	10/30/2024	20:18	LB133235
	Potassium	24400	25000	98	90 - 110	P	10/30/2024	20:18	LB133235
	Selenium	5010	5000	100	90 - 110	P	10/30/2024	20:18	LB133235
	Silver	1210	1250	97	90 - 110	P	10/30/2024	20:18	LB133235
	Sodium	23300	25000	93	90 - 110	P	10/30/2024	20:18	LB133235
	Thallium	5020	5000	100	90 - 110	P	10/30/2024	20:18	LB133235
	Vanadium	2660	2500	106	90 - 110	P	10/30/2024	20:18	LB133235
	Zinc	2460	2500	98	90 - 110	P	10/30/2024	20:18	LB133235
CCV05	Aluminum	9120	10000	91	90 - 110	P	10/30/2024	21:10	LB133235
	Antimony	4730	5000	94	90 - 110	P	10/30/2024	21:10	LB133235
	Arsenic	4670	5000	94	90 - 110	P	10/30/2024	21:10	LB133235
	Barium	10300	10000	103	90 - 110	P	10/30/2024	21:10	LB133235
	Beryllium	243	250	97	90 - 110	P	10/30/2024	21:10	LB133235
	Cadmium	2260	2500	90	90 - 110	P	10/30/2024	21:10	LB133235
	Calcium	26500	25000	106	90 - 110	P	10/30/2024	21:10	LB133235
	Chromium	1040	1000	104	90 - 110	P	10/30/2024	21:10	LB133235
	Cobalt	2270	2500	91	90 - 110	P	10/30/2024	21:10	LB133235
	Copper	1180	1250	94	90 - 110	P	10/30/2024	21:10	LB133235
	Iron	4500	5000	90	90 - 110	P	10/30/2024	21:10	LB133235
	Lead	4530	5000	91	90 - 110	P	10/30/2024	21:10	LB133235
	Magnesium	23600	25000	95	90 - 110	P	10/30/2024	21:10	LB133235
	Manganese	2560	2500	102	90 - 110	P	10/30/2024	21:10	LB133235

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc. **SDG No.:** P4595

Contract: TULL02 **Lab Code:** CHEM **Case No.:** P4595 **SAS No.:** P4595

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2290	2500	91	90 - 110	P	10/30/2024	21:10	LB133235
	Potassium	23800	25000	95	90 - 110	P	10/30/2024	21:10	LB133235
	Selenium	4780	5000	96	90 - 110	P	10/30/2024	21:10	LB133235
	Silver	1160	1250	93	90 - 110	P	10/30/2024	21:10	LB133235
	Sodium	23100	25000	92	90 - 110	P	10/30/2024	21:10	LB133235
	Thallium	4780	5000	96	90 - 110	P	10/30/2024	21:10	LB133235
	Vanadium	2590	2500	104	90 - 110	P	10/30/2024	21:10	LB133235
	Zinc	2360	2500	94	90 - 110	P	10/30/2024	21:10	LB133235
CCV06	Aluminum	9480	10000	95	90 - 110	P	10/30/2024	21:59	LB133235
	Antimony	4940	5000	99	90 - 110	P	10/30/2024	21:59	LB133235
	Arsenic	4880	5000	98	90 - 110	P	10/30/2024	21:59	LB133235
	Barium	9250	10000	92	90 - 110	P	10/30/2024	21:59	LB133235
	Beryllium	254	250	102	90 - 110	P	10/30/2024	21:59	LB133235
	Cadmium	2350	2500	94	90 - 110	P	10/30/2024	21:59	LB133235
	Calcium	22700	25000	91	90 - 110	P	10/30/2024	21:59	LB133235
	Chromium	921	1000	92	90 - 110	P	10/30/2024	21:59	LB133235
	Cobalt	2370	2500	95	90 - 110	P	10/30/2024	21:59	LB133235
	Copper	1230	1250	98	90 - 110	P	10/30/2024	21:59	LB133235
	Iron	4700	5000	94	90 - 110	P	10/30/2024	21:59	LB133235
	Lead	4730	5000	95	90 - 110	P	10/30/2024	21:59	LB133235
	Magnesium	24700	25000	99	90 - 110	P	10/30/2024	21:59	LB133235
	Manganese	2660	2500	106	90 - 110	P	10/30/2024	21:59	LB133235
	Nickel	2380	2500	95	90 - 110	P	10/30/2024	21:59	LB133235
	Potassium	25400	25000	101	90 - 110	P	10/30/2024	21:59	LB133235
	Selenium	4990	5000	100	90 - 110	P	10/30/2024	21:59	LB133235
	Silver	1220	1250	97	90 - 110	P	10/30/2024	21:59	LB133235
	Sodium	24600	25000	98	90 - 110	P	10/30/2024	21:59	LB133235
	Thallium	4960	5000	99	90 - 110	P	10/30/2024	21:59	LB133235
	Vanadium	2690	2500	108	90 - 110	P	10/30/2024	21:59	LB133235
	Zinc	2470	2500	99	90 - 110	P	10/30/2024	21:59	LB133235
CCV07	Aluminum	9250	10000	92	90 - 110	P	10/30/2024	22:53	LB133235
	Antimony	4900	5000	98	90 - 110	P	10/30/2024	22:53	LB133235
	Arsenic	4830	5000	97	90 - 110	P	10/30/2024	22:53	LB133235
	Barium	9230	10000	92	90 - 110	P	10/30/2024	22:53	LB133235

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc. **SDG No.:** P4595

Contract: TULL02 **Lab Code:** CHEM **Case No.:** P4595 **SAS No.:** P4595

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	246	250	98	90 - 110	P	10/30/2024	22:53	LB133235
	Cadmium	2300	2500	92	90 - 110	P	10/30/2024	22:53	LB133235
	Calcium	26500	25000	106	90 - 110	P	10/30/2024	22:53	LB133235
	Chromium	1050	1000	105	90 - 110	P	10/30/2024	22:53	LB133235
	Cobalt	2320	2500	93	90 - 110	P	10/30/2024	22:53	LB133235
	Copper	1220	1250	97	90 - 110	P	10/30/2024	22:53	LB133235
	Iron	4510	5000	90	90 - 110	P	10/30/2024	22:53	LB133235
	Lead	4630	5000	93	90 - 110	P	10/30/2024	22:53	LB133235
	Magnesium	23600	25000	94	90 - 110	P	10/30/2024	22:53	LB133235
	Manganese	2580	2500	103	90 - 110	P	10/30/2024	22:53	LB133235
	Nickel	2340	2500	94	90 - 110	P	10/30/2024	22:53	LB133235
	Potassium	25300	25000	101	90 - 110	P	10/30/2024	22:53	LB133235
	Selenium	4960	5000	99	90 - 110	P	10/30/2024	22:53	LB133235
	Silver	1180	1250	94	90 - 110	P	10/30/2024	22:53	LB133235
	Sodium	24500	25000	98	90 - 110	P	10/30/2024	22:53	LB133235
	Thallium	4780	5000	96	90 - 110	P	10/30/2024	22:53	LB133235
	Vanadium	2620	2500	105	90 - 110	P	10/30/2024	22:53	LB133235
	Zinc	2350	2500	94	90 - 110	P	10/30/2024	22:53	LB133235
CCV08	Aluminum	9490	10000	95	90 - 110	P	10/30/2024	23:44	LB133235
	Antimony	4940	5000	99	90 - 110	P	10/30/2024	23:44	LB133235
	Arsenic	4850	5000	97	90 - 110	P	10/30/2024	23:44	LB133235
	Barium	9370	10000	94	90 - 110	P	10/30/2024	23:44	LB133235
	Beryllium	227	250	91	90 - 110	P	10/30/2024	23:44	LB133235
	Cadmium	2330	2500	93	90 - 110	P	10/30/2024	23:44	LB133235
	Calcium	22500	25000	90	90 - 110	P	10/30/2024	23:44	LB133235
	Chromium	916	1000	92	90 - 110	P	10/30/2024	23:44	LB133235
	Cobalt	2350	2500	94	90 - 110	P	10/30/2024	23:44	LB133235
	Copper	1230	1250	98	90 - 110	P	10/30/2024	23:44	LB133235
	Iron	4670	5000	93	90 - 110	P	10/30/2024	23:44	LB133235
	Lead	4680	5000	94	90 - 110	P	10/30/2024	23:44	LB133235
	Magnesium	24100	25000	96	90 - 110	P	10/30/2024	23:44	LB133235
	Manganese	2630	2500	105	90 - 110	P	10/30/2024	23:44	LB133235
	Nickel	2360	2500	95	90 - 110	P	10/30/2024	23:44	LB133235
	Potassium	26200	25000	105	90 - 110	P	10/30/2024	23:44	LB133235

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc. **SDG No.:** P4595
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P4595 **SAS No.:** P4595
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	4990	5000	100	90 - 110	P	10/30/2024	23:44	LB133235
	Silver	1220	1250	98	90 - 110	P	10/30/2024	23:44	LB133235
	Sodium	25100	25000	100	90 - 110	P	10/30/2024	23:44	LB133235
	Thallium	4890	5000	98	90 - 110	P	10/30/2024	23:44	LB133235
	Vanadium	2670	2500	107	90 - 110	P	10/30/2024	23:44	LB133235
	Zinc	2470	2500	99	90 - 110	P	10/30/2024	23:44	LB133235
CCV09	Aluminum	9200	10000	92	90 - 110	P	10/31/2024	00:13	LB133235
	Antimony	4850	5000	97	90 - 110	P	10/31/2024	00:13	LB133235
	Arsenic	4770	5000	95	90 - 110	P	10/31/2024	00:13	LB133235
	Barium	9140	10000	91	90 - 110	P	10/31/2024	00:13	LB133235
	Beryllium	251	250	100	90 - 110	P	10/31/2024	00:13	LB133235
	Cadmium	2270	2500	91	90 - 110	P	10/31/2024	00:13	LB133235
	Calcium	26500	25000	106	90 - 110	P	10/31/2024	00:13	LB133235
	Chromium	1060	1000	106	90 - 110	P	10/31/2024	00:13	LB133235
	Cobalt	2290	2500	92	90 - 110	P	10/31/2024	00:13	LB133235
	Copper	1200	1250	96	90 - 110	P	10/31/2024	00:13	LB133235
	Iron	4520	5000	90	90 - 110	P	10/31/2024	00:13	LB133235
	Lead	4580	5000	92	90 - 110	P	10/31/2024	00:13	LB133235
	Magnesium	23800	25000	95	90 - 110	P	10/31/2024	00:13	LB133235
	Manganese	2610	2500	104	90 - 110	P	10/31/2024	00:13	LB133235
	Nickel	2310	2500	92	90 - 110	P	10/31/2024	00:13	LB133235
	Potassium	25400	25000	102	90 - 110	P	10/31/2024	00:13	LB133235
	Selenium	4890	5000	98	90 - 110	P	10/31/2024	00:13	LB133235
	Silver	1180	1250	95	90 - 110	P	10/31/2024	00:13	LB133235
	Sodium	23800	25000	95	90 - 110	P	10/31/2024	00:13	LB133235
	Thallium	4770	5000	95	90 - 110	P	10/31/2024	00:13	LB133235
	Vanadium	2630	2500	105	90 - 110	P	10/31/2024	00:13	LB133235
	Zinc	2360	2500	94	90 - 110	P	10/31/2024	00:13	LB133235