



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Construction Co., Inc. **SDG No.:** P4595
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P4595 **SAS No.:** P4595
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.20	0.2	98	40 - 160	CV	10/29/2024	10:41	LB133185
CRI01	Aluminum	93.0	100	93	40 - 160	P	10/30/2024	17:12	LB133235
	Antimony	47.7	50.0	95	40 - 160	P	10/30/2024	17:12	LB133235
	Arsenic	18.4	20.0	92	40 - 160	P	10/30/2024	17:12	LB133235
	Barium	88.3	100	88	40 - 160	P	10/30/2024	17:12	LB133235
	Beryllium	5.33	6.0	89	40 - 160	P	10/30/2024	17:12	LB133235
	Cadmium	5.71	6.0	95	40 - 160	P	10/30/2024	17:12	LB133235
	Calcium	1830	2000	91	40 - 160	P	10/30/2024	17:12	LB133235
	Chromium	9.37	10.0	94	40 - 160	P	10/30/2024	17:12	LB133235
	Cobalt	28.3	30.0	94	40 - 160	P	10/30/2024	17:12	LB133235
	Copper	21.4	20.0	107	40 - 160	P	10/30/2024	17:12	LB133235
	Iron	96.5	100	96	40 - 160	P	10/30/2024	17:12	LB133235
	Lead	12.2	12.0	102	40 - 160	P	10/30/2024	17:12	LB133235
	Magnesium	1730	2000	86	40 - 160	P	10/30/2024	17:12	LB133235
	Manganese	18.0	20.0	90	40 - 160	P	10/30/2024	17:12	LB133235
	Nickel	37.9	40.0	95	40 - 160	P	10/30/2024	17:12	LB133235
	Potassium	1770	2000	89	40 - 160	P	10/30/2024	17:12	LB133235
	Selenium	19.6	20.0	98	40 - 160	P	10/30/2024	17:12	LB133235
	Silver	10.2	10.0	102	40 - 160	P	10/30/2024	17:12	LB133235
	Sodium	1680	2000	84	40 - 160	P	10/30/2024	17:12	LB133235
	Thallium	38.6	40.0	96	40 - 160	P	10/30/2024	17:12	LB133235
	Vanadium	34.6	40.0	86	40 - 160	P	10/30/2024	17:12	LB133235
	Zinc	41.6	40.0	104	40 - 160	P	10/30/2024	17:12	LB133235