

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Tully Construction Co., Inc. **SDG No.:** P4595
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P4595 **SAS No.:** P4595
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	277000	255000	109	216000	294000	10/30/2024	17:16	LB133235
	Antimony	-2.51			-50	50	10/30/2024	17:16	LB133235
	Arsenic	4.39			-20	20	10/30/2024	17:16	LB133235
	Barium	2.58	6.0	43	-94	106	10/30/2024	17:16	LB133235
	Beryllium	1.35			-6	6	10/30/2024	17:16	LB133235
	Cadmium	2.53	1.0	253	-5	7	10/30/2024	17:16	LB133235
	Calcium	252000	245000	103	208000	282000	10/30/2024	17:16	LB133235
	Chromium	60.7	52.0	117	42	62	10/30/2024	17:16	LB133235
	Cobalt	2.03			-30	30	10/30/2024	17:16	LB133235
	Copper	-2.35	2.0	118	-18	22	10/30/2024	17:16	LB133235
	Iron	115000	101000	114	85600	116500	10/30/2024	17:16	LB133235
	Lead	11.8			-12	12	10/30/2024	17:16	LB133235
	Magnesium	272000	255000	107	216000	294000	10/30/2024	17:16	LB133235
	Manganese	4.48	7.0	64	-13	27	10/30/2024	17:16	LB133235
	Nickel	2.81	2.0	140	-38	42	10/30/2024	17:16	LB133235
	Potassium	-162			0	0	10/30/2024	17:16	LB133235
	Selenium	-16.0			-20	20	10/30/2024	17:16	LB133235
	Silver	1.85			-10	10	10/30/2024	17:16	LB133235
	Sodium	-118			0	0	10/30/2024	17:16	LB133235
	Thallium	3.43			-40	40	10/30/2024	17:16	LB133235
	Vanadium	7.02			-40	40	10/30/2024	17:16	LB133235
	Zinc	5.63			-40	40	10/30/2024	17:16	LB133235
ICSAB01	Aluminum	235000	247000	95	209000	285000	10/30/2024	17:27	LB133235
	Antimony	635	618	103	525	711	10/30/2024	17:27	LB133235
	Arsenic	105	104	101	88.4	120	10/30/2024	17:27	LB133235
	Barium	467	537	87	437	637	10/30/2024	17:27	LB133235
	Beryllium	472	495	95	420	570	10/30/2024	17:27	LB133235
	Cadmium	956	972	98	826	1120	10/30/2024	17:27	LB133235
	Calcium	215000	235000	92	199000	271000	10/30/2024	17:27	LB133235
	Chromium	547	542	101	460	624	10/30/2024	17:27	LB133235
	Cobalt	516	476	108	404	548	10/30/2024	17:27	LB133235
	Copper	501	511	98	434	588	10/30/2024	17:27	LB133235
	Iron	95300	99300	96	84400	114500	10/30/2024	17:27	LB133235
	Lead	57.3	49.0	117	37	61	10/30/2024	17:27	LB133235
	Magnesium	227000	248000	92	210000	286000	10/30/2024	17:27	LB133235
	Manganese	451	507	89	430	584	10/30/2024	17:27	LB133235
	Nickel	1020	954	107	810	1100	10/30/2024	17:27	LB133235
	Potassium	-183			0	0	10/30/2024	17:27	LB133235
	Selenium	36.7	46.0	80	26	66	10/30/2024	17:27	LB133235
	Silver	207	201	103	170	232	10/30/2024	17:27	LB133235
	Sodium	-147			0	0	10/30/2024	17:27	LB133235
	Thallium	93.9	108	87	68	148	10/30/2024	17:27	LB133235

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tully Construction Co., Inc.

SDG No.: P4595

Contract: TULL02

Lab Code: CHEM

Case No.: P4595

SAS No.: P4595

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	442	491	90	417	565	10/30/2024	17:27	LB133235
	Zinc	898	952	94	809	1095	10/30/2024	17:27	LB133235