

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Construction Co., Inc.

SDG No.: P4595

Contract: TULL02

Lab Code: CHEM

Case No.: P4595

SAS No.: P4595

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164487BS Mercury	mg/Kg	0.27	0.24		87	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Construction Co., Inc. **SDG No.:** P4595
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P4595 **SAS No.:** P4595

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164499BS							
Aluminum	mg/Kg	99.0	87.1		88	80 - 120	P
Antimony	mg/Kg	39.6	36.0		91	80 - 120	P
Arsenic	mg/Kg	39.6	33.7		85	80 - 120	P
Barium	mg/Kg	9.9	8.23		83	80 - 120	P
Beryllium	mg/Kg	9.9	9.51		96	80 - 120	P
Cadmium	mg/Kg	9.9	8.18		83	80 - 120	P
Calcium	mg/Kg	49.5	44.1	J	89	80 - 120	P
Chromium	mg/Kg	19.8	17.6		89	80 - 120	P
Cobalt	mg/Kg	9.9	8.86		90	80 - 120	P
Copper	mg/Kg	14.9	14.2		95	80 - 120	P
Iron	mg/Kg	150	130		87	80 - 120	P
Lead	mg/Kg	49.5	41.9		85	80 - 120	P
Magnesium	mg/Kg	99.0	91.4	J	92	80 - 120	P
Manganese	mg/Kg	9.9	10.1		102	80 - 120	P
Nickel	mg/Kg	24.8	22.4		90	80 - 120	P
Potassium	mg/Kg	500	449		90	80 - 120	P
Selenium	mg/Kg	99.0	86.6		88	80 - 120	P
Silver	mg/Kg	3.7	3.46		94	80 - 120	P
Sodium	mg/Kg	150	121		81	80 - 120	P
Thallium	mg/Kg	99.0	87.2		88	80 - 120	P
Vanadium	mg/Kg	14.9	14.9		100	80 - 120	P
Zinc	mg/Kg	9.9	9.56		97	80 - 120	P