

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB133216

Review By	jaswal	Review On	10/31/2024 10:27:53 PM
Supervise By	mohan	Supervise On	10/31/2024 10:28:42 PM

STD. NAME	STD REF.#
ICAL Standard	MP82962,MP82963,MP82964,MP82965,MP82966,MP82967
ICV Standard	MP82968
CCV Standard	MP82970
ICSA Standard	
CRI Standard	MP82972
LCS Standard	
Chk Standard	MP82969,MP82971,MP82973,MP82975

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/30/24 15:33		Mohan	OK
2	S0.2	S0.2	CAL2	10/30/24 15:35		Mohan	OK
3	S2.5	S2.5	CAL3	10/30/24 15:37		Mohan	OK
4	S5	S5	CAL4	10/30/24 15:40		Mohan	OK
5	S7.5	S7.5	CAL5	10/30/24 15:42		Mohan	OK
6	S10	S10	CAL6	10/30/24 15:44		Mohan	OK
7	ICV59	ICV59	ICV	10/30/24 15:47		Mohan	OK
8	ICB59	ICB59	ICB	10/30/24 15:50		Mohan	OK
9	CCV94	CCV94	CCV	10/30/24 15:52		Mohan	OK
10	CCB94	CCB94	CCB	10/30/24 15:54		Mohan	OK
11	CRA	CRA	CRDL	10/30/24 15:56		Mohan	OK
12	HighStd	HighStd	HIGH STD	10/30/24 15:59		Mohan	OK
13	ChkStd	ChkStd	SAM	10/30/24 16:01		Mohan	OK
14	PB164520BL	PB164520BL	MB	10/30/24 16:06		Mohan	OK
15	PB164520BS	PB164520BS	LCS	10/30/24 16:08		Mohan	OK
16	P4594-04	TP-4	SAM	10/30/24 16:11		Mohan	OK
17	P4594-08	BP-F17	SAM	10/30/24 16:13		Mohan	OK
18	P4594-12	BP-F16	SAM	10/30/24 16:15		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB133216

Review By	jaswal	Review On	10/31/2024 10:27:53 PM
Supervise By	mohan	Supervise On	10/31/2024 10:28:42 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82962,MP82963,MP82964,MP82965,MP82966,MP82967		
ICV Standard	MP82968		
CCV Standard	MP82970		
ICSA Standard			
CRI Standard	MP82972		
LCS Standard			
Chk Standard	MP82969,MP82971,MP82973,MP82975		

19	P4594-16	TP-5	SAM	10/30/24 16:17		Mohan	OK
20	P4594-20	BP-F15	SAM	10/30/24 16:20		Mohan	OK
21	CCV95	CCV95	CCV	10/30/24 16:22		Mohan	OK
22	CCB95	CCB95	CCB	10/30/24 16:24		Mohan	OK
23	P4595-02	PITKIN-COMP	SAM	10/30/24 16:27		Mohan	OK
24	P4597-03	RED-1-1	SAM	10/30/24 16:29		Mohan	OK
25	P4597-06	RED-1-2	SAM	10/30/24 16:31		Mohan	OK
26	P4597-09	BLUE-2-1	SAM	10/30/24 16:33		Mohan	OK
27	P4597-12	BLUE-2-2	SAM	10/30/24 16:36		Mohan	OK
28	P4598-04	BP-F12	SAM	10/30/24 16:38		Mohan	OK
29	P4598-08	BP-F11	SAM	10/30/24 16:40		Mohan	OK
30	P4598-12	TP-8	SAM	10/30/24 16:42		Mohan	OK
31	P4598-12DUP	TP-8DUP	DUP	10/30/24 16:54		Mohan	OK
32	P4598-12MS	TP-8MS	MS	10/30/24 16:56		Mohan	OK
33	CCV96	CCV96	CCV	10/30/24 16:59		Mohan	OK
34	CCB96	CCB96	CCB	10/30/24 17:01		Mohan	OK
35	P4598-12MSD	TP-8MSD	MSD	10/30/24 17:03		Mohan	OK
36	PB164478TB	PB164478TB	MB	10/30/24 17:05		Mohan	OK
37	P4598-12L	TP-8L	SD	10/30/24 17:08		Mohan	OK
38	P4598-12A	TP-8A	PS	10/30/24 17:10		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB133216

Review By	jaswal	Review On	10/31/2024 10:27:53 PM
Supervise By	mohan	Supervise On	10/31/2024 10:28:42 PM

STD. NAME	STD REF.#
ICAL Standard	MP82962,MP82963,MP82964,MP82965,MP82966,MP82967
ICV Standard	MP82968
CCV Standard	MP82970
ICSA Standard	
CRI Standard	MP82972
LCS Standard	
Chk Standard	MP82969,MP82971,MP82973,MP82975

39	CCV97	CCV97	CCV	10/30/24 17:12		Mohan	OK
40	CCB97	CCB97	CCB	10/30/24 17:15		Mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133235

Review By	kareem	Review On	11/6/2024 11:05:31 PM
Supervise By	mohan	Supervise On	11/7/2024 10:04:01 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086,MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091,MP83092		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/30/24 16:34		Kareem	OK
2	S1	S1	CAL2	10/30/24 16:38		Kareem	OK
3	S2	S2	CAL3	10/30/24 16:42		Kareem	OK
4	S3	S3	CAL4	10/30/24 16:46		Kareem	OK
5	S4	S4	CAL5	10/30/24 16:51		Kareem	OK
6	S5	S5	CAL6	10/30/24 16:55		Kareem	OK
7	ICV01	ICV01	ICV	10/30/24 16:59		Kareem	OK
8	LLICV01	LLICV01	LLICV	10/30/24 17:03		Kareem	OK
9	ICB01	ICB01	ICB	10/30/24 17:07		Kareem	OK
10	CRI01	CRI01	CRDL	10/30/24 17:12		Kareem	OK
11	ICSA01	ICSA01	ICSA	10/30/24 17:16		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	10/30/24 17:27		Kareem	OK
13	CCV01	CCV01	CCV	10/30/24 17:33		Kareem	OK
14	CCB01	CCB01	CCB	10/30/24 17:37		Kareem	OK
15	P4596-07	TOTE-WATER-COMP	SAM	10/30/24 17:46		Kareem	OK
16	P4596-07DUP	TOTE-WATER-COMP	DUP	10/30/24 17:50		Kareem	OK
17	P4596-07L	TOTE-WATER-COMP	SD	10/30/24 17:54		Kareem	OK
18	P4596-07MS	TOTE-WATER-COMP	MS	10/30/24 17:59	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133235

Review By	kareem	Review On	11/6/2024 11:05:31 PM
Supervise By	mohan	Supervise On	11/7/2024 10:04:01 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

19	P4596-07MSD	TOTE-WATER-COMP	MSD	10/30/24 18:03	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
20	P4596-07A	TOTE-WATER-COMP	PS	10/30/24 18:07	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
21	P4594-04	TP-4	SAM	10/30/24 18:11		Kareem	OK
22	P4594-08	BP-F17	SAM	10/30/24 18:15		Kareem	OK
23	P4594-12	BP-F16	SAM	10/30/24 18:20		Kareem	OK
24	P4594-16	TP-5	SAM	10/30/24 18:24		Kareem	OK
25	CCV02	CCV02	CCV	10/30/24 18:29		Kareem	OK
26	CCB02	CCB02	CCB	10/30/24 18:33		Kareem	OK
27	P4594-20	BP-F15	SAM	10/30/24 18:39		Kareem	OK
28	P4595-02	PITKIN-COMP	SAM	10/30/24 18:43		Kareem	OK
29	P4597-03	RED-1-1	SAM	10/30/24 18:47		Kareem	OK
30	P4597-06	RED-1-2	SAM	10/30/24 18:52		Kareem	OK
31	P4597-09	BLUE-2-1	SAM	10/30/24 18:56		Kareem	OK
32	P4597-12	BLUE-2-2	SAM	10/30/24 19:01		Kareem	OK
33	P4598-04	BP-F12	SAM	10/30/24 19:05		Kareem	OK
34	P4598-08	BP-F11	SAM	10/30/24 19:10		Kareem	OK
35	P4598-12	TP-8	SAM	10/30/24 19:14		Kareem	OK
36	P4598-12DUP	TP-8DUP	DUP	10/30/24 19:19		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133235

Review By	kareem	Review On	11/6/2024 11:05:31 PM
Supervise By	mohan	Supervise On	11/7/2024 10:04:01 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

37	CCV03	CCV03	CCV	10/30/24 19:23		Kareem	OK
38	CCB03	CCB03	CCB	10/30/24 19:27		Kareem	OK
39	P4598-12L	TP-8L	SD	10/30/24 19:32		Kareem	OK
40	P4598-12MS	TP-8MS	MS	10/30/24 19:36	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
41	P4598-12MSD	TP-8MSD	MSD	10/30/24 19:40	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
42	P4598-12A	TP-8A	PS	10/30/24 19:45	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
43	PB164478TB	PB164478TB	MB	10/30/24 19:49		Kareem	OK
44	PB164508BS	PB164508BS	LCS	10/30/24 19:57	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
45	PB164508BL	PB164508BL	MB	10/30/24 20:01		Kareem	OK
46	P4594-01	TP-4	SAM	10/30/24 20:06		Kareem	OK
47	P4594-05	BP-F17	SAM	10/30/24 20:10		Kareem	OK
48	P4594-09	BP-F16	SAM	10/30/24 20:14		Kareem	OK
49	CCV04	CCV04	CCV	10/30/24 20:18		Kareem	OK
50	CCB04	CCB04	CCB	10/30/24 20:23		Kareem	OK
51	P4594-13	TP-5	SAM	10/30/24 20:28		Kareem	OK
52	P4594-17	BP-F15	SAM	10/30/24 20:32		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133235

Review By	kareem	Review On	11/6/2024 11:05:31 PM
Supervise By	mohan	Supervise On	11/7/2024 10:04:01 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086,MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091,MP83092		

53	P4595-01	PITKIN-COMP	SAM	10/30/24 20:36		Kareem	OK
54	P4597-01	RED-1-1	SAM	10/30/24 20:40		Kareem	OK
55	P4597-04	RED-1-2	SAM	10/30/24 20:44		Kareem	OK
56	P4597-07	BLUE-2-1	SAM	10/30/24 20:49		Kareem	OK
57	P4597-10	BLUE-2-2	SAM	10/30/24 20:53		Kareem	OK
58	P4598-01	BP-F12	SAM	10/30/24 20:57		Kareem	OK
59	P4598-05	BP-F11	SAM	10/30/24 21:01		Kareem	OK
60	P4598-09	TP-8	SAM	10/30/24 21:06		Kareem	OK
61	CCV05	CCV05	CCV	10/30/24 21:10		Kareem	OK
62	CCB05	CCB05	CCB	10/30/24 21:14		Kareem	OK
63	P4598-09DUP	TP-8DUP	DUP	10/30/24 21:18		Kareem	OK
64	P4598-09L	TP-8L	SD	10/30/24 21:22		Kareem	OK
65	P4598-09MS	TP-8MS	MS	10/30/24 21:26	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
66	P4598-09MSD	TP-8MSD	MSD	10/30/24 21:30	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
67	P4598-09A	TP-8A	PS	10/30/24 21:34	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
68	P4598-09RE	TP-8RE	SAM	10/30/24 21:38	Not use	Kareem	Not Ok
69	PB164499BL	PB164499BL	MB	10/30/24 21:42		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133235

Review By	kareem	Review On	11/6/2024 11:05:31 PM
Supervise By	mohan	Supervise On	11/7/2024 10:04:01 PM

STD. NAME	STD REF.#
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084
ICV Standard	MP83085
CCV Standard	MP83088
ICSA Standard	MP83086,MP83087
CRI Standard	MP83084
LCS Standard	
Chk Standard	MP83091,MP83092

70	PB164499BS	PB164499BS	LCS	10/30/24 21:47	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
71	P4611-01	TP-1	SAM	10/30/24 21:51		Kareem	OK
72	P4611-04	TP-2	SAM	10/30/24 21:55		Kareem	OK
73	CCV06	CCV06	CCV	10/30/24 21:59		Kareem	OK
74	CCB06	CCB06	CCB	10/30/24 22:03		Kareem	OK
75	P4611-07	TP-3	SAM	10/30/24 22:11		Kareem	OK
76	P4611-10	TP-4	SAM	10/30/24 22:15		Kareem	OK
77	P4611-13	TP-5	SAM	10/30/24 22:19		Kareem	OK
78	P4611-16	TP-6	SAM	10/30/24 22:24		Kareem	OK
79	P4612-03	MOO-24-00335	SAM	10/30/24 22:28		Kareem	OK
80	P4612-05	CONCRETE-PAD	SAM	10/30/24 22:32		Kareem	OK
81	P4613-01	ARS20-0001	SAM	10/30/24 22:36		Kareem	OK
82	P4614-01	NB-07-102924	SAM	10/30/24 22:41		Kareem	OK
83	P4616-01	BP-F10	SAM	10/30/24 22:45		Kareem	OK
84	P4616-05	BP-F9-MOVED	SAM	10/30/24 22:49		Kareem	OK
85	CCV07	CCV07	CCV	10/30/24 22:53		Kareem	OK
86	CCB07	CCB07	CCB	10/30/24 22:58		Kareem	OK
87	P4617-01	RO-SOIL	SAM	10/30/24 23:03		Kareem	OK
88	P4617-01DUP	RO-SOILDUP	DUP	10/30/24 23:07		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133235

Review By	kareem	Review On	11/6/2024 11:05:31 PM
Supervise By	mohan	Supervise On	11/7/2024 10:04:01 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83078,MP83079,MP73080,MP83081,MP83082,MP83084		
ICV Standard	MP83085		
CCV Standard	MP83088		
ICSA Standard	MP83086,MP83087		
CRI Standard	MP83084		
LCS Standard			
Chk Standard	MP83091,MP83092		

89	P4617-01L	RO-SOILL	SD	10/30/24 23:11		Kareem	OK
90	P4617-01MS	RO-SOILMS	MS	10/30/24 23:15	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
91	P4617-01MSD	RO-SOILMSD	MSD	10/30/24 23:19	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
92	P4617-01A	RO-SOILA	PS	10/30/24 23:23	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
93	P4618-01	HR-01-102924	SAM	10/30/24 23:27		Kareem	OK
94	P4618-03	HR-04-102924	SAM	10/30/24 23:32		Kareem	OK
95	P4619-01	SU-03-102924	SAM	10/30/24 23:36		Kareem	OK
96	PB164523BL	PB164523BL	MB	10/30/24 23:40		Kareem	OK
97	CCV08	CCV08	CCV	10/30/24 23:44		Kareem	OK
98	CCB08	CCB08	CCB	10/30/24 23:48		Kareem	OK
99	LR1	LR1	HIGH STD	10/30/24 23:57		Kareem	OK
100	LR2	LR2	HIGH STD	10/31/24 00:02		Kareem	OK
101	PB164523BS	PB164523BS	LCS	10/31/24 00:09	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
102	CCV09	CCV09	CCV	10/31/24 00:13		Kareem	OK
103	CCB09	CCB09	CCB	10/31/24 00:17		Kareem	OK