



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Construction Co., Inc.

SDG No.: P4595

Contract: TULL02

Lab Code: CHEM

Case No.: P4595

SAS No.: P4595

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.15	0.2	73	40 - 160	CV	10/30/2024	15:56	LB133216
CRI01	Arsenic	18.4	20.0	92	40 - 160	P	10/30/2024	17:12	LB133235
	Barium	88.3	100	88	40 - 160	P	10/30/2024	17:12	LB133235
	Cadmium	5.71	6.0	95	40 - 160	P	10/30/2024	17:12	LB133235
	Chromium	9.37	10.0	94	40 - 160	P	10/30/2024	17:12	LB133235
	Lead	12.2	12.0	102	40 - 160	P	10/30/2024	17:12	LB133235
	Selenium	19.6	20.0	98	40 - 160	P	10/30/2024	17:12	LB133235
	Silver	10.2	10.0	102	40 - 160	P	10/30/2024	17:12	LB133235