



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Tully Construction Co., Inc.				SDG No.:	P4595			
Contract:	TULL02	Lab Code:	CHEM		Case No.:	P4595	SAS No.:	P4595	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB59	Mercury	0.20	+/-0.20	U	0.20	CV	10/30/2024	15:50	LB133216

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Construction Co., Inc.

SDG No.: P4595

Contract: TULL02

Lab Code: CHEM

Case No.: P4595

SAS No.: P4595

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB94	Mercury	0.20	+/-0.20	U	0.20	CV	10/30/2024	15:54	LB133216
CCB95	Mercury	0.20	+/-0.20	U	0.20	CV	10/30/2024	16:24	LB133216
CCB96	Mercury	0.20	+/-0.20	U	0.20	CV	10/30/2024	17:01	LB133216
CCB97	Mercury	0.20	+/-0.20	U	0.20	CV	10/30/2024	17:15	LB133216

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Construction Co., Inc.

SDG No.: P4595

Contract: TULL02

Lab Code: CHEM

Case No.: P4595

SAS No.: P4595

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	10/30/2024	17:07	LB133235
	Barium	100	+/-100	U	100	P	10/30/2024	17:07	LB133235
	Cadmium	6.00	+/-6.00	U	6.00	P	10/30/2024	17:07	LB133235
	Chromium	10.0	+/-10.0	U	10.0	P	10/30/2024	17:07	LB133235
	Lead	12.0	+/-12.0	U	12.0	P	10/30/2024	17:07	LB133235
	Selenium	20.0	+/-20.0	U	20.0	P	10/30/2024	17:07	LB133235
	Silver	10.0	+/-10.0	U	10.0	P	10/30/2024	17:07	LB133235

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Construction Co., Inc.		SDG No.: P4595							
Contract: TULL02	Lab Code: CHEM	Case No.: P4595	SAS No.: P4595						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	10/30/2024	17:37	LB133235
	Barium	100	+/-100	U	100	P	10/30/2024	17:37	LB133235
	Cadmium	6.00	+/-6.00	U	6.00	P	10/30/2024	17:37	LB133235
	Chromium	10.0	+/-10.0	U	10.0	P	10/30/2024	17:37	LB133235
	Lead	12.0	+/-12.0	U	12.0	P	10/30/2024	17:37	LB133235
	Selenium	20.0	+/-20.0	U	20.0	P	10/30/2024	17:37	LB133235
	Silver	10.0	+/-10.0	U	10.0	P	10/30/2024	17:37	LB133235
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	10/30/2024	18:33	LB133235
	Barium	100	+/-100	U	100	P	10/30/2024	18:33	LB133235
	Cadmium	6.00	+/-6.00	U	6.00	P	10/30/2024	18:33	LB133235
	Chromium	10.0	+/-10.0	U	10.0	P	10/30/2024	18:33	LB133235
	Lead	12.0	+/-12.0	U	12.0	P	10/30/2024	18:33	LB133235
	Selenium	20.0	+/-20.0	U	20.0	P	10/30/2024	18:33	LB133235
	Silver	10.0	+/-10.0	U	10.0	P	10/30/2024	18:33	LB133235
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	10/30/2024	19:27	LB133235
	Barium	100	+/-100	U	100	P	10/30/2024	19:27	LB133235
	Cadmium	6.00	+/-6.00	U	6.00	P	10/30/2024	19:27	LB133235
	Chromium	10.0	+/-10.0	U	10.0	P	10/30/2024	19:27	LB133235
	Lead	12.0	+/-12.0	U	12.0	P	10/30/2024	19:27	LB133235
	Selenium	20.0	+/-20.0	U	20.0	P	10/30/2024	19:27	LB133235
	Silver	10.0	+/-10.0	U	10.0	P	10/30/2024	19:27	LB133235
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	10/30/2024	20:23	LB133235
	Barium	100	+/-100	U	100	P	10/30/2024	20:23	LB133235
	Cadmium	6.00	+/-6.00	U	6.00	P	10/30/2024	20:23	LB133235
	Chromium	10.0	+/-10.0	U	10.0	P	10/30/2024	20:23	LB133235
	Lead	12.0	+/-12.0	U	12.0	P	10/30/2024	20:23	LB133235
	Selenium	20.0	+/-20.0	U	20.0	P	10/30/2024	20:23	LB133235
	Silver	10.0	+/-10.0	U	10.0	P	10/30/2024	20:23	LB133235
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	10/30/2024	21:14	LB133235
	Barium	100	+/-100	U	100	P	10/30/2024	21:14	LB133235
	Cadmium	6.00	+/-6.00	U	6.00	P	10/30/2024	21:14	LB133235
	Chromium	10.0	+/-10.0	U	10.0	P	10/30/2024	21:14	LB133235
	Lead	12.0	+/-12.0	U	12.0	P	10/30/2024	21:14	LB133235
	Selenium	20.0	+/-20.0	U	20.0	P	10/30/2024	21:14	LB133235
	Silver	10.0	+/-10.0	U	10.0	P	10/30/2024	21:14	LB133235
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	10/30/2024	22:03	LB133235
	Barium	100	+/-100	U	100	P	10/30/2024	22:03	LB133235
	Cadmium	6.00	+/-6.00	U	6.00	P	10/30/2024	22:03	LB133235
	Chromium	10.0	+/-10.0	U	10.0	P	10/30/2024	22:03	LB133235

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tully Construction Co., Inc.</u>		SDG No.: <u>P4595</u>							
Contract: <u>TULL02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4595</u>	SAS No.: <u>P4595</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	12.0	+/-12.0	U	12.0	P	10/30/2024	22:03	LB133235
	Selenium	20.0	+/-20.0	U	20.0	P	10/30/2024	22:03	LB133235
	Silver	10.0	+/-10.0	U	10.0	P	10/30/2024	22:03	LB133235
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	10/30/2024	22:58	LB133235
	Barium	100	+/-100	U	100	P	10/30/2024	22:58	LB133235
	Cadmium	6.00	+/-6.00	U	6.00	P	10/30/2024	22:58	LB133235
	Chromium	10.0	+/-10.0	U	10.0	P	10/30/2024	22:58	LB133235
	Lead	12.0	+/-12.0	U	12.0	P	10/30/2024	22:58	LB133235
	Selenium	20.0	+/-20.0	U	20.0	P	10/30/2024	22:58	LB133235
	Silver	10.0	+/-10.0	U	10.0	P	10/30/2024	22:58	LB133235
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	10/30/2024	23:48	LB133235
	Barium	100	+/-100	U	100	P	10/30/2024	23:48	LB133235
	Cadmium	6.00	+/-6.00	U	6.00	P	10/30/2024	23:48	LB133235
	Chromium	10.0	+/-10.0	U	10.0	P	10/30/2024	23:48	LB133235
	Lead	12.0	+/-12.0	U	12.0	P	10/30/2024	23:48	LB133235
	Selenium	20.0	+/-20.0	U	20.0	P	10/30/2024	23:48	LB133235
	Silver	10.0	+/-10.0	U	10.0	P	10/30/2024	23:48	LB133235
CCB09	Arsenic	20.0	+/-20.0	U	20.0	P	10/31/2024	00:17	LB133235
	Barium	100	+/-100	U	100	P	10/31/2024	00:17	LB133235
	Cadmium	6.00	+/-6.00	U	6.00	P	10/31/2024	00:17	LB133235
	Chromium	10.0	+/-10.0	U	10.0	P	10/31/2024	00:17	LB133235
	Lead	12.0	+/-12.0	U	12.0	P	10/31/2024	00:17	LB133235
	Selenium	20.0	+/-20.0	U	20.0	P	10/31/2024	00:17	LB133235
	Silver	10.0	+/-10.0	U	10.0	P	10/31/2024	00:17	LB133235