

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Tully Construction Co., Inc. **SDG No.:** P4595
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P4595 **SAS No.:** P4595
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	4.39			-20	20	10/30/2024	17:16	LB133235
	Barium	2.58	6.0	43	-94	106	10/30/2024	17:16	LB133235
	Cadmium	2.53	1.0	253	-5	7	10/30/2024	17:16	LB133235
	Chromium	60.7	52.0	117	42	62	10/30/2024	17:16	LB133235
	Lead	11.8			-12	12	10/30/2024	17:16	LB133235
	Selenium	-16.0			-20	20	10/30/2024	17:16	LB133235
	Silver	1.85			-10	10	10/30/2024	17:16	LB133235
ICSAB01	Arsenic	105	104	101	88.4	120	10/30/2024	17:27	LB133235
	Barium	467	537	87	437	637	10/30/2024	17:27	LB133235
	Cadmium	956	972	98	826	1120	10/30/2024	17:27	LB133235
	Chromium	547	542	101	460	624	10/30/2024	17:27	LB133235
	Lead	57.3	49.0	117	37	61	10/30/2024	17:27	LB133235
	Selenium	36.7	46.0	80	26	66	10/30/2024	17:27	LB133235
	Silver	207	201	103	170	232	10/30/2024	17:27	LB133235