

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tully Construction Co., Inc. **Level:** LOW **SDG No.:** P4595
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P4595 **SAS No.:** P4595
Matrix: Water **Sample ID:** P4598-12 **Client ID:** TP-8DUP
Percent Solids for Sample: NA **Duplicate ID** P4598-12DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	1730		1720		1		P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	50.0	U	50.0	U			P
Lead	ug/L	20	60.0	U	60.0	U			P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tully Construction Co., Inc. **Level:** LOW **SDG No.:** P4595
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P4595 **SAS No.:** P4595
Matrix: Water **Sample ID:** P4598-12MS **Client ID:** TP-8MSD
Percent Solids for Sample: NA **Duplicate ID** P4598-12MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	3880		3850		1		P
Barium	ug/L	20	2860		2940		3		P
Cadmium	ug/L	20	919		909		1		P
Chromium	ug/L	20	1900		1890		1		P
Lead	ug/L	20	4470		4420		1		P
Mercury	ug/L	20	36.0		38.7		7		CV
Selenium	ug/L	20	9690		9660		0		P
Silver	ug/L	20	396		397		0		P