

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4595

Client: Tully Construction Co., Inc.

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	P4594-05MS	Client Sample ID:	BP-F17MS					DataFile:	BF140117.D		
Naphthalene	1100	0	990	ug/Kg	90				72	110	
Acenaphthylene	1100	0	1100	ug/Kg	100				79	118	
Acenaphthene	1100	0	1200	ug/Kg	109				70	121	
Fluorene	1100	0	1000	ug/Kg	91				68	116	
Phenanthrene	1100	0	1000	ug/Kg	91				52	128	
Anthracene	1100	0	1100	ug/Kg	100				62	124	
Fluoranthene	1100	0	960	ug/Kg	87				44	125	
Pyrene	1100	0	1000	ug/Kg	91				26	142	
Benzo(a)anthracene	1100	0	1100	ug/Kg	100				71	114	
Chrysene	1100	0	1100	ug/Kg	100				57	121	
Benzo(b)fluoranthene	1100	0	980	ug/Kg	89				67	121	
Benzo(k)fluoranthene	1100	0	1100	ug/Kg	100				57	134	
Benzo(a)pyrene	1100	0	1100	ug/Kg	100				70	142	
Indeno(1,2,3-cd)pyrene	1100	0	1100	ug/Kg	100				40	129	
Dibenz(a,h)anthracene	1100	0	1100	ug/Kg	100				43	123	
Benzo(g,h,i)perylene	1100	0	1000	ug/Kg	91				24	125	