

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4595 SAS No.: P4595 SDG NO.: P4595

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/31/2024

Lab File ID: BE101425.D Time Analyzed: 10:18

Instrument ID: BNA_E GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	136647	7.573	695845	10.34	498092	14.18
UPPER LIMIT	273294	8.073	1391690	10.84	996184	14.683
LOWER LIMIT	68323.5	7.073	347923	9.84	249046	13.683
EPA SAMPLE NO.						
01 PITKIN-COMP	148112	7.58	698047	10.35	479391	14.19

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.