

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4595 SAS No.: P4595 SDG NO.: P4595

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/31/2024

Lab File ID: BE101425.D Time Analyzed: 10:18

Instrument ID: BNA_E GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1107610	16.915	1206960	21.081	1484760	23.366
UPPER LIMIT	2215220	17.415	2413920	21.581	2969520	23.866
LOWER LIMIT	553805	16.415	603480	20.581	742380	22.866
EPA SAMPLE NO.						
01 PITKIN-COMP	1031180	16.93	1039490	21.09	1358300	23.38

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.