

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TULL02

Lab Code: CHEM Case No.: P4595 SAS No.: P4595 SDG No.: P4595

Instrument ID: MSVOA_Y Calibration Date(s): 10/09/2024 10/09/2024

Heated Purge: (Y/N) Y Calibration Time(s): 10:18 12:11

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:	RRF005 = VY019827.D			RRF010 = VY019828.D		RRF020 = VY019829.D		
	RRF050 = VY019830.D			RRF100 = VY019831.D		RRF150 = VY019832.D		
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Dichlorodifluoromethane	0.500	0.498	0.410	0.458	0.434	0.492	0.465	8.1
Chloromethane	0.643	0.637	0.517	0.616	0.550	0.671	0.606	9.8
Vinyl Chloride	0.707	0.685	0.579	0.688	0.624	0.716	0.667	8
Bromomethane	0.458	0.447	0.357	0.433	0.390	0.447	0.422	9.4
Chloroethane	0.493	0.469	0.390	0.456	0.408	0.472	0.448	8.9
Trichlorofluoromethane	1.101	1.016	0.868	1.000	0.915	1.051	0.992	8.7
1,1,2-Trichlorotrifluoroethane	0.619	0.611	0.509	0.581	0.535	0.606	0.577	7.8
1,1-Dichloroethene	0.588	0.558	0.453	0.550	0.499	0.570	0.536	9.4
Acetone	0.196	0.156	0.134	0.170	0.152	0.153	0.160	12.9
Carbon Disulfide	1.501	1.392	1.193	1.552	1.404	1.586	1.438	10
Methyl tert-butyl Ether	1.719	1.559	1.334	1.574	1.443	1.617	1.541	8.8
Methyl Acetate	0.393	0.334	0.301	0.350	0.321	0.372	0.345	9.7
Methylene Chloride	0.869	0.695	0.551	0.619	0.543	0.603	0.647	18.9
trans-1,2-Dichloroethene	0.630	0.597	0.511	0.605	0.547	0.612	0.584	7.7
1,1-Dichloroethane	1.288	1.219	1.025	1.202	1.084	1.214	1.172	8.3
Cyclohexane	1.262	1.098	0.863	1.009	0.915	1.024	1.029	13.8
2-Butanone	0.251	0.217	0.186	0.222	0.201	0.211	0.215	10.2
Carbon Tetrachloride	0.525	0.505	0.446	0.527	0.495	0.559	0.510	7.5
cis-1,2-Dichloroethene	0.777	0.766	0.628	0.740	0.669	0.749	0.721	8.2
Bromochloromethane	0.607	0.470	0.502	0.529	0.483	0.504	0.516	9.5
Chloroform	1.318	1.238	1.055	1.225	1.102	1.229	1.195	8.1
1,1,1-Trichloroethane	1.118	1.091	0.915	1.067	0.978	1.112	1.047	7.9
Methylcyclohexane	0.629	0.584	0.512	0.629	0.581	0.658	0.599	8.6
Benzene	1.512	1.474	1.283	1.488	1.361	1.514	1.439	6.6
1,2-Dichloroethane	0.448	0.399	0.368	0.430	0.398	0.436	0.413	7.3
Trichloroethene	0.381	0.340	0.300	0.366	0.333	0.370	0.348	8.7
1,2-Dichloropropane	0.390	0.375	0.322	0.366	0.337	0.371	0.360	7.1
Bromodichloromethane	0.549	0.524	0.456	0.543	0.501	0.560	0.522	7.4
4-Methyl-2-Pentanone	0.281	0.247	0.220	0.269	0.254	0.275	0.258	8.7
Toluene	0.930	0.892	0.800	0.940	0.866	0.960	0.898	6.5

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TULL02
Lab Code: CHEM Case No.: P4595 SAS No.: P4595 SDG No.: P4595
Instrument ID: MSVOA_Y Calibration Date(s): 10/09/2024 10/09/2024
Heated Purge: (Y/N) Y Calibration Time(s): 10:18 12:11
GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VY019827.D		RRF010 = VY019828.D		RRF020 = VY019829.D			
		RRF050 = VY019830.D		RRF100 = VY019831.D		RRF150 = VY019832.D			
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
t-1,3-Dichloropropene	0.477	0.448	0.408	0.497	0.464	0.524	0.470	8.6	
cis-1,3-Dichloropropene	0.577	0.537	0.489	0.579	0.542	0.603	0.554	7.3	
1,1,2-Trichloroethane	0.280	0.257	0.230	0.269	0.248	0.271	0.259	6.9	
2-Hexanone	0.193	0.173	0.160	0.197	0.185	0.197	0.184	8.1	
Dibromochloromethane	0.348	0.317	0.294	0.344	0.328	0.364	0.333	7.5	
1,2-Dibromoethane	0.244	0.234	0.207	0.246	0.226	0.248	0.234	6.7	
Tetrachloroethene	0.379	0.366	0.316	0.368	0.328	0.377	0.356	7.6	
Chlorobenzene	1.226	1.167	1.032	1.171	1.064	1.202	1.144	6.8	
Ethyl Benzene	2.177	2.100	1.861	2.164	1.953	2.209	2.077	6.7	
m/p-Xylenes	0.802	0.781	0.691	0.796	0.720	0.813	0.767	6.5	
o-Xylene	0.750	0.756	0.665	0.769	0.700	0.783	0.737	6.1	
Styrene	1.276	1.238	1.120	1.304	1.191	1.333	1.244	6.3	
Bromoform	0.208	0.198	0.180	0.220	0.204	0.232	0.207	8.7	
Isopropylbenzene	4.420	4.371	3.794	4.268	3.912	4.470	4.206	6.7	
1,1,2,2-Tetrachloroethane	0.796	0.737	0.664	0.763	0.719	0.805	0.747	7	
1,3-Dichlorobenzene	1.968	1.892	1.598	1.819	1.651	1.887	1.802	8.1	
1,4-Dichlorobenzene	1.928	1.831	1.572	1.801	1.635	1.856	1.771	7.8	
1,2-Dichlorobenzene	1.716	1.620	1.416	1.612	1.474	1.663	1.584	7.3	
1,2-Dibromo-3-Chloropropane	0.133	0.115	0.100	0.121	0.118	0.129	0.119	9.7	
1,2,4-Trichlorobenzene	0.860	0.832	0.758	0.965	0.886	1.027	0.888	10.8	
1,2,3-Trichlorobenzene	0.716	0.698	0.634	0.823	0.761	0.872	0.751	11.5	
1,2-Dichloroethane-d4	0.701	0.656	0.566	0.581	0.583	0.607	0.616	8.5	
Dibromofluoromethane	0.356	0.335	0.300	0.321	0.326	0.341	0.330	5.8	
Toluene-d8	1.299	1.254	1.134	1.185	1.191	1.246	1.218	4.9	
4-Bromofluorobenzene	0.498	0.438	0.402	0.426	0.427	0.450	0.440	7.4	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.