

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4595

Client: Tully Construction Co., Inc.

Analytical Method: SW8260D

Datafile : VY020072.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1029SBSD01	Dichlorodifluoromethane	17.7	14.2	ug/Kg	80	6		64	136	20
	Chloromethane	17.7	15.9	ug/Kg	90	13		70	130	20
	Vinyl chloride	17.7	15.8	ug/Kg	89	0		72	129	20
	Bromomethane	17.7	18.4	ug/Kg	104	11		58	141	20
	Chloroethane	17.7	17.9	ug/Kg	101	2		69	130	20
	Trichlorofluoromethane	17.7	17.6	ug/Kg	99	2		69	134	20
	1,1,2-Trichlorotrifluoroethane	17.7	18.4	ug/Kg	104	2		81	123	20
	1,1-Dichloroethene	17.7	15.2	ug/Kg	86	1		79	121	20
	Acetone	88.7	110	ug/Kg	124	12		60	131	20
	Carbon disulfide	17.7	9.00	ug/Kg	51	0		45	154	20
	Methyl tert-butyl Ether	17.7	20.9	ug/Kg	118	11		77	129	20
	Methyl Acetate	17.7	24.0	ug/Kg	136	23	*	69	149	20
	Methylene Chloride	17.7	22.8	ug/Kg	129	17		56	174	20
	trans-1,2-Dichloroethene	17.7	15.9	ug/Kg	90	2		80	123	20
	1,1-Dichloroethane	17.7	20.5	ug/Kg	116	4		82	123	20
	Cyclohexane	17.7	15.3	ug/Kg	86	1		76	122	20
	2-Butanone	88.7	120	ug/Kg	135	4	*	69	131	20
	Carbon Tetrachloride	17.7	17.8	ug/Kg	101	1		76	129	20
	cis-1,2-Dichloroethene	17.7	19.2	ug/Kg	108	5		82	123	20
	Bromochloromethane	17.7	20.6	ug/Kg	116	1		80	127	20
	Chloroform	17.7	21.7	ug/Kg	123	5		82	125	20
	1,1,1-Trichloroethane	17.7	20.1	ug/Kg	114	4		80	126	20
	Methylcyclohexane	17.7	14.2	ug/Kg	80	1		77	123	20
	Benzene	17.7	18.2	ug/Kg	103	3		84	121	20
	1,2-Dichloroethane	17.7	19.5	ug/Kg	110	5		81	126	20
	Trichloroethene	17.7	17.8	ug/Kg	101	8		83	122	20
	1,2-Dichloropropane	17.7	20.4	ug/Kg	115	2		83	122	20
	Bromodichloromethane	17.7	21.7	ug/Kg	123	8		82	123	20
	4-Methyl-2-Pentanone	88.7	110	ug/Kg	124	3		70	135	20
	Toluene	17.7	18.4	ug/Kg	104	2		83	122	20
	t-1,3-Dichloropropene	17.7	18.6	ug/Kg	105	3		78	124	20
	cis-1,3-Dichloropropene	17.7	19.0	ug/Kg	107	4		81	122	20
	1,1,2-Trichloroethane	17.7	21.9	ug/Kg	124	8		82	125	20
	2-Hexanone	88.7	110	ug/Kg	124	5		66	138	20
	Dibromochloromethane	17.7	20.8	ug/Kg	118	8		79	125	20
	1,2-Dibromoethane	17.7	20.5	ug/Kg	116	11		80	125	20
	Tetrachloroethene	17.7	19.5	ug/Kg	110	21	*	83	125	20
	Chlorobenzene	17.7	18.8	ug/Kg	106	4		84	122	20
	Ethyl Benzene	17.7	18.3	ug/Kg	103	0		82	124	20
	m/p-Xylenes	35.5	36.2	ug/Kg	102	0		83	124	20
	o-Xylene	17.7	18.9	ug/Kg	107	6		83	123	20
	Styrene	17.7	19.2	ug/Kg	108	2		82	124	20
	Bromoform	17.7	20.7	ug/Kg	117	7		75	127	20
	Isopropylbenzene	17.7	18.3	ug/Kg	103	4		82	124	20
	1,1,2,2-Tetrachloroethane	17.7	22.1	ug/Kg	125	5		77	127	20
	1,3-Dichlorobenzene	17.7	18.8	ug/Kg	106	0		83	122	20
	1,4-Dichlorobenzene	17.7	19.5	ug/Kg	110	3		84	121	20
	1,2-Dichlorobenzene	17.7	19.4	ug/Kg	110	2		83	124	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4595

Client: Tully Construction Co., Inc.

Analytical Method: SW8260D

Datafile : VY020072.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VY1029SBSD01	1,2-Dibromo-3-Chloropropane	17.7	22.7	ug/Kg	128	13		66	134	20
	1,2,4-Trichlorobenzene	17.7	18.3	ug/Kg	103	7		78	127	20
	1,2,3-Trichlorobenzene	17.7	19.1	ug/Kg	108	13		70	137	20