

**DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY**

Lab Name: Chemtech Contract: ENTA05  
 ProjectID: North Point  
 Lab Code: CHEM Case No.: P4599 SAS No.: P4599 SDG No.: P4599

| Calibration Sequence : FF102124 |            |                  | Test : Diesel Range Organics |                  |
|---------------------------------|------------|------------------|------------------------------|------------------|
| Concentration (PPM)             | Area Count | Reference Factor | File ID                      |                  |
| 1000                            | 140350994  | 140351           | FF014706.D                   |                  |
| 500                             | 71252825   | 142506           | FF014707.D                   |                  |
| 200                             | 29731497   | 148657           | FF014708.D                   |                  |
| 100                             | 14736133   | 147361           | FF014709.D                   |                  |
| 50                              | 7756416    | 155128           | FF014710.D                   |                  |
| AVG RF : 146801                 |            | % RSD : 3.93     |                              | AVG RT : 14.9988 |