

DIESEL RANGE ORGANICS CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: ENTA05
ProjectID: North Point
Lab Code: CHEM Case No.: P4599 SAS No.: P4599 SDG No.: P4599
DataFile: FF014861.D Analyst Name: YP\AJ Analyst Date: 10-31-2024

Conc. (PPM)	Area Count	RF	Average RF	%D
500	78838190	157676	146801	7.408