

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** ENTACT
Lab Code: CHEM **Cas No:** P4599 **SAS No :** P4599 **SDG No:** P4599
Sample No : PB164525BS **Datafile:** FD048624.D
Client ID : PB164525BS

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aromatic C10-C12	6.7	3.53	53		(40-140)
Aromatic C12-C16	10.0	7.16	72		(40-140)
Aromatic C16-C21	16.7	14.1	85		(40-140)
Aromatic C21-C36	26.7	22.2	83		(40-140)