

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4600 SAS No.: P4600 SDG No.: P4600

Instrument ID: MSVOA_X Calibration Date/Time: 10/29/2024 09:29

Lab File ID: VX043583.D Init. Calib. Date(s): 10/28/2024 10/28/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:59 12:54

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.684	0.597	0.1	-12.72	20
Vinyl Chloride	0.634	0.614		-3.31	20
Bromomethane	0.259	0.236		-8.88	20
Chloroethane	0.223	0.209		-6.28	20
Trichlorofluoromethane	0.775	0.759		-2.07	20
1,1,2-Trichlorotrifluoroethane	0.524	0.542		3.43	20
1,1-Dichloroethene	0.536	0.536		0	20
Acetone	0.317	0.297		-6.31	20
Carbon Disulfide	1.120	1.146		2.32	20
Methyl tert-butyl Ether	2.037	1.814		-10.95	20
Methylene Chloride	0.682	0.601		-11.88	20
trans-1,2-Dichloroethene	0.566	0.555		-1.94	20
1,1-Dichloroethane	1.101	1.028	0.1	-6.63	20
2-Butanone	0.462	0.403		-12.77	20
Carbon Tetrachloride	0.444	0.466		4.95	20
cis-1,2-Dichloroethene	0.723	0.687		-4.98	20
Chloroform	1.154	1.065		-7.71	20
1,1,1-Trichloroethane	0.969	0.948		-2.17	20
Methylcyclohexane	0.499	0.574		15.03	20
Benzene	1.341	1.333		-0.6	20
1,2-Dichloroethane	0.482	0.452		-6.22	20
Trichloroethene	0.333	0.351		5.41	20
1,2-Dichloropropane	0.330	0.327		-0.91	20
Bromodichloromethane	0.446	0.461		3.36	20
4-Methyl-2-Pentanone	0.492	0.451		-8.33	20
Toluene	0.813	0.854		5.04	20
t-1,3-Dichloropropene	0.479	0.490		2.3	20
cis-1,3-Dichloropropene	0.510	0.537		5.29	20
1,1,2-Trichloroethane	0.339	0.329		-2.95	20
2-Hexanone	0.374	0.352		-5.88	20
Dibromochloromethane	0.335	0.358		6.87	20
Tetrachloroethene	0.314	0.333		6.05	20
Chlorobenzene	1.076	1.092	0.3	1.49	20
Ethyl Benzene	1.754	1.860		6.04	20
m/p-Xylenes	0.671	0.725		8.05	20
o-Xylene	0.674	0.708		5.05	20
Styrene	1.120	1.213		8.3	20
Bromoform	0.253	0.265	0.1	4.74	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4600 SAS No.: P4600 SDG No.: P4600

Instrument ID: MSVOA_X Calibration Date/Time: 10/29/2024 09:29

Lab File ID: VX043583.D Init. Calib. Date(s): 10/28/2024 10/28/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:59 12:54

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.542	3.725		5.17	20
1,1,2,2-Tetrachloroethane	1.274	1.176	0.3	-7.69	20
1,3-Dichlorobenzene	1.639	1.707		4.15	20
1,4-Dichlorobenzene	1.699	1.695		-0.23	20
1,2-Dichlorobenzene	1.676	1.698		1.31	20
1,2-Dichloroethane-d4	0.797	0.708		-11.17	20
Dibromofluoromethane	0.339	0.345		1.77	20
Toluene-d8	1.174	1.260		7.32	20
4-Bromofluorobenzene	0.435	0.471		8.28	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.